

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 04/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 04/16/2025 18:33:18

Diagnostics Check	OK	18:26
Air Blank	0.000	18:27
Control Test	0.082	18:27
Air Blank	0.000	18:28
Subject Sample # 1	0.000	18:29
Breath Volume	2.472	18:29
Air Blank	0.000	18:29
Air Blank	0.000	18:31
Subject Sample # 2	0.000	18:31
Breath Volume	2.308	18:31
Air Blank	0.000	18:32
Control Test	0.082	18:32
Air Blank	0.000	18:33
Diagnostics Check	OK	18:33

Data Download Date/Time: 05/12/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: RODGERS,WILLIAM
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 04/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GADDIS

DATE/TIME OF TEST: 04/18/2025 04:04:53

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.082	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.177	03:59
Breath Volume	1.476	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.157	04:03
Breath Volume	2.355	04:03
Air Blank	0.000	04:03
Control Test	0.083	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 05/12/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 04/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 04/30/2025 18:30:59

Diagnostics Check	OK	18:25
Air Blank	0.000	18:25
Control Test	0.082	18:25
Air Blank	0.000	18:26
Subject Sample # 1	0.000	18:26
Breath Volume	1.843	18:26
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.000	18:29
Breath Volume	2.277	18:29
Air Blank	0.000	18:30
Control Test	0.081	18:30
Air Blank	0.000	18:30
Diagnostics Check	OK	18:30

Data Download Date/Time: 05/12/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ,JORGE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MESA,ALEX
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: BOBBITT

DATE/TIME OF TEST: 04/27/2025 07:14:56

Diagnostics Check	OK	07:10	
Air Blank	0.000	07:10	
Control Test	0.078	07:10	
Air Blank	0.000	07:11	
Subject Sample	# 1	07:13	Subject Test Refused
Breath Volume	0.000	07:13	
Air Blank	0.000	07:14	
Control Test	0.078	07:14	
Air Blank	0.000	07:14	
Diagnostics Check	OK	07:14	

Data Download Date/Time: 05/22/2025 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAZARYAN,LILIT,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: BARRETT,DEVON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: GRUBB

DATE/TIME OF TEST: 04/28/2025 00:02:11

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.076	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:00	Subject Test Refused
Breath Volume	0.000	00:00	
Air Blank	0.000	00:01	
Control Test	0.076	00:01	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 05/22/2025 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: WESSELING,SCOTT
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/30/2025 03:27:21

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:25	
Control Test	0.078	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Control Test	0.078	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 05/22/2025 11:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CABRERA,KACEY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: DE LA MATA

DATE/TIME OF TEST: 04/08/2025 20:41:08

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:38	
Control Test	0.081	20:39	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:40	
Control Test	0.081	20:40	
Air Blank	0.000	20:41	
Diagnostics Check	OK	20:41	

Data Download Date/Time: 04/18/2025 01:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ISIAIAH,
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: RODRIGUEZ,ISIAIAH
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:19

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 04/09/2025 11:55:40

Diagnostics Check	OK	11:49
Air Blank	0.000	11:49
Control Test	0.079	11:49
Air Blank	0.000	11:50
Subject Sample # 1	0.106	11:51
Breath Volume	1.972	11:51
Air Blank	0.000	11:51
Air Blank	0.000	11:53
Subject Sample # 2	0.102	11:54
Breath Volume	1.585	11:54
Air Blank	0.000	11:54
Control Test	0.080	11:55
Air Blank	0.000	11:55
Diagnostics Check	OK	11:55

Data Download Date/Time: 04/18/2025 01:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,KENY,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GONZALEZ,KENY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: HIGGS

DATE/TIME OF TEST: 04/12/2025 00:40:01

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.250	00:35
Breath Volume	3.250	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.258	00:38
Breath Volume	1.257	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:40

Data Download Date/Time: 04/18/2025 01:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,RASHAUN,K
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: HALL,RASHAUN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/12/2025 06:15:08

Diagnostics Check	OK	06:11	
Air Blank	0.000	06:11	
Control Test	0.079	06:11	
Air Blank	0.000	06:12	
Subject Sample	# 1	06:13	Subject Test Refused
Breath Volume	0.000	06:13	
Air Blank	0.000	06:14	
Control Test	0.079	06:14	
Air Blank	0.000	06:15	
Diagnostics Check	OK	06:15	

Data Download Date/Time: 04/18/2025 01:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ',CELINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/21/2025 09:31:37

Diagnostics Check	OK	09:28	
Air Blank	0.000	09:29	
Control Test	0.081	09:29	
Air Blank	0.000	09:30	
Subject Sample	# 1	09:30	Subject Test Refused
Breath Volume	0.000	09:30	
Air Blank	0.000	09:30	
Control Test	0.080	09:31	
Air Blank	0.000	09:31	
Diagnostics Check	OK	09:31	

Data Download Date/Time: 05/01/2025 03:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKEWELL,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: BAKEWELL,ZACHARY
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 04/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: SUME

DATE/TIME OF TEST: 04/27/2025 23:38:47

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.176	23:34
Breath Volume	1.312	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.180	23:37
Breath Volume	2.511	23:37
Air Blank	0.000	23:37
Control Test	0.081	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 05/01/2025 03:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BRICENO

DATE/TIME OF TEST: 04/04/2025 03:12:35

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:11	
Control Test	0.082	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank		03:12	RFI Detect
Air Blank		03:12	RFI Detect

Data Download Date/Time: 05/01/2025 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BRICENO

DATE/TIME OF TEST: 04/04/2025 03:17:18

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.082	03:15	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:16	
Control Test	0.081	03:16	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 05/01/2025 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 04/05/2025 03:27:17

Diagnostics Check	OK	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:25	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.082	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 05/01/2025 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALTUNAY,MUHAMMED
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: CHAVEZ LOPEZ

DATE/TIME OF TEST: 04/11/2025 23:48:56

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.150	23:44
Breath Volume	1.964	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:46
Subject Sample # 2	0.148	23:47
Breath Volume	2.214	23:47
Air Blank	0.000	23:48
Control Test	0.076	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 05/01/2025 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 04/12/2025 03:44:38

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.146	03:40
Breath Volume	2.351	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.151	03:43
Breath Volume	2.316	03:43
Air Blank	0.000	03:43
Control Test	0.078	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 05/01/2025 03:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BUFALINO

DATE/TIME OF TEST: 04/20/2025 02:16:25

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:11	
Control Test	0.082	02:11	
Air Blank	0.000	02:12	
Subject Sample # 1	0.135	02:12	
Breath Volume	1.636	02:12	
Air Blank	0.000	02:13	
Air Blank		02:15	RFI Detect
Air Blank	0.000	02:15	
Control Test	0.082	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 05/01/2025 03:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BUFALINO

DATE/TIME OF TEST: 04/20/2025 02:24:21

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.136	02:19
Breath Volume	1.375	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.125	02:22
Breath Volume	2.453	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 05/01/2025 03:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: PERALTA RODRIGUEZ

DATE/TIME OF TEST: 04/20/2025 08:11:51

Diagnostics Check	OK	08:05
Air Blank	0.000	08:05
Control Test	0.083	08:06
Air Blank	0.000	08:06
Subject Sample # 1	0.268	08:07
Breath Volume	2.449	08:07
Air Blank	0.000	08:08
Air Blank	0.000	08:09
Subject Sample # 2	0.265	08:10
Breath Volume	1.601	08:10
Air Blank	0.000	08:10
Control Test	0.080	08:11
Air Blank	0.000	08:11
Diagnostics Check	OK	08:11

Data Download Date/Time: 05/01/2025 03:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,CRISTIAN,X
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CERNADAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 09:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:13

SUBJECT'S LAST NAME: PAONE

DATE/TIME OF TEST: 04/05/2025 10:43:44

Diagnostics Check	OK	10:33
Air Blank	0.000	10:34
Control Test	0.080	10:34
Air Blank	0.000	10:35
Subject Sample # 1	0.225	10:35
Breath Volume	1.394	10:35
Air Blank	0.000	10:36
Air Blank	0.000	10:38
Subject Sample # 2	0.191	10:38
Breath Volume	1.417	10:38
Air Blank	0.000	10:39
Air Blank	0.000	10:41
Subject Sample # 3	0.214	10:42
Breath Volume	2.316	10:42
Air Blank	0.000	10:42
Control Test	0.079	10:43
Air Blank	0.000	10:43
Diagnostics Check	OK	10:43

Data Download Date/Time: 05/01/2025 05:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SULLIVAN,SHEYANNE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: MONTALVAN ARTEAGA

DATE/TIME OF TEST: 04/19/2025 16:41:46

Diagnostics Check	OK	16:33
Air Blank	0.000	16:33
Control Test	0.077	16:34
Air Blank	0.000	16:34
Subject Sample # 1	0.254	16:35
Breath Volume	2.941	16:35
Air Blank	0.000	16:36
Air Blank	0.000	16:37
Subject Sample # 2	0.247	16:39
Breath Volume	2.148	16:39
Air Blank	0.000	16:40
Control Test	0.081	16:41
Air Blank	0.000	16:41
Diagnostics Check	OK	16:41

Data Download Date/Time: 05/01/2025 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON III,EMANUEL,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WALTON III,EMANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/23/2025 23:47:07

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.082	23:40
Air Blank	0.000	23:40
Subject Sample # 1	0.273	23:42
Breath Volume	2.496	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.273	23:45
Breath Volume	2.433	23:45
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 05/01/2025 05:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORRIS,J
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:19

SUBJECT'S LAST NAME: COREA RIVAS

DATE/TIME OF TEST: 04/02/2025 09:46:44

Diagnostics Check	OK	09:39
Air Blank	0.000	09:39
Control Test	0.082	09:40
Air Blank	0.000	09:40
Subject Sample # 1	0.161	09:41
Breath Volume	2.308	09:41
Air Blank	0.000	09:42
Air Blank	0.000	09:44
Subject Sample # 2	0.157	09:45
Breath Volume	3.371	09:45
Air Blank	0.000	09:45
Control Test	0.083	09:46
Air Blank	0.000	09:46
Diagnostics Check	OK	09:46

Data Download Date/Time: 04/03/2025 03:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORTEGA,WILLFREDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/20/2025 08:04:09

Diagnostics Check	OK	08:01	
Air Blank	0.000	08:01	
Control Test	0.083	08:02	
Air Blank	0.000	08:02	
Subject Sample	# 1	08:02	Subject Test Refused
Breath Volume	0.000	08:02	
Air Blank	0.000	08:03	
Control Test	0.083	08:03	
Air Blank	0.000	08:04	
Diagnostics Check	OK	08:04	

Data Download Date/Time: 05/01/2025 05:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORTEGA,WILFREDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:09

SUBJECT'S LAST NAME: LOPEZ PEREZ

DATE/TIME OF TEST: 04/20/2025 10:36:35

Diagnostics Check	OK	10:33	
Air Blank	0.000	10:34	
Control Test	0.082	10:34	
Air Blank	0.000	10:35	
Subject Sample	# 1	10:35	Subject Test Refused
Breath Volume	0.000	10:35	
Air Blank	0.000	10:35	
Control Test	0.082	10:35	
Air Blank	0.000	10:36	
Diagnostics Check	OK	10:36	

Data Download Date/Time: 05/01/2025 05:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MENDEZ RODRIGUEZ

DATE/TIME OF TEST: 04/24/2025 23:12:50

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.081	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.181	23:08
Breath Volume	2.687	23:08
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.184	23:11
Breath Volume	2.992	23:11
Air Blank	0.000	23:11
Control Test	0.082	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 05/01/2025 05:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON III,EMANUEL,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WALTON III,EMANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: FIGUEROA VARGAS

DATE/TIME OF TEST: 04/27/2025 19:43:43

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.081	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.204	19:39
Breath Volume	4.476	19:39
Air Blank	0.000	19:39
Air Blank	0.000	19:41
Subject Sample # 2	0.206	19:42
Breath Volume	4.175	19:42
Air Blank	0.000	19:42
Control Test	0.077	19:43
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 05/01/2025 05:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: CARBONELL

DATE/TIME OF TEST: 04/29/2025 03:14:23

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.082	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.082	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 05/01/2025 05:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: TOFT

DATE/TIME OF TEST: 04/03/2025 01:12:31

Diagnostics Check	OK	01:06
Air Blank	0.000	01:06
Control Test	0.077	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.141	01:07
Breath Volume	1.714	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.138	01:11
Breath Volume	1.175	01:11
Air Blank	0.000	01:11
Control Test	0.076	01:11
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 04/23/2025 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: TREHARN

DATE/TIME OF TEST: 04/04/2025 03:58:50

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.077	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.139	03:54
Breath Volume	1.648	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:56
Subject Sample # 2	0.136	03:57
Breath Volume	1.128	03:57
Air Blank	0.000	03:57
Control Test	0.077	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 04/23/2025 11:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: WHITTLETON

DATE/TIME OF TEST: 04/18/2025 00:02:13

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.000	00:01	RFI Detect
Air Blank	0.000	00:02	

Data Download Date/Time: 04/23/2025 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 04/26/2025 00:44:16

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.076	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.065	00:39
Breath Volume	1.214	00:39
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.064	00:42
Breath Volume	1.925	00:42
Air Blank	0.000	00:43
Control Test	0.077	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 05/21/2025 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WARF,KYLE,J

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: REEVES,JAMES

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402783528

DRY GAS EXPIRY DATE: 06/27/2026

VIOLATION CODE: BUI

ARREST TIME: 15:29

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/17/2025 16:04:24

Diagnostics Check	OK	16:00	
Air Blank	0.000	16:00	
Control Test	0.080	16:01	
Air Blank	0.000	16:01	
Subject Sample	# 1	16:03	Subject Test Refused
Breath Volume	0.000	16:03	
Air Blank	0.000	16:03	
Control Test	0.081	16:03	
Air Blank	0.000	16:04	
Diagnostics Check	OK	16:04	

Data Download Date/Time: 04/21/2025 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402783528
BREATH TEST OPERATOR: GERBER,OLIVER,B	DRY GAS EXPIRY DATE: 06/27/2026
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: GERBER,OLIVER	ARREST TIME: 18:30
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 04/21/2025	OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: LOCKAMY	DATE/TIME OF TEST: 04/26/2025 19:11:50
------------------------------	--

Diagnostics Check	OK	19:05
Air Blank	0.000	19:05
Control Test	0.081	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.114	19:07
Breath Volume	3.390	19:07
Air Blank	0.000	19:07
Air Blank	0.000	19:09
Subject Sample # 2	0.111	19:10
Breath Volume	2.312	19:10
Air Blank	0.000	19:10
Control Test	0.080	19:11
Air Blank	0.000	19:11
Diagnostics Check	OK	19:11

Data Download Date/Time: 07/01/2025 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASEILLE,BRYANT,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: GONZALEZ,LUIS
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: OCANDO

DATE/TIME OF TEST: 04/25/2025 06:37:51

Diagnostics Check	OK	06:26	
Air Blank	0.000	06:26	
Control Test	0.083	06:26	
Air Blank	0.000	06:27	
Subject Sample	# 1 0.000	06:30	Volume Not Met
Breath Volume	0.082	06:30	
Air Blank	0.000	06:31	
Air Blank	0.000	06:33	
Subject Sample	# 2 0.169	06:36	Volume Not Met
Breath Volume	0.031	06:36	
Air Blank	0.000	06:36	
Control Test	0.080	06:37	
Air Blank	0.000	06:37	
Diagnostics Check	OK	06:37	

Data Download Date/Time: 05/30/2025 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASEILLE,BRYANT,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: GONZALEZ,LUIS
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 04/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: OCANDO

DATE/TIME OF TEST: 04/25/2025 06:46:34

Diagnostics Check	OK	06:43	
Air Blank	0.000	06:44	
Control Test	0.080	06:44	
Air Blank	0.000	06:45	
Subject Sample	# 1	06:45	Subject Test Refused
Breath Volume	0.000	06:45	
Air Blank	0.000	06:45	
Control Test	0.081	06:45	
Air Blank	0.000	06:46	
Diagnostics Check	OK	06:46	

Data Download Date/Time: 05/30/2025 20:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CAMILLERI, ZACHARIAS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: STAAL

DATE/TIME OF TEST: 04/06/2025 00:33:44

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:25	
Control Test	0.083	00:25	
Air Blank	0.000	00:26	
Subject Sample	# 1 0.169	00:27	
Breath Volume	2.761	00:27	
Air Blank	0.000	00:27	
Air Blank	0.000	00:29	
Subject Sample	# 2	00:32	Subject Test Refused
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Control Test	0.084	00:33	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 04/21/2025 09:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLUESNER,D,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MILLIGAN,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:18

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/07/2025 09:47:30

Diagnostics Check	OK	09:40
Air Blank	0.000	09:41
Control Test	0.083	09:41
Air Blank	0.000	09:42
Subject Sample # 1	0.183	09:43
Breath Volume	1.625	09:43
Air Blank	0.000	09:43
Air Blank	0.000	09:45
Subject Sample # 2	0.179	09:45
Breath Volume	1.175	09:45
Air Blank	0.000	09:46
Control Test	0.082	09:46
Air Blank	0.000	09:47
Diagnostics Check	OK	09:47

Data Download Date/Time: 04/21/2025 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: REYNOLDS,C
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: TOMAS GARCIA

DATE/TIME OF TEST: 04/13/2025 00:49:13

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.083	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.200	00:44
Breath Volume	1.582	00:44
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.202	00:47
Breath Volume	1.226	00:47
Air Blank	0.000	00:48
Control Test	0.083	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 04/21/2025 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEE,BRYCE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: CRANDALL

DATE/TIME OF TEST: 04/18/2025 22:37:28

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:27	
Control Test	0.083	22:28	
Air Blank	0.000	22:28	
Subject Sample	# 1 0.111	22:31	Volume Not Met
Breath Volume	0.093	22:31	
Air Blank	0.000	22:32	
Air Blank	0.000	22:34	
Subject Sample	# 2	22:35	Subject Test Refused
Breath Volume	0.097	22:35	
Air Blank	0.000	22:36	
Control Test	0.081	22:36	
Air Blank	0.000	22:37	
Diagnostics Check	OK	22:37	

Data Download Date/Time: 04/21/2025 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRARO,PAUL,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TIRADO,B
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: CORNWELL

DATE/TIME OF TEST: 04/19/2025 23:36:12

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.084	23:34	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:34	Subject Test Refused
Breath Volume	0.000	23:34	
Air Blank	0.000	23:35	
Control Test	0.082	23:35	
Air Blank		23:35	RFI Detect
Air Blank	0.000	23:36	

Data Download Date/Time: 04/21/2025 09:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: SOMARRIBA

DATE/TIME OF TEST: 04/12/2025 02:34:13

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.161	02:29
Breath Volume	2.738	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.141	02:32
Breath Volume	2.050	02:32
Air Blank	0.000	02:33
Control Test	0.076	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: URBINA TORRES

DATE/TIME OF TEST: 04/13/2025 05:36:25

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.081	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.148	05:31
Breath Volume	2.843	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.142	05:34
Breath Volume	1.992	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: MURILLO SINISTERRA

DATE/TIME OF TEST: 04/14/2025 05:39:53

Diagnostics Check	OK	05:33
Air Blank	0.000	05:34
Control Test	0.080	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.156	05:35
Breath Volume	2.386	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:37
Subject Sample # 2	0.149	05:38
Breath Volume	2.761	05:38
Air Blank	0.000	05:38
Control Test	0.080	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORTUGUES,GABRIEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PORTUGUES,GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 04/19/2025 00:40:27

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.091	00:35
Breath Volume	1.554	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.093	00:38
Breath Volume	1.429	00:38
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: MORENO HERNANDEZ

DATE/TIME OF TEST: 04/19/2025 23:55:16

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.182	23:50
Breath Volume	1.753	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.171	23:53
Breath Volume	1.621	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: LUMBI ROMERO

DATE/TIME OF TEST: 04/21/2025 00:23:32

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.081	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.269	00:19
Breath Volume	2.988	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.276	00:21
Breath Volume	2.687	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/22/2025 04:08:25

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.277	04:03
Breath Volume	2.765	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.265	04:06
Breath Volume	2.277	04:06
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: DEL BARRIO LOPEZ

DATE/TIME OF TEST: 04/24/2025 22:30:22

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.081	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.197	22:25
Breath Volume	2.390	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.199	22:28
Breath Volume	1.574	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLOUGHEN,KORY,I
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MCCLOUGHEN,KORY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 03/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: PENA CAZARES

DATE/TIME OF TEST: 04/01/2025 21:26:51

Diagnostics Check	OK	21:20
Air Blank	0.000	21:21
Control Test	0.081	21:21
Air Blank	0.000	21:21
Subject Sample # 1	0.009	21:22
Breath Volume	3.214	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.008	21:25
Breath Volume	2.867	21:25
Air Blank	0.000	21:25
Control Test	0.081	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 04/02/2025 02:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLOUGHEN,KORY,I
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MCCLOUGHEN,KORY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GRAY KELLY

DATE/TIME OF TEST: 04/06/2025 03:45:16

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.082	03:40	
Air Blank	0.000	03:40	
Subject Sample # 1	0.073	03:41	
Breath Volume	3.339	03:41	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample # 2		03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 05/13/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALFONSO,CHRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: KEYTE

DATE/TIME OF TEST: 04/06/2025 04:35:29

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.159	04:30
Breath Volume	2.117	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.152	04:33
Breath Volume	1.500	04:33
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 05/13/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/08/2025 22:17:32

Diagnostics Check	OK	22:16	
Air Blank		22:17	RFI Detect
Air Blank	0.000	22:17	

Data Download Date/Time: 05/13/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/08/2025 22:25:08

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.081	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.139	22:20
Breath Volume	3.390	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.131	22:23
Breath Volume	3.707	22:23
Air Blank	0.000	22:24
Control Test	0.081	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 05/13/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANKOSKY,DON,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STANKOSKY,DON
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 04/12/2025 19:57:50

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:43	
Control Test	0.082	19:43	
Air Blank	0.000	19:44	
Subject Sample # 1	0.000	19:45	
Breath Volume	3.332	19:45	
Air Blank	0.000	19:45	
Air Blank	0.000	19:47	
Subject Sample # 2		19:50	No Sample Provided
Breath Volume	0.000	19:50	
Air Blank	0.000	19:51	
Air Blank	0.000	19:53	
Subject Sample # 3		19:56	No Sample Provided
Breath Volume	0.000	19:56	
Air Blank	0.000	19:56	
Control Test	0.081	19:57	
Air Blank	0.000	19:57	
Diagnostics Check	OK	19:57	

Data Download Date/Time: 05/13/2025 02:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KERN, PETER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 04/13/2025 03:09:37

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.082	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.076	03:04
Breath Volume	3.062	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.074	03:08
Breath Volume	3.027	03:08
Air Blank	0.000	03:08
Control Test	0.081	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 05/13/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VEGA, JAHMIL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BENECHÉ

DATE/TIME OF TEST: 04/14/2025 00:24:21

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.082	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.309	00:17
Breath Volume	1.585	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.302	00:22
Breath Volume	1.128	00:22
Air Blank	0.000	00:23
Control Test	0.082	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 05/13/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: DOANE

DATE/TIME OF TEST: 04/17/2025 00:30:36

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.081	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.166	00:26
Breath Volume	1.449	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.166	00:29
Breath Volume	1.468	00:29
Air Blank	0.000	00:29
Control Test	0.081	00:30
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 05/13/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,ERIC,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALMENDAREZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: SALCEDO RODRIGUEZ

DATE/TIME OF TEST: 04/17/2025 23:20:23

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.082	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.171	23:15
Breath Volume	3.117	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.159	23:18
Breath Volume	2.621	23:18
Air Blank	0.000	23:19
Control Test	0.080	23:19
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: DAGNILLO

DATE/TIME OF TEST: 04/18/2025 17:36:21

Diagnostics Check	OK	17:29
Air Blank	0.000	17:30
Control Test	0.082	17:30
Air Blank	0.000	17:31
Subject Sample # 1	0.126	17:31
Breath Volume	2.179	17:31
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample # 2	0.125	17:34
Breath Volume	2.152	17:34
Air Blank	0.000	17:35
Control Test	0.080	17:35
Air Blank	0.000	17:36
Diagnostics Check	OK	17:36

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIZARDO, RAILY, J
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ARD, RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ZNAJMIECKI

DATE/TIME OF TEST: 04/19/2025 01:31:02

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.082	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.196	01:26
Breath Volume	1.230	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.187	01:29
Breath Volume	1.660	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:31

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLOUGHEN,KORY,I
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MCCLOUGHEN,KORY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 04/20/2025 00:39:22

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.081	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.225	00:34
Breath Volume	2.460	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.214	00:37
Breath Volume	2.890	00:37
Air Blank	0.000	00:38
Control Test	0.082	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALFONSO,CHRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ARIAS CRUZ

DATE/TIME OF TEST: 04/20/2025 03:44:04

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.082	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.200	03:39
Breath Volume	1.894	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.191	03:42
Breath Volume	3.843	03:42
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:44

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRITZ, TYLER, S
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALFONSO, CHRISTIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: MENDEZ SANCHEZ

DATE/TIME OF TEST: 04/24/2025 18:11:30

Diagnostics Check	OK	18:05
Air Blank	0.000	18:05
Control Test	0.082	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.000	18:07
Breath Volume	1.109	18:07
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.000	18:10
Breath Volume	4.242	18:10
Air Blank	0.000	18:10
Control Test	0.081	18:10
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: DOSCH

DATE/TIME OF TEST: 04/24/2025 18:31:19

Diagnostics Check	OK	18:25
Air Blank	0.000	18:25
Control Test	0.081	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.000	18:27
Breath Volume	1.972	18:27
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.000	18:29
Breath Volume	2.132	18:29
Air Blank	0.000	18:30
Control Test	0.081	18:30
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: CARRERA

DATE/TIME OF TEST: 04/26/2025 01:18:07

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.082	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.242	01:13
Breath Volume	2.042	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.223	01:16
Breath Volume	2.617	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,KWESI,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,KWESI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VELA

DATE/TIME OF TEST: 04/29/2025 23:36:34

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.082	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.230	23:32
Breath Volume	2.628	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.227	23:34
Breath Volume	2.652	23:34
Air Blank	0.000	23:35
Control Test	0.081	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,AVERY,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,AVERY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/05/2025 23:37:42

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.079	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.000	23:33
Breath Volume	3.664	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.000	23:36
Breath Volume	3.535	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 05/13/2025 14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVACK,BRADLEY,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: NOVACK,BRADLEY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/18/2025 07:00:48

Diagnostics Check	OK	07:00	
Air Blank		07:00	RFI Detect
Air Blank		07:00	RFI Detect

Data Download Date/Time: 05/13/2025 14:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: BARTON,LORA
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BREEN

DATE/TIME OF TEST: 04/04/2025 22:22:12

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.076	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.222	22:17
Breath Volume	2.667	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:19
Subject Sample # 2	0.217	22:20
Breath Volume	2.203	22:20
Air Blank	0.000	22:21
Control Test	0.075	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 07/07/2025 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: EPPERS,DYLAN
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 04/04/2025 22:40:37

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.078	22:38	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:39	Subject Test Refused
Breath Volume	0.000	22:39	
Air Blank	0.000	22:39	
Control Test	0.080	22:40	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 07/07/2025 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: BARTON,LORA
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BREEN

DATE/TIME OF TEST: 04/04/2025 21:46:43

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:45	
Control Test	0.074	21:46	Control Outside Tolerance
Air Blank	0.000	21:46	

Data Download Date/Time: 04/27/2025 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: BARTON,LORA
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: BREEN

DATE/TIME OF TEST: 04/04/2025 21:50:11

Diagnostics Check	OK	21:48	
Air Blank	0.000	21:49	
Control Test	0.071	21:49	Control Outside Tolerance
Air Blank		21:50	Ambient Fail

Data Download Date/Time: 04/27/2025 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HUAPILLA,MIGUEL
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 04/09/2025 01:07:17

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.078	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.152	01:02
Breath Volume	1.445	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.158	01:05
Breath Volume	2.066	01:05
Air Blank	0.000	01:06
Control Test	0.075	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 04/27/2025 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: EPPERS,DYLAN
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 04/18/2025 01:44:17

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.078	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:42	Subject Test Refused
Breath Volume	0.000	01:42	
Air Blank	0.000	01:43	
Control Test	0.078	01:43	
Air Blank	0.000	01:44	
Diagnostics Check	OK	01:44	

Data Download Date/Time: 04/27/2025 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HERNANDEZ,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BOOKNES

DATE/TIME OF TEST: 04/19/2025 03:04:24

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:01	
Control Test	0.077	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:03	Subject Test Refused
Breath Volume	0.000	03:03	
Air Blank	0.000	03:03	
Control Test	0.078	03:03	
Air Blank	0.000	03:04	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 04/27/2025 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNTER,CURTIS,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HUNTER,CURTIS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: HEATHCOE

DATE/TIME OF TEST: 04/01/2025 16:45:32

Diagnostics Check	OK	16:42	
Air Blank	0.000	16:43	
Control Test	0.079	16:43	
Air Blank	0.000	16:43	
Subject Sample	# 1	16:44	Subject Test Refused
Breath Volume	0.000	16:44	
Air Blank	0.000	16:44	
Control Test	0.079	16:44	
Air Blank	0.000	16:45	
Diagnostics Check	OK	16:45	

Data Download Date/Time: 04/30/2025 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,ROLF,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SHIVER,ROBERT
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: REYNESO

DATE/TIME OF TEST: 04/12/2025 22:46:35

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.220	22:41
Breath Volume	2.359	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:43
Subject Sample # 2	0.212	22:44
Breath Volume	2.894	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 04/30/2025 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROGERSON,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 04/05/2025 21:32:47

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.080	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.124	21:28
Breath Volume	2.324	21:28
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.117	21:31
Breath Volume	1.101	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 04/29/2025 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: NEWELL,CHRISTINE
ARRESTING OFFICER AGENCY: Monticello PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: EVENS

DATE/TIME OF TEST: 04/11/2025 01:26:58

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.226	01:22
Breath Volume	2.105	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.225	01:25
Breath Volume	2.281	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 04/29/2025 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA,COLBY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROSMINI,JOE
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 04/21/2025 17:45:45

Diagnostics Check	OK	17:42	
Air Blank	0.000	17:42	
Control Test	0.080	17:43	
Air Blank	0.000	17:43	
Subject Sample	# 1	17:44	Subject Test Refused
Breath Volume	0.000	17:44	
Air Blank	0.000	17:44	
Control Test	0.080	17:45	
Air Blank	0.000	17:45	
Diagnostics Check	OK	17:45	

Data Download Date/Time: 04/29/2025 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA,COLBY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BETHEA,COLBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: DEXTER

DATE/TIME OF TEST: 04/25/2025 20:47:12

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:44	
Control Test	0.080	20:45	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:46	
Control Test	0.080	20:46	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 04/29/2025 15:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL, EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/03/2025 01:40:57

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.081	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.184	01:36
Breath Volume	1.914	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.187	01:39
Breath Volume	1.320	01:39
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:40
Diagnostics Check	OK	01:40

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SMITH,ALEXANDER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 04/05/2025 00:20:18

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.081	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.050	00:14
Breath Volume	2.191	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.045	00:18
Breath Volume	1.726	00:18
Air Blank	0.000	00:19
Control Test	0.081	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: HOFFMAN,AUSTIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/06/2025 05:40:02

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.080	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.093	05:35
Breath Volume	3.558	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.088	05:38
Breath Volume	1.851	05:38
Air Blank	0.000	05:39
Control Test	0.081	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:40

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN,DANIEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 04/06/2025 22:25:11

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.082	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.286	22:20
Breath Volume	1.527	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:22
Subject Sample # 2	0.271	22:23
Breath Volume	1.531	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL, EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: VILLALON SANCHEZ

DATE/TIME OF TEST: 04/07/2025 00:50:07

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.119	00:45
Breath Volume	1.503	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.110	00:48
Breath Volume	1.402	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSH, TYLER, C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MOULTON, MAURICE
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 04/11/2025 05:46:20

Diagnostics Check	OK	05:39
Air Blank	0.000	05:40
Control Test	0.081	05:40
Air Blank	0.000	05:41
Subject Sample # 1	0.236	05:41
Breath Volume	2.003	05:41
Air Blank	0.000	05:42
Air Blank	0.000	05:44
Subject Sample # 2	0.242	05:44
Breath Volume	2.734	05:44
Air Blank	0.000	05:45
Control Test	0.081	05:45
Air Blank	0.000	05:46
Diagnostics Check	OK	05:46

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DARBY

DATE/TIME OF TEST: 04/13/2025 02:20:43

Diagnostics Check	OK	02:19	
Air Blank		02:20	RFI Detect
Air Blank	0.000	02:20	

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DARBY

DATE/TIME OF TEST: 04/13/2025 02:24:48

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:23	
Control Test	0.081	02:23	
Air Blank	0.000	02:24	
		02:24	RFI Detect
Air Blank	0.000	02:24	

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DARBY

DATE/TIME OF TEST: 04/13/2025 02:32:02

Diagnostics Check	OK	02:31	
Air Blank		02:31	RFI Detect
Air Blank		02:32	RFI Detect

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DARBY

DATE/TIME OF TEST: 04/13/2025 02:42:17

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.082	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.150	02:37
Breath Volume	1.718	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.153	02:40
Breath Volume	2.593	02:40
Air Blank	0.000	02:41
Control Test	0.081	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILCOX,DAVID,B
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PHILCOX,DAVID
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/15/2025 22:32:45

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:29	
Control Test	0.081	22:29	
Air Blank	0.000	22:29	
Subject Sample	# 1 0.194	22:30	
Breath Volume	2.687	22:30	
Air Blank		22:31	RFI Detect
Air Blank	0.000	22:31	
Control Test	0.082	22:32	
Air Blank	0.000	22:32	
Diagnostics Check	OK	22:32	

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILCOX,DAVID,B
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PHILCOX,DAVID
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/15/2025 22:44:22

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.081	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.192	22:36
Breath Volume	2.269	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:38
Subject Sample # 2	0.168	22:39
Breath Volume	1.906	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 3	0.181	22:42
Breath Volume	2.519	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/16/2025 18:06:23

Diagnostics Check	OK	17:59
Air Blank	0.000	18:00
Control Test	0.082	18:00
Air Blank	0.000	18:01
Subject Sample # 1	0.399	18:01
Breath Volume	2.445	18:01
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.391	18:04
Breath Volume	2.628	18:04
Air Blank	0.000	18:05
Control Test	0.081	18:05
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERGEN,PATRICK,S
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: BEAUBRUM,ULNA
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GOLDENBERG

DATE/TIME OF TEST: 04/22/2025 01:07:25

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:55	
Control Test	0.079	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:59	No Sample Provided
Breath Volume	0.000	00:59	
Air Blank	0.000	00:59	
Air Blank	0.000	01:01	
Subject Sample	# 2	01:02	
Breath Volume	2.515	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 3	01:05	
Breath Volume	2.468	01:05	
Air Blank	0.000	01:06	
Control Test	0.079	01:06	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MCCABE,PATRICK
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 04/23/2025 02:03:44

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.081	01:52	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:56	No Sample Provided
Breath Volume	0.000	01:56	
Air Blank	0.000	01:56	
Air Blank	0.000	01:58	
Subject Sample	# 2	02:01	Volume Not Met
Breath Volume	0.195	02:01	
Air Blank	0.000	02:02	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Diagnostics Check	OK	02:03	

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MCCABE,PATRICK
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 04/23/2025 02:17:01

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.079	02:06	
Air Blank	0.000	02:06	
Subject Sample	# 1 0.219	02:09	Volume Not Met
Breath Volume	0.257	02:09	
Air Blank	0.000	02:10	
Air Blank	0.000	02:12	
Subject Sample	# 2 0.185	02:15	Volume Not Met
Breath Volume	0.019	02:15	
Air Blank	0.000	02:16	
Control Test	0.078	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TICE

DATE/TIME OF TEST: 04/27/2025 03:46:10

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.081	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1 0.126	03:39	Volume Not Met
Breath Volume	0.160	03:39	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample	# 2 0.119	03:44	Volume Not Met
Breath Volume	0.304	03:44	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TICE

DATE/TIME OF TEST: 04/27/2025 03:55:47

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:48	
Control Test	0.080	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1 0.114	03:52	Volume Not Met
Breath Volume	0.097	03:52	
Air Blank	0.000	03:53	
Air Blank		03:55	RFI Detect
Air Blank		03:55	RFI Detect
Air Blank	0.000	03:55	

Data Download Date/Time: 04/28/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENAFIELD,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BENAFIELD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 04/14/2025 21:02:50

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Subject Sample # 1	0.184	20:58
Breath Volume	3.308	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:00
Subject Sample # 2	0.168	21:01
Breath Volume	1.851	21:01
Air Blank	0.000	21:01
Control Test	0.079	21:02
Air Blank	0.000	21:02
Diagnostics Check	OK	21:02

Data Download Date/Time: 04/29/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ, RESEAN, R
BREATH TEST OPERATOR AGENCY: Midway PD
ARRESTING OFFICER: KIMBLE, PARADISE
ARRESTING OFFICER AGENCY: Midway PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: DUBOSE

DATE/TIME OF TEST: 04/16/2025 21:14:49

Diagnostics Check	OK	21:08
Air Blank	0.000	21:09
Control Test	0.082	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.297	21:10
Breath Volume	1.949	21:10
Air Blank	0.000	21:11
Air Blank	0.000	21:12
Subject Sample # 2	0.293	21:13
Breath Volume	2.023	21:13
Air Blank	0.000	21:13
Control Test	0.081	21:14
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 04/29/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALDAY,SETH,A
BREATH TEST OPERATOR AGENCY: Chattahoochee PD
ARRESTING OFFICER: ALDAY,SETH
ARRESTING OFFICER AGENCY: Chattahoochee PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: STREATER

DATE/TIME OF TEST: 04/19/2025 03:47:41

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.124	03:43
Breath Volume	2.339	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.128	03:46
Breath Volume	2.343	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 04/29/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALDAY,SETH,A
BREATH TEST OPERATOR AGENCY: Chattahoochee PD
ARRESTING OFFICER: ALDAY,SETH
ARRESTING OFFICER AGENCY: Chattahoochee PD
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: LOCKWOOD

DATE/TIME OF TEST: 04/20/2025 05:40:52

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.081	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.100	05:36
Breath Volume	1.367	05:36
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.100	05:39
Breath Volume	1.410	05:39
Air Blank	0.000	05:39
Control Test	0.082	05:40
Air Blank	0.000	05:40
Diagnostics Check	OK	05:40

Data Download Date/Time: 04/29/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENAFIELD,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BENAFIELD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: MENDOZA MOCTEZUMA

DATE/TIME OF TEST: 04/20/2025 22:45:48

Diagnostics Check	OK	22:45	
Air Blank		22:45	RFI Detect
Air Blank	0.000	22:45	

Data Download Date/Time: 04/29/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

INSTRUMENT REGISTERED TO: GADSDEN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENAFIELD,RYAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BENAFIELD,RYAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MENDOZA MOCTEZUMA

DATE/TIME OF TEST: 04/20/2025 22:53:25

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.115	22:48
Breath Volume	2.765	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.117	22:51
Breath Volume	2.835	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 04/29/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALLOW, CHARLES, S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MALLOW, CHARLES
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: BUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 04/12/2025 18:44:33

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.082	18:39
Air Blank	0.000	18:39
Subject Sample # 1	0.122	18:40
Breath Volume	3.187	18:40
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.118	18:42
Breath Volume	3.343	18:42
Air Blank	0.000	18:43
Control Test	0.080	18:43
Air Blank	0.000	18:44
Diagnostics Check	OK	18:44

Data Download Date/Time: 04/30/2025 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILL,HEATHER,L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: WILL,HEATHER
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: HOLLEY

DATE/TIME OF TEST: 04/12/2025 19:52:05

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.081	19:46
Air Blank	0.000	19:46
Subject Sample # 1	0.109	19:47
Breath Volume	1.324	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.108	19:50
Breath Volume	1.671	19:50
Air Blank	0.000	19:51
Control Test	0.082	19:51
Air Blank	0.000	19:51
Diagnostics Check	OK	19:52

Data Download Date/Time: 04/30/2025 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILL, HEATHER, L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: PARKER, JOSH
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: GLASS

DATE/TIME OF TEST: 04/13/2025 02:45:53

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.000	02:45	RFI Detect
Air Blank		02:45	RFI Detect

Data Download Date/Time: 04/30/2025 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILL, HEATHER, L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: PARKER, JOSH
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: GLASS

DATE/TIME OF TEST: 04/13/2025 02:50:14

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.081	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Control Test	0.081	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 04/30/2025 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: ALEXANDER,MICHAEL
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 04/16/2025 20:18:19

Diagnostics Check	OK	20:15	
Air Blank	0.000	20:15	
Control Test	0.082	20:15	
Air Blank	0.000	20:16	
Subject Sample	# 1 0.251	20:16	
Breath Volume	2.597	20:16	
Air Blank		20:17	Purge Fail
Air Blank	0.000	20:18	

Data Download Date/Time: 04/30/2025 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDER,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: ALEXANDER,MICHAEL
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 03/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 04/16/2025 20:28:13

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.081	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.250	20:23
Breath Volume	1.687	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.242	20:26
Breath Volume	1.300	20:26
Air Blank	0.000	20:27
Control Test	0.081	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 04/30/2025 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,JASON,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BASS,JASON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: CALLEJAS

DATE/TIME OF TEST: 04/05/2025 22:54:40

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.200	22:49
Breath Volume	2.148	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.195	22:52
Breath Volume	1.656	22:52
Air Blank	0.000	22:53
Control Test	0.079	22:54
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 04/29/2025 15:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: PEREZQUINONES

DATE/TIME OF TEST: 04/05/2025 00:22:44

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.081	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.140	00:18
Breath Volume	2.382	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.124	00:21
Breath Volume	1.453	00:21
Air Blank	0.000	00:21
Control Test	0.081	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 04/22/2025 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: MC GLAMORY

DATE/TIME OF TEST: 04/06/2025 05:06:36

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.081	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.150	05:01
Breath Volume	3.660	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.148	05:05
Breath Volume	3.863	05:05
Air Blank	0.000	05:05
Control Test	0.081	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 04/22/2025 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARC,VUKANSON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: VILLENA

DATE/TIME OF TEST: 04/06/2025 21:56:58

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.081	21:50
Air Blank	0.000	21:51
Subject Sample # 1	0.156	21:52
Breath Volume	1.316	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.161	21:55
Breath Volume	1.296	21:55
Air Blank	0.000	21:56
Control Test	0.081	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 04/22/2025 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ MARTINEZ,HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 04/07/2025 02:04:05

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.082	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.188	01:57
Breath Volume	1.164	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.207	02:02
Breath Volume	3.000	02:02
Air Blank	0.000	02:03
Control Test	0.081	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:04

Data Download Date/Time: 04/22/2025 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,DANIELLI
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: GULINO

DATE/TIME OF TEST: 04/10/2025 04:11:03

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.106	04:06
Breath Volume	2.050	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.107	04:09
Breath Volume	1.765	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:11

Data Download Date/Time: 04/22/2025 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: RODRIGUEZ BOYET

DATE/TIME OF TEST: 04/22/2025 22:25:18

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.079	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.133	22:21
Breath Volume	3.183	22:21
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.122	22:23
Breath Volume	1.222	22:23
Air Blank	0.000	22:24
Control Test	0.078	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECKER, THOMAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 04/23/2025 02:08:36

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.079	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Control Test	0.080	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSHUA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: GALT

DATE/TIME OF TEST: 04/23/2025 18:43:58

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.079	18:38
Air Blank	0.000	18:39
Subject Sample # 1	0.101	18:39
Breath Volume	1.703	18:39
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.100	18:42
Breath Volume	2.648	18:42
Air Blank	0.000	18:43
Control Test	0.079	18:43
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARTER,DAVID
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: JAUDON

DATE/TIME OF TEST: 04/23/2025 19:18:22

Diagnostics Check	OK	19:08
Air Blank	0.000	19:09
Control Test	0.080	19:09
Air Blank	0.000	19:10
Subject Sample # 1	0.177	19:11
Breath Volume	1.339	19:11
Air Blank	0.000	19:11
Air Blank	0.000	19:13
Subject Sample # 2	0.201	19:13
Breath Volume	1.453	19:13
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 3	0.217	19:16
Breath Volume	1.441	19:16
Air Blank	0.000	19:17
Control Test	0.078	19:17
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TRENDAFILOV, NICHOLAS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: HERNANDEZ MORALES

DATE/TIME OF TEST: 04/23/2025 20:52:07

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Subject Sample # 1	0.157	20:47
Breath Volume	2.390	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.159	20:50
Breath Volume	3.347	20:50
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: RIVERA GARCIA

DATE/TIME OF TEST: 04/25/2025 04:05:47

Diagnostics Check	OK	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.217	04:01
Breath Volume	2.285	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.219	04:04
Breath Volume	2.074	04:04
Air Blank	0.000	04:04
Control Test	0.079	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARRISON,ELIJAH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MURRIETAMANNIN

DATE/TIME OF TEST: 04/25/2025 04:43:20

Diagnostics Check	OK	04:35
Air Blank	0.000	04:35
Control Test	0.080	04:36
Air Blank	0.000	04:36
Subject Sample # 1	0.190	04:39
Breath Volume	2.164	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.193	04:41
Breath Volume	1.816	04:41
Air Blank	0.000	04:42
Control Test	0.078	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 04/25/2025 09:42:37

Diagnostics Check	OK	09:35
Air Blank	0.000	09:35
Control Test	0.079	09:35
Air Blank	0.000	09:36
Subject Sample # 1	0.187	09:37
Breath Volume	1.207	09:37
Air Blank	0.000	09:38
Air Blank	0.000	09:40
Subject Sample # 2	0.195	09:41
Breath Volume	1.644	09:41
Air Blank	0.000	09:41
Control Test	0.079	09:42
Air Blank	0.000	09:42
Diagnostics Check	OK	09:42

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TUKES,JUSTIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: VOLLEAU DIAZ

DATE/TIME OF TEST: 04/26/2025 01:06:00

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.079	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1 0.083	00:56	Volume Not Met
Breath Volume	1.082	00:56	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample	# 2 0.083	00:59	
Breath Volume	1.542	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 3 0.085	01:04	
Breath Volume	1.246	01:04	
Air Blank	0.000	01:05	
Control Test	0.080	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA PASTORIZA,ARNALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BISPO,STEPHANIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WRIGHT II

DATE/TIME OF TEST: 04/26/2025 02:03:51

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.046	01:59
Breath Volume	3.785	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:01
Subject Sample # 2	0.044	02:02
Breath Volume	3.960	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GARCIARUJANO

DATE/TIME OF TEST: 04/26/2025 02:59:18

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.080	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.171	02:54
Breath Volume	2.050	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.177	02:57
Breath Volume	2.832	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALFRED,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: STILLMAN

DATE/TIME OF TEST: 04/26/2025 04:41:12

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.206	04:33
Breath Volume	1.433	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.183	04:36
Breath Volume	1.777	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 3	0.191	04:39
Breath Volume	1.574	04:39
Air Blank	0.000	04:40
Control Test	0.079	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARDY,BRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 04/26/2025 22:54:19

Diagnostics Check	OK	22:48
Air Blank	0.000	22:48
Control Test	0.079	22:49
Air Blank	0.000	22:49
Subject Sample # 1	0.047	22:50
Breath Volume	1.570	22:50
Air Blank	0.000	22:50
Air Blank	0.000	22:52
Subject Sample # 2	0.047	22:52
Breath Volume	2.789	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECKER,THOMAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: JEUNE

DATE/TIME OF TEST: 04/27/2025 00:54:07

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.026	00:49
Breath Volume	1.683	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.029	00:52
Breath Volume	1.363	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: VAUGHAN

DATE/TIME OF TEST: 04/27/2025 03:45:43

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.223	03:41
Breath Volume	1.464	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.210	03:44
Breath Volume	1.937	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSA,BRANDON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: PAYAMPSALMONTE

DATE/TIME OF TEST: 04/27/2025 15:40:58

Diagnostics Check	OK	15:35
Air Blank	0.000	15:35
Control Test	0.080	15:35
Air Blank	0.000	15:36
Subject Sample # 1	0.008	15:36
Breath Volume	3.691	15:36
Air Blank	0.000	15:37
Air Blank	0.000	15:39
Subject Sample # 2	0.007	15:39
Breath Volume	3.296	15:39
Air Blank	0.000	15:40
Control Test	0.080	15:40
Air Blank	0.000	15:40
Diagnostics Check	OK	15:40

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARDY,BRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 04/28/2025 01:46:17

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.000	01:42
Breath Volume	4.085	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.000	01:44
Breath Volume	4.046	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIELS,DAMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 04/28/2025 02:51:08

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.145	02:46
Breath Volume	1.531	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.142	02:49
Breath Volume	1.171	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORDOVA,ODELIA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/30/2025 02:58:29

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.179	02:53
Breath Volume	2.394	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.178	02:57
Breath Volume	2.363	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ MARTINEZ,HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 04/30/2025 03:45:05

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1 0.127	03:36	Volume Not Met
Breath Volume	0.968	03:36	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2 0.126	03:39	
Breath Volume	1.441	03:39	
Air Blank	0.000	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 3 0.129	03:43	
Breath Volume	1.507	03:43	
Air Blank	0.000	03:44	
Control Test	0.079	03:44	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERARDI,DEPUTY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 04/01/2025 23:50:22

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.077	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.184	23:45
Breath Volume	2.113	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.176	23:48
Breath Volume	1.753	23:48
Air Blank	0.000	23:49
Control Test	0.076	23:49
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 04/20/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 04/04/2025 22:55:34

Diagnostics Check	OK	22:49
Air Blank	0.000	22:49
Control Test	0.077	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.099	22:50
Breath Volume	2.023	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.099	22:54
Breath Volume	2.113	22:54
Air Blank	0.000	22:54
Control Test	0.077	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 04/20/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: LAMERE

DATE/TIME OF TEST: 04/05/2025 20:38:34

Diagnostics Check	OK	20:31
Air Blank	0.000	20:32
Control Test	0.077	20:32
Air Blank	0.000	20:33
Subject Sample # 1	0.149	20:33
Breath Volume	1.675	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.152	20:36
Breath Volume	1.585	20:36
Air Blank	0.000	20:37
Control Test	0.077	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 04/20/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LOMAGLIO

DATE/TIME OF TEST: 04/05/2025 23:07:01

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.077	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.090	23:02
Breath Volume	2.628	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.087	23:05
Breath Volume	1.726	23:05
Air Blank	0.000	23:06
Control Test	0.077	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:07

Data Download Date/Time: 04/20/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,WILLIAM,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: RAMOS,KEARSTIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: AMES

DATE/TIME OF TEST: 04/06/2025 19:41:54

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.077	19:36
Air Blank	0.000	19:36
Subject Sample # 1	0.214	19:37
Breath Volume	1.792	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:39
Subject Sample # 2	0.202	19:40
Breath Volume	1.722	19:40
Air Blank	0.000	19:40
Control Test	0.076	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 04/20/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LARocca

DATE/TIME OF TEST: 04/20/2025 02:29:45

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.076	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.145	02:25
Breath Volume	3.230	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.141	02:28
Breath Volume	2.976	02:28
Air Blank	0.000	02:28
Control Test	0.077	02:29
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 04/20/2025 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LAMANDIA,V
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: FLAIM

DATE/TIME OF TEST: 04/20/2025 04:19:24

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.076	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:10	No Sample Provided
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	0.197	04:14
Breath Volume	1.863	04:14	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 3	0.187	04:17
Breath Volume	1.324	04:17	
Air Blank	0.000	04:18	
Control Test	0.076	04:18	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

Data Download Date/Time: 04/20/2025 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIXSON,JEREMY,C
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HIXSON,JEREMY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/25/2025 02:27:47

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.078	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Control Test	0.078	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 05/16/2025 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: NESS

DATE/TIME OF TEST: 04/29/2025 02:53:41

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.077	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.127	02:49
Breath Volume	1.890	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.128	02:52
Breath Volume	2.074	02:52
Air Blank	0.000	02:52
Control Test	0.075	02:53
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 05/16/2025 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ALEXIOU

DATE/TIME OF TEST: 04/30/2025 01:44:56

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.078	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.000	01:40
Breath Volume	2.535	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.000	01:43
Breath Volume	2.578	01:43
Air Blank	0.000	01:44
Control Test	0.078	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 05/16/2025 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,CORINA,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: MERCADO,RUBEN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RUIZ GONZALEZ

DATE/TIME OF TEST: 04/02/2025 04:03:06

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.070	03:58
Breath Volume	2.886	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.064	04:01
Breath Volume	1.960	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWEENEY,MICHAEL,W
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SWEENEY,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 04/04/2025 21:57:11

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.079	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.208	21:49
Breath Volume	3.339	21:49
Air Blank	0.000	21:50
Air Blank	0.000	21:51
Subject Sample # 2	0.187	21:52
Breath Volume	1.792	21:52
Air Blank	0.000	21:53
Air Blank	0.000	21:54
Subject Sample # 3	0.205	21:55
Breath Volume	1.632	21:55
Air Blank	0.000	21:56
Control Test	0.077	21:56
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: AULTMAN,SHAWN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 04/06/2025 17:48:59

Diagnostics Check	OK	17:42
Air Blank	0.000	17:43
Control Test	0.079	17:43
Air Blank	0.000	17:44
Subject Sample # 1	0.236	17:44
Breath Volume	2.984	17:44
Air Blank	0.000	17:45
Air Blank	0.000	17:47
Subject Sample # 2	0.233	17:47
Breath Volume	2.269	17:47
Air Blank	0.000	17:48
Control Test	0.079	17:48
Air Blank	0.000	17:48
Diagnostics Check	OK	17:48

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMALDI, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BARR, ASHLEY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: GASS

DATE/TIME OF TEST: 04/10/2025 03:06:08

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.252	03:01
Breath Volume	1.484	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.239	03:04
Breath Volume	2.835	03:04
Air Blank	0.000	03:05
Control Test	0.077	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 04/10/2025 23:38:13

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.078	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.155	23:33
Breath Volume	2.320	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.158	23:36
Breath Volume	2.121	23:36
Air Blank	0.000	23:37
Control Test	0.077	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AULTMAN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VICKERS,SETH
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: LIRIANO

DATE/TIME OF TEST: 04/11/2025 02:43:38

Diagnostics Check	OK	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.177	02:38
Breath Volume	2.304	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.169	02:41
Breath Volume	1.984	02:41
Air Blank	0.000	02:42
Control Test	0.076	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMACHO,FRANCES
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BURITICA

DATE/TIME OF TEST: 04/12/2025 04:17:37

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.078	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.174	04:13
Breath Volume	1.441	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.166	04:16
Breath Volume	1.648	04:16
Air Blank	0.000	04:16
Control Test	0.078	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORCELY,NIXON,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMACHO,FRANCES
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 04/13/2025 01:08:16

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.129	01:02
Breath Volume	2.082	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.136	01:06
Breath Volume	2.164	01:06
Air Blank	0.000	01:07
Control Test	0.077	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSELL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,AVERY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: RIVERA SANTIAGO

DATE/TIME OF TEST: 04/13/2025 06:34:58

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.079	06:28
Air Blank	0.000	06:29
Subject Sample # 1	0.190	06:30
Breath Volume	2.363	06:30
Air Blank	0.000	06:31
Air Blank	0.000	06:32
Subject Sample # 2	0.179	06:33
Breath Volume	1.980	06:33
Air Blank	0.000	06:34
Control Test	0.078	06:34
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AL-ANBKI,HAEDER,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VAZQUEZ,ELIEZER
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MORALES ARREAGA

DATE/TIME OF TEST: 04/14/2025 01:50:50

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:41
Subject Sample # 1	0.278	01:42
Breath Volume	1.589	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.253	01:46
Breath Volume	2.128	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 3	0.277	01:49
Breath Volume	3.253	01:49
Air Blank	0.000	01:49
Control Test	0.078	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO-PEREZ,MICHAEL,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRUZ-ROSARIO,ANTHONY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: IRRIZARY

DATE/TIME OF TEST: 04/18/2025 03:43:57

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:28	
Subject Sample	# 1 0.088	03:31	
Breath Volume	2.105	03:31	
Air Blank	0.000	03:32	
Air Blank	0.000	03:34	
Subject Sample	# 2 0.085	03:37	Volume Not Met
Breath Volume	0.937	03:37	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 3 0.084	03:43	Volume Not Met
Breath Volume	0.136	03:43	
Air Blank		03:43	RFI Detect
Air Blank	0.000	03:43	

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCANLAN,JAMES,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: RODRIGUZ,SELENA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: MUNIZ-ARILL

DATE/TIME OF TEST: 04/20/2025 02:21:44

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:07	
Control Test	0.079	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1 0.145	02:11	Volume Not Met
Breath Volume	0.140	02:11	
Air Blank	0.000	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 2 0.147	02:15	
Breath Volume	1.351	02:15	
Air Blank	0.000	02:16	
Air Blank	0.000	02:18	
Subject Sample	# 3 0.152	02:19	
Breath Volume	1.300	02:19	
Air Blank	0.000	02:20	
Control Test	0.077	02:21	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,JACOB,S
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ZULUAGA,SANTIAGO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: PEREZ RIVERO

DATE/TIME OF TEST: 04/21/2025 16:39:40

Diagnostics Check	OK	16:31
Air Blank	0.000	16:31
Control Test	0.079	16:32
Air Blank	0.000	16:32
Subject Sample # 1	0.227	16:34
Breath Volume	3.265	16:34
Air Blank	0.000	16:35
Air Blank	0.000	16:37
Subject Sample # 2	0.223	16:38
Breath Volume	2.218	16:38
Air Blank	0.000	16:38
Control Test	0.079	16:39
Air Blank	0.000	16:39
Diagnostics Check	OK	16:39

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,CHARLES,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCONNELL,CHARLES
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:51

SUBJECT'S LAST NAME: LOPEZ URBINA

DATE/TIME OF TEST: 04/21/2025 18:24:27

Diagnostics Check	OK	18:11	
Air Blank	0.000	18:12	
Control Test	0.079	18:12	
Air Blank	0.000	18:13	
Subject Sample	# 1 0.259	18:16	Volume Not Met
Breath Volume	0.113	18:16	
Air Blank	0.000	18:17	
Air Blank	0.000	18:18	
Subject Sample	# 2 0.243	18:19	
Breath Volume	2.757	18:19	
Air Blank	0.000	18:20	
Air Blank	0.000	18:22	
Subject Sample	# 3 0.226	18:22	
Breath Volume	1.953	18:22	
Air Blank	0.000	18:23	
Control Test	0.077	18:23	
Air Blank	0.000	18:24	
Diagnostics Check	OK	18:24	

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AL-ANBKI,HAEDER,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,AVERY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SANTAMARIA DELA CRUZ

DATE/TIME OF TEST: 04/22/2025 03:31:12

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:20	
Control Test	0.079	03:20	
Air Blank	0.000	03:20	
Subject Sample	# 1 0.125	03:23	Volume Not Met
Breath Volume	0.066	03:23	
Air Blank	0.000	03:24	
Air Blank	0.000	03:26	
Subject Sample	# 2 0.143	03:29	Volume Not Met
Breath Volume	0.019	03:29	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AL-ANBKI,HAEDER,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,AVERY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SANTAMARIA DELA CRUZ

DATE/TIME OF TEST: 04/22/2025 04:03:55

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:52	
Control Test	0.078	03:53	
Air Blank	0.000	03:53	
Subject Sample	# 1 0.130	03:56	Volume Not Met
Breath Volume	0.871	03:56	
Air Blank	0.000	03:57	
Air Blank	0.000	03:59	
Subject Sample	# 2 0.137	04:02	Volume Not Met
Breath Volume	0.476	04:02	
Air Blank	0.000	04:03	
Control Test	0.078	04:03	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: NIEVES, CHARLIE
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SANCHEZ OROZCO

DATE/TIME OF TEST: 04/23/2025 04:10:03

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.143	04:05
Breath Volume	2.394	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample # 2	0.140	04:08
Breath Volume	1.562	04:08
Air Blank	0.000	04:09
Control Test	0.076	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:10

Data Download Date/Time: 04/24/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INGERSOLL,NICHOLAS,F
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: INGERSOLL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: HAARS

DATE/TIME OF TEST: 04/08/2025 23:04:55

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.078	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.212	23:00
Breath Volume	1.750	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample # 2	0.215	23:03
Breath Volume	2.394	23:03
Air Blank	0.000	23:04
Control Test	0.078	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 05/12/2025 17:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FINE

DATE/TIME OF TEST: 04/11/2025 22:46:47

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.079	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.034	22:42
Breath Volume	2.226	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:44
Subject Sample # 2	0.027	22:45
Breath Volume	1.507	22:45
Air Blank	0.000	22:45
Control Test	0.078	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 05/12/2025 17:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: LOCHTE

DATE/TIME OF TEST: 04/12/2025 03:19:56

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.077	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.083	03:15
Breath Volume	1.878	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:17
Subject Sample # 2	0.064	03:18
Breath Volume	1.390	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/12/2025 22:24:36

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.079	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.031	22:19
Breath Volume	2.386	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.027	22:23
Breath Volume	2.191	22:23
Air Blank	0.000	22:23
Control Test	0.079	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MOSHE

DATE/TIME OF TEST: 04/13/2025 03:09:28

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.059	03:03
Breath Volume	1.855	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.058	03:08
Breath Volume	2.074	03:08
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: VOYE

DATE/TIME OF TEST: 04/14/2025 01:36:01

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.124	01:31
Breath Volume	4.617	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.122	01:34
Breath Volume	4.656	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:36

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLSEN, IAN, N
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: HERNANDEZ, MANUEL
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: BUHLER

DATE/TIME OF TEST: 04/14/2025 18:21:03

Diagnostics Check	OK	18:14
Air Blank	0.000	18:15
Control Test	0.079	18:15
Air Blank	0.000	18:15
Subject Sample # 1	0.129	18:16
Breath Volume	3.238	18:16
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 2	0.117	18:19
Breath Volume	3.003	18:19
Air Blank	0.000	18:20
Control Test	0.079	18:20
Air Blank	0.000	18:20
Diagnostics Check	OK	18:21

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLBRIGHT,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BARNES,NOAH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: BUEHNER-DULUCA

DATE/TIME OF TEST: 04/23/2025 22:32:17

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.000	22:28
Breath Volume	1.894	22:28
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.000	22:30
Breath Volume	2.265	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CHERIZOL

DATE/TIME OF TEST: 04/26/2025 02:07:05

Diagnostics Check	OK	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Subject Sample # 1	0.155	02:00
Breath Volume	1.535	02:00
Air Blank	0.000	02:01
Air Blank	0.000	02:02
Subject Sample # 2	0.152	02:05
Breath Volume	1.242	02:05
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: PETIT-HOMME,CARLO
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: JUAREZ MORALES

DATE/TIME OF TEST: 04/26/2025 03:51:15

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.079	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1 0.197	03:41	Volume Not Met
Breath Volume	0.156	03:41	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2 0.210	03:45	
Breath Volume	1.261	03:45	
Air Blank	0.000	03:46	
Air Blank	0.000	03:48	
Subject Sample	# 3 0.202	03:49	
Breath Volume	1.835	03:49	
Air Blank	0.000	03:50	
Control Test	0.076	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: STENKAMP

DATE/TIME OF TEST: 04/27/2025 02:30:13

Diagnostics Check	OK	02:24
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.128	02:25
Breath Volume	3.074	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.120	02:28
Breath Volume	2.335	02:28
Air Blank	0.000	02:29
Control Test	0.079	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 04/28/2025 00:42:05

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:32
Air Blank	0.000	00:33
Subject Sample # 1	0.169	00:35
Breath Volume	2.152	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.165	00:40
Breath Volume	2.121	00:40
Air Blank	0.000	00:41
Control Test	0.077	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN, TYRONE, P

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: LOGAN, TYRONE

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 231592

DRY GAS EXPIRY DATE: 11/05/2025

VIOLATION CODE: DUI

ARREST TIME: 15:49

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/01/2025 16:25:09

Diagnostics Check	OK	16:18
Air Blank	0.000	16:19
Control Test	0.081	16:19
Air Blank	0.000	16:19
Subject Sample # 1	0.019	16:20
Breath Volume	3.316	16:20
Air Blank	0.000	16:21
Air Blank	0.000	16:22
Subject Sample # 2	0.018	16:23
Breath Volume	3.003	16:23
Air Blank	0.000	16:24
Control Test	0.081	16:24
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAEUTER,DARIAN,R
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: KRAEUTER,DARIAN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 04/01/2025 18:10:07

Diagnostics Check	OK	18:03
Air Blank	0.000	18:03
Control Test	0.081	18:04
Air Blank	0.000	18:04
Subject Sample # 1	0.000	18:05
Breath Volume	1.738	18:05
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.000	18:08
Breath Volume	2.222	18:08
Air Blank	0.000	18:09
Control Test	0.081	18:09
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAEUTER,DARIAN,R
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: NEGRON,JAVIER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:34

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 04/04/2025 15:02:37

Diagnostics Check	OK	14:55
Air Blank	0.000	14:56
Control Test	0.082	14:56
Air Blank	0.000	14:56
Subject Sample # 1	0.030	14:58
Breath Volume	2.804	14:58
Air Blank	0.000	14:58
Air Blank	0.000	15:00
Subject Sample # 2	0.029	15:01
Breath Volume	2.878	15:01
Air Blank	0.000	15:01
Control Test	0.081	15:02
Air Blank	0.000	15:02
Diagnostics Check	OK	15:02

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,KAITLYN,V
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SILVESTRI,CODY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 04/05/2025 08:16:36

Diagnostics Check	OK	08:10
Air Blank	0.000	08:10
Control Test	0.082	08:10
Air Blank	0.000	08:11
Subject Sample # 1	0.128	08:11
Breath Volume	3.316	08:11
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.126	08:15
Breath Volume	2.890	08:15
Air Blank	0.000	08:15
Control Test	0.081	08:16
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN, TYRONE, P

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: DUMMER, LUKE

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 231592

DRY GAS EXPIRY DATE: 11/05/2025

VIOLATION CODE: DUI

ARREST TIME: 15:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:07

SUBJECT'S LAST NAME: WILDES

DATE/TIME OF TEST: 04/09/2025 15:43:16

Diagnostics Check	OK	15:37
Air Blank	0.000	15:37
Control Test	0.081	15:37
Air Blank	0.000	15:38
Subject Sample # 1	0.000	15:38
Breath Volume	4.468	15:38
Air Blank	0.000	15:39
Air Blank	0.000	15:41
Subject Sample # 2	0.000	15:41
Breath Volume	4.398	15:41
Air Blank	0.000	15:42
Control Test	0.081	15:42
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN, TYRONE, P

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: LOGAN, TYRONE

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 231592

DRY GAS EXPIRY DATE: 11/05/2025

VIOLATION CODE: DUI

ARREST TIME: 15:43

GENDER: Female

OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/12/2025 16:20:18

Diagnostics Check	OK	16:08	
Air Blank	0.000	16:09	
Control Test	0.081	16:09	
Air Blank	0.000	16:09	
Subject Sample	# 1 0.133	16:13	Volume Not Met
Breath Volume	1.238	16:13	
Air Blank	0.000	16:13	
Air Blank	0.000	16:15	
Subject Sample	# 2 0.142	16:18	Volume Not Met
Breath Volume	0.261	16:18	
Air Blank	0.000	16:19	
Control Test	0.080	16:19	
Air Blank	0.000	16:20	
Diagnostics Check	OK	16:20	

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN, TYRONE, P

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: LOGAN, TYRONE

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 231592

DRY GAS EXPIRY DATE: 11/05/2025

VIOLATION CODE: DUI

ARREST TIME: 14:43

GENDER: Female

OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/12/2025 16:59:11

Diagnostics Check	OK	16:46
Air Blank	0.000	16:46
Control Test	0.081	16:46
Air Blank	0.000	16:47
Subject Sample # 1	0.120	16:50
Breath Volume	1.726	16:50
Air Blank	0.000	16:50
Air Blank	0.000	16:52
Subject Sample # 2	0.142	16:54
Breath Volume	2.734	16:54
Air Blank	0.000	16:55
Air Blank	0.000	16:57
Subject Sample # 3	0.132	16:57
Breath Volume	2.355	16:57
Air Blank	0.000	16:58
Control Test	0.080	16:58
Air Blank	0.000	16:59
Diagnostics Check	OK	16:59

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: EDMONSON

DATE/TIME OF TEST: 04/17/2025 00:02:28

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.081	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.122	23:57
Breath Volume	2.199	23:57
Air Blank	0.000	23:58
Air Blank	0.000	23:59
Subject Sample # 1	0.118	00:00
Breath Volume	1.585	00:00
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: NIGH,ALEXANDER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: STRANGE

DATE/TIME OF TEST: 04/19/2025 20:56:32

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:47	
Control Test	0.081	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 1 0.213	20:51	Volume Not Met
Breath Volume	0.113	20:51	
Air Blank	0.000	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 2	20:54	Subject Test Refused
Breath Volume	0.554	20:54	
Air Blank	0.000	20:55	
Control Test	0.078	20:55	
Air Blank	0.000	20:56	
Diagnostics Check	OK	20:56	

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SILVESTRI,CODY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/24/2025 19:20:48

Diagnostics Check	OK	19:13	
Air Blank	0.000	19:14	
Control Test	0.082	19:14	
Air Blank	0.000	19:14	
Subject Sample # 1	0.000	19:15	
Breath Volume	1.253	19:15	
Air Blank	0.000	19:16	
Air Blank	0.000	19:18	
Subject Sample # 2	0.000	19:18	
Breath Volume	2.269	19:18	
Air Blank	0.000	19:19	
Control Test	0.082	19:19	
Air Blank	0.000	19:20	
Diagnostics Check	Fail	19:20	Temperature Regulation Test
Air Blank	0.000	19:20	

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SILVESTRI,CODY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/24/2025 19:31:45

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.081	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.000	19:27
Breath Volume	1.652	19:27
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.000	19:30
Breath Volume	2.542	19:30
Air Blank	0.000	19:30
Control Test	0.082	19:31
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ORTIZ,AXEL
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: PEFFLY

DATE/TIME OF TEST: 04/26/2025 02:10:39

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.081	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.094	02:06
Breath Volume	1.480	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.095	02:09
Breath Volume	1.894	02:09
Air Blank	0.000	02:09
Control Test	0.081	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/27/2025 00:14:33

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:06	
Control Test	0.082	00:06	
Air Blank	0.000	00:07	
Subject Sample # 1	0.098	00:09	
Breath Volume	2.031	00:09	
Air Blank	0.000	00:09	
Air Blank	0.000	00:11	
Subject Sample # 2	0.095	00:12	
Breath Volume	1.246	00:12	
Air Blank	0.000	00:13	
Control Test	0.081	00:13	
Air Blank	0.000	00:13	
Diagnostics Check	Fail	00:14	Temperature Regulation Test
Air Blank	0.000	00:14	

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,JESSE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 04/27/2025 01:28:19

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.000	01:24
Breath Volume	1.296	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.000	01:26
Breath Volume	1.457	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: COTTRELL,CORY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: METOTT

DATE/TIME OF TEST: 04/27/2025 05:07:47

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.081	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.000	05:02
Breath Volume	1.796	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.000	05:06
Breath Volume	2.085	05:06
Air Blank	0.000	05:06
Control Test	0.081	05:07
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SILVESTRI,CODY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 04/30/2025 00:56:53

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.081	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.214	00:52
Breath Volume	3.750	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.213	00:55
Breath Volume	4.011	00:55
Air Blank	0.000	00:55
Control Test	0.081	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 04/30/2025 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: MC GINNIS

DATE/TIME OF TEST: 04/30/2025 23:35:23

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.081	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1 0.194	23:27	
Breath Volume	2.929	23:27	
Air Blank	0.000	23:28	
Air Blank	0.000	23:30	
Subject Sample	# 2 0.179	23:33	
Breath Volume	1.824	23:33	
Air Blank	0.000	23:33	
Control Test	0.081	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	Fail	23:34	Temperature Regulation Test
Air Blank	0.000	23:35	

Data Download Date/Time: 05/28/2025 20:07