Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TALLAHASSEE PD
Time of Inspection: 07:57 Date of Inspection: 06/20/2022 Serial Number: 80-006926
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34520080A1 Exp: 02/05/2023
0.000	0.048	0.078	0.192	0.080
0.000	0.049	0.080	0.196	0.000
0.000	0.050	0.079	0.197	0.000

Number of Simulators Used: 5

Remarks:
08: Control Outside Tolerance. Non-compliance:DRYGASS LOW.

Removed dry gas cylinder, re-seated to regulator
No issues on subsequent inspection JJH TPD 34

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JJH TPD 34 JAMES J HARRIS
Signature and Printed Name

06/20/2022
Date

FDLE/ATP Form 40 – March 2004

X

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TALLAHASSEE PD
Time of Inspection: 07:58 Date of Inspection: 10/27/2022
Serial Number: 80-006926
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) ~~Lot#:3450080A1~~ Exp: 02/05/2023
0.000	0.047	0.076	0.195	0.082
0.000	0.048	0.077	0.196	0.082
0.000	0.048	0.077	0.198	0.083

Number of Simulators Used: 5

Remarks:

lot 345200801 JJH 10/27/22

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JJH TP0321 JAMES J HARRIS

Signature and Printed Name

10/27/2022
Date

FDLE/ATP Form 40 – March 2004