Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ZEPHYRHILLS PD
Time of Inspection: 09:22
Date of Inspection: 01/31/2022
Serial Number: 80-001076
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202005A Exp: 05/12/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13921080A2 Exp: 07/05/2023
0.000	0.049	0.079	0.201	0.081
0.000	0.050	0.080	0.203	0.081
0.000	0.050	0.081	0.203	0.081

Number of Simulators Used: 4

Remarks:

Modem slot broke
& retracted in to
machine

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ailisa J Stanley — AILISA J STANLEY
Signature and Printed Name

01/31/2022
Date

FDLE/ATP Form 40 - March 2004

Amended

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ZEPHYRHILLS PD
Time of Inspection: 09:13 Date of Inspection: 12/28/2022
Serial Number: 80-001076
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13921080A2 Exp: 07/05/2023
0.000	0.048	0.079	0.199	0.077
0.000	0.048	0.079	0.200	0.077
0.000	0.048	0.079	0.200	0.077

Number of Simulators Used: 4

Remarks:
COUPLE AMB*, PUR*, RGE*
Ambient Fails Kept clearing area and checking hose. finally just moved machine.
Purge mouth piece was still attached, removed
Rge - Blew to much air 2 once

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

AILISA J STARNES
Signature and Printed Name

12/28/2022
Date

FDLE/ATP Form 40 – March 2004