

Alcohol Testing Program - Instrument Processing Sheet

Agency: PUTNAM COUNTY SO

Instrument Serial Number: 80-001648

Date In: 8/6/2026

DI Completion Date: 09/01/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>8/6/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>08/11/2026</u>		Flow Adjustment By: <u>SLH</u>																
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>238</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.128</u> (.139-.169) 36 mm <u>0.152</u> (.156-.190) 53 mm <u>0.226</u> (.228-.278) 103 mm <u>0.507</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1012</u> Instrument: <u>1015</u> ☒ Stability Checks		Flow Column #: <u>ATP 102</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value: <u>239</u> ☒ Post Adjustment Verification (L/S) Flow Column #: <u>ATP 105</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.539</u> (.447-.547)																
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202603A 03/03/2028</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B 03/03/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202603D 03/04/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/05/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202603A 03/03/2028	0.080	MP6292	202603B 03/03/2028	0.200	MP6293	202603D 03/04/2028	0.080 DGS	N/A	23825080A1 10/05/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	
Simulator	Serial #	Lot#/Exp																		
0.050	MP6291	202603A 03/03/2028																		
0.080	MP6292	202603B 03/03/2028																		
0.200	MP6293	202603D 03/04/2028																		
0.080 DGS	N/A	23825080A1 10/05/2027																		
Optical Bench Adjustment By: <u>SLH</u>		Department Inspection By: <u>SLH</u>																		
Barometric Pressure Gauge: <u>1012</u> ID#: <u>34420</u>		Barometric Pressure ID#: <u>28427</u>																		
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1012</u> Instrument: <u>1012</u>																
0.000	MP6294	N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u>																
0.040	MP6295	25400	12/03/2027	Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>																
0.100	MP6296	24390	10/29/2026	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>0.200</td> <td>MP5087</td> </tr> <tr> <td>0.300</td> <td>MP5088</td> </tr> <tr> <td>0.080 DGS</td> <td>MP5089</td> </tr> <tr> <td></td> <td>MP5090</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	0.200	MP5087	0.300	MP5088	0.080 DGS	MP5089		MP5090			
Simulator	Serial Number																			
0.000	MP5086																			
0.200	MP5087																			
0.300	MP5088																			
0.080 DGS	MP5089																			
	MP5090																			
0.200	MP6297	25020	01/14/2027																	
0.300	MP6298	24430	12/10/2026																	
0.080 DGS	N/A	AG526603	09/23/2027																	
☒ Post Optical Bench Adjustment Stability Checks																				
Simulator	Serial #	Lot #	Expiration	Attachments																
0.050	MP5088	202603A	03/03/2028	☒ Form 41																
0.080	MP5089	202603B	03/03/2028	☒ Stability Checks																
0.200	MP5090	202603D	03/04/2028	☒ Calibration Certificate																
0.080 DGS	N/A	23725080A1	10/05/2027	☒ Optical Bench Adjustment																
Gauge ID #: <u>28427</u>				☒ Post-Stability Checks																
Gauge: <u>1010</u> Instrument: <u>1011</u>				☒ Flow Adjustment																
Notes/Suggested Service:				☒ Form 40																
				☒ Other: Calculation Form																
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
Digitally signed by Wen-Chi Pierson Date: 2026.09.09 08:42:59 -04'00'				Digitally signed by Kaitlyn Spearin Date: 2026.09.10 10:24:58 -04'00'																
Tech Review				Admin Review																

PUTNAM COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

08/11/2026

Software: 8100.27

SN 80-001648 ✓

SW

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

·SQRT(Diff)) = 7.414

2: Rate (Liters/min) = 15

SQRT(Diff)) = 11.746

3: Rate (Liters/min) = 30

SQRT(Diff)) = 20.441

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 739

Rounded Intercept = -829780

Correlation = 0.99711

Stability Checks

80-00 1648

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.077 to 0.083 g/210L 0.003 g/210L of Wet Sub	DGS 0.080 g/210L																																																																																																																																				
PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2026 Software: 8100.27 SN 80-001648 11:24 08:12 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:14</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:15</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.1000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.1000</td><td></td></tr> </table> Operator's Signature <i>Muzzel</i>	Air Blank	0.000	15:14	Control Test	0.049	15:14	Air Blank	0.000	15:15	Control Test	0.049	15:16	Air Blank	0.000	15:16	Control Test	0.049	15:17	Air Blank	0.000	15:17	Control Test Stats			Average	0.0490		Std Dev	0.1000		Rel Std Dev(%)	0.1000		PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2026 Software: 3100.27 SN 80-001648 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:08</td></tr> <tr><td>Control Test</td><td>0.077</td><td>15:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:09</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:10</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0106</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </table> Operator's Signature <i>Muzzel</i>	Air Blank	0.000	15:08	Control Test	0.077	15:09	Air Blank	0.000	15:09	Control Test	0.078	15:10	Air Blank	0.000	15:10	Control Test	0.078	15:11	Air Blank	0.000	15:11	Control Test Stats			Average	0.0777		Std Dev	0.0106		Rel Std Dev(%)	0.7434		PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/12/2026 Software: 8100.27 SN 80-001648 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.100</td><td>15:18</td></tr> <tr><td>Control Test</td><td>0.197</td><td>15:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:20</td></tr> <tr><td>Control Test</td><td>0.197</td><td>15:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:21</td></tr> <tr><td>Control Test</td><td>0.196</td><td>15:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1967</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2936</td><td></td></tr> </table> Operator's Signature <i>Muzzel</i>	Air Blank	0.100	15:18	Control Test	0.197	15:19	Air Blank	0.000	15:20	Control Test	0.197	15:20	Air Blank	0.000	15:21	Control Test	0.196	15:22	Air Blank	0.000	15:22	Control Test Stats			Average	0.1967		Std Dev	0.0016		Rel Std Dev(%)	0.2936		PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/11/2026 Software: 8100.27 SN 80-001648 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>16:25</td></tr> <tr><td>Control Test</td><td>0.081</td><td>16:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:26</td></tr> <tr><td>Control Test</td><td>0.081</td><td>16:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:26</td></tr> <tr><td>Control Test</td><td>0.081</td><td>16:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature <i>Muzzel</i>	Air Blank	0.000	16:25	Control Test	0.081	16:25	Air Blank	0.000	16:26	Control Test	0.081	16:26	Air Blank	0.000	16:26	Control Test	0.081	16:27	Air Blank	0.000	16:27	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000		*performed 8/11/26 DGS	Verified instrument was able to read gas pressure. ☒
Air Blank	0.000	15:14																																																																																																																																							
Control Test	0.049	15:14																																																																																																																																							
Air Blank	0.000	15:15																																																																																																																																							
Control Test	0.049	15:16																																																																																																																																							
Air Blank	0.000	15:16																																																																																																																																							
Control Test	0.049	15:17																																																																																																																																							
Air Blank	0.000	15:17																																																																																																																																							
Control Test Stats																																																																																																																																									
Average	0.0490																																																																																																																																								
Std Dev	0.1000																																																																																																																																								
Rel Std Dev(%)	0.1000																																																																																																																																								
Air Blank	0.000	15:08																																																																																																																																							
Control Test	0.077	15:09																																																																																																																																							
Air Blank	0.000	15:09																																																																																																																																							
Control Test	0.078	15:10																																																																																																																																							
Air Blank	0.000	15:10																																																																																																																																							
Control Test	0.078	15:11																																																																																																																																							
Air Blank	0.000	15:11																																																																																																																																							
Control Test Stats																																																																																																																																									
Average	0.0777																																																																																																																																								
Std Dev	0.0106																																																																																																																																								
Rel Std Dev(%)	0.7434																																																																																																																																								
Air Blank	0.100	15:18																																																																																																																																							
Control Test	0.197	15:19																																																																																																																																							
Air Blank	0.000	15:20																																																																																																																																							
Control Test	0.197	15:20																																																																																																																																							
Air Blank	0.000	15:21																																																																																																																																							
Control Test	0.196	15:22																																																																																																																																							
Air Blank	0.000	15:22																																																																																																																																							
Control Test Stats																																																																																																																																									
Average	0.1967																																																																																																																																								
Std Dev	0.0016																																																																																																																																								
Rel Std Dev(%)	0.2936																																																																																																																																								
Air Blank	0.000	16:25																																																																																																																																							
Control Test	0.081	16:25																																																																																																																																							
Air Blank	0.000	16:26																																																																																																																																							
Control Test	0.081	16:26																																																																																																																																							
Air Blank	0.000	16:26																																																																																																																																							
Control Test	0.081	16:27																																																																																																																																							
Air Blank	0.000	16:27																																																																																																																																							
Control Test Stats																																																																																																																																									
Average	0.0810																																																																																																																																								
Std Dev	0.0000																																																																																																																																								
Rel Std Dev(%)	0.0000																																																																																																																																								

requires optical Bench adjustment. Sub

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001648
08/18/2026 12:53:41

SUK

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.1390	(-0.0170)	0.1200	(-0.0190)	
Sample #2	0.1120	(0.0310)	0.1190	(-0.0110)	
Sample #3	0.1000	(0.0480)	0.1300	(-0.0070)	
Sample #4	0.1090	(0.0570)	0.1220	(0.0100)	
Avg % Abs	0.1070	(0.0453)	0.1237	(-0.0027)	
STD DEV	0.0062	(0.0132)	0.0057	(0.0112)	
REL STD DEV	5.836	(29.125)	4.598	(418.143)	

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.8310	(0.0000)	1.4900	(-0.0160)	
Sample #2	0.8360	(0.0110)	1.4820	(-0.0110)	
Sample #3	0.8360	(0.0110)	1.4630	(-0.0100)	
Sample #4	0.8430	(0.0240)	1.4860	(-0.0040)	
Avg % Abs	0.8383	(0.0153)	1.4770	(-0.0083)	
STD DEV	0.0040	(0.0075)	0.0123	(0.0038)	
REL STD DEV	0.482	(48.949)	0.832	(45.431)	

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	1.9090	(-0.0150)	3.4480	(-0.0110)	
Sample #2	1.9080	(-0.0010)	3.4740	(-0.0090)	
Sample #3	1.8970	(0.0180)	3.4710	(0.0020)	
Sample #4	1.9030	(0.0280)	3.4590	(0.0180)	
Avg % Abs	1.9027	(0.0150)	3.4680	(0.0037)	
STD DEV	0.0055	(0.0147)	0.0079	(0.0136)	
REL STD DEV	0.289	(98.206)	0.229	(370.280)	

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	3.6770	(-0.0170)	6.6480	(-0.0200)	
Sample #2	3.6660	(-0.0010)	6.6470	(0.0000)	
Sample #3	3.6780	(0.0050)	6.6460	(-0.0030)	
Sample #4	3.6670	(0.0240)	6.6350	(0.0080)	
Avg % Abs	3.6703	(0.0093)	6.6427	(0.0017)	
STD DEV	0.0067	(0.0131)	0.0067	(0.0057)	
REL STD DEV	0.181	(139.834)	0.100	(341.174)	

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	5.3560	(-0.0190)	9.6360	(0.0000)	
Sample #2	5.3720	(-0.0060)	9.6650	(0.0200)	
Sample #3	5.3750	(-0.0010)	9.6590	(0.0160)	
Sample #4	5.3490	(0.0100)	9.6400	(0.0210)	
Avg % Abs	5.3653	(0.0010)	9.6547	(0.0190)	
STD DEV	0.0142	(0.0082)	0.0131	(0.0026)	
REL STD DEV	0.265	(818.535)	0.135	(13.925)	

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001648 *SUB*
08/18/2026 12:53:41

Auto Calibration

pg 2 of 2

<<<< 3um >>>>			<<<< 9um >>>>		
-----			-----		
Zero Order Coef	-276.28			-164.76	
First Order Coef	2593.52			1377.57	
Second Order Coef	22.25			12.31	
-----			-----		
Act	Fit	Residual	Act	Fit	Residual
(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)	(g/210L)
0.000	0.000	-0.0000	0.000	0.000	-0.0001
0.040	0.040	-0.0002	0.040	0.040	0.0002
0.100	0.100	0.0005	0.100	0.100	0.0000
0.200	0.200	-0.0004	0.200	0.200	-0.0001
0.300	0.300	0.0001	0.300	0.300	0.0001
-----			-----		

<<<< 3um >>>>		<<<< 9um >>>>	
-----		-----	
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1			
Sample			
Sample #1	3310.00	3439.00	
Sample #2	3289.00	3445.00	
Sample #3	3269.00	3449.00	
Sample #4	3283.00	3442.00	
Avg	3280.3333	3445.3333	
STD DEV	10.2632	3.5119	
REL STD DEV	0.313	0.102	
H2O adjust (mg/l*10k)	529	364	

Barometric Pressure = 1011

*****CALIBRATION SUCCESSFUL*****

Post Optical Bench Adjustment Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	* DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																			
PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/19/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:47</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:48</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:49</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>1.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.0000</td><td></td></tr> </table> Operator's Signature: <i>Mesial</i>	Air Blank	0.000	15:47	Control Test	0.049	15:47	Air Blank	0.000	15:48	Control Test	0.049	15:49	Air Blank	0.000	15:49	Control Test	0.049	15:50	Air Blank	0.000	15:50	Control Test Stats			Average	0.0490		Std Dev	1.0000		Rel Std Dev(%)	1.0000		PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/19/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:37</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:38</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:40</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature: <i>Mesial</i>	Air Blank	0.000	15:37	Control Test	0.078	15:38	Air Blank	0.000	15:38	Control Test	0.078	15:39	Air Blank	0.000	15:40	Control Test	0.078	15:40	Air Blank	0.000	15:41	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PUTNAM COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/19/2026 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:42</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:43</td></tr> <tr><td>Control Test</td><td>0.197</td><td>15:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:45</td></tr> <tr><td>Control Test</td><td>0.196</td><td>15:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1970</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5076</td><td></td></tr> </table> Operator's Signature: <i>Mesial</i>	Air Blank	0.000	15:42	Control Test	0.198	15:43	Air Blank	0.000	15:43	Control Test	0.197	15:44	Air Blank	0.000	15:45	Control Test	0.196	15:45	Air Blank	0.000	15:46	Control Test Stats			Average	0.1970		Std Dev	0.0010		Rel Std Dev(%)	0.5076		see next page
Air Blank	0.000	15:47																																																																																																				
Control Test	0.049	15:47																																																																																																				
Air Blank	0.000	15:48																																																																																																				
Control Test	0.049	15:49																																																																																																				
Air Blank	0.000	15:49																																																																																																				
Control Test	0.049	15:50																																																																																																				
Air Blank	0.000	15:50																																																																																																				
Control Test Stats																																																																																																						
Average	0.0490																																																																																																					
Std Dev	1.0000																																																																																																					
Rel Std Dev(%)	1.0000																																																																																																					
Air Blank	0.000	15:37																																																																																																				
Control Test	0.078	15:38																																																																																																				
Air Blank	0.000	15:38																																																																																																				
Control Test	0.078	15:39																																																																																																				
Air Blank	0.000	15:40																																																																																																				
Control Test	0.078	15:40																																																																																																				
Air Blank	0.000	15:41																																																																																																				
Control Test Stats																																																																																																						
Average	0.0780																																																																																																					
Std Dev	0.0000																																																																																																					
Rel Std Dev(%)	0.0000																																																																																																					
Air Blank	0.000	15:42																																																																																																				
Control Test	0.198	15:43																																																																																																				
Air Blank	0.000	15:43																																																																																																				
Control Test	0.197	15:44																																																																																																				
Air Blank	0.000	15:45																																																																																																				
Control Test	0.196	15:45																																																																																																				
Air Blank	0.000	15:46																																																																																																				
Control Test Stats																																																																																																						
Average	0.1970																																																																																																					
Std Dev	0.0010																																																																																																					
Rel Std Dev(%)	0.5076																																																																																																					

80-001648
* performed 8/18/2026
SUT

PUTNAM COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001648
08/18/2026
Software: 8100.27

post-stability
DGS

Test	g/210L	Time
Air Blank	0.000	14:00
Control Test	0.079	14:00
Air Blank	0.000	14:01
Control Test	0.079	14:01
Air Blank	0.000	14:02
Control Test	0.080	14:02
Air Blank	0.000	14:02
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PUTNAM COUNTY SO
Time of Inspection: 13:38

Date of Inspection: 09/01/2026

Serial Number: 80-001648
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

09/01/2026
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PUTNAM COUNTY SO
Time of Inspection: 17:06

Date of Inspection: 09/01/2026

Serial Number: 80-001648
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23725080A1 Exp: 10/05/2027
0.000	0.049	0.077	0.196	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.048	0.077	0.196	0.079
0.000	0.048	0.078	0.196	0.079
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.077	0.196	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.050	0.078	0.195	0.080
0.000	0.049	0.078	0.196	0.079

Standard Deviations	0.0005	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM
Signature and Printed Name

09/01/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: SLH

Serial Number: 80-001648 Date: 9/1/2026

Solution:

0.05 g/210 L Test: 0.489 / 10 = 0.049 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.777 / 10 = 0.078 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 1.959 / 10 = 0.196 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.792 / 10 = 0.079 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001648, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001648</u>	Average of $y \pm$ Uncertainty*
Owning Agency:	<u>PUTNAM COUNTY SO</u>	0.049g/ 210 L 0.004
Calibration Date:	<u>09/01/2026</u>	0.078g/ 210 L 0.004
Calibration Time:	<u>17:06</u>	0.196g/ 210 L 0.008
		0.079g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

S. LeAndra
Higginbotham
Digitally signed by S. LeAndra
Higginbotham
Date: 2026.09.04 15:28:16 -04'00'

09/01/2026

Date

LEANDRA HIGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 August 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality