

Alcohol Testing Program - Instrument Processing Sheet

Agency: PANAMA CITY BEACH PD Instrument Serial Number: 80-000838

Date In: 7/23/2026 DI Completion Date: 8/6/2026 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>7/30/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>8/6/2026</u>		Flow Adjustment By: _____																													
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>200</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1017</u> Instrument: <u>1021</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																													
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23725080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202603A 3/3/2028	0.080	MP6292	202603B 3/3/2028	0.200	MP6293	202603D 3/4/2028	0.080 DGS	N/A	23725080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:														
Simulator	Serial #	Lot#/Exp																															
0.050	MP6291	202603A 3/3/2028																															
0.080	MP6292	202603B 3/3/2028																															
0.200	MP6293	202603D 3/4/2028																															
0.080 DGS	N/A	23725080A1 10/5/2027																															
Optical Bench Adjustment By: _____		Department Inspection By: <u>WKP</u>																															
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>28427</u>																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1017</u> Instrument: <u>1021</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>			
Simulator	Serial #	Lot #	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
☐ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A										
Simulator	Serial #	Lot #	Expiration																														
0.050																																	
0.080																																	
0.200																																	
0.080 DGS	N/A																																
Gauge ID #: _____		Gauge: <u>1017</u> Instrument: <u>1021</u>																															
Gauge: _____ Instrument: _____																																	
Notes/Suggested Service: CMI loaded missing forms on instrument. 7/30/2026 Tech Correction: Added signature to Form 40. WKP 8/10/2026		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
		<table style="width: 100%;"> <tr> <td style="width: 50%;"> LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.08.10 10:04:15 -04'00' </td> <td style="width: 50%;"> Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.10 12:54:05 -04'00' </td> </tr> </table>				LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.08.10 10:04:15 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.10 12:54:05 -04'00'																										
LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.08.10 10:04:15 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.10 12:54:05 -04'00'																																
		<table style="width: 100%;"> <tr> <td style="width: 50%;">Tech Review</td> <td style="width: 50%;">Admin Review</td> </tr> </table>				Tech Review	Admin Review																										
Tech Review	Admin Review																																

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY BEACH PD

Time of Inspection: 13:39

Date of Inspection: 07/30/2026

Serial Number: 80-000838

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi Pierson

WEN-CHI K PIERSON

Signature and Printed Name

07/30/2026
Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
PANAMA CITY BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 09/15/2026 SN 80-000838 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:02</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </table> Operator's Signature <i>menpi</i>	Air Blank	0.000	10:02	Control Test	0.050	10:03	Air Blank	0.000	10:03	Control Test	0.050	10:04	Air Blank	0.000	10:04	Control Test	0.049	10:05	Air Blank	0.000	10:05	Control Test Stats			Average	0.0497		Std Dev	0.0005		Rel Std Dev(%)	1.1625		PANAMA CITY BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 08/06/2026 SN 80-000838 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:15</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature <i>menpi</i>	Air Blank	0.000	10:12	Control Test	0.079	10:13	Air Blank	0.000	10:14	Control Test	0.079	10:14	Air Blank	0.000	10:15	Control Test	0.079	10:16	Air Blank	0.000	10:16	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PANAMA CITY BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 08/05/2026 SN 80-000838 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr> <tr><td>Control Test</td><td>0.203</td><td>10:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test</td><td>0.202</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.202</td><td>10:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2023</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2853</td><td></td></tr> </table> Operator's Signature <i>menpi</i>	Air Blank	0.000	10:18	Control Test	0.203	10:19	Air Blank	0.000	10:19	Control Test	0.202	10:20	Air Blank	0.000	10:21	Control Test	0.202	10:21	Air Blank	0.000	10:22	Control Test Stats			Average	0.2023		Std Dev	0.0006		Rel Std Dev(%)	0.2853		PANAMA CITY BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 08/16/2026 SN 80-000838 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:36</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:37</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:38</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:38</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature <i>menpi</i>	Air Blank	0.000	09:36	Control Test	0.080	09:36	Air Blank	0.000	09:37	Control Test	0.080	09:37	Air Blank	0.000	09:38	Control Test	0.080	09:38	Air Blank	0.000	09:38	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	10:02																																																																																																																																					
Control Test	0.050	10:03																																																																																																																																					
Air Blank	0.000	10:03																																																																																																																																					
Control Test	0.050	10:04																																																																																																																																					
Air Blank	0.000	10:04																																																																																																																																					
Control Test	0.049	10:05																																																																																																																																					
Air Blank	0.000	10:05																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0497																																																																																																																																						
Std Dev	0.0005																																																																																																																																						
Rel Std Dev(%)	1.1625																																																																																																																																						
Air Blank	0.000	10:12																																																																																																																																					
Control Test	0.079	10:13																																																																																																																																					
Air Blank	0.000	10:14																																																																																																																																					
Control Test	0.079	10:14																																																																																																																																					
Air Blank	0.000	10:15																																																																																																																																					
Control Test	0.079	10:16																																																																																																																																					
Air Blank	0.000	10:16																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	10:18																																																																																																																																					
Control Test	0.203	10:19																																																																																																																																					
Air Blank	0.000	10:19																																																																																																																																					
Control Test	0.202	10:20																																																																																																																																					
Air Blank	0.000	10:21																																																																																																																																					
Control Test	0.202	10:21																																																																																																																																					
Air Blank	0.000	10:22																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.2023																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2853																																																																																																																																						
Air Blank	0.000	09:36																																																																																																																																					
Control Test	0.080	09:36																																																																																																																																					
Air Blank	0.000	09:37																																																																																																																																					
Control Test	0.080	09:37																																																																																																																																					
Air Blank	0.000	09:38																																																																																																																																					
Control Test	0.080	09:38																																																																																																																																					
Air Blank	0.000	09:38																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY BEACH PD
Time of Inspection: 13:56

Date of Inspection: 08/06/2026

Serial Number: 80-000838
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#: 202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#: 202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 23725080A1 Exp: 10/05/2027
0.000	0.048	0.079	0.201	0.080
0.000	0.048	0.079	0.201	0.080
0.000	0.048	0.079	0.202	0.081
0.000	0.048	0.079	0.202	0.079
0.000	0.048	0.079	0.202	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.048	0.079	0.203	0.080
0.000	0.049	0.079	0.202	0.081
0.000	0.049	0.080	0.202	0.080
0.000	0.049	0.079	0.202	0.080
Standard Deviations	0.0005	0.0003	0.0005	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON

Signature and Printed Name

08/06/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: WKP

Serial Number: 80-000838

Date: 8/6/2026

Solution:

0.05 g/210 L Test: 0.484 / 10 = 0.048 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.791 / 10 = 0.079 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 2.019 / 10 = 0.202 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.801 / 10 = 0.080 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000838, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000838</u>	Average of y ± Uncertainty*
Owning Agency:	<u>PANAMA CITY BEACH PD</u>	0.048 g/ 210 L 0.004
Calibration Date:	<u>08/06/2026</u>	0.079 g/ 210 L 0.004
Calibration Time:	<u>13:56</u>	0.202 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi

Chi Pierson
Date: 2026.08.06
16:06:21 -04'00'

Pierson

08/06/2026

Date

WEN-CHI K PIERSON,

Department Inspector

FDLE/ATP Form 69 August 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Alcohol Testing Program - Instrument Processing Sheet

Agency: Panama City Beach PD

Instrument Serial Number: 80-000838

Date In: 6/3/2026

DI Completion Date: 6/9/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>SLH</u> Date: <u>6/3/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>6/5/2028</u>		Flow Adjustment By: _____																															
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>200</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.148</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.496</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28662</u> Gauge: <u>1016</u> Instrument: <u>1019</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP5088	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP5089	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP5090	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	23825080A1		10/5/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP5088	202603A	☐ Battery Replacement																										
								3/3/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP5089	202603B	☐ Tank Sensor Tare																																
		3/3/2028	☐ Breath Tube Replacement																																
0.200	MP5090	202603D	☐ Other:																																
		3/4/2028																																	
0.080 DGS	N/A	23825080A1																																	
		10/5/2027																																	
Optical Bench Adjustment By: _____		Department Inspection By: <u>WKP</u>																																	
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>34419</u>																																	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1018</u> Instrument: <u>1021</u>																															
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>9/25/2027</u>																														
0.040				Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>9/22/2027</u>																														
0.100				Simulator	Serial Number																														
0.200				0.000	MP5086																														
0.300				Interferent	MP5087																														
0.080 DGS	N/A			0.050	MP5088																														
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP5089																														
Simulator	Serial #	Lot #	Expiration	0.200	MP5090																														
0.050				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:																															
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Gauge ID #: _____ Gauge: _____ Instrument: _____																																			
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																															
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																															
				☒ Conduct an Agency Inspection Before Evidentiary Use																															
LeAndra Higginbotham Tech Review				Kaitlyn Spearin Admin Review																															

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY BEACH PD

Time of Inspection: 12:10

Date of Inspection: 06/05/2026

Serial Number: 80-000838

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON
Signature and Printed Name

06/05/2026
Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
PANAMA CITY BEACH PD Intoxilyzer - Alcotest Analyzer Model: 8000 SN 80-000838 06/05/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:28 Control Test 0.051 12:28 Air Blank 0.000 12:29 Control Test 0.050 12:30 Air Blank 0.000 12:31 Control Test 0.053 12:31 Air Blank 0.000 12:31 Control Test Stats Average 0.0503 Std Dev 0.0006 Rel Std Dev(%) 1.1471 Operator's Signature 	PANAMA CITY BEACH PD Intoxilyzer - Alcotest Analyzer Model: 8000 SN 80-000838 06/05/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:17 Control Test 0.080 12:18 Air Blank 0.000 12:19 Control Test 0.080 12:19 Air Blank 0.000 12:20 Control Test 0.080 12:20 Air Blank 0.000 12:21 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	PANAMA CITY BEACH PD Intoxilyzer - Alcotest Analyzer Model: 8000 SN 80-000838 06/05/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:10 Control Test 0.203 12:11 Air Blank 0.000 12:12 Control Test 0.203 12:12 Air Blank 0.000 12:13 Control Test 0.203 12:13 Air Blank 0.000 12:14 Control Test Stats Average 0.2030 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	PANAMA CITY BEACH PD Intoxilyzer - Alcotest Analyzer Model: 8000 SN 80-000838 06/05/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:24 Control Test 0.081 12:24 Air Blank 0.000 12:24 Control Test 0.080 12:25 Air Blank 0.000 12:25 Control Test 0.081 12:25 Air Blank 0.000 12:26 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157 Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY BEACH PD
Time of Inspection: 15:15

Date of Inspection: 06/09/2026

Serial Number: 80-000838
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.081


Standard Deviations	0.0000	0.0000	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON
Signature and Printed Name

06/09/2026
Date



Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000838, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000838</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PANAMA CITY BEACH PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/09/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:15</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. **Wen-Chi**
This document shall not be reproduced except in full, Digitally signed by Wen-Chi Pierson
without written approval of the Florida Department of Law Enforcement Alcohol Testing Program. Date: 2026.06.10 09:17:43 -04'00'

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Date

06/09/2026

WEN-CHI K PIERSON,

Department Inspector

Service • Integrity • Respect • Quality