

Alcohol Testing Program - Instrument Processing Sheet

Agency: LAFAYETTE COUNTY SO Instrument Serial Number: 80-000773

Date In: 8/19/2026 DI Completion Date: 8/19/2026 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>8/19/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>8/19/2026</u>		Flow Adjustment By: _____																																							
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>144</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.515</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1007</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																							
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6291</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6292</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6293</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>28424080A3</td> <td></td> </tr> <tr> <td>11/5/2026</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6291	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP6292	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP6293	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	28424080A3		11/5/2026									
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																																		
						0.050	MP6291	202603A	☐ Battery Replacement																																		
								3/3/2028	☐ Dry Gas Regulator Replacement																																		
0.080	MP6292	202603B	☐ Tank Sensor Tare																																								
		3/3/2028	☐ Breath Tube Replacement																																								
0.200	MP6293	202603D	☐ Other:																																								
		3/4/2028																																									
0.080 DGS	N/A	28424080A3																																									
		11/5/2026																																									
Optical Bench Adjustment By: _____ Department Inspection By: <u>WKP</u> Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: <u>34420</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1013</u> Instrument: <u>1007</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 50%;">Simulator</th> <th style="width: 50%;">Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000	MP6289																																										
Interferent	MP6290																																										
0.050	MP6291																																										
0.080	MP6292																																										
0.200	MP6293																																										
☐ Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Optical Bench Adjustment ☒ Average Calculations ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																					
Simulator	Serial #	Lot #	Expiration																																								
0.050																																											
0.080																																											
0.200																																											
0.080 DGS	N/A																																										
Gauge ID #: _____ Gauge: _____ Instrument: _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																									
Notes/Suggested Service: Tech Correction: Average Calculation form Issue Date: August 4th, 2026. WKP 9/9/2026		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2026.09.14 14:27:23 -0400 S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.09.17 09:11:22 -0400																																									
		Tech Review Admin Review																																									

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤ 0.003 of Wet ☒																																																																																																																																																												
LABETTE CO SO Intoxilyzer - Alcohol Analyzer Model: 8100 08/19/2026 Software: 8100.27 SN 81-000773 Test: 9/21/26 Time <table border="1"> <thead> <tr> <th>Test</th> <th>9/21/26</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:55</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.049</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.000</td><td></td></tr> </tbody> </table> Operator's Signature: <u>Wen</u>	Test	9/21/26	Time	Air Blank	0.000	10:51	Control Test	0.049	10:52	Air Blank	0.000	10:53	Control Test	0.048	10:53	Air Blank	0.000	10:54	Control Test	0.049	10:54	Air Blank	0.000	10:55	Control Test	0.048	10:55	Control Test Stats			Average	0.049		Std Dev	0.000		Rel Std Dev(%)	0.000		LABETTE CO SO Intoxilyzer - Alcohol Analyzer Model: 8100 08/19/2026 Software: 8100.27 SN 81-000773 Test: 9/21/26 Time <table border="1"> <thead> <tr> <th>Test</th> <th>9/21/26</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test</td><td>0.076</td><td>10:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:59</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.077</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.000</td><td></td></tr> </tbody> </table> Operator's Signature: <u>Wen</u>	Test	9/21/26	Time	Air Blank	0.000	10:56	Control Test	0.076	10:57	Air Blank	0.000	10:58	Control Test	0.077	10:58	Air Blank	0.000	10:59	Control Test	0.077	10:59	Air Blank	0.000	11:00	Control Test	0.077	11:00	Control Test Stats			Average	0.077		Std Dev	0.000		Rel Std Dev(%)	0.000		LABETTE CO SO Intoxilyzer - Alcohol Analyzer Model: 8100 08/19/2026 Software: 8100.27 SN 81-000773 Test: 9/21/26 Time <table border="1"> <thead> <tr> <th>Test</th> <th>9/21/26</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:05</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.196</td><td></td></tr> <tr><td>Std Dev</td><td>0.001</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.587</td><td></td></tr> </tbody> </table> Operator's Signature: <u>Wen</u>	Test	9/21/26	Time	Air Blank	0.000	11:02	Control Test	0.196	11:02	Air Blank	0.000	11:03	Control Test	0.196	11:03	Air Blank	0.000	11:04	Control Test	0.198	11:05	Air Blank	0.000	11:05	Control Test	0.196	11:05	Control Test Stats			Average	0.196		Std Dev	0.001		Rel Std Dev(%)	0.587		LABETTE CO SO Intoxilyzer - Alcohol Analyzer Model: 8100 08/19/2026 Software: 8100.27 SN 81-000773 Test: 9/21/26 Time <table border="1"> <thead> <tr> <th>Test</th> <th>9/21/26</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.080</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.718</td><td></td></tr> </tbody> </table> Operator's Signature: <u>Wen</u>	Test	9/21/26	Time	Air Blank	0.000	10:22	Control Test	0.081	10:22	Air Blank	0.000	10:23	Control Test	0.081	10:23	Air Blank	0.000	10:24	Control Test	0.080	10:24	Air Blank	0.000	10:25	Control Test	0.080	10:25	Control Test Stats			Average	0.080		Std Dev	0.000		Rel Std Dev(%)	0.718	
Test	9/21/26	Time																																																																																																																																																													
Air Blank	0.000	10:51																																																																																																																																																													
Control Test	0.049	10:52																																																																																																																																																													
Air Blank	0.000	10:53																																																																																																																																																													
Control Test	0.048	10:53																																																																																																																																																													
Air Blank	0.000	10:54																																																																																																																																																													
Control Test	0.049	10:54																																																																																																																																																													
Air Blank	0.000	10:55																																																																																																																																																													
Control Test	0.048	10:55																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.049																																																																																																																																																														
Std Dev	0.000																																																																																																																																																														
Rel Std Dev(%)	0.000																																																																																																																																																														
Test	9/21/26	Time																																																																																																																																																													
Air Blank	0.000	10:56																																																																																																																																																													
Control Test	0.076	10:57																																																																																																																																																													
Air Blank	0.000	10:58																																																																																																																																																													
Control Test	0.077	10:58																																																																																																																																																													
Air Blank	0.000	10:59																																																																																																																																																													
Control Test	0.077	10:59																																																																																																																																																													
Air Blank	0.000	11:00																																																																																																																																																													
Control Test	0.077	11:00																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.077																																																																																																																																																														
Std Dev	0.000																																																																																																																																																														
Rel Std Dev(%)	0.000																																																																																																																																																														
Test	9/21/26	Time																																																																																																																																																													
Air Blank	0.000	11:02																																																																																																																																																													
Control Test	0.196	11:02																																																																																																																																																													
Air Blank	0.000	11:03																																																																																																																																																													
Control Test	0.196	11:03																																																																																																																																																													
Air Blank	0.000	11:04																																																																																																																																																													
Control Test	0.198	11:05																																																																																																																																																													
Air Blank	0.000	11:05																																																																																																																																																													
Control Test	0.196	11:05																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.196																																																																																																																																																														
Std Dev	0.001																																																																																																																																																														
Rel Std Dev(%)	0.587																																																																																																																																																														
Test	9/21/26	Time																																																																																																																																																													
Air Blank	0.000	10:22																																																																																																																																																													
Control Test	0.081	10:22																																																																																																																																																													
Air Blank	0.000	10:23																																																																																																																																																													
Control Test	0.081	10:23																																																																																																																																																													
Air Blank	0.000	10:24																																																																																																																																																													
Control Test	0.080	10:24																																																																																																																																																													
Air Blank	0.000	10:25																																																																																																																																																													
Control Test	0.080	10:25																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.080																																																																																																																																																														
Std Dev	0.000																																																																																																																																																														
Rel Std Dev(%)	0.718																																																																																																																																																														

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAFAYETTE CO SO
Time of Inspection: 12:59

Date of Inspection: 08/19/2026

Serial Number: 80-000773
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28424080A3 Exp: 11/05/2026
0.000	0.049	0.076	0.197	0.000 / 0.080
0.000	0.048	0.076	0.197	0.000 / 0.080
0.000	0.049	0.076	0.197	0.000 / 0.080
0.000	0.049	0.077	0.197	0.000 / 0.080
0.000	0.048	0.076	0.197	0.000 / 0.080
0.000	0.049	0.077	0.197	0.000 / 0.080
0.000	0.049	0.076	0.197	0.000 / 0.080
0.000	0.049	0.077	0.197	0.000 / 0.080
0.000	0.049	0.077	0.198	0.000 / 0.080
0.000	0.049	0.077	0.198	0.000 / 0.079

Standard Deviations	0.0004	0.0005	0.0004	0.0000 / 0.0003
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. *inadvertently did not name any gas fully
screwed into the regulator. ensured secure attachment
and then repeated.* WKP 8/19/2026

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi K Pierson

WEN-CHI K PIERSON

Signature and Printed Name

08/19/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: WKP

Serial Number: 80-000773

Date: 8/19/2026

Solution:

0.05 g/210 L Test: 0.488 / 10 = 0.049 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.765 / 10 = 0.077 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 1.972 / 10 = 0.197 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.799 / 10 = 0.080 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000773, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000773</u>	Average of y ± Uncertainty*
Owning Agency:	<u>LAFAYETTE CO SO</u>	0.049 g/ 210 L 0.004
Calibration Date:	<u>08/19/2026</u>	0.077 g/ 210 L 0.004
Calibration Time:	<u>12:59</u>	0.197 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson

Digitally signed by Wen-Chi
Pierson
Date: 2026.08.20 09:22:10
-04'00'

08/19/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 August 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality