Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 10/10/2025 01:19:44

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.081	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.195	01:15
Breath Volume		1.312	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.182	01:18
Breath Volume		1.371	01:18
Air Blank		0.000	01:18
Control Test		0.077	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 10/17/2025 09:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 10/17/2025 22:15:43

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.081	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.065	22:10
Breath Volume		1.269	22:10
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.065	22:13
Breath Volume		1.503	22:13
Air Blank		0.000	22:14
Control Test		0.079	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 11/10/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GARCIA MAYE

DATE/TIME OF TEST: 10/18/2025 04:12:18

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.081	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.138	04:06
Breath Volume		1.625	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:08
Subject Sample	# 2	0.130	04:10
Breath Volume		1.121	04:10
Air Blank		0.000	04:11
Control Test		0.078	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 11/10/2025 13:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: HOUGHTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: JACOBO PRIETO

DATE/TIME OF TEST: 10/20/2025 02:03:45

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:54	
Control Test		0.080	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.157	01:58	Volume Not Met
Breath Volume		0.066	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:00	
Subject Sample	# 2		02:01	Subject Test Refused
Breath Volume		0.531	02:01	
Air Blank		0.000	02:02	
Control Test		0.076	02:03	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 11/10/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: JEFFERSON

DATE/TIME OF TEST: 10/20/2025 02:37:51

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.079	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1	0.173	02:29	Volume Not Met
Breath Volume		0.378	02:29	
Air Blank		0.000	02:30	
Air Blank		0.000	02:32	
Subject Sample	# 2	0.228	02:32	
Breath Volume		1.710	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 3	0.219	02:36	
Breath Volume		1.160	02:36	
Air Blank		0.000	02:36	
Control Test		0.076	02:37	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 11/10/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CARIDDI

DATE/TIME OF TEST: 10/24/2025 01:08:20

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.188	01:00
Breath Volume		1.257	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.167	01:02
Breath Volume		1.300	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 3	0.160	01:06
Breath Volume		1.414	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 11/10/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 10/25/2025 01:08:27

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.081	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.205	01:03
Breath Volume		1.500	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.190	01:06
Breath Volume		1.371	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 11/10/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ALLEGRUCCI

DATE/TIME OF TEST: 10/26/2025 18:03:26

Diagnostics Check		OK	17:52	
Air Blank		0.000	17:52	
Control Test		0.081	17:52	
Air Blank		0.000	17:53	
Subject Sample	# 1		17:56	No Sample Provided
Breath Volume		0.000	17:56	
Air Blank		0.000	17:56	
Air Blank		0.000	17:58	
Subject Sample	# 2	0.000	18:02	Volume Not Met
Breath Volume		0.000	18:02	
Air Blank		0.000	18:02	
Control Test		0.081	18:02	
Air Blank		0.000	18:03	
Diagnostics Check		OK	18:03	

Data Download Date/Time: 11/10/2025 13:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ALLEGRUCCI

DATE/TIME OF TEST: 10/26/2025 18:18:51

Diagnostics Check		OK	18:07	
Air Blank		0.000	18:07	
Control Test		0.081	18:07	
Air Blank		0.000	18:08	
Subject Sample	# 1	0.153	18:11	Volume Not Met
Breath Volume		0.152	18:11	
Air Blank		0.000	18:12	
Air Blank		0.000	18:13	
Subject Sample	# 2	0.118	18:17	Volume Not Met
Breath Volume		0.203	18:17	
Air Blank		0.000	18:17	
Control Test		0.077	18:18	
Air Blank		0.000	18:18	
Diagnostics Check		OK	18:18	

Data Download Date/Time: 11/10/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA-MEJIA,JAVIER,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MORENO-BARRIOS

DATE/TIME OF TEST: 10/30/2025 00:09:40

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.081	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1	0.116	00:01	Volume Not Met
Breath Volume		0.585	00:01	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2	0.130	00:04	
Breath Volume		1.343	00:04	
Air Blank		0.000	00:05	
Air Blank		0.000	00:07	
Subject Sample	# 3	0.131	00:07	
Breath Volume		1.437	00:07	
Air Blank		0.000	00:08	
Control Test		0.078	00:09	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 11/10/2025 13:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOGLUND,WALKER,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BLAKE,PAIGE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: GRUPI

DATE/TIME OF TEST: 10/01/2025 00:40:18

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.000	00:35
Breath Volume		2.398	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.000	00:38
Breath Volume		2.718	00:38
Air Blank		0.000	00:39
Control Test		0.081	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 10/28/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINSON,RYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WILKINSON,RYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: LOHMAN

DATE/TIME OF TEST: 10/02/2025 01:27:35

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:14	
Control Test		0.080	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 1	0.206	01:18	Volume Not Met
Breath Volume		0.406	01:18	
Air Blank		0.000	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 2	0.202	01:21	
Breath Volume		1.789	01:21	
Air Blank		0.000	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 3	0.198	01:25	
Breath Volume		1.808	01:25	
Air Blank		0.000	01:26	
Control Test		0.076	01:26	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 10/28/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/04/2025 04:42:39

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.081	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.149	04:37
Breath Volume		1.535	04:37
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.142	04:41
Breath Volume		1.742	04:41
Air Blank		0.000	04:41
Control Test		0.080	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 10/28/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,SCOTTIE,A
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SNYDER,MICHAEL
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/19/2025 18:49:27

Diagnostics Check		OK	18:43
Air Blank		0.000	18:43
Control Test		0.081	18:44
Air Blank		0.000	18:44
Subject Sample	# 1	0.000	18:45
Breath Volume		1.507	18:45
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.000	18:48
Breath Volume		1.261	18:48
Air Blank		0.000	18:48
Control Test		0.081	18:48
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 10/28/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,TYLER,S
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MEDINA,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: DIMUCCIO

DATE/TIME OF TEST: 10/19/2025 23:46:06

Diagnostics Check		OK	23:39
Air Blank		0.000	23:39
Control Test		0.081	23:40
Air Blank		0.000	23:40
Subject Sample	# 1	0.226	23:41
Breath Volume		3.460	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:43
Subject Sample	# 2	0.217	23:44
Breath Volume		2.593	23:44
Air Blank		0.000	23:45
Control Test		0.080	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 10/28/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERICKSON,NATHAN,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ANDERSON,MICHAEL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 10/21/2025 03:42:58

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.081	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.241	03:38
Breath Volume		2.128	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.232	03:41
Breath Volume		1.367	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 10/28/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 10/23/2025 19:21:29

Diagnostics Check	OK	19:20	
Air Blank	0.000	19:20	
Control Test	0.081	19:20	
Air Blank		19:21	RFI Detect
Air Blank	0.000	19:21	

Data Download Date/Time: 10/28/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 10/23/2025 19:31:05

Diagnostics Check		OK	19:24
Air Blank		0.000	19:24
Control Test		0.081	19:25
Air Blank		0.000	19:25
Subject Sample	# 1	0.207	19:26
Breath Volume		2.238	19:26
Air Blank		0.000	19:27
Air Blank		0.000	19:28
Subject Sample	# 2	0.211	19:29
Breath Volume		2.410	19:29
Air Blank		0.000	19:30
Control Test		0.080	19:30
Air Blank		0.000	19:31
Diagnostics Check		OK	19:31

Data Download Date/Time: 10/28/2025 16:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REXFORD,MARK,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BLAKE,PAIGE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 10/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: DOMINICK

DATE/TIME OF TEST: 10/31/2025 19:11:20

Diagnostics Check		OK	19:04
Air Blank		0.000	19:04
Control Test		0.080	19:05
Air Blank		0.000	19:05
Subject Sample	# 1	0.103	19:06
Breath Volume		1.566	19:06
Air Blank		0.000	19:06
Air Blank		0.000	19:08
Subject Sample	# 2	0.106	19:09
Breath Volume		2.964	19:09
Air Blank		0.000	19:10
Control Test		0.080	19:10
Air Blank		0.000	19:11
Diagnostics Check		OK	19:11

Data Download Date/Time: 11/14/2025 09:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMUNNI,FRANCESCO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RAMUNNI,FRANCESCO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/01/2025 04:32:42

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:30	
Control Test		0.078	04:30	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Control Test		0.079	04:32	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,JONATHAN,G
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: MEYER,TREVOR
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SEBASTIAN PABLO

DATE/TIME OF TEST: 10/08/2025 02:03:53

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.079	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.185	01:59
Breath Volume		1.765	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:01
Subject Sample	# 2	0.182	02:02
Breath Volume		2.136	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASAS,JAVIER
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VLACHOS

DATE/TIME OF TEST: 10/10/2025 02:43:15

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.078	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMUNNI,FRANCESCO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: TEMLAK

DATE/TIME OF TEST: 10/11/2025 21:26:18

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:23	
Control Test		0.078	21:24	
Air Blank		0.000	21:24	
Subject Sample	# 1		21:24	Subject Test Refused
Breath Volume		0.000	21:24	
Air Blank		0.000	21:25	
Control Test		0.079	21:25	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,ANDREW,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CRAWFORD,A
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: BORISOV

DATE/TIME OF TEST: 10/13/2025 11:31:49

Diagnostics Check		OK	11:24
Air Blank		0.000	11:24
Control Test		0.078	11:25
Air Blank		0.000	11:25
Subject Sample	# 1	0.197	11:26
Breath Volume		2.535	11:26
Air Blank		0.000	11:27
Air Blank		0.000	11:28
Subject Sample	# 2	0.181	11:30
Breath Volume		1.742	11:30
Air Blank		0.000	11:30
Control Test		0.079	11:31
Air Blank		0.000	11:31
Diagnostics Check		OK	11:31

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADOTTA,LIONEL,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAVVIDIS,GEORGIOS
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: BENZENBERG

DATE/TIME OF TEST: 10/16/2025 21:21:54

Diagnostics Check		OK	21:09
Air Blank		0.000	21:10
Control Test		0.078	21:10
Air Blank		0.000	21:11
Subject Sample	# 1	0.199	21:13
Breath Volume		1.109	21:13
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.169	21:16
Breath Volume		1.808	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:18
Subject Sample	# 3	0.169	21:20
Breath Volume		1.351	21:20
Air Blank		0.000	21:20
Control Test		0.078	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:17

SUBJECT'S LAST NAME: HEVERT

DATE/TIME OF TEST: 10/17/2025 13:08:48

Diagnostics Check		OK	13:05	
Air Blank		0.000	13:06	
Control Test		0.079	13:06	
Air Blank		0.000	13:07	
Subject Sample	# 1		13:07	Subject Test Refused
Breath Volume		0.000	13:07	
Air Blank		0.000	13:07	
Control Test		0.079	13:08	
Air Blank		0.000	13:08	
Diagnostics Check		OK	13:08	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAVVIDIS,GEORGIOS
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: ESTABA AGUILERA

DATE/TIME OF TEST: 10/17/2025 23:40:30

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:38	
Control Test		0.078	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.078	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: MEYER,TREVOR
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: TADLA

DATE/TIME OF TEST: 10/18/2025 01:55:57

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.078	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1	0.190	01:51	Volume Not Met
Breath Volume		0.250	01:51	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 2		01:54	Subject Test Refused
Breath Volume		0.000	01:54	
Air Blank		0.000	01:55	
Control Test		0.076	01:55	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: DIEUVEILLE

DATE/TIME OF TEST: 10/20/2025 02:31:00

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.078	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:30	
Control Test		0.078	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMUNNI,FRANCESCO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: KERSHOW,RICHARD
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 10/22/2025 01:19:43

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.078	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.151	01:14
Breath Volume		2.855	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.146	01:17
Breath Volume		2.921	01:17
Air Blank		0.000	01:18
Control Test		0.077	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: KERSHOW,RICHARD
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: PAVAO

DATE/TIME OF TEST: 10/23/2025 02:11:13

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.078	02:09	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Control Test		0.079	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: KERSHOW,RICHARD
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: CAVASINI

DATE/TIME OF TEST: 10/24/2025 00:22:47

Diagnostics Check		OK	00:19	
Air Blank		0.000	00:20	
Control Test		0.078	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:21	Subject Test Refused
Breath Volume		0.000	00:21	
Air Blank		0.000	00:21	
Control Test		0.079	00:22	
Air Blank		0.000	00:22	
Diagnostics Check		OK	00:22	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SHIVER,JOEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: GOSS

DATE/TIME OF TEST: 10/26/2025 23:31:12

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.078	23:29	
Air Blank		0.000	23:29	
Subject Sample	# 1		23:29	Subject Test Refused
Breath Volume		0.000	23:29	
Air Blank		0.000	23:30	
Control Test		0.079	23:30	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:31	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMUNNI,FRANCESCO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RAMUNNI,FRANCESCO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: VALERA

DATE/TIME OF TEST: 10/27/2025 05:37:08

Diagnostics Check		OK	05:23	
Air Blank		0.000	05:23	
Control Test		0.079	05:23	
Air Blank		0.000	05:24	
Subject Sample	# 1		05:27	No Sample Provided
Breath Volume		0.000	05:27	
Air Blank		0.000	05:28	
Air Blank		0.000	05:30	
Subject Sample	# 2	0.252	05:30	
Breath Volume		3.289	05:30	
Air Blank		0.000	05:32	
Air Blank		0.000	05:33	
Subject Sample	# 3	0.252	05:35	
Breath Volume		2.917	05:35	
Air Blank		0.000	05:36	
Control Test		0.078	05:36	
Air Blank		0.000	05:37	
Diagnostics Check		OK	05:37	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMUNNI,FRANCESCO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: TORO,MASON
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 10/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ORLANDO

DATE/TIME OF TEST: 10/29/2025 03:28:32

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.079	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.079	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 11/25/2025 09:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:16

SUBJECT'S LAST NAME: FAVARD

DATE/TIME OF TEST: 10/03/2025 08:46:58

Diagnostics Check		OK	08:37
Air Blank		0.000	08:38
Control Test		0.081	08:38
Air Blank		0.000	08:39
Subject Sample	# 1	0.111	08:41
Breath Volume		1.250	08:41
Air Blank		0.000	08:41
Air Blank		0.000	08:43
Subject Sample	# 2	0.120	08:45
Breath Volume		2.269	08:45
Air Blank		0.000	08:46
Control Test		0.079	08:46
Air Blank		0.000	08:46
Diagnostics Check		OK	08:46

Data Download Date/Time: 10/22/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICARDO,EDWARD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: PATINO SUAREZ

DATE/TIME OF TEST: 10/06/2025 01:42:35

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.213	01:37
Breath Volume		1.105	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.217	01:40
Breath Volume		1.425	01:40
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 10/22/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 10/11/2025 04:07:59

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:05	
Control Test		0.080	04:05	
Air Blank		0.000	04:06	
Subject Sample	# 1		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:07	
Control Test		0.081	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 10/22/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: PENA ROMERO

DATE/TIME OF TEST: 10/13/2025 01:07:48

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1	0.155	00:55	
Breath Volume		2.636	00:55	
Air Blank		0.000	00:55	
Air Blank		0.000	00:57	
Subject Sample	# 2		01:00	No Sample Provided
Breath Volume		0.000	01:00	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 3		01:06	No Sample Provided
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Control Test		0.080	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 10/22/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PENA ROMERO

DATE/TIME OF TEST: 10/13/2025 01:36:59

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.145	01:32
Breath Volume		2.531	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.132	01:35
Breath Volume		1.199	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 10/22/2025 19:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEMORE,DESTINEY,S
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: LOPEZ SANCHEZ

DATE/TIME OF TEST: 10/17/2025 20:43:10

Diagnostics Check	OK	20:42	
Air Blank		20:42	RFI Detect
Air Blank		20:43	RFI Detect

Data Download Date/Time: 10/22/2025 19:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEMORE,DESTINEY,S
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: LOPEZ SANCHEZ

DATE/TIME OF TEST: 10/17/2025 20:45:30

Diagnostics Check	OK	20:44	
Air Blank		20:45	RFI Detect
Air Blank	0.000	20:45	

Data Download Date/Time: 10/22/2025 19:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEMORE,DESTINEY,S
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: LOPEZ SANCHEZ

DATE/TIME OF TEST: 10/17/2025 20:49:30

Diagnostics Check		OK	20:47	
Air Blank		0.000	20:47	
Control Test		0.080	20:48	
Air Blank		0.000	20:48	
Subject Sample	# 1		20:49	RFI Detect
Breath Volume		0.000	20:49	
Air Blank			20:49	RFI Detect
Air Blank			20:49	RFI Detect

Data Download Date/Time: 10/22/2025 19:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: LOPEZ SANCHEZ

DATE/TIME OF TEST: 10/17/2025 21:00:52

Diagnostics Check		OK	20:52
Air Blank		0.000	20:53
Control Test		0.081	20:53
Air Blank		0.000	20:54
Subject Sample	# 1	0.221	20:56
Breath Volume		2.441	20:56
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.211	20:59
Breath Volume		2.480	20:59
Air Blank		0.000	20:59
Control Test		0.079	21:00
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 10/22/2025 19:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: ALMONTE

DATE/TIME OF TEST: 10/18/2025 05:27:28

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.081	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.141	05:22
Breath Volume		1.167	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.136	05:25
Breath Volume		1.218	05:25
Air Blank		0.000	05:26
Control Test		0.080	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 10/22/2025 19:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ESCOBAR CORDERO

DATE/TIME OF TEST: 10/19/2025 01:54:29

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.081	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.119	01:49
Breath Volume		1.539	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:51
Subject Sample	# 2	0.118	01:52
Breath Volume		2.617	01:52
Air Blank		0.000	01:53
Control Test		0.081	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 10/22/2025 19:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/25/2025 01:17:38

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.081	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.241	01:13
Breath Volume		1.621	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.242	01:16
Breath Volume		2.554	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 11/25/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: JARQUIN MARADIAGA

DATE/TIME OF TEST: 10/25/2025 05:58:26

Diagnostics Check		OK	05:44
Air Blank		0.000	05:45
Control Test		0.080	05:45
Air Blank		0.000	05:45
Subject Sample	# 1	0.188	05:48
Breath Volume		1.300	05:48
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 2	0.153	05:54
Breath Volume		2.117	05:54
Air Blank		0.000	05:54
Air Blank		0.000	05:56
Subject Sample	# 3	0.162	05:56
Breath Volume		1.562	05:56
Air Blank		0.000	05:57
Control Test		0.078	05:57
Air Blank		0.000	05:58
Diagnostics Check		OK	05:58

Data Download Date/Time: 11/25/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MUZEAU

DATE/TIME OF TEST: 10/26/2025 04:07:38

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.081	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.188	04:03
Breath Volume		2.386	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.184	04:06
Breath Volume		1.429	04:06
Air Blank		0.000	04:06
Control Test		0.078	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 11/25/2025 14:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,VIJAY,N
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCKENZIE,VIJAY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/28/2025 02:38:22

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.075	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.154	02:33
Breath Volume		1.839	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.161	02:36
Breath Volume		2.625	02:36
Air Blank		0.000	02:37
Control Test		0.075	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 11/25/2025 14:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 10/02/2025 21:55:03

Diagnostics Check		OK	21:48
Air Blank		0.000	21:49
Control Test		0.079	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.171	21:50
Breath Volume		2.324	21:50
Air Blank		0.000	21:51
Air Blank		0.000	21:52
Subject Sample	# 2	0.163	21:53
Breath Volume		2.171	21:53
Air Blank		0.000	21:54
Control Test		0.079	21:54
Air Blank		0.000	21:54
Diagnostics Check		OK	21:55

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/05/2025 02:55:56

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.217	02:51
Breath Volume		1.742	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:53
Subject Sample	# 2	0.223	02:54
Breath Volume		1.906	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: AWAD

DATE/TIME OF TEST: 10/12/2025 00:30:51

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.080	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.093	00:26
Breath Volume		3.320	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.095	00:29
Breath Volume		4.007	00:29
Air Blank		0.000	00:29
Control Test		0.080	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: STREETER J

DATE/TIME OF TEST: 10/12/2025 04:26:41

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.037	04:21
Breath Volume		2.394	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.030	04:25
Breath Volume		1.218	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLAHAN,TRACI,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HASS,MALIK
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: HALLAHAN

DATE/TIME OF TEST: 10/03/2025 23:33:47

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:18	
Control Test		0.080	23:18	
Air Blank		0.000	23:19	
Subject Sample	# 1	0.160	23:22	Volume Not Met
Breath Volume		0.406	23:22	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 2	0.160	23:26	
Breath Volume		2.191	23:26	
Air Blank		0.000	23:27	
Air Blank		0.000	23:29	
Subject Sample	# 3	0.153	23:32	Volume Not Met
Breath Volume		0.164	23:32	
Air Blank			23:33	Purge Fail
Air Blank		0.000	23:33	

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MORA ANGULO

DATE/TIME OF TEST: 10/04/2025 02:43:14

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.168	02:37
Breath Volume		1.281	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.175	02:41
Breath Volume		1.812	02:41
Air Blank		0.000	02:42
Control Test		0.075	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SALGADO

DATE/TIME OF TEST: 10/05/2025 02:31:31

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.067	02:26
Breath Volume		2.144	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.063	02:29
Breath Volume		2.125	02:29
Air Blank		0.000	02:30
Control Test		0.078	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MARTINEZ ESCOBAR

DATE/TIME OF TEST: 10/05/2025 03:31:11

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.166	03:25
Breath Volume		1.441	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.157	03:29
Breath Volume		1.515	03:29
Air Blank		0.000	03:30
Control Test		0.077	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PETRUSAK

DATE/TIME OF TEST: 10/05/2025 04:32:33

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.038	04:27
Breath Volume		4.003	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.036	04:30
Breath Volume		4.359	04:30
Air Blank		0.000	04:31
Control Test		0.077	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: PICAUT

DATE/TIME OF TEST: 10/05/2025 18:20:43

Diagnostics Check		OK	18:14
Air Blank		0.000	18:14
Control Test		0.078	18:14
Air Blank		0.000	18:15
Subject Sample	# 1	0.172	18:15
Breath Volume		4.199	18:15
Air Blank		0.000	18:16
Air Blank		0.000	18:18
Subject Sample	# 2	0.170	18:19
Breath Volume		4.546	18:19
Air Blank		0.000	18:19
Control Test		0.077	18:20
Air Blank		0.000	18:20
Diagnostics Check		OK	18:20

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATT,DEREK,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PLATT,DEREK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:29

SUBJECT'S LAST NAME: MARTINEZHERNANDEZ

DATE/TIME OF TEST: 10/05/2025 18:55:47

Diagnostics Check		OK	18:49
Air Blank		0.000	18:49
Control Test		0.079	18:50
Air Blank		0.000	18:50
Subject Sample	# 1	0.177	18:51
Breath Volume		1.433	18:51
Air Blank		0.000	18:51
Air Blank		0.000	18:53
Subject Sample	# 2	0.185	18:54
Breath Volume		2.511	18:54
Air Blank		0.000	18:54
Control Test		0.077	18:55
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 10/06/2025 02:26:49

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:11	
Control Test		0.079	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1	0.072	02:14	
Breath Volume		1.164	02:14	
Air Blank		0.000	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 2	0.056	02:20	Volume Not Met
Breath Volume		0.390	02:20	
Air Blank		0.000	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 3	0.056	02:25	
Breath Volume		1.199	02:25	
Air Blank		0.000	02:25	
Control Test		0.075	02:26	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SCHMITT,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: GRAY JR

DATE/TIME OF TEST: 10/06/2025 11:40:29

Diagnostics Check		OK	11:33
Air Blank		0.000	11:34
Control Test		0.079	11:34
Air Blank		0.000	11:35
Subject Sample	# 1	0.063	11:35
Breath Volume		1.734	11:35
Air Blank		0.000	11:36
Air Blank		0.000	11:38
Subject Sample	# 2	0.069	11:38
Breath Volume		1.191	11:38
Air Blank		0.000	11:39
Control Test		0.078	11:39
Air Blank		0.000	11:40
Diagnostics Check		OK	11:40

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: JIMENEZ ROMERO

DATE/TIME OF TEST: 10/07/2025 01:27:02

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.000	01:22
Breath Volume		2.433	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.000	01:25
Breath Volume		2.171	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:27

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: EADES,EMILY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: GLANZ

DATE/TIME OF TEST: 10/07/2025 11:04:20

Diagnostics Check		OK	10:57
Air Blank		0.000	10:58
Control Test		0.080	10:58
Air Blank		0.000	10:59
Subject Sample	# 1	0.000	10:59
Breath Volume		4.257	10:59
Air Blank		0.000	11:00
Air Blank		0.000	11:02
Subject Sample	# 2	0.000	11:02
Breath Volume		2.847	11:02
Air Blank		0.000	11:03
Control Test		0.079	11:03
Air Blank		0.000	11:04
Diagnostics Check		OK	11:04

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY,QUINN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: MORIARTY,QUINN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: LINDNAL

DATE/TIME OF TEST: 10/10/2025 02:41:09

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.251	02:36
Breath Volume		3.472	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:38
Subject Sample	# 2	0.253	02:39
Breath Volume		4.265	02:39
Air Blank		0.000	02:40
Control Test		0.078	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: SOUTH

DATE/TIME OF TEST: 10/10/2025 22:13:22

Diagnostics Check		OK	22:06
Air Blank		0.000	22:07
Control Test		0.079	22:07
Air Blank		0.000	22:08
Subject Sample	# 1	0.252	22:08
Breath Volume		3.066	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:10
Subject Sample	# 2	0.245	22:11
Breath Volume		3.253	22:11
Air Blank		0.000	22:12
Control Test		0.077	22:12
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BONNET,GREGORY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 10/10/2025 22:31:31

Diagnostics Check		OK	22:24
Air Blank		0.000	22:25
Control Test		0.078	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.130	22:26
Breath Volume		2.488	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.125	22:29
Breath Volume		2.117	22:29
Air Blank		0.000	22:30
Control Test		0.076	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MERIDA-TEMAJ

DATE/TIME OF TEST: 10/10/2025 23:25:20

Diagnostics Check		OK	23:18
Air Blank		0.000	23:19
Control Test		0.079	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.223	23:20
Breath Volume		2.070	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.211	23:23
Breath Volume		2.347	23:23
Air Blank		0.000	23:24
Control Test		0.077	23:24
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: BARAJAS MURILLO

DATE/TIME OF TEST: 10/11/2025 18:39:32

Diagnostics Check		OK	18:30
Air Blank		0.000	18:31
Control Test		0.079	18:31
Air Blank		0.000	18:31
Subject Sample	# 1	0.261	18:32
Breath Volume		1.800	18:32
Air Blank		0.000	18:33
Air Blank		0.000	18:34
Subject Sample	# 2	0.246	18:37
Breath Volume		1.808	18:37
Air Blank		0.000	18:38
Control Test		0.077	18:38
Air Blank		0.000	18:39
Diagnostics Check		OK	18:39

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SARAH,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BROWN,SARAH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: RAYNER

DATE/TIME OF TEST: 10/11/2025 20:18:19

Diagnostics Check		OK	20:11
Air Blank		0.000	20:12
Control Test		0.079	20:12
Air Blank		0.000	20:13
Subject Sample	# 1	0.162	20:13
Breath Volume		2.226	20:13
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.155	20:16
Breath Volume		1.730	20:16
Air Blank		0.000	20:17
Control Test		0.078	20:17
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERYCKEN,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: HAMEL

DATE/TIME OF TEST: 10/11/2025 21:35:13

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.176	21:29
Breath Volume		1.683	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.179	21:33
Breath Volume		1.660	21:33
Air Blank		0.000	21:34
Control Test		0.078	21:34
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: TROVATO

DATE/TIME OF TEST: 10/12/2025 00:53:39

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.079	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.173	00:47
Breath Volume		2.300	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.154	00:51
Breath Volume		1.488	00:51
Air Blank		0.000	00:52
Control Test		0.079	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: XILOJ PELICO

DATE/TIME OF TEST: 10/12/2025 18:00:29

Diagnostics Check		OK	17:48	
Air Blank		0.000	17:49	
Control Test		0.079	17:49	
Air Blank		0.000	17:50	
Subject Sample	# 1		17:53	No Sample Provided
Breath Volume		0.000	17:53	
Air Blank		0.000	17:54	
Air Blank		0.000	17:55	
Subject Sample	# 2		17:58	No Sample Provided
Breath Volume		0.000	17:58	
Air Blank		0.000	17:59	
Control Test		0.066	17:59	Control Outside Tolerance
Air Blank			18:00	Purge Fail

Data Download Date/Time: 10/17/2025 14:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JOEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GONZALEZ,JOEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 10/13/2025 01:09:51

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:03	
Control Test		0.079	01:03	
Air Blank		0.000	01:04	
Subject Sample	# 1		01:07	No Sample Provided
Breath Volume		0.000	01:07	
Air Blank			01:08	Improper Sample
Air Blank		0.000	01:08	
Control Test		0.076	01:09	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: VILLEGAS DE LA PENA

DATE/TIME OF TEST: 10/13/2025 01:45:14

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.184	01:40
Breath Volume		1.941	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.182	01:43
Breath Volume		1.675	01:43
Air Blank		0.000	01:44
Control Test		0.075	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/13/2025 04:25:11

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.157	04:18
Breath Volume		2.050	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.158	04:23
Breath Volume		1.562	04:23
Air Blank		0.000	04:24
Control Test		0.077	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: YAX AJTZOC

DATE/TIME OF TEST: 10/14/2025 01:53:17

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.134	01:48
Breath Volume		3.640	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 2	0.129	01:51
Breath Volume		3.394	01:51
Air Blank		0.000	01:52
Control Test		0.077	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/14/2025 03:56:44

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:41	
Control Test		0.079	03:42	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:45	No Sample Provided
Breath Volume		0.000	03:45	
Air Blank		0.000	03:46	
Air Blank		0.000	03:48	
Subject Sample	# 2	0.198	03:49	
Breath Volume		2.027	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 3		03:54	Subject Test Refused
Breath Volume		0.632	03:54	
Air Blank		0.000	03:55	
Control Test		0.075	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/14/2025 21:58:34

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.079	21:52
Air Blank		0.000	21:52
Subject Sample	# 1	0.230	21:53
Breath Volume		2.753	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.236	21:56
Breath Volume		2.601	21:56
Air Blank		0.000	21:57
Control Test		0.078	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SHUMATE,TODD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 08:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:37

SUBJECT'S LAST NAME: SHILTS

DATE/TIME OF TEST: 10/15/2025 10:12:42

Diagnostics Check		OK	10:06
Air Blank		0.000	10:06
Control Test		0.079	10:07
Air Blank		0.000	10:07
Subject Sample	# 1	0.121	10:08
Breath Volume		1.257	10:08
Air Blank		0.000	10:08
Air Blank		0.000	10:10
Subject Sample	# 2	0.129	10:11
Breath Volume		1.300	10:11
Air Blank		0.000	10:11
Control Test		0.078	10:12
Air Blank		0.000	10:12
Diagnostics Check		OK	10:12

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SARAH,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BROWN,SARAH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VOLLMER

DATE/TIME OF TEST: 10/16/2025 01:44:00

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.080	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.134	01:39
Breath Volume		2.300	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.130	01:42
Breath Volume		2.199	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:44

Data Download Date/Time: 10/17/2025 14:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: CHINAULT

DATE/TIME OF TEST: 10/25/2025 02:08:53

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.122	02:04
Breath Volume		2.867	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.126	02:07
Breath Volume		2.886	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 11/10/2025 09:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135738
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: WETTELAND

DATE/TIME OF TEST: 10/03/2025 02:39:56

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.105	02:34
Breath Volume		2.015	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:36
Subject Sample	# 2	0.117	02:38
Breath Volume		1.722	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 10/15/2025 18:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135738
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: JEAN LOUIS

DATE/TIME OF TEST: 10/11/2025 02:34:36

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.180	02:29
Breath Volume		3.144	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.176	02:32
Breath Volume		3.515	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 10/15/2025 18:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135738
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HARDISON

DATE/TIME OF TEST: 10/17/2025 02:51:29

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.102	02:46
Breath Volume		2.585	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.099	02:49
Breath Volume		3.023	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 11/19/2025 10:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135738
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 10/25/2025 01:10:18

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.081	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.282	01:02
Breath Volume		2.957	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.253	01:05
Breath Volume		1.425	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 3	0.284	01:08
Breath Volume		3.132	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 11/19/2025 10:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: URTUBIA ZURITA

DATE/TIME OF TEST: 10/01/2025 00:58:23

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.000	00:54
Breath Volume		1.707	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.000	00:57
Breath Volume		1.535	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CORTES,CHRISTIAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/03/2025 22:12:33

Diagnostics Check		OK	22:09	
Air Blank		0.000	22:10	
Control Test		0.080	22:10	
Air Blank		0.000	22:10	
Subject Sample	# 1		22:11	Subject Test Refused
Breath Volume		0.000	22:11	
Air Blank		0.000	22:11	
Control Test		0.079	22:11	
Air Blank		0.000	22:12	
Diagnostics Check		OK	22:12	

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,OSCAR,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ESPINOZA,OSCAR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: RUDOLPH

DATE/TIME OF TEST: 10/03/2025 22:32:51

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.080	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.000	22:28
Breath Volume		1.156	22:28
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.000	22:31
Breath Volume		1.515	22:31
Air Blank		0.000	22:31
Control Test		0.077	22:32
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 10/04/2025 00:52:07

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.079	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:51	
Control Test		0.079	00:51	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GONZALEZ CARVELO

DATE/TIME OF TEST: 10/04/2025 22:23:11

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.079	22:16
Air Blank		0.000	22:17
Subject Sample	# 1	0.164	22:18
Breath Volume		2.105	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.167	22:21
Breath Volume		1.683	22:21
Air Blank		0.000	22:22
Control Test		0.079	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRADY,JOSHUA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/07/2027
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 10/04/2025 22:34:09

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:32	
Control Test	0.078	22:32	
Air Blank		22:33	Purge Fail
Air Blank		22:34	Purge Fail

Data Download Date/Time: 10/17/2025 14:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRADY,JOSHUA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/07/2027
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 10/04/2025 22:38:17

Diagnostics Check	OK	22:37	
Air Blank		22:37	Ambient Fail
Air Blank	0.000	22:38	

Data Download Date/Time: 10/17/2025 14:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRADY,JOSHUA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 10/04/2025 22:52:21

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.079	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.079	22:47
Breath Volume		1.113	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.076	22:50
Breath Volume		1.273	22:50
Air Blank		0.000	22:51
Control Test		0.078	22:51
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 10/17/2025 14:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/05/2025 02:15:03

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:12	
Control Test		0.079	02:12	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:13	Subject Test Refused
Breath Volume		0.000	02:13	
Air Blank		0.000	02:14	
Control Test		0.080	02:14	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 10/17/2025 14:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,OSCAR,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARIN,MARIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 10/05/2025 20:34:47

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:32	
Control Test		0.078	20:32	
Air Blank		0.000	20:33	
Subject Sample	# 1		20:33	Subject Test Refused
Breath Volume		0.000	20:33	
Air Blank		0.000	20:33	
Control Test		0.079	20:34	
Air Blank		0.000	20:34	
Diagnostics Check		OK	20:34	

Data Download Date/Time: 10/17/2025 14:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HERNANDEZ,MATTHEW
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 10/12/2025 05:33:50

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:25	
Control Test		0.079	05:25	
Air Blank		0.000	05:26	
Subject Sample	# 1		05:29	No Sample Provided
Breath Volume		0.000	05:29	
Air Blank		0.000	05:30	
Air Blank		0.000	05:32	
Subject Sample	# 2		05:32	Subject Test Refused
Breath Volume		0.000	05:32	
Air Blank		0.000	05:32	
Control Test		0.079	05:33	
Air Blank		0.000	05:33	
Diagnostics Check		OK	05:33	

Data Download Date/Time: 10/17/2025 14:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINOSO,PEDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: REINOSO,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 09:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:24

SUBJECT'S LAST NAME: REIS

DATE/TIME OF TEST: 10/12/2025 09:59:35

Diagnostics Check		OK	09:52
Air Blank		0.000	09:53
Control Test		0.080	09:53
Air Blank		0.000	09:53
Subject Sample	# 1	0.060	09:54
Breath Volume		1.921	09:54
Air Blank		0.000	09:55
Air Blank		0.000	09:56
Subject Sample	# 2	0.055	09:57
Breath Volume		1.535	09:57
Air Blank		0.000	09:58
Control Test		0.078	09:58
Air Blank		0.000	09:59
Diagnostics Check		OK	09:59

Data Download Date/Time: 10/17/2025 14:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTYRE,DELLVON,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MCINTYRE,DELLVON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CAMINAS

DATE/TIME OF TEST: 10/13/2025 00:49:44

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.078	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:45	No Sample Provided
Breath Volume		0.000	00:45	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2		00:48	Subject Test Refused
Breath Volume		0.000	00:48	
Air Blank		0.000	00:48	
Control Test		0.079	00:49	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 10/17/2025 14:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: NADLER

DATE/TIME OF TEST: 10/13/2025 22:35:45

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.079	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.077	22:31
Breath Volume		1.472	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.075	22:34
Breath Volume		1.855	22:34
Air Blank		0.000	22:34
Control Test		0.078	22:35
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 10/17/2025 14:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,VICTOR,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MORALES,VICTOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 10/02/2027
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 10/16/2025 17:24:52

Diagnostics Check		OK	17:18
Air Blank		0.000	17:19
Control Test		0.079	17:19
Air Blank		0.000	17:19
Subject Sample	# 1	0.136	17:20
Breath Volume		1.750	17:20
Air Blank		0.000	17:21
Air Blank		0.000	17:22
Subject Sample	# 2	0.133	17:23
Breath Volume		1.558	17:23
Air Blank		0.000	17:23
Control Test		0.079	17:24
Air Blank		0.000	17:24
Diagnostics Check		OK	17:24

Data Download Date/Time: 10/17/2025 14:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/17/2025 00:29:35

Diagnostics Check		OK	00:16	
Air Blank		0.000	00:16	
Control Test		0.079	00:17	
Air Blank		0.000	00:17	
Subject Sample	# 1	0.000	00:19	
Breath Volume		1.269	00:19	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 2		00:25	No Sample Provided
Breath Volume		0.000	00:25	
Air Blank		0.000	00:25	
Air Blank		0.000	00:27	
Subject Sample	# 3		00:28	Subject Test Refused
Breath Volume		0.000	00:28	
Air Blank		0.000	00:28	
Control Test		0.079	00:28	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 10/17/2025 14:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: IBARRA

DATE/TIME OF TEST: 10/18/2025 01:04:34

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.152	00:59
Breath Volume		4.187	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:01
Subject Sample	# 2	0.149	01:03
Breath Volume		3.984	01:03
Air Blank		0.000	01:03
Control Test		0.078	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 11/17/2025 15:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PECOR

DATE/TIME OF TEST: 10/18/2025 01:58:10

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.079	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1		01:56	Subject Test Refused
Breath Volume		0.000	01:56	
Air Blank		0.000	01:57	
Control Test		0.079	01:57	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 11/17/2025 15:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTYRE,DELLVON,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MOLINEUX

DATE/TIME OF TEST: 10/20/2025 23:43:49

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:41	
Control Test		0.078	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:42	Subject Test Refused
Breath Volume		0.000	23:42	
Air Blank		0.000	23:42	
Control Test		0.078	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 11/17/2025 15:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:47

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/22/2025 10:16:23

Diagnostics Check		OK	10:07
Air Blank		0.000	10:07
Control Test		0.079	10:07
Air Blank		0.000	10:08
Subject Sample	# 1	0.256	10:08
Breath Volume		1.957	10:08
Air Blank		0.000	10:09
Air Blank		0.000	10:11
Subject Sample	# 2	0.235	10:11
Breath Volume		1.742	10:11
Air Blank		0.000	10:12
Air Blank		0.000	10:14
Subject Sample	# 3	0.242	10:14
Breath Volume		1.691	10:14
Air Blank		0.000	10:15
Control Test		0.079	10:15
Air Blank		0.000	10:16
Diagnostics Check		OK	10:16

Data Download Date/Time: 11/17/2025 15:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORTES,CHRISTIAN,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARTINEZ,JESSICA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: NEWELL

DATE/TIME OF TEST: 10/23/2025 02:08:55

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.079	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:08	
Control Test		0.079	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 11/17/2025 15:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARIN,MARIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 10/23/2025 22:13:29

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:10	
Control Test		0.079	22:11	
Air Blank		0.000	22:11	
Subject Sample	# 1		22:12	Subject Test Refused
Breath Volume		0.000	22:12	
Air Blank		0.000	22:12	
Control Test		0.079	22:12	
Air Blank		0.000	22:13	
Diagnostics Check		OK	22:13	

Data Download Date/Time: 11/17/2025 15:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: ATWOOD

DATE/TIME OF TEST: 10/25/2025 22:51:42

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.079	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.210	22:47
Breath Volume		2.750	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.215	22:50
Breath Volume		2.656	22:50
Air Blank		0.000	22:50
Control Test		0.078	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 11/17/2025 15:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: DAWES

DATE/TIME OF TEST: 10/18/2025 02:30:53

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.078	02:29	
Air Blank		02:30	Purge Fail
Air Blank	0.000	02:30	

Data Download Date/Time: 11/10/2025 13:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 10/19/2025 23:22:21

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.077	23:20	
Air Blank		23:21	Purge Fail
Air Blank	0.000	23:22	

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 10/19/2025 23:26:44

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:25	
Control Test	0.071	23:26	Control Outside Tolerance
Air Blank	0.000	23:26	

Data Download Date/Time: 11/10/2025 13:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 10/19/2025 23:36:24

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.199	23:31
Breath Volume		2.500	23:31
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.180	23:34
Breath Volume		1.308	23:34
Air Blank		0.000	23:35
Control Test		0.076	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: JOHNS,JUSTIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MARCOZ

DATE/TIME OF TEST: 10/21/2025 00:12:40

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.009	00:06
Breath Volume		3.625	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.000	00:11
Breath Volume		3.902	00:11
Air Blank		0.000	00:11
Control Test		0.075	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WEAVER,SEAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/24/2025 04:34:59

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.000	04:30
Breath Volume		1.210	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.000	04:33
Breath Volume		1.183	04:33
Air Blank		0.000	04:34
Control Test		0.077	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402419647
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: SULLENS

DATE/TIME OF TEST: 10/24/2025 04:48:27

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.079	04:42
Air Blank		0.000	04:42
Subject Sample	# 1	0.174	04:43
Breath Volume		3.824	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.169	04:46
Breath Volume		2.062	04:46
Air Blank		0.000	04:47
Control Test		0.078	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SARAH,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BROWN,SARAH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 10/24/2025 21:31:37

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.080	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.148	21:26
Breath Volume		2.828	21:26
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.149	21:29
Breath Volume		2.894	21:29
Air Blank		0.000	21:30
Control Test		0.079	21:30
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: JOHNS,JUSTIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HILLS

DATE/TIME OF TEST: 10/25/2025 03:28:02

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.079	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:27	
Control Test		0.080	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASS,MALIK,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HASS,MALIK
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/25/2025 04:09:30

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.164	04:03
Breath Volume		2.128	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.174	04:07
Breath Volume		1.269	04:07
Air Blank		0.000	04:08
Control Test		0.076	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SPEARS

DATE/TIME OF TEST: 10/26/2025 01:24:05

Diagnostics Check		OK	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.161	01:19
Breath Volume		1.175	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.153	01:22
Breath Volume		1.945	01:22
Air Blank		0.000	01:23
Control Test		0.078	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: JOHNS,JUSTIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MORALES RAMIREZ

DATE/TIME OF TEST: 10/26/2025 02:58:21

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.078	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.182	02:53
Breath Volume		2.394	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.187	02:56
Breath Volume		1.500	02:56
Air Blank		0.000	02:57
Control Test		0.076	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: BONNET,GREG
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 10/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: CAESAR

DATE/TIME OF TEST: 10/27/2025 00:31:54

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.203	00:27
Breath Volume		1.750	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.198	00:30
Breath Volume		1.617	00:30
Air Blank		0.000	00:30
Control Test		0.077	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 11/10/2025 13:54

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,LEON,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: JOHNSON,LEON
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/01/2025 01:03:50

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.082	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.218	00:59
Breath Volume		2.550	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:01
Subject Sample	# 2	0.211	01:02
Breath Volume		2.207	01:02
Air Blank		0.000	01:02
Control Test		0.079	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICARDO,EDWARD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: ESPONDA

DATE/TIME OF TEST: 10/03/2025 04:51:08

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:39	
Control Test		0.082	04:39	
Air Blank		0.000	04:40	
Subject Sample	# 1	0.111	04:43	Volume Not Met
Breath Volume		0.542	04:43	
Air Blank		0.000	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 2	0.127	04:46	
Breath Volume		1.882	04:46	
Air Blank		0.000	04:47	
Air Blank		0.000	04:48	
Subject Sample	# 3	0.118	04:49	
Breath Volume		1.324	04:49	
Air Blank		0.000	04:50	
Control Test		0.077	04:50	
Air Blank		0.000	04:51	
Diagnostics Check		OK	04:51	

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: VINDEL DIAZ

DATE/TIME OF TEST: 10/04/2025 06:38:08

Diagnostics Check		OK	06:30
Air Blank		0.000	06:31
Control Test		0.081	06:31
Air Blank		0.000	06:31
Subject Sample	# 1	0.163	06:33
Breath Volume		3.222	06:33
Air Blank		0.000	06:34
Air Blank		0.000	06:35
Subject Sample	# 2	0.158	06:36
Breath Volume		2.398	06:36
Air Blank		0.000	06:37
Control Test		0.078	06:37
Air Blank		0.000	06:38
Diagnostics Check		OK	06:38

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FERNANDEZ,MICHEAL
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/04/2025 08:53:35

Diagnostics Check		OK	08:47
Air Blank		0.000	08:47
Control Test		0.082	08:47
Air Blank		0.000	08:48
Subject Sample	# 1	0.163	08:48
Breath Volume		1.738	08:48
Air Blank		0.000	08:49
Air Blank		0.000	08:51
Subject Sample	# 2	0.157	08:51
Breath Volume		1.234	08:51
Air Blank		0.000	08:52
Control Test		0.081	08:52
Air Blank		0.000	08:53
Diagnostics Check		OK	08:53

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEPESTRE,GREGORY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DEPESTRE,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 14:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: CYPRESS

DATE/TIME OF TEST: 10/04/2025 15:33:56

Diagnostics Check		OK	15:31	
Air Blank		0.000	15:31	
Control Test		0.082	15:31	
Air Blank		0.000	15:32	
Subject Sample	# 1		15:32	Subject Test Refused
Breath Volume		0.000	15:32	
Air Blank		0.000	15:33	
Control Test		0.082	15:33	
Air Blank		0.000	15:33	
Diagnostics Check		OK	15:33	

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: TAMAYO FERNANDEZ

DATE/TIME OF TEST: 10/06/2025 00:04:43

Diagnostics Check		OK	23:57
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.129	00:00
Breath Volume		1.847	00:00
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 2	0.127	00:03
Breath Volume		1.433	00:03
Air Blank		0.000	00:03
Control Test		0.080	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANIO,PIERRE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RODRIGUEZ,ALEXEIS
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 10:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: FERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 10/06/2025 11:46:02

Diagnostics Check		OK	11:43	
Air Blank		0.000	11:43	
Control Test		0.082	11:43	
Air Blank		0.000	11:44	
Subject Sample	# 1		11:44	Subject Test Refused
Breath Volume		0.000	11:44	
Air Blank		0.000	11:45	
Control Test		0.083	11:45	
Air Blank		0.000	11:45	
Diagnostics Check		OK	11:46	

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,MICAH,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LITTLE,MICAH
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: BORGES

DATE/TIME OF TEST: 10/11/2025 20:13:24

Diagnostics Check		OK	20:03
Air Blank		0.000	20:03
Control Test		0.082	20:04
Air Blank		0.000	20:04
Subject Sample	# 1	0.158	20:05
Breath Volume		2.257	20:05
Air Blank		0.000	20:06
Air Blank		0.000	20:07
Subject Sample	# 2	0.137	20:08
Breath Volume		1.628	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 3	0.154	20:11
Breath Volume		2.546	20:11
Air Blank		0.000	20:12
Control Test		0.079	20:12
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,KAREN,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FERNANDEZ,KAREN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: OLIVERO GONZALEZ

DATE/TIME OF TEST: 10/16/2025 03:07:01

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.082	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.102	03:02
Breath Volume		1.644	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.096	03:05
Breath Volume		1.378	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:07

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLEMORE,DESTINEY,S
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: LOPEZ SANCHEZ

DATE/TIME OF TEST: 10/17/2025 20:38:44

Diagnostics Check	OK	20:38	
Air Blank		20:38	RFI Detect
Air Blank		20:38	RFI Detect

Data Download Date/Time: 10/22/2025 19:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: PEREZ SANCHEZ

DATE/TIME OF TEST: 10/20/2025 16:19:28

Diagnostics Check		OK	16:11
Air Blank		0.000	16:12
Control Test		0.082	16:12
Air Blank		0.000	16:13
Subject Sample	# 1	0.238	16:14
Breath Volume		1.914	16:14
Air Blank		0.000	16:15
Air Blank		0.000	16:17
Subject Sample	# 2	0.244	16:17
Breath Volume		1.910	16:17
Air Blank		0.000	16:18
Control Test		0.079	16:18
Air Blank		0.000	16:19
Diagnostics Check		OK	16:19

Data Download Date/Time: 10/22/2025 19:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/26/2025 03:23:23

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.081	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.117	03:18
Breath Volume		2.398	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.109	03:21
Breath Volume		2.511	03:21
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 11/25/2025 14:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 10/31/2025 03:40:20

Diagnostics Check		OK	03:33
Air Blank		0.000	03:34
Control Test		0.082	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.170	03:35
Breath Volume		3.304	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.172	03:38
Breath Volume		1.132	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 11/25/2025 14:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JONES,DALLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: GUILFORD

DATE/TIME OF TEST: 10/03/2025 23:56:18

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:52
Breath Volume		1.777	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:54
Breath Volume		1.250	23:54
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 11/07/2025 07:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: DIAZ RAMIREZ

DATE/TIME OF TEST: 10/05/2025 03:26:10

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:20	
Control Test		0.080	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 1	0.172	03:24	
Breath Volume		1.277	03:24	
Air Blank			03:25	Purge Fail
Air Blank			03:26	Purge Fail

Data Download Date/Time: 11/07/2025 07:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,DALLAS,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEWIS,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: DIAZ RAMIREZ

DATE/TIME OF TEST: 10/05/2025 03:36:25

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.191	03:31
Breath Volume		2.367	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.191	03:34
Breath Volume		2.539	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 11/07/2025 07:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 10/25/2025 00:35:50

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.081	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.180	00:31
Breath Volume		2.292	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.170	00:34
Breath Volume		2.304	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 11/07/2025 07:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: REAVES

DATE/TIME OF TEST: 10/25/2025 01:22:20

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.081	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.268	01:16
Breath Volume		1.214	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:18
Subject Sample	# 2	0.260	01:20
Breath Volume		1.320	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 11/07/2025 07:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: SISTRUNK

DATE/TIME OF TEST: 10/25/2025 18:36:35

Diagnostics Check		OK	18:26
Air Blank		0.000	18:26
Control Test		0.080	18:26
Air Blank		0.000	18:27
Subject Sample	# 1	0.166	18:30
Breath Volume		2.648	18:30
Air Blank		0.000	18:30
Air Blank		0.000	18:32
Subject Sample	# 2	0.155	18:34
Breath Volume		1.800	18:34
Air Blank		0.000	18:35
Control Test		0.079	18:35
Air Blank		0.000	18:36
Diagnostics Check		OK	18:36

Data Download Date/Time: 11/07/2025 07:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JENSON,MATTHEW
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 10/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:01

SUBJECT'S LAST NAME: GEISER

DATE/TIME OF TEST: 10/30/2025 14:30:28

Diagnostics Check		OK	14:24
Air Blank		0.000	14:25
Control Test		0.081	14:25
Air Blank		0.000	14:25
Subject Sample	# 1	0.000	14:26
Breath Volume		2.414	14:26
Air Blank		0.000	14:26
Air Blank		0.000	14:28
Subject Sample	# 2	0.000	14:29
Breath Volume		2.578	14:29
Air Blank		0.000	14:29
Control Test		0.080	14:29
Air Blank		0.000	14:30
Diagnostics Check		OK	14:30

Data Download Date/Time: 11/07/2025 07:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: RESENDIZ-PALACIOS

DATE/TIME OF TEST: 10/02/2025 02:12:06

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.110	02:07
Breath Volume		1.761	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:09
Subject Sample	# 2	0.107	02:10
Breath Volume		1.355	02:10
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:12

Data Download Date/Time: 10/17/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/10/2025 03:17:15

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.088	03:11
Breath Volume		2.679	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.082	03:15
Breath Volume		1.238	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 10/17/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 10/03/2025 21:32:31

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:27	
Control Test		0.079	21:27	
Air Blank		0.000	21:28	
Subject Sample	# 1	0.124	21:28	
Breath Volume		1.500	21:28	
Air Blank		0.000	21:29	
Air Blank			21:31	RFI Detect
Air Blank		0.000	21:31	
Control Test		0.082	21:31	
Air Blank		0.000	21:32	
Diagnostics Check		OK	21:32	

Data Download Date/Time: 10/22/2025 17:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 10/03/2025 21:40:08

Diagnostics Check		OK	21:33	
Air Blank		0.000	21:34	
Control Test		0.082	21:34	
Air Blank		0.000	21:35	
Subject Sample	# 1	0.114	21:35	
Breath Volume		2.199	21:35	
Air Blank		0.000	21:36	
Air Blank		0.000	21:38	
Subject Sample	# 2		21:38	Interferent Detect
Breath Volume		2.175	21:38	
Air Blank		0.000	21:39	
Control Test		0.079	21:39	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 10/22/2025 17:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 10/03/2025 21:48:07

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:42	
Control Test		0.079	21:42	
Air Blank		0.000	21:43	
Subject Sample	# 1	0.111	21:43	
Breath Volume		2.167	21:43	
Air Blank		0.000	21:44	
Air Blank		0.000	21:46	
Subject Sample	# 2		21:46	Interferent Detect
Breath Volume		2.117	21:46	
Air Blank		0.000	21:47	
Control Test		0.077	21:47	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 10/22/2025 17:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 10/03/2025 21:55:54

Diagnostics Check		OK	21:49
Air Blank		0.000	21:50
Control Test		0.080	21:50
Air Blank		0.000	21:51
Subject Sample	# 1	0.110	21:51
Breath Volume		1.777	21:51
Air Blank		0.000	21:52
Air Blank		0.000	21:53
Subject Sample	# 2	0.108	21:54
Breath Volume		1.820	21:54
Air Blank		0.000	21:54
Control Test		0.079	21:55
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 10/22/2025 17:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/04/2025 03:49:12

Diagnostics Check		OK	03:39	
Air Blank		0.000	03:40	
Control Test		0.077	03:40	
Air Blank		0.000	03:40	
Subject Sample	# 1	0.150	03:43	
Breath Volume		1.273	03:43	
Air Blank		0.000	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 2	0.156	03:46	
Breath Volume		1.914	03:46	
Air Blank		0.000	03:47	
Control Test		0.085	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		Fail	03:48	Analytical Stability Test
Air Blank		0.000	03:49	

Data Download Date/Time: 10/22/2025 17:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/04/2025 03:54:07

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.079	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:52	Interferent Detect
Breath Volume		1.632	03:52	
Air Blank		0.000	03:53	
Control Test		0.070	03:53	Interferent Detect
Air Blank		0.000	03:54	

Data Download Date/Time: 10/22/2025 17:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/04/2025 03:57:03

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.072	03:56	Control Outside Tolerance
Air Blank	0.000	03:57	

Data Download Date/Time: 10/22/2025 17:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 10/15/2025 23:43:07

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:36	
Control Test		0.077	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 1	0.143	23:37	
Breath Volume		1.382	23:37	
Air Blank		0.000	23:38	
Air Blank		0.000	23:40	
Subject Sample	# 2	0.139	23:41	
Breath Volume		2.261	23:41	
Air Blank		0.000	23:42	
Control Test		0.078	23:42	
Air Blank			23:42	RFI Detect
Air Blank		0.000	23:43	

Data Download Date/Time: 10/22/2025 17:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 10/15/2025 23:51:08

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.081	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.147	23:46
Breath Volume		3.015	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.142	23:49
Breath Volume		3.007	23:49
Air Blank		0.000	23:50
Control Test		0.077	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 10/22/2025 17:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BRAS

DATE/TIME OF TEST: 10/17/2025 00:31:14

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.078	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.184	00:26
Breath Volume		2.488	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:28
Subject Sample	# 2	0.179	00:29
Breath Volume		1.703	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 10/22/2025 17:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: PARIS

DATE/TIME OF TEST: 10/18/2025 02:21:30

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.243	02:16
Breath Volume		3.593	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.245	02:19
Breath Volume		3.027	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 10/22/2025 17:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 10/20/2025 00:18:13

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.080	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.245	00:13
Breath Volume		3.621	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:15
Subject Sample	# 2	0.244	00:16
Breath Volume		4.027	00:16
Air Blank		0.000	00:17
Control Test		0.078	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 10/22/2025 17:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,CORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SAPP,CORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/23/2025 22:18:37

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.078	22:13
Air Blank		0.000	22:13
Subject Sample	# 1	0.187	22:14
Breath Volume		1.972	22:14
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.185	22:16
Breath Volume		1.882	22:16
Air Blank		0.000	22:17
Control Test		0.078	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 11/20/2025 19:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: CHAMPION

DATE/TIME OF TEST: 10/28/2025 17:21:01

Diagnostics Check	OK	17:20	
Air Blank		17:20	Ambient Fail
Air Blank		17:21	Ambient Fail

Data Download Date/Time: 11/20/2025 19:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: CHAMPION

DATE/TIME OF TEST: 10/28/2025 17:25:09

Diagnostics Check		OK	17:22	
Air Blank		0.000	17:22	
Control Test		0.077	17:23	
Air Blank		0.000	17:23	
Subject Sample	# 1		17:24	Slope Not Met
Breath Volume		1.562	17:24	
Air Blank			17:24	Ambient Fail
Air Blank		0.000	17:25	

Data Download Date/Time: 11/20/2025 19:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: CHAMPION

DATE/TIME OF TEST: 10/28/2025 17:28:17

Diagnostics Check	OK	17:26	
Air Blank	0.000	17:27	
Control Test	0.088	17:27	Control Outside Tolerance
Air Blank		17:28	Ambient Fail

Data Download Date/Time: 11/20/2025 19:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/24/2025 02:03:57

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.060	01:59
Breath Volume		1.378	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.063	02:02
Breath Volume		1.585	02:02
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 11/13/2025 03:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GARCIA,ALEXIS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ACOSTA RAMOS

DATE/TIME OF TEST: 10/26/2025 04:57:51

Diagnostics Check		OK	04:54	
Air Blank		0.000	04:55	
Control Test		0.080	04:55	
Air Blank		0.000	04:56	
Subject Sample	# 1		04:56	Subject Test Refused
Breath Volume		0.000	04:56	
Air Blank		0.000	04:56	
Control Test		0.080	04:57	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 11/13/2025 03:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/02/2025 00:10:13

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.116	00:05
Breath Volume		1.699	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.109	00:08
Breath Volume		1.609	00:08
Air Blank		0.000	00:09
Control Test		0.078	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GRAZIOSE,MADISON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MCFARLANE

DATE/TIME OF TEST: 10/03/2025 03:29:59

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.044	03:25
Breath Volume		1.992	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.044	03:28
Breath Volume		1.558	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GILL,RAMON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 10/03/2025 04:52:21

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.079	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.082	04:46
Breath Volume		2.015	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:48
Subject Sample	# 2	0.074	04:50
Breath Volume		1.367	04:50
Air Blank		0.000	04:51
Control Test		0.077	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROSE,JEROME
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: TORRENEGARA BARCELO

DATE/TIME OF TEST: 10/04/2025 19:52:55

Diagnostics Check		OK	19:46
Air Blank		0.000	19:46
Control Test		0.079	19:47
Air Blank		0.000	19:47
Subject Sample	# 1	0.255	19:48
Breath Volume		2.585	19:48
Air Blank		0.000	19:49
Air Blank		0.000	19:50
Subject Sample	# 2	0.236	19:51
Breath Volume		2.347	19:51
Air Blank		0.000	19:51
Control Test		0.077	19:52
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SOUZA ROCHA

DATE/TIME OF TEST: 10/05/2025 03:37:04

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.078	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.212	03:32
Breath Volume		1.246	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:34
Subject Sample	# 2	0.232	03:35
Breath Volume		2.453	03:35
Air Blank		0.000	03:36
Control Test		0.077	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:37

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: WHALEY

DATE/TIME OF TEST: 10/08/2025 16:49:11

Diagnostics Check		OK	16:42
Air Blank		0.000	16:43
Control Test		0.079	16:43
Air Blank		0.000	16:43
Subject Sample	# 1	0.000	16:44
Breath Volume		1.171	16:44
Air Blank		0.000	16:44
Air Blank		0.000	16:46
Subject Sample	# 2	0.000	16:47
Breath Volume		1.226	16:47
Air Blank		0.000	16:48
Control Test		0.079	16:48
Air Blank		0.000	16:49
Diagnostics Check		OK	16:49

Data Download Date/Time: 11/06/2025 00:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: THOMAS,TREVOR
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: VANN

DATE/TIME OF TEST: 10/08/2025 23:46:56

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.230	23:42
Breath Volume		1.949	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.248	23:45
Breath Volume		2.058	23:45
Air Blank		0.000	23:46
Control Test		0.077	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RAMCHARAN,KYRAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 14:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: BOLANOS CESPEDES

DATE/TIME OF TEST: 10/09/2025 16:16:52

Diagnostics Check		OK	16:09
Air Blank		0.000	16:10
Control Test		0.079	16:10
Air Blank		0.000	16:11
Subject Sample	# 1	0.278	16:11
Breath Volume		1.453	16:11
Air Blank		0.000	16:12
Air Blank		0.000	16:14
Subject Sample	# 2	0.275	16:15
Breath Volume		1.503	16:15
Air Blank		0.000	16:15
Control Test		0.079	16:16
Air Blank		0.000	16:16
Diagnostics Check		OK	16:16

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEY
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/10/2025 02:53:12

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.091	02:48
Breath Volume		1.859	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:50
Subject Sample	# 2	0.091	02:51
Breath Volume		2.441	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PARRIS

DATE/TIME OF TEST: 10/11/2025 03:35:51

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.129	03:31
Breath Volume		1.160	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:33
Subject Sample	# 2	0.129	03:34
Breath Volume		1.414	03:34
Air Blank		0.000	03:34
Control Test		0.078	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WATSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:52

SUBJECT'S LAST NAME: SUMAR

DATE/TIME OF TEST: 10/14/2025 13:23:53

Diagnostics Check		OK	13:17
Air Blank		0.000	13:17
Control Test		0.079	13:18
Air Blank		0.000	13:18
Subject Sample	# 1	0.221	13:19
Breath Volume		1.742	13:19
Air Blank		0.000	13:19
Air Blank		0.000	13:21
Subject Sample	# 2	0.240	13:22
Breath Volume		1.449	13:22
Air Blank		0.000	13:22
Control Test		0.079	13:23
Air Blank		0.000	13:23
Diagnostics Check		OK	13:23

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 10/18/2025 00:18:04

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.224	00:12
Breath Volume		1.625	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.216	00:16
Breath Volume		1.886	00:16
Air Blank		0.000	00:17
Control Test		0.076	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:18

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WRIGHT,KIERSTEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: OWENS PEREZ

DATE/TIME OF TEST: 10/18/2025 02:33:01

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.179	02:27
Breath Volume		1.820	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.162	02:31
Breath Volume		2.058	02:31
Air Blank		0.000	02:32
Control Test		0.078	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:33

Data Download Date/Time: 11/06/2025 00:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOMEZ,DESTINEE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: CINQUINO

DATE/TIME OF TEST: 10/18/2025 04:52:32

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.151	04:47
Breath Volume		2.515	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.154	04:50
Breath Volume		1.277	04:50
Air Blank		0.000	04:51
Control Test		0.077	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: MUNNINGS

DATE/TIME OF TEST: 10/20/2025 02:21:22

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:18	
Control Test		0.080	02:19	
Air Blank		0.000	02:19	
Subject Sample	# 1		02:19	Subject Test Refused
Breath Volume		0.000	02:19	
Air Blank		0.000	02:20	
Control Test		0.079	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 11/06/2025 00:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FERNANDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 11:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:03

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 10/22/2025 13:35:09

Diagnostics Check		OK	13:28
Air Blank		0.000	13:29
Control Test		0.079	13:29
Air Blank		0.000	13:29
Subject Sample	# 1	0.281	13:30
Breath Volume		1.863	13:30
Air Blank		0.000	13:30
Air Blank		0.000	13:32
Subject Sample	# 2	0.291	13:33
Breath Volume		2.433	13:33
Air Blank		0.000	13:34
Control Test		0.080	13:34
Air Blank		0.000	13:35
Diagnostics Check		OK	13:35

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VAZQUEZ,JUSTIN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: TALABERT

DATE/TIME OF TEST: 10/24/2025 06:00:57

Diagnostics Check		OK	05:54
Air Blank		0.000	05:54
Control Test		0.079	05:55
Air Blank		0.000	05:55
Subject Sample	# 1	0.041	05:56
Breath Volume		1.757	05:56
Air Blank		0.000	05:57
Air Blank		0.000	05:58
Subject Sample	# 2	0.041	05:59
Breath Volume		1.625	05:59
Air Blank		0.000	06:00
Control Test		0.080	06:00
Air Blank		0.000	06:00
Diagnostics Check		OK	06:00

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 10/25/2025 01:12:59

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.241	01:08
Breath Volume		3.648	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:10
Subject Sample	# 2	0.226	01:11
Breath Volume		3.425	01:11
Air Blank		0.000	01:12
Control Test		0.077	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BLANKENSHIP,CHRISTOPHER
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 10/25/2025 03:14:24

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.077	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.212	03:09
Breath Volume		1.636	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.211	03:12
Breath Volume		2.167	03:12
Air Blank		0.000	03:13
Control Test		0.077	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 10/25/2025 04:18:44

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:04	
Control Test		0.078	04:05	
Air Blank		0.000	04:05	
Subject Sample	# 1	0.240	04:08	Volume Not Met
Breath Volume		0.371	04:08	
Air Blank		0.000	04:09	
Air Blank		0.000	04:11	
Subject Sample	# 2	0.256	04:12	
Breath Volume		1.898	04:12	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 3		04:17	Subject Test Refused
Breath Volume		0.000	04:17	
Air Blank		0.000	04:17	
Control Test		0.075	04:18	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: STARKS JORDAN

DATE/TIME OF TEST: 10/25/2025 04:55:39

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.078	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.147	04:51
Breath Volume		1.925	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.148	04:54
Breath Volume		2.242	04:54
Air Blank		0.000	04:54
Control Test		0.077	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMUDEZ,DANIEL
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 10/26/2025 23:50:14

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.080	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.000	23:46
Breath Volume		2.121	23:46
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.000	23:48
Breath Volume		1.132	23:48
Air Blank		0.000	23:49
Control Test		0.079	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: FLEURINOR

DATE/TIME OF TEST: 10/27/2025 00:40:58

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.078	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.142	00:35
Breath Volume		1.171	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.143	00:39
Breath Volume		1.332	00:39
Air Blank		0.000	00:40
Control Test		0.078	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GRANOVSKY,ROBERT
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: WARE

DATE/TIME OF TEST: 10/27/2025 02:30:36

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.086	02:26
Breath Volume		1.457	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.082	02:29
Breath Volume		1.199	02:29
Air Blank		0.000	02:29
Control Test		0.078	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MARCANO

DATE/TIME OF TEST: 10/27/2025 04:23:01

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.102	04:18
Breath Volume		1.281	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.104	04:21
Breath Volume		1.355	04:21
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:23

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARBERIO,MADISON
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: SUAREZ OSPINA

DATE/TIME OF TEST: 10/27/2025 05:30:53

Diagnostics Check		OK	05:24
Air Blank		0.000	05:24
Control Test		0.080	05:25
Air Blank		0.000	05:25
Subject Sample	# 1	0.165	05:26
Breath Volume		2.082	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:28
Subject Sample	# 2	0.163	05:29
Breath Volume		1.542	05:29
Air Blank		0.000	05:29
Control Test		0.077	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GALLAGHER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/30/2025 01:40:22

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.252	01:35
Breath Volume		1.132	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.258	01:38
Breath Volume		1.523	01:38
Air Blank		0.000	01:39
Control Test		0.077	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MASUCCI,LOUIS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/30/2025 05:02:29

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.079	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.111	04:57
Breath Volume		1.988	04:57
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.116	05:00
Breath Volume		2.082	05:00
Air Blank		0.000	05:01
Control Test		0.079	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROMANO,FABIOLA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: WOODBURY

DATE/TIME OF TEST: 10/31/2025 05:05:49

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.079	04:59
Air Blank		0.000	05:00
Subject Sample	# 1	0.164	05:00
Breath Volume		2.386	05:00
Air Blank		0.000	05:01
Air Blank		0.000	05:03
Subject Sample	# 2	0.162	05:04
Breath Volume		1.941	05:04
Air Blank		0.000	05:04
Control Test		0.079	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 11/06/2025 00:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOPEZ,LUIS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:31

SUBJECT'S LAST NAME: FANDINO JR

DATE/TIME OF TEST: 10/31/2025 10:07:14

Diagnostics Check		OK	10:00
Air Blank		0.000	10:01
Control Test		0.079	10:01
Air Blank		0.000	10:01
Subject Sample	# 1	0.192	10:02
Breath Volume		1.238	10:02
Air Blank		0.000	10:03
Air Blank		0.000	10:05
Subject Sample	# 2	0.187	10:05
Breath Volume		1.347	10:05
Air Blank		0.000	10:06
Control Test		0.077	10:06
Air Blank		0.000	10:07
Diagnostics Check		OK	10:07

Data Download Date/Time: 11/06/2025 00:23

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,GARY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,GARY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: SEGURA DIAZ

DATE/TIME OF TEST: 10/07/2025 21:56:07

Diagnostics Check		OK	21:49
Air Blank		0.000	21:50
Control Test		0.080	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.189	21:51
Breath Volume		2.082	21:51
Air Blank		0.000	21:52
Air Blank		0.000	21:53
Subject Sample	# 2	0.190	21:54
Breath Volume		2.292	21:54
Air Blank		0.000	21:55
Control Test		0.080	21:55
Air Blank		0.000	21:56
Diagnostics Check		OK	21:56

Data Download Date/Time: 10/15/2025 18:09

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: GUERRIER

DATE/TIME OF TEST: 10/16/2025 23:06:04

Diagnostics Check		OK	22:59
Air Blank		0.000	23:00
Control Test		0.078	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.000	23:01
Breath Volume		1.406	23:01
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.000	23:04
Breath Volume		3.078	23:04
Air Blank		0.000	23:05
Control Test		0.078	23:05
Air Blank		0.000	23:05
Diagnostics Check		OK	23:06

Data Download Date/Time: 11/01/2025 18:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: RUIZ SANTIAGO

DATE/TIME OF TEST: 10/24/2025 01:32:41

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.078	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.192	01:27
Breath Volume		1.601	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.204	01:30
Breath Volume		2.460	01:30
Air Blank		0.000	01:31
Control Test		0.077	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 11/01/2025 18:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MARTINEZ PINEDA

DATE/TIME OF TEST: 10/08/2025 02:13:06

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.219	02:08
Breath Volume		4.132	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.231	02:11
Breath Volume		2.996	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 11/01/2025 18:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: WILLIAMS JR

DATE/TIME OF TEST: 10/26/2025 04:04:28

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.110	04:00
Breath Volume		2.031	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.104	04:02
Breath Volume		2.234	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 11/01/2025 18:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 10/01/2025 00:54:51

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:43	
Control Test		0.081	00:43	
Air Blank		0.000	00:44	
Subject Sample	# 1	0.081	00:47	Volume Not Met
Breath Volume		0.304	00:47	
Air Blank		0.000	00:48	
Air Blank		0.000	00:49	
Subject Sample	# 2	0.094	00:50	
Breath Volume		1.429	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 3	0.093	00:53	
Breath Volume		1.191	00:53	
Air Blank		0.000	00:53	
Control Test		0.081	00:54	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 10/01/2025 10:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,TORRIE,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BAGLEY,ANDREW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CHUN DE LA CRUZ

DATE/TIME OF TEST: 10/02/2025 02:10:20

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:58	
Control Test		0.082	01:58	
Air Blank		0.000	01:59	
Subject Sample	# 1		02:02	No Sample Provided
Breath Volume		0.000	02:02	
Air Blank		0.000	02:02	
Air Blank		0.000	02:04	
Subject Sample	# 2	0.224	02:05	
Breath Volume		1.222	02:05	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 3	0.226	02:08	
Breath Volume		1.292	02:08	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 10/16/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GODWIN,TRISTAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 10/02/2025 23:56:58

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.083	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:52
Breath Volume		2.203	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:55
Breath Volume		3.718	23:55
Air Blank		0.000	23:56
Control Test		0.083	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 10/16/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,TORRIE,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BAGLEY,ANDREW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 10/03/2025 01:19:24

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.083	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.038	01:14
Breath Volume		1.375	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.035	01:17
Breath Volume		3.593	01:17
Air Blank		0.000	01:18
Control Test		0.081	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: FLYNN,REBECCA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 10/03/2025 03:20:24

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:15	
Control Test		0.082	03:15	
Air Blank		0.000	03:16	
Subject Sample	# 1	0.172	03:16	
Breath Volume		1.652	03:16	
Air Blank		0.000	03:17	
Air Blank			03:19	RFI Detect
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:20	

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: FLYNN,REBECCA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 10/03/2025 03:28:28

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.082	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.182	03:23
Breath Volume		1.554	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.174	03:26
Breath Volume		1.343	03:26
Air Blank		0.000	03:27
Control Test		0.082	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: VARONOS,EVANGELOS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/05/2025 04:35:18

Diagnostics Check		OK	04:27
Air Blank		0.000	04:27
Control Test		0.083	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.160	04:30
Breath Volume		2.046	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.159	04:33
Breath Volume		1.433	04:33
Air Blank		0.000	04:34
Control Test		0.079	04:34
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSEN,CODY,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: YOUNG,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:19

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 10/05/2025 11:48:07

Diagnostics Check		OK	11:39
Air Blank		0.000	11:40
Control Test		0.082	11:40
Air Blank		0.000	11:41
Subject Sample	# 1	0.069	11:42
Breath Volume		1.250	11:42
Air Blank		0.000	11:43
Air Blank		0.000	11:44
Subject Sample	# 2	0.073	11:46
Breath Volume		2.179	11:46
Air Blank		0.000	11:47
Control Test		0.083	11:47
Air Blank		0.000	11:48
Diagnostics Check		OK	11:48

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: THOMPSON,GREGORY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/05/2025 23:03:15

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.082	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.099	22:58
Breath Volume		2.132	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.101	23:01
Breath Volume		2.218	23:01
Air Blank		0.000	23:02
Control Test		0.082	23:02
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CHASE,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GAFKJEN,THOMAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: CAVALIERE

DATE/TIME OF TEST: 10/08/2025 15:03:52

Diagnostics Check		OK	14:57
Air Blank		0.000	14:58
Control Test		0.083	14:58
Air Blank		0.000	14:58
Subject Sample	# 1	0.000	14:59
Breath Volume		2.496	14:59
Air Blank		0.000	14:59
Air Blank		0.000	15:01
Subject Sample	# 2	0.000	15:02
Breath Volume		1.664	15:02
Air Blank		0.000	15:02
Control Test		0.080	15:03
Air Blank		0.000	15:03
Diagnostics Check		OK	15:03

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIINANEN,JOSIAH,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WIINANEN,JOSIAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: KEENAN

DATE/TIME OF TEST: 10/08/2025 21:18:16

Diagnostics Check		OK	21:12
Air Blank		0.000	21:12
Control Test		0.083	21:12
Air Blank		0.000	21:13
Subject Sample	# 1	0.000	21:13
Breath Volume		1.804	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.000	21:16
Breath Volume		2.460	21:16
Air Blank		0.000	21:17
Control Test		0.081	21:17
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLEY,RASHELL,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WRIGHT,DUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: STORMO

DATE/TIME OF TEST: 10/10/2025 03:27:47

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.083	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.146	03:23
Breath Volume		2.523	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.146	03:26
Breath Volume		2.402	03:26
Air Blank		0.000	03:26
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GODWIN,TRISTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: LYNCAN

DATE/TIME OF TEST: 10/12/2025 01:15:58

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:12	
Control Test		0.082	01:12	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:15	
Control Test		0.080	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZKAU,AARON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PRITZKAU,AARON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 10/12/2025 03:59:57

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.083	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.193	03:55
Breath Volume		1.988	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:57
Subject Sample	# 2	0.195	03:58
Breath Volume		1.890	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: THOMPSON,GREGORY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:23

SUBJECT'S LAST NAME: BOURGEOIS BEAVER

DATE/TIME OF TEST: 10/14/2025 18:55:50

Diagnostics Check		OK	18:44
Air Blank		0.000	18:45
Control Test		0.083	18:45
Air Blank		0.000	18:46
Subject Sample	# 1	0.142	18:48
Breath Volume		1.187	18:48
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.150	18:54
Breath Volume		1.234	18:54
Air Blank		0.000	18:54
Control Test		0.081	18:55
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 10/16/2025 14:03

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIINANEN,JOSIAH,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WIINANEN,JOSIAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: ABRAHAM

DATE/TIME OF TEST: 10/19/2025 22:46:08

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.084	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:44	Subject Test Refused
Breath Volume		0.000	22:44	
Air Blank		0.000	22:45	
Control Test		0.082	22:45	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 11/12/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,TORRIE,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: THOMPSON,TORRIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HARGROVE

DATE/TIME OF TEST: 10/20/2025 22:29:00

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.083	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.170	22:24
Breath Volume		1.652	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:26
Subject Sample	# 2	0.165	22:27
Breath Volume		1.449	22:27
Air Blank		0.000	22:28
Control Test		0.082	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:29

Data Download Date/Time: 11/12/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTNER,STEVEN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILSON,WILLIAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:51

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 10/24/2025 13:18:27

Diagnostics Check		OK	13:11	
Air Blank		0.000	13:11	
Control Test		0.084	13:12	
Air Blank		0.000	13:12	
Subject Sample	# 1	0.000	13:15	Volume Not Met
Breath Volume		0.269	13:15	
Air Blank		0.000	13:16	
Air Blank			13:17	Sequence Aborted
Air Blank		0.000	13:18	

Data Download Date/Time: 11/12/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTNER,STEVEN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILSON,WILLIAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:19

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 10/24/2025 13:49:31

Diagnostics Check		OK	13:39	
Air Blank		0.000	13:40	
Control Test		0.083	13:40	
Air Blank		0.000	13:40	
Subject Sample	# 1	0.000	13:43	Volume Not Met
Breath Volume		0.218	13:43	
Air Blank		0.000	13:44	
Air Blank		0.000	13:46	
Subject Sample	# 2		13:48	Subject Test Refused
Breath Volume		0.128	13:48	
Air Blank		0.000	13:48	
Control Test		0.082	13:48	
Air Blank		0.000	13:49	
Diagnostics Check		OK	13:49	

Data Download Date/Time: 11/12/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTNER,STEVEN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILSON,WILLIAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:52

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 10/24/2025 14:17:59

Diagnostics Check		OK	14:13	
Air Blank		0.000	14:13	
Control Test		0.083	14:13	
Air Blank		0.000	14:14	
Subject Sample	# 1		14:16	Subject Test Refused
Breath Volume		0.320	14:16	
Air Blank		0.000	14:17	
Control Test		0.082	14:17	
Air Blank		0.000	14:17	
Diagnostics Check		OK	14:17	

Data Download Date/Time: 11/12/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: GODWIN,TRISTAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: PENALOSA

DATE/TIME OF TEST: 10/26/2025 04:55:57

Diagnostics Check		OK	04:48
Air Blank		0.000	04:49
Control Test		0.083	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.215	04:51
Breath Volume		2.000	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.222	04:54
Breath Volume		2.625	04:54
Air Blank		0.000	04:55
Control Test		0.081	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 11/12/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZKAU,AARON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: FLYNN,REBECCA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 10/27/2025 00:37:12

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.082	00:26	
Air Blank		0.000	00:26	
Subject Sample	# 1	0.070	00:29	Volume Not Met
Breath Volume		0.941	00:29	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 2	0.073	00:32	
Breath Volume		1.359	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:35	
Subject Sample	# 3	0.077	00:35	
Breath Volume		1.875	00:35	
Air Blank		0.000	00:36	
Control Test		0.079	00:36	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 11/12/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZKAU,AARON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PRITZKAU,AARON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 10/30/2025 03:34:48

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.084	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.133	03:29
Breath Volume		2.152	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.131	03:33
Breath Volume		2.457	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 11/12/2025 09:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIINANEN,JOSIAH,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WIINANEN,JOSIAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250133
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: CHITTY

DATE/TIME OF TEST: 10/31/2025 20:07:18

Diagnostics Check		OK	19:59
Air Blank		0.000	19:59
Control Test		0.083	20:00
Air Blank		0.000	20:00
Subject Sample	# 1	0.194	20:02
Breath Volume		1.652	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:04
Subject Sample	# 2	0.194	20:05
Breath Volume		1.968	20:05
Air Blank		0.000	20:06
Control Test		0.081	20:06
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 11/12/2025 09:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,SOLEY,G
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: PATTERSON,CHANDLER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/06/2025 04:18:34

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.079	04:16	
Air Blank		0.000	04:16	
Subject Sample	# 1		04:17	Subject Test Refused
Breath Volume		0.000	04:17	
Air Blank		0.000	04:17	
Control Test		0.079	04:17	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANDY,JAMIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOVELL,A
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 10/08/2025 16:06:07

Diagnostics Check		OK	16:00
Air Blank		0.000	16:00
Control Test		0.079	16:00
Air Blank		0.000	16:01
Subject Sample	# 1	0.000	16:01
Breath Volume		1.535	16:01
Air Blank		0.000	16:02
Air Blank		0.000	16:04
Subject Sample	# 2	0.000	16:04
Breath Volume		2.050	16:04
Air Blank		0.000	16:05
Control Test		0.079	16:05
Air Blank		0.000	16:06
Diagnostics Check		OK	16:06

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANDY,JAMIE,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: JACKSON,J
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 09:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: DIAZROSARIO

DATE/TIME OF TEST: 10/09/2025 11:00:04

Diagnostics Check		OK	10:44	
Air Blank		0.000	10:44	
Control Test		0.079	10:45	
Air Blank		0.000	10:45	
Subject Sample	# 1	0.185	10:48	Volume Not Met
Breath Volume		0.058	10:48	
Air Blank		0.000	10:49	
Air Blank		0.000	10:51	
Subject Sample	# 2	0.194	10:52	
Breath Volume		1.316	10:52	
Air Blank		0.000	10:53	
Air Blank		0.000	10:55	
Subject Sample	# 3	0.171	10:58	Volume Not Met
Breath Volume		0.500	10:58	
Air Blank		0.000	10:59	
Control Test		0.077	10:59	
Air Blank		0.000	10:59	
Diagnostics Check		OK	11:00	

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SKINNER,OWEN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 10/11/2025 02:51:19

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.080	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.116	02:46
Breath Volume		1.320	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.119	02:49
Breath Volume		1.683	02:49
Air Blank		0.000	02:50
Control Test		0.077	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WALKER,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: RODRIQUEZ

DATE/TIME OF TEST: 10/12/2025 21:53:03

Diagnostics Check		OK	21:45
Air Blank		0.000	21:46
Control Test		0.080	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.260	21:48
Breath Volume		2.750	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:50
Subject Sample	# 2	0.256	21:51
Breath Volume		2.757	21:51
Air Blank		0.000	21:52
Control Test		0.078	21:52
Air Blank		0.000	21:52
Diagnostics Check		OK	21:53

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SKINNER,OWEN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 10/16/2025 21:05:40

Diagnostics Check		OK	20:54	
Air Blank		0.000	20:54	
Control Test		0.080	20:54	
Air Blank		0.000	20:55	
Subject Sample	# 1	0.094	20:58	Volume Not Met
Breath Volume		0.000	20:58	
Air Blank		0.000	20:59	
Air Blank		0.000	21:00	
Subject Sample	# 2	0.089	21:04	Volume Not Met
Breath Volume		0.769	21:04	
Air Blank		0.000	21:04	
Control Test		0.075	21:05	
Air Blank		0.000	21:05	
Diagnostics Check		OK	21:05	

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,SOLEY,G
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WATERS,BRODY
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PAULINO

DATE/TIME OF TEST: 10/19/2025 04:15:04

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.165	04:08
Breath Volume		1.339	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.166	04:13
Breath Volume		1.820	04:13
Air Blank		0.000	04:14
Control Test		0.076	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:15

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,SOLEY,G
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: JACKSON,JAMAR
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 12:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: ALBRITTON

DATE/TIME OF TEST: 10/19/2025 18:03:44

Diagnostics Check		OK	18:00	
Air Blank		0.000	18:01	
Control Test		0.079	18:01	
Air Blank		0.000	18:02	
Subject Sample	# 1		18:02	Subject Test Refused
Breath Volume		0.000	18:02	
Air Blank		0.000	18:02	
Control Test		0.079	18:03	
Air Blank		0.000	18:03	
Diagnostics Check		OK	18:03	

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,SOLEY,G
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SERGEYEV,PAVEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GOODRICH

DATE/TIME OF TEST: 10/19/2025 23:21:36

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.079	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.192	23:17
Breath Volume		2.433	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.179	23:20
Breath Volume		1.765	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: TAYLOR,OCTAVIAS
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/29/2025 03:05:38

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.216	03:00
Breath Volume		3.617	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.208	03:03
Breath Volume		1.929	03:03
Air Blank		0.000	03:04
Control Test		0.077	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELHOMME,MAXON,N
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WATERS,BRODY
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 10/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/29/2025 03:18:08

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.125	03:13
Breath Volume		1.347	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.131	03:16
Breath Volume		2.054	03:16
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 12/01/2025 03:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: RESTREPO

DATE/TIME OF TEST: 10/04/2025 02:01:16

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.000	01:56
Breath Volume		1.437	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.000	01:59
Breath Volume		1.878	01:59
Air Blank		0.000	02:00
Control Test		0.078	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/04/2025 03:33:10

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.262	03:28
Breath Volume		1.488	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.261	03:31
Breath Volume		1.339	03:31
Air Blank		0.000	03:32
Control Test		0.077	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: VAZQUEZ HERNANDEZ

DATE/TIME OF TEST: 10/04/2025 22:36:35

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.079	22:30
Air Blank		0.000	22:31
Subject Sample	# 1	0.218	22:31
Breath Volume		1.742	22:31
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.204	22:34
Breath Volume		1.457	22:34
Air Blank		0.000	22:35
Control Test		0.077	22:35
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEPPARD,S
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: HICKSON

DATE/TIME OF TEST: 10/05/2025 04:07:08

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.117	04:02
Breath Volume		3.007	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.118	04:05
Breath Volume		2.574	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: RODRIGUEZ GONZALEZ

DATE/TIME OF TEST: 10/05/2025 05:37:53

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.078	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.071	05:33
Breath Volume		1.417	05:33
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.077	05:36
Breath Volume		1.730	05:36
Air Blank		0.000	05:36
Control Test		0.079	05:37
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURAN,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: LOPEZ PEREZ

DATE/TIME OF TEST: 10/05/2025 09:07:19

Diagnostics Check		OK	09:00
Air Blank		0.000	09:01
Control Test		0.078	09:01
Air Blank		0.000	09:02
Subject Sample	# 1	0.000	09:02
Breath Volume		1.628	09:02
Air Blank		0.000	09:03
Air Blank		0.000	09:05
Subject Sample	# 2	0.000	09:05
Breath Volume		1.203	09:05
Air Blank		0.000	09:06
Control Test		0.078	09:06
Air Blank		0.000	09:07
Diagnostics Check		OK	09:07

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: CRIST

DATE/TIME OF TEST: 10/06/2025 00:34:21

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.078	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.121	00:29
Breath Volume		1.292	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.125	00:32
Breath Volume		2.000	00:32
Air Blank		0.000	00:33
Control Test		0.078	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DIAZ SANTIZ

DATE/TIME OF TEST: 10/06/2025 02:37:41

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.210	02:32
Breath Volume		1.714	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.201	02:35
Breath Volume		2.691	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORRIS,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LOGAN

DATE/TIME OF TEST: 10/08/2025 00:33:25

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.295	00:27
Breath Volume		1.753	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.292	00:31
Breath Volume		1.597	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: LEON ARANDA

DATE/TIME OF TEST: 10/08/2025 02:49:06

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.078	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.184	02:44
Breath Volume		1.523	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.183	02:47
Breath Volume		1.804	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/08/2025 23:23:47

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:13	
Control Test		0.078	23:13	
Air Blank		0.000	23:13	
Subject Sample	# 1	0.059	23:17	
Breath Volume		1.218	23:17	
Air Blank		0.000	23:17	
Air Blank		0.000	23:19	
Subject Sample	# 2		23:22	Subject Test Refused
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Control Test		0.081	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: TIMBY

DATE/TIME OF TEST: 10/09/2025 01:47:14

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.078	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.065	01:42
Breath Volume		2.011	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:44
Subject Sample	# 2	0.066	01:45
Breath Volume		1.859	01:45
Air Blank		0.000	01:46
Control Test		0.081	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 10/11/2025 03:29:24

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.078	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.160	03:24
Breath Volume		2.656	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:26
Subject Sample	# 2	0.142	03:27
Breath Volume		1.320	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 10/12/2025 04:38:40

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.164	04:33
Breath Volume		2.421	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.162	04:36
Breath Volume		1.375	04:36
Air Blank		0.000	04:37
Control Test		0.077	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 10/16/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: HALL JR

DATE/TIME OF TEST: 10/12/2025 05:01:30

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:49	
Control Test		0.078	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 1	0.183	04:53	Volume Not Met
Breath Volume		0.488	04:53	
Air Blank		0.000	04:54	
Air Blank		0.000	04:55	
Subject Sample	# 2	0.174	04:56	
Breath Volume		1.101	04:56	
Air Blank		0.000	04:57	
Air Blank		0.000	04:59	
Subject Sample	# 3	0.176	04:59	
Breath Volume		1.429	04:59	
Air Blank		0.000	05:00	
Control Test		0.080	05:00	
Air Blank		0.000	05:01	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 10/16/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JACOBS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: WIMER

DATE/TIME OF TEST: 10/12/2025 05:31:04

Diagnostics Check		OK	05:24
Air Blank		0.000	05:24
Control Test		0.078	05:24
Air Blank		0.000	05:25
Subject Sample	# 1	0.130	05:26
Breath Volume		1.156	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:28
Subject Sample	# 2	0.128	05:29
Breath Volume		1.949	05:29
Air Blank		0.000	05:30
Control Test		0.078	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:31

Data Download Date/Time: 10/16/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: BERBER

DATE/TIME OF TEST: 10/12/2025 22:36:35

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.079	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.150	22:31
Breath Volume		2.242	22:31
Air Blank		0.000	22:32
Air Blank		0.000	22:33
Subject Sample	# 2	0.158	22:34
Breath Volume		2.144	22:34
Air Blank		0.000	22:35
Control Test		0.078	22:35
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 10/16/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: BEJARANO HERNANDEZ

DATE/TIME OF TEST: 10/13/2025 01:53:24

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.152	01:48
Breath Volume		1.441	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.150	01:51
Breath Volume		1.382	01:51
Air Blank		0.000	01:52
Control Test		0.078	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 10/16/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/13/2025 03:55:16

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.079	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.204	03:50
Breath Volume		1.398	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.196	03:53
Breath Volume		1.101	03:53
Air Blank		0.000	03:54
Control Test		0.078	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 10/16/2025 12:31

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 10/16/2025 20:25:11

Diagnostics Check		OK	20:12	
Air Blank		0.000	20:13	
Control Test		0.079	20:13	
Air Blank		0.000	20:14	
Subject Sample	# 1	0.193	20:17	Volume Not Met
Breath Volume		0.617	20:17	
Air Blank		0.000	20:18	
Air Blank		0.000	20:19	
Subject Sample	# 2	0.243	20:20	
Breath Volume		1.519	20:20	
Air Blank		0.000	20:21	
Air Blank		0.000	20:22	
Subject Sample	# 3	0.242	20:23	
Breath Volume		1.691	20:23	
Air Blank		0.000	20:24	
Control Test		0.079	20:24	
Air Blank		0.000	20:25	
Diagnostics Check		OK	20:25	

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: GOMEZ GOMEZ

DATE/TIME OF TEST: 10/16/2025 23:12:58

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.079	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.200	23:08
Breath Volume		1.171	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:10
Subject Sample	# 2	0.204	23:11
Breath Volume		3.027	23:11
Air Blank		0.000	23:12
Control Test		0.078	23:12
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 11/06/2025 12:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: VERA MACEDO

DATE/TIME OF TEST: 10/17/2025 02:29:35

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.078	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.131	02:24
Breath Volume		2.386	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.127	02:27
Breath Volume		2.886	02:27
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RICHARDSON,G
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/17/2025 08:37:27

Diagnostics Check		OK	08:31
Air Blank		0.000	08:31
Control Test		0.079	08:31
Air Blank		0.000	08:32
Subject Sample	# 1	0.189	08:32
Breath Volume		4.156	08:32
Air Blank		0.000	08:33
Air Blank		0.000	08:35
Subject Sample	# 2	0.180	08:35
Breath Volume		3.746	08:35
Air Blank		0.000	08:36
Control Test		0.081	08:36
Air Blank		0.000	08:37
Diagnostics Check		OK	08:37

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ-ARROYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 10/17/2025 18:33:37

Diagnostics Check		OK	18:23
Air Blank		0.000	18:23
Control Test		0.079	18:24
Air Blank		0.000	18:24
Subject Sample	# 1	0.382	18:27
Breath Volume		1.406	18:27
Air Blank		0.000	18:28
Air Blank		0.000	18:30
Subject Sample	# 2	0.364	18:31
Breath Volume		2.074	18:31
Air Blank		0.000	18:32
Control Test		0.080	18:32
Air Blank		0.000	18:33
Diagnostics Check		OK	18:33

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: HUDGINS

DATE/TIME OF TEST: 10/17/2025 22:37:20

Diagnostics Check		OK	22:29
Air Blank		0.000	22:30
Control Test		0.078	22:30
Air Blank		0.000	22:31
Subject Sample	# 1	0.150	22:31
Breath Volume		2.210	22:31
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.147	22:35
Breath Volume		2.019	22:35
Air Blank		0.000	22:36
Control Test		0.078	22:36
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SAUREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 10/18/2025 01:01:21

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:49	
Control Test		0.078	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1	0.103	00:53	Volume Not Met
Breath Volume		0.550	00:53	
Air Blank		0.000	00:53	
Air Blank		0.000	00:55	
Subject Sample	# 2	0.127	00:56	
Breath Volume		1.222	00:56	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 3	0.130	00:59	
Breath Volume		1.363	00:59	
Air Blank		0.000	01:00	
Control Test		0.081	01:00	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/19/2025 02:34:05

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.078	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.187	02:29
Breath Volume		2.015	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.187	02:32
Breath Volume		1.390	02:32
Air Blank		0.000	02:33
Control Test		0.078	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:34

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MERINO,M
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: OLUTOYE

DATE/TIME OF TEST: 10/19/2025 02:57:38

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:45	
Control Test		0.078	02:45	
Air Blank		0.000	02:46	
Subject Sample	# 1	0.130	02:46	
Breath Volume		1.500	02:46	
Air Blank		0.000	02:47	
Air Blank		0.000	02:49	
Subject Sample	# 2		02:52	No Sample Provided
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Air Blank		0.000	02:54	
Subject Sample	# 3		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:56	
Control Test		0.078	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 11/06/2025 12:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST, WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: DAMIAN

DATE/TIME OF TEST: 10/19/2025 05:58:10

Diagnostics Check		OK	05:51
Air Blank		0.000	05:51
Control Test		0.078	05:52
Air Blank		0.000	05:52
Subject Sample	# 1	0.151	05:53
Breath Volume		1.773	05:53
Air Blank		0.000	05:54
Air Blank		0.000	05:55
Subject Sample	# 2	0.148	05:56
Breath Volume		1.613	05:56
Air Blank		0.000	05:57
Control Test		0.078	05:57
Air Blank		0.000	05:58
Diagnostics Check		OK	05:58

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MERCADO MONTANEZ

DATE/TIME OF TEST: 10/21/2025 03:52:13

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.084	03:47
Breath Volume		1.539	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.078	03:50
Breath Volume		1.347	03:50
Air Blank		0.000	03:51
Control Test		0.078	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMOND,JENNY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOTHEM,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:59

SUBJECT'S LAST NAME: SCUDDER

DATE/TIME OF TEST: 10/21/2025 11:25:50

Diagnostics Check		OK	11:19
Air Blank		0.000	11:19
Control Test		0.079	11:20
Air Blank		0.000	11:20
Subject Sample	# 1	0.197	11:21
Breath Volume		2.140	11:21
Air Blank		0.000	11:22
Air Blank		0.000	11:23
Subject Sample	# 2	0.177	11:24
Breath Volume		1.261	11:24
Air Blank		0.000	11:24
Control Test		0.078	11:25
Air Blank		0.000	11:25
Diagnostics Check		OK	11:25

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: TAPIA NIEVES

DATE/TIME OF TEST: 10/21/2025 23:51:14

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:37	
Control Test		0.078	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1	0.147	23:40	
Breath Volume		1.546	23:40	
Air Blank		0.000	23:40	
Air Blank		0.000	23:42	
Subject Sample	# 2	0.143	23:45	Volume Not Met
Breath Volume		0.042	23:45	
Air Blank		0.000	23:46	
Air Blank		0.000	23:48	
Subject Sample	# 3	0.131	23:49	
Breath Volume		1.320	23:49	
Air Blank		0.000	23:50	
Control Test		0.080	23:50	
Air Blank		0.000	23:51	
Diagnostics Check		OK	23:51	

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: DREW

DATE/TIME OF TEST: 10/22/2025 02:31:00

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.173	02:25
Breath Volume		1.378	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.169	02:29
Breath Volume		1.609	02:29
Air Blank		0.000	02:30
Control Test		0.078	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:31

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CASAUBON

DATE/TIME OF TEST: 10/24/2025 02:37:45

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.079	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.161	02:33
Breath Volume		2.214	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.148	02:36
Breath Volume		2.683	02:36
Air Blank		0.000	02:36
Control Test		0.079	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 11/06/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 10/24/2025 18:29:10

Diagnostics Check		OK	18:22
Air Blank		0.000	18:22
Control Test		0.079	18:22
Air Blank		0.000	18:23
Subject Sample	# 1	0.253	18:24
Breath Volume		2.691	18:24
Air Blank		0.000	18:25
Air Blank		0.000	18:26
Subject Sample	# 2	0.249	18:27
Breath Volume		2.710	18:27
Air Blank		0.000	18:28
Control Test		0.077	18:28
Air Blank		0.000	18:29
Diagnostics Check		OK	18:29

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIBBINS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MENDEZ DIAZ

DATE/TIME OF TEST: 10/24/2025 23:59:39

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.079	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.114	23:55
Breath Volume		1.605	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.114	23:57
Breath Volume		2.414	23:57
Air Blank		0.000	23:58
Control Test		0.078	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: ODROWSKI

DATE/TIME OF TEST: 10/25/2025 21:53:55

Diagnostics Check		OK	21:46
Air Blank		0.000	21:47
Control Test		0.079	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.097	21:49
Breath Volume		3.355	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:51
Subject Sample	# 2	0.096	21:52
Breath Volume		3.515	21:52
Air Blank		0.000	21:52
Control Test		0.079	21:53
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,B
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 10/26/2025 01:13:47

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.079	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:12	RFI Detect
Breath Volume		0.000	01:12	
Air Blank		0.000	01:12	
Control Test		0.079	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JOHNSON,B
ARRESTING OFFICER AGENCY: Macdill Air Force Ba
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 10/26/2025 01:21:10

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.078	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.177	01:16
Breath Volume		2.238	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:18
Subject Sample	# 2	0.170	01:19
Breath Volume		1.859	01:19
Air Blank		0.000	01:20
Control Test		0.078	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: CABA

DATE/TIME OF TEST: 10/26/2025 02:21:15

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:09	
Control Test		0.079	02:10	
Air Blank		0.000	02:10	
Subject Sample	# 1	0.064	02:13	Volume Not Met
Breath Volume		0.093	02:13	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 2	0.055	02:19	Volume Not Met
Breath Volume		0.453	02:19	
Air Blank		0.000	02:20	
Control Test		0.082	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 10/26/2025 04:30:09

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.141	04:25
Breath Volume		1.179	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:27
Subject Sample	# 2	0.146	04:28
Breath Volume		1.171	04:28
Air Blank		0.000	04:29
Control Test		0.078	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: GODINEZ MORALES

DATE/TIME OF TEST: 10/26/2025 05:47:46

Diagnostics Check		OK	05:34	
Air Blank		0.000	05:34	
Control Test		0.079	05:34	
Air Blank		0.000	05:35	
Subject Sample	# 1	0.155	05:38	Volume Not Met
Breath Volume		0.390	05:38	
Air Blank		0.000	05:39	
Air Blank		0.000	05:41	
Subject Sample	# 2	0.192	05:42	
Breath Volume		1.250	05:42	
Air Blank		0.000	05:43	
Air Blank		0.000	05:45	
Subject Sample	# 3	0.185	05:45	
Breath Volume		1.710	05:45	
Air Blank		0.000	05:46	
Control Test		0.079	05:47	
Air Blank		0.000	05:47	
Diagnostics Check		OK	05:47	

Data Download Date/Time: 11/06/2025 12:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: ESALOMI

DATE/TIME OF TEST: 10/26/2025 23:39:05

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.079	23:32
Air Blank		0.000	23:32
Subject Sample	# 1	0.337	23:34
Breath Volume		1.675	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:36
Subject Sample	# 2	0.327	23:37
Breath Volume		1.921	23:37
Air Blank		0.000	23:38
Control Test		0.078	23:38
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: COBAS ALVAREZ

DATE/TIME OF TEST: 10/27/2025 03:03:59

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.078	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.174	02:59
Breath Volume		1.343	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:01
Subject Sample	# 2	0.174	03:02
Breath Volume		1.628	03:02
Air Blank		0.000	03:03
Control Test		0.076	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: LEON LEON

DATE/TIME OF TEST: 10/27/2025 07:37:13

Diagnostics Check		OK	07:30
Air Blank		0.000	07:30
Control Test		0.078	07:31
Air Blank		0.000	07:31
Subject Sample	# 1	0.133	07:32
Breath Volume		2.980	07:32
Air Blank		0.000	07:33
Air Blank		0.000	07:34
Subject Sample	# 2	0.140	07:35
Breath Volume		1.609	07:35
Air Blank		0.000	07:36
Control Test		0.082	07:36
Air Blank		0.000	07:37
Diagnostics Check		OK	07:37

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORSEN,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: PAULINO

DATE/TIME OF TEST: 10/28/2025 00:48:33

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:37	
Control Test		0.078	00:37	
Air Blank		0.000	00:38	
Subject Sample	# 1	0.000	00:41	Volume Not Met
Breath Volume		0.035	00:41	
Air Blank		0.000	00:41	
Air Blank		0.000	00:43	
Subject Sample	# 2	0.000	00:47	Volume Not Met
Breath Volume		0.703	00:47	
Air Blank		0.000	00:47	
Control Test		0.082	00:47	
Air Blank		0.000	00:48	
Diagnostics Check		OK	00:48	

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HIJAZ

DATE/TIME OF TEST: 10/31/2025 00:42:38

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.026	00:37
Breath Volume		1.253	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.025	00:40
Breath Volume		1.613	00:40
Air Blank		0.000	00:41
Control Test		0.079	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/31/2025 01:18:15

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:10	
Control Test		0.081	01:10	
Air Blank		0.000	01:10	
Subject Sample	# 1		01:14	No Sample Provided
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Air Blank		0.000	01:16	
Subject Sample	# 2		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BONILLA,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 10/31/2025 14:47:38

Diagnostics Check		OK	14:38
Air Blank		0.000	14:38
Control Test		0.080	14:38
Air Blank		0.000	14:39
Subject Sample	# 1	0.227	14:42
Breath Volume		1.750	14:42
Air Blank		0.000	14:43
Air Blank		0.000	14:44
Subject Sample	# 2	0.218	14:45
Breath Volume		1.531	14:45
Air Blank		0.000	14:46
Control Test		0.079	14:46
Air Blank		0.000	14:47
Diagnostics Check		OK	14:47

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUEVAS,L
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: DE LEON MELENDEZ

DATE/TIME OF TEST: 10/31/2025 21:11:24

Diagnostics Check		OK	21:04
Air Blank		0.000	21:05
Control Test		0.080	21:05
Air Blank		0.000	21:06
Subject Sample	# 1	0.077	21:06
Breath Volume		1.402	21:06
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.076	21:09
Breath Volume		1.105	21:09
Air Blank		0.000	21:10
Control Test		0.079	21:10
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: PORTALATIN

DATE/TIME OF TEST: 10/31/2025 22:34:00

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.079	22:27
Air Blank		0.000	22:28
Subject Sample	# 1	0.090	22:29
Breath Volume		1.878	22:29
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.081	22:32
Breath Volume		1.839	22:32
Air Blank		0.000	22:33
Control Test		0.079	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:34

Data Download Date/Time: 11/06/2025 12:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LASHLEY,ADAM
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PHAM

DATE/TIME OF TEST: 10/01/2025 01:03:49

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.079	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Control Test		0.079	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN-REYES,JUAN,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DORESTANT,MARKENS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: GEBERT

DATE/TIME OF TEST: 10/01/2025 18:08:20

Diagnostics Check		OK	18:01
Air Blank		0.000	18:02
Control Test		0.078	18:02
Air Blank		0.000	18:03
Subject Sample	# 1	0.330	18:03
Breath Volume		2.152	18:03
Air Blank		0.000	18:04
Air Blank		0.000	18:06
Subject Sample	# 2	0.312	18:06
Breath Volume		1.687	18:06
Air Blank		0.000	18:07
Control Test		0.078	18:07
Air Blank		0.000	18:08
Diagnostics Check		OK	18:08

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANDERHOOF,COREY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: SHARP,ELIOT
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 10/02/2025 03:04:35

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.078	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:03	Subject Test Refused
Breath Volume		0.000	03:03	
Air Blank		0.000	03:03	
Control Test		0.079	03:03	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ROSS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HUGHES,ROSS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SCHOFILL

DATE/TIME OF TEST: 10/06/2025 22:04:48

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.078	21:59
Air Blank		0.000	21:59
Subject Sample	# 1	0.243	22:00
Breath Volume		1.867	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.233	22:03
Breath Volume		2.910	22:03
Air Blank		0.000	22:03
Control Test		0.077	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ADORNO,SAMUEL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 10/07/2025 05:48:14

Diagnostics Check		OK	05:43	
Air Blank		0.000	05:44	
Control Test		0.079	05:44	
Air Blank		0.000	05:45	
Subject Sample	# 1		05:46	Subject Test Refused
Breath Volume		0.000	05:46	
Air Blank		0.000	05:47	
Control Test		0.078	05:47	
Air Blank		0.000	05:48	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: OSBORNE,JEREMIAH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: PURCELL JR

DATE/TIME OF TEST: 10/08/2025 02:08:29

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.250	02:00
Breath Volume		2.839	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:02
Subject Sample	# 2	0.226	02:03
Breath Volume		2.222	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 3	0.257	02:06
Breath Volume		3.457	02:06
Air Blank		0.000	02:07
Control Test		0.076	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: KAUFMAN

DATE/TIME OF TEST: 10/08/2025 21:22:27

Diagnostics Check		OK	21:19	
Air Blank		0.000	21:19	
Control Test		0.079	21:20	
Air Blank		0.000	21:20	
Subject Sample	# 1		21:21	Subject Test Refused
Breath Volume		0.000	21:21	
Air Blank		0.000	21:21	
Control Test		0.079	21:21	
Air Blank		0.000	21:22	
Diagnostics Check		OK	21:22	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JAMES
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: BOLTON

DATE/TIME OF TEST: 10/08/2025 21:32:52

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.054	21:27
Breath Volume		2.042	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:29
Subject Sample	# 2	0.052	21:31
Breath Volume		2.449	21:31
Air Blank		0.000	21:31
Control Test		0.079	21:32
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: OSBORNE,JEREMIAH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: KUBAT

DATE/TIME OF TEST: 10/10/2025 20:15:49

Diagnostics Check		OK	20:09
Air Blank		0.000	20:09
Control Test		0.078	20:10
Air Blank		0.000	20:10
Subject Sample	# 1	0.217	20:11
Breath Volume		2.105	20:11
Air Blank		0.000	20:11
Air Blank		0.000	20:13
Subject Sample	# 2	0.213	20:14
Breath Volume		3.246	20:14
Air Blank		0.000	20:14
Control Test		0.077	20:15
Air Blank		0.000	20:15
Diagnostics Check		OK	20:15

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ROJAS BURGOS,JESUS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 10/11/2025 03:08:30

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.078	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.140	03:03
Breath Volume		1.855	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.128	03:06
Breath Volume		1.394	03:06
Air Blank		0.000	03:07
Control Test		0.076	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: UA,CAMERON
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: CRANE

DATE/TIME OF TEST: 10/11/2025 04:49:41

Diagnostics Check		OK	04:43
Air Blank		0.000	04:43
Control Test		0.079	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.037	04:45
Breath Volume		2.269	04:45
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.036	04:48
Breath Volume		1.820	04:48
Air Blank		0.000	04:48
Control Test		0.079	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: NOA BENTANCOURT

DATE/TIME OF TEST: 10/12/2025 04:13:00

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.079	04:04	
Air Blank		0.000	04:04	
Subject Sample	# 1	0.210	04:05	
Breath Volume		1.734	04:05	
Air Blank		0.000	04:06	
Air Blank		0.000	04:08	
Subject Sample	# 2	0.142	04:11	Volume Not Met
Breath Volume		0.421	04:11	
Air Blank			04:12	Purge Fail
Air Blank		0.000	04:13	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: NOA BENTANCOURT

DATE/TIME OF TEST: 10/12/2025 04:24:01

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.075	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.189	04:16
Breath Volume		1.144	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.192	04:22
Breath Volume		1.640	04:22
Air Blank		0.000	04:23
Control Test		0.075	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:24

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ONEILL-COTTLE,PRESTON,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ONEILL-COTTLE,PRESTON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 13:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 10/13/2025 15:10:25

Diagnostics Check		OK	15:03
Air Blank		0.000	15:04
Control Test		0.079	15:04
Air Blank		0.000	15:05
Subject Sample	# 1	0.116	15:05
Breath Volume		1.402	15:05
Air Blank		0.000	15:06
Air Blank		0.000	15:08
Subject Sample	# 2	0.116	15:08
Breath Volume		1.917	15:08
Air Blank		0.000	15:09
Control Test		0.078	15:09
Air Blank		0.000	15:10
Diagnostics Check		OK	15:10

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,BRANDON,H
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ROBINSON,TYREE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: JESUS ABDIEL

DATE/TIME OF TEST: 10/15/2025 06:21:19

Diagnostics Check		OK	06:14
Air Blank		0.000	06:15
Control Test		0.079	06:15
Air Blank		0.000	06:15
Subject Sample	# 1	0.067	06:16
Breath Volume		3.027	06:16
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.064	06:19
Breath Volume		3.609	06:19
Air Blank		0.000	06:20
Control Test		0.077	06:20
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANDERHOOF,COREY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: VANDERHOOF,COREY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: WALSER

DATE/TIME OF TEST: 10/15/2025 21:31:36

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:28	
Control Test		0.079	21:29	
Air Blank		0.000	21:29	
Subject Sample	# 1		21:30	Subject Test Refused
Breath Volume		0.000	21:30	
Air Blank		0.000	21:30	
Control Test		0.079	21:30	
Air Blank		0.000	21:31	
Diagnostics Check		OK	21:31	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANDERHOOF,COREY,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: VANDERHOOF,COREY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BISHOP

DATE/TIME OF TEST: 10/17/2025 03:35:13

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:32	
Control Test		0.079	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:33	Subject Test Refused
Breath Volume		0.000	03:33	
Air Blank		0.000	03:34	
Control Test		0.079	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: BUFF

DATE/TIME OF TEST: 10/17/2025 20:36:43

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.079	20:31
Air Blank		0.000	20:31
Subject Sample	# 1	0.246	20:32
Breath Volume		2.117	20:32
Air Blank		0.000	20:33
Air Blank		0.000	20:34
Subject Sample	# 2	0.250	20:35
Breath Volume		1.554	20:35
Air Blank		0.000	20:35
Control Test		0.077	20:36
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WRIGHT,DANIEL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BADALAMENTI

DATE/TIME OF TEST: 10/18/2025 03:22:02

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.195	03:17
Breath Volume		2.957	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.194	03:20
Breath Volume		3.015	03:20
Air Blank		0.000	03:21
Control Test		0.077	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:22

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: DEFELICE

DATE/TIME OF TEST: 10/18/2025 20:32:13

Diagnostics Check		OK	20:26
Air Blank		0.000	20:26
Control Test		0.079	20:26
Air Blank		0.000	20:27
Subject Sample	# 1	0.052	20:27
Breath Volume		2.066	20:27
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.057	20:30
Breath Volume		2.726	20:30
Air Blank		0.000	20:31
Control Test		0.078	20:31
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WRIGHT,DANIEL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: DIMIC

DATE/TIME OF TEST: 10/18/2025 20:54:48

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.079	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.123	20:50
Breath Volume		2.863	20:50
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.128	20:53
Breath Volume		3.550	20:53
Air Blank		0.000	20:53
Control Test		0.078	20:54
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WRIGHT,DANIEL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CZARNECKI

DATE/TIME OF TEST: 10/18/2025 23:42:34

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:39	
Control Test		0.079	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:41	
Control Test		0.078	23:41	
Air Blank		0.000	23:42	
Diagnostics Check		OK	23:42	

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: NEIFORD

DATE/TIME OF TEST: 10/19/2025 02:11:22

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.079	02:09	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Control Test		0.078	02:10	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 10/19/2025 20:31:45

Diagnostics Check		OK	20:25
Air Blank		0.000	20:25
Control Test		0.079	20:26
Air Blank		0.000	20:26
Subject Sample	# 1	0.077	20:27
Breath Volume		2.324	20:27
Air Blank		0.000	20:27
Air Blank		0.000	20:29
Subject Sample	# 2	0.079	20:30
Breath Volume		3.128	20:30
Air Blank		0.000	20:30
Control Test		0.078	20:31
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: FLORES,MADILYN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: JOHNSON JR

DATE/TIME OF TEST: 10/19/2025 21:19:00

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:16	
Control Test		0.078	21:16	
Air Blank		0.000	21:17	
Subject Sample	# 1		21:17	Subject Test Refused
Breath Volume		0.000	21:17	
Air Blank		0.000	21:18	
Control Test		0.079	21:18	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:19	

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ANDRADE,ESTEBAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 10/22/2025 01:34:13

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.079	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1	0.193	01:26	Volume Not Met
Breath Volume		0.000	01:26	
Air Blank		0.000	01:27	
Air Blank		0.000	01:29	
Subject Sample	# 2	0.176	01:32	Volume Not Met
Breath Volume		0.019	01:32	
Air Blank		0.000	01:33	
Control Test		0.075	01:33	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 11/01/2025 13:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LANE,JONATHAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 14:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/23/2025 15:08:04

Diagnostics Check		OK	15:00
Air Blank		0.000	15:01
Control Test		0.078	15:01
Air Blank		0.000	15:02
Subject Sample	# 1	0.144	15:03
Breath Volume		1.902	15:03
Air Blank		0.000	15:04
Air Blank		0.000	15:05
Subject Sample	# 2	0.143	15:06
Breath Volume		1.500	15:06
Air Blank		0.000	15:07
Control Test		0.077	15:07
Air Blank		0.000	15:07
Diagnostics Check		OK	15:08

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: CARON

DATE/TIME OF TEST: 10/23/2025 15:56:16

Diagnostics Check		OK	15:49
Air Blank		0.000	15:49
Control Test		0.079	15:49
Air Blank		0.000	15:50
Subject Sample	# 1	0.362	15:51
Breath Volume		3.218	15:51
Air Blank		0.000	15:52
Air Blank		0.000	15:53
Subject Sample	# 2	0.355	15:54
Breath Volume		3.164	15:54
Air Blank		0.000	15:55
Control Test		0.078	15:55
Air Blank		0.000	15:56
Diagnostics Check		OK	15:56

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WRIGHT,DANIEL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: ANDREASSON

DATE/TIME OF TEST: 10/23/2025 21:08:22

Diagnostics Check		OK	21:02
Air Blank		0.000	21:02
Control Test		0.076	21:02
Air Blank		0.000	21:03
Subject Sample	# 1	0.198	21:03
Breath Volume		2.523	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.198	21:06
Breath Volume		2.613	21:06
Air Blank		0.000	21:07
Control Test		0.077	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: LOPEZ,JORGE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: ROJAS MATAMOROS

DATE/TIME OF TEST: 10/23/2025 21:21:06

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.077	21:14
Air Blank		0.000	21:14
Subject Sample	# 1	0.223	21:15
Breath Volume		1.527	21:15
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.212	21:19
Breath Volume		1.597	21:19
Air Blank		0.000	21:20
Control Test		0.076	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: NEUFELD

DATE/TIME OF TEST: 10/23/2025 21:59:38

Diagnostics Check		OK	21:56	
Air Blank		0.000	21:57	
Control Test		0.076	21:57	
Air Blank		0.000	21:58	
Subject Sample	# 1		21:58	Subject Test Refused
Breath Volume		0.000	21:58	
Air Blank		0.000	21:58	
Control Test		0.077	21:59	
Air Blank		0.000	21:59	
Diagnostics Check		OK	21:59	

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/23/2025 22:13:59

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.077	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.113	22:08
Breath Volume		2.269	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:11
Subject Sample	# 2	0.106	22:12
Breath Volume		1.886	22:12
Air Blank		0.000	22:13
Control Test		0.075	22:13
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO PEREZ,RODOLFO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: JEAN-BAPTISTE,STANLEY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 10/23/2025 23:12:23

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.077	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.064	23:07
Breath Volume		1.679	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.059	23:10
Breath Volume		1.285	23:10
Air Blank		0.000	23:11
Control Test		0.076	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 10/25/2025 03:36:46

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.079	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1	0.130	03:27	Volume Not Met
Breath Volume		0.070	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2	0.145	03:32	
Breath Volume		2.230	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 3	0.147	03:34	
Breath Volume		1.281	03:34	
Air Blank		0.000	03:35	
Control Test		0.072	03:36	Control Outside Tolerance
Air Blank		0.000	03:36	

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 10/25/2025 03:44:34

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.076	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1	0.139	03:43	
Breath Volume		1.839	03:43	
Air Blank			03:43	Purge Fail
Air Blank		0.000	03:44	

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 10/25/2025 03:53:19

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.076	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.143	03:48
Breath Volume		2.230	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.134	03:51
Breath Volume		1.179	03:51
Air Blank		0.000	03:52
Control Test		0.076	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: FLORYANCE,CODY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 10/26/2025 00:30:08

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.153	00:25
Breath Volume		1.953	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.148	00:28
Breath Volume		1.863	00:28
Air Blank		0.000	00:29
Control Test		0.077	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: FLORYANCE,CODY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: GUADARRAMA

DATE/TIME OF TEST: 10/26/2025 05:23:10

Diagnostics Check		OK	05:20	
Air Blank		0.000	05:20	
Control Test		0.079	05:21	
Air Blank		0.000	05:21	
Subject Sample	# 1		05:21	Subject Test Refused
Breath Volume		0.000	05:21	
Air Blank		0.000	05:22	
Control Test		0.080	05:22	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:23	

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WRIGHT,DANIEL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: BRASHER

DATE/TIME OF TEST: 10/29/2025 03:23:44

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.193	03:19
Breath Volume		3.050	03:19
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.191	03:22
Breath Volume		1.679	03:22
Air Blank		0.000	03:22
Control Test		0.079	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ANDRADE,ESTEBAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 10/29/2025 23:54:47

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.079	23:49
Air Blank		0.000	23:49
Subject Sample	# 1	0.146	23:50
Breath Volume		2.902	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:52
Subject Sample	# 2	0.147	23:53
Breath Volume		3.171	23:53
Air Blank		0.000	23:53
Control Test		0.078	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 11/01/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,TY,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,TY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 10/31/2025 20:39:54

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.079	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.222	20:35
Breath Volume		1.910	20:35
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.224	20:38
Breath Volume		1.285	20:38
Air Blank		0.000	20:38
Control Test		0.078	20:39
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 11/01/2025 13:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASLESON,MARSHALL,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ASLESON,MARSHALL
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 10/05/2025 02:38:29

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.159	02:33
Breath Volume		3.769	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.144	02:36
Breath Volume		1.750	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUPPY,CALEB,J
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: NAVARRO,BRYAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BRADSHAW

DATE/TIME OF TEST: 10/05/2025 03:44:36

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:30	
Control Test		0.079	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 1	0.192	03:31	
Breath Volume		1.386	03:31	
Air Blank		0.000	03:32	
Air Blank		0.000	03:34	
Subject Sample	# 2		03:37	No Sample Provided
Breath Volume		0.000	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 3		03:43	No Sample Provided
Breath Volume		0.000	03:43	
Air Blank		0.000	03:43	
Control Test		0.079	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASNICK,MORGAN,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RASNICK,MORGAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: EVERSON

DATE/TIME OF TEST: 10/05/2025 05:18:20

Diagnostics Check		OK	05:11
Air Blank		0.000	05:12
Control Test		0.080	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.134	05:13
Breath Volume		1.898	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.135	05:16
Breath Volume		2.179	05:16
Air Blank		0.000	05:17
Control Test		0.079	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PEREZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: CLEAREZIA

DATE/TIME OF TEST: 10/05/2025 07:13:12

Diagnostics Check		OK	07:10	
Air Blank		0.000	07:10	
Control Test		0.079	07:10	
Air Blank		0.000	07:11	
Subject Sample	# 1		07:11	RFI Detect
Breath Volume		0.000	07:11	
Air Blank		0.000	07:12	
Control Test		0.079	07:12	
Air Blank		0.000	07:13	
Diagnostics Check		OK	07:13	

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,PEREZ,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PEREZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: CLEAREZIA

DATE/TIME OF TEST: 10/05/2025 07:21:53

Diagnostics Check		OK	07:18	
Air Blank		0.000	07:19	
Control Test		0.080	07:19	
Air Blank		0.000	07:20	
Subject Sample	# 1		07:20	Subject Test Refused
Breath Volume		0.000	07:20	
Air Blank		0.000	07:20	
Control Test		0.080	07:21	
Air Blank		0.000	07:21	
Diagnostics Check		OK	07:21	

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERT,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PEREZ,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: CLEAREZIA

DATE/TIME OF TEST: 10/05/2025 07:34:14

Diagnostics Check		OK	07:26
Air Blank		0.000	07:26
Control Test		0.080	07:27
Air Blank		0.000	07:27
Subject Sample	# 1	0.128	07:29
Breath Volume		1.523	07:29
Air Blank		0.000	07:30
Air Blank		0.000	07:31
Subject Sample	# 2	0.130	07:32
Breath Volume		2.199	07:32
Air Blank		0.000	07:33
Control Test		0.077	07:33
Air Blank		0.000	07:34
Diagnostics Check		OK	07:34

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,JESSICA,V
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,JESSICA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: LAMOTTE

DATE/TIME OF TEST: 10/11/2025 05:48:03

Diagnostics Check		OK	05:41
Air Blank		0.000	05:41
Control Test		0.080	05:42
Air Blank		0.000	05:42
Subject Sample	# 1	0.129	05:43
Breath Volume		1.703	05:43
Air Blank		0.000	05:44
Air Blank		0.000	05:45
Subject Sample	# 2	0.131	05:46
Breath Volume		1.304	05:46
Air Blank		0.000	05:47
Control Test		0.079	05:47
Air Blank		0.000	05:47
Diagnostics Check		OK	05:48

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,COLT,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HATCH,VIRGINIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 10/11/2025 22:53:20

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:50	
Control Test		0.080	22:50	
Air Blank		0.000	22:51	
Subject Sample	# 1	0.055	22:51	
Breath Volume		2.320	22:51	
Air Blank			22:51	Improper Sample
Air Blank		0.000	22:52	
Control Test		0.080	22:52	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:53	

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,COLT,N
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HATCH,VIRGINIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 10/11/2025 23:02:52

Diagnostics Check		OK	22:56
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.055	22:58
Breath Volume		1.578	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:00
Subject Sample	# 2	0.053	23:01
Breath Volume		1.687	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEASMAN,MONTGOMERY,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PEREZ,ALEXANDER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 10:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:27

SUBJECT'S LAST NAME: DELBROOK

DATE/TIME OF TEST: 10/14/2025 12:01:48

Diagnostics Check		OK	11:54
Air Blank		0.000	11:55
Control Test		0.080	11:55
Air Blank		0.000	11:55
Subject Sample	# 1	0.203	11:56
Breath Volume		2.542	11:56
Air Blank		0.000	11:57
Air Blank		0.000	11:59
Subject Sample	# 2	0.189	11:59
Breath Volume		1.441	11:59
Air Blank		0.000	12:00
Control Test		0.077	12:01
Air Blank		0.000	12:01
Diagnostics Check		OK	12:01

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVARRO,BRYAN,
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: NAVARRO,BRYAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/18/2025 03:41:10

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.189	03:36
Breath Volume		2.054	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:38
Subject Sample	# 2	0.193	03:39
Breath Volume		3.046	03:39
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUDER,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SCHUDER,JUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/19/2025 19:25:24

Diagnostics Check		OK	19:18
Air Blank		0.000	19:19
Control Test		0.080	19:19
Air Blank		0.000	19:20
Subject Sample	# 1	0.186	19:20
Breath Volume		2.050	19:20
Air Blank		0.000	19:21
Air Blank		0.000	19:23
Subject Sample	# 2	0.188	19:23
Breath Volume		1.859	19:23
Air Blank		0.000	19:24
Control Test		0.078	19:24
Air Blank		0.000	19:25
Diagnostics Check		OK	19:25

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SURAJBALI,KYLE,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SURAJBALI,KYLE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 10/20/2025 04:00:24

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:55
Subject Sample	# 1	0.159	03:55
Breath Volume		2.851	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:57
Subject Sample	# 2	0.160	03:58
Breath Volume		3.441	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRISCOE,TERRY,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: BRISCOE,TERRY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/20/2025 21:17:07

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.080	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1	0.225	21:12	
Breath Volume		4.359	21:12	
Air Blank		0.000	21:13	
Air Blank		0.000	21:15	
Subject Sample	# 2	0.199	21:15	
Breath Volume		3.015	21:15	
Air Blank			21:16	Purge Fail
Air Blank		0.000	21:17	

Data Download Date/Time: 10/21/2025 12:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABBE,KYLER,S
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LABBE,KYLER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/20/2025 21:42:09

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.074	21:41	Control Outside Tolerance
Air Blank	0.000	21:42	

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABBE,KYLER,S
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LABBE,KYLER
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/20/2025 21:50:54

Diagnostics Check		OK	21:44
Air Blank		0.000	21:44
Control Test		0.078	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.134	21:46
Breath Volume		2.058	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:48
Subject Sample	# 2	0.129	21:49
Breath Volume		1.988	21:49
Air Blank		0.000	21:49
Control Test		0.078	21:50
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RIDDLE,BRIANNA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 10/20/2025 22:41:30

Diagnostics Check		OK	22:33	
Air Blank		0.000	22:33	
Control Test		0.079	22:34	
Air Blank		0.000	22:34	
Subject Sample	# 1	0.143	22:36	
Breath Volume		2.003	22:36	
Air Blank		0.000	22:37	
Air Blank		0.000	22:38	
Subject Sample	# 2	0.136	22:40	
Breath Volume		1.253	22:40	
Air Blank			22:41	Purge Fail
Air Blank		0.000	22:41	

Data Download Date/Time: 10/21/2025 12:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDDLE,BRIANNA,E
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: RIDDLE,BRIANNA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 10/20/2025 22:59:42

Diagnostics Check		OK	22:52	
Air Blank		0.000	22:53	
Control Test		0.077	22:53	
Air Blank		0.000	22:54	
Subject Sample	# 1	0.140	22:54	
Breath Volume		1.640	22:54	
Air Blank		0.000	22:55	
Air Blank		0.000	22:57	
Subject Sample	# 2	0.136	22:57	
Breath Volume		1.660	22:57	
Air Blank		0.000	22:58	
Control Test		0.078	22:58	
Air Blank			22:59	RFI Detect
Air Blank		0.000	22:59	

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOCKETT-WRIGHT,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: CARDOSO,C
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: TOOLE

DATE/TIME OF TEST: 10/02/2025 20:17:16

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.079	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.000	20:12
Breath Volume		2.531	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.000	20:15
Breath Volume		3.941	20:15
Air Blank		0.000	20:16
Control Test		0.080	20:16
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ANTHONY,M
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BAKSH

DATE/TIME OF TEST: 10/04/2025 02:16:19

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.079	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:14	Subject Test Refused
Breath Volume		0.000	02:14	
Air Blank		0.000	02:15	
Control Test		0.078	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,TARA,E
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: LUNA,R
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 10/06/2025 01:20:05

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:16	
Control Test		0.080	01:17	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:18	
Control Test		0.079	01:19	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUISSAINT,ROBERT,P
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: VARDAMAN,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: MEJIA ANCHETA

DATE/TIME OF TEST: 10/12/2025 01:54:26

Diagnostics Check		OK	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.179	01:49
Breath Volume		2.222	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:51
Subject Sample	# 2	0.182	01:52
Breath Volume		1.917	01:52
Air Blank		0.000	01:53
Control Test		0.077	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,BRIANNA,M
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BARAJAS,CARINA
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CERVANTES GARCIA

DATE/TIME OF TEST: 10/12/2025 03:59:30

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.075	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 1	0.164	03:50	Volume Not Met
Breath Volume		0.218	03:50	
Air Blank		0.000	03:51	
Air Blank		0.000	03:53	
Subject Sample	# 2	0.164	03:54	
Breath Volume		2.683	03:54	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 3	0.171	03:57	
Breath Volume		1.785	03:57	
Air Blank		0.000	03:58	
Control Test		0.077	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,BRIANNA,M
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LOWE,MAYCEE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: SHIFFLETT

DATE/TIME OF TEST: 10/12/2025 04:40:33

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:35	
Control Test		0.077	04:35	
Air Blank		0.000	04:36	
Subject Sample	# 1	0.208	04:39	
Breath Volume		2.617	04:39	
Air Blank			04:40	Purge Fail
Air Blank		0.000	04:40	

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,BRIANNA,M
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LOWE,MAYCEE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: SHIFFLETT

DATE/TIME OF TEST: 10/12/2025 04:57:49

Diagnostics Check		OK	04:49	
Air Blank		0.000	04:49	
Control Test		0.075	04:50	
Air Blank		0.000	04:50	
Subject Sample	# 1	0.210	04:53	Volume Not Met
Breath Volume		0.031	04:53	
Air Blank		0.000	04:54	
Air Blank		0.000	04:56	
Subject Sample	# 2		04:56	Subject Test Refused
Breath Volume		0.000	04:56	
Air Blank		0.000	04:56	
Control Test		0.076	04:57	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: LUNA,R
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/15/2025 01:08:51

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.079	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:07	
Control Test		0.078	01:07	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: VARDAMAN,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: TOTH

DATE/TIME OF TEST: 10/15/2025 21:37:16

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.080	21:30
Air Blank		0.000	21:31
Subject Sample	# 1	0.216	21:32
Breath Volume		1.535	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:34
Subject Sample	# 2	0.218	21:35
Breath Volume		1.507	21:35
Air Blank		0.000	21:36
Control Test		0.075	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACAVOY,J,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: WILLIS,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 10/19/2025 01:08:46

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.309	01:03
Breath Volume		1.460	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.292	01:06
Breath Volume		1.972	01:06
Air Blank		0.000	01:07
Control Test		0.075	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: VIANA,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: SKIPPER

DATE/TIME OF TEST: 10/22/2025 03:03:35

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.000	02:59
Breath Volume		2.187	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.000	03:02
Breath Volume		2.078	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: VIANA,A
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: RUDOLPH

DATE/TIME OF TEST: 10/24/2025 21:18:29

Diagnostics Check		OK	21:10
Air Blank		0.000	21:11
Control Test		0.079	21:11
Air Blank		0.000	21:12
Subject Sample	# 1	0.095	21:13
Breath Volume		2.453	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:15
Subject Sample	# 2	0.090	21:16
Breath Volume		2.167	21:16
Air Blank		0.000	21:17
Control Test		0.078	21:17
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: LOWE,M
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/25/2025 04:56:47

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.078	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.080	04:51
Breath Volume		1.808	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.076	04:54
Breath Volume		1.757	04:54
Air Blank		0.000	04:55
Control Test		0.079	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELS,JESSICA,N
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: VARDAMAN,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MAYO

DATE/TIME OF TEST: 10/26/2025 04:24:25

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.079	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.045	04:19
Breath Volume		2.710	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.046	04:22
Breath Volume		2.750	04:22
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELS,JESSICA,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ZOLLWEG,B
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: ABDOOL

DATE/TIME OF TEST: 10/26/2025 05:34:57

Diagnostics Check		OK	05:26
Air Blank		0.000	05:27
Control Test		0.080	05:27
Air Blank		0.000	05:28
Subject Sample	# 1	0.108	05:30
Breath Volume		2.710	05:30
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.113	05:33
Breath Volume		2.710	05:33
Air Blank		0.000	05:33
Control Test		0.080	05:34
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: LEWIS,B
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MARTINEZ PAGOADA

DATE/TIME OF TEST: 10/26/2025 18:40:19

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.079	18:31
Air Blank		0.000	18:31
Subject Sample	# 1	0.172	18:34
Breath Volume		1.265	18:34
Air Blank		0.000	18:35
Air Blank		0.000	18:36
Subject Sample	# 2	0.176	18:38
Breath Volume		1.667	18:38
Air Blank		0.000	18:39
Control Test		0.076	18:39
Air Blank		0.000	18:40
Diagnostics Check		OK	18:40

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: WILLIS,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GALLEGOS ACOSTA

DATE/TIME OF TEST: 10/26/2025 23:20:45

Diagnostics Check		OK	23:12
Air Blank		0.000	23:13
Control Test		0.079	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.150	23:16
Breath Volume		2.007	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.155	23:18
Breath Volume		1.675	23:18
Air Blank		0.000	23:19
Control Test		0.077	23:20
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 10/28/2025 12:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MANGO,BILL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 10/04/2025 03:20:01

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.113	03:14
Breath Volume		2.449	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.121	03:18
Breath Volume		1.714	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:20

Data Download Date/Time: 10/29/2025 18:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 10/15/2025 22:08:16

Diagnostics Check		OK	22:01	
Air Blank		0.000	22:01	
Control Test		0.081	22:02	
Air Blank		0.000	22:02	
Subject Sample	# 1	0.114	22:03	
Breath Volume		2.152	22:03	
Air Blank		0.000	22:03	
Air Blank		0.000	22:05	
Subject Sample	# 2	0.117	22:06	
Breath Volume		2.636	22:06	
Air Blank		0.000	22:06	
Control Test		0.080	22:07	
Air Blank		0.000	22:07	
Diagnostics Check		Fail	22:07	Internal Printer Test
Air Blank		0.000	22:08	

Data Download Date/Time: 10/29/2025 18:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 10/15/2025 22:16:35

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.081	22:10
Air Blank		0.000	22:11
Subject Sample	# 1	0.114	22:12
Breath Volume		1.796	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.109	22:15
Breath Volume		2.117	22:15
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 10/29/2025 18:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 10/22/2025 20:42:31

Diagnostics Check	Fail	20:42	Internal Printer Test
Air Blank	0.000	20:42	

Data Download Date/Time: 10/29/2025 18:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 10/22/2025 20:50:06

Diagnostics Check		OK	20:43
Air Blank		0.000	20:44
Control Test		0.081	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.111	20:45
Breath Volume		3.074	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:47
Subject Sample	# 2	0.131	20:48
Breath Volume		2.898	20:48
Air Blank		0.000	20:49
Control Test		0.080	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:50

Data Download Date/Time: 10/29/2025 18:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAHNSEN,A,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: KISHEL,M
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 10/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 10/29/2025 19:39:08

Diagnostics Check		OK	19:32
Air Blank		0.000	19:33
Control Test		0.080	19:33
Air Blank		0.000	19:33
Subject Sample	# 1	0.000	19:34
Breath Volume		1.984	19:34
Air Blank		0.000	19:34
Air Blank		0.000	19:36
Subject Sample	# 2	0.000	19:37
Breath Volume		1.718	19:37
Air Blank		0.000	19:38
Control Test		0.079	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 11/27/2025 22:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POLO,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: DELUCIA

DATE/TIME OF TEST: 10/03/2025 20:22:21

Diagnostics Check		OK	20:14
Air Blank		0.000	20:15
Control Test		0.080	20:15
Air Blank		0.000	20:16
Subject Sample	# 1	0.179	20:16
Breath Volume		2.078	20:16
Air Blank		0.000	20:17
Air Blank		0.000	20:19
Subject Sample	# 2	0.166	20:20
Breath Volume		1.835	20:20
Air Blank		0.000	20:21
Control Test		0.076	20:21
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 11/05/2025 09:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BACA

DATE/TIME OF TEST: 10/04/2025 03:01:11

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.221	02:55
Breath Volume		1.621	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.220	02:59
Breath Volume		4.128	02:59
Air Blank		0.000	03:00
Control Test		0.075	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 11/05/2025 09:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DE LA CAMPA,BRYANT
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: KOMAROV

DATE/TIME OF TEST: 10/04/2025 04:46:01

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.079	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.078	04:41
Breath Volume		1.765	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:43
Subject Sample	# 2	0.076	04:44
Breath Volume		1.507	04:44
Air Blank		0.000	04:45
Control Test		0.076	04:45
Air Blank		0.000	04:45
Diagnostics Check		OK	04:46

Data Download Date/Time: 11/05/2025 10:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARON,WILL
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MORENO SANCHEZ CONCH

DATE/TIME OF TEST: 10/04/2025 05:08:34

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:02	
Control Test		0.078	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 1	0.200	05:03	
Breath Volume		3.675	05:03	
Air Blank		0.000	05:04	
Air Blank		0.000	05:06	
Subject Sample	# 2	0.199	05:07	
Breath Volume		3.519	05:07	
Air Blank		0.000	05:07	
			05:08	RFI Detect
Air Blank		0.000	05:08	

Data Download Date/Time: 11/05/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARON,WILL
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: MORENO SANCHEZ CONCH

DATE/TIME OF TEST: 10/04/2025 05:17:07

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.078	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.208	05:12
Breath Volume		3.867	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.195	05:15
Breath Volume		3.589	05:15
Air Blank		0.000	05:16
Control Test		0.077	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 11/05/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARON,WILL
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: SALAVERRY GALLARDO

DATE/TIME OF TEST: 10/05/2025 07:21:03

Diagnostics Check		OK	07:13
Air Blank		0.000	07:13
Control Test		0.080	07:13
Air Blank		0.000	07:14
Subject Sample	# 1	0.160	07:16
Breath Volume		2.085	07:16
Air Blank		0.000	07:17
Air Blank		0.000	07:18
Subject Sample	# 2	0.167	07:19
Breath Volume		1.828	07:19
Air Blank		0.000	07:20
Control Test		0.075	07:20
Air Blank		0.000	07:20
Diagnostics Check		OK	07:21

Data Download Date/Time: 11/05/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 10/05/2025 22:54:10

Diagnostics Check		OK	22:45	
Air Blank		0.000	22:45	
Control Test		0.079	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1	0.142	22:48	
Breath Volume		1.347	22:48	
Air Blank		0.000	22:49	
Air Blank		0.000	22:51	
Subject Sample	# 2	0.140	22:52	
Breath Volume		1.523	22:52	
Air Blank		0.000	22:53	
Control Test		0.072	22:53	Control Outside Tolerance
Air Blank		0.000	22:54	

Data Download Date/Time: 11/05/2025 10:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WILKINSON

DATE/TIME OF TEST: 10/05/2025 23:03:47

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.076	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.139	22:58
Breath Volume		1.125	22:58
Air Blank		0.000	22:59
Air Blank		0.000	23:00
Subject Sample	# 2	0.145	23:01
Breath Volume		1.750	23:01
Air Blank		0.000	23:02
Control Test		0.077	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 11/05/2025 10:01

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOWE,RACHEL
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/06/2025 02:54:22

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:38	
Control Test		0.079	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1	0.174	02:42	Volume Not Met
Breath Volume		0.660	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 2	0.180	02:47	
Breath Volume		1.515	02:47	
Air Blank		0.000	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 3	0.178	02:52	
Breath Volume		1.246	02:52	
Air Blank		0.000	02:53	
Control Test		0.074	02:53	Control Outside Tolerance
Air Blank		0.000	02:54	

Data Download Date/Time: 11/05/2025 10:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOWE,RACHEL
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/06/2025 03:09:29

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.076	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1	0.164	03:00	Volume Not Met
Breath Volume		0.453	03:00	
Air Blank		0.000	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 2	0.179	03:03	
Breath Volume		1.718	03:03	
Air Blank		0.000	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 3	0.187	03:07	
Breath Volume		1.449	03:07	
Air Blank		0.000	03:08	
Control Test		0.077	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 11/05/2025 10:02

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TINDALL,TKEYAH
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ST SAUVEUR

DATE/TIME OF TEST: 10/08/2025 00:29:51

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.080	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.190	00:25
Breath Volume		1.503	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:27
Subject Sample	# 2	0.190	00:28
Breath Volume		1.761	00:28
Air Blank		0.000	00:28
Control Test		0.078	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 11/05/2025 10:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: UTER

DATE/TIME OF TEST: 10/11/2025 06:07:40

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.080	06:00
Air Blank		0.000	06:00
Subject Sample	# 1	0.074	06:02
Breath Volume		1.203	06:02
Air Blank		0.000	06:03
Air Blank		0.000	06:04
Subject Sample	# 2	0.072	06:05
Breath Volume		1.437	06:05
Air Blank		0.000	06:06
Control Test		0.077	06:07
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

Data Download Date/Time: 11/06/2025 03:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ANDRODE,DAVID
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 10/11/2025 09:03:26

Diagnostics Check		OK	08:55
Air Blank		0.000	08:56
Control Test		0.079	08:56
Air Blank		0.000	08:57
Subject Sample	# 1	0.165	08:58
Breath Volume		1.699	08:58
Air Blank		0.000	08:59
Air Blank		0.000	09:00
Subject Sample	# 2	0.161	09:01
Breath Volume		1.660	09:01
Air Blank		0.000	09:02
Control Test		0.078	09:02
Air Blank		0.000	09:03
Diagnostics Check		OK	09:03

Data Download Date/Time: 11/06/2025 03:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: DACAS

DATE/TIME OF TEST: 10/12/2025 00:13:19

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.080	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.123	00:08
Breath Volume		1.601	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:10
Subject Sample	# 2	0.125	00:11
Breath Volume		1.519	00:11
Air Blank		0.000	00:12
Control Test		0.078	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 11/06/2025 03:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GRAZIOSE,MADISON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/18/2025 19:45:54

Diagnostics Check		OK	19:35
Air Blank		0.000	19:35
Control Test		0.080	19:36
Air Blank		0.000	19:36
Subject Sample	# 1	0.199	19:38
Breath Volume		1.503	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.193	19:44
Breath Volume		2.636	19:44
Air Blank		0.000	19:44
Control Test		0.078	19:45
Air Blank		0.000	19:45
Diagnostics Check		OK	19:45

Data Download Date/Time: 11/06/2025 03:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: PRESSA

DATE/TIME OF TEST: 10/19/2025 01:24:15

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.121	01:19
Breath Volume		1.570	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:21
Subject Sample	# 2	0.120	01:22
Breath Volume		1.339	01:22
Air Blank		0.000	01:23
Control Test		0.078	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 11/06/2025 03:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: RUBIO MORENO

DATE/TIME OF TEST: 10/21/2025 03:28:15

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.089	03:23
Breath Volume		1.984	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:25
Subject Sample	# 2	0.095	03:26
Breath Volume		3.054	03:26
Air Blank		0.000	03:27
Control Test		0.078	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 11/06/2025 03:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007381

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VIGIL,MATTHEW
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 10/28/2025 20:50:20

Diagnostics Check		OK	20:42
Air Blank		0.000	20:42
Control Test		0.080	20:43
Air Blank		0.000	20:43
Subject Sample	# 1	0.282	20:44
Breath Volume		1.125	20:44
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.279	20:48
Breath Volume		1.417	20:48
Air Blank		0.000	20:49
Control Test		0.076	20:49
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 11/06/2025 03:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: FOGLESON

DATE/TIME OF TEST: 10/15/2025 04:40:08

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:37	
Control Test		0.079	04:38	
Air Blank		0.000	04:38	
Subject Sample	# 1		04:38	Subject Test Refused
Breath Volume		0.000	04:38	
Air Blank		0.000	04:39	
Control Test		0.079	04:39	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 10/31/2025 14:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 10/18/2025 03:18:47

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.094	03:14
Breath Volume		3.363	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:16
Subject Sample	# 2	0.096	03:17
Breath Volume		3.066	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 10/31/2025 14:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ROYER

DATE/TIME OF TEST: 10/20/2025 02:28:36

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.079	02:23
Air Blank		0.000	02:23
Subject Sample	# 1	0.135	02:24
Breath Volume		1.890	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.145	02:27
Breath Volume		2.214	02:27
Air Blank		0.000	02:27
Control Test		0.079	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 10/31/2025 14:06

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: SCHULT

DATE/TIME OF TEST: 10/04/2025 21:41:43

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.079	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.228	21:37
Breath Volume		3.507	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.223	21:39
Breath Volume		2.605	21:39
Air Blank		0.000	21:40
Control Test		0.077	21:41
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 11/06/2025 04:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: PICENO MENDEZ

DATE/TIME OF TEST: 10/27/2025 19:07:59

Diagnostics Check		OK	19:01	
Air Blank		0.000	19:02	
Control Test		0.078	19:02	
Air Blank		0.000	19:02	
Subject Sample	# 1	0.206	19:03	
Breath Volume		3.210	19:03	
Air Blank		0.000	19:04	
Air Blank		0.000	19:05	
Subject Sample	# 2	0.202	19:06	
Breath Volume		2.066	19:06	
Air Blank		0.000	19:07	
Control Test		0.074	19:07	Control Outside Tolerance
Air Blank		0.000	19:07	

Data Download Date/Time: 11/06/2025 04:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: PICENO MENDEZ

DATE/TIME OF TEST: 10/27/2025 19:14:34

Diagnostics Check	OK	19:13	
Air Blank		19:14	RFI Detect
Air Blank	0.000	19:14	

Data Download Date/Time: 11/06/2025 04:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: PICEN MENDEZ

DATE/TIME OF TEST: 10/27/2025 19:27:16

Diagnostics Check		OK	19:20	
Air Blank		0.000	19:20	
Control Test		0.076	19:21	
Air Blank		0.000	19:21	
Subject Sample	# 1	0.202	19:22	
Breath Volume		2.304	19:22	
Air Blank		0.000	19:23	
Air Blank		0.000	19:25	
Subject Sample	# 2	0.203	19:25	
Breath Volume		2.359	19:25	
Air Blank		0.000	19:26	
Control Test		0.074	19:26	Control Outside Tolerance
Air Blank		0.000	19:27	

Data Download Date/Time: 11/06/2025 04:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HERNANDEZ LOAIZA

DATE/TIME OF TEST: 10/01/2025 02:04:07

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.263	01:55
Breath Volume		1.757	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.241	01:58
Breath Volume		2.023	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 3	0.229	02:02
Breath Volume		1.445	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/01/2025 03:43:11

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.097	03:38
Breath Volume		2.343	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.096	03:41
Breath Volume		2.089	03:41
Air Blank		0.000	03:42
Control Test		0.080	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: HOLTON

DATE/TIME OF TEST: 10/01/2025 05:28:13

Diagnostics Check		OK	05:21
Air Blank		0.000	05:22
Control Test		0.079	05:22
Air Blank		0.000	05:23
Subject Sample	# 1	0.182	05:23
Breath Volume		3.453	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 2	0.181	05:26
Breath Volume		2.515	05:26
Air Blank		0.000	05:27
Control Test		0.080	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: BUTLER JR

DATE/TIME OF TEST: 10/01/2025 05:54:49

Diagnostics Check		OK	05:47
Air Blank		0.000	05:48
Control Test		0.079	05:48
Air Blank		0.000	05:49
Subject Sample	# 1	0.268	05:49
Breath Volume		1.554	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.275	05:53
Breath Volume		1.847	05:53
Air Blank		0.000	05:53
Control Test		0.077	05:54
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 10/02/2025 00:06:18

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:52	
Control Test		0.079	23:52	
Air Blank		0.000	23:53	
Subject Sample	# 3	0.174	23:56	Volume Not Met
Breath Volume		1.007	23:56	
Air Blank		0.000	23:57	
Air Blank		0.000	23:58	
Subject Sample	# 1	0.183	00:00	
Breath Volume		1.406	00:00	
Air Blank		0.000	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 2	0.181	00:04	
Breath Volume		1.105	00:04	
Air Blank		0.000	00:05	
Control Test		0.077	00:05	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JACOBS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: SWITZER

DATE/TIME OF TEST: 10/02/2025 02:59:22

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.227	02:54
Breath Volume		1.253	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.232	02:57
Breath Volume		1.410	02:57
Air Blank		0.000	02:58
Control Test		0.079	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 10/02/2025 03:43:21

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.172	03:38
Breath Volume		1.273	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.182	03:41
Breath Volume		2.609	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 10/03/2025 03:59:57

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.079	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.141	03:55
Breath Volume		1.652	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.151	03:58
Breath Volume		1.824	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 10/04/2025 01:16:49

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.080	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1	0.184	01:03	
Breath Volume		1.382	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2	0.188	01:09	Volume Not Met
Breath Volume		0.074	01:09	
Air Blank		0.000	01:10	
Air Blank		0.000	01:11	
Subject Sample	# 3	0.157	01:15	Volume Not Met
Breath Volume		0.015	01:15	
Air Blank		0.000	01:15	
Control Test		0.076	01:16	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PIERRE,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HENDRICKS

DATE/TIME OF TEST: 10/04/2025 01:41:34

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.077	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.085	01:36
Breath Volume		3.089	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.084	01:39
Breath Volume		3.269	01:39
Air Blank		0.000	01:40
Control Test		0.078	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: ARA HUET

DATE/TIME OF TEST: 10/04/2025 02:35:12

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.110	02:30
Breath Volume		1.960	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.108	02:33
Breath Volume		2.160	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ERNST II

DATE/TIME OF TEST: 10/04/2025 03:48:22

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.079	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1	0.205	03:39	Volume Not Met
Breath Volume		0.648	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 2	0.216	03:42	
Breath Volume		1.113	03:42	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 3	0.206	03:46	
Breath Volume		1.437	03:46	
Air Blank		0.000	03:47	
Control Test		0.078	03:47	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/04/2025 04:56:13

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.078	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.128	04:51
Breath Volume		1.824	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.123	04:54
Breath Volume		1.464	04:54
Air Blank		0.000	04:55
Control Test		0.078	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: HILERIO-GUTIERREZ

DATE/TIME OF TEST: 10/04/2025 05:23:46

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.079	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.070	05:18
Breath Volume		1.343	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.065	05:22
Breath Volume		1.660	05:22
Air Blank		0.000	05:22
Control Test		0.079	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/04/2025 22:40:07

Diagnostics Check		OK	22:33
Air Blank		0.000	22:33
Control Test		0.082	22:34
Air Blank		0.000	22:34
Subject Sample	# 1	0.227	22:35
Breath Volume		1.738	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.228	22:38
Breath Volume		1.429	22:38
Air Blank		0.000	22:39
Control Test		0.078	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: MARTINEZ LIMA

DATE/TIME OF TEST: 10/05/2025 00:35:43

Diagnostics Check		OK	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.158	00:31
Breath Volume		1.535	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.164	00:34
Breath Volume		1.695	00:34
Air Blank		0.000	00:34
Control Test		0.077	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CRUZ RODRIGUEZ

DATE/TIME OF TEST: 10/05/2025 02:04:43

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.102	01:59
Breath Volume		1.410	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.105	02:03
Breath Volume		1.406	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 10/05/2025 03:04:26

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.079	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.152	02:59
Breath Volume		2.523	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.142	03:02
Breath Volume		2.257	03:02
Air Blank		0.000	03:03
Control Test		0.079	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GREEN JR

DATE/TIME OF TEST: 10/05/2025 03:31:00

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.080	03:23	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:26	No Sample Provided
Breath Volume		0.000	03:26	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 2		03:29	Subject Test Refused
Breath Volume		0.000	03:29	
Air Blank		0.000	03:30	
Control Test		0.079	03:30	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: CRAVEN

DATE/TIME OF TEST: 10/05/2025 03:59:11

Diagnostics Check		OK	03:52
Air Blank		0.000	03:53
Control Test		0.079	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.142	03:54
Breath Volume		2.734	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.139	03:57
Breath Volume		2.722	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: JAHUEY HERNANDEZ

DATE/TIME OF TEST: 10/05/2025 04:25:36

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.183	04:20
Breath Volume		1.753	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.173	04:23
Breath Volume		1.300	04:23
Air Blank		0.000	04:24
Control Test		0.078	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: MCCULLOUGH

DATE/TIME OF TEST: 10/07/2025 06:21:18

Diagnostics Check		OK	06:16	
Air Blank		0.000	06:17	
Control Test		0.079	06:17	
Air Blank		0.000	06:18	
Subject Sample	# 1		06:19	Subject Test Refused
Breath Volume		0.000	06:19	
Air Blank		0.000	06:20	
Control Test		0.079	06:20	
Air Blank		0.000	06:21	
Diagnostics Check		OK	06:21	

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THEODORE,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: POP CUCUL

DATE/TIME OF TEST: 10/07/2025 20:54:08

Diagnostics Check		OK	20:45
Air Blank		0.000	20:46
Control Test		0.079	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.098	20:49
Breath Volume		1.386	20:49
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.094	20:52
Breath Volume		1.140	20:52
Air Blank		0.000	20:53
Control Test		0.078	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: PATRA

DATE/TIME OF TEST: 10/08/2025 03:13:25

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.079	02:59	
Air Blank		0.000	02:59	
Subject Sample	# 1	0.151	03:03	Slope Not Level
Breath Volume		1.378	03:03	
Air Blank		0.000	03:03	
Air Blank		0.000	03:05	
Subject Sample	# 2	0.155	03:05	
Breath Volume		1.347	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 3	0.167	03:11	Volume Not Met
Breath Volume		0.023	03:11	
Air Blank		0.000	03:12	
Control Test		0.078	03:12	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 10/16/2025 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: PATRA

DATE/TIME OF TEST: 10/08/2025 03:25:34

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.076	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.174	03:20
Breath Volume		1.636	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.159	03:23
Breath Volume		1.269	03:23
Air Blank		0.000	03:24
Control Test		0.075	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: COMM VEHIC
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: TEPFENHART

DATE/TIME OF TEST: 10/08/2025 23:18:35

Diagnostics Check		OK	23:11
Air Blank		0.000	23:12
Control Test		0.078	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.135	23:14
Breath Volume		2.847	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.137	23:16
Breath Volume		2.542	23:16
Air Blank		0.000	23:17
Control Test		0.078	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 10/09/2025 03:42:06

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.163	03:37
Breath Volume		1.507	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:39
Subject Sample	# 2	0.177	03:40
Breath Volume		1.714	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: ROBLES ROMERO

DATE/TIME OF TEST: 10/09/2025 21:36:20

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:24	
Control Test		0.079	21:24	
Air Blank		0.000	21:24	
Subject Sample	# 1	0.180	21:28	Volume Not Met
Breath Volume		0.093	21:28	
Air Blank		0.000	21:28	
Air Blank		0.000	21:30	
Subject Sample	# 2	0.189	21:31	
Breath Volume		1.347	21:31	
Air Blank		0.000	21:32	
Air Blank		0.000	21:34	
Subject Sample	# 3	0.192	21:34	
Breath Volume		2.882	21:34	
Air Blank		0.000	21:35	
Control Test		0.077	21:35	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: SCETTINI

DATE/TIME OF TEST: 10/10/2025 03:47:59

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.033	03:43
Breath Volume		2.679	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:45
Subject Sample	# 2	0.031	03:46
Breath Volume		1.640	03:46
Air Blank		0.000	03:47
Control Test		0.078	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUG,Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: OUSLEY

DATE/TIME OF TEST: 10/11/2025 03:25:04

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.079	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1	0.137	03:18	Slope Not Level
Breath Volume		1.179	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 2		03:23	Subject Test Refused
Breath Volume		1.027	03:23	
Air Blank		0.000	03:24	
Control Test		0.077	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: CHOUAIB

DATE/TIME OF TEST: 10/11/2025 04:52:35

Diagnostics Check		OK	04:41
Air Blank		0.000	04:41
Control Test		0.078	04:42
Air Blank		0.000	04:42
Subject Sample	# 1	0.187	04:45
Breath Volume		1.613	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.205	04:50
Breath Volume		1.613	04:50
Air Blank		0.000	04:51
Control Test		0.077	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVON,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:41

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/11/2025 15:13:24

Diagnostics Check		OK	15:01	
Air Blank		0.000	15:02	
Control Test		0.079	15:02	
Air Blank		0.000	15:03	
Subject Sample	# 1		15:06	No Sample Provided
Breath Volume		0.000	15:06	
Air Blank		0.000	15:06	
Air Blank		0.000	15:08	
Subject Sample	# 2		15:11	No Sample Provided
Breath Volume		0.000	15:11	
Air Blank		0.000	15:12	
Control Test		0.079	15:12	
Air Blank		0.000	15:13	
Diagnostics Check		OK	15:13	

Data Download Date/Time: 10/16/2025 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVON,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:41

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/11/2025 15:22:43

Diagnostics Check		OK	15:16
Air Blank		0.000	15:16
Control Test		0.079	15:16
Air Blank		0.000	15:17
Subject Sample	# 1	0.231	15:18
Breath Volume		2.160	15:18
Air Blank		0.000	15:18
Air Blank		0.000	15:20
Subject Sample	# 2	0.233	15:20
Breath Volume		1.960	15:20
Air Blank		0.000	15:21
Control Test		0.078	15:22
Air Blank		0.000	15:22
Diagnostics Check		OK	15:22

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/11/2025 23:44:54

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.079	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.180	23:39
Breath Volume		2.085	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.186	23:43
Breath Volume		2.386	23:43
Air Blank		0.000	23:43
Control Test		0.077	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 10/12/2025 00:40:31

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.078	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.121	00:35
Breath Volume		1.566	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.124	00:38
Breath Volume		1.726	00:38
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MAROCOLA,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: AUGUSTE

DATE/TIME OF TEST: 10/12/2025 03:02:11

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.079	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:55	No Sample Provided
Breath Volume		0.000	02:55	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 2		03:00	No Sample Provided
Breath Volume		0.000	03:00	
Air Blank		0.000	03:01	
Control Test		0.079	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 10/16/2025 12:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/12/2025 03:17:12

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.096	03:12
Breath Volume		1.140	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.092	03:15
Breath Volume		1.601	03:15
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 10/16/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 10/12/2025 06:20:25

Diagnostics Check		OK	06:13
Air Blank		0.000	06:13
Control Test		0.079	06:14
Air Blank		0.000	06:14
Subject Sample	# 1	0.116	06:15
Breath Volume		1.851	06:15
Air Blank		0.000	06:16
Air Blank		0.000	06:18
Subject Sample	# 2	0.112	06:18
Breath Volume		1.593	06:18
Air Blank		0.000	06:19
Control Test		0.079	06:19
Air Blank		0.000	06:20
Diagnostics Check		OK	06:20

Data Download Date/Time: 10/16/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 10/13/2025 01:17:43

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.253	01:12
Breath Volume		1.742	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.256	01:15
Breath Volume		1.839	01:15
Air Blank		0.000	01:16
Control Test		0.078	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 10/16/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SENTMANAT

DATE/TIME OF TEST: 10/13/2025 03:14:23

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.219	03:05
Breath Volume		1.234	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:07
Subject Sample	# 2	0.246	03:08
Breath Volume		1.359	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 3	0.223	03:12
Breath Volume		1.433	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 10/16/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 10/14/2025 04:14:38

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.158	04:09
Breath Volume		1.550	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.157	04:12
Breath Volume		1.582	04:12
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 10/16/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: RAYO ATEHORTUA

DATE/TIME OF TEST: 10/14/2025 06:00:53

Diagnostics Check		OK	05:50
Air Blank		0.000	05:51
Control Test		0.079	05:51
Air Blank		0.000	05:52
Subject Sample	# 1	0.267	05:52
Breath Volume		2.148	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.290	05:55
Breath Volume		1.585	05:55
Air Blank		0.000	05:56
Air Blank		0.000	05:58
Subject Sample	# 3	0.277	05:59
Breath Volume		1.343	05:59
Air Blank		0.000	05:59
Control Test		0.077	06:00
Air Blank		0.000	06:00
Diagnostics Check		OK	06:00

Data Download Date/Time: 10/16/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: JONES JR

DATE/TIME OF TEST: 10/14/2025 20:14:14

Diagnostics Check		OK	20:07
Air Blank		0.000	20:08
Control Test		0.079	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.103	20:09
Breath Volume		3.367	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:11
Subject Sample	# 2	0.104	20:12
Breath Volume		3.515	20:12
Air Blank		0.000	20:13
Control Test		0.078	20:13
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 10/16/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/14/2025 23:02:59

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.079	22:55
Air Blank		0.000	22:56
Subject Sample	# 1	0.153	22:58
Breath Volume		3.195	22:58
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.155	23:01
Breath Volume		3.050	23:01
Air Blank		0.000	23:02
Control Test		0.078	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 10/16/2025 12:22

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 10/30/2025 02:26:44

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.126	02:21
Breath Volume		1.300	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.120	02:25
Breath Volume		1.593	02:25
Air Blank		0.000	02:25
Control Test		0.080	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 11/06/2025 11:07

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BELEN,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: SILVA PEZO

DATE/TIME OF TEST: 10/16/2025 10:42:51

Diagnostics Check		OK	10:36
Air Blank		0.000	10:37
Control Test		0.079	10:37
Air Blank		0.000	10:38
Subject Sample	# 1	0.000	10:38
Breath Volume		2.359	10:38
Air Blank		0.000	10:39
Air Blank		0.000	10:41
Subject Sample	# 2	0.000	10:41
Breath Volume		2.347	10:41
Air Blank		0.000	10:41
Control Test		0.078	10:42
Air Blank		0.000	10:42
Diagnostics Check		OK	10:42

Data Download Date/Time: 11/06/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: LAYMAN

DATE/TIME OF TEST: 10/17/2025 02:10:48

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.148	02:06
Breath Volume		2.898	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.150	02:09
Breath Volume		1.773	02:09
Air Blank		0.000	02:09
Control Test		0.079	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 11/06/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CHANCEY

DATE/TIME OF TEST: 10/18/2025 04:19:51

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.289	04:15
Breath Volume		1.933	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.288	04:18
Breath Volume		2.027	04:18
Air Blank		0.000	04:18
Control Test		0.078	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 11/06/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORSEN,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/18/2025 05:18:54

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.079	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.153	05:14
Breath Volume		1.648	05:14
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.159	05:17
Breath Volume		1.800	05:17
Air Blank		0.000	05:17
Control Test		0.079	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 11/06/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: MARTINEZ VALLEJO

DATE/TIME OF TEST: 10/18/2025 05:53:20

Diagnostics Check		OK	05:46
Air Blank		0.000	05:47
Control Test		0.079	05:47
Air Blank		0.000	05:48
Subject Sample	# 1	0.070	05:48
Breath Volume		1.921	05:48
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 2	0.072	05:51
Breath Volume		3.085	05:51
Air Blank		0.000	05:52
Control Test		0.079	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 11/06/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: ALVAREZ SANCHO

DATE/TIME OF TEST: 10/18/2025 07:06:29

Diagnostics Check		OK	06:58
Air Blank		0.000	06:58
Control Test		0.079	06:58
Air Blank		0.000	06:59
Subject Sample	# 1	0.140	07:01
Breath Volume		1.406	07:01
Air Blank		0.000	07:01
Air Blank		0.000	07:03
Subject Sample	# 2	0.144	07:04
Breath Volume		1.507	07:04
Air Blank		0.000	07:05
Control Test		0.079	07:05
Air Blank		0.000	07:06
Diagnostics Check		OK	07:06

Data Download Date/Time: 11/06/2025 12:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: BELCHER

DATE/TIME OF TEST: 10/18/2025 23:02:42

Diagnostics Check		OK	22:59	
Air Blank		0.000	23:00	
Control Test		0.080	23:00	
Air Blank		0.000	23:00	
Subject Sample	# 1		23:01	Subject Test Refused
Breath Volume		0.000	23:01	
Air Blank		0.000	23:01	
Control Test		0.079	23:02	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 11/06/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: BENITEZ MORALES

DATE/TIME OF TEST: 10/18/2025 23:21:37

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.000	23:17
Breath Volume		3.128	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.000	23:20
Breath Volume		2.562	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 11/06/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: RAMTAHAL SINGH

DATE/TIME OF TEST: 10/18/2025 23:52:37

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.079	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.175	23:47
Breath Volume		1.277	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.175	23:51
Breath Volume		1.925	23:51
Air Blank		0.000	23:51
Control Test		0.079	23:52
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 11/06/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRAME,C
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 10/19/2025 02:18:30

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.107	02:13
Breath Volume		1.722	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.110	02:16
Breath Volume		1.507	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 11/06/2025 12:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: VAZQUEZ GOMEZ

DATE/TIME OF TEST: 10/19/2025 02:38:03

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.079	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.209	02:33
Breath Volume		2.789	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.213	02:36
Breath Volume		2.574	02:36
Air Blank		0.000	02:37
Control Test		0.079	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:38

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELEZ,A
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/19/2025 04:07:26

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.041	04:02
Breath Volume		3.257	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.040	04:05
Breath Volume		3.187	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: DAMBECK

DATE/TIME OF TEST: 10/19/2025 05:01:13

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.079	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.273	04:53
Breath Volume		1.312	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.303	04:56
Breath Volume		2.296	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 3	0.292	04:59
Breath Volume		1.722	04:59
Air Blank		0.000	05:00
Control Test		0.078	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: MENDIVELSO FERNANDEZ

DATE/TIME OF TEST: 10/19/2025 06:21:27

Diagnostics Check		OK	06:14
Air Blank		0.000	06:15
Control Test		0.079	06:15
Air Blank		0.000	06:15
Subject Sample	# 1	0.111	06:16
Breath Volume		1.437	06:16
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.107	06:19
Breath Volume		1.468	06:19
Air Blank		0.000	06:20
Control Test		0.078	06:20
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CRUZ-BONILLA

DATE/TIME OF TEST: 10/20/2025 00:45:01

Diagnostics Check		OK	00:38
Air Blank		0.000	00:38
Control Test		0.080	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.125	00:40
Breath Volume		2.218	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:42
Subject Sample	# 2	0.124	00:43
Breath Volume		1.628	00:43
Air Blank		0.000	00:44
Control Test		0.078	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:45

Data Download Date/Time: 11/06/2025 12:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: BRUTUS

DATE/TIME OF TEST: 10/20/2025 03:45:45

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:34	
Control Test		0.079	03:35	
Air Blank		0.000	03:35	
Subject Sample	# 1		03:38	No Sample Provided
Breath Volume		0.000	03:38	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 2		03:44	No Sample Provided
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.079	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: COMM VEHIC
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/21/2025 00:38:38

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.238	00:34
Breath Volume		2.835	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.234	00:37
Breath Volume		2.089	00:37
Air Blank		0.000	00:37
Control Test		0.078	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 10/21/2025 03:45:45

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.079	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.158	03:41
Breath Volume		2.714	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.156	03:44
Breath Volume		2.683	03:44
Air Blank		0.000	03:44
Control Test		0.078	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 10/23/2025 01:54:46

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.214	01:49
Breath Volume		2.675	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:51
Subject Sample	# 2	0.199	01:53
Breath Volume		1.792	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 10/23/2025 04:16:31

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.079	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.187	04:11
Breath Volume		2.042	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.192	04:14
Breath Volume		1.515	04:14
Air Blank		0.000	04:15
Control Test		0.078	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TORRES,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 10/23/2025 13:32:25

Diagnostics Check		OK	13:26
Air Blank		0.000	13:26
Control Test		0.080	13:26
Air Blank		0.000	13:27
Subject Sample	# 1	0.000	13:28
Breath Volume		2.765	13:28
Air Blank		0.000	13:28
Air Blank		0.000	13:30
Subject Sample	# 2	0.000	13:31
Breath Volume		2.640	13:31
Air Blank		0.000	13:31
Control Test		0.079	13:31
Air Blank		0.000	13:32
Diagnostics Check		OK	13:32

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: AREVALO

DATE/TIME OF TEST: 10/23/2025 20:51:40

Diagnostics Check		OK	20:39	
Air Blank		0.000	20:40	
Control Test		0.079	20:40	
Air Blank		0.000	20:41	
Subject Sample	# 1	0.145	20:44	Volume Not Met
Breath Volume		0.378	20:44	
Air Blank		0.000	20:45	
Air Blank		0.000	20:46	
Subject Sample	# 2	0.149	20:50	Volume Not Met
Breath Volume		0.207	20:50	
Air Blank		0.000	20:50	
Control Test		0.079	20:51	
Air Blank		0.000	20:51	
Diagnostics Check		OK	20:51	

Data Download Date/Time: 11/06/2025 12:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: AREVALO

DATE/TIME OF TEST: 10/23/2025 21:07:04

Diagnostics Check		OK	20:55	
Air Blank		0.000	20:55	
Control Test		0.078	20:56	
Air Blank		0.000	20:56	
Subject Sample	# 1	0.149	20:59	Volume Not Met
Breath Volume		0.207	20:59	
Air Blank		0.000	21:00	
Air Blank		0.000	21:02	
Subject Sample	# 2	0.154	21:05	Volume Not Met
Breath Volume		0.218	21:05	
Air Blank		0.000	21:06	
Control Test		0.078	21:06	
Air Blank		0.000	21:06	
Diagnostics Check		OK	21:07	

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 10/23/2025 22:35:47

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.079	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.010	22:31
Breath Volume		2.000	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.011	22:34
Breath Volume		2.125	22:34
Air Blank		0.000	22:34
Control Test		0.079	22:35
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 10/24/2025 02:30:27

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.157	02:25
Breath Volume		1.691	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.171	02:28
Breath Volume		2.304	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUG,Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: HOLLAND JR

DATE/TIME OF TEST: 10/25/2025 00:04:24

Diagnostics Check		OK	23:49	
Air Blank		0.000	23:49	
Control Test		0.080	23:50	
Air Blank		0.000	23:50	
Subject Sample	# 2	0.130	23:53	Volume Not Met
Breath Volume		0.101	23:53	
Air Blank		0.000	23:54	
Air Blank		0.000	23:56	
Subject Sample	# 3	0.121	23:57	
Breath Volume		1.527	23:57	
Air Blank		0.000	23:57	
Air Blank		0.000	23:59	
Subject Sample	# 1	0.125	00:02	Volume Not Met
Breath Volume		0.023	00:02	
Air Blank		0.000	00:03	
Control Test		0.078	00:03	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: PORTALES HERNANDEZ

DATE/TIME OF TEST: 10/25/2025 01:19:46

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.149	01:14
Breath Volume		1.558	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.150	01:18
Breath Volume		2.710	01:18
Air Blank		0.000	01:18
Control Test		0.078	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 11/06/2025 12:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CALLOWAY

DATE/TIME OF TEST: 10/25/2025 03:36:02

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.157	03:30
Breath Volume		1.113	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.159	03:34
Breath Volume		2.078	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:36

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STANTON,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 10/25/2025 04:14:40

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.204	04:10
Breath Volume		1.511	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.206	04:13
Breath Volume		1.871	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: FRANTZ

DATE/TIME OF TEST: 10/25/2025 21:46:27

Diagnostics Check		OK	21:33	
Air Blank		0.000	21:34	
Control Test		0.080	21:34	
Air Blank		0.000	21:34	
Subject Sample	# 1	0.148	21:37	Slope Not Level
Breath Volume		1.621	21:37	
Air Blank		0.000	21:38	
Air Blank		0.000	21:40	
Subject Sample	# 2	0.164	21:41	
Breath Volume		2.000	21:41	
Air Blank		0.000	21:42	
Air Blank		0.000	21:44	
Subject Sample	# 3	0.162	21:44	
Breath Volume		1.683	21:44	
Air Blank		0.000	21:45	
Control Test		0.079	21:45	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: ESPINOZA COELLO

DATE/TIME OF TEST: 10/25/2025 22:09:27

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.079	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.139	22:05
Breath Volume		2.171	22:05
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.132	22:07
Breath Volume		1.601	22:07
Air Blank		0.000	22:08
Control Test		0.079	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 10/26/2025 00:06:09

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.129	00:01
Breath Volume		3.316	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:03
Subject Sample	# 2	0.132	00:04
Breath Volume		3.582	00:04
Air Blank		0.000	00:05
Control Test		0.078	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FELICIANO LOPEZ

DATE/TIME OF TEST: 10/26/2025 01:07:44

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.210	01:03
Breath Volume		1.988	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.201	01:06
Breath Volume		2.742	01:06
Air Blank		0.000	01:06
Control Test		0.079	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BARAHONA

DATE/TIME OF TEST: 10/26/2025 02:28:41

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.171	02:24
Breath Volume		3.441	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.166	02:27
Breath Volume		3.312	02:27
Air Blank		0.000	02:27
Control Test		0.078	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 10/26/2025 04:07:12

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.079	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:06	
Control Test		0.079	04:06	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: POPPELL

DATE/TIME OF TEST: 10/26/2025 23:42:30

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.079	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.000	23:37
Breath Volume		1.457	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.000	23:41
Breath Volume		1.843	23:41
Air Blank		0.000	23:41
Control Test		0.078	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/27/2025 00:14:15

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.158	00:07
Breath Volume		1.363	00:07
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.160	00:12
Breath Volume		1.589	00:12
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ROJAS SUAREZ

DATE/TIME OF TEST: 10/27/2025 02:46:25

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.080	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.090	02:41
Breath Volume		1.121	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:43
Subject Sample	# 2	0.095	02:44
Breath Volume		2.531	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BALZA RODRIGUEZ

DATE/TIME OF TEST: 10/27/2025 04:27:21

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.149	04:22
Breath Volume		2.085	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.156	04:25
Breath Volume		2.000	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 11/06/2025 12:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: DELGADO QUINONES

DATE/TIME OF TEST: 10/29/2025 01:19:04

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.179	01:14
Breath Volume		2.292	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.180	01:17
Breath Volume		1.675	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:19

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/30/2025 03:17:43

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.309	03:13
Breath Volume		1.214	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.291	03:16
Breath Volume		1.382	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: EHR

DATE/TIME OF TEST: 10/30/2025 07:22:05

Diagnostics Check		OK	07:15
Air Blank		0.000	07:16
Control Test		0.080	07:16
Air Blank		0.000	07:16
Subject Sample	# 1	0.100	07:17
Breath Volume		1.687	07:17
Air Blank		0.000	07:18
Air Blank		0.000	07:19
Subject Sample	# 2	0.089	07:20
Breath Volume		1.742	07:20
Air Blank		0.000	07:21
Control Test		0.079	07:21
Air Blank		0.000	07:21
Diagnostics Check		OK	07:22

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THOMAS,K
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 10/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: LAWRENCE

DATE/TIME OF TEST: 10/31/2025 17:15:57

Diagnostics Check		OK	17:08
Air Blank		0.000	17:09
Control Test		0.080	17:09
Air Blank		0.000	17:09
Subject Sample	# 1	0.206	17:11
Breath Volume		2.015	17:11
Air Blank		0.000	17:11
Air Blank		0.000	17:13
Subject Sample	# 2	0.210	17:14
Breath Volume		2.113	17:14
Air Blank		0.000	17:15
Control Test		0.079	17:15
Air Blank		0.000	17:15
Diagnostics Check		OK	17:15

Data Download Date/Time: 11/06/2025 12:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MCLELLAND

DATE/TIME OF TEST: 10/25/2025 03:40:22

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.078	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.149	03:35
Breath Volume		3.089	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.146	03:38
Breath Volume		2.812	03:38
Air Blank		0.000	03:39
Control Test		0.075	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 10/30/2025 06:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MARTIN,SAMUEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/26/2025 00:47:53

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.079	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.183	00:43
Breath Volume		1.589	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:45
Subject Sample	# 2	0.199	00:46
Breath Volume		1.253	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 10/30/2025 06:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: LOPEZ PERKNS

DATE/TIME OF TEST: 10/27/2025 04:47:36

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.081	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:40	No Sample Provided
Breath Volume		0.000	04:40	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 2		04:46	No Sample Provided
Breath Volume		0.000	04:46	
Air Blank		0.000	04:46	
Control Test		0.080	04:46	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 10/30/2025 06:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MONSIVAIS ARIAS

DATE/TIME OF TEST: 10/27/2025 20:48:19

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:31	
Control Test		0.081	20:32	
Air Blank		0.000	20:32	
Subject Sample	# 1	0.277	20:35	No .020 agreement
Breath Volume		2.671	20:35	
Air Blank		0.000	20:36	
Air Blank		0.000	20:38	
Subject Sample	# 2	0.286	20:41	Volume Not Met
Breath Volume		0.062	20:41	
Air Blank		0.000	20:42	
Air Blank		0.000	20:43	
Subject Sample	# 3	0.303	20:46	No .020 agreement
Breath Volume		2.140	20:46	
Air Blank		0.000	20:47	
Control Test		0.075	20:47	
Air Blank		0.000	20:48	
Diagnostics Check		OK	20:48	

Data Download Date/Time: 10/30/2025 06:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GARCIA,LUIS
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 10/28/2025 21:44:14

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:33	
Control Test		0.081	21:33	
Air Blank		0.000	21:34	
Subject Sample	# 1		21:37	No Sample Provided
Breath Volume		0.000	21:37	
Air Blank		0.000	21:37	
Air Blank		0.000	21:39	
Subject Sample	# 2		21:42	No Sample Provided
Breath Volume		0.000	21:42	
Air Blank		0.000	21:43	
Control Test		0.081	21:43	
Air Blank		0.000	21:44	
Diagnostics Check		OK	21:44	

Data Download Date/Time: 10/30/2025 06:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 10/29/2025 02:20:57

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:17	
Control Test		0.081	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:19	RFI Detect
Breath Volume		0.375	02:19	
Air Blank		0.000	02:20	
Control Test		0.080	02:20	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 10/30/2025 06:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 10/29/2025 02:29:43

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.080	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.164	02:24
Breath Volume		1.304	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:26
Subject Sample	# 2	0.160	02:28
Breath Volume		1.132	02:28
Air Blank		0.000	02:28
Control Test		0.080	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 10/30/2025 06:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JAROENWAI,WASAN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/29/2025 03:55:02

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:44	
Control Test		0.081	03:44	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:48	No Sample Provided
Breath Volume		0.000	03:48	
Air Blank		0.000	03:48	
Air Blank		0.000	03:50	
Subject Sample	# 2		03:53	No Sample Provided
Breath Volume		0.000	03:53	
Air Blank		0.000	03:54	
Control Test		0.080	03:54	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 10/30/2025 06:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 10/29/2025 04:21:24

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.171	04:16
Breath Volume		4.125	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.170	04:19
Breath Volume		3.445	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 10/30/2025 06:21

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,BENALY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLASER,STEVEN
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 10/30/2025 21:33:42

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.203	21:28
Breath Volume		1.621	21:28
Air Blank		0.000	21:29
Air Blank		0.000	21:30
Subject Sample	# 2	0.194	21:32
Breath Volume		1.281	21:32
Air Blank		0.000	21:32
Control Test		0.080	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 11/22/2025 05:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,BENALY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: PAUL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CROWE

DATE/TIME OF TEST: 10/30/2025 23:40:54

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:29	
Control Test		0.081	23:30	
Air Blank		0.000	23:30	
Subject Sample	# 1		23:33	No Sample Provided
Breath Volume		0.000	23:33	
Air Blank		0.000	23:34	
Air Blank		0.000	23:36	
Subject Sample	# 2		23:39	No Sample Provided
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.080	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 11/22/2025 05:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/31/2025 19:06:54

Diagnostics Check		OK	18:59
Air Blank		0.000	18:59
Control Test		0.080	18:59
Air Blank		0.000	19:00
Subject Sample	# 1	0.223	19:01
Breath Volume		1.910	19:01
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.219	19:05
Breath Volume		1.777	19:05
Air Blank		0.000	19:05
Control Test		0.078	19:06
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 11/22/2025 05:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,AARON,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HINSON,AARON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 10/31/2025 21:06:17

Diagnostics Check		OK	20:54	
Air Blank		0.000	20:55	
Control Test		0.080	20:55	
Air Blank		0.000	20:56	
Subject Sample	# 1		20:59	No Sample Provided
Breath Volume		0.000	20:59	
Air Blank		0.000	20:59	
Air Blank		0.000	21:01	
Subject Sample	# 2		21:04	No Sample Provided
Breath Volume		0.000	21:04	
Air Blank		0.000	21:05	
Control Test		0.080	21:05	
Air Blank		0.000	21:06	
Diagnostics Check		OK	21:06	

Data Download Date/Time: 11/22/2025 05:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 10/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: LORA GUTIERREZ

DATE/TIME OF TEST: 10/31/2025 22:45:32

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.080	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.121	22:40
Breath Volume		1.519	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:42
Subject Sample	# 2	0.119	22:43
Breath Volume		1.449	22:43
Air Blank		0.000	22:44
Control Test		0.079	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 11/22/2025 05:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: MONSIVAIS ARIAS

DATE/TIME OF TEST: 10/27/2025 19:37:59

Diagnostics Check	Fail	19:37	Internal Printer Test
Air Blank	0.000	19:37	

Data Download Date/Time: 10/30/2025 06:16

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJA,YOSMANY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 10/01/2025 00:20:13

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.080	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.000	00:16
Breath Volume		1.613	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.000	00:18
Breath Volume		1.742	00:18
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJA,YOSMANY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 10/04/2025 03:34:15

Diagnostics Check		OK	03:31	
Air Blank		0.000	03:31	
Control Test		0.081	03:32	
Air Blank		0.000	03:32	
Subject Sample	# 1		03:32	Subject Test Refused
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:34	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 10:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: STOCKS-GOODRICH

DATE/TIME OF TEST: 10/04/2025 11:28:45

Diagnostics Check		OK	11:22
Air Blank		0.000	11:23
Control Test		0.081	11:23
Air Blank		0.000	11:23
Subject Sample	# 1	0.000	11:24
Breath Volume		1.734	11:24
Air Blank		0.000	11:24
Air Blank		0.000	11:26
Subject Sample	# 2	0.000	11:27
Breath Volume		1.503	11:27
Air Blank		0.000	11:27
Control Test		0.082	11:28
Air Blank		0.000	11:28
Diagnostics Check		OK	11:28

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,WESLEY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: JONES,WESLEY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LUTZ

DATE/TIME OF TEST: 10/05/2025 02:26:47

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.013	02:22
Breath Volume		2.058	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:24
Subject Sample	# 2	0.011	02:25
Breath Volume		2.597	02:25
Air Blank		0.000	02:25
Control Test		0.080	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: STOCKS-GOODRICH

DATE/TIME OF TEST: 10/09/2025 17:13:54

Diagnostics Check		OK	17:07
Air Blank		0.000	17:07
Control Test		0.081	17:08
Air Blank		0.000	17:08
Subject Sample	# 1	0.000	17:09
Breath Volume		1.707	17:09
Air Blank		0.000	17:09
Air Blank		0.000	17:11
Subject Sample	# 2	0.000	17:12
Breath Volume		1.511	17:12
Air Blank		0.000	17:12
Control Test		0.081	17:13
Air Blank		0.000	17:13
Diagnostics Check		OK	17:13

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PIERRE-PAUL

DATE/TIME OF TEST: 10/10/2025 22:50:19

Diagnostics Check		OK	22:43
Air Blank		0.000	22:43
Control Test		0.081	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.097	22:45
Breath Volume		1.417	22:45
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.087	22:48
Breath Volume		1.414	22:48
Air Blank		0.000	22:49
Control Test		0.082	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LISTER,DALTON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/11/2025 02:57:07

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.105	02:52
Breath Volume		2.195	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.108	02:55
Breath Volume		1.750	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 10/12/2025 17:23:56

Diagnostics Check		OK	17:18
Air Blank		0.000	17:18
Control Test		0.081	17:18
Air Blank		0.000	17:19
Subject Sample	# 1	0.000	17:19
Breath Volume		2.070	17:19
Air Blank		0.000	17:20
Air Blank		0.000	17:22
Subject Sample	# 2	0.000	17:22
Breath Volume		2.328	17:22
Air Blank		0.000	17:23
Control Test		0.081	17:23
Air Blank		0.000	17:23
Diagnostics Check		OK	17:23

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,N,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJIA,Y
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 10/13/2025 19:38:04

Diagnostics Check		OK	19:31
Air Blank		0.000	19:31
Control Test		0.081	19:32
Air Blank		0.000	19:32
Subject Sample	# 1	0.000	19:33
Breath Volume		1.101	19:33
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.000	19:36
Breath Volume		1.562	19:36
Air Blank		0.000	19:37
Control Test		0.081	19:37
Air Blank		0.000	19:37
Diagnostics Check		OK	19:38

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,N,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HODGE,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: HERNANDEZ-DIAZ

DATE/TIME OF TEST: 10/14/2025 03:01:19

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.081	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.211	02:56
Breath Volume		2.144	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.213	02:59
Breath Volume		1.601	02:59
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 10/26/2025 21:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: HELMS

DATE/TIME OF TEST: 10/14/2025 11:06:45

Diagnostics Check		OK	11:00
Air Blank		0.000	11:01
Control Test		0.080	11:01
Air Blank		0.000	11:01
Subject Sample	# 1	0.000	11:02
Breath Volume		2.226	11:02
Air Blank		0.000	11:02
Air Blank		0.000	11:04
Subject Sample	# 2	0.000	11:05
Breath Volume		2.257	11:05
Air Blank		0.000	11:05
Control Test		0.080	11:06
Air Blank		0.000	11:06
Diagnostics Check		OK	11:06

Data Download Date/Time: 10/26/2025 21:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,A,K
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BUTLER,A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: STOCKS

DATE/TIME OF TEST: 10/14/2025 17:10:48

Diagnostics Check		OK	17:05
Air Blank		0.000	17:05
Control Test		0.081	17:05
Air Blank		0.000	17:06
Subject Sample	# 1	0.000	17:06
Breath Volume		1.488	17:06
Air Blank		0.000	17:07
Air Blank		0.000	17:09
Subject Sample	# 2	0.000	17:09
Breath Volume		1.292	17:09
Air Blank		0.000	17:09
Control Test		0.079	17:10
Air Blank		0.000	17:10
Diagnostics Check		OK	17:10

Data Download Date/Time: 10/26/2025 21:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,NEAL,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MCROY,I
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BUSSIERE

DATE/TIME OF TEST: 10/14/2025 21:21:08

Diagnostics Check		OK	21:13
Air Blank		0.000	21:13
Control Test		0.081	21:13
Air Blank		0.000	21:14
Subject Sample	# 1	0.000	21:16
Breath Volume		1.347	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.000	21:19
Breath Volume		1.195	21:19
Air Blank		0.000	21:20
Control Test		0.080	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 10/26/2025 21:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJA,YOSMANY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 10/18/2025 04:29:45

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:21	
Control Test		0.080	04:22	
Air Blank		0.000	04:22	
Subject Sample	# 1		04:25	No Sample Provided
Breath Volume		0.000	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:28	
Control Test		0.080	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 10/26/2025 21:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BECERRA,JOSE
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: IBANEZ

DATE/TIME OF TEST: 10/18/2025 22:10:21

Diagnostics Check		OK	22:07	
Air Blank		0.000	22:07	
Control Test		0.081	22:08	
Air Blank		0.000	22:08	
Subject Sample	# 1		22:08	Subject Test Refused
Breath Volume		0.000	22:08	
Air Blank		0.000	22:09	
Control Test		0.081	22:09	
Air Blank		0.000	22:10	
Diagnostics Check		OK	22:10	

Data Download Date/Time: 10/26/2025 21:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJA,YOSMANY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: STANTON

DATE/TIME OF TEST: 10/19/2025 00:31:38

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:29	
Control Test		0.081	00:29	
Air Blank		0.000	00:30	
Subject Sample	# 1		00:30	Subject Test Refused
Breath Volume		0.000	00:30	
Air Blank		0.000	00:30	
Control Test		0.081	00:31	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 10/26/2025 21:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJA,YOSMANY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: ROMAN

DATE/TIME OF TEST: 10/19/2025 02:35:52

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.129	02:31
Breath Volume		2.382	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:33
Subject Sample	# 2	0.120	02:34
Breath Volume		2.058	02:34
Air Blank		0.000	02:34
Control Test		0.079	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 10/26/2025 21:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LISTER,DALTON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/20/2025 00:55:57

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:55	
Control Test		0.080	00:55	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 10/26/2025 21:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: STOCKS-GOODRICH

DATE/TIME OF TEST: 10/21/2025 09:48:36

Diagnostics Check		OK	09:42
Air Blank		0.000	09:42
Control Test		0.081	09:42
Air Blank		0.000	09:43
Subject Sample	# 1	0.000	09:44
Breath Volume		1.531	09:44
Air Blank		0.000	09:44
Air Blank		0.000	09:46
Subject Sample	# 2	0.000	09:47
Breath Volume		1.410	09:47
Air Blank		0.000	09:47
Control Test		0.081	09:47
Air Blank		0.000	09:48
Diagnostics Check		OK	09:48

Data Download Date/Time: 10/26/2025 21:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WALKER,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 10/22/2025 00:21:00

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:18	
Control Test		0.080	00:18	
Air Blank		0.000	00:19	
Subject Sample	# 1		00:19	Subject Test Refused
Breath Volume		0.000	00:19	
Air Blank		0.000	00:20	
Control Test		0.080	00:20	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:21	

Data Download Date/Time: 10/26/2025 21:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: STOCKS-GOODRICH

DATE/TIME OF TEST: 10/24/2025 17:07:52

Diagnostics Check		OK	17:01
Air Blank		0.000	17:02
Control Test		0.082	17:02
Air Blank		0.000	17:02
Subject Sample	# 1	0.000	17:03
Breath Volume		1.503	17:03
Air Blank		0.000	17:04
Air Blank		0.000	17:06
Subject Sample	# 2	0.000	17:06
Breath Volume		1.421	17:06
Air Blank		0.000	17:06
Control Test		0.081	17:07
Air Blank		0.000	17:07
Diagnostics Check		OK	17:07

Data Download Date/Time: 10/26/2025 21:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDSON,A,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PTR,PTR
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 10/26/2025 18:13:50

Diagnostics Check		OK	18:07
Air Blank		0.000	18:08
Control Test		0.081	18:08
Air Blank		0.000	18:09
Subject Sample	# 1	0.000	18:09
Breath Volume		2.253	18:09
Air Blank		0.000	18:10
Air Blank		0.000	18:12
Subject Sample	# 2	0.000	18:12
Breath Volume		1.867	18:12
Air Blank		0.000	18:12
Control Test		0.080	18:13
Air Blank		0.000	18:13
Diagnostics Check		OK	18:13

Data Download Date/Time: 10/26/2025 21:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HARRISON,DUSTIN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: PIERRE-PAUL

DATE/TIME OF TEST: 10/27/2025 21:02:28

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.081	20:57
Air Blank		0.000	20:57
Subject Sample	# 1	0.000	20:58
Breath Volume		1.917	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.000	21:01
Breath Volume		2.117	21:01
Air Blank		0.000	21:01
Control Test		0.079	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 11/10/2025 01:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,CALEB,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RAMOS,L
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: PAPESH

DATE/TIME OF TEST: 10/28/2025 18:59:46

Diagnostics Check		OK	18:53
Air Blank		0.000	18:53
Control Test		0.081	18:54
Air Blank		0.000	18:54
Subject Sample	# 1	0.262	18:55
Breath Volume		2.398	18:55
Air Blank		0.000	18:55
Air Blank		0.000	18:57
Subject Sample	# 2	0.263	18:58
Breath Volume		2.335	18:58
Air Blank		0.000	18:58
Control Test		0.080	18:59
Air Blank		0.000	18:59
Diagnostics Check		OK	18:59

Data Download Date/Time: 11/10/2025 01:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: HELMS

DATE/TIME OF TEST: 10/29/2025 09:30:17

Diagnostics Check		OK	09:23
Air Blank		0.000	09:24
Control Test		0.080	09:24
Air Blank		0.000	09:24
Subject Sample	# 1	0.000	09:25
Breath Volume		1.855	09:25
Air Blank		0.000	09:25
Air Blank		0.000	09:27
Subject Sample	# 2	0.000	09:28
Breath Volume		2.234	09:28
Air Blank		0.000	09:29
Control Test		0.080	09:29
Air Blank		0.000	09:30
Diagnostics Check		OK	09:30

Data Download Date/Time: 11/10/2025 01:49

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 10/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: STOCKS-GOODRICH

DATE/TIME OF TEST: 10/29/2025 17:12:12

Diagnostics Check		OK	17:06
Air Blank		0.000	17:06
Control Test		0.081	17:06
Air Blank		0.000	17:07
Subject Sample	# 1	0.000	17:07
Breath Volume		1.468	17:07
Air Blank		0.000	17:08
Air Blank		0.000	17:10
Subject Sample	# 2	0.000	17:10
Breath Volume		1.324	17:10
Air Blank		0.000	17:11
Control Test		0.081	17:11
Air Blank		0.000	17:12
Diagnostics Check		OK	17:12

Data Download Date/Time: 11/10/2025 01:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOVAR,LUIS,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: QUIGLEY,JEREMIAH
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 10/10/2025 01:59:41

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.192	01:54
Breath Volume		3.167	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:56
Subject Sample	# 2	0.177	01:57
Breath Volume		2.656	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 10/27/2025 18:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: JIMENEZ,ANDREW
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/19/2025 03:15:35

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:12	
Control Test		0.080	03:12	
Air Blank		0.000	03:13	
Subject Sample	# 1		03:14	Subject Test Refused
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Control Test		0.080	03:14	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 10/27/2025 18:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEMHEUER,MICHAEL,W
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL,DOMINIC
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 10/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:16

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 10/29/2025 15:43:05

Diagnostics Check		OK	15:37
Air Blank		0.000	15:37
Control Test		0.080	15:37
Air Blank		0.000	15:38
Subject Sample	# 1	0.000	15:38
Breath Volume		2.273	15:38
Air Blank		0.000	15:39
Air Blank		0.000	15:41
Subject Sample	# 2	0.000	15:41
Breath Volume		1.898	15:41
Air Blank		0.000	15:42
Control Test		0.078	15:42
Air Blank		0.000	15:43
Diagnostics Check		OK	15:43

Data Download Date/Time: 11/11/2025 14:05

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: MYDLYK

DATE/TIME OF TEST: 10/02/2025 15:40:59

Diagnostics Check		OK	15:32	
Air Blank		0.000	15:32	
Control Test		0.079	15:32	
Air Blank		0.000	15:33	
Subject Sample	# 1	0.228	15:36	Volume Not Met
Breath Volume		0.222	15:36	
Air Blank		0.000	15:37	
Air Blank		0.000	15:39	
Subject Sample	# 2		15:39	Subject Test Refused
Breath Volume		0.000	15:39	
Air Blank		0.000	15:40	
Control Test		0.078	15:40	
Air Blank		0.000	15:40	
Diagnostics Check		OK	15:40	

Data Download Date/Time: 10/29/2025 20:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:02

SUBJECT'S LAST NAME: MYDLYK

DATE/TIME OF TEST: 10/02/2025 15:52:48

Diagnostics Check		OK	15:48	
Air Blank		0.000	15:48	
Control Test		0.079	15:49	
Air Blank		0.000	15:49	
Subject Sample	# 1		15:51	Subject Test Refused
Breath Volume		0.925	15:51	
Air Blank		0.000	15:51	
Control Test		0.080	15:52	
Air Blank		0.000	15:52	
Diagnostics Check		OK	15:52	

Data Download Date/Time: 10/29/2025 20:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BENJAMIN,P
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MYERS,BENJAMIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: BOCK-SWEDA

DATE/TIME OF TEST: 10/03/2025 18:02:37

Diagnostics Check		OK	17:54
Air Blank		0.000	17:54
Control Test		0.080	17:55
Air Blank		0.000	17:55
Subject Sample	# 1	0.229	17:57
Breath Volume		1.523	17:57
Air Blank		0.000	17:58
Air Blank		0.000	17:59
Subject Sample	# 2	0.222	18:00
Breath Volume		1.785	18:00
Air Blank		0.000	18:01
Control Test		0.077	18:02
Air Blank		0.000	18:02
Diagnostics Check		OK	18:02

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: GUARINO

DATE/TIME OF TEST: 10/03/2025 21:00:10

Diagnostics Check		OK	20:52
Air Blank		0.000	20:52
Control Test		0.080	20:53
Air Blank		0.000	20:53
Subject Sample	# 1	0.061	20:55
Breath Volume		1.558	20:55
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.059	20:58
Breath Volume		1.886	20:58
Air Blank		0.000	20:59
Control Test		0.078	20:59
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUI,DANIEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RUCCI,MATTHEW
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: VALLAS

DATE/TIME OF TEST: 10/04/2025 01:50:58

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.171	01:44
Breath Volume		1.332	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.174	01:49
Breath Volume		1.851	01:49
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ANZALONE

DATE/TIME OF TEST: 10/04/2025 02:22:07

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.079	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.000	02:17
Breath Volume		2.667	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.000	02:20
Breath Volume		3.023	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MESCH

DATE/TIME OF TEST: 10/04/2025 03:08:03

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:02
Subject Sample	# 1	0.156	03:03
Breath Volume		4.207	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:05
Subject Sample	# 2	0.153	03:06
Breath Volume		4.539	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:07
Diagnostics Check		OK	03:08

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,KEATON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: TBER,TARIQ
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 09:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: WOOTERS

DATE/TIME OF TEST: 10/04/2025 11:04:07

Diagnostics Check		OK	10:56
Air Blank		0.000	10:57
Control Test		0.080	10:57
Air Blank		0.000	10:57
Subject Sample	# 1	0.088	10:59
Breath Volume		3.652	10:59
Air Blank		0.000	10:59
Air Blank		0.000	11:01
Subject Sample	# 2	0.086	11:02
Breath Volume		4.093	11:02
Air Blank		0.000	11:03
Control Test		0.079	11:03
Air Blank		0.000	11:04
Diagnostics Check		OK	11:04

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: FENSTERMACHER

DATE/TIME OF TEST: 10/05/2025 03:13:36

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.126	03:08
Breath Volume		5.183	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.127	03:11
Breath Volume		5.003	03:11
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DA SILVA,ANDRE,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: DA SILVA,ANDRE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: DOMINGO DIAZ

DATE/TIME OF TEST: 10/06/2025 02:33:45

Diagnostics Check		OK	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.159	02:29
Breath Volume		2.910	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.159	02:32
Breath Volume		2.480	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: MCAULIFFE

DATE/TIME OF TEST: 10/06/2025 12:18:21

Diagnostics Check		OK	12:16	
Air Blank		0.000	12:16	
Control Test		0.080	12:16	
Air Blank		0.000	12:17	
Subject Sample	# 1		12:17	RFI Detect
Breath Volume		0.859	12:17	
Air Blank			12:17	RFI Detect
Air Blank		0.000	12:18	

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: MCAULIFFE

DATE/TIME OF TEST: 10/06/2025 12:26:14

Diagnostics Check		OK	12:20
Air Blank		0.000	12:20
Control Test		0.080	12:20
Air Blank		0.000	12:21
Subject Sample	# 1	0.000	12:21
Breath Volume		1.398	12:21
Air Blank		0.000	12:22
Air Blank		0.000	12:24
Subject Sample	# 2	0.000	12:24
Breath Volume		1.503	12:24
Air Blank		0.000	12:25
Control Test		0.079	12:25
Air Blank		0.000	12:26
Diagnostics Check		OK	12:26

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRONSON,TAYLOR,K
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PIERCE,MYREON
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: EASTERLING

DATE/TIME OF TEST: 10/07/2025 05:13:23

Diagnostics Check		OK	05:06
Air Blank		0.000	05:06
Control Test		0.080	05:07
Air Blank		0.000	05:07
Subject Sample	# 1	0.108	05:08
Breath Volume		1.804	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.111	05:11
Breath Volume		2.511	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PIERCE,MYREON
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: BRIGUGLIO

DATE/TIME OF TEST: 10/07/2025 09:12:42

Diagnostics Check		OK	09:06
Air Blank		0.000	09:07
Control Test		0.080	09:07
Air Blank		0.000	09:07
Subject Sample	# 1	0.000	09:08
Breath Volume		2.363	09:08
Air Blank		0.000	09:08
Air Blank		0.000	09:10
Subject Sample	# 2	0.000	09:11
Breath Volume		2.144	09:11
Air Blank		0.000	09:11
Control Test		0.081	09:12
Air Blank		0.000	09:12
Diagnostics Check		OK	09:12

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHWINE,CHARLES,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MENNINGA,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: FELLIS

DATE/TIME OF TEST: 10/07/2025 22:19:43

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.081	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.206	22:15
Breath Volume		2.480	22:15
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.210	22:18
Breath Volume		1.933	22:18
Air Blank		0.000	22:18
Control Test		0.080	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 10/09/2025 22:50:45

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:39	
Control Test		0.080	22:40	
Air Blank		0.000	22:40	
Subject Sample	# 1		22:43	No Sample Provided
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Air Blank		0.000	22:46	
Subject Sample	# 2		22:49	No Sample Provided
Breath Volume		0.000	22:49	
Air Blank		0.000	22:49	
Control Test		0.079	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:50	

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRONSON,TAYLOR,K
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RUCCI,MATTHEW
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: BRIGUGLIO

DATE/TIME OF TEST: 10/10/2025 04:53:48

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.081	04:47
Air Blank		0.000	04:48
Subject Sample	# 1	0.000	04:49
Breath Volume		1.355	04:49
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.000	04:52
Breath Volume		1.320	04:52
Air Blank		0.000	04:52
Control Test		0.081	04:53
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 10/11/2025 00:18:17

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.117	00:13
Breath Volume		4.812	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.112	00:16
Breath Volume		4.042	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOMBARDI,MADISON,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ROCKWELL,NICKOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/11/2025 11:28:28

Diagnostics Check		OK	11:22
Air Blank		0.000	11:22
Control Test		0.081	11:22
Air Blank		0.000	11:23
Subject Sample	# 1	0.108	11:23
Breath Volume		2.003	11:23
Air Blank		0.000	11:24
Air Blank		0.000	11:26
Subject Sample	# 2	0.114	11:26
Breath Volume		1.312	11:26
Air Blank		0.000	11:27
Control Test		0.079	11:27
Air Blank		0.000	11:28
Diagnostics Check		OK	11:28

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANILAG,SAMUEL,C
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PANILAG,SAMUEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: ROMINIK

DATE/TIME OF TEST: 10/11/2025 16:24:21

Diagnostics Check		OK	16:16
Air Blank		0.000	16:17
Control Test		0.080	16:17
Air Blank		0.000	16:17
Subject Sample	# 1	0.313	16:18
Breath Volume		2.472	16:18
Air Blank		0.000	16:19
Air Blank		0.000	16:20
Subject Sample	# 2	0.322	16:22
Breath Volume		3.796	16:22
Air Blank		0.000	16:23
Control Test		0.079	16:23
Air Blank		0.000	16:24
Diagnostics Check		OK	16:24

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/12/2025 00:05:59

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.080	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1	0.000	00:02	
Breath Volume		2.570	00:02	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2	0.000	00:04	
Breath Volume		2.179	00:04	
Air Blank		0.000	00:05	
			00:05	RFI Detect
Air Blank		0.000	00:05	

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 10/12/2025 00:14:09

Diagnostics Check		OK	00:07
Air Blank		0.000	00:08
Control Test		0.080	00:08
Air Blank		0.000	00:09
Subject Sample	# 1	0.000	00:09
Breath Volume		2.589	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.000	00:12
Breath Volume		1.812	00:12
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,MIKO,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: PHILLIPS,DONNIE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 10/12/2025 02:07:20

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.000	02:03
Breath Volume		1.906	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.000	02:05
Breath Volume		2.035	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BLACKMAN,ALBERT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: CLINCH

DATE/TIME OF TEST: 10/14/2025 19:35:09

Diagnostics Check		OK	19:28
Air Blank		0.000	19:29
Control Test		0.080	19:29
Air Blank		0.000	19:30
Subject Sample	# 1	0.000	19:30
Breath Volume		1.222	19:30
Air Blank		0.000	19:31
Air Blank		0.000	19:33
Subject Sample	# 2	0.000	19:33
Breath Volume		1.308	19:33
Air Blank		0.000	19:34
Control Test		0.079	19:34
Air Blank		0.000	19:35
Diagnostics Check		OK	19:35

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,SEAN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CANNING,SEAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 10/18/2025 03:18:33

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.080	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1	0.083	03:09	Volume Not Met
Breath Volume		0.101	03:09	
Air Blank		0.000	03:10	
Air Blank		0.000	03:12	
Subject Sample	# 2	0.111	03:12	
Breath Volume		2.039	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 3	0.113	03:16	
Breath Volume		1.722	03:16	
Air Blank		0.000	03:17	
Control Test		0.079	03:17	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSER,AARON,M
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: FLAHERTY,ZACH
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: SEDBERRY

DATE/TIME OF TEST: 10/18/2025 05:42:16

Diagnostics Check		OK	05:35
Air Blank		0.000	05:36
Control Test		0.080	05:36
Air Blank		0.000	05:36
Subject Sample	# 1	0.144	05:37
Breath Volume		3.527	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.148	05:40
Breath Volume		3.484	05:40
Air Blank		0.000	05:41
Control Test		0.079	05:41
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSER,AARON,M
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MACK,T
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 10/19/2025 05:51:13

Diagnostics Check		OK	05:44
Air Blank		0.000	05:44
Control Test		0.080	05:44
Air Blank		0.000	05:45
Subject Sample	# 1	0.128	05:46
Breath Volume		2.500	05:46
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.125	05:49
Breath Volume		2.445	05:49
Air Blank		0.000	05:50
Control Test		0.078	05:50
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 10/20/2025 17:27:42

Diagnostics Check		OK	17:24	
Air Blank		0.000	17:24	
Control Test		0.080	17:25	
Air Blank		0.000	17:25	
Subject Sample	# 1	0.171	17:26	
Breath Volume		1.636	17:26	
Air Blank			17:26	RFI Detect
Air Blank		0.000	17:26	
Control Test		0.080	17:27	
Air Blank		0.000	17:27	
Diagnostics Check		OK	17:27	

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 10/20/2025 17:35:51

Diagnostics Check		OK	17:29
Air Blank		0.000	17:29
Control Test		0.080	17:29
Air Blank		0.000	17:30
Subject Sample	# 1	0.166	17:31
Breath Volume		2.714	17:31
Air Blank		0.000	17:31
Air Blank		0.000	17:33
Subject Sample	# 2	0.164	17:34
Breath Volume		2.417	17:34
Air Blank		0.000	17:34
Control Test		0.079	17:35
Air Blank		0.000	17:35
Diagnostics Check		OK	17:35

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JERRY,F
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HARVEY,PARKER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: EL-KAMEL

DATE/TIME OF TEST: 10/21/2025 05:31:36

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.079	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.200	05:26
Breath Volume		2.570	05:26
Air Blank		0.000	05:27
Air Blank		0.000	05:28
Subject Sample	# 2	0.204	05:29
Breath Volume		2.605	05:29
Air Blank		0.000	05:30
Control Test		0.079	05:31
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGOLD,K,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: DEAN,C
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: STRUNK

DATE/TIME OF TEST: 10/25/2025 19:42:59

Diagnostics Check		OK	19:36
Air Blank		0.000	19:37
Control Test		0.081	19:37
Air Blank		0.000	19:37
Subject Sample	# 1	0.191	19:38
Breath Volume		5.449	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:40
Subject Sample	# 2	0.192	19:41
Breath Volume		4.871	19:41
Air Blank		0.000	19:42
Control Test		0.079	19:42
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JERRY,F
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MENNINGA,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: HERITAGE

DATE/TIME OF TEST: 10/26/2025 12:55:05

Diagnostics Check		OK	12:52	
Air Blank		0.000	12:52	
Control Test		0.081	12:53	
Air Blank		0.000	12:53	
Subject Sample	# 1		12:53	Subject Test Refused
Breath Volume		0.000	12:53	
Air Blank		0.000	12:54	
Control Test		0.081	12:54	
Air Blank		0.000	12:55	
Diagnostics Check		OK	12:55	

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZ,BRANDON,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BECKER,NATHANIAL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 10/27/2025 21:39:51

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:28	
Control Test		0.080	21:28	
Air Blank		0.000	21:29	
Subject Sample	# 1	0.233	21:32	Volume Not Met
Breath Volume		0.000	21:32	
Air Blank		0.000	21:33	
Air Blank		0.000	21:35	
Subject Sample	# 2	0.229	21:38	Volume Not Met
Breath Volume		0.039	21:38	
Air Blank		0.000	21:38	
Control Test		0.078	21:39	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZ,BRANDON,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BECKER,NATHANIAL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 10/27/2025 22:02:23

Diagnostics Check		OK	21:50	
Air Blank		0.000	21:50	
Control Test		0.079	21:51	
Air Blank		0.000	21:51	
Subject Sample	# 1	0.271	21:54	Volume Not Met
Breath Volume		0.210	21:54	
Air Blank		0.000	21:55	
Air Blank		0.000	21:57	
Subject Sample	# 2	0.225	22:00	Volume Not Met
Breath Volume		0.031	22:00	
Air Blank		0.000	22:01	
Control Test		0.077	22:01	
Air Blank			22:01	RFI Detect
Air Blank		0.000	22:02	

Data Download Date/Time: 10/30/2025 06:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDRE,BOBBY,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CELLA,DOMENICO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: LEHMAN

DATE/TIME OF TEST: 10/06/2025 03:57:10

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:54	
Control Test		0.081	03:55	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.081	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 10/30/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDRE,BOBBY,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: LUNN,JAMES
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 10/13/2025 04:39:22

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.081	04:37	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:38	
Control Test		0.081	04:38	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 10/30/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.