

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:36

SUBJECT'S LAST NAME: ASHWORTH

DATE/TIME OF TEST: 09/03/2025 11:08:02

Diagnostics Check	OK	11:01
Air Blank	0.000	11:01
Control Test	0.080	11:01
Air Blank	0.000	11:02
Subject Sample # 1	0.000	11:03
Breath Volume	2.207	11:03
Air Blank	0.000	11:03
Air Blank	0.000	11:05
Subject Sample # 2	0.000	11:06
Breath Volume	2.785	11:06
Air Blank	0.000	11:07
Control Test	0.080	11:07
Air Blank	0.000	11:07
Diagnostics Check	OK	11:08

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESTO,LUIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: RESTO,LUIS
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/04/2025 00:05:55

Diagnostics Check	OK	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.113	00:01
Breath Volume	1.753	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.111	00:04
Breath Volume	2.050	00:04
Air Blank	0.000	00:04
Control Test	0.077	00:05
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: KLOOSTERMAN, JASON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: PHALON

DATE/TIME OF TEST: 09/04/2025 00:39:48

Diagnostics Check	OK	00:32
Air Blank	0.000	00:33
Control Test	0.080	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.113	00:35
Breath Volume	1.804	00:35
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.113	00:38
Breath Volume	1.757	00:38
Air Blank	0.000	00:38
Control Test	0.078	00:39
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,BRYAN,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MARTINI,BRYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/04/2025 01:14:00

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.078	01:06
Air Blank	0.000	01:07
Subject Sample # 1	0.167	01:09
Breath Volume	2.347	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.173	01:12
Breath Volume	1.988	01:12
Air Blank	0.000	01:13
Control Test	0.078	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:14

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:56

SUBJECT'S LAST NAME: SCHAGRIN

DATE/TIME OF TEST: 09/04/2025 10:39:17

Diagnostics Check	OK	10:29	
Air Blank	0.000	10:30	
Control Test	0.079	10:30	
Air Blank	0.000	10:31	
Subject Sample	# 1	10:34	No Sample Provided
Breath Volume	0.000	10:34	
Air Blank	0.000	10:34	
Air Blank	0.000	10:36	
Subject Sample	# 2	10:37	Subject Test Refused
Breath Volume	0.000	10:37	
Air Blank	0.000	10:38	
Control Test	0.079	10:38	
Air Blank	0.000	10:39	
Diagnostics Check	OK	10:39	

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/04/2025 22:03:19

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.079	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.159	21:58
Breath Volume	2.324	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.151	22:01
Breath Volume	3.710	22:01
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAW, COLLIN, R
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: YAW, COLLIN
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 09/05/2025 01:35:48

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.163	01:27
Breath Volume	1.792	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.199	01:30
Breath Volume	1.699	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 3	0.169	01:34
Breath Volume	2.437	01:34
Air Blank	0.000	01:34
Control Test	0.077	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,ERIC,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LOTT,ERIC
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 09/06/2025 01:30:33

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.079	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.318	01:26
Breath Volume	4.183	01:26
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.312	01:28
Breath Volume	3.722	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNLOP,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DUNLOP,MATTHEW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PLATA

DATE/TIME OF TEST: 09/06/2025 02:16:49

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.139	02:12
Breath Volume	1.597	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.137	02:15
Breath Volume	1.710	02:15
Air Blank	0.000	02:15
Control Test	0.078	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,D
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: CERNUTO,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ROWLEY

DATE/TIME OF TEST: 09/06/2025 20:41:42

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.079	20:35
Air Blank	0.000	20:36
Subject Sample # 1	0.193	20:36
Breath Volume	1.507	20:36
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.202	20:40
Breath Volume	1.730	20:40
Air Blank	0.000	20:40
Control Test	0.079	20:41
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADY,REBECCA,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRADY,REBECCA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 09/07/2025 17:55:00

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.079	17:47
Air Blank	0.000	17:47
Subject Sample # 1	0.065	17:50
Breath Volume	3.445	17:50
Air Blank	0.000	17:51
Air Blank	0.000	17:53
Subject Sample # 2	0.064	17:53
Breath Volume	3.222	17:53
Air Blank	0.000	17:54
Control Test	0.079	17:54
Air Blank	0.000	17:54
Diagnostics Check	OK	17:55

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERKINS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: PERKINS,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: GALVAN SALINAS

DATE/TIME OF TEST: 09/07/2025 20:58:09

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.079	20:52
Air Blank	0.000	20:52
Subject Sample # 1	0.240	20:53
Breath Volume	1.949	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:55
Subject Sample # 2	0.248	20:56
Breath Volume	2.593	20:56
Air Blank	0.000	20:57
Control Test	0.077	20:57
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 09/17/2025 21:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENOVIL,GEDEON,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BRENOVIL,GEDEON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: MURRY

DATE/TIME OF TEST: 09/12/2025 10:16:57

Diagnostics Check	OK	10:10
Air Blank	0.000	10:11
Control Test	0.079	10:11
Air Blank	0.000	10:12
Subject Sample # 1	0.000	10:12
Breath Volume	3.195	10:12
Air Blank	0.000	10:13
Air Blank	0.000	10:15
Subject Sample # 2	0.000	10:15
Breath Volume	2.546	10:15
Air Blank	0.000	10:16
Control Test	0.080	10:16
Air Blank	0.000	10:16
Diagnostics Check	OK	10:16

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,BRYAN,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MARTINI,BRYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 09/12/2025 22:53:28

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.160	22:48
Breath Volume	1.535	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.171	22:51
Breath Volume	3.082	22:51
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: ALVAREZ LOPEZ

DATE/TIME OF TEST: 09/13/2025 05:15:38

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.080	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.153	05:11
Breath Volume	2.226	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.157	05:14
Breath Volume	3.335	05:14
Air Blank	0.000	05:14
Control Test	0.079	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: GOLINO, ANTHONY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MENDEZ DOMINGUEZ

DATE/TIME OF TEST: 09/14/2025 00:14:11

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.271	00:09
Breath Volume	1.875	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.273	00:12
Breath Volume	2.128	00:12
Air Blank	0.000	00:13
Control Test	0.078	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEGANY,JOSHUA,M
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: BEGANY,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 09/14/2025 06:13:41

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.080	06:02
Air Blank	0.000	06:02
Subject Sample # 1	0.151	06:04
Breath Volume	1.507	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample # 2	0.124	06:08
Breath Volume	1.691	06:08
Air Blank	0.000	06:08
Air Blank	0.000	06:10
Subject Sample # 3	0.130	06:12
Breath Volume	1.625	06:12
Air Blank	0.000	06:12
Control Test	0.076	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNALLY, TYLER, S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE, LOGAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: VAN

DATE/TIME OF TEST: 09/15/2025 00:41:29

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.153	00:36
Breath Volume	2.167	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.161	00:39
Breath Volume	1.757	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/15/2025 05:58:01

Diagnostics Check	OK	05:50
Air Blank	0.000	05:51
Control Test	0.079	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.000	05:52
Breath Volume	2.464	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.000	05:56
Breath Volume	2.101	05:56
Air Blank	0.000	05:57
Control Test	0.080	05:57
Air Blank	0.000	05:57
Diagnostics Check	OK	05:58

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: BESZE

DATE/TIME OF TEST: 09/16/2025 20:58:56

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.080	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.000	20:54
Breath Volume	1.542	20:54
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.000	20:57
Breath Volume	1.265	20:57
Air Blank	0.000	20:57
Control Test	0.079	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:58

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:55

SUBJECT'S LAST NAME: SINGER

DATE/TIME OF TEST: 09/17/2025 09:29:07

Diagnostics Check	OK	09:22
Air Blank	0.000	09:22
Control Test	0.080	09:23
Air Blank	0.000	09:23
Subject Sample # 1	0.000	09:24
Breath Volume	2.183	09:24
Air Blank	0.000	09:24
Air Blank	0.000	09:26
Subject Sample # 2	0.000	09:27
Breath Volume	2.167	09:27
Air Blank	0.000	09:28
Control Test	0.079	09:28
Air Blank	0.000	09:29
Diagnostics Check	OK	09:29

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,NICHOLAS
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/17/2025 19:13:36

Diagnostics Check	OK	19:06
Air Blank	0.000	19:07
Control Test	0.079	19:07
Air Blank	0.000	19:08
Subject Sample # 1	0.150	19:08
Breath Volume	1.582	19:08
Air Blank	0.000	19:09
Air Blank	0.000	19:11
Subject Sample # 2	0.144	19:12
Breath Volume	2.535	19:12
Air Blank	0.000	19:12
Control Test	0.079	19:12
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 09/17/2025 21:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTTLE, ADAM, R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TUTTLE, ADAM
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: MARTINEZ CARBAJAL

DATE/TIME OF TEST: 09/18/2025 01:10:28

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.198	01:05
Breath Volume	1.257	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.186	01:08
Breath Volume	1.164	01:08
Air Blank	0.000	01:09
Control Test	0.078	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 10/15/2025 19:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YAW, COLLIN, R
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: YAW, COLLIN
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 09/18/2025 21:27:33

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.079	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.200	21:23
Breath Volume	5.386	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.204	21:25
Breath Volume	2.367	21:25
Air Blank	0.000	21:26
Control Test	0.077	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,BRYAN,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MARTINI,BRYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: CSASZAR

DATE/TIME OF TEST: 09/18/2025 22:11:00

Diagnostics Check	OK	21:59	
Air Blank	0.000	21:59	
Control Test	0.078	22:00	
Air Blank	0.000	22:00	
Subject Sample	# 1 0.135	22:03	Volume Not Met
Breath Volume	0.398	22:03	
Air Blank	0.000	22:04	
Air Blank	0.000	22:06	
Subject Sample	# 2 0.119	22:09	Volume Not Met
Breath Volume	0.109	22:09	
Air Blank	0.000	22:10	
Control Test	0.077	22:10	
Air Blank	0.000	22:10	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEVENAR,TATIANA,R
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: LEVENAR,TATIANA
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MATUL ROBLERO

DATE/TIME OF TEST: 09/19/2025 00:27:28

Diagnostics Check	OK	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Subject Sample # 1	0.235	00:21
Breath Volume	2.410	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:23
Subject Sample # 2	0.221	00:25
Breath Volume	1.378	00:25
Air Blank	0.000	00:26
Control Test	0.078	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: KRESTAN

DATE/TIME OF TEST: 09/19/2025 21:33:31

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.079	21:27
Air Blank	0.000	21:27
Subject Sample # 1	0.210	21:28
Breath Volume	2.375	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.205	21:31
Breath Volume	1.992	21:31
Air Blank	0.000	21:32
Control Test	0.078	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURENE,BIKENDY,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: TURENE,BIKENDY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CZERNOWSKI

DATE/TIME OF TEST: 09/20/2025 05:16:17

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Subject Sample # 1	0.139	05:11
Breath Volume	2.085	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.131	05:14
Breath Volume	1.105	05:14
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEBLAKER,JOSHUA,N
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: DEBLAKER,JOSHUA
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: ALEMAN

DATE/TIME OF TEST: 09/20/2025 22:17:17

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.190	22:12
Breath Volume	1.542	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.192	22:15
Breath Volume	1.394	22:15
Air Blank	0.000	22:16
Control Test	0.078	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUFFNER,PRESTON,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: KUFFNER,PRESTON
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MONTANEZ

DATE/TIME OF TEST: 09/21/2025 03:32:49

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.079	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.161	03:28
Breath Volume	2.535	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.157	03:31
Breath Volume	2.621	03:31
Air Blank	0.000	03:31
Control Test	0.078	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTEUCCI,DANIELLE,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MATTEUCCI,DANIELLE
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 09/21/2025 22:24:14

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.079	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.000	22:19
Breath Volume	4.273	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.000	22:22
Breath Volume	4.605	22:22
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,ERIC,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LOTT,ERIC
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/24/2025 23:24:37

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.079	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1 0.161	23:13	
Breath Volume	2.605	23:13	
Air Blank	0.000	23:13	
Air Blank	0.000	23:15	
Subject Sample	# 2 0.144	23:18	Volume Not Met
Breath Volume	0.105	23:18	
Air Blank	0.000	23:19	
Air Blank	0.000	23:21	
Subject Sample	# 3 0.141	23:23	
Breath Volume	1.960	23:23	
Air Blank	0.000	23:23	
Control Test	0.078	23:23	
Air Blank	0.000	23:24	
Diagnostics Check	OK	23:24	

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,ERIC,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LOTT,ERIC
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: COE

DATE/TIME OF TEST: 09/25/2025 23:30:52

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.182	23:26
Breath Volume	2.496	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:28
Subject Sample # 2	0.185	23:29
Breath Volume	2.585	23:29
Air Blank	0.000	23:29
Control Test	0.078	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,BRYAN,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MARTINI,BRYAN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: IACONO

DATE/TIME OF TEST: 09/27/2025 18:35:00

Diagnostics Check	OK	18:27
Air Blank	0.000	18:27
Control Test	0.079	18:27
Air Blank	0.000	18:28
Subject Sample # 1	0.097	18:30
Breath Volume	2.210	18:30
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.100	18:33
Breath Volume	2.519	18:33
Air Blank	0.000	18:34
Control Test	0.079	18:34
Air Blank	0.000	18:34
Diagnostics Check	OK	18:35

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOURNE,RUSSELL,L
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: BURKE,LOGAN
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 09/28/2025 02:07:10

Diagnostics Check	OK	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.203	02:01
Breath Volume	1.578	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.195	02:05
Breath Volume	2.003	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNALLY,TYLER,S
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MCNALLY,TYLER
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: CICCARELLI

DATE/TIME OF TEST: 09/29/2025 00:41:43

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.079	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1 0.064	00:35	Volume Not Met
Breath Volume	0.089	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample	# 2	00:40	Subject Test Refused
Breath Volume	0.140	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:41	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,ERIC,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LOTT,ERIC
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: WHERRY

DATE/TIME OF TEST: 09/30/2025 00:38:28

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.000	00:34
Breath Volume	2.011	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.000	00:37
Breath Volume	2.375	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 10/15/2025 19:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE, OSCAR, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROQUE, OSCAR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: CHAVEZ SALINAS

DATE/TIME OF TEST: 09/01/2025 00:12:48

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.181	00:08
Breath Volume	3.339	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 2	0.194	00:11
Breath Volume	3.230	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 09/25/2025 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: HILSON,RANDY
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: TURNBULL

DATE/TIME OF TEST: 09/09/2025 03:11:13

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.081	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:07	No Sample Provided
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 2	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 09/25/2025 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ALTIDOR,JEFFLEY
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SIMON MENDEZ

DATE/TIME OF TEST: 09/09/2025 23:21:44

Diagnostics Check	OK	23:14
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.140	23:17
Breath Volume	2.750	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.135	23:20
Breath Volume	2.121	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 09/25/2025 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 09/12/2025 00:42:50

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.081	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:35	No Sample Provided
Breath Volume	0.000	00:35	
Air Blank	0.000	00:36	
Air Blank	0.000	00:38	
Subject Sample	# 2	00:41	No Sample Provided
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.079	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 09/25/2025 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMN,ELVIS,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 09/12/2025 00:49:29

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.081	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 09/25/2025 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: BRAVO GODINEZ

DATE/TIME OF TEST: 09/19/2025 19:34:17

Diagnostics Check	OK	19:26
Air Blank	0.000	19:27
Control Test	0.080	19:27
Air Blank	0.000	19:28
Subject Sample # 1	0.278	19:28
Breath Volume	2.156	19:28
Air Blank	0.000	19:29
Air Blank	0.000	19:31
Subject Sample # 2	0.266	19:32
Breath Volume	1.507	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 09/25/2025 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE, OSCAR, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: CELLA, DANIEL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: LEVIS SUAREZ

DATE/TIME OF TEST: 09/21/2025 18:53:44

Diagnostics Check	OK	18:47
Air Blank	0.000	18:47
Control Test	0.079	18:48
Air Blank	0.000	18:48
Subject Sample # 1	0.177	18:49
Breath Volume	1.988	18:49
Air Blank	0.000	18:49
Air Blank	0.000	18:51
Subject Sample # 2	0.169	18:52
Breath Volume	1.898	18:52
Air Blank	0.000	18:52
Control Test	0.079	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:53

Data Download Date/Time: 09/25/2025 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGUIRRE,HENRY
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 09/24/2025 17:22:41

Diagnostics Check	OK	17:15
Air Blank	0.000	17:16
Control Test	0.080	17:16
Air Blank	0.000	17:17
Subject Sample # 1	0.084	17:17
Breath Volume	1.230	17:17
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.087	17:21
Breath Volume	1.316	17:21
Air Blank	0.000	17:21
Control Test	0.079	17:22
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 09/25/2025 16:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: BENITEZ VELASQUEZ

DATE/TIME OF TEST: 09/26/2025 21:46:48

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.082	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.207	21:42
Breath Volume	1.832	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.222	21:45
Breath Volume	1.175	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 10/30/2025 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE, OSCAR, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MORUA, JESUS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402623828
DRY GAS EXPIRY DATE: 12/22/2025
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GABRIEL

DATE/TIME OF TEST: 09/27/2025 22:21:36

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:10	
Control Test	0.080	22:10	
Air Blank	0.000	22:11	
Subject Sample	# 1	22:14	No Sample Provided
Breath Volume	0.000	22:14	
Air Blank	0.000	22:15	
Air Blank	0.000	22:17	
Subject Sample	# 2	22:20	No Sample Provided
Breath Volume	0.000	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:20	
Air Blank	0.000	22:21	
Diagnostics Check	OK	22:21	

Data Download Date/Time: 10/30/2025 05:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: VALLES HERNANDEZ

DATE/TIME OF TEST: 09/01/2025 04:13:33

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.169	04:08
Breath Volume	1.640	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.164	04:11
Breath Volume	1.437	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASS,MALIK,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HASS,MALIK
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: GONZALEZ AREVALO

DATE/TIME OF TEST: 09/01/2025 19:13:46

Diagnostics Check	OK	19:07
Air Blank	0.000	19:07
Control Test	0.079	19:08
Air Blank	0.000	19:08
Subject Sample # 1	0.204	19:09
Breath Volume	2.589	19:09
Air Blank	0.000	19:10
Air Blank	0.000	19:11
Subject Sample # 2	0.209	19:12
Breath Volume	1.937	19:12
Air Blank	0.000	19:12
Control Test	0.077	19:13
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: TAYLOR,JONATHAN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: CLISBY

DATE/TIME OF TEST: 09/04/2025 23:38:21

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.000	23:33
Breath Volume	2.203	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.000	23:36
Breath Volume	3.113	23:36
Air Blank	0.000	23:37
Control Test	0.079	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: ROMERO PORTILLO

DATE/TIME OF TEST: 09/06/2025 00:53:35

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.134	00:48
Breath Volume	1.886	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.136	00:51
Breath Volume	1.839	00:51
Air Blank	0.000	00:52
Control Test	0.078	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASS,MALIK,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HASS,MALIK
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: MANDERSON

DATE/TIME OF TEST: 09/06/2025 18:21:37

Diagnostics Check	OK	18:14
Air Blank	0.000	18:15
Control Test	0.080	18:15
Air Blank	0.000	18:16
Subject Sample # 1	0.244	18:16
Breath Volume	2.476	18:16
Air Blank	0.000	18:17
Air Blank	0.000	18:19
Subject Sample # 2	0.241	18:19
Breath Volume	2.199	18:19
Air Blank	0.000	18:20
Control Test	0.078	18:21
Air Blank	0.000	18:21
Diagnostics Check	OK	18:21

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASS,MALIK,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HASS,MALIK
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/07/2025 17:11:41

Diagnostics Check	OK	17:04
Air Blank	0.000	17:04
Control Test	0.081	17:05
Air Blank	0.000	17:05
Subject Sample # 1	0.125	17:06
Breath Volume	2.105	17:06
Air Blank	0.000	17:07
Air Blank	0.000	17:09
Subject Sample # 2	0.122	17:09
Breath Volume	2.488	17:09
Air Blank	0.000	17:10
Control Test	0.077	17:11
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: GAMEZ-GAMEZ

DATE/TIME OF TEST: 09/11/2025 01:34:47

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.104	01:30
Breath Volume	1.449	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.102	01:33
Breath Volume	1.582	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: LAGINESS

DATE/TIME OF TEST: 09/11/2025 02:22:41

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.240	02:17
Breath Volume	2.171	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.247	02:20
Breath Volume	1.445	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNUTSON,TRISTAN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KNUTSON,TRISTAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: CESPEDES PULGAR

DATE/TIME OF TEST: 09/14/2025 01:43:07

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.148	01:38
Breath Volume	3.367	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.141	01:41
Breath Volume	3.468	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: CASTRO AGIS,J
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 09/14/2025 04:07:34

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.081	04:05	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:06	Subject Test Refused
Breath Volume	0.000	04:06	
Air Blank	0.000	04:06	
Control Test	0.081	04:06	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SARAH,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BROWN,GEORGE
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 09/14/2025 14:52:46

Diagnostics Check	OK	14:46
Air Blank	0.000	14:46
Control Test	0.081	14:46
Air Blank	0.000	14:47
Subject Sample # 1	0.023	14:48
Breath Volume	2.609	14:48
Air Blank	0.000	14:48
Air Blank	0.000	14:50
Subject Sample # 2	0.019	14:51
Breath Volume	1.636	14:51
Air Blank	0.000	14:51
Control Test	0.079	14:52
Air Blank	0.000	14:52
Diagnostics Check	OK	14:52

Data Download Date/Time: 09/15/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 09/01/2025 21:03:24

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.099	20:58
Breath Volume	1.671	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.099	21:01
Breath Volume	1.968	21:01
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 09/02/2025 00:09:52

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.079	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.000	00:02	Volume Not Met
Breath Volume	0.074	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2 0.000	00:08	Volume Not Met
Breath Volume	0.066	00:08	
Air Blank	0.000	00:08	
Control Test	0.080	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: DEBRITO

DATE/TIME OF TEST: 09/02/2025 04:59:36

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.138	04:54
Breath Volume	2.425	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.138	04:57
Breath Volume	2.707	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MCNEES

DATE/TIME OF TEST: 09/02/2025 23:35:20

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.136	23:30
Breath Volume	1.109	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:32
Subject Sample # 2	0.135	23:33
Breath Volume	1.257	23:33
Air Blank	0.000	23:34
Control Test	0.079	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,M,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 09/04/2025 21:10:53

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:02	
Control Test	0.080	21:03	
Air Blank	0.000	21:03	
Subject Sample	# 1	21:06	No Sample Provided
Breath Volume	0.000	21:06	
Air Blank	0.000	21:07	
Air Blank	0.000	21:09	
Subject Sample	# 2	21:09	Subject Test Refused
Breath Volume	0.000	21:09	
Air Blank	0.000	21:09	
Control Test	0.080	21:10	
Air Blank	0.000	21:10	
Diagnostics Check	OK	21:10	

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 09/06/2025 02:20:59

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.000	02:16
Breath Volume	2.078	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.000	02:19
Breath Volume	1.519	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ORTIZ RAMOS

DATE/TIME OF TEST: 09/06/2025 03:33:43

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2	03:32	Volume Not Met
Breath Volume	0.761	03:32	
Air Blank	0.000	03:32	
Control Test	0.080	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DYKE,T,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/06/2025 04:11:46

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.079	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.173	04:07
Breath Volume	3.082	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.172	04:10
Breath Volume	2.753	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: HALTER

DATE/TIME OF TEST: 09/06/2025 20:18:20

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.223	20:13
Breath Volume	2.675	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.223	20:16
Breath Volume	2.417	20:16
Air Blank	0.000	20:17
Control Test	0.077	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ORBE

DATE/TIME OF TEST: 09/07/2025 01:05:26

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.097	01:00
Breath Volume	1.359	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.086	01:03
Breath Volume	1.218	01:03
Air Blank	0.000	01:04
Control Test	0.078	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MARTIN,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WYATT

DATE/TIME OF TEST: 09/07/2025 03:54:17

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.108	03:49
Breath Volume	2.394	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.098	03:52
Breath Volume	1.308	03:52
Air Blank	0.000	03:53
Control Test	0.078	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 09/29/2025 00:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 09/07/2025 23:31:12

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.118	23:26
Breath Volume	1.687	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.121	23:29
Breath Volume	1.121	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY, K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HINOJOSA

DATE/TIME OF TEST: 09/08/2025 01:34:15

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:30	No Sample Provided
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Air Blank	0.000	01:32	
Subject Sample	# 2	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.080	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CAMPBELL,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: ROZEWICZ

DATE/TIME OF TEST: 09/08/2025 12:59:41

Diagnostics Check	OK	12:53
Air Blank	0.000	12:53
Control Test	0.080	12:54
Air Blank	0.000	12:54
Subject Sample # 1	0.000	12:55
Breath Volume	2.253	12:55
Air Blank	0.000	12:55
Air Blank	0.000	12:57
Subject Sample # 2	0.000	12:58
Breath Volume	2.988	12:58
Air Blank	0.000	12:58
Control Test	0.080	12:59
Air Blank	0.000	12:59
Diagnostics Check	OK	12:59

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 09/13/2025 02:39:50

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.080	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:35	No Sample Provided
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Air Blank	0.000	02:38	
Subject Sample	# 2	02:38	Subject Test Refused
Breath Volume	0.000	02:38	
Air Blank	0.000	02:38	
Control Test	0.080	02:39	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BASOV,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: RINGER

DATE/TIME OF TEST: 09/13/2025 21:00:23

Diagnostics Check	OK	20:51	
Air Blank	0.000	20:52	
Control Test	0.080	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 1	20:56	No Sample Provided
Breath Volume	0.000	20:56	
Air Blank	0.000	20:56	
Air Blank	0.000	20:58	
Subject Sample	# 2	20:59	Subject Test Refused
Breath Volume	0.000	20:59	
Air Blank	0.000	20:59	
Control Test	0.080	20:59	
Air Blank	0.000	21:00	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MORANDO,B
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 09/14/2025 02:53:32

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.080	02:45	
Air Blank	0.000	02:46	
Subject Sample	# 1	02:49	No Sample Provided
Breath Volume	0.000	02:49	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 2	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ROY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RILEY,R
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: DICKERSON

DATE/TIME OF TEST: 09/14/2025 06:03:21

Diagnostics Check	OK	05:56
Air Blank	0.000	05:57
Control Test	0.080	05:57
Air Blank	0.000	05:58
Subject Sample # 1	0.211	05:58
Breath Volume	1.820	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.199	06:01
Breath Volume	2.277	06:01
Air Blank	0.000	06:02
Control Test	0.079	06:02
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: NUNEZ MOYA

DATE/TIME OF TEST: 09/16/2025 01:04:03

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.080	00:59
Breath Volume	2.054	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.086	01:02
Breath Volume	3.714	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:04

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HURLEY

DATE/TIME OF TEST: 09/17/2025 02:53:11

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.080	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.277	02:48
Breath Volume	1.214	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.294	02:51
Breath Volume	1.589	02:51
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 09/20/2025 01:54:01

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.203	01:49
Breath Volume	1.203	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:51
Subject Sample # 2	0.203	01:52
Breath Volume	1.269	01:52
Air Blank	0.000	01:53
Control Test	0.080	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:54

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WALDROP,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: BELLMAN

DATE/TIME OF TEST: 09/20/2025 03:42:03

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.110	03:37
Breath Volume	4.429	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.104	03:40
Breath Volume	2.921	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:42

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 09/20/2025 04:04:16

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:00	
Air Blank	0.000	04:02	
Subject Sample	# 2	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:03	
Control Test	0.080	04:03	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 09/21/2025 01:58:17

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.080	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:54	No Sample Provided
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 2	01:56	Subject Test Refused
Breath Volume	0.000	01:56	
Air Blank	0.000	01:57	
Control Test	0.079	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 09/21/2025 04:17:21

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.210	04:12
Breath Volume	2.226	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.207	04:15
Breath Volume	1.988	04:15
Air Blank	0.000	04:16
Control Test	0.079	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ERNEST,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: COLON,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 09/21/2025 08:31:30

Diagnostics Check	OK	08:27	
Air Blank	0.000	08:28	
Control Test	0.080	08:28	
Air Blank	0.000	08:29	
Subject Sample	# 1	08:30	Subject Test Refused
Breath Volume	0.000	08:30	
Air Blank	0.000	08:30	
Control Test	0.080	08:30	
Air Blank	0.000	08:31	
Diagnostics Check	OK	08:31	

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MARTINI

DATE/TIME OF TEST: 09/22/2025 01:30:53

Diagnostics Check	OK	01:24
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.086	01:25
Breath Volume	2.011	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.101	01:29
Breath Volume	4.164	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEUBAUER,J,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MINKOFF

DATE/TIME OF TEST: 09/24/2025 23:19:35

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:11	
Control Test	0.080	23:11	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:15	No Sample Provided
Breath Volume	0.000	23:15	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample	# 2	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:18	
Control Test	0.080	23:18	
Air Blank	0.000	23:19	
Diagnostics Check	OK	23:19	

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON, RONALD, R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MOSLEY

DATE/TIME OF TEST: 09/26/2025 03:01:44

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.080	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:57	No Sample Provided
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Air Blank	0.000	02:59	
Subject Sample	# 2	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.080	03:01	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTON,GARY,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GELSOMINO,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HEFFNER

DATE/TIME OF TEST: 09/26/2025 03:15:01

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.132	03:09
Breath Volume	1.500	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.123	03:13
Breath Volume	1.230	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:15

Data Download Date/Time: 09/29/2025 00:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SHEAFFER

DATE/TIME OF TEST: 09/05/2025 02:43:04

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:42	
Control Test	0.073	02:42	Control Outside Tolerance
Air Blank	0.000	02:43	

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SHEAFFER

DATE/TIME OF TEST: 09/05/2025 02:46:24

Diagnostics Check	OK	02:45	
Air Blank	0.000	02:45	
Control Test	0.074	02:45	Control Outside Tolerance
Air Blank	0.000	02:46	

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SHEAFFER

DATE/TIME OF TEST: 09/05/2025 02:49:20

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:48	
Control Test	0.072	02:48	Control Outside Tolerance
Air Blank	0.000	02:49	

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SHEAFFER

DATE/TIME OF TEST: 09/05/2025 03:05:30

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.075	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.126	03:00
Breath Volume	3.687	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.133	03:03
Breath Volume	3.628	03:03
Air Blank	0.000	03:04
Control Test	0.075	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LAGUARDIA

DATE/TIME OF TEST: 09/05/2025 22:27:00

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.077	22:21	
Air Blank	0.000	22:22	
Subject Sample # 1	0.067	22:22	
Breath Volume	4.339	22:22	
Air Blank	0.000	22:23	
Air Blank	0.000	22:24	
Subject Sample # 2	0.062	22:25	
Breath Volume	4.074	22:25	
Air Blank	0.000	22:26	
Control Test	0.074	22:26	Control Outside Tolerance
Air Blank	0.000	22:27	

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LAGUARDIA

DATE/TIME OF TEST: 09/05/2025 22:33:31

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:32	
Control Test	0.073	22:32	Control Outside Tolerance
Air Blank	0.000	22:33	

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LAGUARDIA

DATE/TIME OF TEST: 09/05/2025 22:48:03

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.076	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.058	22:43
Breath Volume	4.230	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.061	22:46
Breath Volume	3.949	22:46
Air Blank	0.000	22:47
Control Test	0.077	22:47
Air Blank	0.000	22:47
Diagnostics Check	OK	22:48

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: OSTAPUK

DATE/TIME OF TEST: 09/06/2025 01:43:27

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.075	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.069	01:38
Breath Volume	1.617	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.076	01:41
Breath Volume	2.195	01:41
Air Blank	0.000	01:42
Control Test	0.076	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 09/07/2025 00:01:51

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.076	23:55	
Air Blank	0.000	23:56	
Subject Sample # 2	0.120	23:57	
Breath Volume	3.687	23:57	
Air Blank	0.000	23:57	
Air Blank	0.000	23:59	
Subject Sample # 1	0.107	00:00	
Breath Volume	1.718	00:00	
Air Blank	0.000	00:00	
Control Test	0.074	00:01	Control Outside Tolerance
Air Blank	0.000	00:01	

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 09/07/2025 00:04:44

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.072	00:04	Control Outside Tolerance
Air Blank	0.000	00:04	

Data Download Date/Time: 09/10/2025 18:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 09/07/2025 00:19:32

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.075	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.102	00:14
Breath Volume	1.703	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.113	00:17
Breath Volume	2.277	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 09/10/2025 18:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVEY,ELIZABETH,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEWART,TYREN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231294
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: HEARD

DATE/TIME OF TEST: 09/11/2025 23:24:17

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.262	23:19
Breath Volume	2.152	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.247	23:22
Breath Volume	2.046	23:22
Air Blank	0.000	23:23
Control Test	0.077	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 09/28/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,JASON,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BOYETT,KATIE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CARRASCO DAZA

DATE/TIME OF TEST: 09/14/2025 03:34:11

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1 0.214	03:24	
Breath Volume	1.785	03:24	
Air Blank	0.000	03:25	
Air Blank	0.000	03:27	
Subject Sample	# 2	03:30	No Sample Provided
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Air Blank		03:32	RFI Detect
Air Blank	0.000	03:33	
Control Test	0.076	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 09/28/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON II,AUTIS,E
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JIN,KEVIN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: FIALA

DATE/TIME OF TEST: 09/26/2025 04:26:25

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.075	04:21
Breath Volume	2.425	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample # 2	0.077	04:24
Breath Volume	3.312	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 09/28/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,DAVIAN,L
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BLANCO,DAVIAN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CIATTEO

DATE/TIME OF TEST: 09/05/2025 22:27:06

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.246	22:22
Breath Volume	2.605	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:24
Subject Sample # 2	0.233	22:25
Breath Volume	2.804	22:25
Air Blank	0.000	22:26
Control Test	0.077	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 09/23/2025 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BRUTUS,FRANTZ
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 08/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 09/17/2025 21:00:12

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.233	20:55
Breath Volume	1.417	20:55
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.238	20:58
Breath Volume	1.734	20:58
Air Blank	0.000	20:59
Control Test	0.076	20:59
Air Blank	0.000	21:00
Diagnostics Check	OK	21:00

Data Download Date/Time: 09/23/2025 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEE,BRENNON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 09/02/2025 00:56:57

Diagnostics Check	OK	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Subject Sample # 1	0.303	00:52
Breath Volume	2.398	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.298	00:55
Breath Volume	1.949	00:55
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 09/16/2025 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINFREE,SAIGE,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WINFREE,SAIGE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: ROLOPH

DATE/TIME OF TEST: 09/05/2025 04:10:28

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:03	
Control Test	0.082	04:04	
Air Blank	0.000	04:04	
Subject Sample # 1	0.141	04:05	
Breath Volume	3.085	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:07	
Subject Sample # 2	0.135	04:08	
Breath Volume	2.871	04:08	
Air Blank	0.000	04:09	
Control Test	0.079	04:09	
Air Blank	0.000	04:09	
Diagnostics Check	Fail	04:09	Temperature Regulation Test
Air Blank	0.000	04:10	

Data Download Date/Time: 09/16/2025 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASH,JUSTIN,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NASH,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WOLTZ

DATE/TIME OF TEST: 09/05/2025 04:24:26

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.145	04:19
Breath Volume	1.261	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample # 2	0.127	04:22
Breath Volume	1.109	04:22
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 09/16/2025 12:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABEL,NICHOLAS,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BABEL,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:49

SUBJECT'S LAST NAME: VINES

DATE/TIME OF TEST: 09/05/2025 15:19:42

Diagnostics Check	OK	15:11
Air Blank	0.000	15:11
Control Test	0.080	15:12
Air Blank	0.000	15:12
Subject Sample # 1	0.076	15:14
Breath Volume	1.285	15:14
Air Blank	0.000	15:14
Air Blank	0.000	15:16
Subject Sample # 2	0.078	15:18
Breath Volume	1.109	15:18
Air Blank	0.000	15:18
Control Test	0.080	15:19
Air Blank	0.000	15:19
Diagnostics Check	OK	15:19

Data Download Date/Time: 09/16/2025 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEE,BRENNON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: HUFF-LEDET

DATE/TIME OF TEST: 09/06/2025 05:56:35

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.082	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.031	05:52
Breath Volume	1.882	05:52
Air Blank	0.000	05:52
Air Blank	0.000	05:54
Subject Sample # 2	0.028	05:55
Breath Volume	1.828	05:55
Air Blank	0.000	05:55
Control Test	0.082	05:55
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 09/16/2025 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTICE,JOSEPH,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JUSTICE,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 08:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:51

SUBJECT'S LAST NAME: SORTO BONILLA

DATE/TIME OF TEST: 09/07/2025 09:23:50

Diagnostics Check	OK	09:11	
Air Blank	0.000	09:12	
Control Test	0.081	09:12	
Air Blank	0.000	09:13	
Subject Sample	# 1 0.193	09:16	Volume Not Met
Breath Volume	0.398	09:16	
Air Blank	0.000	09:17	
Air Blank	0.000	09:18	
Subject Sample	# 2 0.190	09:19	
Breath Volume	2.093	09:19	
Air Blank	0.000	09:20	
Air Blank	0.000	09:21	
Subject Sample	# 3 0.187	09:22	
Breath Volume	1.968	09:22	
Air Blank	0.000	09:22	
Control Test	0.077	09:23	
Air Blank	0.000	09:23	
Diagnostics Check	OK	09:23	

Data Download Date/Time: 09/16/2025 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLTON,JUSTIN,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOLTON,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 09/14/2025 02:00:01

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.082	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.211	01:55
Breath Volume	1.230	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.203	01:58
Breath Volume	1.421	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	02:00

Data Download Date/Time: 09/16/2025 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDD,RACHEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: RUDD,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: COURSEY

DATE/TIME OF TEST: 09/14/2025 18:25:21

Diagnostics Check	OK	18:18
Air Blank	0.000	18:18
Control Test	0.082	18:18
Air Blank	0.000	18:19
Subject Sample # 1	0.228	18:20
Breath Volume	2.367	18:20
Air Blank	0.000	18:21
Air Blank	0.000	18:22
Subject Sample # 2	0.214	18:23
Breath Volume	1.570	18:23
Air Blank	0.000	18:24
Control Test	0.080	18:24
Air Blank	0.000	18:25
Diagnostics Check	OK	18:25

Data Download Date/Time: 09/16/2025 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRENNON,L
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEE,BRENNON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402987895
DRY GAS EXPIRY DATE: 03/06/2027
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: VARGAS TEJEDA

DATE/TIME OF TEST: 09/15/2025 01:37:48

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.082	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.170	01:33
Breath Volume	1.812	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.160	01:36
Breath Volume	1.531	01:36
Air Blank	0.000	01:36
Control Test	0.082	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 09/16/2025 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JULIAN,
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: WADDELL,JENEFFER
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 12:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:06

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 09/07/2025 13:44:33

Diagnostics Check	OK	13:34	
Air Blank	0.000	13:34	
Control Test	0.082	13:34	
Air Blank	0.000	13:35	
Subject Sample	# 1 0.390	13:36	No .020 agreement
Breath Volume	1.816	13:36	
Air Blank	0.000	13:37	
Air Blank	0.000	13:38	
Subject Sample	# 2 0.434	13:39	No .020 agreement
Breath Volume	3.445	13:39	
Air Blank	0.000	13:40	
Air Blank	0.000	13:41	
Subject Sample	# 3 0.412	13:42	No .020 agreement
Breath Volume	3.414	13:42	
Air Blank	0.000	13:43	
Control Test	0.079	13:43	
Air Blank	0.000	13:44	
Diagnostics Check	OK	13:44	

Data Download Date/Time: 10/31/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUENTES, BIANCA, M
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: REDWAN, NADER
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 09/19/2025 00:22:42

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.082	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:21	
Control Test	0.081	00:22	
Air Blank	0.000	00:22	
Diagnostics Check	OK	00:22	

Data Download Date/Time: 10/31/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUENTES, BIANCA, M
BREATH TEST OPERATOR AGENCY: PINECREST PD
ARRESTING OFFICER: POVEDA, REYNALDO
ARRESTING OFFICER AGENCY: PINECREST PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: QUIROZ OROZCO

DATE/TIME OF TEST: 09/19/2025 00:55:35

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.235	00:50
Breath Volume	3.343	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.236	00:53
Breath Volume	3.843	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 10/31/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUZMAN,JOSE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 09/03/2025 02:02:45

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:56
Air Blank	0.000	01:57
Subject Sample # 1	0.154	01:58
Breath Volume	3.027	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.148	02:01
Breath Volume	3.070	02:01
Air Blank	0.000	02:01
Control Test	0.077	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 10/01/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: IBARRA PARRALES

DATE/TIME OF TEST: 09/01/2025 04:29:45

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.248	04:25
Breath Volume	2.394	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.248	04:27
Breath Volume	2.156	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MEJIA SAUCEDO

DATE/TIME OF TEST: 09/02/2025 02:06:23

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.146	02:01
Breath Volume	2.613	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.145	02:04
Breath Volume	2.468	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MATTHEWS,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: NAVARRO JR

DATE/TIME OF TEST: 09/04/2025 21:28:09

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:22
Subject Sample # 1	0.152	21:23
Breath Volume	1.343	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:25
Subject Sample # 2	0.162	21:26
Breath Volume	2.218	21:26
Air Blank	0.000	21:27
Control Test	0.078	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 09/05/2025 01:22:54

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.206	01:17
Breath Volume	1.382	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.207	01:21
Breath Volume	1.410	01:21
Air Blank	0.000	01:21
Control Test	0.079	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AGAOGLU,YASEMIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MENDEZ SALCEDO

DATE/TIME OF TEST: 09/05/2025 04:35:48

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:31	
Control Test	0.079	04:31	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:34	Subject Test Refused
Breath Volume	0.000	04:34	
Air Blank	0.000	04:34	
Control Test	0.080	04:35	
Air Blank	0.000	04:35	
Diagnostics Check	OK	04:35	

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: VASQUEZ VASQUEZ

DATE/TIME OF TEST: 09/05/2025 20:30:37

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.080	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.169	20:24
Breath Volume	1.394	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:26
Subject Sample # 2	0.177	20:28
Breath Volume	1.386	20:28
Air Blank	0.000	20:29
Control Test	0.079	20:29
Air Blank	0.000	20:30
Diagnostics Check	OK	20:30

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BADOLATI, NICHOLAS
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/06/2025 02:09:32

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.080	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1 0.142	01:58	
Breath Volume	1.453	01:58	
Air Blank	0.000	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 2 0.127	02:03	Volume Not Met
Breath Volume	0.980	02:03	
Air Blank	0.000	02:04	
Air Blank	0.000	02:06	
Subject Sample	# 3 0.138	02:07	
Breath Volume	1.574	02:07	
Air Blank	0.000	02:08	
Control Test	0.076	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: MURILLO VELIZ

DATE/TIME OF TEST: 09/06/2025 04:42:58

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.166	04:35
Breath Volume	1.101	04:35
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.188	04:38
Breath Volume	1.820	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:40
Subject Sample # 3	0.186	04:41
Breath Volume	1.945	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ISASI FAVIE,DUANY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: TORSIELLO

DATE/TIME OF TEST: 09/06/2025 15:32:40

Diagnostics Check	OK	15:26
Air Blank	0.000	15:26
Control Test	0.079	15:26
Air Blank	0.000	15:27
Subject Sample # 1	0.091	15:28
Breath Volume	1.558	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.092	15:30
Breath Volume	2.164	15:30
Air Blank	0.000	15:31
Control Test	0.078	15:32
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JUAN ESTEBAN

DATE/TIME OF TEST: 09/06/2025 22:36:01

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.079	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.228	22:31
Breath Volume	1.890	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:33
Subject Sample # 2	0.243	22:34
Breath Volume	2.171	22:34
Air Blank	0.000	22:35
Control Test	0.077	22:35
Air Blank	0.000	22:35
Diagnostics Check	OK	22:36

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EUTENEUER,BRICE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: KIRTON

DATE/TIME OF TEST: 09/07/2025 00:57:04

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.000	00:52
Breath Volume	4.468	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.000	00:55
Breath Volume	4.765	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:57

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STROUD,ELIZABETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SMITH JR

DATE/TIME OF TEST: 09/07/2025 02:56:10

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.216	02:50
Breath Volume	2.968	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.205	02:54
Breath Volume	1.484	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/08/2025 05:49:31

Diagnostics Check	OK	05:43
Air Blank	0.000	05:43
Control Test	0.080	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.000	05:45
Breath Volume	2.855	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.000	05:48
Breath Volume	3.253	05:48
Air Blank	0.000	05:48
Control Test	0.080	05:48
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ORTEGA,JUAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: AGENORD

DATE/TIME OF TEST: 09/08/2025 19:12:39

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.080	19:06
Air Blank	0.000	19:07
Subject Sample # 1	0.140	19:07
Breath Volume	1.695	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.141	19:10
Breath Volume	1.644	19:10
Air Blank	0.000	19:11
Control Test	0.079	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: TETA

DATE/TIME OF TEST: 09/09/2025 17:10:00

Diagnostics Check	OK	17:03
Air Blank	0.000	17:03
Control Test	0.080	17:04
Air Blank	0.000	17:04
Subject Sample # 1	0.000	17:05
Breath Volume	1.164	17:05
Air Blank	0.000	17:06
Air Blank	0.000	17:08
Subject Sample # 2	0.000	17:08
Breath Volume	1.503	17:08
Air Blank	0.000	17:09
Control Test	0.080	17:09
Air Blank	0.000	17:09
Diagnostics Check	OK	17:10

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FREDERICK

DATE/TIME OF TEST: 09/10/2025 04:16:20

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.251	04:10
Breath Volume	1.417	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.262	04:14
Breath Volume	2.351	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO, JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: JANSEN

DATE/TIME OF TEST: 09/10/2025 19:30:10

Diagnostics Check	OK	19:23
Air Blank	0.000	19:24
Control Test	0.080	19:24
Air Blank	0.000	19:25
Subject Sample # 1	0.189	19:25
Breath Volume	1.425	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:27
Subject Sample # 2	0.181	19:28
Breath Volume	2.656	19:28
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MARTINEZ NOTARIO

DATE/TIME OF TEST: 09/11/2025 02:00:45

Diagnostics Check	OK	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.174	01:55
Breath Volume	2.636	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.174	01:58
Breath Volume	3.738	01:58
Air Blank	0.000	01:59
Control Test	0.079	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: JUSTIN

DATE/TIME OF TEST: 09/12/2025 02:44:09

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.111	02:39
Breath Volume	2.144	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:41
Subject Sample # 2	0.111	02:42
Breath Volume	1.574	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 09/12/2025 14:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FARDELLA,BLAKE
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:04

SUBJECT'S LAST NAME: ANDRES FELIX

DATE/TIME OF TEST: 09/12/2025 17:41:00

Diagnostics Check	OK	17:34
Air Blank	0.000	17:34
Control Test	0.080	17:35
Air Blank	0.000	17:35
Subject Sample # 1	0.158	17:36
Breath Volume	1.785	17:36
Air Blank	0.000	17:37
Air Blank	0.000	17:38
Subject Sample # 2	0.165	17:39
Breath Volume	2.085	17:39
Air Blank	0.000	17:40
Control Test	0.078	17:40
Air Blank	0.000	17:40
Diagnostics Check	OK	17:41

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUND,PARIS,D
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MELLERS,KEVIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 09/12/2025 19:29:28

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.081	19:23
Air Blank	0.000	19:23
Subject Sample # 1	0.000	19:25
Breath Volume	1.257	19:25
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.000	19:28
Breath Volume	1.335	19:28
Air Blank	0.000	19:28
Control Test	0.081	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: VICKERS

DATE/TIME OF TEST: 09/12/2025 20:38:07

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.080	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.184	20:32
Breath Volume	1.515	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.187	20:36
Breath Volume	1.933	20:36
Air Blank	0.000	20:37
Control Test	0.078	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MANNING, WESLEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/12/2025 22:16:00

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.079	22:08
Air Blank	0.000	22:08
Subject Sample # 1	0.180	22:11
Breath Volume	1.414	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:13
Subject Sample # 2	0.181	22:14
Breath Volume	1.957	22:14
Air Blank	0.000	22:15
Control Test	0.076	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:16

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 09/13/2025 01:30:17

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.193	01:25
Breath Volume	2.523	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:27
Subject Sample # 2	0.178	01:28
Breath Volume	2.109	01:28
Air Blank	0.000	01:29
Control Test	0.078	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NORMIL,RITCHY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SCOTT JR

DATE/TIME OF TEST: 09/13/2025 03:26:17

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.116	03:21
Breath Volume	1.429	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:23
Subject Sample # 2	0.119	03:24
Breath Volume	1.359	03:24
Air Blank	0.000	03:25
Control Test	0.078	03:25
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 09/13/2025 04:57:30

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.153	04:52
Breath Volume	1.468	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.165	04:55
Breath Volume	2.218	04:55
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: LOPEZ ORTIZ

DATE/TIME OF TEST: 09/14/2025 02:39:15

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.191	02:34
Breath Volume	2.097	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.195	02:37
Breath Volume	2.074	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZIMMERMAN,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 09/14/2025 02:56:21

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.191	02:51
Breath Volume	2.437	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.198	02:54
Breath Volume	1.593	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: GODINEZ PEREZ

DATE/TIME OF TEST: 09/15/2025 01:22:47

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.235	01:18
Breath Volume	3.050	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.234	01:20
Breath Volume	2.871	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RODRIGUEZ,ANDRES
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: OLVERA CASAS

DATE/TIME OF TEST: 09/15/2025 04:14:38

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.157	04:09
Breath Volume	1.886	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.160	04:12
Breath Volume	2.589	04:12
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HUGHES,RONNELLE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/16/2025 23:57:41

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.196	23:51
Breath Volume	1.429	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:53
Subject Sample # 2	0.198	23:55
Breath Volume	1.406	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GREEN, TRAVIS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GROBLER

DATE/TIME OF TEST: 09/19/2025 03:07:06

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.151	03:01
Breath Volume	2.277	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.151	03:05
Breath Volume	1.652	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PIDGEON

DATE/TIME OF TEST: 09/19/2025 03:59:56

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.187	03:55
Breath Volume	2.519	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:57
Subject Sample # 2	0.177	03:58
Breath Volume	2.546	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HACHE,RANDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: PADILLA ANILLO

DATE/TIME OF TEST: 09/19/2025 15:41:07

Diagnostics Check	OK	15:34
Air Blank	0.000	15:34
Control Test	0.080	15:34
Air Blank	0.000	15:35
Subject Sample # 1	0.209	15:36
Breath Volume	1.546	15:36
Air Blank	0.000	15:37
Air Blank	0.000	15:38
Subject Sample # 2	0.204	15:39
Breath Volume	1.746	15:39
Air Blank	0.000	15:40
Control Test	0.078	15:40
Air Blank	0.000	15:41
Diagnostics Check	OK	15:41

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MARTINEZ MARTINEZ

DATE/TIME OF TEST: 09/19/2025 22:23:40

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.206	22:18
Breath Volume	1.957	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.207	22:21
Breath Volume	1.425	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 09/20/2025 00:52:01

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.175	00:47
Breath Volume	2.105	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.177	00:50
Breath Volume	2.117	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:52

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 09/20/2025 23:34:34

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.000	23:30
Breath Volume	2.125	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.000	23:33
Breath Volume	1.179	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: FRANCISCO RAMON

DATE/TIME OF TEST: 09/21/2025 00:31:48

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.080	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.194	00:27
Breath Volume	2.226	00:27
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.213	00:30
Breath Volume	2.929	00:30
Air Blank	0.000	00:30
Control Test	0.079	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCDONOUGH,M
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MONTES BONILLA

DATE/TIME OF TEST: 09/21/2025 02:16:39

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.081	02:10
Air Blank	0.000	02:11
Subject Sample # 1	0.108	02:12
Breath Volume	2.382	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.115	02:14
Breath Volume	1.992	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DICKERSON,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SANCHEZ LOPEZ

DATE/TIME OF TEST: 09/21/2025 02:55:33

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.101	02:50
Breath Volume	1.523	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.091	02:53
Breath Volume	1.992	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MARTIN RAMIREZ,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: GABELLO RODRIGUEZ

DATE/TIME OF TEST: 09/21/2025 18:51:08

Diagnostics Check	OK	18:43
Air Blank	0.000	18:44
Control Test	0.081	18:44
Air Blank	0.000	18:45
Subject Sample # 1	0.164	18:46
Breath Volume	1.367	18:46
Air Blank	0.000	18:47
Air Blank	0.000	18:48
Subject Sample # 2	0.160	18:49
Breath Volume	1.386	18:49
Air Blank	0.000	18:50
Control Test	0.079	18:50
Air Blank	0.000	18:51
Diagnostics Check	OK	18:51

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: HAYE

DATE/TIME OF TEST: 09/22/2025 22:59:52

Diagnostics Check	OK	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.161	22:55
Breath Volume	1.167	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:57
Subject Sample # 2	0.160	22:58
Breath Volume	1.156	22:58
Air Blank	0.000	22:58
Control Test	0.077	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MADELIN,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: SEYMORE

DATE/TIME OF TEST: 09/23/2025 05:58:03

Diagnostics Check	OK	05:50
Air Blank	0.000	05:50
Control Test	0.081	05:51
Air Blank	0.000	05:51
Subject Sample # 1	0.207	05:53
Breath Volume	1.914	05:53
Air Blank	0.000	05:54
Air Blank	0.000	05:55
Subject Sample # 2	0.213	05:56
Breath Volume	1.585	05:56
Air Blank	0.000	05:57
Control Test	0.079	05:57
Air Blank	0.000	05:57
Diagnostics Check	OK	05:58

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, THOMASINA, E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 09/23/2025 07:36:24

Diagnostics Check	OK	07:28
Air Blank	0.000	07:29
Control Test	0.080	07:29
Air Blank	0.000	07:30
Subject Sample # 1	0.157	07:30
Breath Volume	2.421	07:30
Air Blank	0.000	07:31
Air Blank	0.000	07:33
Subject Sample # 2	0.148	07:34
Breath Volume	2.187	07:34
Air Blank	0.000	07:35
Control Test	0.079	07:35
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RAMIREZ,JAMIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: LOWRY

DATE/TIME OF TEST: 09/24/2025 10:07:14

Diagnostics Check	OK	09:57
Air Blank	0.000	09:57
Control Test	0.080	09:58
Air Blank	0.000	09:58
Subject Sample # 1	0.416	09:59
Breath Volume	1.992	09:59
Air Blank	0.000	10:00
Air Blank	0.000	10:01
Subject Sample # 2	0.388	10:02
Breath Volume	2.500	10:02
Air Blank	0.000	10:03
Air Blank	0.000	10:04
Subject Sample # 3	0.393	10:05
Breath Volume	2.601	10:05
Air Blank	0.000	10:06
Control Test	0.077	10:06
Air Blank	0.000	10:07
Diagnostics Check	OK	10:07

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOUNG,ALEX
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:06

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 09/24/2025 16:09:09

Diagnostics Check	OK	16:03
Air Blank	0.000	16:03
Control Test	0.079	16:04
Air Blank	0.000	16:04
Subject Sample # 1	0.000	16:04
Breath Volume	1.363	16:04
Air Blank	0.000	16:05
Air Blank	0.000	16:07
Subject Sample # 2	0.000	16:07
Breath Volume	1.574	16:07
Air Blank	0.000	16:08
Control Test	0.079	16:08
Air Blank	0.000	16:09
Diagnostics Check	OK	16:09

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN,RENEE,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EUTENEUER,BRICE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/24/2025 23:56:05

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1 0.167	23:47	Volume Not Met
Breath Volume	0.957	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:50	
Subject Sample	# 2 0.175	23:51	
Breath Volume	1.253	23:51	
Air Blank	0.000	23:51	
Air Blank	0.000	23:53	
Subject Sample	# 3 0.171	23:54	
Breath Volume	1.265	23:54	
Air Blank	0.000	23:55	
Control Test	0.077	23:55	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 09/25/2025 04:47:29

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.198	04:36
Breath Volume	1.730	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.177	04:41
Breath Volume	1.371	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 3	0.168	04:45
Breath Volume	1.160	04:45
Air Blank	0.000	04:46
Control Test	0.077	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 10/17/2025 14:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 09/26/2025 04:32:56

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.231	04:27
Breath Volume	1.585	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.230	04:31
Breath Volume	1.347	04:31
Air Blank	0.000	04:31
Control Test	0.078	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: RAMIREZ PASCUAL

DATE/TIME OF TEST: 09/28/2025 00:53:53

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.028	00:49
Breath Volume	1.621	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:51
Subject Sample # 2	0.026	00:52
Breath Volume	1.261	00:52
Air Blank	0.000	00:52
Control Test	0.080	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 09/28/2025 02:37:57

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.081	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.119	02:33
Breath Volume	1.558	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.116	02:36
Breath Volume	2.281	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: GARCIA PEREZ

DATE/TIME OF TEST: 09/28/2025 03:24:02

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.076	03:18
Breath Volume	1.355	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.077	03:22
Breath Volume	1.488	03:22
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:24

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TOMAS ANTONIO,ANIBAL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:12

SUBJECT'S LAST NAME: FRANCISCO DIEGO

DATE/TIME OF TEST: 09/28/2025 09:50:39

Diagnostics Check	OK	09:43
Air Blank	0.000	09:44
Control Test	0.080	09:44
Air Blank	0.000	09:45
Subject Sample # 1	0.218	09:46
Breath Volume	2.816	09:46
Air Blank	0.000	09:46
Air Blank	0.000	09:48
Subject Sample # 2	0.210	09:48
Breath Volume	1.843	09:48
Air Blank	0.000	09:49
Control Test	0.078	09:49
Air Blank	0.000	09:50
Diagnostics Check	OK	09:50

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NUNEZ,BRENDA
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: HERNANDEZ HERNANDEZ

DATE/TIME OF TEST: 09/28/2025 10:19:04

Diagnostics Check	OK	10:11
Air Blank	0.000	10:12
Control Test	0.079	10:12
Air Blank	0.000	10:13
Subject Sample # 1	0.221	10:13
Breath Volume	2.316	10:13
Air Blank	0.000	10:14
Air Blank	0.000	10:16
Subject Sample # 2	0.213	10:17
Breath Volume	2.289	10:17
Air Blank	0.000	10:18
Control Test	0.078	10:18
Air Blank	0.000	10:18
Diagnostics Check	OK	10:19

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: RAMIREZ OLIVERA

DATE/TIME OF TEST: 09/28/2025 19:04:14

Diagnostics Check	OK	18:57
Air Blank	0.000	18:57
Control Test	0.079	18:58
Air Blank	0.000	18:58
Subject Sample # 1	0.306	18:59
Breath Volume	2.320	18:59
Air Blank	0.000	19:00
Air Blank	0.000	19:01
Subject Sample # 2	0.308	19:02
Breath Volume	3.179	19:02
Air Blank	0.000	19:03
Control Test	0.077	19:03
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON,JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: ARCHILA GARCIA

DATE/TIME OF TEST: 09/28/2025 23:29:40

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.185	23:24
Breath Volume	1.394	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.181	23:27
Breath Volume	1.339	23:27
Air Blank	0.000	23:28
Control Test	0.079	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:29

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JOSEPH,WADSON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 09/30/2025 00:29:06

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.226	00:24
Breath Volume	1.835	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:26
Subject Sample # 2	0.220	00:27
Breath Volume	1.152	00:27
Air Blank	0.000	00:28
Control Test	0.077	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS,KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: SANCHEZ BARCENA

DATE/TIME OF TEST: 09/30/2025 20:50:04

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:42
Air Blank	0.000	20:42
Subject Sample # 1	0.293	20:45
Breath Volume	1.992	20:45
Air Blank	0.000	20:46
Air Blank	0.000	20:47
Subject Sample # 2	0.286	20:48
Breath Volume	1.550	20:48
Air Blank	0.000	20:49
Control Test	0.075	20:49
Air Blank	0.000	20:49
Diagnostics Check	OK	20:50

Data Download Date/Time: 10/17/2025 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ORTEGA,JUAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: PASCUAL PEDRO

DATE/TIME OF TEST: 09/01/2025 02:47:25

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.176	02:42
Breath Volume	1.566	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.177	02:45
Breath Volume	2.542	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: FELIZ II

DATE/TIME OF TEST: 09/01/2025 04:14:15

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.079	04:08	
Air Blank	0.000	04:09	
Subject Sample # 1	0.324	04:09	
Breath Volume	2.140	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample # 2		04:12	Slope Not Met
Breath Volume	0.843	04:12	
Air Blank	0.000	04:13	
Control Test	0.078	04:13	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: FELIZ II

DATE/TIME OF TEST: 09/01/2025 04:44:00

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.298	04:39
Breath Volume	1.632	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:41
Subject Sample # 2	0.285	04:42
Breath Volume	1.707	04:42
Air Blank	0.000	04:43
Control Test	0.078	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:44

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EDWARDS, DERRICK
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: STRONG

DATE/TIME OF TEST: 09/01/2025 11:34:39

Diagnostics Check	OK	11:27
Air Blank	0.000	11:28
Control Test	0.079	11:28
Air Blank	0.000	11:29
Subject Sample # 1	0.317	11:29
Breath Volume	1.222	11:29
Air Blank	0.000	11:30
Air Blank	0.000	11:32
Subject Sample # 2	0.337	11:32
Breath Volume	2.683	11:32
Air Blank	0.000	11:33
Control Test	0.077	11:34
Air Blank	0.000	11:34
Diagnostics Check	OK	11:34

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARVER, CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/01/2025 18:38:44

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.079	18:31
Air Blank	0.000	18:32
Subject Sample # 1	0.248	18:33
Breath Volume	1.605	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.243	18:37
Breath Volume	1.457	18:37
Air Blank	0.000	18:37
Control Test	0.076	18:38
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: NAVAS

DATE/TIME OF TEST: 09/01/2025 20:38:23

Diagnostics Check	OK	20:26	
Air Blank	0.000	20:27	
Control Test	0.079	20:27	
Air Blank	0.000	20:27	
Subject Sample	# 1 0.301	20:31	Volume Not Met
Breath Volume	0.308	20:31	
Air Blank	0.000	20:31	
Air Blank	0.000	20:33	
Subject Sample	# 2 0.013	20:36	Volume Not Met
Breath Volume	0.187	20:36	
Air Blank	0.000	20:37	
Control Test	0.075	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN, JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 09/02/2025 16:05:51

Diagnostics Check	OK	15:51	
Air Blank	0.000	15:52	
Control Test	0.079	15:52	
Air Blank	0.000	15:53	
Subject Sample	# 1 0.319	15:54	
Breath Volume	1.105	15:54	
Air Blank	0.000	15:55	
Air Blank	0.000	15:57	
Subject Sample	# 2 0.305	16:00	Volume Not Met
Breath Volume	0.070	16:00	
Air Blank	0.000	16:01	
Air Blank	0.000	16:02	
Subject Sample	# 3 0.304	16:04	
Breath Volume	1.101	16:04	
Air Blank	0.000	16:04	
Control Test	0.072	16:05	Control Outside Tolerance
Air Blank	0.000	16:05	

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN, JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 09/02/2025 16:38:48

Diagnostics Check	OK	16:29	
Air Blank	0.000	16:30	
Control Test	0.076	16:30	
Air Blank	0.000	16:30	
Subject Sample	# 1 0.287	16:34	Volume Not Met
Breath Volume	0.601	16:34	
Air Blank	0.000	16:34	
Air Blank	0.000	16:36	
Subject Sample	# 2	16:37	Subject Test Refused
Breath Volume	0.000	16:37	
Air Blank	0.000	16:37	
Control Test	0.075	16:38	
Air Blank	0.000	16:38	
Diagnostics Check	OK	16:38	

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HECTOR,BRADY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: GILLIAND

DATE/TIME OF TEST: 09/03/2025 23:05:19

Diagnostics Check	OK	22:58
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.068	23:00
Breath Volume	3.289	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.067	23:03
Breath Volume	3.070	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THOMASINA,E
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: HILLOCK

DATE/TIME OF TEST: 09/04/2025 03:27:17

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:20
Air Blank	0.000	03:21
Subject Sample # 1	0.123	03:22
Breath Volume	1.914	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.126	03:25
Breath Volume	1.457	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: SZEWCZYK

DATE/TIME OF TEST: 09/04/2025 23:46:09

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.080	23:42	
Air Blank	0.000	23:43	
Subject Sample	# 1	23:44	Subject Test Refused
Breath Volume	0.000	23:44	
Air Blank	0.000	23:45	
Control Test	0.080	23:45	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAFLEUR,ROBENSON
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: LAFALAISE

DATE/TIME OF TEST: 09/05/2025 23:52:13

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.191	23:47
Breath Volume	1.117	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.202	23:50
Breath Volume	1.746	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAPPIN,KEVIN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: ALONZO MARTINEZ

DATE/TIME OF TEST: 09/06/2025 01:41:18

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.244	01:36
Breath Volume	2.351	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:38
Subject Sample # 2	0.250	01:39
Breath Volume	3.273	01:39
Air Blank	0.000	01:40
Control Test	0.078	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONDONO, JULIAN
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: SAINT-JULIEN

DATE/TIME OF TEST: 09/06/2025 23:31:58

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:20	
Control Test	0.079	23:21	
Air Blank	0.000	23:21	
Subject Sample	# 1 0.165	23:24	Volume Not Met
Breath Volume	0.015	23:24	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 2 0.186	23:30	Slope Not Level
Breath Volume	1.875	23:30	
Air Blank	0.000	23:31	
Control Test	0.077	23:31	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 09/12/2025 17:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL, KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 09/07/2025 01:37:32

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.215	01:32
Breath Volume	2.843	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.213	01:35
Breath Volume	2.054	01:35
Air Blank	0.000	01:36
Control Test	0.077	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLACKMAN, VICTOR
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/07/2025 02:17:30

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.077	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.175	02:10
Breath Volume	1.335	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.172	02:15
Breath Volume	1.664	02:15
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FLINK, ANDREW
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CHAJ-GONZALEZ

DATE/TIME OF TEST: 09/07/2025 04:11:02

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.129	04:06
Breath Volume	1.632	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.128	04:09
Breath Volume	1.234	04:09
Air Blank	0.000	04:10
Control Test	0.077	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:11

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ORTIZ,JOSE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: ERIL

DATE/TIME OF TEST: 09/07/2025 05:08:12

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.079	05:01
Air Blank	0.000	05:01
Subject Sample # 1	0.147	05:02
Breath Volume	1.382	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:04
Subject Sample # 2	0.140	05:06
Breath Volume	2.183	05:06
Air Blank	0.000	05:07
Control Test	0.078	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT, ALDOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: MERCHANT

DATE/TIME OF TEST: 09/07/2025 20:00:06

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:55	
Control Test	0.078	19:55	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:58	Subject Test Refused
Breath Volume	0.695	19:58	
Air Blank	0.000	19:59	
Control Test	0.079	19:59	
Air Blank	0.000	20:00	
Diagnostics Check	OK	20:00	

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CANTU RAMIREZ

DATE/TIME OF TEST: 09/08/2025 02:26:50

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.116	02:21
Breath Volume	2.070	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.112	02:25
Breath Volume	1.621	02:25
Air Blank	0.000	02:25
Control Test	0.078	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARRERAS, KIMBERLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: DESIR

DATE/TIME OF TEST: 09/09/2025 01:43:47

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.165	01:37
Breath Volume	1.246	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.163	01:42
Breath Volume	1.769	01:42
Air Blank	0.000	01:42
Control Test	0.078	01:43
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEAHEY, THOMAS, H
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: RAMOS LOPEZ

DATE/TIME OF TEST: 09/09/2025 22:25:03

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.173	22:20
Breath Volume	1.390	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:22
Subject Sample # 2	0.178	22:23
Breath Volume	2.164	22:23
Air Blank	0.000	22:24
Control Test	0.077	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:25

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALSTON,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ZAVALETA-VISOSO

DATE/TIME OF TEST: 09/10/2025 03:31:27

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:26
Subject Sample # 1	0.113	03:26
Breath Volume	1.425	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.114	03:29
Breath Volume	2.261	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: CALLWOOD

DATE/TIME OF TEST: 09/10/2025 21:03:29

Diagnostics Check	OK	20:57
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.000	20:59
Breath Volume	1.214	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.000	21:02
Breath Volume	1.171	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GRAMAJO DE LEON

DATE/TIME OF TEST: 09/11/2025 00:53:15

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.079	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1 0.121	00:46	Volume Not Met
Breath Volume	0.375	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2 0.115	00:51	Volume Not Met
Breath Volume	0.332	00:51	
Air Blank	0.000	00:52	
Control Test	0.078	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,TIA,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KAHN,JUSTIN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: GRAMAJO DE LEON

DATE/TIME OF TEST: 09/11/2025 01:01:48

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.077	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.132	00:57
Breath Volume	1.972	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.128	01:00
Breath Volume	1.800	01:00
Air Blank	0.000	01:00
Control Test	0.078	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: SAXON

DATE/TIME OF TEST: 09/11/2025 20:28:19

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.275	20:23
Breath Volume	1.976	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.272	20:26
Breath Volume	2.046	20:26
Air Blank	0.000	20:27
Control Test	0.078	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIER, MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: DE LA CRUZ GARCIA

DATE/TIME OF TEST: 09/11/2025 21:28:04

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.078	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.196	21:22
Breath Volume	1.433	21:22
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.207	21:26
Breath Volume	1.207	21:26
Air Blank	0.000	21:27
Control Test	0.078	21:27
Air Blank	0.000	21:27
Diagnostics Check	OK	21:28

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CISSON, JACKIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: LOPEZ DIAZ

DATE/TIME OF TEST: 09/11/2025 22:15:27

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.078	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.241	22:10
Breath Volume	2.546	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.236	22:13
Breath Volume	2.250	22:13
Air Blank	0.000	22:14
Control Test	0.077	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 09/12/2025 17:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGIN, RENEE, M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE, EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: LUX IXCOTOYAC

DATE/TIME OF TEST: 09/12/2025 02:59:03

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.254	02:54
Breath Volume	1.207	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:56
Subject Sample # 2	0.249	02:57
Breath Volume	1.601	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:59

Data Download Date/Time: 09/12/2025 17:08