

Alcohol Testing Program - Instrument Processing Sheet

Agency: SARASOTA COUNTY SO Instrument Serial Number: 80-001346

Date In: 8/21/2026 DI Completion Date: 8/28/26 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>WKP</u> Date: <u>8/21/2026</u>		Quality Checks By: <u>IAS</u> Date: <u>8/26/2026</u>		Flow Adjustment By: _____																																											
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>196</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP102</u> 32 mm .148 (.139-.169) 36 mm .160 (.156-.190) 53 mm .238 (.228-.278) 103 mm .488 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1012</u> Instrument: <u>1015</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																											
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23725080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP5088	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP5089	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP5090	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	23725080A1		10/5/2027													
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																																						
						0.050	MP5088	202603A	☐ Battery Replacement																																						
								3/3/2028	☐ Dry Gas Regulator Replacement																																						
0.080	MP5089	202603B	☐ Tank Sensor Tare																																												
		3/3/2028	☐ Breath Tube Replacement																																												
0.200	MP5090	202603D	☐ Other:																																												
		3/4/2028																																													
0.080 DGS	N/A	23725080A1																																													
		10/5/2027																																													
Optical Bench Adjustment By: _____ Department Inspection By: <u>KTS</u> Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: <u>28427</u>																																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1011</u> Instrument: <u>1014</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>																	
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
☐ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☒ Average Calculations ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:	
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Gauge ID #: _____ Gauge: _____ Instrument: _____		Notes/Suggested Service:		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																											
Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.09.03 10:24:39 -04'00'		S. LeAndra Higginbotham Digitally signed by S. LeAndra Higginbotham Date: 2026.09.03 10:43:12 -04'00'		Tech Review Admin Review																																											

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
SARASOTA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001346 18/26/2026 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>19:37</td></tr><tr><td>Control Test</td><td>0.050</td><td>19:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:38</td></tr><tr><td>Control Test</td><td>0.050</td><td>19:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:39</td></tr><tr><td>Control Test</td><td>0.050</td><td>19:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:41</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	19:37	Control Test	0.050	19:38	Air Blank	0.000	19:38	Control Test	0.050	19:39	Air Blank	0.000	19:39	Control Test	0.050	19:40	Air Blank	0.000	19:41	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		SARASOTA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001346 18/26/2026 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>19:43</td></tr><tr><td>Control Test</td><td>0.079</td><td>19:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:44</td></tr><tr><td>Control Test</td><td>0.079</td><td>19:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:45</td></tr><tr><td>Control Test</td><td>0.079</td><td>19:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:47</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0791</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	19:43	Control Test	0.079	19:44	Air Blank	0.000	19:44	Control Test	0.079	19:45	Air Blank	0.000	19:45	Control Test	0.079	19:46	Air Blank	0.000	19:47	Control Test Stats			Average	0.0791		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		SARASOTA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001346 18/26/2026 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>19:19</td></tr><tr><td>Control Test</td><td>0.200</td><td>19:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:20</td></tr><tr><td>Control Test</td><td>0.200</td><td>19:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:21</td></tr><tr><td>Control Test</td><td>0.200</td><td>19:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:22</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2000</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev(%)</td><td>0.2882</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	19:19	Control Test	0.200	19:20	Air Blank	0.000	19:20	Control Test	0.200	19:21	Air Blank	0.000	19:21	Control Test	0.200	19:22	Air Blank	0.000	19:22	Control Test Stats			Average	0.2000		Std Dev	0.0000		Rel. Std Dev(%)	0.2882		SARASOTA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001346 18/26/2026 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>19:27</td></tr><tr><td>Control Test</td><td>0.078</td><td>19:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:28</td></tr><tr><td>Control Test</td><td>0.079</td><td>19:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:29</td></tr><tr><td>Control Test</td><td>0.079</td><td>19:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>19:29</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev(%)</td><td>0.7339</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	19:27	Control Test	0.078	19:27	Air Blank	0.000	19:28	Control Test	0.079	19:28	Air Blank	0.000	19:29	Control Test	0.079	19:29	Air Blank	0.000	19:29	Control Test Stats			Average	0.0787		Std Dev	0.0000		Rel. Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:37																																																																																																																																																	
Control Test	0.050	19:38																																																																																																																																																	
Air Blank	0.000	19:38																																																																																																																																																	
Control Test	0.050	19:39																																																																																																																																																	
Air Blank	0.000	19:39																																																																																																																																																	
Control Test	0.050	19:40																																																																																																																																																	
Air Blank	0.000	19:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:43																																																																																																																																																	
Control Test	0.079	19:44																																																																																																																																																	
Air Blank	0.000	19:44																																																																																																																																																	
Control Test	0.079	19:45																																																																																																																																																	
Air Blank	0.000	19:45																																																																																																																																																	
Control Test	0.079	19:46																																																																																																																																																	
Air Blank	0.000	19:47																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0791																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:19																																																																																																																																																	
Control Test	0.200	19:20																																																																																																																																																	
Air Blank	0.000	19:20																																																																																																																																																	
Control Test	0.200	19:21																																																																																																																																																	
Air Blank	0.000	19:21																																																																																																																																																	
Control Test	0.200	19:22																																																																																																																																																	
Air Blank	0.000	19:22																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2000																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.2882																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	19:27																																																																																																																																																	
Control Test	0.078	19:27																																																																																																																																																	
Air Blank	0.000	19:28																																																																																																																																																	
Control Test	0.079	19:28																																																																																																																																																	
Air Blank	0.000	19:29																																																																																																																																																	
Control Test	0.079	19:29																																																																																																																																																	
Air Blank	0.000	19:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.7339																																																																																																																																																		
Anabella Anyaforwa Operator's Signature	Anabella Anyaforwa Operator's Signature	Anabella Anyaforwa Operator's Signature	Anabella Anyaforwa Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 10:36

Date of Inspection: 08/28/2026

Serial Number: 80-001346
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.081
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.051	0.079	0.200	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.200	0.080


Standard Deviations	0.0005	0.0005	0.0007	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN
Signature and Printed Name

08/28/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001346, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001346</u>	Average of y $\pm$ Uncertainty*
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/28/2026</u>	0.079 g/ 210 L 0.004
Calibration Time:	<u>10:36</u>	0.199g/ 210 L 0.008
		0.080g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Digitally signed by Kaitlyn
Spearin
Date: 2026.08.31 13:06:56
-04'00'

Kaitlyn Spearin

08/28/2026

Date

KATIE T SPEARIN,

Department Inspector

FDLE/ATP Form 69 August 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality