# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LAMPARTER

DATE/TIME OF TEST: 05/02/2026 01:56:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Control Test | | 0.081 | 01:51 |
| Air Blank | | 0.000 | 01:52 |
| Subject Sample | # 1 | 0.207 | 01:52 |
| Breath Volume | | 2.039 | 01:52 |
| Air Blank | | 0.000 | 01:53 |
| Air Blank | | 0.000 | 01:54 |
| Subject Sample | # 2 | 0.196 | 01:55 |
| Breath Volume | | 2.066 | 01:55 |
| Air Blank | | 0.000 | 01:55 |
| Control Test | | 0.079 | 01:56 |
| Air Blank | | 0.000 | 01:56 |
| Diagnostics Check | | OK | 01:56 |

Data Download Date/Time: 05/22/2026 14:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONNET,GREGORY,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: PALACIOS,GLORIA
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: CARINO

DATE/TIME OF TEST: 05/02/2026 07:27:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 07:20 |
| Air Blank | | 0.000 | 07:21 |
| Control Test | | 0.081 | 07:21 |
| Air Blank | | 0.000 | 07:22 |
| Subject Sample | # 1 | 0.000 | 07:22 |
| Breath Volume | | 1.433 | 07:22 |
| Air Blank | | 0.000 | 07:23 |
| Air Blank | | 0.000 | 07:25 |
| Subject Sample | # 2 | 0.000 | 07:25 |
| Breath Volume | | 1.710 | 07:25 |
| Air Blank | | 0.000 | 07:26 |
| Control Test | | 0.081 | 07:26 |
| Air Blank | | 0.000 | 07:27 |
| Diagnostics Check | | OK | 07:27 |

Data Download Date/Time: 05/22/2026 14:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASS,MALIK,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HASANBASIC,AMAR
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 05/03/2026 01:25:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:18 |
| Air Blank | | 0.000 | 01:19 |
| Control Test | | 0.081 | 01:19 |
| Air Blank | | 0.000 | 01:20 |
| Subject Sample | # 1 | 0.075 | 01:21 |
| Breath Volume | | 3.042 | 01:21 |
| Air Blank | | 0.000 | 01:22 |
| Air Blank | | 0.000 | 01:23 |
| Subject Sample | # 2 | 0.073 | 01:24 |
| Breath Volume | | 3.222 | 01:24 |
| Air Blank | | 0.000 | 01:24 |
| Control Test | | 0.079 | 01:25 |
| Air Blank | | 0.000 | 01:25 |
| Diagnostics Check | | OK | 01:25 |

Data Download Date/Time: 05/22/2026 14:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASS,MALIK,
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: HASS,MALIK
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: SHRIMPLIN

DATE/TIME OF TEST: 05/03/2026 22:55:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:47 |
| Air Blank | | 0.000 | 22:48 |
| Control Test | | 0.081 | 22:48 |
| Air Blank | | 0.000 | 22:48 |
| Subject Sample | # 1 | 0.104 | 22:50 |
| Breath Volume | | 2.195 | 22:50 |
| Air Blank | | 0.000 | 22:51 |
| Air Blank | | 0.000 | 22:53 |
| Subject Sample | # 2 | 0.105 | 22:53 |
| Breath Volume | | 1.859 | 22:53 |
| Air Blank | | 0.000 | 22:54 |
| Control Test | | 0.079 | 22:54 |
| Air Blank | | 0.000 | 22:55 |
| Diagnostics Check | | OK | 22:55 |

Data Download Date/Time: 05/22/2026 14:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRINGMAN,JASON,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: LAHURD

DATE/TIME OF TEST: 05/08/2026 18:01:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 17:55 |
| Air Blank | | 0.000 | 17:56 |
| Control Test | | 0.080 | 17:56 |
| Air Blank | | 0.000 | 17:56 |
| Subject Sample | # 1 | 0.000 | 17:57 |
| Breath Volume | | 1.839 | 17:57 |
| Air Blank | | 0.000 | 17:57 |
| Air Blank | | 0.000 | 17:59 |
| Subject Sample | # 2 | 0.000 | 18:00 |
| Breath Volume | | 1.652 | 18:00 |
| Air Blank | | 0.000 | 18:00 |
| Control Test | | 0.080 | 18:01 |
| Air Blank | | 0.000 | 18:01 |
| Diagnostics Check | | OK | 18:01 |

Data Download Date/Time: 05/22/2026 14:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: OLIVAS TELLEZ

DATE/TIME OF TEST: 05/09/2026 03:14:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:07 |
| Air Blank | | 0.000 | 03:08 |
| Control Test | | 0.077 | 03:08 |
| Air Blank | | 0.000 | 03:08 |
| Subject Sample | # 1 | 0.175 | 03:09 |
| Breath Volume | | 1.265 | 03:09 |
| Air Blank | | 0.000 | 03:10 |
| Air Blank | | 0.000 | 03:12 |
| Subject Sample | # 2 | 0.171 | 03:12 |
| Breath Volume | | 1.800 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Control Test | | 0.078 | 03:13 |
| Air Blank | | 0.000 | 03:14 |
| Diagnostics Check | | OK | 03:14 |

Data Download Date/Time: 05/22/2026 14:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRINGMAN,JASON,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: US US

DATE/TIME OF TEST: 05/10/2026 22:54:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:47 |
| Air Blank | | 0.000 | 22:48 |
| Control Test | | 0.081 | 22:48 |
| Air Blank | | 0.000 | 22:49 |
| Subject Sample | # 1 | 0.120 | 22:49 |
| Breath Volume | | 3.117 | 22:49 |
| Air Blank | | 0.000 | 22:50 |
| Air Blank | | 0.000 | 22:52 |
| Subject Sample | # 2 | 0.115 | 22:53 |
| Breath Volume | | 2.859 | 22:53 |
| Air Blank | | 0.000 | 22:53 |
| Control Test | | 0.081 | 22:54 |
| Air Blank | | 0.000 | 22:54 |
| Diagnostics Check | | OK | 22:54 |

Data Download Date/Time: 05/22/2026 14:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: JOHNS,JUSTIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 05/13/2026 01:39:42

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:36 | |
| Air Blank | | 0.000 | 01:37 | |
| Control Test | | 0.081 | 01:37 | |
| Air Blank | | 0.000 | 01:38 | |
| Subject Sample | # 1 | | 01:38 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:38 | |
| Air Blank | | 0.000 | 01:38 | |
| Control Test | | 0.082 | 01:39 | |
| Air Blank | | 0.000 | 01:39 | |
| Diagnostics Check | | OK | 01:39 | |

Data Download Date/Time: 05/22/2026 14:22


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,WILFREDO,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: HERNANDEZ,WILFREDO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 10:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:21

SUBJECT'S LAST NAME: KURUTZ

DATE/TIME OF TEST: 05/13/2026 11:48:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 11:42 |
| Air Blank | | 0.000 | 11:42 |
| Control Test | | 0.081 | 11:42 |
| Air Blank | | 0.000 | 11:43 |
| Subject Sample | # 1 | 0.000 | 11:43 |
| Breath Volume | | 2.257 | 11:43 |
| Air Blank | | 0.000 | 11:44 |
| Air Blank | | 0.000 | 11:46 |
| Subject Sample | # 2 | 0.000 | 11:47 |
| Breath Volume | | 2.546 | 11:47 |
| Air Blank | | 0.000 | 11:47 |
| Control Test | | 0.080 | 11:47 |
| Air Blank | | 0.000 | 11:48 |
| Diagnostics Check | | OK | 11:48 |

Data Download Date/Time: 05/22/2026 14:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: JOHNS,JUSTIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MORENO OSORIO

DATE/TIME OF TEST: 05/20/2026 21:42:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:35 |
| Air Blank | | 0.000 | 21:35 |
| Control Test | | 0.082 | 21:36 |
| Air Blank | | 0.000 | 21:36 |
| Subject Sample | # 1 | 0.189 | 21:38 |
| Breath Volume | | 2.320 | 21:38 |
| Air Blank | | 0.000 | 21:38 |
| Air Blank | | 0.000 | 21:40 |
| Subject Sample | # 2 | 0.189 | 21:41 |
| Breath Volume | | 2.300 | 21:41 |
| Air Blank | | 0.000 | 21:41 |
| Control Test | | 0.078 | 21:42 |
| Air Blank | | 0.000 | 21:42 |
| Diagnostics Check | | OK | 21:42 |

Data Download Date/Time: 05/22/2026 14:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: HASANBASIC,AMAR
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: BEARD

DATE/TIME OF TEST: 05/01/2026 19:41:27

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:35 |
| Air Blank | | 0.000 | 19:35 |
| Control Test | | 0.080 | 19:36 |
| Air Blank | | 0.000 | 19:36 |
| Subject Sample | # 1 | 0.000 | 19:37 |
| Breath Volume | | 3.636 | 19:37 |
| Air Blank | | 0.000 | 19:37 |
| Air Blank | | 0.000 | 19:39 |
| Subject Sample | # 2 | 0.000 | 19:40 |
| Breath Volume | | 3.957 | 19:40 |
| Air Blank | | 0.000 | 19:40 |
| Control Test | | 0.078 | 19:40 |
| Air Blank | | 0.000 | 19:41 |
| Diagnostics Check | | OK | 19:41 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: REYES-PEREZ

DATE/TIME OF TEST: 05/03/2026 01:27:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:21 |
| Air Blank | | 0.000 | 01:21 |
| Control Test | | 0.080 | 01:21 |
| Air Blank | | 0.000 | 01:22 |
| Subject Sample | # 1 | 0.138 | 01:22 |
| Breath Volume | | 2.230 | 01:22 |
| Air Blank | | 0.000 | 01:23 |
| Air Blank | | 0.000 | 01:25 |
| Subject Sample | # 2 | 0.138 | 01:25 |
| Breath Volume | | 2.695 | 01:25 |
| Air Blank | | 0.000 | 01:26 |
| Control Test | | 0.077 | 01:26 |
| Air Blank | | 0.000 | 01:27 |
| Diagnostics Check | | OK | 01:27 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: STACH

DATE/TIME OF TEST: 05/03/2026 02:36:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:30 |
| Air Blank | | 0.000 | 02:30 |
| Control Test | | 0.080 | 02:31 |
| Air Blank | | 0.000 | 02:31 |
| Subject Sample | # 1 | 0.198 | 02:32 |
| Breath Volume | | 1.792 | 02:32 |
| Air Blank | | 0.000 | 02:32 |
| Air Blank | | 0.000 | 02:34 |
| Subject Sample | # 2 | 0.200 | 02:35 |
| Breath Volume | | 1.746 | 02:35 |
| Air Blank | | 0.000 | 02:35 |
| Control Test | | 0.079 | 02:36 |
| Air Blank | | 0.000 | 02:36 |
| Diagnostics Check | | OK | 02:36 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVR,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WEAVER,SEAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 05/03/2026 03:02:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:55 |
| Air Blank | | 0.000 | 02:55 |
| Control Test | | 0.080 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Subject Sample | # 1 | 0.000 | 02:57 |
| Breath Volume | | 1.609 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Air Blank | | 0.000 | 03:00 |
| Subject Sample | # 2 | 0.000 | 03:00 |
| Breath Volume | | 2.003 | 03:00 |
| Air Blank | | 0.000 | 03:01 |
| Control Test | | 0.078 | 03:01 |
| Air Blank | | 0.000 | 03:02 |
| Diagnostics Check | | OK | 03:02 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLEMAN,WILLIAM
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 12:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: HABERLAND

DATE/TIME OF TEST: 05/06/2026 14:02:11

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 13:59 | |
| Air Blank | | 0.000 | 13:59 | |
| Control Test | | 0.080 | 13:59 | |
| Air Blank | | 0.000 | 14:00 | |
| Subject Sample | # 1 | | 14:00 | Subject Test Refused |
| Breath Volume | | 0.000 | 14:00 | |
| Air Blank | | 0.000 | 14:01 | |
| Control Test | | 0.081 | 14:01 | |
| Air Blank | | 0.000 | 14:02 | |
| Diagnostics Check | | OK | 14:02 | |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNS,JUSTIN,A
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: JOHNS,JUSTIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ZIRKELBACH

DATE/TIME OF TEST: 05/08/2026 00:51:59

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:48 | |
| Air Blank | | 0.000 | 00:49 | |
| Control Test | | 0.079 | 00:49 | |
| Air Blank | | 0.000 | 00:50 | |
| Subject Sample | # 1 | | 00:50 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:50 | |
| Air Blank | | 0.000 | 00:51 | |
| Control Test | | 0.080 | 00:51 | |
| Air Blank | | 0.000 | 00:51 | |
| Diagnostics Check | | OK | 00:51 | |

Data Download Date/Time: 05/22/2026 14:19


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLE,RODNEY,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EARLE,RODNEY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: LOY

DATE/TIME OF TEST: 05/11/2026 01:57:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Control Test | | 0.080 | 01:51 |
| Air Blank | | 0.000 | 01:52 |
| Subject Sample | # 1 | 0.053 | 01:52 |
| Breath Volume | | 1.449 | 01:52 |
| Air Blank | | 0.000 | 01:53 |
| Air Blank | | 0.000 | 01:55 |
| Subject Sample | # 2 | 0.057 | 01:56 |
| Breath Volume | | 2.742 | 01:56 |
| Air Blank | | 0.000 | 01:56 |
| Control Test | | 0.079 | 01:57 |
| Air Blank | | 0.000 | 01:57 |
| Diagnostics Check | | OK | 01:57 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREYRA JR,ROBERT,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: HOUGHTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: HARPE

DATE/TIME OF TEST: 05/11/2026 22:02:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:55 |
| Air Blank | | 0.000 | 21:55 |
| Control Test | | 0.080 | 21:55 |
| Air Blank | | 0.000 | 21:56 |
| Subject Sample | # 1 | 0.216 | 21:57 |
| Breath Volume | | 3.039 | 21:57 |
| Air Blank | | 0.000 | 21:57 |
| Air Blank | | 0.000 | 21:59 |
| Subject Sample | # 2 | 0.196 | 22:00 |
| Breath Volume | | 1.257 | 22:00 |
| Air Blank | | 0.000 | 22:01 |
| Control Test | | 0.077 | 22:01 |
| Air Blank | | 0.000 | 22:02 |
| Diagnostics Check | | OK | 22:02 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BERLINGERI

DATE/TIME OF TEST: 05/13/2026 00:06:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:58 |
| Air Blank | | 0.000 | 23:59 |
| Control Test | | 0.079 | 23:59 |
| Air Blank | | 0.000 | 00:00 |
| Subject Sample | # 1 | 0.148 | 00:01 |
| Breath Volume | | 2.546 | 00:01 |
| Air Blank | | 0.000 | 00:01 |
| Air Blank | | 0.000 | 00:03 |
| Subject Sample | # 2 | 0.145 | 00:04 |
| Breath Volume | | 3.039 | 00:04 |
| Air Blank | | 0.000 | 00:05 |
| Control Test | | 0.078 | 00:05 |
| Air Blank | | 0.000 | 00:06 |
| Diagnostics Check | | OK | 00:06 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WEAVER,SEAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RIVERA DE LA CRUZ

DATE/TIME OF TEST: 05/16/2026 02:19:07

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:14 | |
| Air Blank | | 0.000 | 02:15 | |
| Control Test | | 0.080 | 02:15 | |
| Air Blank | | 0.000 | 02:16 | |
| Subject Sample | # 1 | 0.204 | 02:17 | |
| Breath Volume | | 1.925 | 02:17 | |
| Air Blank | | | 02:18 | Purge Fail |
| Air Blank | | 0.000 | 02:19 | |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WEAVER,SEAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RIVERA DE LA CRUZ

DATE/TIME OF TEST: 05/16/2026 02:29:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:22 |
| Air Blank | | 0.000 | 02:22 |
| Control Test | | 0.079 | 02:23 |
| Air Blank | | 0.000 | 02:23 |
| Subject Sample | # 1 | 0.212 | 02:24 |
| Breath Volume | | 1.722 | 02:24 |
| Air Blank | | 0.000 | 02:25 |
| Air Blank | | 0.000 | 02:27 |
| Subject Sample | # 2 | 0.200 | 02:27 |
| Breath Volume | | 1.835 | 02:27 |
| Air Blank | | 0.000 | 02:28 |
| Control Test | | 0.079 | 02:28 |
| Air Blank | | 0.000 | 02:29 |
| Diagnostics Check | | OK | 02:29 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GAUNT

DATE/TIME OF TEST: 05/16/2026 02:59:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Control Test | | 0.079 | 02:53 |
| Air Blank | | 0.000 | 02:54 |
| Subject Sample | # 1 | 0.223 | 02:54 |
| Breath Volume | | 2.703 | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Air Blank | | 0.000 | 02:57 |
| Subject Sample | # 2 | 0.219 | 02:57 |
| Breath Volume | | 2.214 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Control Test | | 0.079 | 02:58 |
| Air Blank | | 0.000 | 02:59 |
| Diagnostics Check | | OK | 02:59 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: VAN DER MEER

DATE/TIME OF TEST: 05/16/2026 23:55:27

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:49 |
| Air Blank | | 0.000 | 23:49 |
| Control Test | | 0.080 | 23:50 |
| Air Blank | | 0.000 | 23:50 |
| Subject Sample | # 1 | 0.133 | 23:50 |
| Breath Volume | | 1.617 | 23:50 |
| Air Blank | | 0.000 | 23:51 |
| Air Blank | | 0.000 | 23:53 |
| Subject Sample | # 2 | 0.138 | 23:53 |
| Breath Volume | | 1.839 | 23:53 |
| Air Blank | | 0.000 | 23:54 |
| Control Test | | 0.078 | 23:54 |
| Air Blank | | 0.000 | 23:55 |
| Diagnostics Check | | OK | 23:55 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: HAROLD

DATE/TIME OF TEST: 05/17/2026 05:48:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:41 |
| Air Blank | | 0.000 | 05:42 |
| Control Test | | 0.080 | 05:42 |
| Air Blank | | 0.000 | 05:43 |
| Subject Sample | # 1 | 0.085 | 05:43 |
| Breath Volume | | 2.660 | 05:43 |
| Air Blank | | 0.000 | 05:44 |
| Air Blank | | 0.000 | 05:46 |
| Subject Sample | # 2 | 0.073 | 05:47 |
| Breath Volume | | 1.136 | 05:47 |
| Air Blank | | 0.000 | 05:47 |
| Control Test | | 0.078 | 05:48 |
| Air Blank | | 0.000 | 05:48 |
| Diagnostics Check | | OK | 05:48 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,CODY,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 09:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: BLANCHARD

DATE/TIME OF TEST: 05/17/2026 10:34:24

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 10:25 | |
| Air Blank | | 0.000 | 10:26 | |
| Control Test | | 0.081 | 10:26 | |
| Air Blank | | 0.000 | 10:27 | |
| Subject Sample | # 1 | | 10:30 | No Sample Provided |
| Breath Volume | | 0.000 | 10:30 | |
| Air Blank | | 0.000 | 10:30 | |
| Air Blank | | 0.000 | 10:32 | |
| Subject Sample | # 2 | | 10:33 | Subject Test Refused |
| Breath Volume | | 0.000 | 10:33 | |
| Air Blank | | 0.000 | 10:33 | |
| Control Test | | 0.081 | 10:33 | |
| Air Blank | | 0.000 | 10:34 | |
| Diagnostics Check | | OK | 10:34 | |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: MENNA

DATE/TIME OF TEST: 05/17/2026 18:10:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:04 |
| Air Blank | | 0.000 | 18:04 |
| Control Test | | 0.080 | 18:05 |
| Air Blank | | 0.000 | 18:05 |
| Subject Sample | # 1 | 0.000 | 18:06 |
| Breath Volume | | 1.136 | 18:06 |
| Air Blank | | 0.000 | 18:06 |
| Air Blank | | 0.000 | 18:08 |
| Subject Sample | # 2 | 0.000 | 18:09 |
| Breath Volume | | 1.101 | 18:09 |
| Air Blank | | 0.000 | 18:09 |
| Control Test | | 0.080 | 18:09 |
| Air Blank | | 0.000 | 18:10 |
| Diagnostics Check | | OK | 18:10 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WEAVER,SEAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SARABIA

DATE/TIME OF TEST: 05/17/2026 23:16:28

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:04 | |
| Air Blank | | 0.000 | 23:04 | |
| Control Test | | 0.080 | 23:05 | |
| Air Blank | | 0.000 | 23:05 | |
| Subject Sample | # 1 | 0.165 | 23:08 | Volume Not Met |
| Breath Volume | | 0.605 | 23:08 | |
| Air Blank | | 0.000 | 23:09 | |
| Air Blank | | 0.000 | 23:11 | |
| Subject Sample | # 2 | 0.180 | 23:14 | Volume Not Met |
| Breath Volume | | 0.140 | 23:14 | |
| Air Blank | | 0.000 | 23:15 | |
| Control Test | | 0.078 | 23:15 | |
| Air Blank | | 0.000 | 23:16 | |
| Diagnostics Check | | OK | 23:16 | |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WEAVER,SEAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SARABIA

DATE/TIME OF TEST: 05/17/2026 23:26:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:18 |
| Air Blank | | 0.000 | 23:19 |
| Control Test | | 0.078 | 23:19 |
| Air Blank | | 0.000 | 23:20 |
| Subject Sample | # 1 | 0.181 | 23:21 |
| Breath Volume | | 2.640 | 23:21 |
| Air Blank | | 0.000 | 23:22 |
| Air Blank | | 0.000 | 23:23 |
| Subject Sample | # 2 | 0.177 | 23:24 |
| Breath Volume | | 2.312 | 23:24 |
| Air Blank | | 0.000 | 23:25 |
| Control Test | | 0.078 | 23:25 |
| Air Blank | | 0.000 | 23:25 |
| Diagnostics Check | | OK | 23:26 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HERNANDEZ-JIMENEZ

DATE/TIME OF TEST: 05/18/2026 00:32:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:25 |
| Air Blank | | 0.000 | 00:25 |
| Control Test | | 0.080 | 00:26 |
| Air Blank | | 0.000 | 00:26 |
| Subject Sample | # 1 | 0.155 | 00:27 |
| Breath Volume | | 1.535 | 00:27 |
| Air Blank | | 0.000 | 00:28 |
| Air Blank | | 0.000 | 00:29 |
| Subject Sample | # 2 | 0.152 | 00:30 |
| Breath Volume | | 1.535 | 00:30 |
| Air Blank | | 0.000 | 00:31 |
| Control Test | | 0.078 | 00:31 |
| Air Blank | | 0.000 | 00:31 |
| Diagnostics Check | | OK | 00:32 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GALLINA,JOSEPH
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LALOR

DATE/TIME OF TEST: 05/18/2026 00:59:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:52 |
| Air Blank | | 0.000 | 00:53 |
| Control Test | | 0.080 | 00:53 |
| Air Blank | | 0.000 | 00:54 |
| Subject Sample | # 1 | 0.106 | 00:54 |
| Breath Volume | | 1.402 | 00:54 |
| Air Blank | | 0.000 | 00:55 |
| Air Blank | | 0.000 | 00:56 |
| Subject Sample | # 2 | 0.108 | 00:57 |
| Breath Volume | | 1.304 | 00:57 |
| Air Blank | | 0.000 | 00:58 |
| Control Test | | 0.079 | 00:58 |
| Air Blank | | 0.000 | 00:59 |
| Diagnostics Check | | OK | 00:59 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTE,SHANE,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COTE,SHANE
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 11:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: POLOUS

DATE/TIME OF TEST: 05/21/2026 14:41:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 14:34 |
| Air Blank | | 0.000 | 14:34 |
| Control Test | | 0.080 | 14:35 |
| Air Blank | | 0.000 | 14:35 |
| Subject Sample | # 1 | 0.000 | 14:36 |
| Breath Volume | | 1.859 | 14:36 |
| Air Blank | | 0.000 | 14:37 |
| Air Blank | | 0.000 | 14:39 |
| Subject Sample | # 2 | 0.000 | 14:40 |
| Breath Volume | | 1.218 | 14:40 |
| Air Blank | | 0.000 | 14:40 |
| Control Test | | 0.080 | 14:40 |
| Air Blank | | 0.000 | 14:41 |
| Diagnostics Check | | OK | 14:41 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: KNUCK

DATE/TIME OF TEST: 05/21/2026 23:48:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:41 |
| Air Blank | | 0.000 | 23:41 |
| Control Test | | 0.080 | 23:42 |
| Air Blank | | 0.000 | 23:42 |
| Subject Sample | # 1 | 0.243 | 23:43 |
| Breath Volume | | 1.515 | 23:43 |
| Air Blank | | 0.000 | 23:44 |
| Air Blank | | 0.000 | 23:46 |
| Subject Sample | # 2 | 0.238 | 23:46 |
| Breath Volume | | 1.378 | 23:46 |
| Air Blank | | 0.000 | 23:47 |
| Control Test | | 0.078 | 23:47 |
| Air Blank | | 0.000 | 23:48 |
| Diagnostics Check | | OK | 23:48 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: VILLALOBOS SOLANO

DATE/TIME OF TEST: 05/22/2026 00:51:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:42 |
| Air Blank | | 0.000 | 00:43 |
| Control Test | | 0.079 | 00:43 |
| Air Blank | | 0.000 | 00:43 |
| Subject Sample | # 1 | 0.249 | 00:46 |
| Breath Volume | | 1.867 | 00:46 |
| Air Blank | | 0.000 | 00:47 |
| Air Blank | | 0.000 | 00:48 |
| Subject Sample | # 2 | 0.252 | 00:49 |
| Breath Volume | | 1.804 | 00:49 |
| Air Blank | | 0.000 | 00:50 |
| Control Test | | 0.078 | 00:50 |
| Air Blank | | 0.000 | 00:51 |
| Diagnostics Check | | OK | 00:51 |

Data Download Date/Time: 05/22/2026 14:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: JACQUES,A
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 04/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 01/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: KENNELLY

DATE/TIME OF TEST: 05/02/2026 19:52:37

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 19:52 | |
| Air Blank | | 19:52 | RFI Detect |
| Air Blank | 0.000 | 19:52 | |

Data Download Date/Time: 06/02/2026 15:09

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEYMOUR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: GARDNER,TREVOR
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A1
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 05/07/2026 19:50:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:43 |
| Air Blank | | 0.000 | 19:44 |
| Control Test | | 0.079 | 19:44 |
| Air Blank | | 0.000 | 19:44 |
| Subject Sample | # 1 | 0.177 | 19:45 |
| Breath Volume | | 3.117 | 19:45 |
| Air Blank | | 0.000 | 19:46 |
| Air Blank | | 0.000 | 19:47 |
| Subject Sample | # 2 | 0.176 | 19:48 |
| Breath Volume | | 3.734 | 19:48 |
| Air Blank | | 0.000 | 19:49 |
| Control Test | | 0.076 | 19:49 |
| Air Blank | | 0.000 | 19:50 |
| Diagnostics Check | | OK | 19:50 |

Data Download Date/Time: 05/24/2026 00:05

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEYMOUR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: GARDNER,TREVOR
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A1
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/31/2026 16:19:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:12 |
| Air Blank | | 0.000 | 16:12 |
| Control Test | | 0.079 | 16:12 |
| Air Blank | | 0.000 | 16:13 |
| Subject Sample | # 1 | 0.116 | 16:14 |
| Breath Volume | | 1.253 | 16:14 |
| Air Blank | | 0.000 | 16:14 |
| Air Blank | | 0.000 | 16:16 |
| Subject Sample | # 2 | 0.122 | 16:17 |
| Breath Volume | | 1.921 | 16:17 |
| Air Blank | | 0.000 | 16:18 |
| Control Test | | 0.076 | 16:18 |
| Air Blank | | 0.000 | 16:19 |
| Diagnostics Check | | OK | 16:19 |

Data Download Date/Time: 06/25/2026 12:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: DELGADO,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SABATINI

DATE/TIME OF TEST: 05/03/2026 03:47:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:40 |
| Air Blank | | 0.000 | 03:40 |
| Control Test | | 0.077 | 03:40 |
| Air Blank | | 0.000 | 03:41 |
| Subject Sample | # 1 | 0.173 | 03:42 |
| Breath Volume | | 1.558 | 03:42 |
| Air Blank | | 0.000 | 03:43 |
| Air Blank | | 0.000 | 03:45 |
| Subject Sample | # 2 | 0.167 | 03:45 |
| Breath Volume | | 1.484 | 03:45 |
| Air Blank | | 0.000 | 03:46 |
| Control Test | | 0.078 | 03:46 |
| Air Blank | | 0.000 | 03:47 |
| Diagnostics Check | | OK | 03:47 |

Data Download Date/Time: 05/21/2026 16:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 05/12/2026 00:15:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:09 |
| Air Blank | | 0.000 | 00:09 |
| Control Test | | 0.077 | 00:10 |
| Air Blank | | 0.000 | 00:10 |
| Subject Sample | # 1 | 0.233 | 00:11 |
| Breath Volume | | 2.054 | 00:11 |
| Air Blank | | 0.000 | 00:11 |
| Air Blank | | 0.000 | 00:13 |
| Subject Sample | # 2 | 0.235 | 00:14 |
| Breath Volume | | 2.031 | 00:14 |
| Air Blank | | 0.000 | 00:14 |
| Control Test | | 0.077 | 00:15 |
| Air Blank | | 0.000 | 00:15 |
| Diagnostics Check | | OK | 00:15 |

Data Download Date/Time: 05/21/2026 16:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ZEPEDA ALVARADO

DATE/TIME OF TEST: 05/12/2026 02:14:34

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:11 | |
| Air Blank | | 0.000 | 02:11 | |
| Control Test | | 0.077 | 02:11 | |
| Air Blank | | 0.000 | 02:12 | |
| Subject Sample | # 1 | 0.244 | 02:13 | |
| Breath Volume | | 2.636 | 02:13 | |
| Air Blank | | | 02:14 | Purge Fail |
| Air Blank | | 0.000 | 02:14 | |

Data Download Date/Time: 05/21/2026 16:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ZEPEDA ALVARADO

DATE/TIME OF TEST: 05/12/2026 02:23:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:16 |
| Air Blank | | 0.000 | 02:16 |
| Control Test | | 0.075 | 02:16 |
| Air Blank | | 0.000 | 02:17 |
| Subject Sample | # 1 | 0.236 | 02:17 |
| Breath Volume | | 1.750 | 02:17 |
| Air Blank | | 0.000 | 02:18 |
| Air Blank | | 0.000 | 02:20 |
| Subject Sample | # 2 | 0.235 | 02:21 |
| Breath Volume | | 1.980 | 02:21 |
| Air Blank | | 0.000 | 02:22 |
| Control Test | | 0.077 | 02:22 |
| Air Blank | | 0.000 | 02:23 |
| Diagnostics Check | | OK | 02:23 |

Data Download Date/Time: 05/21/2026 16:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANGER,MOOKIE,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: DE NICOLO

DATE/TIME OF TEST: 05/19/2026 19:36:14

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:31 | |
| Air Blank | | 0.000 | 19:32 | |
| Control Test | | 0.077 | 19:32 | |
| Air Blank | | 0.000 | 19:33 | |
| Subject Sample | # 1 | | 19:34 | Subject Test Refused |
| Breath Volume | | 0.000 | 19:34 | |
| Air Blank | | 0.000 | 19:35 | |
| Control Test | | 0.078 | 19:35 | |
| Air Blank | | 0.000 | 19:36 | |
| Diagnostics Check | | OK | 19:36 | |

Data Download Date/Time: 05/21/2026 16:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CARABALLO,JOSE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: BADRU

DATE/TIME OF TEST: 05/20/2026 19:54:24

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:51 | |
| Air Blank | | 0.000 | 19:51 | |
| Control Test | | 0.076 | 19:52 | |
| Air Blank | | 0.000 | 19:52 | |
| Subject Sample | # 1 | | 19:53 | Subject Test Refused |
| Breath Volume | | 0.000 | 19:53 | |
| Air Blank | | 0.000 | 19:53 | |
| Control Test | | 0.077 | 19:53 | |
| Air Blank | | 0.000 | 19:54 | |
| Diagnostics Check | | OK | 19:54 | |

Data Download Date/Time: 05/21/2026 16:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: TAYLOR,ZACHARY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: OHANA

DATE/TIME OF TEST: 05/21/2026 00:43:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:35 |
| Air Blank | | 0.000 | 00:35 |
| Control Test | | 0.078 | 00:36 |
| Air Blank | | 0.000 | 00:36 |
| Subject Sample | # 1 | 0.105 | 00:37 |
| Breath Volume | | 1.714 | 00:37 |
| Air Blank | | 0.000 | 00:38 |
| Air Blank | | 0.000 | 00:40 |
| Subject Sample | # 2 | 0.111 | 00:42 |
| Breath Volume | | 2.476 | 00:42 |
| Air Blank | | 0.000 | 00:42 |
| Control Test | | 0.078 | 00:43 |
| Air Blank | | 0.000 | 00:43 |
| Diagnostics Check | | OK | 00:43 |

Data Download Date/Time: 05/21/2026 16:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,DYLAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: TAYLOR,ZACHARY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 05/23/2026 01:55:42

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:52 | |
| Air Blank | | 0.000 | 01:53 | |
| Control Test | | 0.076 | 01:53 | |
| Air Blank | | 0.000 | 01:54 | |
| Subject Sample | # 1 | | 01:54 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:54 | |
| Air Blank | | 0.000 | 01:54 | |
| Control Test | | 0.078 | 01:55 | |
| Air Blank | | 0.000 | 01:55 | |
| Diagnostics Check | | OK | 01:55 | |

Data Download Date/Time: 06/25/2026 14:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,REBECCA,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: WILLIAMS,DAVID
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:36

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/23/2026 14:05:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 13:56 |
| Air Blank | | 0.000 | 13:56 |
| Control Test | | 0.077 | 13:57 |
| Air Blank | | 0.000 | 13:57 |
| Subject Sample | # 1 | 0.000 | 14:00 |
| Breath Volume | | 2.472 | 14:00 |
| Air Blank | | 0.000 | 14:00 |
| Air Blank | | 0.000 | 14:02 |
| Subject Sample | # 2 | 0.000 | 14:03 |
| Breath Volume | | 1.351 | 14:03 |
| Air Blank | | 0.000 | 14:04 |
| Control Test | | 0.075 | 14:04 |
| Air Blank | | 0.000 | 14:05 |
| Diagnostics Check | | OK | 14:05 |

Data Download Date/Time: 06/25/2026 14:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: MEYER,TREVOR
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: THYBULLE

DATE/TIME OF TEST: 05/26/2026 02:31:08

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 02:30 | |
| Air Blank | | 02:30 | RFI Detect |
| Air Blank | 0.000 | 02:31 | |

Data Download Date/Time: 06/25/2026 14:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DAVID,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: MEYER,TREVOR
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: THYBULLE

DATE/TIME OF TEST: 05/26/2026 02:36:13

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:33 | |
| Air Blank | | 0.000 | 02:33 | |
| Control Test | | 0.078 | 02:34 | |
| Air Blank | | 0.000 | 02:34 | |
| Subject Sample | # 1 | | 02:34 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:34 | |
| Air Blank | | 0.000 | 02:35 | |
| Control Test | | 0.078 | 02:35 | |
| Air Blank | | 0.000 | 02:36 | |
| Diagnostics Check | | OK | 02:36 | |

Data Download Date/Time: 06/25/2026 14:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMUNNI,FRANCESCO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GOLDEN,LISA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: GOLDMAN

DATE/TIME OF TEST: 05/27/2026 12:07:03

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 11:58 |
| Air Blank | | 0.000 | 11:58 |
| Control Test | | 0.078 | 11:58 |
| Air Blank | | 0.000 | 11:59 |
| Subject Sample | # 1 | 0.000 | 12:01 |
| Breath Volume | | 2.253 | 12:01 |
| Air Blank | | 0.000 | 12:02 |
| Air Blank | | 0.000 | 12:04 |
| Subject Sample | # 2 | 0.000 | 12:05 |
| Breath Volume | | 2.273 | 12:05 |
| Air Blank | | 0.000 | 12:06 |
| Control Test | | 0.076 | 12:06 |
| Air Blank | | 0.000 | 12:06 |
| Diagnostics Check | | OK | 12:07 |

Data Download Date/Time: 06/25/2026 14:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALMEIDA,TIMOTHY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A2
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: DELGADILLO

DATE/TIME OF TEST: 05/30/2026 03:06:16

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:03 | |
| Air Blank | | 0.000 | 03:03 | |
| Control Test | | 0.076 | 03:04 | |
| Air Blank | | 0.000 | 03:04 | |
| Subject Sample | # 1 | | 03:04 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:04 | |
| Air Blank | | 0.000 | 03:05 | |
| Control Test | | 0.079 | 03:05 | |
| Air Blank | | 0.000 | 03:06 | |
| Diagnostics Check | | OK | 03:06 | |

Data Download Date/Time: 06/25/2026 14:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MADERAS MOREIRA

DATE/TIME OF TEST: 05/10/2026 03:13:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:06 |
| Air Blank | | 0.000 | 03:06 |
| Control Test | | 0.079 | 03:07 |
| Air Blank | | 0.000 | 03:07 |
| Subject Sample | # 1 | 0.175 | 03:08 |
| Breath Volume | | 1.527 | 03:08 |
| Air Blank | | 0.000 | 03:08 |
| Air Blank | | 0.000 | 03:10 |
| Subject Sample | # 2 | 0.176 | 03:11 |
| Breath Volume | | 1.449 | 03:11 |
| Air Blank | | 0.000 | 03:12 |
| Control Test | | 0.080 | 03:12 |
| Air Blank | | 0.000 | 03:12 |
| Diagnostics Check | | OK | 03:13 |

Data Download Date/Time: 05/22/2026 16:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: PAGE

DATE/TIME OF TEST: 05/22/2026 21:34:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:28 |
| Air Blank | | 0.000 | 21:28 |
| Control Test | | 0.081 | 21:28 |
| Air Blank | | 0.000 | 21:29 |
| Subject Sample | # 1 | 0.236 | 21:29 |
| Breath Volume | | 3.402 | 21:29 |
| Air Blank | | 0.000 | 21:30 |
| Air Blank | | 0.000 | 21:32 |
| Subject Sample | # 2 | 0.220 | 21:32 |
| Breath Volume | | 1.667 | 21:32 |
| Air Blank | | 0.000 | 21:33 |
| Control Test | | 0.080 | 21:33 |
| Air Blank | | 0.000 | 21:34 |
| Diagnostics Check | | OK | 21:34 |

Data Download Date/Time: 06/28/2026 17:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 05/24/2026 00:09:59

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:05 | |
| Air Blank | | 0.000 | 00:06 | |
| Control Test | | 0.081 | 00:06 | |
| Air Blank | | 0.000 | 00:07 | |
| Subject Sample | # 1 | | 00:08 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:08 | |
| Air Blank | | 0.000 | 00:09 | |
| Control Test | | 0.081 | 00:09 | |
| Air Blank | | 0.000 | 00:09 | |
| Diagnostics Check | | OK | 00:09 | |

Data Download Date/Time: 06/28/2026 17:26


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBO,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ALBO,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: NSHIMIYIMANA

DATE/TIME OF TEST: 05/24/2026 04:50:27

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:38 | |
| Air Blank | | 0.000 | 04:39 | |
| Control Test | | 0.081 | 04:39 | |
| Air Blank | | 0.000 | 04:40 | |
| Subject Sample | # 1 | 0.189 | 04:43 | Volume Not Met |
| Breath Volume | | 0.093 | 04:43 | |
| Air Blank | | 0.000 | 04:43 | |
| Air Blank | | 0.000 | 04:45 | |
| Subject Sample | # 2 | 0.180 | 04:48 | Volume Not Met |
| Breath Volume | | 0.039 | 04:48 | |
| Air Blank | | 0.000 | 04:49 | |
| Control Test | | 0.080 | 04:49 | |
| Air Blank | | 0.000 | 04:50 | |
| Diagnostics Check | | OK | 04:50 | |

Data Download Date/Time: 06/28/2026 17:26


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: CALERO

DATE/TIME OF TEST: 05/24/2026 22:20:53

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:14 | |
| Air Blank | | 0.000 | 22:14 | |
| Control Test | | 0.080 | 22:15 | |
| Air Blank | | 0.000 | 22:15 | |
| Subject Sample | # 1 | 0.214 | 22:16 | |
| Breath Volume | | 1.406 | 22:16 | |
| Air Blank | | 0.000 | 22:16 | |
| Air Blank | | 0.000 | 22:18 | |
| Subject Sample | # 2 | | 22:19 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:19 | |
| Air Blank | | 0.000 | 22:19 | |
| Control Test | | 0.079 | 22:20 | |
| Air Blank | | 0.000 | 22:20 | |
| Diagnostics Check | | OK | 22:20 | |

Data Download Date/Time: 06/28/2026 17:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ARIEL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HERNANDEZ,ARIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: FRICE

DATE/TIME OF TEST: 05/26/2026 17:54:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 17:44 |
| Air Blank | | 0.000 | 17:44 |
| Control Test | | 0.081 | 17:44 |
| Air Blank | | 0.000 | 17:45 |
| Subject Sample | # 1 | 0.199 | 17:48 |
| Breath Volume | | 1.550 | 17:48 |
| Air Blank | | 0.000 | 17:49 |
| Air Blank | | 0.000 | 17:50 |
| Subject Sample | # 2 | 0.191 | 17:52 |
| Breath Volume | | 1.550 | 17:52 |
| Air Blank | | 0.000 | 17:53 |
| Control Test | | 0.079 | 17:53 |
| Air Blank | | 0.000 | 17:53 |
| Diagnostics Check | | OK | 17:54 |

Data Download Date/Time: 06/28/2026 17:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDOVAL,LUIS,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CARRERA,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 05/27/2026 18:12:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:06 |
| Air Blank | | 0.000 | 18:06 |
| Control Test | | 0.081 | 18:06 |
| Air Blank | | 0.000 | 18:07 |
| Subject Sample | # 1 | 0.000 | 18:08 |
| Breath Volume | | 1.480 | 18:08 |
| Air Blank | | 0.000 | 18:09 |
| Air Blank | | 0.000 | 18:11 |
| Subject Sample | # 2 | 0.000 | 18:11 |
| Breath Volume | | 1.480 | 18:11 |
| Air Blank | | 0.000 | 18:12 |
| Control Test | | 0.079 | 18:12 |
| Air Blank | | 0.000 | 18:12 |
| Diagnostics Check | | OK | 18:12 |

Data Download Date/Time: 06/28/2026 17:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: CARTER GONZALEZ

DATE/TIME OF TEST: 05/30/2026 03:17:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:14 | |
| Air Blank | | 0.000 | 03:15 | |
| Control Test | | 0.082 | 03:15 | |
| Air Blank | | 0.000 | 03:15 | |
| Subject Sample | # 1 | | 03:16 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:16 | |
| Air Blank | | 0.000 | 03:17 | |
| Control Test | | 0.082 | 03:17 | |
| Air Blank | | 0.000 | 03:17 | |
| Diagnostics Check | | OK | 03:17 | |

Data Download Date/Time: 06/28/2026 17:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DENNIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: COLE,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 05/31/2026 00:28:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:26 | |
| Air Blank | | 0.000 | 00:26 | |
| Control Test | | 0.081 | 00:26 | |
| Air Blank | | 0.000 | 00:27 | |
| Subject Sample | # 1 | | 00:27 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:27 | |
| Air Blank | | 0.000 | 00:28 | |
| Control Test | | 0.081 | 00:28 | |
| Air Blank | | 0.000 | 00:28 | |
| Diagnostics Check | | OK | 00:28 | |

Data Download Date/Time: 06/28/2026 17:26


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 05/31/2026 02:04:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:58 |
| Air Blank | | 0.000 | 01:58 |
| Control Test | | 0.081 | 01:59 |
| Air Blank | | 0.000 | 01:59 |
| Subject Sample | # 1 | 0.046 | 02:00 |
| Breath Volume | | 1.785 | 02:00 |
| Air Blank | | 0.000 | 02:00 |
| Air Blank | | 0.000 | 02:02 |
| Subject Sample | # 2 | 0.053 | 02:03 |
| Breath Volume | | 1.964 | 02:03 |
| Air Blank | | 0.000 | 02:03 |
| Control Test | | 0.080 | 02:04 |
| Air Blank | | 0.000 | 02:04 |
| Diagnostics Check | | OK | 02:04 |

Data Download Date/Time: 06/28/2026 17:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: RUIZ HERNANDEZ

DATE/TIME OF TEST: 05/31/2026 23:59:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:52 |
| Air Blank | | 0.000 | 23:53 |
| Control Test | | 0.080 | 23:53 |
| Air Blank | | 0.000 | 23:53 |
| Subject Sample | # 1 | 0.168 | 23:54 |
| Breath Volume | | 2.140 | 23:54 |
| Air Blank | | 0.000 | 23:54 |
| Air Blank | | 0.000 | 23:56 |
| Subject Sample | # 2 | 0.161 | 23:57 |
| Breath Volume | | 2.878 | 23:57 |
| Air Blank | | 0.000 | 23:58 |
| Control Test | | 0.081 | 23:58 |
| Air Blank | | 0.000 | 23:59 |
| Diagnostics Check | | OK | 23:59 |

Data Download Date/Time: 06/28/2026 17:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRINGMAN,JASON,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 05/01/2026 00:27:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:20 |
| Air Blank | | 0.000 | 00:21 |
| Control Test | | 0.081 | 00:21 |
| Air Blank | | 0.000 | 00:22 |
| Subject Sample | # 1 | 0.087 | 00:22 |
| Breath Volume | | 1.589 | 00:22 |
| Air Blank | | 0.000 | 00:23 |
| Air Blank | | 0.000 | 00:25 |
| Subject Sample | # 2 | 0.089 | 00:25 |
| Breath Volume | | 1.640 | 00:25 |
| Air Blank | | 0.000 | 00:26 |
| Control Test | | 0.081 | 00:26 |
| Air Blank | | 0.000 | 00:26 |
| Diagnostics Check | | OK | 00:27 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WEAVER,SEAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 05/01/2026 16:55:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:48 |
| Air Blank | | 0.000 | 16:48 |
| Control Test | | 0.082 | 16:48 |
| Air Blank | | 0.000 | 16:49 |
| Subject Sample | # 1 | 0.188 | 16:50 |
| Breath Volume | | 2.816 | 16:50 |
| Air Blank | | 0.000 | 16:51 |
| Air Blank | | 0.000 | 16:52 |
| Subject Sample | # 2 | 0.191 | 16:54 |
| Breath Volume | | 2.542 | 16:54 |
| Air Blank | | 0.000 | 16:55 |
| Control Test | | 0.081 | 16:55 |
| Air Blank | | 0.000 | 16:55 |
| Diagnostics Check | | OK | 16:55 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: MOORE PAPPEL

DATE/TIME OF TEST: 05/01/2026 23:20:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:13 |
| Air Blank | | 0.000 | 23:14 |
| Control Test | | 0.081 | 23:14 |
| Air Blank | | 0.000 | 23:15 |
| Subject Sample | # 1 | 0.000 | 23:15 |
| Breath Volume | | 1.210 | 23:15 |
| Air Blank | | 0.000 | 23:16 |
| Air Blank | | 0.000 | 23:18 |
| Subject Sample | # 2 | 0.000 | 23:18 |
| Breath Volume | | 1.164 | 23:18 |
| Air Blank | | 0.000 | 23:19 |
| Control Test | | 0.082 | 23:19 |
| Air Blank | | 0.000 | 23:19 |
| Diagnostics Check | | OK | 23:20 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TEDESCO

DATE/TIME OF TEST: 05/02/2026 22:29:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:22 |
| Air Blank | | 0.000 | 22:23 |
| Control Test | | 0.081 | 22:23 |
| Air Blank | | 0.000 | 22:24 |
| Subject Sample | # 1 | 0.035 | 22:24 |
| Breath Volume | | 1.914 | 22:24 |
| Air Blank | | 0.000 | 22:25 |
| Air Blank | | 0.000 | 22:27 |
| Subject Sample | # 2 | 0.032 | 22:27 |
| Breath Volume | | 1.363 | 22:27 |
| Air Blank | | 0.000 | 22:28 |
| Control Test | | 0.078 | 22:28 |
| Air Blank | | 0.000 | 22:29 |
| Diagnostics Check | | OK | 22:29 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/03/2026 20:48:12

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:33 | |
| Air Blank | | 0.000 | 20:34 | |
| Control Test | | 0.082 | 20:34 | |
| Air Blank | | 0.000 | 20:34 | |
| Subject Sample | # 1 | 0.103 | 20:38 | Volume Not Met |
| Breath Volume | | 1.234 | 20:38 | |
| Air Blank | | 0.000 | 20:38 | |
| Air Blank | | 0.000 | 20:40 | |
| Subject Sample | # 2 | 0.121 | 20:42 | |
| Breath Volume | | 1.269 | 20:42 | |
| Air Blank | | 0.000 | 20:43 | |
| Air Blank | | 0.000 | 20:45 | |
| Subject Sample | # 3 | 0.102 | 20:46 | |
| Breath Volume | | 1.390 | 20:46 | |
| Air Blank | | 0.000 | 20:47 | |
| Control Test | | 0.081 | 20:47 | |
| Air Blank | | 0.000 | 20:48 | |
| Diagnostics Check | | OK | 20:48 | |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRINGMAN,JASON,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: SWINDELL

DATE/TIME OF TEST: 05/04/2026 19:55:27

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:47 |
| Air Blank | | 0.000 | 19:48 |
| Control Test | | 0.081 | 19:48 |
| Air Blank | | 0.000 | 19:49 |
| Subject Sample | # 1 | 0.214 | 19:50 |
| Breath Volume | | 2.183 | 19:50 |
| Air Blank | | 0.000 | 19:51 |
| Air Blank | | 0.000 | 19:53 |
| Subject Sample | # 2 | 0.200 | 19:53 |
| Breath Volume | | 1.214 | 19:53 |
| Air Blank | | 0.000 | 19:54 |
| Control Test | | 0.079 | 19:54 |
| Air Blank | | 0.000 | 19:55 |
| Diagnostics Check | | OK | 19:55 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRINGMAN,JASON,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BARRAGAN

DATE/TIME OF TEST: 05/06/2026 01:48:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:40 |
| Air Blank | | 0.000 | 01:40 |
| Control Test | | 0.082 | 01:41 |
| Air Blank | | 0.000 | 01:41 |
| Subject Sample | # 1 | 0.173 | 01:42 |
| Breath Volume | | 2.089 | 01:42 |
| Air Blank | | 0.000 | 01:43 |
| Air Blank | | 0.000 | 01:45 |
| Subject Sample | # 2 | 0.166 | 01:47 |
| Breath Volume | | 2.148 | 01:47 |
| Air Blank | | 0.000 | 01:47 |
| Control Test | | 0.081 | 01:48 |
| Air Blank | | 0.000 | 01:48 |
| Diagnostics Check | | OK | 01:48 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAER,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WEAVER,SEAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GALOSKY

DATE/TIME OF TEST: 05/08/2026 03:42:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:34 |
| Air Blank | | 0.000 | 03:34 |
| Control Test | | 0.082 | 03:35 |
| Air Blank | | 0.000 | 03:35 |
| Subject Sample | # 1 | 0.111 | 03:36 |
| Breath Volume | | 2.175 | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Air Blank | | 0.000 | 03:39 |
| Subject Sample | # 2 | 0.113 | 03:40 |
| Breath Volume | | 1.992 | 03:40 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.079 | 03:41 |
| Air Blank | | 0.000 | 03:42 |
| Diagnostics Check | | OK | 03:42 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GUEVARAMONTES

DATE/TIME OF TEST: 05/08/2026 03:53:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:45 |
| Air Blank | | 0.000 | 03:46 |
| Control Test | | 0.081 | 03:46 |
| Air Blank | | 0.000 | 03:47 |
| Subject Sample | # 1 | 0.104 | 03:48 |
| Breath Volume | | 1.335 | 03:48 |
| Air Blank | | 0.000 | 03:49 |
| Air Blank | | 0.000 | 03:51 |
| Subject Sample | # 2 | 0.089 | 03:52 |
| Breath Volume | | 1.523 | 03:52 |
| Air Blank | | 0.000 | 03:52 |
| Control Test | | 0.075 | 03:53 |
| Air Blank | | 0.000 | 03:53 |
| Diagnostics Check | | OK | 03:53 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRINGMAN,JASON,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: ESQUIVEL

DATE/TIME OF TEST: 05/09/2026 02:14:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Control Test | | 0.081 | 02:08 |
| Air Blank | | 0.000 | 02:08 |
| Subject Sample | # 1 | 0.176 | 02:09 |
| Breath Volume | | 2.207 | 02:09 |
| Air Blank | | 0.000 | 02:09 |
| Air Blank | | 0.000 | 02:11 |
| Subject Sample | # 2 | 0.186 | 02:12 |
| Breath Volume | | 2.132 | 02:12 |
| Air Blank | | 0.000 | 02:13 |
| Control Test | | 0.081 | 02:13 |
| Air Blank | | 0.000 | 02:14 |
| Diagnostics Check | | OK | 02:14 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRINGMAN,JASON,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: GEIGER

DATE/TIME OF TEST: 05/11/2026 01:53:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:47 |
| Air Blank | | 0.000 | 01:47 |
| Control Test | | 0.082 | 01:47 |
| Air Blank | | 0.000 | 01:48 |
| Subject Sample | # 1 | 0.041 | 01:48 |
| Breath Volume | | 2.292 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Air Blank | | 0.000 | 01:51 |
| Subject Sample | # 2 | 0.041 | 01:51 |
| Breath Volume | | 2.398 | 01:51 |
| Air Blank | | 0.000 | 01:52 |
| Control Test | | 0.082 | 01:52 |
| Air Blank | | 0.000 | 01:53 |
| Diagnostics Check | | OK | 01:53 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: DE MERCURIO

DATE/TIME OF TEST: 05/14/2026 16:44:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:35 |
| Air Blank | | 0.000 | 16:36 |
| Control Test | | 0.080 | 16:36 |
| Air Blank | | 0.000 | 16:37 |
| Subject Sample | # 1 | 0.279 | 16:38 |
| Breath Volume | | 1.488 | 16:38 |
| Air Blank | | 0.000 | 16:39 |
| Air Blank | | 0.000 | 16:40 |
| Subject Sample | # 2 | 0.276 | 16:42 |
| Breath Volume | | 1.457 | 16:42 |
| Air Blank | | 0.000 | 16:43 |
| Control Test | | 0.079 | 16:43 |
| Air Blank | | 0.000 | 16:44 |
| Diagnostics Check | | OK | 16:44 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,SEAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WEAVER,SEAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/17/2026 02:43:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:35 |
| Air Blank | | 0.000 | 02:35 |
| Control Test | | 0.081 | 02:36 |
| Air Blank | | 0.000 | 02:36 |
| Subject Sample | # 1 | 0.137 | 02:38 |
| Breath Volume | | 1.746 | 02:38 |
| Air Blank | | 0.000 | 02:38 |
| Air Blank | | 0.000 | 02:40 |
| Subject Sample | # 2 | 0.139 | 02:41 |
| Breath Volume | | 2.578 | 02:41 |
| Air Blank | | 0.000 | 02:42 |
| Control Test | | 0.080 | 02:42 |
| Air Blank | | 0.000 | 02:43 |
| Diagnostics Check | | OK | 02:43 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: HERNANDEZ SALES

DATE/TIME OF TEST: 05/22/2026 04:18:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:11 |
| Air Blank | | 0.000 | 04:11 |
| Control Test | | 0.080 | 04:11 |
| Air Blank | | 0.000 | 04:12 |
| Subject Sample | # 1 | 0.175 | 04:13 |
| Breath Volume | | 2.062 | 04:13 |
| Air Blank | | 0.000 | 04:14 |
| Air Blank | | 0.000 | 04:16 |
| Subject Sample | # 2 | 0.185 | 04:16 |
| Breath Volume | | 2.601 | 04:16 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.079 | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Diagnostics Check | | OK | 04:18 |

Data Download Date/Time: 05/22/2026 15:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,CORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MORGAN,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: LEO

DATE/TIME OF TEST: 05/02/2026 19:42:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:36 |
| Air Blank | | 0.000 | 19:36 |
| Control Test | | 0.080 | 19:36 |
| Air Blank | | 0.000 | 19:37 |
| Subject Sample | # 1 | 0.155 | 19:37 |
| Breath Volume | | 2.449 | 19:37 |
| Air Blank | | 0.000 | 19:38 |
| Air Blank | | 0.000 | 19:40 |
| Subject Sample | # 2 | 0.157 | 19:40 |
| Breath Volume | | 2.546 | 19:40 |
| Air Blank | | 0.000 | 19:41 |
| Control Test | | 0.078 | 19:41 |
| Air Blank | | 0.000 | 19:42 |
| Diagnostics Check | | OK | 19:42 |

Data Download Date/Time: 05/06/2026 20:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,CORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SAPP,CORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: BARLEY

DATE/TIME OF TEST: 05/03/2026 03:06:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:00 |
| Air Blank | | 0.000 | 03:00 |
| Control Test | | 0.080 | 03:01 |
| Air Blank | | 0.000 | 03:01 |
| Subject Sample | # 1 | 0.199 | 03:02 |
| Breath Volume | | 1.332 | 03:02 |
| Air Blank | | 0.000 | 03:02 |
| Air Blank | | 0.000 | 03:04 |
| Subject Sample | # 2 | 0.181 | 03:05 |
| Breath Volume | | 1.250 | 03:05 |
| Air Blank | | 0.000 | 03:05 |
| Control Test | | 0.078 | 03:06 |
| Air Blank | | 0.000 | 03:06 |
| Diagnostics Check | | OK | 03:06 |

Data Download Date/Time: 05/06/2026 20:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VERSTEEG

DATE/TIME OF TEST: 05/08/2026 01:39:45

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:36 | |
| Air Blank | | 0.000 | 01:36 | |
| Control Test | | 0.080 | 01:36 | |
| Air Blank | | 0.000 | 01:37 | |
| Subject Sample | # 1 | | 01:38 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:38 | |
| Air Blank | | 0.000 | 01:38 | |
| Control Test | | 0.081 | 01:39 | |
| Air Blank | | 0.000 | 01:39 | |
| Diagnostics Check | | OK | 01:39 | |

Data Download Date/Time: 05/14/2026 09:11

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: ADKINS,ANDREW
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/09/2026 18:55:00

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 18:42 | |
| Air Blank | | 0.000 | 18:42 | |
| Control Test | | 0.080 | 18:43 | |
| Air Blank | | 0.000 | 18:43 | |
| Subject Sample | # 1 | 0.111 | 18:46 | Volume Not Met |
| Breath Volume | | 0.253 | 18:46 | |
| Air Blank | | 0.000 | 18:47 | |
| Air Blank | | 0.000 | 18:49 | |
| Subject Sample | # 2 | 0.136 | 18:50 | |
| Breath Volume | | 1.277 | 18:50 | |
| Air Blank | | 0.000 | 18:50 | |
| Air Blank | | 0.000 | 18:52 | |
| Subject Sample | # 3 | 0.141 | 18:53 | |
| Breath Volume | | 1.562 | 18:53 | |
| Air Blank | | 0.000 | 18:54 | |
| Control Test | | 0.079 | 18:54 | |
| Air Blank | | 0.000 | 18:54 | |
| Diagnostics Check | | OK | 18:55 | |

Data Download Date/Time: 05/14/2026 09:11

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADKINS,ANDREW,R
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: GAGLIANO,RYAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: CLARKE RING

DATE/TIME OF TEST: 05/09/2026 23:33:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:26 |
| Air Blank | | 0.000 | 23:26 |
| Control Test | | 0.080 | 23:26 |
| Air Blank | | 0.000 | 23:27 |
| Subject Sample | # 1 | 0.177 | 23:29 |
| Breath Volume | | 1.875 | 23:29 |
| Air Blank | | 0.000 | 23:29 |
| Air Blank | | 0.000 | 23:31 |
| Subject Sample | # 2 | 0.175 | 23:32 |
| Breath Volume | | 1.472 | 23:32 |
| Air Blank | | 0.000 | 23:32 |
| Control Test | | 0.077 | 23:33 |
| Air Blank | | 0.000 | 23:33 |
| Diagnostics Check | | OK | 23:33 |

Data Download Date/Time: 05/14/2026 09:12

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA-MEJIA,JAVIER,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ZUNIGA-MEJIA,JAVIER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 05/23/2026 04:55:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:48 |
| Air Blank | | 0.000 | 04:48 |
| Control Test | | 0.079 | 04:49 |
| Air Blank | | 0.000 | 04:49 |
| Subject Sample | # 1 | 0.175 | 04:51 |
| Breath Volume | | 1.351 | 04:51 |
| Air Blank | | 0.000 | 04:52 |
| Air Blank | | 0.000 | 04:53 |
| Subject Sample | # 2 | 0.193 | 04:54 |
| Breath Volume | | 2.289 | 04:54 |
| Air Blank | | 0.000 | 04:55 |
| Control Test | | 0.077 | 04:55 |
| Air Blank | | 0.000 | 04:55 |
| Diagnostics Check | | OK | 04:55 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: RODRIGUEZ FLORES

DATE/TIME OF TEST: 05/24/2026 00:21:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:13 |
| Air Blank | | 0.000 | 00:14 |
| Control Test | | 0.078 | 00:14 |
| Air Blank | | 0.000 | 00:15 |
| Subject Sample | # 1 | 0.238 | 00:16 |
| Breath Volume | | 1.765 | 00:16 |
| Air Blank | | 0.000 | 00:17 |
| Air Blank | | 0.000 | 00:18 |
| Subject Sample | # 2 | 0.244 | 00:19 |
| Breath Volume | | 3.226 | 00:19 |
| Air Blank | | 0.000 | 00:20 |
| Control Test | | 0.078 | 00:20 |
| Air Blank | | 0.000 | 00:21 |
| Diagnostics Check | | OK | 00:21 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRINGMAN,JASON,L
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/24/2026 03:33:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:27 |
| Air Blank | | 0.000 | 03:28 |
| Control Test | | 0.078 | 03:28 |
| Air Blank | | 0.000 | 03:29 |
| Subject Sample | # 1 | 0.144 | 03:29 |
| Breath Volume | | 2.503 | 03:29 |
| Air Blank | | 0.000 | 03:30 |
| Air Blank | | 0.000 | 03:31 |
| Subject Sample | # 2 | 0.141 | 03:32 |
| Breath Volume | | 2.792 | 03:32 |
| Air Blank | | 0.000 | 03:33 |
| Control Test | | 0.077 | 03:33 |
| Air Blank | | 0.000 | 03:33 |
| Diagnostics Check | | OK | 03:33 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,SARAH,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BROWN,SARAH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 05/24/2026 17:47:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 17:41 |
| Air Blank | | 0.000 | 17:41 |
| Control Test | | 0.078 | 17:42 |
| Air Blank | | 0.000 | 17:42 |
| Subject Sample | # 1 | 0.171 | 17:43 |
| Breath Volume | | 1.894 | 17:43 |
| Air Blank | | 0.000 | 17:43 |
| Air Blank | | 0.000 | 17:45 |
| Subject Sample | # 2 | 0.157 | 17:46 |
| Breath Volume | | 1.339 | 17:46 |
| Air Blank | | 0.000 | 17:46 |
| Control Test | | 0.077 | 17:47 |
| Air Blank | | 0.000 | 17:47 |
| Diagnostics Check | | OK | 17:47 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: POWELL JOHNSON

DATE/TIME OF TEST: 05/24/2026 22:54:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:47 |
| Air Blank | | 0.000 | 22:47 |
| Control Test | | 0.079 | 22:47 |
| Air Blank | | 0.000 | 22:48 |
| Subject Sample | # 1 | 0.216 | 22:49 |
| Breath Volume | | 1.753 | 22:49 |
| Air Blank | | 0.000 | 22:49 |
| Air Blank | | 0.000 | 22:51 |
| Subject Sample | # 2 | 0.219 | 22:52 |
| Breath Volume | | 2.816 | 22:52 |
| Air Blank | | 0.000 | 22:53 |
| Control Test | | 0.077 | 22:53 |
| Air Blank | | 0.000 | 22:53 |
| Diagnostics Check | | OK | 22:54 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: SYED-FISHER

DATE/TIME OF TEST: 05/25/2026 03:34:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:27 |
| Air Blank | | 0.000 | 03:27 |
| Control Test | | 0.078 | 03:28 |
| Air Blank | | 0.000 | 03:28 |
| Subject Sample | # 1 | 0.032 | 03:29 |
| Breath Volume | | 3.382 | 03:29 |
| Air Blank | | 0.000 | 03:30 |
| Air Blank | | 0.000 | 03:31 |
| Subject Sample | # 2 | 0.031 | 03:32 |
| Breath Volume | | 2.511 | 03:32 |
| Air Blank | | 0.000 | 03:33 |
| Control Test | | 0.078 | 03:33 |
| Air Blank | | 0.000 | 03:34 |
| Diagnostics Check | | OK | 03:34 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ELIASSAINT

DATE/TIME OF TEST: 05/26/2026 01:33:25

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:24 | |
| Air Blank | | 0.000 | 01:25 | |
| Control Test | | 0.079 | 01:25 | |
| Air Blank | | 0.000 | 01:25 | |
| Subject Sample | # 1 | | 01:28 | No Sample Provided |
| Breath Volume | | 0.000 | 01:28 | |
| Air Blank | | 0.000 | 01:29 | |
| Air Blank | | 0.000 | 01:31 | |
| Subject Sample | # 2 | | 01:32 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:32 | |
| Air Blank | | 0.000 | 01:32 | |
| Control Test | | 0.079 | 01:32 | |
| Air Blank | | 0.000 | 01:33 | |
| Diagnostics Check | | OK | 01:33 | |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 05/26/2026 04:10:44

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:58 | |
| Air Blank | | 0.000 | 03:58 | |
| Control Test | | 0.079 | 03:59 | |
| Air Blank | | 0.000 | 03:59 | |
| Subject Sample | # 1 | 0.160 | 04:02 | Volume Not Met |
| Breath Volume | | 0.308 | 04:02 | |
| Air Blank | | 0.000 | 04:03 | |
| Air Blank | | 0.000 | 04:05 | |
| Subject Sample | # 2 | 0.170 | 04:06 | |
| Breath Volume | | 2.691 | 04:06 | |
| Air Blank | | 0.000 | 04:06 | |
| Air Blank | | 0.000 | 04:08 | |
| Subject Sample | # 3 | 0.169 | 04:09 | |
| Breath Volume | | 1.816 | 04:09 | |
| Air Blank | | 0.000 | 04:09 | |
| Control Test | | 0.076 | 04:10 | |
| Air Blank | | 0.000 | 04:10 | |
| Diagnostics Check | | OK | 04:10 | |

Data Download Date/Time: 06/19/2026 13:19


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COBB,JAMEICA,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: ECHEVERRIA

DATE/TIME OF TEST: 05/30/2026 03:27:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:21 |
| Air Blank | | 0.000 | 03:22 |
| Control Test | | 0.079 | 03:22 |
| Air Blank | | 0.000 | 03:22 |
| Subject Sample | # 1 | 0.214 | 03:23 |
| Breath Volume | | 1.578 | 03:23 |
| Air Blank | | 0.000 | 03:24 |
| Air Blank | | 0.000 | 03:25 |
| Subject Sample | # 2 | 0.200 | 03:26 |
| Breath Volume | | 1.300 | 03:26 |
| Air Blank | | 0.000 | 03:26 |
| Control Test | | 0.078 | 03:27 |
| Air Blank | | 0.000 | 03:27 |
| Diagnostics Check | | OK | 03:27 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUGHTON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: HOUGHTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/30/2026 04:12:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:05 |
| Air Blank | | 0.000 | 04:06 |
| Control Test | | 0.079 | 04:06 |
| Air Blank | | 0.000 | 04:07 |
| Subject Sample | # 1 | 0.099 | 04:07 |
| Breath Volume | | 1.957 | 04:07 |
| Air Blank | | 0.000 | 04:08 |
| Air Blank | | 0.000 | 04:10 |
| Subject Sample | # 2 | 0.092 | 04:11 |
| Breath Volume | | 1.878 | 04:11 |
| Air Blank | | 0.000 | 04:11 |
| Control Test | | 0.076 | 04:12 |
| Air Blank | | 0.000 | 04:12 |
| Diagnostics Check | | OK | 04:12 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMANDL,REAGHN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: SGAMBATO

DATE/TIME OF TEST: 05/31/2026 02:38:23

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:31 |
| Air Blank | | 0.000 | 02:32 |
| Control Test | | 0.078 | 02:32 |
| Air Blank | | 0.000 | 02:33 |
| Subject Sample | # 1 | 0.121 | 02:33 |
| Breath Volume | | 3.785 | 02:33 |
| Air Blank | | 0.000 | 02:34 |
| Air Blank | | 0.000 | 02:36 |
| Subject Sample | # 2 | 0.120 | 02:36 |
| Breath Volume | | 3.089 | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Control Test | | 0.078 | 02:37 |
| Air Blank | | 0.000 | 02:38 |
| Diagnostics Check | | OK | 02:38 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOGLE,ERIC,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FOGLE,ERIC
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 05/31/2026 19:57:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:51 |
| Air Blank | | 0.000 | 19:51 |
| Control Test | | 0.079 | 19:52 |
| Air Blank | | 0.000 | 19:52 |
| Subject Sample | # 1 | 0.212 | 19:53 |
| Breath Volume | | 4.390 | 19:53 |
| Air Blank | | 0.000 | 19:54 |
| Air Blank | | 0.000 | 19:55 |
| Subject Sample | # 2 | 0.194 | 19:56 |
| Breath Volume | | 2.808 | 19:56 |
| Air Blank | | 0.000 | 19:56 |
| Control Test | | 0.078 | 19:57 |
| Air Blank | | 0.000 | 19:57 |
| Diagnostics Check | | OK | 19:57 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOGLE,ERIC,T
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FOGLE,ERIC
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PELLETIER

DATE/TIME OF TEST: 05/31/2026 23:23:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:17 |
| Air Blank | | 0.000 | 23:17 |
| Control Test | | 0.079 | 23:17 |
| Air Blank | | 0.000 | 23:18 |
| Subject Sample | # 1 | 0.246 | 23:18 |
| Breath Volume | | 2.535 | 23:18 |
| Air Blank | | 0.000 | 23:19 |
| Air Blank | | 0.000 | 23:21 |
| Subject Sample | # 2 | 0.235 | 23:21 |
| Breath Volume | | 1.343 | 23:21 |
| Air Blank | | 0.000 | 23:22 |
| Control Test | | 0.077 | 23:22 |
| Air Blank | | 0.000 | 23:23 |
| Diagnostics Check | | OK | 23:23 |

Data Download Date/Time: 06/19/2026 13:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,CELESTE,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANCHEZ,CELESTE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ESQUIVEL

DATE/TIME OF TEST: 05/04/2026 03:23:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:10 |
| Air Blank | | 0.000 | 03:10 |
| Control Test | | 0.077 | 03:10 |
| Air Blank | | 0.000 | 03:11 |
| Subject Sample | # 1 | 0.165 | 03:12 |
| Breath Volume | | 1.492 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Air Blank | | 0.000 | 03:14 |
| Subject Sample | # 2 | 0.141 | 03:17 |
| Breath Volume | | 1.683 | 03:17 |
| Air Blank | | 0.000 | 03:18 |
| Air Blank | | 0.000 | 03:19 |
| Subject Sample | # 3 | 0.147 | 03:21 |
| Breath Volume | | 1.460 | 03:21 |
| Air Blank | | 0.000 | 03:22 |
| Control Test | | 0.075 | 03:22 |
| Air Blank | | 0.000 | 03:23 |
| Diagnostics Check | | OK | 03:23 |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,MICAH,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LITTLE,MICAH
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: TAWEL

DATE/TIME OF TEST: 05/07/2026 05:34:42

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:32 | |
| Air Blank | | 0.000 | 05:33 | |
| Control Test | | 0.082 | 05:33 | |
| Air Blank | | 0.000 | 05:34 | |
| Subject Sample | # 1 | | 05:34 | Sequence Aborted |
| Breath Volume | | 0.000 | 05:34 | |
| Air Blank | | 0.000 | 05:34 | |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,MICAH,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LITTLE,MICAH
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 05/10/2026 20:50:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:44 |
| Air Blank | | 0.000 | 20:44 |
| Control Test | | 0.081 | 20:44 |
| Air Blank | | 0.000 | 20:45 |
| Subject Sample | # 1 | 0.216 | 20:46 |
| Breath Volume | | 1.890 | 20:46 |
| Air Blank | | 0.000 | 20:46 |
| Air Blank | | 0.000 | 20:48 |
| Subject Sample | # 2 | 0.203 | 20:49 |
| Breath Volume | | 1.820 | 20:49 |
| Air Blank | | 0.000 | 20:49 |
| Control Test | | 0.079 | 20:50 |
| Air Blank | | 0.000 | 20:50 |
| Diagnostics Check | | OK | 20:50 |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,MICAH,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LITTLE,MICAH
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: VALDES CORDERO

DATE/TIME OF TEST: 05/15/2026 20:26:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:20 |
| Air Blank | | 0.000 | 20:20 |
| Control Test | | 0.082 | 20:21 |
| Air Blank | | 0.000 | 20:21 |
| Subject Sample | # 1 | 0.000 | 20:22 |
| Breath Volume | | 3.664 | 20:22 |
| Air Blank | | 0.000 | 20:22 |
| Air Blank | | 0.000 | 20:24 |
| Subject Sample | # 2 | 0.000 | 20:24 |
| Breath Volume | | 1.597 | 20:24 |
| Air Blank | | 0.000 | 20:25 |
| Control Test | | 0.080 | 20:25 |
| Air Blank | | 0.000 | 20:26 |
| Diagnostics Check | | OK | 20:26 |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/15/2026 21:57:20

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:53 | |
| Air Blank | | 0.000 | 21:54 | |
| Control Test | | 0.081 | 21:54 | |
| Air Blank | | 0.000 | 21:55 | |
| Subject Sample | # 1 | | 21:55 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:55 | |
| Air Blank | | 0.000 | 21:56 | |
| Control Test | | 0.082 | 21:56 | |
| Air Blank | | 0.000 | 21:57 | |
| Diagnostics Check | | OK | 21:57 | |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 05/16/2026 02:28:22

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:16 | |
| Air Blank | | 0.000 | 02:16 | |
| Control Test | | 0.082 | 02:17 | |
| Air Blank | | 0.000 | 02:17 | |
| Subject Sample | # 1 | 0.158 | 02:20 | Volume Not Met |
| Breath Volume | | 0.031 | 02:20 | |
| Air Blank | | 0.000 | 02:21 | |
| Air Blank | | 0.000 | 02:23 | |
| Subject Sample | # 2 | 0.154 | 02:26 | Volume Not Met |
| Breath Volume | | 0.121 | 02:26 | |
| Air Blank | | 0.000 | 02:27 | |
| Control Test | | 0.081 | 02:27 | |
| Air Blank | | 0.000 | 02:28 | |
| Diagnostics Check | | OK | 02:28 | |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 05/16/2026 02:37:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:30 |
| Air Blank | | 0.000 | 02:31 |
| Control Test | | 0.079 | 02:31 |
| Air Blank | | 0.000 | 02:31 |
| Subject Sample | # 1 | 0.147 | 02:32 |
| Breath Volume | | 2.078 | 02:32 |
| Air Blank | | 0.000 | 02:33 |
| Air Blank | | 0.000 | 02:35 |
| Subject Sample | # 2 | 0.136 | 02:35 |
| Breath Volume | | 2.324 | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Control Test | | 0.079 | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Diagnostics Check | | OK | 02:37 |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: ALVAREZDELPINO

DATE/TIME OF TEST: 05/16/2026 21:45:40

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:37 | |
| Air Blank | | 0.000 | 21:37 | |
| Control Test | | 0.082 | 21:38 | |
| Air Blank | | 0.000 | 21:38 | |
| Subject Sample | # 1 | | 21:41 | No Sample Provided |
| Breath Volume | | 0.000 | 21:41 | |
| Air Blank | | 0.000 | 21:42 | |
| Air Blank | | 0.000 | 21:44 | |
| Subject Sample | # 2 | | 21:44 | Subject Test Refused |
| Breath Volume | | 0.000 | 21:44 | |
| Air Blank | | 0.000 | 21:44 | |
| Control Test | | 0.081 | 21:45 | |
| Air Blank | | 0.000 | 21:45 | |
| Diagnostics Check | | OK | 21:45 | |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: CARDONNE

DATE/TIME OF TEST: 05/17/2026 04:56:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:50 |
| Air Blank | | 0.000 | 04:50 |
| Control Test | | 0.082 | 04:50 |
| Air Blank | | 0.000 | 04:51 |
| Subject Sample | # 1 | 0.196 | 04:51 |
| Breath Volume | | 3.554 | 04:51 |
| Air Blank | | 0.000 | 04:52 |
| Air Blank | | 0.000 | 04:54 |
| Subject Sample | # 2 | 0.197 | 04:54 |
| Breath Volume | | 4.023 | 04:54 |
| Air Blank | | 0.000 | 04:55 |
| Control Test | | 0.081 | 04:55 |
| Air Blank | | 0.000 | 04:56 |
| Diagnostics Check | | OK | 04:56 |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: DINKINS

DATE/TIME OF TEST: 05/17/2026 05:48:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:41 |
| Air Blank | | 0.000 | 05:42 |
| Control Test | | 0.081 | 05:42 |
| Air Blank | | 0.000 | 05:43 |
| Subject Sample | # 1 | 0.144 | 05:43 |
| Breath Volume | | 1.894 | 05:43 |
| Air Blank | | 0.000 | 05:44 |
| Air Blank | | 0.000 | 05:46 |
| Subject Sample | # 2 | 0.139 | 05:47 |
| Breath Volume | | 2.179 | 05:47 |
| Air Blank | | 0.000 | 05:47 |
| Control Test | | 0.081 | 05:48 |
| Air Blank | | 0.000 | 05:48 |
| Diagnostics Check | | OK | 05:48 |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICARDO,EDWARD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RICARDO,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/22/2026 01:58:51

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:55 | |
| Air Blank | | 0.000 | 01:56 | |
| Control Test | | 0.082 | 01:56 | |
| Air Blank | | 0.000 | 01:57 | |
| Subject Sample | # 1 | | 01:57 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:57 | |
| Air Blank | | 0.000 | 01:57 | |
| Control Test | | 0.081 | 01:58 | |
| Air Blank | | 0.000 | 01:58 | |
| Diagnostics Check | | OK | 01:58 | |

Data Download Date/Time: 05/22/2026 16:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/02/2026 04:00:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:53 |
| Air Blank | | 0.000 | 03:54 |
| Control Test | | 0.080 | 03:54 |
| Air Blank | | 0.000 | 03:54 |
| Subject Sample | # 1 | 0.178 | 03:55 |
| Breath Volume | | 2.734 | 03:55 |
| Air Blank | | 0.000 | 03:56 |
| Air Blank | | 0.000 | 03:57 |
| Subject Sample | # 2 | 0.175 | 03:58 |
| Breath Volume | | 2.726 | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Control Test | | 0.080 | 03:59 |
| Air Blank | | 0.000 | 04:00 |
| Diagnostics Check | | OK | 04:00 |

Data Download Date/Time: 05/05/2026 06:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: MOTEL

DATE/TIME OF TEST: 05/06/2026 01:25:26

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:19 | |
| Air Blank | | 0.000 | 01:20 | |
| Control Test | | 0.080 | 01:20 | |
| Air Blank | | 0.000 | 01:20 | |
| Subject Sample | # 1 | 0.204 | 01:21 | |
| Breath Volume | | 3.246 | 01:21 | |
| Air Blank | | 0.000 | 01:22 | |
| Air Blank | | 0.000 | 01:23 | |
| Subject Sample | # 2 | 0.197 | 01:24 | |
| Breath Volume | | 3.824 | 01:24 | |
| Air Blank | | | 01:24 | Purge Fail |
| Air Blank | | 0.000 | 01:25 | |

Data Download Date/Time: 05/08/2026 15:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: MOTEL

DATE/TIME OF TEST: 05/06/2026 01:34:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:27 |
| Air Blank | | 0.000 | 01:28 |
| Control Test | | 0.080 | 01:28 |
| Air Blank | | 0.000 | 01:29 |
| Subject Sample | # 1 | 0.210 | 01:29 |
| Breath Volume | | 2.902 | 01:29 |
| Air Blank | | 0.000 | 01:30 |
| Air Blank | | 0.000 | 01:32 |
| Subject Sample | # 2 | 0.197 | 01:32 |
| Breath Volume | | 4.242 | 01:32 |
| Air Blank | | 0.000 | 01:33 |
| Control Test | | 0.079 | 01:33 |
| Air Blank | | 0.000 | 01:34 |
| Diagnostics Check | | OK | 01:34 |

Data Download Date/Time: 05/08/2026 15:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: SYS CHECK
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/23/2026 20:19:29

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 20:18 | |
| Air Blank | 0.000 | 20:18 | |
| Control Test | 0.000 | 20:19 | Control Outside Tolerance |
| Air Blank | | 20:19 | Improper Sample |

Data Download Date/Time: 05/23/2026 22:15


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3024028434
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: SYS CHECK
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 05/23/2026 20:23:10

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:20 | |
| Air Blank | | 0.000 | 20:21 | |
| Control Test | | 0.077 | 20:21 | |
| Air Blank | | 0.000 | 20:22 | |
| Subject Sample | # 1 | 0.000 | 20:22 | |
| Breath Volume | | 1.554 | 20:22 | |
| Air Blank | | | 20:22 | Sequence Aborted |
| Air Blank | | 0.000 | 20:23 | |

Data Download Date/Time: 05/23/2026 22:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 05/02/2026 00:33:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:27 |
| Air Blank | | 0.000 | 00:27 |
| Control Test | | 0.080 | 00:28 |
| Air Blank | | 0.000 | 00:28 |
| Subject Sample | # 1 | 0.199 | 00:29 |
| Breath Volume | | 2.050 | 00:29 |
| Air Blank | | 0.000 | 00:29 |
| Air Blank | | 0.000 | 00:31 |
| Subject Sample | # 2 | 0.201 | 00:32 |
| Breath Volume | | 1.949 | 00:32 |
| Air Blank | | 0.000 | 00:32 |
| Control Test | | 0.078 | 00:33 |
| Air Blank | | 0.000 | 00:33 |
| Diagnostics Check | | OK | 00:33 |

Data Download Date/Time: 05/23/2026 22:10

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: GARY

DATE/TIME OF TEST: 05/02/2026 02:34:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:28 |
| Air Blank | | 0.000 | 02:28 |
| Control Test | | 0.078 | 02:29 |
| Air Blank | | 0.000 | 02:29 |
| Subject Sample | # 1 | 0.077 | 02:30 |
| Breath Volume | | 1.863 | 02:30 |
| Air Blank | | 0.000 | 02:30 |
| Air Blank | | 0.000 | 02:32 |
| Subject Sample | # 2 | 0.073 | 02:33 |
| Breath Volume | | 1.125 | 02:33 |
| Air Blank | | 0.000 | 02:33 |
| Control Test | | 0.079 | 02:34 |
| Air Blank | | 0.000 | 02:34 |
| Diagnostics Check | | OK | 02:34 |

Data Download Date/Time: 05/23/2026 22:10

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: CORRIGAN

DATE/TIME OF TEST: 05/08/2026 21:49:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:42 |
| Air Blank | | 0.000 | 21:43 |
| Control Test | | 0.080 | 21:43 |
| Air Blank | | 0.000 | 21:43 |
| Subject Sample | # 1 | 0.137 | 21:44 |
| Breath Volume | | 3.324 | 21:44 |
| Air Blank | | 0.000 | 21:45 |
| Air Blank | | 0.000 | 21:47 |
| Subject Sample | # 2 | 0.124 | 21:48 |
| Breath Volume | | 1.949 | 21:48 |
| Air Blank | | 0.000 | 21:48 |
| Control Test | | 0.078 | 21:49 |
| Air Blank | | 0.000 | 21:49 |
| Diagnostics Check | | OK | 21:49 |

Data Download Date/Time: 05/23/2026 22:10

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 05/10/2026 03:08:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:02 |
| Air Blank | | 0.000 | 03:02 |
| Control Test | | 0.079 | 03:02 |
| Air Blank | | 0.000 | 03:03 |
| Subject Sample | # 1 | 0.088 | 03:03 |
| Breath Volume | | 3.398 | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Air Blank | | 0.000 | 03:06 |
| Subject Sample | # 2 | 0.088 | 03:07 |
| Breath Volume | | 3.460 | 03:07 |
| Air Blank | | 0.000 | 03:07 |
| Control Test | | 0.079 | 03:08 |
| Air Blank | | 0.000 | 03:08 |
| Diagnostics Check | | OK | 03:08 |

Data Download Date/Time: 05/23/2026 22:10

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/20/2026 21:49:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:43 |
| Air Blank | | 0.000 | 21:43 |
| Control Test | | 0.078 | 21:43 |
| Air Blank | | 0.000 | 21:44 |
| Subject Sample | # 1 | 0.072 | 21:45 |
| Breath Volume | | 3.437 | 21:45 |
| Air Blank | | 0.000 | 21:45 |
| Air Blank | | 0.000 | 21:47 |
| Subject Sample | # 2 | 0.064 | 21:48 |
| Breath Volume | | 2.792 | 21:48 |
| Air Blank | | 0.000 | 21:48 |
| Control Test | | 0.079 | 21:49 |
| Air Blank | | 0.000 | 21:49 |
| Diagnostics Check | | OK | 21:49 |

Data Download Date/Time: 05/23/2026 22:10

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: JEREZANO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: BASTIEN

DATE/TIME OF TEST: 05/21/2026 03:45:04

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:38 |
| Air Blank | | 0.000 | 03:39 |
| Control Test | | 0.079 | 03:39 |
| Air Blank | | 0.000 | 03:40 |
| Subject Sample | # 1 | 0.143 | 03:40 |
| Breath Volume | | 2.335 | 03:40 |
| Air Blank | | 0.000 | 03:41 |
| Air Blank | | 0.000 | 03:43 |
| Subject Sample | # 2 | 0.143 | 03:43 |
| Breath Volume | | 2.582 | 03:43 |
| Air Blank | | 0.000 | 03:44 |
| Control Test | | 0.078 | 03:44 |
| Air Blank | | 0.000 | 03:44 |
| Diagnostics Check | | OK | 03:45 |

Data Download Date/Time: 05/23/2026 22:10

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 05/23/2026 02:36:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:30 |
| Air Blank | | 0.000 | 02:30 |
| Control Test | | 0.078 | 02:31 |
| Air Blank | | 0.000 | 02:31 |
| Subject Sample | # 1 | 0.302 | 02:32 |
| Breath Volume | | 2.019 | 02:32 |
| Air Blank | | 0.000 | 02:32 |
| Air Blank | | 0.000 | 02:34 |
| Subject Sample | # 2 | 0.302 | 02:35 |
| Breath Volume | | 2.699 | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Control Test | | 0.079 | 02:36 |
| Air Blank | | 0.000 | 02:36 |
| Diagnostics Check | | OK | 02:36 |

Data Download Date/Time: 05/23/2026 22:10

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 05/29/2026 00:06:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:59 |
| Air Blank | | 0.000 | 00:00 |
| Control Test | | 0.078 | 00:00 |
| Air Blank | | 0.000 | 00:01 |
| Subject Sample | # 1 | 0.146 | 00:01 |
| Breath Volume | | 2.878 | 00:01 |
| Air Blank | | 0.000 | 00:02 |
| Air Blank | | 0.000 | 00:04 |
| Subject Sample | # 2 | 0.137 | 00:04 |
| Breath Volume | | 2.921 | 00:04 |
| Air Blank | | 0.000 | 00:05 |
| Control Test | | 0.079 | 00:05 |
| Air Blank | | 0.000 | 00:06 |
| Diagnostics Check | | OK | 00:06 |

Data Download Date/Time: 06/17/2026 23:44

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: ORELIEN

DATE/TIME OF TEST: 05/30/2026 23:06:14

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:54 | |
| Air Blank | | 0.000 | 22:55 | |
| Control Test | | 0.078 | 22:55 | |
| Air Blank | | 0.000 | 22:56 | |
| Subject Sample | # 1 | | 22:59 | No Sample Provided |
| Breath Volume | | 0.000 | 22:59 | |
| Air Blank | | 0.000 | 22:59 | |
| Air Blank | | 0.000 | 23:01 | |
| Subject Sample | # 2 | | 23:04 | No Sample Provided |
| Breath Volume | | 0.000 | 23:04 | |
| Air Blank | | 0.000 | 23:05 | |
| Control Test | | 0.078 | 23:05 | |
| Air Blank | | 0.000 | 23:06 | |
| Diagnostics Check | | OK | 23:06 | |

Data Download Date/Time: 06/17/2026 23:44

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006767

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: IZQUIERDO-RODRIGUEZ

DATE/TIME OF TEST: 05/31/2026 02:57:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:49 |
| Air Blank | | 0.000 | 02:50 |
| Control Test | | 0.079 | 02:50 |
| Air Blank | | 0.000 | 02:51 |
| Subject Sample | # 1 | 0.161 | 02:52 |
| Breath Volume | | 2.664 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 2 | 0.162 | 02:55 |
| Breath Volume | | 3.546 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Control Test | | 0.078 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Diagnostics Check | | OK | 02:57 |

Data Download Date/Time: 06/17/2026 23:44

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOME,PATRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 11:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:24

SUBJECT'S LAST NAME: FINCHUM

DATE/TIME OF TEST: 05/19/2026 11:46:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 11:44 | |
| Air Blank | | 0.000 | 11:45 | |
| Control Test | | 0.078 | 11:45 | |
| Air Blank | | 0.000 | 11:46 | |
| Subject Sample | # 1 | | 11:46 | Sequence Aborted |
| Breath Volume | | 0.000 | 11:46 | |
| Air Blank | | 0.000 | 11:46 | |

Data Download Date/Time: 06/02/2026 07:01

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOME,PATRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROOME,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 12/30/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:47

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 05/27/2026 14:12:01

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 14:08 | |
| Air Blank | | 0.000 | 14:08 | |
| Control Test | | 0.078 | 14:09 | |
| Air Blank | | 0.000 | 14:09 | |
| Subject Sample | # 1 | | 14:10 | Subject Test Refused |
| Breath Volume | | 0.000 | 14:10 | |
| Air Blank | | 0.000 | 14:11 | |
| Control Test | | 0.077 | 14:11 | |
| Air Blank | | 0.000 | 14:11 | |
| Diagnostics Check | | OK | 14:12 | |

Data Download Date/Time: 06/02/2026 07:01

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOME,PATRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROOME,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/05/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: ESPOSITO

DATE/TIME OF TEST: 05/28/2026 08:20:19

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 08:08 | |
| Air Blank | | 0.000 | 08:08 | |
| Control Test | | 0.077 | 08:09 | |
| Air Blank | | 0.000 | 08:09 | |
| Subject Sample | # 1 | 0.159 | 08:12 | Volume Not Met |
| Breath Volume | | 0.015 | 08:12 | |
| Air Blank | | 0.000 | 08:13 | |
| Air Blank | | 0.000 | 08:15 | |
| Subject Sample | # 2 | 0.219 | 08:15 | |
| Breath Volume | | 2.906 | 08:15 | |
| Air Blank | | 0.000 | 08:16 | |
| Air Blank | | 0.000 | 08:18 | |
| Subject Sample | # 3 | 0.202 | 08:18 | |
| Breath Volume | | 2.394 | 08:18 | |
| Air Blank | | 0.000 | 08:19 | |
| Control Test | | 0.077 | 08:19 | |
| Air Blank | | 0.000 | 08:20 | |
| Diagnostics Check | | OK | 08:20 | |

Data Download Date/Time: 06/02/2026 07:01

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,CORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SAPP,CORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: CAPPILLINO

DATE/TIME OF TEST: 05/18/2026 04:36:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:29 |
| Air Blank | | 0.000 | 04:30 |
| Control Test | | 0.080 | 04:30 |
| Air Blank | | 0.000 | 04:30 |
| Subject Sample | # 1 | 0.149 | 04:31 |
| Breath Volume | | 1.988 | 04:31 |
| Air Blank | | 0.000 | 04:32 |
| Air Blank | | 0.000 | 04:34 |
| Subject Sample | # 2 | 0.144 | 04:34 |
| Breath Volume | | 1.496 | 04:34 |
| Air Blank | | 0.000 | 04:35 |
| Control Test | | 0.079 | 04:35 |
| Air Blank | | 0.000 | 04:36 |
| Diagnostics Check | | OK | 04:36 |

Data Download Date/Time: 06/24/2026 12:21

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SANGRINO,A
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: MENDES BARBOSA

DATE/TIME OF TEST: 05/10/2026 06:36:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:28 |
| Air Blank | | 0.000 | 06:29 |
| Control Test | | 0.081 | 06:29 |
| Air Blank | | 0.000 | 06:30 |
| Subject Sample | # 1 | 0.158 | 06:31 |
| Breath Volume | | 2.457 | 06:31 |
| Air Blank | | 0.000 | 06:32 |
| Air Blank | | 0.000 | 06:33 |
| Subject Sample | # 2 | 0.148 | 06:34 |
| Breath Volume | | 1.597 | 06:34 |
| Air Blank | | 0.000 | 06:35 |
| Control Test | | 0.080 | 06:35 |
| Air Blank | | 0.000 | 06:36 |
| Diagnostics Check | | OK | 06:36 |

Data Download Date/Time: 06/03/2026 23:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: VASQUEZ RAMRIEZ

DATE/TIME OF TEST: 05/17/2026 02:41:40

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:31 | |
| Air Blank | | 0.000 | 02:31 | |
| Control Test | | 0.080 | 02:32 | |
| Air Blank | | 0.000 | 02:32 | |
| Subject Sample | # 1 | 0.247 | 02:33 | |
| Breath Volume | | 1.167 | 02:33 | |
| Air Blank | | 0.000 | 02:33 | |
| Air Blank | | 0.000 | 02:35 | |
| Subject Sample | # 2 | 0.226 | 02:37 | |
| Breath Volume | | 1.972 | 02:37 | |
| Air Blank | | 0.000 | 02:37 | |
| Air Blank | | 0.000 | 02:39 | |
| Subject Sample | # 3 | | 02:40 | Slope Not Met |
| Breath Volume | | 1.332 | 02:40 | |
| Air Blank | | 0.000 | 02:40 | |
| Control Test | | 0.080 | 02:41 | |
| Air Blank | | 0.000 | 02:41 | |
| Diagnostics Check | | OK | 02:41 | |

Data Download Date/Time: 06/03/2026 23:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: VASQUEZ RAMRIEZ

DATE/TIME OF TEST: 05/17/2026 03:10:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:04 |
| Air Blank | | 0.000 | 03:05 |
| Control Test | | 0.081 | 03:05 |
| Air Blank | | 0.000 | 03:05 |
| Subject Sample | # 1 | 0.238 | 03:06 |
| Breath Volume | | 1.250 | 03:06 |
| Air Blank | | 0.000 | 03:07 |
| Air Blank | | 0.000 | 03:09 |
| Subject Sample | # 2 | 0.237 | 03:09 |
| Breath Volume | | 1.296 | 03:09 |
| Air Blank | | 0.000 | 03:10 |
| Control Test | | 0.079 | 03:10 |
| Air Blank | | 0.000 | 03:10 |
| Diagnostics Check | | OK | 03:10 |

Data Download Date/Time: 06/03/2026 23:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ALTUNAY,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: PEDROSO RIVERA

DATE/TIME OF TEST: 05/20/2026 22:48:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:41 |
| Air Blank | | 0.000 | 22:41 |
| Control Test | | 0.083 | 22:42 |
| Air Blank | | 0.000 | 22:42 |
| Subject Sample | # 1 | 0.113 | 22:43 |
| Breath Volume | | 1.332 | 22:43 |
| Air Blank | | 0.000 | 22:44 |
| Air Blank | | 0.000 | 22:46 |
| Subject Sample | # 2 | 0.115 | 22:47 |
| Breath Volume | | 1.218 | 22:47 |
| Air Blank | | 0.000 | 22:47 |
| Control Test | | 0.080 | 22:48 |
| Air Blank | | 0.000 | 22:48 |
| Diagnostics Check | | OK | 22:48 |

Data Download Date/Time: 06/03/2026 23:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LEVY,A
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:31

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 05/21/2026 07:08:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 07:01 |
| Air Blank | | 0.000 | 07:02 |
| Control Test | | 0.081 | 07:02 |
| Air Blank | | 0.000 | 07:03 |
| Subject Sample | # 1 | 0.147 | 07:03 |
| Breath Volume | | 2.285 | 07:03 |
| Air Blank | | 0.000 | 07:04 |
| Air Blank | | 0.000 | 07:06 |
| Subject Sample | # 2 | 0.134 | 07:06 |
| Breath Volume | | 1.492 | 07:06 |
| Air Blank | | 0.000 | 07:07 |
| Control Test | | 0.081 | 07:07 |
| Air Blank | | 0.000 | 07:08 |
| Diagnostics Check | | OK | 07:08 |

Data Download Date/Time: 06/03/2026 23:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006895

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: BANEGAS ACEITUNO

DATE/TIME OF TEST: 05/23/2026 00:44:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:36 |
| Air Blank | | 0.000 | 00:37 |
| Control Test | | 0.080 | 00:37 |
| Air Blank | | 0.000 | 00:38 |
| Subject Sample | # 1 | 0.029 | 00:39 |
| Breath Volume | | 1.562 | 00:39 |
| Air Blank | | 0.000 | 00:40 |
| Air Blank | | 0.000 | 00:41 |
| Subject Sample | # 2 | 0.028 | 00:43 |
| Breath Volume | | 1.367 | 00:43 |
| Air Blank | | 0.000 | 00:43 |
| Control Test | | 0.080 | 00:44 |
| Air Blank | | 0.000 | 00:44 |
| Diagnostics Check | | OK | 00:44 |

Data Download Date/Time: 06/03/2026 23:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GARCIA,CAROLYN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: BURROWS

DATE/TIME OF TEST: 05/09/2026 01:33:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:27 |
| Air Blank | | 0.000 | 01:27 |
| Control Test | | 0.079 | 01:27 |
| Air Blank | | 0.000 | 01:28 |
| Subject Sample | # 1 | 0.178 | 01:28 |
| Breath Volume | | 2.644 | 01:28 |
| Air Blank | | 0.000 | 01:29 |
| Air Blank | | 0.000 | 01:31 |
| Subject Sample | # 2 | 0.186 | 01:31 |
| Breath Volume | | 2.664 | 01:31 |
| Air Blank | | 0.000 | 01:32 |
| Control Test | | 0.078 | 01:32 |
| Air Blank | | 0.000 | 01:33 |
| Diagnostics Check | | OK | 01:33 |

Data Download Date/Time: 06/10/2026 07:39

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOPEZ,ALYSSA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ACOSTA RODRIGUEZ

DATE/TIME OF TEST: 05/09/2026 02:30:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:23 |
| Air Blank | | 0.000 | 02:24 |
| Control Test | | 0.078 | 02:24 |
| Air Blank | | 0.000 | 02:25 |
| Subject Sample | # 1 | 0.149 | 02:26 |
| Breath Volume | | 1.117 | 02:26 |
| Air Blank | | 0.000 | 02:26 |
| Air Blank | | 0.000 | 02:28 |
| Subject Sample | # 2 | 0.159 | 02:28 |
| Breath Volume | | 1.355 | 02:28 |
| Air Blank | | 0.000 | 02:29 |
| Control Test | | 0.078 | 02:29 |
| Air Blank | | 0.000 | 02:30 |
| Diagnostics Check | | OK | 02:30 |

Data Download Date/Time: 06/10/2026 07:39

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ENGEL,ANTHONY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: DRISDOM

DATE/TIME OF TEST: 05/10/2026 21:21:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:14 |
| Air Blank | | 0.000 | 21:15 |
| Control Test | | 0.077 | 21:15 |
| Air Blank | | 0.000 | 21:16 |
| Subject Sample | # 1 | 0.154 | 21:16 |
| Breath Volume | | 2.144 | 21:16 |
| Air Blank | | 0.000 | 21:17 |
| Air Blank | | 0.000 | 21:19 |
| Subject Sample | # 2 | 0.153 | 21:19 |
| Breath Volume | | 1.511 | 21:19 |
| Air Blank | | 0.000 | 21:20 |
| Control Test | | 0.076 | 21:20 |
| Air Blank | | 0.000 | 21:21 |
| Diagnostics Check | | OK | 21:21 |

Data Download Date/Time: 06/10/2026 07:39

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DAVIS,LUKE
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: COMFORT

DATE/TIME OF TEST: 05/11/2026 00:43:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:35 |
| Air Blank | | 0.000 | 00:36 |
| Control Test | | 0.078 | 00:36 |
| Air Blank | | 0.000 | 00:37 |
| Subject Sample | # 1 | 0.404 | 00:37 |
| Breath Volume | | 1.242 | 00:37 |
| Air Blank | | 0.000 | 00:38 |
| Air Blank | | 0.000 | 00:40 |
| Subject Sample | # 2 | 0.384 | 00:41 |
| Breath Volume | | 1.347 | 00:41 |
| Air Blank | | 0.000 | 00:42 |
| Control Test | | 0.077 | 00:42 |
| Air Blank | | 0.000 | 00:43 |
| Diagnostics Check | | OK | 00:43 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KENNY,MICHAEL
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 05/11/2026 05:08:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:01 |
| Air Blank | | 0.000 | 05:01 |
| Control Test | | 0.078 | 05:01 |
| Air Blank | | 0.000 | 05:02 |
| Subject Sample | # 1 | 0.191 | 05:03 |
| Breath Volume | | 1.277 | 05:03 |
| Air Blank | | 0.000 | 05:03 |
| Air Blank | | 0.000 | 05:05 |
| Subject Sample | # 2 | 0.191 | 05:06 |
| Breath Volume | | 1.890 | 05:06 |
| Air Blank | | 0.000 | 05:07 |
| Control Test | | 0.078 | 05:07 |
| Air Blank | | 0.000 | 05:08 |
| Diagnostics Check | | OK | 05:08 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NORRITO,GIUSEPPE
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: PALACIOS TORREZ

DATE/TIME OF TEST: 05/14/2026 17:57:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 17:50 |
| Air Blank | | 0.000 | 17:50 |
| Control Test | | 0.079 | 17:51 |
| Air Blank | | 0.000 | 17:51 |
| Subject Sample | # 1 | 0.136 | 17:52 |
| Breath Volume | | 1.984 | 17:52 |
| Air Blank | | 0.000 | 17:53 |
| Air Blank | | 0.000 | 17:55 |
| Subject Sample | # 2 | 0.143 | 17:56 |
| Breath Volume | | 2.609 | 17:56 |
| Air Blank | | 0.000 | 17:56 |
| Control Test | | 0.080 | 17:57 |
| Air Blank | | 0.000 | 17:57 |
| Diagnostics Check | | OK | 17:57 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RADUTIU,TOMMY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: KISHMAN

DATE/TIME OF TEST: 05/16/2026 02:18:56

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:12 |
| Air Blank | | 0.000 | 02:12 |
| Control Test | | 0.078 | 02:13 |
| Air Blank | | 0.000 | 02:13 |
| Subject Sample | # 1 | 0.171 | 02:14 |
| Breath Volume | | 3.007 | 02:14 |
| Air Blank | | 0.000 | 02:15 |
| Air Blank | | 0.000 | 02:16 |
| Subject Sample | # 2 | 0.164 | 02:17 |
| Breath Volume | | 2.468 | 02:17 |
| Air Blank | | 0.000 | 02:18 |
| Control Test | | 0.078 | 02:18 |
| Air Blank | | 0.000 | 02:18 |
| Diagnostics Check | | OK | 02:18 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PURRONE,ZACHARY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: LAFFITE

DATE/TIME OF TEST: 05/17/2026 07:45:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 07:39 |
| Air Blank | | 0.000 | 07:39 |
| Control Test | | 0.080 | 07:40 |
| Air Blank | | 0.000 | 07:40 |
| Subject Sample | # 1 | 0.113 | 07:41 |
| Breath Volume | | 2.179 | 07:41 |
| Air Blank | | 0.000 | 07:41 |
| Air Blank | | 0.000 | 07:43 |
| Subject Sample | # 2 | 0.111 | 07:44 |
| Breath Volume | | 1.398 | 07:44 |
| Air Blank | | 0.000 | 07:44 |
| Control Test | | 0.078 | 07:45 |
| Air Blank | | 0.000 | 07:45 |
| Diagnostics Check | | OK | 07:45 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SHOLOMOV,RAFAEL
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: WALTERS III

DATE/TIME OF TEST: 05/18/2026 04:23:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:16 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.078 | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Subject Sample | # 1 | 0.103 | 04:18 |
| Breath Volume | | 1.265 | 04:18 |
| Air Blank | | 0.000 | 04:19 |
| Air Blank | | 0.000 | 04:21 |
| Subject Sample | # 2 | 0.113 | 04:21 |
| Breath Volume | | 3.023 | 04:21 |
| Air Blank | | 0.000 | 04:22 |
| Control Test | | 0.077 | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Diagnostics Check | | OK | 04:23 |

Data Download Date/Time: 06/10/2026 07:40


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAPOTA,JOSHUA
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:38

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 05/19/2026 15:12:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:06 |
| Air Blank | | 0.000 | 15:06 |
| Control Test | | 0.079 | 15:07 |
| Air Blank | | 0.000 | 15:07 |
| Subject Sample | # 1 | 0.000 | 15:08 |
| Breath Volume | | 3.503 | 15:08 |
| Air Blank | | 0.000 | 15:08 |
| Air Blank | | 0.000 | 15:10 |
| Subject Sample | # 2 | 0.000 | 15:11 |
| Breath Volume | | 3.613 | 15:11 |
| Air Blank | | 0.000 | 15:11 |
| Control Test | | 0.079 | 15:12 |
| Air Blank | | 0.000 | 15:12 |
| Diagnostics Check | | OK | 15:12 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GARCIA,CAROLYN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: LOPEZ GODINEZ

DATE/TIME OF TEST: 05/22/2026 00:31:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:23 |
| Air Blank | | 0.000 | 00:24 |
| Control Test | | 0.079 | 00:24 |
| Air Blank | | 0.000 | 00:25 |
| Subject Sample | # 1 | 0.186 | 00:26 |
| Breath Volume | | 1.203 | 00:26 |
| Air Blank | | 0.000 | 00:26 |
| Air Blank | | 0.000 | 00:28 |
| Subject Sample | # 2 | 0.197 | 00:29 |
| Breath Volume | | 2.171 | 00:29 |
| Air Blank | | 0.000 | 00:30 |
| Control Test | | 0.077 | 00:30 |
| Air Blank | | 0.000 | 00:31 |
| Diagnostics Check | | OK | 00:31 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROMANO,FABIOLA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 05/22/2026 20:14:42

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:02 | |
| Air Blank | | 0.000 | 20:03 | |
| Control Test | | 0.078 | 20:03 | |
| Air Blank | | 0.000 | 20:04 | |
| Subject Sample | # 1 | 0.203 | 20:07 | Volume Not Met |
| Breath Volume | | 0.628 | 20:07 | |
| Air Blank | | 0.000 | 20:08 | |
| Air Blank | | 0.000 | 20:09 | |
| Subject Sample | # 2 | 0.212 | 20:12 | Volume Not Met |
| Breath Volume | | 1.085 | 20:12 | |
| Air Blank | | 0.000 | 20:13 | |
| Control Test | | 0.075 | 20:14 | |
| Air Blank | | 0.000 | 20:14 | |
| Diagnostics Check | | OK | 20:14 | |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROMANO,FABIOLA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 05/22/2026 20:24:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:17 |
| Air Blank | | 0.000 | 20:18 |
| Control Test | | 0.075 | 20:18 |
| Air Blank | | 0.000 | 20:18 |
| Subject Sample | # 1 | 0.223 | 20:19 |
| Breath Volume | | 1.183 | 20:19 |
| Air Blank | | 0.000 | 20:20 |
| Air Blank | | 0.000 | 20:22 |
| Subject Sample | # 2 | 0.207 | 20:22 |
| Breath Volume | | 1.105 | 20:22 |
| Air Blank | | 0.000 | 20:23 |
| Control Test | | 0.076 | 20:23 |
| Air Blank | | 0.000 | 20:24 |
| Diagnostics Check | | OK | 20:24 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: HMEIDAN

DATE/TIME OF TEST: 05/22/2026 23:57:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:50 |
| Air Blank | | 0.000 | 23:50 |
| Control Test | | 0.077 | 23:51 |
| Air Blank | | 0.000 | 23:51 |
| Subject Sample | # 1 | 0.194 | 23:52 |
| Breath Volume | | 1.402 | 23:52 |
| Air Blank | | 0.000 | 23:53 |
| Air Blank | | 0.000 | 23:55 |
| Subject Sample | # 2 | 0.196 | 23:55 |
| Breath Volume | | 2.066 | 23:55 |
| Air Blank | | 0.000 | 23:56 |
| Control Test | | 0.077 | 23:56 |
| Air Blank | | 0.000 | 23:57 |
| Diagnostics Check | | OK | 23:57 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 05/23/2026 01:52:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:45 |
| Air Blank | | 0.000 | 01:45 |
| Control Test | | 0.078 | 01:46 |
| Air Blank | | 0.000 | 01:46 |
| Subject Sample | # 1 | 0.124 | 01:47 |
| Breath Volume | | 1.535 | 01:47 |
| Air Blank | | 0.000 | 01:48 |
| Air Blank | | 0.000 | 01:50 |
| Subject Sample | # 2 | 0.128 | 01:50 |
| Breath Volume | | 1.238 | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Control Test | | 0.079 | 01:51 |
| Air Blank | | 0.000 | 01:52 |
| Diagnostics Check | | OK | 01:52 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: AN

DATE/TIME OF TEST: 05/23/2026 02:27:27

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:13 | |
| Air Blank | | 0.000 | 02:13 | |
| Control Test | | 0.079 | 02:14 | |
| Air Blank | | 0.000 | 02:14 | |
| Subject Sample | # 1 | 0.216 | 02:16 | |
| Breath Volume | | 1.371 | 02:16 | |
| Air Blank | | 0.000 | 02:17 | |
| Air Blank | | 0.000 | 02:18 | |
| Subject Sample | # 2 | 0.215 | 02:22 | Volume Not Met |
| Breath Volume | | 0.382 | 02:22 | |
| Air Blank | | 0.000 | 02:23 | |
| Air Blank | | 0.000 | 02:24 | |
| Subject Sample | # 3 | 0.229 | 02:25 | |
| Breath Volume | | 2.054 | 02:25 | |
| Air Blank | | 0.000 | 02:26 | |
| Control Test | | 0.073 | 02:26 | Control Outside Tolerance |
| Air Blank | | 0.000 | 02:27 | |

Data Download Date/Time: 06/10/2026 07:40


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: AN

DATE/TIME OF TEST: 05/23/2026 02:37:39

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:30 |
| Air Blank | | 0.000 | 02:31 |
| Control Test | | 0.077 | 02:31 |
| Air Blank | | 0.000 | 02:31 |
| Subject Sample | # 1 | 0.207 | 02:32 |
| Breath Volume | | 1.402 | 02:32 |
| Air Blank | | 0.000 | 02:33 |
| Air Blank | | 0.000 | 02:35 |
| Subject Sample | # 2 | 0.216 | 02:36 |
| Breath Volume | | 1.566 | 02:36 |
| Air Blank | | 0.000 | 02:36 |
| Control Test | | 0.075 | 02:37 |
| Air Blank | | 0.000 | 02:37 |
| Diagnostics Check | | OK | 02:37 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOWE,RACHEL
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: FELISME

DATE/TIME OF TEST: 05/23/2026 04:18:28

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:11 |
| Air Blank | | 0.000 | 04:11 |
| Control Test | | 0.078 | 04:11 |
| Air Blank | | 0.000 | 04:12 |
| Subject Sample | # 1 | 0.167 | 04:13 |
| Breath Volume | | 1.867 | 04:13 |
| Air Blank | | 0.000 | 04:14 |
| Air Blank | | 0.000 | 04:15 |
| Subject Sample | # 2 | 0.159 | 04:16 |
| Breath Volume | | 1.789 | 04:16 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.077 | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Diagnostics Check | | OK | 04:18 |

Data Download Date/Time: 06/10/2026 07:40

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GOMEZ,DESTINEE
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/23/2026 05:12:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:05 |
| Air Blank | | 0.000 | 05:05 |
| Control Test | | 0.078 | 05:05 |
| Air Blank | | 0.000 | 05:06 |
| Subject Sample | # 1 | 0.153 | 05:07 |
| Breath Volume | | 1.156 | 05:07 |
| Air Blank | | 0.000 | 05:08 |
| Air Blank | | 0.000 | 05:10 |
| Subject Sample | # 2 | 0.155 | 05:11 |
| Breath Volume | | 1.257 | 05:11 |
| Air Blank | | 0.000 | 05:11 |
| Control Test | | 0.077 | 05:12 |
| Air Blank | | 0.000 | 05:12 |
| Diagnostics Check | | OK | 05:12 |

Data Download Date/Time: 06/10/2026 07:41

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HOWE,RACHEL
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ALFONSE

DATE/TIME OF TEST: 05/25/2026 02:37:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:29 |
| Air Blank | | 0.000 | 02:30 |
| Control Test | | 0.078 | 02:30 |
| Air Blank | | 0.000 | 02:31 |
| Subject Sample | # 1 | 0.254 | 02:32 |
| Breath Volume | | 1.132 | 02:32 |
| Air Blank | | 0.000 | 02:32 |
| Air Blank | | 0.000 | 02:34 |
| Subject Sample | # 2 | 0.265 | 02:35 |
| Breath Volume | | 1.660 | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Control Test | | 0.076 | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Diagnostics Check | | OK | 02:37 |

Data Download Date/Time: 06/10/2026 07:41

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PURRONE,ZACHARY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 05/25/2026 03:03:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:57 |
| Air Blank | | 0.000 | 02:57 |
| Control Test | | 0.078 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Subject Sample | # 1 | 0.142 | 02:58 |
| Breath Volume | | 1.667 | 02:58 |
| Air Blank | | 0.000 | 02:59 |
| Air Blank | | 0.000 | 03:01 |
| Subject Sample | # 2 | 0.144 | 03:01 |
| Breath Volume | | 1.765 | 03:01 |
| Air Blank | | 0.000 | 03:02 |
| Control Test | | 0.077 | 03:02 |
| Air Blank | | 0.000 | 03:03 |
| Diagnostics Check | | OK | 03:03 |

Data Download Date/Time: 06/10/2026 07:41

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOPEZ,BRYAN
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 07:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: VERGARA VERGARA

DATE/TIME OF TEST: 05/26/2026 09:06:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:59 |
| Air Blank | | 0.000 | 09:00 |
| Control Test | | 0.079 | 09:00 |
| Air Blank | | 0.000 | 09:01 |
| Subject Sample | # 1 | 0.177 | 09:01 |
| Breath Volume | | 3.128 | 09:01 |
| Air Blank | | 0.000 | 09:02 |
| Air Blank | | 0.000 | 09:04 |
| Subject Sample | # 2 | 0.166 | 09:05 |
| Breath Volume | | 2.960 | 09:05 |
| Air Blank | | 0.000 | 09:05 |
| Control Test | | 0.078 | 09:06 |
| Air Blank | | 0.000 | 09:06 |
| Diagnostics Check | | OK | 09:06 |

Data Download Date/Time: 06/10/2026 07:41

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GARRIDO,PIERO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:11

SUBJECT'S LAST NAME: CORDOBA BERMEO

DATE/TIME OF TEST: 05/31/2026 09:52:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 09:39 | |
| Air Blank | | 0.000 | 09:39 | |
| Control Test | | 0.079 | 09:40 | |
| Air Blank | | 0.000 | 09:40 | |
| Subject Sample | # 1 | 0.110 | 09:43 | Volume Not Met |
| Breath Volume | | 0.476 | 09:43 | |
| Air Blank | | 0.000 | 09:44 | |
| Air Blank | | 0.000 | 09:46 | |
| Subject Sample | # 2 | 0.112 | 09:47 | |
| Breath Volume | | 2.847 | 09:47 | |
| Air Blank | | 0.000 | 09:48 | |
| Air Blank | | 0.000 | 09:50 | |
| Subject Sample | # 3 | 0.114 | 09:51 | |
| Breath Volume | | 1.558 | 09:51 | |
| Air Blank | | 0.000 | 09:51 | |
| Control Test | | 0.077 | 09:52 | |
| Air Blank | | 0.000 | 09:52 | |
| Diagnostics Check | | OK | 09:52 | |

Data Download Date/Time: 06/10/2026 07:41

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DELACAMPA,BRYANT
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: QUINCE

DATE/TIME OF TEST: 05/01/2026 03:21:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:14 |
| Air Blank | | 0.000 | 03:14 |
| Control Test | | 0.078 | 03:14 |
| Air Blank | | 0.000 | 03:15 |
| Subject Sample | # 1 | 0.146 | 03:16 |
| Breath Volume | | 1.695 | 03:16 |
| Air Blank | | 0.000 | 03:16 |
| Air Blank | | 0.000 | 03:18 |
| Subject Sample | # 2 | 0.136 | 03:19 |
| Breath Volume | | 1.808 | 03:19 |
| Air Blank | | 0.000 | 03:20 |
| Control Test | | 0.082 | 03:20 |
| Air Blank | | 0.000 | 03:21 |
| Diagnostics Check | | OK | 03:21 |

Data Download Date/Time: 05/06/2026 07:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WRIGHT,KIERSTEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: POVEDA

DATE/TIME OF TEST: 05/02/2026 03:53:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:46 |
| Air Blank | | 0.000 | 03:46 |
| Control Test | | 0.079 | 03:47 |
| Air Blank | | 0.000 | 03:47 |
| Subject Sample | # 1 | 0.186 | 03:48 |
| Breath Volume | | 1.820 | 03:48 |
| Air Blank | | 0.000 | 03:49 |
| Air Blank | | 0.000 | 03:50 |
| Subject Sample | # 2 | 0.191 | 03:51 |
| Breath Volume | | 1.867 | 03:51 |
| Air Blank | | 0.000 | 03:52 |
| Control Test | | 0.078 | 03:52 |
| Air Blank | | 0.000 | 03:53 |
| Diagnostics Check | | OK | 03:53 |

Data Download Date/Time: 05/06/2026 07:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MOISE

DATE/TIME OF TEST: 05/05/2026 02:26:52

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:20 |
| Air Blank | | 0.000 | 02:20 |
| Control Test | | 0.079 | 02:20 |
| Air Blank | | 0.000 | 02:21 |
| Subject Sample | # 1 | 0.160 | 02:22 |
| Breath Volume | | 1.335 | 02:22 |
| Air Blank | | 0.000 | 02:22 |
| Air Blank | | 0.000 | 02:24 |
| Subject Sample | # 2 | 0.159 | 02:24 |
| Breath Volume | | 1.445 | 02:24 |
| Air Blank | | 0.000 | 02:25 |
| Control Test | | 0.079 | 02:26 |
| Air Blank | | 0.000 | 02:26 |
| Diagnostics Check | | OK | 02:26 |

Data Download Date/Time: 05/06/2026 07:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 05/10/2026 03:42:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:34 |
| Air Blank | | 0.000 | 03:35 |
| Control Test | | 0.079 | 03:35 |
| Air Blank | | 0.000 | 03:36 |
| Subject Sample | # 1 | 0.210 | 03:37 |
| Breath Volume | | 1.683 | 03:37 |
| Air Blank | | 0.000 | 03:38 |
| Air Blank | | 0.000 | 03:39 |
| Subject Sample | # 2 | 0.203 | 03:40 |
| Breath Volume | | 2.027 | 03:40 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.078 | 03:41 |
| Air Blank | | 0.000 | 03:42 |
| Diagnostics Check | | OK | 03:42 |

Data Download Date/Time: 06/10/2026 09:55

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: QUINTERO,THOMAS
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 05/14/2026 00:48:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:42 |
| Air Blank | | 0.000 | 00:42 |
| Control Test | | 0.079 | 00:43 |
| Air Blank | | 0.000 | 00:43 |
| Subject Sample | # 1 | 0.201 | 00:44 |
| Breath Volume | | 2.855 | 00:44 |
| Air Blank | | 0.000 | 00:44 |
| Air Blank | | 0.000 | 00:46 |
| Subject Sample | # 2 | 0.212 | 00:47 |
| Breath Volume | | 2.402 | 00:47 |
| Air Blank | | 0.000 | 00:47 |
| Control Test | | 0.078 | 00:48 |
| Air Blank | | 0.000 | 00:48 |
| Diagnostics Check | | OK | 00:48 |

Data Download Date/Time: 06/10/2026 09:55

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: ZEITOUNE

DATE/TIME OF TEST: 05/14/2026 04:09:28

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:03 | |
| Air Blank | | 0.000 | 04:04 | |
| Control Test | | 0.079 | 04:04 | |
| Air Blank | | 0.000 | 04:05 | |
| Subject Sample | # 1 | 0.120 | 04:07 | |
| Breath Volume | | 1.523 | 04:07 | |
| Air Blank | | | 04:08 | RFI Detect |
| Air Blank | | 0.000 | 04:08 | |
| Control Test | | 0.085 | 04:08 | |
| Air Blank | | 0.000 | 04:09 | |
| Diagnostics Check | | OK | 04:09 | |

Data Download Date/Time: 06/10/2026 09:55

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: ZEITOUNE

DATE/TIME OF TEST: 05/14/2026 04:18:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:11 |
| Air Blank | | 0.000 | 04:12 |
| Control Test | | 0.079 | 04:12 |
| Air Blank | | 0.000 | 04:13 |
| Subject Sample | # 1 | 0.122 | 04:14 |
| Breath Volume | | 1.679 | 04:14 |
| Air Blank | | 0.000 | 04:14 |
| Air Blank | | 0.000 | 04:16 |
| Subject Sample | # 2 | 0.108 | 04:17 |
| Breath Volume | | 1.144 | 04:17 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.080 | 04:18 |
| Air Blank | | 0.000 | 04:18 |
| Diagnostics Check | | OK | 04:18 |

Data Download Date/Time: 06/10/2026 09:55

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,JEREMY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: TORRACA SANTIAGO

DATE/TIME OF TEST: 05/15/2026 01:13:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:06 |
| Air Blank | | 0.000 | 01:06 |
| Control Test | | 0.080 | 01:06 |
| Air Blank | | 0.000 | 01:07 |
| Subject Sample | # 1 | 0.187 | 01:08 |
| Breath Volume | | 1.406 | 01:08 |
| Air Blank | | 0.000 | 01:09 |
| Air Blank | | 0.000 | 01:10 |
| Subject Sample | # 2 | 0.190 | 01:11 |
| Breath Volume | | 2.316 | 01:11 |
| Air Blank | | 0.000 | 01:12 |
| Control Test | | 0.081 | 01:12 |
| Air Blank | | 0.000 | 01:13 |
| Diagnostics Check | | OK | 01:13 |

Data Download Date/Time: 06/10/2026 09:55

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOGAN,JOHNATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: JIMENEZ SANCHEZ

DATE/TIME OF TEST: 05/15/2026 03:14:07

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:02 | |
| Air Blank | | 0.000 | 03:03 | |
| Control Test | | 0.080 | 03:03 | |
| Air Blank | | 0.000 | 03:03 | |
| Subject Sample | # 1 | 0.217 | 03:07 | Volume Not Met |
| Breath Volume | | 0.058 | 03:07 | |
| Air Blank | | 0.000 | 03:08 | |
| Air Blank | | 0.000 | 03:09 | |
| Subject Sample | # 2 | | 03:12 | No Sample Provided |
| Breath Volume | | 0.000 | 03:12 | |
| Air Blank | | 0.000 | 03:13 | |
| Control Test | | 0.078 | 03:13 | |
| Air Blank | | 0.000 | 03:14 | |
| Diagnostics Check | | OK | 03:14 | |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: COULTON

DATE/TIME OF TEST: 05/15/2026 04:05:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Control Test | | 0.080 | 03:59 |
| Air Blank | | 0.000 | 04:00 |
| Subject Sample | # 1 | 0.167 | 04:00 |
| Breath Volume | | 1.390 | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Air Blank | | 0.000 | 04:03 |
| Subject Sample | # 2 | 0.169 | 04:03 |
| Breath Volume | | 2.332 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Control Test | | 0.080 | 04:04 |
| Air Blank | | 0.000 | 04:05 |
| Diagnostics Check | | OK | 04:05 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: CARRANZA MARTINEZ

DATE/TIME OF TEST: 05/17/2026 06:23:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:16 |
| Air Blank | | 0.000 | 06:16 |
| Control Test | | 0.080 | 06:17 |
| Air Blank | | 0.000 | 06:17 |
| Subject Sample | # 1 | 0.260 | 06:18 |
| Breath Volume | | 1.589 | 06:18 |
| Air Blank | | 0.000 | 06:19 |
| Air Blank | | 0.000 | 06:21 |
| Subject Sample | # 2 | 0.254 | 06:21 |
| Breath Volume | | 1.972 | 06:21 |
| Air Blank | | 0.000 | 06:22 |
| Control Test | | 0.078 | 06:22 |
| Air Blank | | 0.000 | 06:23 |
| Diagnostics Check | | OK | 06:23 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: POLIKANDRIOTIS

DATE/TIME OF TEST: 05/23/2026 23:03:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:56 |
| Air Blank | | 0.000 | 22:57 |
| Control Test | | 0.080 | 22:57 |
| Air Blank | | 0.000 | 22:58 |
| Subject Sample | # 1 | 0.140 | 22:58 |
| Breath Volume | | 1.105 | 22:58 |
| Air Blank | | 0.000 | 22:59 |
| Air Blank | | 0.000 | 23:00 |
| Subject Sample | # 2 | 0.142 | 23:01 |
| Breath Volume | | 1.609 | 23:01 |
| Air Blank | | 0.000 | 23:02 |
| Control Test | | 0.080 | 23:02 |
| Air Blank | | 0.000 | 23:03 |
| Diagnostics Check | | OK | 23:03 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PIERRE,JAYSON
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: GARVIN

DATE/TIME OF TEST: 05/23/2026 23:57:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:49 |
| Air Blank | | 0.000 | 23:50 |
| Control Test | | 0.079 | 23:50 |
| Air Blank | | 0.000 | 23:51 |
| Subject Sample | # 1 | 0.136 | 23:52 |
| Breath Volume | | 2.644 | 23:52 |
| Air Blank | | 0.000 | 23:52 |
| Air Blank | | 0.000 | 23:54 |
| Subject Sample | # 2 | 0.133 | 23:55 |
| Breath Volume | | 3.781 | 23:55 |
| Air Blank | | 0.000 | 23:56 |
| Control Test | | 0.081 | 23:56 |
| Air Blank | | 0.000 | 23:57 |
| Diagnostics Check | | OK | 23:57 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: PEREZ MARTINEZ

DATE/TIME OF TEST: 05/24/2026 01:19:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:10 |
| Air Blank | | 0.000 | 01:10 |
| Control Test | | 0.078 | 01:11 |
| Air Blank | | 0.000 | 01:11 |
| Subject Sample | # 1 | 0.193 | 01:14 |
| Breath Volume | | 2.343 | 01:14 |
| Air Blank | | 0.000 | 01:15 |
| Air Blank | | 0.000 | 01:17 |
| Subject Sample | # 2 | 0.194 | 01:18 |
| Breath Volume | | 1.425 | 01:18 |
| Air Blank | | 0.000 | 01:18 |
| Control Test | | 0.081 | 01:19 |
| Air Blank | | 0.000 | 01:19 |
| Diagnostics Check | | OK | 01:19 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRICENO,OSVALDO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/29/2026 05:42:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:35 |
| Air Blank | | 0.000 | 05:36 |
| Control Test | | 0.079 | 05:36 |
| Air Blank | | 0.000 | 05:37 |
| Subject Sample | # 1 | 0.096 | 05:37 |
| Breath Volume | | 1.281 | 05:37 |
| Air Blank | | 0.000 | 05:38 |
| Air Blank | | 0.000 | 05:40 |
| Subject Sample | # 2 | 0.092 | 05:40 |
| Breath Volume | | 1.214 | 05:40 |
| Air Blank | | 0.000 | 05:41 |
| Control Test | | 0.081 | 05:42 |
| Air Blank | | 0.000 | 05:42 |
| Diagnostics Check | | OK | 05:42 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: MOHAMMED

DATE/TIME OF TEST: 05/30/2026 01:54:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:48 |
| Air Blank | | 0.000 | 01:48 |
| Control Test | | 0.080 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Subject Sample | # 1 | 0.202 | 01:50 |
| Breath Volume | | 2.527 | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Air Blank | | 0.000 | 01:52 |
| Subject Sample | # 2 | 0.204 | 01:53 |
| Breath Volume | | 1.507 | 01:53 |
| Air Blank | | 0.000 | 01:53 |
| Control Test | | 0.079 | 01:54 |
| Air Blank | | 0.000 | 01:54 |
| Diagnostics Check | | OK | 01:54 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 05/30/2026 04:01:35

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:48 | |
| Air Blank | | 0.000 | 03:48 | |
| Control Test | | 0.078 | 03:48 | |
| Air Blank | | 0.000 | 03:49 | |
| Subject Sample | # 1 | | 03:52 | No Sample Provided |
| Breath Volume | | 0.000 | 03:52 | |
| Air Blank | | 0.000 | 03:53 | |
| Air Blank | | 0.000 | 03:55 | |
| Subject Sample | # 2 | 0.202 | 03:55 | |
| Breath Volume | | 1.214 | 03:55 | |
| Air Blank | | 0.000 | 03:56 | |
| Air Blank | | 0.000 | 03:58 | |
| Subject Sample | # 3 | 0.206 | 03:59 | |
| Breath Volume | | 1.187 | 03:59 | |
| Air Blank | | 0.000 | 04:00 | |
| Control Test | | 0.086 | 04:00 | Control Outside Tolerance |
| Air Blank | | 0.000 | 04:01 | |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 05/30/2026 04:11:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:03 |
| Air Blank | | 0.000 | 04:03 |
| Control Test | | 0.079 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Subject Sample | # 1 | 0.203 | 04:05 |
| Breath Volume | | 1.207 | 04:05 |
| Air Blank | | 0.000 | 04:05 |
| Air Blank | | 0.000 | 04:07 |
| Subject Sample | # 2 | 0.214 | 04:09 |
| Breath Volume | | 2.121 | 04:09 |
| Air Blank | | 0.000 | 04:10 |
| Control Test | | 0.085 | 04:10 |
| Air Blank | | 0.000 | 04:11 |
| Diagnostics Check | | OK | 04:11 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: MORFEY

DATE/TIME OF TEST: 05/30/2026 22:34:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:27 |
| Air Blank | | 0.000 | 22:28 |
| Control Test | | 0.079 | 22:28 |
| Air Blank | | 0.000 | 22:28 |
| Subject Sample | # 1 | 0.258 | 22:29 |
| Breath Volume | | 3.062 | 22:29 |
| Air Blank | | 0.000 | 22:30 |
| Air Blank | | 0.000 | 22:32 |
| Subject Sample | # 2 | 0.248 | 22:32 |
| Breath Volume | | 1.906 | 22:32 |
| Air Blank | | 0.000 | 22:33 |
| Control Test | | 0.079 | 22:33 |
| Air Blank | | 0.000 | 22:34 |
| Diagnostics Check | | OK | 22:34 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: ARTIS

DATE/TIME OF TEST: 05/30/2026 23:16:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:09 |
| Air Blank | | 0.000 | 23:09 |
| Control Test | | 0.078 | 23:09 |
| Air Blank | | 0.000 | 23:10 |
| Subject Sample | # 1 | 0.335 | 23:11 |
| Breath Volume | | 1.480 | 23:11 |
| Air Blank | | 0.000 | 23:12 |
| Air Blank | | 0.000 | 23:13 |
| Subject Sample | # 2 | 0.331 | 23:14 |
| Breath Volume | | 1.320 | 23:14 |
| Air Blank | | 0.000 | 23:15 |
| Control Test | | 0.076 | 23:15 |
| Air Blank | | 0.000 | 23:16 |
| Diagnostics Check | | OK | 23:16 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CHRISTIAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: COBO

DATE/TIME OF TEST: 05/31/2026 02:37:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:29 |
| Air Blank | | 0.000 | 02:30 |
| Control Test | | 0.080 | 02:30 |
| Air Blank | | 0.000 | 02:31 |
| Subject Sample | # 1 | 0.117 | 02:31 |
| Breath Volume | | 1.406 | 02:31 |
| Air Blank | | 0.000 | 02:32 |
| Air Blank | | 0.000 | 02:34 |
| Subject Sample | # 2 | 0.107 | 02:35 |
| Breath Volume | | 1.464 | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Control Test | | 0.083 | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Diagnostics Check | | OK | 02:37 |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006925

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PIERRE,RICHARDSON
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 05/31/2026 23:32:20

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:18 | |
| Air Blank | | 0.000 | 23:19 | |
| Control Test | | 0.080 | 23:19 | |
| Air Blank | | 0.000 | 23:20 | |
| Subject Sample | # 1 | 0.218 | 23:21 | |
| Breath Volume | | 1.187 | 23:21 | |
| Air Blank | | 0.000 | 23:22 | |
| Air Blank | | 0.000 | 23:24 | |
| Subject Sample | # 2 | 0.194 | 23:27 | Volume Not Met |
| Breath Volume | | 1.007 | 23:27 | |
| Air Blank | | 0.000 | 23:28 | |
| Air Blank | | 0.000 | 23:29 | |
| Subject Sample | # 3 | 0.223 | 23:30 | |
| Breath Volume | | 2.007 | 23:30 | |
| Air Blank | | 0.000 | 23:31 | |
| Control Test | | 0.078 | 23:31 | |
| Air Blank | | 0.000 | 23:32 | |
| Diagnostics Check | | OK | 23:32 | |

Data Download Date/Time: 06/10/2026 09:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/01/2026 02:32:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:26 |
| Air Blank | | 0.000 | 02:27 |
| Control Test | | 0.080 | 02:27 |
| Air Blank | | 0.000 | 02:27 |
| Subject Sample | # 1 | 0.174 | 02:28 |
| Breath Volume | | 1.371 | 02:28 |
| Air Blank | | 0.000 | 02:28 |
| Air Blank | | 0.000 | 02:30 |
| Subject Sample | # 2 | 0.194 | 02:30 |
| Breath Volume | | 1.277 | 02:30 |
| Air Blank | | 0.000 | 02:31 |
| Control Test | | 0.081 | 02:31 |
| Air Blank | | 0.000 | 02:32 |
| Diagnostics Check | | OK | 02:32 |

Data Download Date/Time: 05/21/2026 10:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MARQUES,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 05/02/2026 02:36:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:29 |
| Air Blank | | 0.000 | 02:29 |
| Control Test | | 0.082 | 02:30 |
| Air Blank | | 0.000 | 02:30 |
| Subject Sample | # 1 | 0.251 | 02:31 |
| Breath Volume | | 2.488 | 02:31 |
| Air Blank | | 0.000 | 02:32 |
| Air Blank | | 0.000 | 02:33 |
| Subject Sample | # 2 | 0.253 | 02:34 |
| Breath Volume | | 2.125 | 02:34 |
| Air Blank | | 0.000 | 02:35 |
| Control Test | | 0.081 | 02:35 |
| Air Blank | | 0.000 | 02:35 |
| Diagnostics Check | | OK | 02:36 |

Data Download Date/Time: 05/21/2026 10:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GLUESENKAMP

DATE/TIME OF TEST: 05/07/2026 03:51:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:46 |
| Air Blank | | 0.000 | 03:46 |
| Control Test | | 0.079 | 03:46 |
| Air Blank | | 0.000 | 03:47 |
| Subject Sample | # 1 | 0.000 | 03:47 |
| Breath Volume | | 1.785 | 03:47 |
| Air Blank | | 0.000 | 03:48 |
| Air Blank | | 0.000 | 03:50 |
| Subject Sample | # 2 | 0.000 | 03:50 |
| Breath Volume | | 1.210 | 03:50 |
| Air Blank | | 0.000 | 03:50 |
| Control Test | | 0.080 | 03:51 |
| Air Blank | | 0.000 | 03:51 |
| Diagnostics Check | | OK | 03:51 |

Data Download Date/Time: 05/21/2026 10:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CAMPBELL,EDWARD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/09/2026 01:29:39

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:21 |
| Air Blank | | 0.000 | 01:21 |
| Control Test | | 0.082 | 01:21 |
| Air Blank | | 0.000 | 01:22 |
| Subject Sample | # 1 | 0.228 | 01:23 |
| Breath Volume | | 1.535 | 01:23 |
| Air Blank | | 0.000 | 01:24 |
| Air Blank | | 0.000 | 01:26 |
| Subject Sample | # 2 | 0.226 | 01:27 |
| Breath Volume | | 1.585 | 01:27 |
| Air Blank | | 0.000 | 01:28 |
| Control Test | | 0.080 | 01:28 |
| Air Blank | | 0.000 | 01:29 |
| Diagnostics Check | | OK | 01:29 |

Data Download Date/Time: 05/21/2026 10:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006926

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HALL,NOAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: HERRON

DATE/TIME OF TEST: 05/09/2026 23:56:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:49 |
| Air Blank | | 0.000 | 23:49 |
| Control Test | | 0.083 | 23:50 |
| Air Blank | | 0.000 | 23:50 |
| Subject Sample | # 1 | 0.113 | 23:51 |
| Breath Volume | | 1.460 | 23:51 |
| Air Blank | | 0.000 | 23:52 |
| Air Blank | | 0.000 | 23:54 |
| Subject Sample | # 2 | 0.113 | 23:54 |
| Breath Volume | | 1.468 | 23:54 |
| Air Blank | | 0.000 | 23:55 |
| Control Test | | 0.081 | 23:56 |
| Air Blank | | 0.000 | 23:56 |
| Diagnostics Check | | OK | 23:56 |

Data Download Date/Time: 05/21/2026 10:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 05/04/2026 23:17:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:11 |
| Air Blank | | 0.000 | 23:11 |
| Control Test | | 0.078 | 23:11 |
| Air Blank | | 0.000 | 23:12 |
| Subject Sample | # 1 | 0.165 | 23:13 |
| Breath Volume | | 1.199 | 23:13 |
| Air Blank | | 0.000 | 23:13 |
| Air Blank | | 0.000 | 23:15 |
| Subject Sample | # 2 | 0.159 | 23:16 |
| Breath Volume | | 1.109 | 23:16 |
| Air Blank | | 0.000 | 23:16 |
| Control Test | | 0.077 | 23:17 |
| Air Blank | | 0.000 | 23:17 |
| Diagnostics Check | | OK | 23:17 |

Data Download Date/Time: 05/29/2026 22:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: HADLEY

DATE/TIME OF TEST: 05/07/2026 00:08:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:59 |
| Air Blank | | 0.000 | 00:00 |
| Control Test | | 0.078 | 00:00 |
| Air Blank | | 0.000 | 00:01 |
| Subject Sample | # 1 | 0.141 | 00:02 |
| Breath Volume | | 2.480 | 00:02 |
| Air Blank | | 0.000 | 00:03 |
| Air Blank | | 0.000 | 00:05 |
| Subject Sample | # 2 | 0.130 | 00:06 |
| Breath Volume | | 1.652 | 00:06 |
| Air Blank | | 0.000 | 00:07 |
| Control Test | | 0.078 | 00:07 |
| Air Blank | | 0.000 | 00:08 |
| Diagnostics Check | | OK | 00:08 |

Data Download Date/Time: 05/29/2026 22:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: ANABITARTE

DATE/TIME OF TEST: 05/07/2026 02:06:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:59 |
| Air Blank | | 0.000 | 02:00 |
| Control Test | | 0.079 | 02:00 |
| Air Blank | | 0.000 | 02:00 |
| Subject Sample | # 1 | 0.035 | 02:01 |
| Breath Volume | | 1.144 | 02:01 |
| Air Blank | | 0.000 | 02:02 |
| Air Blank | | 0.000 | 02:03 |
| Subject Sample | # 2 | 0.034 | 02:04 |
| Breath Volume | | 1.726 | 02:04 |
| Air Blank | | 0.000 | 02:05 |
| Control Test | | 0.079 | 02:05 |
| Air Blank | | 0.000 | 02:05 |
| Diagnostics Check | | OK | 02:06 |

Data Download Date/Time: 05/29/2026 22:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/15/2026 02:22:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:15 |
| Air Blank | | 0.000 | 02:16 |
| Control Test | | 0.079 | 02:16 |
| Air Blank | | 0.000 | 02:16 |
| Subject Sample | # 1 | 0.146 | 02:17 |
| Breath Volume | | 4.289 | 02:17 |
| Air Blank | | 0.000 | 02:18 |
| Air Blank | | 0.000 | 02:20 |
| Subject Sample | # 2 | 0.149 | 02:21 |
| Breath Volume | | 5.007 | 02:21 |
| Air Blank | | 0.000 | 02:21 |
| Control Test | | 0.078 | 02:22 |
| Air Blank | | 0.000 | 02:22 |
| Diagnostics Check | | OK | 02:22 |

Data Download Date/Time: 05/29/2026 22:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PALUMBO

DATE/TIME OF TEST: 05/23/2026 03:00:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:54 |
| Air Blank | | 0.000 | 02:54 |
| Control Test | | 0.079 | 02:55 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 1 | 0.169 | 02:55 |
| Breath Volume | | 1.570 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Air Blank | | 0.000 | 02:58 |
| Subject Sample | # 2 | 0.165 | 02:58 |
| Breath Volume | | 2.187 | 02:58 |
| Air Blank | | 0.000 | 02:59 |
| Control Test | | 0.078 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Diagnostics Check | | OK | 03:00 |

Data Download Date/Time: 05/29/2026 22:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,TREVOR,
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: SHEKHAWAT

DATE/TIME OF TEST: 05/28/2026 02:04:03

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:56 |
| Air Blank | | 0.000 | 01:57 |
| Control Test | | 0.077 | 01:57 |
| Air Blank | | 0.000 | 01:58 |
| Subject Sample | # 1 | 0.244 | 01:59 |
| Breath Volume | | 1.320 | 01:59 |
| Air Blank | | 0.000 | 01:59 |
| Air Blank | | 0.000 | 02:01 |
| Subject Sample | # 2 | 0.244 | 02:02 |
| Breath Volume | | 1.347 | 02:02 |
| Air Blank | | 0.000 | 02:03 |
| Control Test | | 0.078 | 02:03 |
| Air Blank | | 0.000 | 02:03 |
| Diagnostics Check | | OK | 02:04 |

Data Download Date/Time: 05/29/2026 22:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: COAKLEY

DATE/TIME OF TEST: 05/29/2026 01:32:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:25 |
| Air Blank | | 0.000 | 01:25 |
| Control Test | | 0.078 | 01:26 |
| Air Blank | | 0.000 | 01:26 |
| Subject Sample | # 1 | 0.286 | 01:27 |
| Breath Volume | | 1.910 | 01:27 |
| Air Blank | | 0.000 | 01:28 |
| Air Blank | | 0.000 | 01:30 |
| Subject Sample | # 2 | 0.272 | 01:31 |
| Breath Volume | | 1.503 | 01:31 |
| Air Blank | | 0.000 | 01:31 |
| Control Test | | 0.078 | 01:32 |
| Air Blank | | 0.000 | 01:32 |
| Diagnostics Check | | OK | 01:32 |

Data Download Date/Time: 05/29/2026 22:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WHITFIEL JR

DATE/TIME OF TEST: 05/29/2026 04:23:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:16 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.078 | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Subject Sample | # 1 | 0.191 | 04:18 |
| Breath Volume | | 2.953 | 04:18 |
| Air Blank | | 0.000 | 04:19 |
| Air Blank | | 0.000 | 04:21 |
| Subject Sample | # 2 | 0.189 | 04:21 |
| Breath Volume | | 2.640 | 04:21 |
| Air Blank | | 0.000 | 04:22 |
| Control Test | | 0.077 | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Diagnostics Check | | OK | 04:23 |

Data Download Date/Time: 05/29/2026 22:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: CLINTON

DATE/TIME OF TEST: 05/31/2026 00:51:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:44 |
| Air Blank | | 0.000 | 00:44 |
| Control Test | | 0.078 | 00:45 |
| Air Blank | | 0.000 | 00:45 |
| Subject Sample | # 1 | 0.190 | 00:46 |
| Breath Volume | | 2.566 | 00:46 |
| Air Blank | | 0.000 | 00:47 |
| Air Blank | | 0.000 | 00:48 |
| Subject Sample | # 2 | 0.187 | 00:49 |
| Breath Volume | | 1.960 | 00:49 |
| Air Blank | | 0.000 | 00:50 |
| Control Test | | 0.078 | 00:50 |
| Air Blank | | 0.000 | 00:51 |
| Diagnostics Check | | OK | 00:51 |

Data Download Date/Time: 06/27/2026 21:29

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007007

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONER,NATHAN,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: STONER,NATHAN
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: STRARODUB

DATE/TIME OF TEST: 05/31/2026 23:43:59

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:37 | |
| Air Blank | | 0.000 | 23:38 | |
| Control Test | | 0.079 | 23:38 | |
| Air Blank | | 0.000 | 23:39 | |
| Subject Sample | # 1 | | 23:41 | Subject Test Refused |
| Breath Volume | | 0.035 | 23:41 | |
| Air Blank | | 0.000 | 23:42 | |
| Control Test | | 0.079 | 23:42 | |
| Air Blank | | 0.000 | 23:43 | |
| Diagnostics Check | | Fail | 23:43 | Internal Printer Test |
| Air Blank | | 0.000 | 23:43 | |

Data Download Date/Time: 06/27/2026 21:29

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,C,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ENRRIQUEZ,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: GIRONPEREZ

DATE/TIME OF TEST: 05/03/2026 22:04:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:58 |
| Air Blank | | 0.000 | 21:58 |
| Control Test | | 0.080 | 21:58 |
| Air Blank | | 0.000 | 21:59 |
| Subject Sample | # 1 | 0.238 | 21:59 |
| Breath Volume | | 2.003 | 21:59 |
| Air Blank | | 0.000 | 22:00 |
| Air Blank | | 0.000 | 22:02 |
| Subject Sample | # 2 | 0.241 | 22:02 |
| Breath Volume | | 1.812 | 22:02 |
| Air Blank | | 0.000 | 22:03 |
| Control Test | | 0.079 | 22:03 |
| Air Blank | | 0.000 | 22:04 |
| Diagnostics Check | | OK | 22:04 |

Data Download Date/Time: 05/30/2026 14:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EMILIUS,ETHIN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: EMILIUS,ETHIN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: JEAN BAPTISTE

DATE/TIME OF TEST: 05/07/2026 04:00:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:50 |
| Air Blank | | 0.000 | 03:50 |
| Control Test | | 0.081 | 03:51 |
| Air Blank | | 0.000 | 03:51 |
| Subject Sample | # 1 | 0.173 | 03:52 |
| Breath Volume | | 1.632 | 03:52 |
| Air Blank | | 0.000 | 03:53 |
| Air Blank | | 0.000 | 03:55 |
| Subject Sample | # 2 | 0.186 | 03:58 |
| Breath Volume | | 2.062 | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Control Test | | 0.081 | 03:59 |
| Air Blank | | 0.000 | 04:00 |
| Diagnostics Check | | OK | 04:00 |

Data Download Date/Time: 05/30/2026 14:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARIMON,KARLHEINZ,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARMONA,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: PEDEMONTE

DATE/TIME OF TEST: 05/09/2026 06:29:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:22 |
| Air Blank | | 0.000 | 06:23 |
| Control Test | | 0.081 | 06:23 |
| Air Blank | | 0.000 | 06:24 |
| Subject Sample | # 1 | 0.165 | 06:24 |
| Breath Volume | | 2.316 | 06:24 |
| Air Blank | | 0.000 | 06:25 |
| Air Blank | | 0.000 | 06:27 |
| Subject Sample | # 2 | 0.158 | 06:27 |
| Breath Volume | | 1.703 | 06:27 |
| Air Blank | | 0.000 | 06:28 |
| Control Test | | 0.080 | 06:28 |
| Air Blank | | 0.000 | 06:29 |
| Diagnostics Check | | OK | 06:29 |

Data Download Date/Time: 05/30/2026 14:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,GATLIN,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,TINA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: LINGELBACH

DATE/TIME OF TEST: 05/10/2026 20:02:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:56 |
| Air Blank | | 0.000 | 19:56 |
| Control Test | | 0.080 | 19:56 |
| Air Blank | | 0.000 | 19:57 |
| Subject Sample | # 1 | 0.186 | 19:57 |
| Breath Volume | | 1.707 | 19:57 |
| Air Blank | | 0.000 | 19:58 |
| Air Blank | | 0.000 | 20:00 |
| Subject Sample | # 2 | 0.189 | 20:01 |
| Breath Volume | | 1.847 | 20:01 |
| Air Blank | | 0.000 | 20:01 |
| Control Test | | 0.081 | 20:02 |
| Air Blank | | 0.000 | 20:02 |
| Diagnostics Check | | OK | 20:02 |

Data Download Date/Time: 05/30/2026 14:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYMAN,PATRICK,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAIL,MATTHEW
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 05/15/2026 01:44:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:38 |
| Air Blank | | 0.000 | 01:38 |
| Control Test | | 0.081 | 01:38 |
| Air Blank | | 0.000 | 01:39 |
| Subject Sample | # 1 | 0.117 | 01:40 |
| Breath Volume | | 2.601 | 01:40 |
| Air Blank | | 0.000 | 01:40 |
| Air Blank | | 0.000 | 01:42 |
| Subject Sample | # 2 | 0.121 | 01:43 |
| Breath Volume | | 2.832 | 01:43 |
| Air Blank | | 0.000 | 01:43 |
| Control Test | | 0.080 | 01:44 |
| Air Blank | | 0.000 | 01:44 |
| Diagnostics Check | | OK | 01:44 |

Data Download Date/Time: 05/30/2026 14:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLELLAND,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GUZMAN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 05/26/2026 00:14:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:07 |
| Air Blank | | 0.000 | 00:08 |
| Control Test | | 0.081 | 00:08 |
| Air Blank | | 0.000 | 00:09 |
| Subject Sample | # 1 | 0.176 | 00:09 |
| Breath Volume | | 4.085 | 00:09 |
| Air Blank | | 0.000 | 00:10 |
| Air Blank | | 0.000 | 00:12 |
| Subject Sample | # 2 | 0.177 | 00:12 |
| Breath Volume | | 3.839 | 00:12 |
| Air Blank | | 0.000 | 00:13 |
| Control Test | | 0.080 | 00:13 |
| Air Blank | | 0.000 | 00:14 |
| Diagnostics Check | | OK | 00:14 |

Data Download Date/Time: 05/30/2026 14:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,MEGAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAIL,MATTHEW
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 05/28/2026 02:57:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:50 |
| Air Blank | | 0.000 | 02:50 |
| Control Test | | 0.081 | 02:51 |
| Air Blank | | 0.000 | 02:51 |
| Subject Sample | # 1 | 0.100 | 02:52 |
| Breath Volume | | 2.835 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 2 | 0.097 | 02:55 |
| Breath Volume | | 1.667 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Control Test | | 0.080 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Diagnostics Check | | OK | 02:57 |

Data Download Date/Time: 05/30/2026 14:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PALMER,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 05/01/2026 03:40:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:33 |
| Air Blank | | 0.000 | 03:33 |
| Control Test | | 0.080 | 03:34 |
| Air Blank | | 0.000 | 03:34 |
| Subject Sample | # 1 | 0.200 | 03:35 |
| Breath Volume | | 3.160 | 03:35 |
| Air Blank | | 0.000 | 03:36 |
| Air Blank | | 0.000 | 03:38 |
| Subject Sample | # 2 | 0.187 | 03:39 |
| Breath Volume | | 3.277 | 03:39 |
| Air Blank | | 0.000 | 03:39 |
| Control Test | | 0.079 | 03:40 |
| Air Blank | | 0.000 | 03:40 |
| Diagnostics Check | | OK | 03:40 |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNN,CDFC,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JOHNSON,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 10:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:58

SUBJECT'S LAST NAME: SHEEN

DATE/TIME OF TEST: 05/01/2026 11:36:11

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 11:19 | |
| Air Blank | | 0.000 | 11:19 | |
| Control Test | | 0.080 | 11:19 | |
| Air Blank | | 0.000 | 11:20 | |
| Subject Sample | # 1 | 0.241 | 11:23 | |
| Breath Volume | | 2.015 | 11:23 | |
| Air Blank | | 0.000 | 11:24 | |
| Air Blank | | 0.000 | 11:25 | |
| Subject Sample | # 2 | 0.253 | 11:29 | Volume Not Met |
| Breath Volume | | 0.031 | 11:29 | |
| Air Blank | | 0.000 | 11:29 | |
| Air Blank | | 0.000 | 11:31 | |
| Subject Sample | # 3 | | 11:34 | No Sample Provided |
| Breath Volume | | 0.000 | 11:34 | |
| Air Blank | | 0.000 | 11:35 | |
| Control Test | | 0.076 | 11:35 | |
| Air Blank | | 0.000 | 11:36 | |
| Diagnostics Check | | OK | 11:36 | |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LETT,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/01/2026 17:28:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 17:22 |
| Air Blank | | 0.000 | 17:22 |
| Control Test | | 0.080 | 17:22 |
| Air Blank | | 0.000 | 17:23 |
| Subject Sample | # 1 | 0.165 | 17:23 |
| Breath Volume | | 2.546 | 17:23 |
| Air Blank | | 0.000 | 17:24 |
| Air Blank | | 0.000 | 17:26 |
| Subject Sample | # 2 | 0.164 | 17:26 |
| Breath Volume | | 3.519 | 17:26 |
| Air Blank | | 0.000 | 17:27 |
| Control Test | | 0.080 | 17:27 |
| Air Blank | | 0.000 | 17:28 |
| Diagnostics Check | | OK | 17:28 |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 05/02/2026 01:38:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:31 |
| Air Blank | | 0.000 | 01:32 |
| Control Test | | 0.080 | 01:32 |
| Air Blank | | 0.000 | 01:33 |
| Subject Sample | # 1 | 0.083 | 01:33 |
| Breath Volume | | 2.480 | 01:33 |
| Air Blank | | 0.000 | 01:34 |
| Air Blank | | 0.000 | 01:36 |
| Subject Sample | # 2 | 0.078 | 01:36 |
| Breath Volume | | 1.593 | 01:36 |
| Air Blank | | 0.000 | 01:37 |
| Control Test | | 0.079 | 01:37 |
| Air Blank | | 0.000 | 01:38 |
| Diagnostics Check | | OK | 01:38 |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRADY,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: VAN-BODEGON

DATE/TIME OF TEST: 05/02/2026 01:54:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:47 |
| Air Blank | | 0.000 | 01:48 |
| Control Test | | 0.080 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Subject Sample | # 1 | 0.128 | 01:49 |
| Breath Volume | | 2.250 | 01:49 |
| Air Blank | | 0.000 | 01:50 |
| Air Blank | | 0.000 | 01:52 |
| Subject Sample | # 2 | 0.129 | 01:52 |
| Breath Volume | | 3.078 | 01:52 |
| Air Blank | | 0.000 | 01:53 |
| Control Test | | 0.080 | 01:53 |
| Air Blank | | 0.000 | 01:54 |
| Diagnostics Check | | OK | 01:54 |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: STORIE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 05/02/2026 23:38:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:31 |
| Air Blank | | 0.000 | 23:32 |
| Control Test | | 0.080 | 23:32 |
| Air Blank | | 0.000 | 23:33 |
| Subject Sample | # 1 | 0.187 | 23:34 |
| Breath Volume | | 1.761 | 23:34 |
| Air Blank | | 0.000 | 23:34 |
| Air Blank | | 0.000 | 23:36 |
| Subject Sample | # 2 | 0.189 | 23:37 |
| Breath Volume | | 1.617 | 23:37 |
| Air Blank | | 0.000 | 23:38 |
| Control Test | | 0.078 | 23:38 |
| Air Blank | | 0.000 | 23:38 |
| Diagnostics Check | | OK | 23:38 |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ALVAREZ,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 05/03/2026 20:00:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:52 |
| Air Blank | | 0.000 | 19:53 |
| Control Test | | 0.080 | 19:53 |
| Air Blank | | 0.000 | 19:54 |
| Subject Sample | # 1 | 0.149 | 19:55 |
| Breath Volume | | 1.214 | 19:55 |
| Air Blank | | 0.000 | 19:56 |
| Air Blank | | 0.000 | 19:57 |
| Subject Sample | # 2 | 0.131 | 19:58 |
| Breath Volume | | 1.722 | 19:58 |
| Air Blank | | 0.000 | 19:59 |
| Control Test | | 0.079 | 19:59 |
| Air Blank | | 0.000 | 20:00 |
| Diagnostics Check | | OK | 20:00 |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWOCHO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KEITER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/07/2026 16:59:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:52 |
| Air Blank | | 0.000 | 16:52 |
| Control Test | | 0.080 | 16:53 |
| Air Blank | | 0.000 | 16:53 |
| Subject Sample | # 1 | 0.000 | 16:54 |
| Breath Volume | | 3.269 | 16:54 |
| Air Blank | | 0.000 | 16:54 |
| Air Blank | | 0.000 | 16:56 |
| Subject Sample | # 2 | 0.000 | 16:57 |
| Breath Volume | | 3.070 | 16:57 |
| Air Blank | | 0.000 | 16:58 |
| Control Test | | 0.080 | 16:58 |
| Air Blank | | 0.000 | 16:59 |
| Diagnostics Check | | OK | 16:59 |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WEIHS,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BEAVER

DATE/TIME OF TEST: 05/09/2026 00:21:23

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:13 |
| Air Blank | | 0.000 | 00:14 |
| Control Test | | 0.080 | 00:14 |
| Air Blank | | 0.000 | 00:15 |
| Subject Sample | # 1 | 0.125 | 00:16 |
| Breath Volume | | 1.683 | 00:16 |
| Air Blank | | 0.000 | 00:17 |
| Air Blank | | 0.000 | 00:19 |
| Subject Sample | # 2 | 0.124 | 00:19 |
| Breath Volume | | 1.703 | 00:19 |
| Air Blank | | 0.000 | 00:20 |
| Control Test | | 0.079 | 00:20 |
| Air Blank | | 0.000 | 00:21 |
| Diagnostics Check | | OK | 00:21 |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007083

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWTHORNE,OLIVIA,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WADLINGTON,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A2
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 05/09/2026 00:48:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:42 |
| Air Blank | | 0.000 | 00:42 |
| Control Test | | 0.079 | 00:43 |
| Air Blank | | 0.000 | 00:43 |
| Subject Sample | # 1 | 0.083 | 00:44 |
| Breath Volume | | 1.562 | 00:44 |
| Air Blank | | 0.000 | 00:44 |
| Air Blank | | 0.000 | 00:46 |
| Subject Sample | # 2 | 0.081 | 00:47 |
| Breath Volume | | 1.613 | 00:47 |
| Air Blank | | 0.000 | 00:47 |
| Control Test | | 0.079 | 00:47 |
| Air Blank | | 0.000 | 00:48 |
| Diagnostics Check | | OK | 00:48 |

Data Download Date/Time: 05/19/2026 16:07

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,CRISTIAN,X
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: STRINGER,A
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: REYNOSO

DATE/TIME OF TEST: 05/19/2026 22:10:06

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:01 | |
| Air Blank | | 0.000 | 22:01 | |
| Control Test | | 0.080 | 22:01 | |
| Air Blank | | 0.000 | 22:02 | |
| Subject Sample | # 1 | | 22:05 | No Sample Provided |
| Breath Volume | | 0.000 | 22:05 | |
| Air Blank | | 0.000 | 22:06 | |
| Air Blank | | 0.000 | 22:08 | |
| Subject Sample | # 2 | | 22:08 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:08 | |
| Air Blank | | 0.000 | 22:08 | |
| Control Test | | 0.080 | 22:08 | |
| Air Blank | | 0.000 | 22:09 | |
| Diagnostics Check | | Fail | 22:09 | Internal Printer Test |
| Air Blank | | 0.000 | 22:10 | |

Data Download Date/Time: 06/04/2026 01:58

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ROBLEDO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: CORDOBA

DATE/TIME OF TEST: 05/22/2026 16:49:02

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 16:37 | |
| Air Blank | | 0.000 | 16:38 | |
| Control Test | | 0.080 | 16:38 | |
| Air Blank | | 0.000 | 16:38 | |
| Subject Sample | # 1 | | 16:42 | No Sample Provided |
| Breath Volume | | 0.000 | 16:42 | |
| Air Blank | | 0.000 | 16:42 | |
| Air Blank | | 0.000 | 16:44 | |
| Subject Sample | # 2 | | 16:47 | No Sample Provided |
| Breath Volume | | 0.000 | 16:47 | |
| Air Blank | | 0.000 | 16:48 | |
| Control Test | | 0.080 | 16:48 | |
| Air Blank | | 0.000 | 16:48 | |
| Diagnostics Check | | OK | 16:49 | |

Data Download Date/Time: 06/04/2026 01:58

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ROBLEDO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: CORDOBA

DATE/TIME OF TEST: 05/22/2026 16:59:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:50 |
| Air Blank | | 0.000 | 16:51 |
| Control Test | | 0.080 | 16:51 |
| Air Blank | | 0.000 | 16:52 |
| Subject Sample | # 1 | 0.176 | 16:54 |
| Breath Volume | | 1.937 | 16:54 |
| Air Blank | | 0.000 | 16:55 |
| Air Blank | | 0.000 | 16:56 |
| Subject Sample | # 2 | 0.171 | 16:58 |
| Breath Volume | | 1.402 | 16:58 |
| Air Blank | | 0.000 | 16:58 |
| Control Test | | 0.077 | 16:59 |
| Air Blank | | 0.000 | 16:59 |
| Diagnostics Check | | OK | 16:59 |

Data Download Date/Time: 06/04/2026 01:58

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007084

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OROL,MANUEL,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: OROL,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: RODRIGUEZ CRUZ

DATE/TIME OF TEST: 05/24/2026 06:23:45

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 06:19 | |
| Air Blank | | 0.000 | 06:19 | |
| Control Test | | 0.080 | 06:20 | |
| Air Blank | | 0.000 | 06:20 | |
| Subject Sample | # 1 | | 06:22 | Subject Test Refused |
| Breath Volume | | 0.000 | 06:22 | |
| Air Blank | | 0.000 | 06:22 | |
| Control Test | | 0.080 | 06:23 | |
| Air Blank | | 0.000 | 06:23 | |
| Diagnostics Check | | OK | 06:23 | |

Data Download Date/Time: 06/04/2026 01:58

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DE LA CAMPA,ERICK
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: MONDESTIN

DATE/TIME OF TEST: 05/03/2026 03:07:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:54 | |
| Air Blank | | 0.000 | 02:54 | |
| Control Test | | 0.079 | 02:55 | |
| Air Blank | | 0.000 | 02:55 | |
| Subject Sample | # 1 | 0.198 | 02:56 | |
| Breath Volume | | 1.343 | 02:56 | |
| Air Blank | | 0.000 | 02:57 | |
| Air Blank | | 0.000 | 02:59 | |
| Subject Sample | # 2 | 0.185 | 03:02 | Volume Not Met |
| Breath Volume | | 0.144 | 03:02 | |
| Air Blank | | 0.000 | 03:03 | |
| Air Blank | | 0.000 | 03:04 | |
| Subject Sample | # 3 | 0.188 | 03:06 | |
| Breath Volume | | 1.582 | 03:06 | |
| Air Blank | | 0.000 | 03:06 | |
| Control Test | | 0.077 | 03:07 | |
| Air Blank | | 0.000 | 03:07 | |
| Diagnostics Check | | OK | 03:07 | |

Data Download Date/Time: 05/06/2026 07:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LAGE

DATE/TIME OF TEST: 05/03/2026 04:05:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Control Test | | 0.078 | 03:59 |
| Air Blank | | 0.000 | 04:00 |
| Subject Sample | # 1 | 0.210 | 04:00 |
| Breath Volume | | 2.164 | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Air Blank | | 0.000 | 04:03 |
| Subject Sample | # 2 | 0.208 | 04:03 |
| Breath Volume | | 3.183 | 04:03 |
| Air Blank | | 0.000 | 04:04 |
| Control Test | | 0.078 | 04:04 |
| Air Blank | | 0.000 | 04:05 |
| Diagnostics Check | | OK | 04:05 |

Data Download Date/Time: 05/06/2026 07:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: KRAMERS

DATE/TIME OF TEST: 05/03/2026 23:03:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:56 |
| Air Blank | | 0.000 | 22:57 |
| Control Test | | 0.079 | 22:57 |
| Air Blank | | 0.000 | 22:58 |
| Subject Sample | # 1 | 0.169 | 22:58 |
| Breath Volume | | 3.539 | 22:58 |
| Air Blank | | 0.000 | 22:59 |
| Air Blank | | 0.000 | 23:01 |
| Subject Sample | # 2 | 0.166 | 23:01 |
| Breath Volume | | 1.812 | 23:01 |
| Air Blank | | 0.000 | 23:02 |
| Control Test | | 0.078 | 23:02 |
| Air Blank | | 0.000 | 23:03 |
| Diagnostics Check | | OK | 23:03 |

Data Download Date/Time: 05/06/2026 07:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 05/04/2026 01:37:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:30 |
| Air Blank | | 0.000 | 01:31 |
| Control Test | | 0.080 | 01:31 |
| Air Blank | | 0.000 | 01:32 |
| Subject Sample | # 1 | 0.164 | 01:32 |
| Breath Volume | | 2.058 | 01:32 |
| Air Blank | | 0.000 | 01:33 |
| Air Blank | | 0.000 | 01:35 |
| Subject Sample | # 2 | 0.161 | 01:36 |
| Breath Volume | | 1.519 | 01:36 |
| Air Blank | | 0.000 | 01:36 |
| Control Test | | 0.079 | 01:37 |
| Air Blank | | 0.000 | 01:37 |
| Diagnostics Check | | OK | 01:37 |

Data Download Date/Time: 05/06/2026 07:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PIERRE,RICHARDSON
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: DUREUS

DATE/TIME OF TEST: 05/05/2026 22:09:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:02 |
| Air Blank | | 0.000 | 22:02 |
| Control Test | | 0.078 | 22:02 |
| Air Blank | | 0.000 | 22:03 |
| Subject Sample | # 1 | 0.340 | 22:04 |
| Breath Volume | | 1.285 | 22:04 |
| Air Blank | | 0.000 | 22:05 |
| Air Blank | | 0.000 | 22:07 |
| Subject Sample | # 2 | 0.347 | 22:07 |
| Breath Volume | | 1.507 | 22:07 |
| Air Blank | | 0.000 | 22:08 |
| Control Test | | 0.077 | 22:08 |
| Air Blank | | 0.000 | 22:09 |
| Diagnostics Check | | OK | 22:09 |

Data Download Date/Time: 05/06/2026 07:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007107

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERDINAND,GAETANE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RAPALO

DATE/TIME OF TEST: 05/06/2026 02:28:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:20 |
| Air Blank | | 0.000 | 02:21 |
| Control Test | | 0.079 | 02:21 |
| Air Blank | | 0.000 | 02:21 |
| Subject Sample | # 1 | 0.121 | 02:22 |
| Breath Volume | | 1.960 | 02:22 |
| Air Blank | | 0.000 | 02:23 |
| Air Blank | | 0.000 | 02:25 |
| Subject Sample | # 2 | 0.122 | 02:26 |
| Breath Volume | | 1.492 | 02:26 |
| Air Blank | | 0.000 | 02:27 |
| Control Test | | 0.080 | 02:27 |
| Air Blank | | 0.000 | 02:28 |
| Diagnostics Check | | OK | 02:28 |

Data Download Date/Time: 05/06/2026 07:06

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZKAU,AARON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PRITZKAU,AARON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TELLER

DATE/TIME OF TEST: 05/01/2026 02:11:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:05 |
| Air Blank | | 0.000 | 02:05 |
| Control Test | | 0.081 | 02:06 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 1 | 0.134 | 02:07 |
| Breath Volume | | 2.742 | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Air Blank | | 0.000 | 02:09 |
| Subject Sample | # 2 | 0.133 | 02:10 |
| Breath Volume | | 3.339 | 02:10 |
| Air Blank | | 0.000 | 02:10 |
| Control Test | | 0.079 | 02:11 |
| Air Blank | | 0.000 | 02:11 |
| Diagnostics Check | | OK | 02:11 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,NOAH,I
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: JARA,JONATHAN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: RAMIREZ DOMINGO

DATE/TIME OF TEST: 05/02/2026 17:46:15

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 17:34 | |
| Air Blank | | 0.000 | 17:34 | |
| Control Test | | 0.081 | 17:34 | |
| Air Blank | | 0.000 | 17:35 | |
| Subject Sample | # 1 | 0.253 | 17:38 | Volume Not Met |
| Breath Volume | | 0.976 | 17:38 | |
| Air Blank | | 0.000 | 17:39 | |
| Air Blank | | 0.000 | 17:41 | |
| Subject Sample | # 2 | 0.256 | 17:41 | |
| Breath Volume | | 1.132 | 17:41 | |
| Air Blank | | 0.000 | 17:42 | |
| Air Blank | | 0.000 | 17:44 | |
| Subject Sample | # 3 | 0.251 | 17:44 | |
| Breath Volume | | 1.171 | 17:44 | |
| Air Blank | | 0.000 | 17:45 | |
| Control Test | | 0.076 | 17:45 | |
| Air Blank | | 0.000 | 17:46 | |
| Diagnostics Check | | OK | 17:46 | |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZKAU,AARON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PETERSEN,CODY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FRADETTE

DATE/TIME OF TEST: 05/05/2026 01:57:56

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Control Test | | 0.080 | 01:51 |
| Air Blank | | 0.000 | 01:52 |
| Subject Sample | # 1 | 0.151 | 01:53 |
| Breath Volume | | 2.183 | 01:53 |
| Air Blank | | 0.000 | 01:54 |
| Air Blank | | 0.000 | 01:55 |
| Subject Sample | # 2 | 0.146 | 01:56 |
| Breath Volume | | 2.414 | 01:56 |
| Air Blank | | 0.000 | 01:57 |
| Control Test | | 0.077 | 01:57 |
| Air Blank | | 0.000 | 01:57 |
| Diagnostics Check | | OK | 01:57 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZKAU,AARON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WRIGHT,DUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LACK

DATE/TIME OF TEST: 05/05/2026 22:56:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:49 |
| Air Blank | | 0.000 | 22:49 |
| Control Test | | 0.080 | 22:49 |
| Air Blank | | 0.000 | 22:50 |
| Subject Sample | # 1 | 0.133 | 22:51 |
| Breath Volume | | 3.281 | 22:51 |
| Air Blank | | 0.000 | 22:52 |
| Air Blank | | 0.000 | 22:54 |
| Subject Sample | # 2 | 0.132 | 22:54 |
| Breath Volume | | 1.359 | 22:54 |
| Air Blank | | 0.000 | 22:55 |
| Control Test | | 0.077 | 22:55 |
| Air Blank | | 0.000 | 22:56 |
| Diagnostics Check | | OK | 22:56 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZKAU,AARON,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PRITZKAU,AARON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MCPHERSON

DATE/TIME OF TEST: 05/06/2026 02:39:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:33 |
| Air Blank | | 0.000 | 02:33 |
| Control Test | | 0.080 | 02:34 |
| Air Blank | | 0.000 | 02:34 |
| Subject Sample | # 1 | 0.182 | 02:35 |
| Breath Volume | | 1.378 | 02:35 |
| Air Blank | | 0.000 | 02:35 |
| Air Blank | | 0.000 | 02:37 |
| Subject Sample | # 2 | 0.178 | 02:38 |
| Breath Volume | | 1.613 | 02:38 |
| Air Blank | | 0.000 | 02:39 |
| Control Test | | 0.080 | 02:39 |
| Air Blank | | 0.000 | 02:39 |
| Diagnostics Check | | OK | 02:39 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: THOMPSON,GREGORY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/07/2026 20:36:58

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:30 |
| Air Blank | | 0.000 | 20:30 |
| Control Test | | 0.081 | 20:31 |
| Air Blank | | 0.000 | 20:31 |
| Subject Sample | # 1 | 0.057 | 20:32 |
| Breath Volume | | 2.945 | 20:32 |
| Air Blank | | 0.000 | 20:32 |
| Air Blank | | 0.000 | 20:34 |
| Subject Sample | # 2 | 0.061 | 20:35 |
| Breath Volume | | 2.882 | 20:35 |
| Air Blank | | 0.000 | 20:36 |
| Control Test | | 0.081 | 20:36 |
| Air Blank | | 0.000 | 20:36 |
| Diagnostics Check | | OK | 20:36 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CUNNINGHAM,BENJAMIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 05/09/2026 02:36:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:30 |
| Air Blank | | 0.000 | 02:30 |
| Control Test | | 0.081 | 02:30 |
| Air Blank | | 0.000 | 02:31 |
| Subject Sample | # 1 | 0.180 | 02:32 |
| Breath Volume | | 2.414 | 02:32 |
| Air Blank | | 0.000 | 02:32 |
| Air Blank | | 0.000 | 02:34 |
| Subject Sample | # 2 | 0.177 | 02:35 |
| Breath Volume | | 2.125 | 02:35 |
| Air Blank | | 0.000 | 02:35 |
| Control Test | | 0.079 | 02:36 |
| Air Blank | | 0.000 | 02:36 |
| Diagnostics Check | | OK | 02:36 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CULPAN,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SESSOMS

DATE/TIME OF TEST: 05/10/2026 00:12:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:04 |
| Air Blank | | 0.000 | 00:05 |
| Control Test | | 0.081 | 00:05 |
| Air Blank | | 0.000 | 00:06 |
| Subject Sample | # 1 | 0.280 | 00:07 |
| Breath Volume | | 1.515 | 00:07 |
| Air Blank | | 0.000 | 00:08 |
| Air Blank | | 0.000 | 00:09 |
| Subject Sample | # 2 | 0.271 | 00:10 |
| Breath Volume | | 1.316 | 00:10 |
| Air Blank | | 0.000 | 00:11 |
| Control Test | | 0.078 | 00:11 |
| Air Blank | | 0.000 | 00:12 |
| Diagnostics Check | | OK | 00:12 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTNER,STEVEN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTNER,STEVEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:14

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/10/2026 06:41:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:34 |
| Air Blank | | 0.000 | 06:35 |
| Control Test | | 0.079 | 06:35 |
| Air Blank | | 0.000 | 06:35 |
| Subject Sample | # 1 | 0.157 | 06:36 |
| Breath Volume | | 1.128 | 06:36 |
| Air Blank | | 0.000 | 06:37 |
| Air Blank | | 0.000 | 06:39 |
| Subject Sample | # 2 | 0.155 | 06:39 |
| Breath Volume | | 1.371 | 06:39 |
| Air Blank | | 0.000 | 06:40 |
| Control Test | | 0.077 | 06:40 |
| Air Blank | | 0.000 | 06:41 |
| Diagnostics Check | | OK | 06:41 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CUNNINGHAM,BENJAMIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BERRIAN

DATE/TIME OF TEST: 05/14/2026 02:55:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:52 | |
| Air Blank | | 0.000 | 02:52 | |
| Control Test | | 0.081 | 02:52 | |
| Air Blank | | 0.000 | 02:53 | |
| Subject Sample | # 1 | 0.141 | 02:54 | |
| Breath Volume | | 1.718 | 02:54 | |
| Air Blank | | | 02:54 | RFI Detect |
| Air Blank | | 0.000 | 02:54 | |
| Control Test | | 0.080 | 02:55 | |
| Air Blank | | 0.000 | 02:55 | |
| Diagnostics Check | | OK | 02:55 | |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CUNNINGHAM,BENJAMIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BERRIAN

DATE/TIME OF TEST: 05/14/2026 03:05:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:58 |
| Air Blank | | 0.000 | 02:59 |
| Control Test | | 0.080 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Subject Sample | # 1 | 0.138 | 03:01 |
| Breath Volume | | 1.910 | 03:01 |
| Air Blank | | 0.000 | 03:01 |
| Air Blank | | 0.000 | 03:03 |
| Subject Sample | # 2 | 0.140 | 03:04 |
| Breath Volume | | 2.007 | 03:04 |
| Air Blank | | 0.000 | 03:05 |
| Control Test | | 0.080 | 03:05 |
| Air Blank | | 0.000 | 03:05 |
| Diagnostics Check | | OK | 03:05 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLEY,RASHELL,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BAGLEY,ANDREW
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: GUITERREZ CUARTAS

DATE/TIME OF TEST: 05/16/2026 22:10:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:03 |
| Air Blank | | 0.000 | 22:04 |
| Control Test | | 0.081 | 22:04 |
| Air Blank | | 0.000 | 22:04 |
| Subject Sample | # 1 | 0.000 | 22:05 |
| Breath Volume | | 2.503 | 22:05 |
| Air Blank | | 0.000 | 22:06 |
| Air Blank | | 0.000 | 22:08 |
| Subject Sample | # 2 | 0.000 | 22:09 |
| Breath Volume | | 2.769 | 22:09 |
| Air Blank | | 0.000 | 22:09 |
| Control Test | | 0.078 | 22:09 |
| Air Blank | | 0.000 | 22:10 |
| Diagnostics Check | | OK | 22:10 |

Data Download Date/Time: 05/18/2026 09:52

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CUNNINGHAM,BENJAMIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BOURNE

DATE/TIME OF TEST: 05/19/2026 04:36:07

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:33 | |
| Air Blank | | 0.000 | 04:34 | |
| Control Test | | 0.081 | 04:34 | |
| Air Blank | | 0.000 | 04:35 | |
| Subject Sample | # 1 | | 04:35 | RFI Detect |
| Breath Volume | | 0.894 | 04:35 | |
| Air Blank | | | 04:35 | RFI Detect |
| Air Blank | | 0.000 | 04:36 | |

Data Download Date/Time: 06/29/2026 10:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CUNNINGHAM,BENJAMIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: BOURNE

DATE/TIME OF TEST: 05/19/2026 04:44:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:38 |
| Air Blank | | 0.000 | 04:38 |
| Control Test | | 0.081 | 04:38 |
| Air Blank | | 0.000 | 04:39 |
| Subject Sample | # 1 | 0.208 | 04:39 |
| Breath Volume | | 3.839 | 04:39 |
| Air Blank | | 0.000 | 04:40 |
| Air Blank | | 0.000 | 04:42 |
| Subject Sample | # 2 | 0.199 | 04:42 |
| Breath Volume | | 3.562 | 04:42 |
| Air Blank | | 0.000 | 04:43 |
| Control Test | | 0.080 | 04:43 |
| Air Blank | | 0.000 | 04:44 |
| Diagnostics Check | | OK | 04:44 |

Data Download Date/Time: 06/29/2026 10:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTNER,STEVEN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILLIAMS,EARL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: SHOWS

DATE/TIME OF TEST: 05/24/2026 17:10:00

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 16:56 | |
| Air Blank | | 0.000 | 16:57 | |
| Control Test | | 0.080 | 16:57 | |
| Air Blank | | 0.000 | 16:57 | |
| Subject Sample | # 1 | 0.080 | 17:00 | Volume Not Met |
| Breath Volume | | 0.156 | 17:00 | |
| Air Blank | | 0.000 | 17:01 | |
| Air Blank | | 0.000 | 17:03 | |
| Subject Sample | # 2 | 0.080 | 17:04 | |
| Breath Volume | | 1.675 | 17:04 | |
| Air Blank | | 0.000 | 17:05 | |
| Air Blank | | 0.000 | 17:07 | |
| Subject Sample | # 3 | 0.078 | 17:08 | |
| Breath Volume | | 1.167 | 17:08 | |
| Air Blank | | 0.000 | 17:09 | |
| Control Test | | 0.076 | 17:09 | |
| Air Blank | | 0.000 | 17:09 | |
| Diagnostics Check | | OK | 17:10 | |

Data Download Date/Time: 06/29/2026 10:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CULPAN,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 05/24/2026 19:17:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:10 |
| Air Blank | | 0.000 | 19:11 |
| Control Test | | 0.080 | 19:11 |
| Air Blank | | 0.000 | 19:11 |
| Subject Sample | # 1 | 0.041 | 19:12 |
| Breath Volume | | 1.769 | 19:12 |
| Air Blank | | 0.000 | 19:13 |
| Air Blank | | 0.000 | 19:15 |
| Subject Sample | # 2 | 0.043 | 19:15 |
| Breath Volume | | 2.292 | 19:15 |
| Air Blank | | 0.000 | 19:16 |
| Control Test | | 0.079 | 19:16 |
| Air Blank | | 0.000 | 19:17 |
| Diagnostics Check | | OK | 19:17 |

Data Download Date/Time: 06/29/2026 10:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,DUSTIN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILLIAMS,BENJAMIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 05/27/2026 01:17:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:10 |
| Air Blank | | 0.000 | 01:11 |
| Control Test | | 0.080 | 01:11 |
| Air Blank | | 0.000 | 01:12 |
| Subject Sample | # 1 | 0.000 | 01:12 |
| Breath Volume | | 2.507 | 01:12 |
| Air Blank | | 0.000 | 01:13 |
| Air Blank | | 0.000 | 01:15 |
| Subject Sample | # 2 | 0.000 | 01:15 |
| Breath Volume | | 1.652 | 01:15 |
| Air Blank | | 0.000 | 01:16 |
| Control Test | | 0.079 | 01:16 |
| Air Blank | | 0.000 | 01:17 |
| Diagnostics Check | | OK | 01:17 |

Data Download Date/Time: 06/29/2026 10:54


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,DUSTIN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WRIGHT,DUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WILLKOMM

DATE/TIME OF TEST: 05/27/2026 03:44:07

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:32 | |
| Air Blank | | 0.000 | 03:32 | |
| Control Test | | 0.080 | 03:32 | |
| Air Blank | | 0.000 | 03:33 | |
| Subject Sample | # 1 | 0.162 | 03:36 | Volume Not Met |
| Breath Volume | | 0.156 | 03:36 | |
| Air Blank | | 0.000 | 03:37 | |
| Air Blank | | 0.000 | 03:39 | |
| Subject Sample | # 2 | 0.183 | 03:42 | Volume Not Met |
| Breath Volume | | 0.113 | 03:42 | |
| Air Blank | | 0.000 | 03:43 | |
| Control Test | | 0.076 | 03:43 | |
| Air Blank | | 0.000 | 03:44 | |
| Diagnostics Check | | OK | 03:44 | |

Data Download Date/Time: 06/29/2026 10:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,DUSTIN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WRIGHT,DUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WILLKOMM

DATE/TIME OF TEST: 05/27/2026 04:04:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:51 | |
| Air Blank | | 0.000 | 03:51 | |
| Control Test | | 0.077 | 03:51 | |
| Air Blank | | 0.000 | 03:52 | |
| Subject Sample | # 1 | 0.164 | 03:54 | |
| Breath Volume | | 1.250 | 03:54 | |
| Air Blank | | 0.000 | 03:55 | |
| Air Blank | | 0.000 | 03:57 | |
| Subject Sample | # 2 | 0.195 | 03:58 | |
| Breath Volume | | 1.296 | 03:58 | |
| Air Blank | | 0.000 | 04:00 | |
| Air Blank | | 0.000 | 04:01 | |
| Subject Sample | # 3 | 0.159 | 04:03 | |
| Breath Volume | | 1.507 | 04:03 | |
| Air Blank | | 0.000 | 04:03 | |
| Control Test | | 0.074 | 04:04 | Control Outside Tolerance |
| Air Blank | | 0.000 | 04:04 | |

Data Download Date/Time: 06/29/2026 10:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,DUSTIN,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WRIGHT,DUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: WILLKOMM

DATE/TIME OF TEST: 05/27/2026 04:27:39

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:17 | |
| Air Blank | | 0.000 | 04:18 | |
| Control Test | | 0.077 | 04:18 | |
| Air Blank | | 0.000 | 04:19 | |
| Subject Sample | # 1 | 0.177 | 04:22 | Volume Not Met |
| Breath Volume | | 0.171 | 04:22 | |
| Air Blank | | 0.000 | 04:23 | |
| Air Blank | | 0.000 | 04:24 | |
| Subject Sample | # 2 | | 04:25 | Subject Test Refused |
| Breath Volume | | 0.082 | 04:25 | |
| Air Blank | | 0.000 | 04:26 | |
| Control Test | | 0.074 | 04:27 | Control Outside Tolerance |
| Air Blank | | 0.000 | 04:27 | |

Data Download Date/Time: 06/29/2026 10:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: KHANG,ADAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: HINOJOSA

DATE/TIME OF TEST: 05/02/2026 07:52:08

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 07:49 | |
| Air Blank | | 0.000 | 07:49 | |
| Control Test | | 0.080 | 07:49 | |
| Air Blank | | 0.000 | 07:50 | |
| Subject Sample | # 1 | | 07:50 | Subject Test Refused |
| Breath Volume | | 0.000 | 07:50 | |
| Air Blank | | 0.000 | 07:51 | |
| Control Test | | 0.081 | 07:51 | |
| Air Blank | | 0.000 | 07:52 | |
| Diagnostics Check | | OK | 07:52 | |

Data Download Date/Time: 05/31/2026 20:28

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,SOLEY,G
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: CUSICK,WILLIAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: REGINO CARRILLO

DATE/TIME OF TEST: 05/02/2026 22:52:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:45 |
| Air Blank | | 0.000 | 22:45 |
| Control Test | | 0.080 | 22:45 |
| Air Blank | | 0.000 | 22:46 |
| Subject Sample | # 1 | 0.253 | 22:47 |
| Breath Volume | | 2.351 | 22:47 |
| Air Blank | | 0.000 | 22:48 |
| Air Blank | | 0.000 | 22:49 |
| Subject Sample | # 2 | 0.243 | 22:50 |
| Breath Volume | | 2.964 | 22:50 |
| Air Blank | | 0.000 | 22:51 |
| Control Test | | 0.077 | 22:51 |
| Air Blank | | 0.000 | 22:52 |
| Diagnostics Check | | OK | 22:52 |

Data Download Date/Time: 05/31/2026 20:28

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: EARLE,RODNEY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 05/08/2026 21:48:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:41 |
| Air Blank | | 0.000 | 21:42 |
| Control Test | | 0.080 | 21:42 |
| Air Blank | | 0.000 | 21:43 |
| Subject Sample | # 1 | 0.124 | 21:43 |
| Breath Volume | | 2.429 | 21:43 |
| Air Blank | | 0.000 | 21:44 |
| Air Blank | | 0.000 | 21:46 |
| Subject Sample | # 2 | 0.121 | 21:46 |
| Breath Volume | | 2.312 | 21:46 |
| Air Blank | | 0.000 | 21:47 |
| Control Test | | 0.078 | 21:47 |
| Air Blank | | 0.000 | 21:47 |
| Diagnostics Check | | OK | 21:48 |

Data Download Date/Time: 05/31/2026 20:28

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: GUINN,JALYNN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 05/09/2026 23:00:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:58 | |
| Air Blank | | 0.000 | 22:58 | |
| Control Test | | 0.081 | 22:58 | |
| Air Blank | | 0.000 | 22:59 | |
| Subject Sample | # 1 | | 22:59 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:59 | |
| Air Blank | | 0.000 | 23:00 | |
| Control Test | | 0.080 | 23:00 | |
| Air Blank | | 0.000 | 23:00 | |
| Diagnostics Check | | OK | 23:00 | |

Data Download Date/Time: 05/31/2026 20:28

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: DAWSON,TYLER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SAMBRANO

DATE/TIME OF TEST: 05/16/2026 02:08:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:05 | |
| Air Blank | | 0.000 | 02:06 | |
| Control Test | | 0.079 | 02:06 | |
| Air Blank | | 0.000 | 02:07 | |
| Subject Sample | # 1 | | 02:07 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:07 | |
| Air Blank | | 0.000 | 02:07 | |
| Control Test | | 0.080 | 02:08 | |
| Air Blank | | 0.000 | 02:08 | |
| Diagnostics Check | | OK | 02:08 | |

Data Download Date/Time: 05/31/2026 20:28

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: CUSICK,WILLIAM
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 04/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: ALAMIA

DATE/TIME OF TEST: 05/17/2026 01:56:01

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:53 | |
| Air Blank | | 0.000 | 01:53 | |
| Control Test | | 0.081 | 01:53 | |
| Air Blank | | 0.000 | 01:54 | |
| Subject Sample | # 1 | | 01:54 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:54 | |
| Air Blank | | 0.000 | 01:55 | |
| Control Test | | 0.080 | 01:55 | |
| Air Blank | | 0.000 | 01:55 | |
| Diagnostics Check | | OK | 01:56 | |

Data Download Date/Time: 05/31/2026 20:28

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: REANO GARCIA

DATE/TIME OF TEST: 05/02/2026 00:16:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:10 |
| Air Blank | | 0.000 | 00:10 |
| Control Test | | 0.079 | 00:11 |
| Air Blank | | 0.000 | 00:11 |
| Subject Sample | # 1 | 0.168 | 00:12 |
| Breath Volume | | 1.761 | 00:12 |
| Air Blank | | 0.000 | 00:12 |
| Air Blank | | 0.000 | 00:14 |
| Subject Sample | # 2 | 0.170 | 00:15 |
| Breath Volume | | 2.484 | 00:15 |
| Air Blank | | 0.000 | 00:15 |
| Control Test | | 0.078 | 00:16 |
| Air Blank | | 0.000 | 00:16 |
| Diagnostics Check | | OK | 00:16 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PINCAY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: SETTE

DATE/TIME OF TEST: 05/02/2026 03:08:56

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:02 |
| Air Blank | | 0.000 | 03:02 |
| Control Test | | 0.078 | 03:03 |
| Air Blank | | 0.000 | 03:03 |
| Subject Sample | # 1 | 0.209 | 03:04 |
| Breath Volume | | 2.519 | 03:04 |
| Air Blank | | 0.000 | 03:04 |
| Air Blank | | 0.000 | 03:06 |
| Subject Sample | # 2 | 0.209 | 03:07 |
| Breath Volume | | 1.531 | 03:07 |
| Air Blank | | 0.000 | 03:07 |
| Control Test | | 0.077 | 03:08 |
| Air Blank | | 0.000 | 03:08 |
| Diagnostics Check | | OK | 03:08 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PINCAY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: VIRAMONTES

DATE/TIME OF TEST: 05/03/2026 02:23:28

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:11 | |
| Air Blank | | 0.000 | 02:12 | |
| Control Test | | 0.079 | 02:12 | |
| Air Blank | | 0.000 | 02:13 | |
| Subject Sample | # 1 | | 02:16 | No Sample Provided |
| Breath Volume | | 0.000 | 02:16 | |
| Air Blank | | 0.000 | 02:16 | |
| Air Blank | | 0.000 | 02:18 | |
| Subject Sample | # 2 | | 02:22 | No Sample Provided |
| Breath Volume | | 0.000 | 02:22 | |
| Air Blank | | 0.000 | 02:22 | |
| Control Test | | 0.079 | 02:22 | |
| Air Blank | | 0.000 | 02:23 | |
| Diagnostics Check | | OK | 02:23 | |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PINCAY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: VIRAMONTES

DATE/TIME OF TEST: 05/03/2026 02:44:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Control Test | | 0.079 | 02:37 |
| Air Blank | | 0.000 | 02:38 |
| Subject Sample | # 1 | 0.110 | 02:39 |
| Breath Volume | | 2.625 | 02:39 |
| Air Blank | | 0.000 | 02:39 |
| Air Blank | | 0.000 | 02:41 |
| Subject Sample | # 2 | 0.115 | 02:42 |
| Breath Volume | | 3.804 | 02:42 |
| Air Blank | | 0.000 | 02:43 |
| Control Test | | 0.081 | 02:43 |
| Air Blank | | 0.000 | 02:44 |
| Diagnostics Check | | OK | 02:44 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DISLA GARCIA

DATE/TIME OF TEST: 05/03/2026 03:22:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:15 |
| Air Blank | | 0.000 | 03:15 |
| Control Test | | 0.078 | 03:15 |
| Air Blank | | 0.000 | 03:16 |
| Subject Sample | # 1 | 0.024 | 03:17 |
| Breath Volume | | 1.613 | 03:17 |
| Air Blank | | 0.000 | 03:18 |
| Air Blank | | 0.000 | 03:19 |
| Subject Sample | # 2 | 0.024 | 03:20 |
| Breath Volume | | 1.792 | 03:20 |
| Air Blank | | 0.000 | 03:21 |
| Control Test | | 0.079 | 03:21 |
| Air Blank | | 0.000 | 03:22 |
| Diagnostics Check | | OK | 03:22 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BRANTLEY JR

DATE/TIME OF TEST: 05/03/2026 04:10:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:04 |
| Air Blank | | 0.000 | 04:04 |
| Control Test | | 0.079 | 04:05 |
| Air Blank | | 0.000 | 04:05 |
| Subject Sample | # 1 | 0.103 | 04:06 |
| Breath Volume | | 2.617 | 04:06 |
| Air Blank | | 0.000 | 04:06 |
| Air Blank | | 0.000 | 04:08 |
| Subject Sample | # 2 | 0.104 | 04:09 |
| Breath Volume | | 1.511 | 04:09 |
| Air Blank | | 0.000 | 04:09 |
| Control Test | | 0.079 | 04:10 |
| Air Blank | | 0.000 | 04:10 |
| Diagnostics Check | | OK | 04:10 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 05/04/2026 00:02:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:55 |
| Air Blank | | 0.000 | 23:55 |
| Control Test | | 0.080 | 23:56 |
| Air Blank | | 0.000 | 23:56 |
| Subject Sample | # 2 | 0.099 | 23:57 |
| Breath Volume | | 1.464 | 23:57 |
| Air Blank | | 0.000 | 23:58 |
| Air Blank | | 0.000 | 23:59 |
| Subject Sample | # 1 | 0.101 | 00:00 |
| Breath Volume | | 1.429 | 00:00 |
| Air Blank | | 0.000 | 00:01 |
| Control Test | | 0.078 | 00:01 |
| Air Blank | | 0.000 | 00:01 |
| Diagnostics Check | | OK | 00:02 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: COLLETT

DATE/TIME OF TEST: 05/04/2026 02:35:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:28 |
| Air Blank | | 0.000 | 02:28 |
| Control Test | | 0.079 | 02:29 |
| Air Blank | | 0.000 | 02:29 |
| Subject Sample | # 1 | 0.150 | 02:30 |
| Breath Volume | | 1.234 | 02:30 |
| Air Blank | | 0.000 | 02:31 |
| Air Blank | | 0.000 | 02:33 |
| Subject Sample | # 2 | 0.155 | 02:33 |
| Breath Volume | | 2.847 | 02:33 |
| Air Blank | | 0.000 | 02:34 |
| Control Test | | 0.078 | 02:34 |
| Air Blank | | 0.000 | 02:35 |
| Diagnostics Check | | OK | 02:35 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BASU

DATE/TIME OF TEST: 05/04/2026 04:11:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:04 |
| Air Blank | | 0.000 | 04:04 |
| Control Test | | 0.078 | 04:05 |
| Air Blank | | 0.000 | 04:05 |
| Subject Sample | # 1 | 0.158 | 04:06 |
| Breath Volume | | 2.375 | 04:06 |
| Air Blank | | 0.000 | 04:07 |
| Air Blank | | 0.000 | 04:09 |
| Subject Sample | # 2 | 0.161 | 04:09 |
| Breath Volume | | 2.609 | 04:09 |
| Air Blank | | 0.000 | 04:10 |
| Control Test | | 0.078 | 04:10 |
| Air Blank | | 0.000 | 04:11 |
| Diagnostics Check | | OK | 04:11 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEACH,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: KRAUS

DATE/TIME OF TEST: 05/05/2026 00:02:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:56 |
| Air Blank | | 0.000 | 23:56 |
| Control Test | | 0.079 | 23:56 |
| Air Blank | | 0.000 | 23:57 |
| Subject Sample | # 2 | 0.254 | 23:58 |
| Breath Volume | | 2.066 | 23:58 |
| Air Blank | | 0.000 | 23:58 |
| Air Blank | | 0.000 | 00:00 |
| Subject Sample | # 1 | 0.252 | 00:01 |
| Breath Volume | | 2.558 | 00:01 |
| Air Blank | | 0.000 | 00:01 |
| Control Test | | 0.078 | 00:02 |
| Air Blank | | 0.000 | 00:02 |
| Diagnostics Check | | OK | 00:02 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: INDOVINA

DATE/TIME OF TEST: 05/09/2026 02:59:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Control Test | | 0.078 | 02:53 |
| Air Blank | | 0.000 | 02:54 |
| Subject Sample | # 1 | 0.131 | 02:54 |
| Breath Volume | | 2.335 | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Air Blank | | 0.000 | 02:57 |
| Subject Sample | # 2 | 0.131 | 02:57 |
| Breath Volume | | 2.113 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Control Test | | 0.078 | 02:58 |
| Air Blank | | 0.000 | 02:59 |
| Diagnostics Check | | OK | 02:59 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SANTIMAURO

DATE/TIME OF TEST: 05/09/2026 03:14:03

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:07 |
| Air Blank | | 0.000 | 03:07 |
| Control Test | | 0.078 | 03:08 |
| Air Blank | | 0.000 | 03:08 |
| Subject Sample | # 1 | 0.122 | 03:09 |
| Breath Volume | | 3.632 | 03:09 |
| Air Blank | | 0.000 | 03:10 |
| Air Blank | | 0.000 | 03:11 |
| Subject Sample | # 2 | 0.123 | 03:12 |
| Breath Volume | | 2.593 | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Control Test | | 0.080 | 03:13 |
| Air Blank | | 0.000 | 03:13 |
| Diagnostics Check | | OK | 03:14 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 05/09/2026 04:48:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:38 |
| Air Blank | | 0.000 | 04:38 |
| Control Test | | 0.079 | 04:39 |
| Air Blank | | 0.000 | 04:39 |
| Subject Sample | # 1 | 0.220 | 04:40 |
| Breath Volume | | 1.902 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Air Blank | | 0.000 | 04:42 |
| Subject Sample | # 2 | 0.244 | 04:43 |
| Breath Volume | | 1.812 | 04:43 |
| Air Blank | | 0.000 | 04:44 |
| Air Blank | | 0.000 | 04:46 |
| Subject Sample | # 3 | 0.245 | 04:47 |
| Breath Volume | | 3.066 | 04:47 |
| Air Blank | | 0.000 | 04:47 |
| Control Test | | 0.081 | 04:48 |
| Air Blank | | 0.000 | 04:48 |
| Diagnostics Check | | OK | 04:48 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: CASAZZA

DATE/TIME OF TEST: 05/10/2026 00:55:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:49 |
| Air Blank | | 0.000 | 00:49 |
| Control Test | | 0.079 | 00:50 |
| Air Blank | | 0.000 | 00:50 |
| Subject Sample | # 1 | 0.194 | 00:51 |
| Breath Volume | | 3.933 | 00:51 |
| Air Blank | | 0.000 | 00:52 |
| Air Blank | | 0.000 | 00:53 |
| Subject Sample | # 2 | 0.204 | 00:54 |
| Breath Volume | | 3.554 | 00:54 |
| Air Blank | | 0.000 | 00:55 |
| Control Test | | 0.082 | 00:55 |
| Air Blank | | 0.000 | 00:55 |
| Diagnostics Check | | OK | 00:55 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MADISON

DATE/TIME OF TEST: 05/12/2026 03:00:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:52 |
| Air Blank | | 0.000 | 02:52 |
| Control Test | | 0.079 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Subject Sample | # 1 | 0.131 | 02:54 |
| Breath Volume | | 1.878 | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Air Blank | | 0.000 | 02:57 |
| Subject Sample | # 2 | 0.137 | 02:58 |
| Breath Volume | | 1.816 | 02:58 |
| Air Blank | | 0.000 | 02:59 |
| Control Test | | 0.082 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Diagnostics Check | | OK | 03:00 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VANTREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ACABA

DATE/TIME OF TEST: 05/14/2026 01:48:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:41 |
| Air Blank | | 0.000 | 01:42 |
| Control Test | | 0.078 | 01:42 |
| Air Blank | | 0.000 | 01:43 |
| Subject Sample | # 1 | 0.130 | 01:43 |
| Breath Volume | | 2.660 | 01:43 |
| Air Blank | | 0.000 | 01:44 |
| Air Blank | | 0.000 | 01:46 |
| Subject Sample | # 2 | 0.122 | 01:46 |
| Breath Volume | | 2.230 | 01:46 |
| Air Blank | | 0.000 | 01:47 |
| Control Test | | 0.078 | 01:47 |
| Air Blank | | 0.000 | 01:48 |
| Diagnostics Check | | OK | 01:48 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: KAISER

DATE/TIME OF TEST: 05/15/2026 01:56:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:49 |
| Air Blank | | 0.000 | 01:50 |
| Control Test | | 0.079 | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Subject Sample | # 1 | 0.032 | 01:52 |
| Breath Volume | | 1.335 | 01:52 |
| Air Blank | | 0.000 | 01:52 |
| Air Blank | | 0.000 | 01:54 |
| Subject Sample | # 2 | 0.033 | 01:54 |
| Breath Volume | | 1.636 | 01:54 |
| Air Blank | | 0.000 | 01:55 |
| Control Test | | 0.079 | 01:55 |
| Air Blank | | 0.000 | 01:56 |
| Diagnostics Check | | OK | 01:56 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EHRHART,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: DEXTER

DATE/TIME OF TEST: 05/15/2026 02:49:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:39 |
| Air Blank | | 0.000 | 02:39 |
| Control Test | | 0.079 | 02:40 |
| Air Blank | | 0.000 | 02:40 |
| Subject Sample | # 1 | 0.118 | 02:43 |
| Breath Volume | | 2.964 | 02:43 |
| Air Blank | | 0.000 | 02:43 |
| Air Blank | | 0.000 | 02:45 |
| Subject Sample | # 2 | 0.115 | 02:47 |
| Breath Volume | | 2.546 | 02:47 |
| Air Blank | | 0.000 | 02:48 |
| Control Test | | 0.079 | 02:48 |
| Air Blank | | 0.000 | 02:49 |
| Diagnostics Check | | OK | 02:49 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/15/2026 04:38:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:31 |
| Air Blank | | 0.000 | 04:31 |
| Control Test | | 0.079 | 04:32 |
| Air Blank | | 0.000 | 04:32 |
| Subject Sample | # 1 | 0.192 | 04:33 |
| Breath Volume | | 1.628 | 04:33 |
| Air Blank | | 0.000 | 04:34 |
| Air Blank | | 0.000 | 04:35 |
| Subject Sample | # 2 | 0.202 | 04:36 |
| Breath Volume | | 2.472 | 04:36 |
| Air Blank | | 0.000 | 04:37 |
| Control Test | | 0.079 | 04:37 |
| Air Blank | | 0.000 | 04:38 |
| Diagnostics Check | | OK | 04:38 |

Data Download Date/Time: 05/15/2026 12:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PINCAY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 05/16/2026 01:20:10

| | | | |
|---|---|---|---|
| Diagnostics Check | Fail | 01:19 | Internal Printer Test |
| Air Blank | 0.000 | 01:20 | |

Data Download Date/Time: 06/24/2026 11:34


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PINCAY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 05/16/2026 01:37:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:30 |
| Air Blank | | 0.000 | 01:31 |
| Control Test | | 0.079 | 01:31 |
| Air Blank | | 0.000 | 01:32 |
| Subject Sample | # 1 | 0.098 | 01:32 |
| Breath Volume | | 2.328 | 01:32 |
| Air Blank | | 0.000 | 01:33 |
| Air Blank | | 0.000 | 01:35 |
| Subject Sample | # 2 | 0.096 | 01:35 |
| Breath Volume | | 2.289 | 01:35 |
| Air Blank | | 0.000 | 01:36 |
| Control Test | | 0.078 | 01:36 |
| Air Blank | | 0.000 | 01:37 |
| Diagnostics Check | | OK | 01:37 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: DELGADO NERIS

DATE/TIME OF TEST: 05/16/2026 02:53:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:46 |
| Air Blank | | 0.000 | 02:47 |
| Control Test | | 0.079 | 02:47 |
| Air Blank | | 0.000 | 02:48 |
| Subject Sample | # 1 | 0.017 | 02:49 |
| Breath Volume | | 2.804 | 02:49 |
| Air Blank | | 0.000 | 02:49 |
| Air Blank | | 0.000 | 02:51 |
| Subject Sample | # 2 | 0.015 | 02:52 |
| Breath Volume | | 2.515 | 02:52 |
| Air Blank | | 0.000 | 02:52 |
| Control Test | | 0.079 | 02:53 |
| Air Blank | | 0.000 | 02:53 |
| Diagnostics Check | | OK | 02:53 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 05/16/2026 03:29:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:23 |
| Air Blank | | 0.000 | 03:23 |
| Control Test | | 0.078 | 03:24 |
| Air Blank | | 0.000 | 03:24 |
| Subject Sample | # 1 | 0.155 | 03:25 |
| Breath Volume | | 1.367 | 03:25 |
| Air Blank | | 0.000 | 03:25 |
| Air Blank | | 0.000 | 03:27 |
| Subject Sample | # 2 | 0.158 | 03:28 |
| Breath Volume | | 1.398 | 03:28 |
| Air Blank | | 0.000 | 03:28 |
| Control Test | | 0.078 | 03:29 |
| Air Blank | | 0.000 | 03:29 |
| Diagnostics Check | | OK | 03:29 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: NAVARRO PEREZ

DATE/TIME OF TEST: 05/16/2026 04:07:23

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Control Test | | 0.078 | 04:01 |
| Air Blank | | 0.000 | 04:02 |
| Subject Sample | # 1 | 0.244 | 04:02 |
| Breath Volume | | 2.746 | 04:02 |
| Air Blank | | 0.000 | 04:03 |
| Air Blank | | 0.000 | 04:05 |
| Subject Sample | # 2 | 0.241 | 04:05 |
| Breath Volume | | 1.750 | 04:05 |
| Air Blank | | 0.000 | 04:06 |
| Control Test | | 0.078 | 04:06 |
| Air Blank | | 0.000 | 04:07 |
| Diagnostics Check | | OK | 04:07 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THEODORE,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:42

SUBJECT'S LAST NAME: TORIBIO

DATE/TIME OF TEST: 05/16/2026 08:09:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:02 |
| Air Blank | | 0.000 | 08:02 |
| Control Test | | 0.078 | 08:03 |
| Air Blank | | 0.000 | 08:03 |
| Subject Sample | # 1 | 0.205 | 08:04 |
| Breath Volume | | 1.832 | 08:04 |
| Air Blank | | 0.000 | 08:05 |
| Air Blank | | 0.000 | 08:07 |
| Subject Sample | # 2 | 0.208 | 08:07 |
| Breath Volume | | 1.839 | 08:07 |
| Air Blank | | 0.000 | 08:08 |
| Control Test | | 0.077 | 08:08 |
| Air Blank | | 0.000 | 08:09 |
| Diagnostics Check | | OK | 08:09 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MAXIME,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 05/16/2026 11:27:48

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 11:16 | |
| Air Blank | | 0.000 | 11:16 | |
| Control Test | | 0.079 | 11:16 | |
| Air Blank | | 0.000 | 11:17 | |
| Subject Sample | # 1 | 0.149 | 11:20 | Volume Not Met |
| Breath Volume | | 0.304 | 11:20 | |
| Air Blank | | 0.000 | 11:21 | |
| Air Blank | | 0.000 | 11:22 | |
| Subject Sample | # 2 | 0.040 | 11:26 | Volume Not Met |
| Breath Volume | | 0.082 | 11:26 | |
| Air Blank | | 0.000 | 11:26 | |
| Control Test | | 0.079 | 11:27 | |
| Air Blank | | 0.000 | 11:27 | |
| Diagnostics Check | | OK | 11:27 | |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MAXIME,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: ASHLEY III

DATE/TIME OF TEST: 05/16/2026 11:38:37

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 11:34 | |
| Air Blank | | 0.000 | 11:35 | |
| Control Test | | 0.077 | 11:35 | |
| Air Blank | | 0.000 | 11:36 | |
| Subject Sample | # 1 | | 11:37 | Subject Test Refused |
| Breath Volume | | 0.000 | 11:37 | |
| Air Blank | | 0.000 | 11:37 | |
| Control Test | | 0.078 | 11:37 | |
| Air Blank | | 0.000 | 11:38 | |
| Diagnostics Check | | OK | 11:38 | |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BERMUDEZ PERALTA

DATE/TIME OF TEST: 05/16/2026 23:56:30

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:48 |
| Air Blank | | 0.000 | 23:49 |
| Control Test | | 0.078 | 23:49 |
| Air Blank | | 0.000 | 23:50 |
| Subject Sample | # 1 | 0.234 | 23:51 |
| Breath Volume | | 2.066 | 23:51 |
| Air Blank | | 0.000 | 23:52 |
| Air Blank | | 0.000 | 23:54 |
| Subject Sample | # 2 | 0.232 | 23:54 |
| Breath Volume | | 2.667 | 23:54 |
| Air Blank | | 0.000 | 23:55 |
| Control Test | | 0.076 | 23:55 |
| Air Blank | | 0.000 | 23:56 |
| Diagnostics Check | | OK | 23:56 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LAROSE,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: JOFFRE

DATE/TIME OF TEST: 05/17/2026 00:36:26

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:23 | |
| Air Blank | | 0.000 | 00:24 | |
| Control Test | | 0.079 | 00:24 | |
| Air Blank | | 0.000 | 00:25 | |
| Subject Sample | # 1 | | 00:28 | No Sample Provided |
| Breath Volume | | 0.000 | 00:28 | |
| Air Blank | | 0.000 | 00:28 | |
| Air Blank | | 0.000 | 00:30 | |
| Subject Sample | # 2 | 0.146 | 00:31 | |
| Breath Volume | | 1.769 | 00:31 | |
| Air Blank | | 0.000 | 00:32 | |
| Air Blank | | 0.000 | 00:34 | |
| Subject Sample | # 3 | 0.139 | 00:34 | |
| Breath Volume | | 1.214 | 00:34 | |
| Air Blank | | 0.000 | 00:35 | |
| Control Test | | 0.078 | 00:35 | |
| Air Blank | | 0.000 | 00:36 | |
| Diagnostics Check | | OK | 00:36 | |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 05/17/2026 02:59:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:52 |
| Air Blank | | 0.000 | 02:52 |
| Control Test | | 0.078 | 02:53 |
| Air Blank | | 0.000 | 02:53 |
| Subject Sample | # 1 | 0.158 | 02:55 |
| Breath Volume | | 1.375 | 02:55 |
| Air Blank | | 0.000 | 02:55 |
| Air Blank | | 0.000 | 02:57 |
| Subject Sample | # 2 | 0.152 | 02:58 |
| Breath Volume | | 1.511 | 02:58 |
| Air Blank | | 0.000 | 02:58 |
| Control Test | | 0.081 | 02:59 |
| Air Blank | | 0.000 | 02:59 |
| Diagnostics Check | | OK | 02:59 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WURSTER,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: ESCOBAR SUAREZ

DATE/TIME OF TEST: 05/17/2026 08:51:23

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:42 |
| Air Blank | | 0.000 | 08:43 |
| Control Test | | 0.078 | 08:43 |
| Air Blank | | 0.000 | 08:43 |
| Subject Sample | # 1 | 0.132 | 08:46 |
| Breath Volume | | 1.296 | 08:46 |
| Air Blank | | 0.000 | 08:47 |
| Air Blank | | 0.000 | 08:49 |
| Subject Sample | # 2 | 0.134 | 08:49 |
| Breath Volume | | 1.320 | 08:49 |
| Air Blank | | 0.000 | 08:50 |
| Control Test | | 0.081 | 08:50 |
| Air Blank | | 0.000 | 08:51 |
| Diagnostics Check | | OK | 08:51 |

Data Download Date/Time: 06/24/2026 11:34


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: NUVIOLA SANCHEZ

DATE/TIME OF TEST: 05/18/2026 01:03:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:55 |
| Air Blank | | 0.000 | 00:55 |
| Control Test | | 0.078 | 00:56 |
| Air Blank | | 0.000 | 00:56 |
| Subject Sample | # 1 | 0.085 | 00:58 |
| Breath Volume | | 2.101 | 00:58 |
| Air Blank | | 0.000 | 00:59 |
| Air Blank | | 0.000 | 01:00 |
| Subject Sample | # 2 | 0.086 | 01:01 |
| Breath Volume | | 2.359 | 01:01 |
| Air Blank | | 0.000 | 01:02 |
| Control Test | | 0.082 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Diagnostics Check | | OK | 01:03 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ACEVEDO,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: FALADE

DATE/TIME OF TEST: 05/18/2026 03:12:36

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:59 | |
| Air Blank | | 0.000 | 03:00 | |
| Control Test | | 0.078 | 03:00 | |
| Air Blank | | 0.000 | 03:00 | |
| Subject Sample | # 1 | 0.032 | 03:04 | Volume Not Met |
| Breath Volume | | 0.031 | 03:04 | |
| Air Blank | | 0.000 | 03:04 | |
| Air Blank | | 0.000 | 03:06 | |
| Subject Sample | # 2 | 0.032 | 03:07 | |
| Breath Volume | | 1.464 | 03:07 | |
| Air Blank | | 0.000 | 03:07 | |
| Air Blank | | 0.000 | 03:09 | |
| Subject Sample | # 3 | 0.032 | 03:10 | |
| Breath Volume | | 1.132 | 03:10 | |
| Air Blank | | 0.000 | 03:11 | |
| Control Test | | 0.083 | 03:11 | |
| Air Blank | | 0.000 | 03:12 | |
| Diagnostics Check | | OK | 03:12 | |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PETERS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: QUEZADA LOPEZ

DATE/TIME OF TEST: 05/20/2026 00:37:11

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:25 | |
| Air Blank | | 0.000 | 00:25 | |
| Control Test | | 0.079 | 00:26 | |
| Air Blank | | 0.000 | 00:26 | |
| Subject Sample | # 1 | 0.179 | 00:29 | Volume Not Met |
| Breath Volume | | 0.406 | 00:29 | |
| Air Blank | | 0.000 | 00:30 | |
| Air Blank | | 0.000 | 00:32 | |
| Subject Sample | # 2 | 0.167 | 00:35 | Volume Not Met |
| Breath Volume | | 0.074 | 00:35 | |
| Air Blank | | 0.000 | 00:36 | |
| Control Test | | 0.077 | 00:36 | |
| Air Blank | | 0.000 | 00:37 | |
| Diagnostics Check | | OK | 00:37 | |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PETERS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: QUEZADA LOPEZ

DATE/TIME OF TEST: 05/20/2026 01:07:32

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:57 | |
| Air Blank | | 0.000 | 00:57 | |
| Control Test | | 0.077 | 00:58 | |
| Air Blank | | 0.000 | 00:58 | |
| Subject Sample | # 1 | 0.186 | 01:02 | Volume Not Met |
| Breath Volume | | 0.144 | 01:02 | |
| Air Blank | | 0.000 | 01:02 | |
| Air Blank | | 0.000 | 01:04 | |
| Subject Sample | # 2 | | 01:05 | Subject Test Refused |
| Breath Volume | | 0.164 | 01:05 | |
| Air Blank | | 0.000 | 01:06 | |
| Control Test | | 0.077 | 01:06 | |
| Air Blank | | 0.000 | 01:07 | |
| Diagnostics Check | | OK | 01:07 | |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: DE LEON

DATE/TIME OF TEST: 05/20/2026 02:25:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:18 |
| Air Blank | | 0.000 | 02:19 |
| Control Test | | 0.078 | 02:19 |
| Air Blank | | 0.000 | 02:20 |
| Subject Sample | # 1 | 0.168 | 02:20 |
| Breath Volume | | 2.343 | 02:20 |
| Air Blank | | 0.000 | 02:21 |
| Air Blank | | 0.000 | 02:23 |
| Subject Sample | # 2 | 0.157 | 02:23 |
| Breath Volume | | 1.476 | 02:23 |
| Air Blank | | 0.000 | 02:24 |
| Control Test | | 0.078 | 02:24 |
| Air Blank | | 0.000 | 02:25 |
| Diagnostics Check | | OK | 02:25 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST.WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MUNOZ CASTIBLANCO

DATE/TIME OF TEST: 05/22/2026 02:28:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:21 |
| Air Blank | | 0.000 | 02:21 |
| Control Test | | 0.078 | 02:22 |
| Air Blank | | 0.000 | 02:22 |
| Subject Sample | # 1 | 0.139 | 02:23 |
| Breath Volume | | 3.433 | 02:23 |
| Air Blank | | 0.000 | 02:24 |
| Air Blank | | 0.000 | 02:25 |
| Subject Sample | # 2 | 0.134 | 02:26 |
| Breath Volume | | 2.640 | 02:26 |
| Air Blank | | 0.000 | 02:27 |
| Control Test | | 0.078 | 02:27 |
| Air Blank | | 0.000 | 02:28 |
| Diagnostics Check | | OK | 02:28 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: TANKERSLEY

DATE/TIME OF TEST: 05/22/2026 04:52:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:45 |
| Air Blank | | 0.000 | 04:45 |
| Control Test | | 0.078 | 04:45 |
| Air Blank | | 0.000 | 04:46 |
| Subject Sample | # 1 | 0.228 | 04:47 |
| Breath Volume | | 2.011 | 04:47 |
| Air Blank | | 0.000 | 04:48 |
| Air Blank | | 0.000 | 04:49 |
| Subject Sample | # 2 | 0.221 | 04:50 |
| Breath Volume | | 1.707 | 04:50 |
| Air Blank | | 0.000 | 04:51 |
| Control Test | | 0.078 | 04:51 |
| Air Blank | | 0.000 | 04:52 |
| Diagnostics Check | | OK | 04:52 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: GONTARSKI

DATE/TIME OF TEST: 05/22/2026 05:33:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:26 |
| Air Blank | | 0.000 | 05:26 |
| Control Test | | 0.078 | 05:26 |
| Air Blank | | 0.000 | 05:27 |
| Subject Sample | # 1 | 0.204 | 05:28 |
| Breath Volume | | 1.789 | 05:28 |
| Air Blank | | 0.000 | 05:28 |
| Air Blank | | 0.000 | 05:30 |
| Subject Sample | # 2 | 0.201 | 05:31 |
| Breath Volume | | 1.859 | 05:31 |
| Air Blank | | 0.000 | 05:32 |
| Control Test | | 0.078 | 05:32 |
| Air Blank | | 0.000 | 05:32 |
| Diagnostics Check | | OK | 05:33 |

Data Download Date/Time: 06/24/2026 11:34

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERRERA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CLARK SR

DATE/TIME OF TEST: 05/23/2026 01:26:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:18 |
| Air Blank | | 0.000 | 01:19 |
| Control Test | | 0.078 | 01:19 |
| Air Blank | | 0.000 | 01:20 |
| Subject Sample | # 1 | 0.123 | 01:21 |
| Breath Volume | | 1.964 | 01:21 |
| Air Blank | | 0.000 | 01:22 |
| Air Blank | | 0.000 | 01:23 |
| Subject Sample | # 2 | 0.119 | 01:24 |
| Breath Volume | | 1.335 | 01:24 |
| Air Blank | | 0.000 | 01:25 |
| Control Test | | 0.079 | 01:25 |
| Air Blank | | 0.000 | 01:26 |
| Diagnostics Check | | OK | 01:26 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEACH,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: COMM VEHIC
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: ECKER

DATE/TIME OF TEST: 05/23/2026 02:02:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:55 |
| Air Blank | | 0.000 | 01:55 |
| Control Test | | 0.078 | 01:55 |
| Air Blank | | 0.000 | 01:56 |
| Subject Sample | # 1 | 0.124 | 01:58 |
| Breath Volume | | 1.148 | 01:58 |
| Air Blank | | 0.000 | 01:58 |
| Air Blank | | 0.000 | 02:00 |
| Subject Sample | # 2 | 0.122 | 02:01 |
| Breath Volume | | 1.101 | 02:01 |
| Air Blank | | 0.000 | 02:01 |
| Control Test | | 0.080 | 02:02 |
| Air Blank | | 0.000 | 02:02 |
| Diagnostics Check | | OK | 02:02 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MAROCOLA,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 05/23/2026 04:13:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:07 |
| Air Blank | | 0.000 | 04:07 |
| Control Test | | 0.078 | 04:07 |
| Air Blank | | 0.000 | 04:08 |
| Subject Sample | # 1 | 0.082 | 04:09 |
| Breath Volume | | 2.371 | 04:09 |
| Air Blank | | 0.000 | 04:09 |
| Air Blank | | 0.000 | 04:11 |
| Subject Sample | # 2 | 0.081 | 04:12 |
| Breath Volume | | 2.542 | 04:12 |
| Air Blank | | 0.000 | 04:12 |
| Control Test | | 0.078 | 04:13 |
| Air Blank | | 0.000 | 04:13 |
| Diagnostics Check | | OK | 04:13 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: PALMA CUPIDO

DATE/TIME OF TEST: 05/23/2026 22:35:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:28 |
| Air Blank | | 0.000 | 22:29 |
| Control Test | | 0.079 | 22:29 |
| Air Blank | | 0.000 | 22:30 |
| Subject Sample | # 1 | 0.167 | 22:30 |
| Breath Volume | | 1.437 | 22:30 |
| Air Blank | | 0.000 | 22:31 |
| Air Blank | | 0.000 | 22:32 |
| Subject Sample | # 2 | 0.164 | 22:33 |
| Breath Volume | | 2.148 | 22:33 |
| Air Blank | | 0.000 | 22:34 |
| Control Test | | 0.078 | 22:34 |
| Air Blank | | 0.000 | 22:34 |
| Diagnostics Check | | OK | 22:35 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: NEVILLE

DATE/TIME OF TEST: 05/23/2026 23:46:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:40 |
| Air Blank | | 0.000 | 23:40 |
| Control Test | | 0.079 | 23:41 |
| Air Blank | | 0.000 | 23:41 |
| Subject Sample | # 1 | 0.013 | 23:42 |
| Breath Volume | | 2.496 | 23:42 |
| Air Blank | | 0.000 | 23:42 |
| Air Blank | | 0.000 | 23:44 |
| Subject Sample | # 2 | 0.012 | 23:45 |
| Breath Volume | | 2.359 | 23:45 |
| Air Blank | | 0.000 | 23:45 |
| Control Test | | 0.078 | 23:45 |
| Air Blank | | 0.000 | 23:46 |
| Diagnostics Check | | OK | 23:46 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERRERA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: EDMONDSON

DATE/TIME OF TEST: 05/24/2026 00:21:11

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:14 |
| Air Blank | | 0.000 | 00:14 |
| Control Test | | 0.078 | 00:15 |
| Air Blank | | 0.000 | 00:15 |
| Subject Sample | # 1 | 0.114 | 00:16 |
| Breath Volume | | 2.101 | 00:16 |
| Air Blank | | 0.000 | 00:16 |
| Air Blank | | 0.000 | 00:18 |
| Subject Sample | # 2 | 0.115 | 00:19 |
| Breath Volume | | 3.140 | 00:19 |
| Air Blank | | 0.000 | 00:20 |
| Control Test | | 0.081 | 00:20 |
| Air Blank | | 0.000 | 00:21 |
| Diagnostics Check | | OK | 00:21 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIBBINS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: TORRANCE

DATE/TIME OF TEST: 05/24/2026 04:39:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:33 |
| Air Blank | | 0.000 | 04:34 |
| Control Test | | 0.078 | 04:34 |
| Air Blank | | 0.000 | 04:35 |
| Subject Sample | # 1 | 0.000 | 04:35 |
| Breath Volume | | 2.019 | 04:35 |
| Air Blank | | 0.000 | 04:36 |
| Air Blank | | 0.000 | 04:38 |
| Subject Sample | # 2 | 0.000 | 04:38 |
| Breath Volume | | 1.597 | 04:38 |
| Air Blank | | 0.000 | 04:38 |
| Control Test | | 0.078 | 04:39 |
| Air Blank | | 0.000 | 04:39 |
| Diagnostics Check | | OK | 04:39 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: OTERO GUZMAN

DATE/TIME OF TEST: 05/25/2026 00:08:30

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:59 | |
| Air Blank | | 0.000 | 00:00 | |
| Control Test | | 0.078 | 00:00 | |
| Air Blank | | 0.000 | 00:01 | |
| Subject Sample | # 1 | | 00:04 | No Sample Provided |
| Breath Volume | | 0.000 | 00:04 | |
| Air Blank | | 0.000 | 00:04 | |
| Air Blank | | 0.000 | 00:06 | |
| Subject Sample | # 2 | | 00:07 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:07 | |
| Air Blank | | 0.000 | 00:07 | |
| Control Test | | 0.078 | 00:07 | |
| Air Blank | | 0.000 | 00:08 | |
| Diagnostics Check | | OK | 00:08 | |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: JOYNER

DATE/TIME OF TEST: 05/25/2026 00:54:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:47 |
| Air Blank | | 0.000 | 00:48 |
| Control Test | | 0.079 | 00:48 |
| Air Blank | | 0.000 | 00:49 |
| Subject Sample | # 1 | 0.034 | 00:49 |
| Breath Volume | | 2.144 | 00:49 |
| Air Blank | | 0.000 | 00:50 |
| Air Blank | | 0.000 | 00:52 |
| Subject Sample | # 2 | 0.034 | 00:52 |
| Breath Volume | | 2.066 | 00:52 |
| Air Blank | | 0.000 | 00:53 |
| Control Test | | 0.079 | 00:53 |
| Air Blank | | 0.000 | 00:54 |
| Diagnostics Check | | OK | 00:54 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HERNANDEZ MONROY

DATE/TIME OF TEST: 05/25/2026 23:20:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:12 |
| Air Blank | | 0.000 | 23:12 |
| Control Test | | 0.079 | 23:12 |
| Air Blank | | 0.000 | 23:13 |
| Subject Sample | # 1 | 0.214 | 23:15 |
| Breath Volume | | 1.468 | 23:15 |
| Air Blank | | 0.000 | 23:15 |
| Air Blank | | 0.000 | 23:17 |
| Subject Sample | # 2 | 0.233 | 23:18 |
| Breath Volume | | 1.398 | 23:18 |
| Air Blank | | 0.000 | 23:19 |
| Control Test | | 0.081 | 23:19 |
| Air Blank | | 0.000 | 23:20 |
| Diagnostics Check | | OK | 23:20 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: PALACIOS

DATE/TIME OF TEST: 05/28/2026 04:11:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:04 |
| Air Blank | | 0.000 | 04:05 |
| Control Test | | 0.079 | 04:05 |
| Air Blank | | 0.000 | 04:06 |
| Subject Sample | # 1 | 0.252 | 04:07 |
| Breath Volume | | 2.101 | 04:07 |
| Air Blank | | 0.000 | 04:07 |
| Air Blank | | 0.000 | 04:09 |
| Subject Sample | # 2 | 0.253 | 04:10 |
| Breath Volume | | 2.304 | 04:10 |
| Air Blank | | 0.000 | 04:10 |
| Control Test | | 0.078 | 04:11 |
| Air Blank | | 0.000 | 04:11 |
| Diagnostics Check | | OK | 04:11 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 05/28/2026 05:47:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:38 |
| Air Blank | | 0.000 | 05:39 |
| Control Test | | 0.079 | 05:39 |
| Air Blank | | 0.000 | 05:40 |
| Subject Sample | # 1 | 0.240 | 05:41 |
| Breath Volume | | 1.390 | 05:41 |
| Air Blank | | 0.000 | 05:42 |
| Air Blank | | 0.000 | 05:44 |
| Subject Sample | # 2 | 0.249 | 05:45 |
| Breath Volume | | 1.773 | 05:45 |
| Air Blank | | 0.000 | 05:46 |
| Control Test | | 0.080 | 05:46 |
| Air Blank | | 0.000 | 05:47 |
| Diagnostics Check | | OK | 05:47 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BLANDING

DATE/TIME OF TEST: 05/29/2026 00:49:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:42 |
| Air Blank | | 0.000 | 00:42 |
| Control Test | | 0.079 | 00:42 |
| Air Blank | | 0.000 | 00:43 |
| Subject Sample | # 1 | 0.029 | 00:44 |
| Breath Volume | | 1.707 | 00:44 |
| Air Blank | | 0.000 | 00:44 |
| Air Blank | | 0.000 | 00:46 |
| Subject Sample | # 2 | 0.029 | 00:47 |
| Breath Volume | | 1.554 | 00:47 |
| Air Blank | | 0.000 | 00:48 |
| Control Test | | 0.079 | 00:48 |
| Air Blank | | 0.000 | 00:49 |
| Diagnostics Check | | OK | 00:49 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GOMENIS

DATE/TIME OF TEST: 05/29/2026 02:46:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:40 |
| Air Blank | | 0.000 | 02:40 |
| Control Test | | 0.079 | 02:41 |
| Air Blank | | 0.000 | 02:41 |
| Subject Sample | # 1 | 0.115 | 02:42 |
| Breath Volume | | 2.156 | 02:42 |
| Air Blank | | 0.000 | 02:42 |
| Air Blank | | 0.000 | 02:44 |
| Subject Sample | # 2 | 0.109 | 02:45 |
| Breath Volume | | 1.839 | 02:45 |
| Air Blank | | 0.000 | 02:45 |
| Control Test | | 0.078 | 02:46 |
| Air Blank | | 0.000 | 02:46 |
| Diagnostics Check | | OK | 02:46 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 05/29/2026 03:41:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:34 |
| Air Blank | | 0.000 | 03:35 |
| Control Test | | 0.079 | 03:35 |
| Air Blank | | 0.000 | 03:36 |
| Subject Sample | # 1 | 0.176 | 03:36 |
| Breath Volume | | 2.621 | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Air Blank | | 0.000 | 03:39 |
| Subject Sample | # 2 | 0.171 | 03:39 |
| Breath Volume | | 2.656 | 03:39 |
| Air Blank | | 0.000 | 03:40 |
| Control Test | | 0.078 | 03:40 |
| Air Blank | | 0.000 | 03:41 |
| Diagnostics Check | | OK | 03:41 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRRIZARY,KATTIE,D
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:07

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 05/29/2026 12:34:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 12:27 |
| Air Blank | | 0.000 | 12:27 |
| Control Test | | 0.079 | 12:28 |
| Air Blank | | 0.000 | 12:28 |
| Subject Sample | # 1 | 0.000 | 12:29 |
| Breath Volume | | 1.335 | 12:29 |
| Air Blank | | 0.000 | 12:30 |
| Air Blank | | 0.000 | 12:32 |
| Subject Sample | # 2 | 0.000 | 12:32 |
| Breath Volume | | 1.441 | 12:32 |
| Air Blank | | 0.000 | 12:33 |
| Control Test | | 0.079 | 12:33 |
| Air Blank | | 0.000 | 12:34 |
| Diagnostics Check | | OK | 12:34 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: RECKLEY JR

DATE/TIME OF TEST: 05/30/2026 03:31:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:24 |
| Air Blank | | 0.000 | 03:25 |
| Control Test | | 0.079 | 03:25 |
| Air Blank | | 0.000 | 03:26 |
| Subject Sample | # 1 | 0.018 | 03:26 |
| Breath Volume | | 1.492 | 03:26 |
| Air Blank | | 0.000 | 03:27 |
| Air Blank | | 0.000 | 03:29 |
| Subject Sample | # 2 | 0.020 | 03:30 |
| Breath Volume | | 2.375 | 03:30 |
| Air Blank | | 0.000 | 03:30 |
| Control Test | | 0.080 | 03:31 |
| Air Blank | | 0.000 | 03:31 |
| Diagnostics Check | | OK | 03:31 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: POMPA

DATE/TIME OF TEST: 05/30/2026 04:46:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:40 |
| Air Blank | | 0.000 | 04:40 |
| Control Test | | 0.079 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Subject Sample | # 1 | 0.093 | 04:42 |
| Breath Volume | | 3.812 | 04:42 |
| Air Blank | | 0.000 | 04:42 |
| Air Blank | | 0.000 | 04:44 |
| Subject Sample | # 2 | 0.091 | 04:45 |
| Breath Volume | | 3.535 | 04:45 |
| Air Blank | | 0.000 | 04:45 |
| Control Test | | 0.082 | 04:46 |
| Air Blank | | 0.000 | 04:46 |
| Diagnostics Check | | OK | 04:46 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SHADLER

DATE/TIME OF TEST: 05/31/2026 01:15:25

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:06 |
| Air Blank | | 0.000 | 01:06 |
| Control Test | | 0.080 | 01:07 |
| Air Blank | | 0.000 | 01:07 |
| Subject Sample | # 1 | 0.096 | 01:10 |
| Breath Volume | | 1.414 | 01:10 |
| Air Blank | | 0.000 | 01:10 |
| Air Blank | | 0.000 | 01:12 |
| Subject Sample | # 2 | 0.095 | 01:13 |
| Breath Volume | | 1.972 | 01:13 |
| Air Blank | | 0.000 | 01:14 |
| Control Test | | 0.082 | 01:14 |
| Air Blank | | 0.000 | 01:15 |
| Diagnostics Check | | OK | 01:15 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TABOADA,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MCDUFFIE

DATE/TIME OF TEST: 05/31/2026 02:57:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:48 |
| Air Blank | | 0.000 | 02:48 |
| Control Test | | 0.079 | 02:49 |
| Air Blank | | 0.000 | 02:49 |
| Subject Sample | # 1 | 0.133 | 02:52 |
| Breath Volume | | 1.832 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Air Blank | | 0.000 | 02:54 |
| Subject Sample | # 2 | 0.150 | 02:55 |
| Breath Volume | | 1.515 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Control Test | | 0.082 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Diagnostics Check | | OK | 02:57 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: PUCKETT

DATE/TIME OF TEST: 05/31/2026 03:22:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:12 |
| Air Blank | | 0.000 | 03:13 |
| Control Test | | 0.078 | 03:13 |
| Air Blank | | 0.000 | 03:13 |
| Subject Sample | # 1 | 0.204 | 03:15 |
| Breath Volume | | 1.417 | 03:15 |
| Air Blank | | 0.000 | 03:16 |
| Air Blank | | 0.000 | 03:17 |
| Subject Sample | # 2 | 0.200 | 03:20 |
| Breath Volume | | 1.828 | 03:20 |
| Air Blank | | 0.000 | 03:21 |
| Control Test | | 0.083 | 03:21 |
| Air Blank | | 0.000 | 03:22 |
| Diagnostics Check | | OK | 03:22 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: ARTHURS

DATE/TIME OF TEST: 05/31/2026 04:41:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:32 |
| Air Blank | | 0.000 | 04:32 |
| Control Test | | 0.079 | 04:32 |
| Air Blank | | 0.000 | 04:33 |
| Subject Sample | # 1 | 0.176 | 04:33 |
| Breath Volume | | 1.457 | 04:33 |
| Air Blank | | 0.000 | 04:34 |
| Air Blank | | 0.000 | 04:36 |
| Subject Sample | # 2 | 0.203 | 04:36 |
| Breath Volume | | 2.429 | 04:36 |
| Air Blank | | 0.000 | 04:37 |
| Air Blank | | 0.000 | 04:39 |
| Subject Sample | # 3 | 0.180 | 04:39 |
| Breath Volume | | 1.402 | 04:39 |
| Air Blank | | 0.000 | 04:40 |
| Control Test | | 0.077 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Diagnostics Check | | OK | 04:41 |

Data Download Date/Time: 06/24/2026 11:35

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADORNO,SAMUEL,D
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: OSBORNE,JEREMIAH
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 05/23/2026 20:32:11

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:29 | |
| Air Blank | | 0.000 | 20:29 | |
| Control Test | | 0.078 | 20:30 | |
| Air Blank | | 0.000 | 20:30 | |
| Subject Sample | # 1 | | 20:30 | Subject Test Refused |
| Breath Volume | | 0.000 | 20:30 | |
| Air Blank | | 0.000 | 20:31 | |
| Control Test | | 0.079 | 20:31 | |
| Air Blank | | 0.000 | 20:32 | |
| Diagnostics Check | | OK | 20:32 | |

Data Download Date/Time: 06/27/2026 11:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ROSS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HUGHES,ROSS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WEIHERT

DATE/TIME OF TEST: 05/24/2026 02:08:57

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:05 | |
| Air Blank | | 0.000 | 02:05 | |
| Control Test | | 0.078 | 02:06 | |
| Air Blank | | 0.000 | 02:06 | |
| Subject Sample | # 1 | | 02:07 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:07 | |
| Air Blank | | 0.000 | 02:08 | |
| Control Test | | 0.079 | 02:08 | |
| Air Blank | | 0.000 | 02:08 | |
| Diagnostics Check | | OK | 02:08 | |

Data Download Date/Time: 06/27/2026 11:22


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGINS,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: JEAN-BAPTISTE,STANLEY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 05/25/2026 04:56:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:50 |
| Air Blank | | 0.000 | 04:50 |
| Control Test | | 0.079 | 04:51 |
| Air Blank | | 0.000 | 04:51 |
| Subject Sample | # 1 | 0.065 | 04:52 |
| Breath Volume | | 2.765 | 04:52 |
| Air Blank | | 0.000 | 04:53 |
| Air Blank | | 0.000 | 04:54 |
| Subject Sample | # 2 | 0.063 | 04:55 |
| Breath Volume | | 1.843 | 04:55 |
| Air Blank | | 0.000 | 04:56 |
| Control Test | | 0.076 | 04:56 |
| Air Blank | | 0.000 | 04:56 |
| Diagnostics Check | | OK | 04:56 |

Data Download Date/Time: 06/27/2026 11:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ROSS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HUGHES,ROSS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/29/2026 00:01:44

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:57 | |
| Air Blank | | 0.000 | 23:58 | |
| Control Test | | 0.078 | 23:58 | |
| Air Blank | | 0.000 | 23:59 | |
| Subject Sample | # 1 | | 00:00 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:00 | |
| Air Blank | | 0.000 | 00:00 | |
| Control Test | | 0.078 | 00:01 | |
| Air Blank | | 0.000 | 00:01 | |
| Diagnostics Check | | OK | 00:01 | |

Data Download Date/Time: 06/27/2026 11:22


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WILLIAMS,JAVON
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: ROCHA JAENZ

DATE/TIME OF TEST: 05/29/2026 19:27:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:21 |
| Air Blank | | 0.000 | 19:21 |
| Control Test | | 0.078 | 19:21 |
| Air Blank | | 0.000 | 19:22 |
| Subject Sample | # 1 | 0.228 | 19:22 |
| Breath Volume | | 1.234 | 19:22 |
| Air Blank | | 0.000 | 19:23 |
| Air Blank | | 0.000 | 19:25 |
| Subject Sample | # 2 | 0.229 | 19:26 |
| Breath Volume | | 1.261 | 19:26 |
| Air Blank | | 0.000 | 19:26 |
| Control Test | | 0.078 | 19:27 |
| Air Blank | | 0.000 | 19:27 |
| Diagnostics Check | | OK | 19:27 |

Data Download Date/Time: 06/27/2026 11:22


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JAMES
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: MADRID

DATE/TIME OF TEST: 05/30/2026 04:42:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:36 |
| Air Blank | | 0.000 | 04:36 |
| Control Test | | 0.078 | 04:36 |
| Air Blank | | 0.000 | 04:37 |
| Subject Sample | # 1 | 0.161 | 04:37 |
| Breath Volume | | 1.539 | 04:37 |
| Air Blank | | 0.000 | 04:38 |
| Air Blank | | 0.000 | 04:40 |
| Subject Sample | # 2 | 0.166 | 04:40 |
| Breath Volume | | 1.160 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Control Test | | 0.079 | 04:41 |
| Air Blank | | 0.000 | 04:42 |
| Diagnostics Check | | OK | 04:42 |

Data Download Date/Time: 06/27/2026 11:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: KNOLES

DATE/TIME OF TEST: 05/30/2026 18:08:39

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 18:05 | |
| Air Blank | | 0.000 | 18:06 | |
| Control Test | | 0.078 | 18:06 | |
| Air Blank | | 0.000 | 18:07 | |
| Subject Sample | # 1 | | 18:07 | Subject Test Refused |
| Breath Volume | | 0.000 | 18:07 | |
| Air Blank | | 0.000 | 18:07 | |
| Control Test | | 0.078 | 18:08 | |
| Air Blank | | 0.000 | 18:08 | |
| Diagnostics Check | | OK | 18:08 | |

Data Download Date/Time: 06/27/2026 11:22

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,DARRYL,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: QUICK,LEANN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MENGULERTI

DATE/TIME OF TEST: 05/01/2026 03:27:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:24 | |
| Air Blank | | 0.000 | 03:24 | |
| Control Test | | 0.080 | 03:24 | |
| Air Blank | | 0.000 | 03:25 | |
| Subject Sample | # 1 | | 03:26 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:26 | |
| Air Blank | | 0.000 | 03:26 | |
| Control Test | | 0.080 | 03:27 | |
| Air Blank | | 0.000 | 03:27 | |
| Diagnostics Check | | OK | 03:27 | |

Data Download Date/Time: 05/21/2026 12:57

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRAW,KAELEE,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DEL CUETO,ARYEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MACKINNON

DATE/TIME OF TEST: 05/03/2026 02:26:01

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:23 | |
| Air Blank | | 0.000 | 02:23 | |
| Control Test | | 0.080 | 02:23 | |
| Air Blank | | 0.000 | 02:24 | |
| Subject Sample | # 1 | | 02:24 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:24 | |
| Air Blank | | 0.000 | 02:25 | |
| Control Test | | 0.080 | 02:25 | |
| Air Blank | | 0.000 | 02:25 | |
| Diagnostics Check | | OK | 02:26 | |

Data Download Date/Time: 05/21/2026 12:57

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: BIGGERSTAFF

DATE/TIME OF TEST: 05/06/2026 02:11:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:04 |
| Air Blank | | 0.000 | 02:04 |
| Control Test | | 0.080 | 02:04 |
| Air Blank | | 0.000 | 02:05 |
| Subject Sample | # 1 | 0.110 | 02:06 |
| Breath Volume | | 1.929 | 02:06 |
| Air Blank | | 0.000 | 02:07 |
| Air Blank | | 0.000 | 02:09 |
| Subject Sample | # 2 | 0.111 | 02:09 |
| Breath Volume | | 1.917 | 02:09 |
| Air Blank | | 0.000 | 02:10 |
| Control Test | | 0.079 | 02:10 |
| Air Blank | | 0.000 | 02:11 |
| Diagnostics Check | | OK | 02:11 |

Data Download Date/Time: 05/21/2026 12:57

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,LEANN,E
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: QUICK,LEANN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SINCIC

DATE/TIME OF TEST: 05/10/2026 03:27:18

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:19 |
| Air Blank | | 0.000 | 03:20 |
| Control Test | | 0.080 | 03:20 |
| Air Blank | | 0.000 | 03:21 |
| Subject Sample | # 1 | 0.232 | 03:22 |
| Breath Volume | | 1.363 | 03:22 |
| Air Blank | | 0.000 | 03:22 |
| Air Blank | | 0.000 | 03:24 |
| Subject Sample | # 2 | 0.227 | 03:25 |
| Breath Volume | | 1.593 | 03:25 |
| Air Blank | | 0.000 | 03:26 |
| Control Test | | 0.080 | 03:26 |
| Air Blank | | 0.000 | 03:27 |
| Diagnostics Check | | OK | 03:27 |

Data Download Date/Time: 05/21/2026 12:57

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/20/2026 03:05:10

| | | | |
|---|---|---|---|
| Diagnostics Check | Fail | 03:04 | Internal Printer Test |
| Air Blank | 0.000 | 03:05 | |

Data Download Date/Time: 05/21/2026 12:57

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARINO,JUAN CARLOS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARINO,JUAN CARLOS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/20/2026 03:15:45

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:07 | |
| Air Blank | | 0.000 | 03:07 | |
| Control Test | | 0.080 | 03:08 | |
| Air Blank | | 0.000 | 03:08 | |
| Subject Sample | # 1 | 0.265 | 03:09 | |
| Breath Volume | | 1.550 | 03:09 | |
| Air Blank | | 0.000 | 03:10 | |
| Air Blank | | 0.000 | 03:12 | |
| Subject Sample | # 2 | 0.253 | 03:13 | |
| Breath Volume | | 1.558 | 03:13 | |
| Air Blank | | 0.000 | 03:14 | |
| Control Test | | 0.080 | 03:14 | |
| Air Blank | | 0.000 | 03:15 | |
| Diagnostics Check | | Fail | 03:15 | Internal Printer Test |
| Air Blank | | 0.000 | 03:15 | |

Data Download Date/Time: 05/21/2026 12:57

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIL,JOSE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: QUICK,LEANN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 05/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: MCDERMOTT

DATE/TIME OF TEST: 05/24/2026 23:35:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:26 |
| Air Blank | | 0.000 | 23:26 |
| Control Test | | 0.081 | 23:27 |
| Air Blank | | 0.000 | 23:27 |
| Subject Sample | # 1 | 0.080 | 23:30 |
| Breath Volume | | 1.457 | 23:30 |
| Air Blank | | 0.000 | 23:31 |
| Air Blank | | 0.000 | 23:32 |
| Subject Sample | # 2 | 0.085 | 23:33 |
| Breath Volume | | 2.585 | 23:33 |
| Air Blank | | 0.000 | 23:34 |
| Control Test | | 0.080 | 23:34 |
| Air Blank | | 0.000 | 23:35 |
| Diagnostics Check | | OK | 23:35 |

Data Download Date/Time: 06/26/2026 10:39

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFENBERG,ASHTON,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WELFENBERG,ASHTON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: JENTSCH

DATE/TIME OF TEST: 05/02/2026 01:24:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:18 |
| Air Blank | | 0.000 | 01:18 |
| Control Test | | 0.080 | 01:19 |
| Air Blank | | 0.000 | 01:19 |
| Subject Sample | # 1 | 0.105 | 01:20 |
| Breath Volume | | 1.214 | 01:20 |
| Air Blank | | 0.000 | 01:20 |
| Air Blank | | 0.000 | 01:22 |
| Subject Sample | # 2 | 0.098 | 01:22 |
| Breath Volume | | 1.222 | 01:22 |
| Air Blank | | 0.000 | 01:23 |
| Control Test | | 0.078 | 01:23 |
| Air Blank | | 0.000 | 01:24 |
| Diagnostics Check | | OK | 01:24 |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRIGGERS,BRANDON,P
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: DRIGGERS,BRANDON
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 05/03/2026 03:15:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:09 |
| Air Blank | | 0.000 | 03:09 |
| Control Test | | 0.080 | 03:10 |
| Air Blank | | 0.000 | 03:10 |
| Subject Sample | # 1 | 0.044 | 03:11 |
| Breath Volume | | 3.347 | 03:11 |
| Air Blank | | 0.000 | 03:11 |
| Air Blank | | 0.000 | 03:13 |
| Subject Sample | # 2 | 0.040 | 03:14 |
| Breath Volume | | 4.691 | 03:14 |
| Air Blank | | 0.000 | 03:14 |
| Control Test | | 0.076 | 03:15 |
| Air Blank | | 0.000 | 03:15 |
| Diagnostics Check | | OK | 03:15 |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 05/03/2026 21:05:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:58 |
| Air Blank | | 0.000 | 20:59 |
| Control Test | | 0.080 | 20:59 |
| Air Blank | | 0.000 | 21:00 |
| Subject Sample | # 1 | 0.171 | 21:00 |
| Breath Volume | | 1.398 | 21:00 |
| Air Blank | | 0.000 | 21:01 |
| Air Blank | | 0.000 | 21:02 |
| Subject Sample | # 2 | 0.188 | 21:03 |
| Breath Volume | | 1.632 | 21:03 |
| Air Blank | | 0.000 | 21:04 |
| Control Test | | 0.079 | 21:04 |
| Air Blank | | 0.000 | 21:05 |
| Diagnostics Check | | OK | 21:05 |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINNERT,AJAY,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MEINNERT,AJAY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BOGER

DATE/TIME OF TEST: 05/03/2026 23:42:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:35 |
| Air Blank | | 0.000 | 23:36 |
| Control Test | | 0.080 | 23:36 |
| Air Blank | | 0.000 | 23:37 |
| Subject Sample | # 1 | 0.189 | 23:37 |
| Breath Volume | | 3.683 | 23:37 |
| Air Blank | | 0.000 | 23:38 |
| Air Blank | | 0.000 | 23:40 |
| Subject Sample | # 2 | 0.185 | 23:40 |
| Breath Volume | | 3.496 | 23:40 |
| Air Blank | | 0.000 | 23:41 |
| Control Test | | 0.077 | 23:41 |
| Air Blank | | 0.000 | 23:42 |
| Diagnostics Check | | OK | 23:42 |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,HUNTER,L
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ECKERT,HUNTER
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DELMAR

DATE/TIME OF TEST: 05/04/2026 22:52:33

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:40 | |
| Air Blank | | 0.000 | 22:41 | |
| Control Test | | 0.080 | 22:41 | |
| Air Blank | | 0.000 | 22:41 | |
| Subject Sample | # 1 | 0.152 | 22:44 | Volume Not Met |
| Breath Volume | | 0.035 | 22:44 | |
| Air Blank | | 0.000 | 22:45 | |
| Air Blank | | 0.000 | 22:47 | |
| Subject Sample | # 2 | | 22:50 | No Sample Provided |
| Breath Volume | | 0.000 | 22:50 | |
| Air Blank | | 0.000 | 22:51 | |
| Control Test | | 0.076 | 22:51 | |
| Air Blank | | 0.000 | 22:52 | |
| Diagnostics Check | | OK | 22:52 | |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,HUNTER,L
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ECKERT,HUNTER
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DELMAR

DATE/TIME OF TEST: 05/04/2026 23:19:02

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:15 | |
| Air Blank | | 0.000 | 23:15 | |
| Control Test | | 0.077 | 23:16 | |
| Air Blank | | 0.000 | 23:16 | |
| Subject Sample | # 1 | | 23:17 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:17 | |
| Air Blank | | 0.000 | 23:18 | |
| Control Test | | 0.078 | 23:18 | |
| Air Blank | | 0.000 | 23:18 | |
| Diagnostics Check | | OK | 23:19 | |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,ASHLEY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: TUCKER,ASHLEY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/05/2026 19:45:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:38 |
| Air Blank | | 0.000 | 19:39 |
| Control Test | | 0.080 | 19:39 |
| Air Blank | | 0.000 | 19:40 |
| Subject Sample | # 1 | 0.154 | 19:41 |
| Breath Volume | | 1.832 | 19:41 |
| Air Blank | | 0.000 | 19:42 |
| Air Blank | | 0.000 | 19:43 |
| Subject Sample | # 2 | 0.149 | 19:44 |
| Breath Volume | | 1.867 | 19:44 |
| Air Blank | | 0.000 | 19:44 |
| Control Test | | 0.076 | 19:45 |
| Air Blank | | 0.000 | 19:45 |
| Diagnostics Check | | OK | 19:45 |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRIGGERS,BRANDON,P
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: DRIGGERS,BRANDON
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: GILBROOK

DATE/TIME OF TEST: 05/06/2026 19:11:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:05 |
| Air Blank | | 0.000 | 19:06 |
| Control Test | | 0.080 | 19:06 |
| Air Blank | | 0.000 | 19:07 |
| Subject Sample | # 1 | 0.000 | 19:07 |
| Breath Volume | | 1.906 | 19:07 |
| Air Blank | | 0.000 | 19:08 |
| Air Blank | | 0.000 | 19:10 |
| Subject Sample | # 2 | 0.000 | 19:10 |
| Breath Volume | | 1.929 | 19:10 |
| Air Blank | | 0.000 | 19:10 |
| Control Test | | 0.077 | 19:11 |
| Air Blank | | 0.000 | 19:11 |
| Diagnostics Check | | OK | 19:11 |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFENBERG,ASHTON,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WELFENBERG,ASHTON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 05/07/2026 00:04:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:57 |
| Air Blank | | 0.000 | 23:57 |
| Control Test | | 0.080 | 23:57 |
| Air Blank | | 0.000 | 23:58 |
| Subject Sample | # 2 | 0.220 | 23:59 |
| Breath Volume | | 1.121 | 23:59 |
| Air Blank | | 0.000 | 00:00 |
| Air Blank | | 0.000 | 00:01 |
| Subject Sample | # 1 | 0.216 | 00:02 |
| Breath Volume | | 1.304 | 00:02 |
| Air Blank | | 0.000 | 00:03 |
| Control Test | | 0.076 | 00:03 |
| Air Blank | | 0.000 | 00:04 |
| Diagnostics Check | | OK | 00:04 |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICKMAN,K,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BRICKMAN,K
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 05/08/2026 01:23:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:16 |
| Air Blank | | 0.000 | 01:17 |
| Control Test | | 0.080 | 01:17 |
| Air Blank | | 0.000 | 01:18 |
| Subject Sample | # 1 | 0.000 | 01:18 |
| Breath Volume | | 2.007 | 01:18 |
| Air Blank | | 0.000 | 01:19 |
| Air Blank | | 0.000 | 01:21 |
| Subject Sample | # 2 | 0.000 | 01:21 |
| Breath Volume | | 1.343 | 01:21 |
| Air Blank | | 0.000 | 01:22 |
| Control Test | | 0.078 | 01:22 |
| Air Blank | | 0.000 | 01:23 |
| Diagnostics Check | | OK | 01:23 |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKERT,HUNTER,L
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: ECKERT,HUNTER
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BERTRAM

DATE/TIME OF TEST: 05/10/2026 02:54:36

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:43 | |
| Air Blank | | 0.000 | 02:43 | |
| Control Test | | 0.080 | 02:44 | |
| Air Blank | | 0.000 | 02:44 | |
| Subject Sample | # 1 | 0.099 | 02:47 | Volume Not Met |
| Breath Volume | | 0.085 | 02:47 | |
| Air Blank | | 0.000 | 02:48 | |
| Air Blank | | 0.000 | 02:50 | |
| Subject Sample | # 2 | | 02:52 | Subject Test Refused |
| Breath Volume | | 0.199 | 02:52 | |
| Air Blank | | 0.000 | 02:53 | |
| Control Test | | 0.076 | 02:53 | |
| Air Blank | | 0.000 | 02:54 | |
| Diagnostics Check | | OK | 02:54 | |

Data Download Date/Time: 05/15/2026 16:36

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERICO,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PERICO,JOSEPH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 05/25/2026 23:15:48

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:09 | |
| Air Blank | | 0.000 | 23:09 | |
| Control Test | | 0.077 | 23:10 | |
| Air Blank | | 0.000 | 23:10 | |
| Subject Sample | # 1 | 0.156 | 23:11 | |
| Breath Volume | | 2.351 | 23:11 | |
| Air Blank | | 0.000 | 23:12 | |
| Air Blank | | 0.000 | 23:13 | |
| Subject Sample | # 2 | 0.144 | 23:14 | |
| Breath Volume | | 1.445 | 23:14 | |
| Air Blank | | 0.000 | 23:14 | |
| Control Test | | 0.074 | 23:15 | Control Outside Tolerance |
| Air Blank | | 0.000 | 23:15 | |

Data Download Date/Time: 06/05/2026 06:08

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERICO,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: PERICO,JOSEPH
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 05/25/2026 23:41:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:35 |
| Air Blank | | 0.000 | 23:35 |
| Control Test | | 0.077 | 23:36 |
| Air Blank | | 0.000 | 23:36 |
| Subject Sample | # 1 | 0.146 | 23:36 |
| Breath Volume | | 1.226 | 23:36 |
| Air Blank | | 0.000 | 23:37 |
| Air Blank | | 0.000 | 23:39 |
| Subject Sample | # 2 | 0.143 | 23:39 |
| Breath Volume | | 2.085 | 23:39 |
| Air Blank | | 0.000 | 23:40 |
| Control Test | | 0.075 | 23:40 |
| Air Blank | | 0.000 | 23:41 |
| Diagnostics Check | | OK | 23:41 |

Data Download Date/Time: 06/05/2026 06:08

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRIGGERS,BRANDON,P
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: DRIGGERS,BRANDON
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/30/2026 00:07:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:01 |
| Air Blank | | 0.000 | 00:01 |
| Control Test | | 0.079 | 00:01 |
| Air Blank | | 0.000 | 00:02 |
| Subject Sample | # 1 | 0.123 | 00:03 |
| Breath Volume | | 2.550 | 00:03 |
| Air Blank | | 0.000 | 00:03 |
| Air Blank | | 0.000 | 00:05 |
| Subject Sample | # 2 | 0.128 | 00:06 |
| Breath Volume | | 2.605 | 00:06 |
| Air Blank | | 0.000 | 00:07 |
| Control Test | | 0.079 | 00:07 |
| Air Blank | | 0.000 | 00:07 |
| Diagnostics Check | | OK | 00:07 |

Data Download Date/Time: 06/05/2026 06:08

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOPER,JORDAN,B
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HOOPER,JORDAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 05/30/2026 03:37:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:30 |
| Air Blank | | 0.000 | 03:31 |
| Control Test | | 0.079 | 03:31 |
| Air Blank | | 0.000 | 03:32 |
| Subject Sample | # 1 | 0.169 | 03:32 |
| Breath Volume | | 2.566 | 03:32 |
| Air Blank | | 0.000 | 03:33 |
| Air Blank | | 0.000 | 03:35 |
| Subject Sample | # 2 | 0.175 | 03:36 |
| Breath Volume | | 2.304 | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Control Test | | 0.078 | 03:37 |
| Air Blank | | 0.000 | 03:37 |
| Diagnostics Check | | OK | 03:37 |

Data Download Date/Time: 06/05/2026 06:08

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 05/30/2026 04:00:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:53 |
| Air Blank | | 0.000 | 03:54 |
| Control Test | | 0.078 | 03:54 |
| Air Blank | | 0.000 | 03:55 |
| Subject Sample | # 1 | 0.186 | 03:55 |
| Breath Volume | | 2.332 | 03:55 |
| Air Blank | | 0.000 | 03:56 |
| Air Blank | | 0.000 | 03:58 |
| Subject Sample | # 2 | 0.179 | 03:58 |
| Breath Volume | | 2.367 | 03:58 |
| Air Blank | | 0.000 | 03:59 |
| Control Test | | 0.078 | 03:59 |
| Air Blank | | 0.000 | 04:00 |
| Diagnostics Check | | OK | 04:00 |

Data Download Date/Time: 06/05/2026 06:08

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINNERT,AJAY,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MEINNERT,AJAY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: DELGADO-MARQUEZ

DATE/TIME OF TEST: 05/31/2026 04:19:28

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:13 |
| Air Blank | | 0.000 | 04:13 |
| Control Test | | 0.079 | 04:13 |
| Air Blank | | 0.000 | 04:14 |
| Subject Sample | # 1 | 0.077 | 04:14 |
| Breath Volume | | 1.816 | 04:14 |
| Air Blank | | 0.000 | 04:15 |
| Air Blank | | 0.000 | 04:17 |
| Subject Sample | # 2 | 0.083 | 04:17 |
| Breath Volume | | 1.757 | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Control Test | | 0.077 | 04:18 |
| Air Blank | | 0.000 | 04:19 |
| Diagnostics Check | | OK | 04:19 |

Data Download Date/Time: 06/05/2026 06:08

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSTANCE,WADE,O
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: CONSTANCE,WADE
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CILLA

DATE/TIME OF TEST: 05/15/2026 23:02:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:55 |
| Air Blank | | 0.000 | 22:56 |
| Control Test | | 0.079 | 22:56 |
| Air Blank | | 0.000 | 22:57 |
| Subject Sample | # 1 | 0.000 | 22:58 |
| Breath Volume | | 2.015 | 22:58 |
| Air Blank | | 0.000 | 22:58 |
| Air Blank | | 0.000 | 23:00 |
| Subject Sample | # 2 | 0.000 | 23:01 |
| Breath Volume | | 2.222 | 23:01 |
| Air Blank | | 0.000 | 23:01 |
| Control Test | | 0.077 | 23:02 |
| Air Blank | | 0.000 | 23:02 |
| Diagnostics Check | | OK | 23:02 |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOCKE,BRIANNA,
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LOCKE,BRIANNA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 05/16/2026 03:08:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:01 |
| Air Blank | | 0.000 | 03:01 |
| Control Test | | 0.078 | 03:02 |
| Air Blank | | 0.000 | 03:02 |
| Subject Sample | # 1 | 0.208 | 03:03 |
| Breath Volume | | 1.589 | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Air Blank | | 0.000 | 03:06 |
| Subject Sample | # 2 | 0.217 | 03:06 |
| Breath Volume | | 1.312 | 03:06 |
| Air Blank | | 0.000 | 03:07 |
| Control Test | | 0.075 | 03:07 |
| Air Blank | | 0.000 | 03:08 |
| Diagnostics Check | | OK | 03:08 |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFENBERG,ASHTON,N
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WELFENBERG,ASHTON
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: PEDERSEN

DATE/TIME OF TEST: 05/17/2026 01:53:00

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:50 | |
| Air Blank | | 0.000 | 01:50 | |
| Control Test | | 0.079 | 01:50 | |
| Air Blank | | 0.000 | 01:51 | |
| Subject Sample | # 1 | | 01:51 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:51 | |
| Air Blank | | 0.000 | 01:52 | |
| Control Test | | 0.079 | 01:52 | |
| Air Blank | | 0.000 | 01:52 | |
| Diagnostics Check | | OK | 01:53 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/17/2026 02:25:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:19 |
| Air Blank | | 0.000 | 02:19 |
| Control Test | | 0.079 | 02:19 |
| Air Blank | | 0.000 | 02:20 |
| Subject Sample | # 1 | 0.097 | 02:20 |
| Breath Volume | | 1.570 | 02:20 |
| Air Blank | | 0.000 | 02:21 |
| Air Blank | | 0.000 | 02:23 |
| Subject Sample | # 2 | 0.095 | 02:24 |
| Breath Volume | | 1.343 | 02:24 |
| Air Blank | | 0.000 | 02:24 |
| Control Test | | 0.078 | 02:25 |
| Air Blank | | 0.000 | 02:25 |
| Diagnostics Check | | OK | 02:25 |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINNERT,AJAY,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MEINNERT,AJAY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/17/2026 04:47:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Control Test | | 0.079 | 04:41 |
| Air Blank | | 0.000 | 04:42 |
| Subject Sample | # 1 | 0.135 | 04:43 |
| Breath Volume | | 2.484 | 04:43 |
| Air Blank | | 0.000 | 04:43 |
| Air Blank | | 0.000 | 04:45 |
| Subject Sample | # 2 | 0.133 | 04:45 |
| Breath Volume | | 3.179 | 04:45 |
| Air Blank | | 0.000 | 04:46 |
| Control Test | | 0.077 | 04:47 |
| Air Blank | | 0.000 | 04:47 |
| Diagnostics Check | | OK | 04:47 |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,RYAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEAVER,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: TROWELL

DATE/TIME OF TEST: 05/20/2026 02:04:40

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 02:02 | |
| Air Blank | 0.000 | 02:03 | |
| Control Test | 0.079 | 02:03 | |
| Air Blank | | 02:04 | Purge Fail |
| Air Blank | 0.000 | 02:04 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,RYAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEAVER,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: TROWELL

DATE/TIME OF TEST: 05/20/2026 02:07:57

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 02:06 | |
| Air Blank | 0.000 | 02:06 | |
| Control Test | 0.071 | 02:06 | Control Outside Tolerance |
| Air Blank | 0.000 | 02:07 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,RYAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEAVER,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: TROWELL

DATE/TIME OF TEST: 05/20/2026 02:10:57

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 02:09 | |
| Air Blank | 0.000 | 02:09 | |
| Control Test | 0.072 | 02:09 | Control Outside Tolerance |
| Air Blank | 0.000 | 02:10 | |

Data Download Date/Time: 06/05/2026 06:02


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,RYAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEAVER,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: TROWELL

DATE/TIME OF TEST: 05/20/2026 02:13:54

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 02:12 | |
| Air Blank | 0.000 | 02:12 | |
| Control Test | 0.073 | 02:13 | Control Outside Tolerance |
| Air Blank | 0.000 | 02:13 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,RYAN,M
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: WEAVER,RYAN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: TROWELL

DATE/TIME OF TEST: 05/20/2026 02:42:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Control Test | | 0.079 | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Subject Sample | # 1 | 0.053 | 02:37 |
| Breath Volume | | 2.507 | 02:37 |
| Air Blank | | 0.000 | 02:38 |
| Air Blank | | 0.000 | 02:40 |
| Subject Sample | # 2 | 0.050 | 02:40 |
| Breath Volume | | 2.027 | 02:40 |
| Air Blank | | 0.000 | 02:41 |
| Control Test | | 0.077 | 02:41 |
| Air Blank | | 0.000 | 02:42 |
| Diagnostics Check | | OK | 02:42 |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRIGGERS,BRANDON,P
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: DRIGGERS,BRANDON
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 05/21/2026 21:11:02

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:04 | |
| Air Blank | | 0.000 | 21:04 | |
| Control Test | | 0.079 | 21:05 | |
| Air Blank | | 0.000 | 21:05 | |
| Subject Sample | # 1 | 0.145 | 21:06 | |
| Breath Volume | | 4.292 | 21:06 | |
| Air Blank | | 0.000 | 21:07 | |
| Air Blank | | 0.000 | 21:08 | |
| Subject Sample | # 2 | 0.140 | 21:09 | |
| Breath Volume | | 3.691 | 21:09 | |
| Air Blank | | 0.000 | 21:10 | |
| Control Test | | 0.073 | 21:10 | Control Outside Tolerance |
| Air Blank | | 0.000 | 21:11 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAM,JAMIE,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: LATHAM,JAMIE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: PEREZ MONTANEZ

DATE/TIME OF TEST: 05/22/2026 23:21:54

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:12 | |
| Air Blank | | 0.000 | 23:13 | |
| Control Test | | 0.078 | 23:13 | |
| Air Blank | | 0.000 | 23:13 | |
| Subject Sample | # 1 | 0.192 | 23:17 | |
| Breath Volume | | 1.406 | 23:17 | |
| Air Blank | | 0.000 | 23:18 | |
| Air Blank | | 0.000 | 23:19 | |
| Subject Sample | # 2 | 0.203 | 23:19 | |
| Breath Volume | | 1.535 | 23:19 | |
| Air Blank | | 0.000 | 23:20 | |
| Control Test | | 0.073 | 23:21 | Control Outside Tolerance |
| Air Blank | | 0.000 | 23:21 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THEIS,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: THEIS,ZACHARY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: BURFORD

DATE/TIME OF TEST: 05/23/2026 10:09:27

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 10:04 | |
| Air Blank | | 0.000 | 10:05 | |
| Control Test | | 0.078 | 10:05 | |
| Air Blank | | 0.000 | 10:06 | |
| Subject Sample | # 1 | | 10:07 | Interferent Detect |
| Breath Volume | | 2.019 | 10:07 | |
| Air Blank | | 0.000 | 10:08 | |
| Control Test | | 0.078 | 10:08 | |
| Air Blank | | 0.000 | 10:09 | |
| Diagnostics Check | | OK | 10:09 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THEIS,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: THEIS,ZACHARY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: BURFORD

DATE/TIME OF TEST: 05/23/2026 10:16:14

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 10:11 | |
| Air Blank | | 0.000 | 10:11 | |
| Control Test | | 0.076 | 10:12 | |
| Air Blank | | 0.000 | 10:12 | |
| Subject Sample | # 1 | | 10:14 | Interferent Detect |
| Breath Volume | | 2.937 | 10:14 | |
| Air Blank | | 0.000 | 10:15 | |
| Control Test | | 0.076 | 10:15 | |
| Air Blank | | 0.000 | 10:16 | |
| Diagnostics Check | | OK | 10:16 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THEIS,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: THEIS,ZACHARY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 09:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: BURFORD

DATE/TIME OF TEST: 05/23/2026 10:46:44

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 10:42 | |
| Air Blank | | 0.000 | 10:43 | |
| Control Test | | 0.075 | 10:43 | |
| Air Blank | | 0.000 | 10:44 | |
| Subject Sample | # 1 | | 10:44 | Interferent Detect |
| Breath Volume | | 1.734 | 10:44 | |
| Air Blank | | 0.000 | 10:45 | |
| Control Test | | 0.077 | 10:46 | |
| Air Blank | | 0.000 | 10:46 | |
| Diagnostics Check | | OK | 10:46 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEASMAN,MONTGOMERY,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: VEASMAN,MONTGOMERY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 05/23/2026 23:18:13

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:14 | |
| Air Blank | | 0.000 | 23:15 | |
| Control Test | | 0.079 | 23:15 | |
| Air Blank | | 0.000 | 23:15 | |
| Subject Sample | # 1 | | 23:16 | Subject Test Refused |
| Breath Volume | | 0.000 | 23:16 | |
| Air Blank | | 0.000 | 23:17 | |
| Control Test | | 0.079 | 23:17 | |
| Air Blank | | 0.000 | 23:18 | |
| Diagnostics Check | | OK | 23:18 | |

Data Download Date/Time: 06/05/2026 06:02


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/25/2026 22:19:46

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:08 | |
| Air Blank | | 0.000 | 22:09 | |
| Control Test | | 0.079 | 22:09 | |
| Air Blank | | 0.000 | 22:10 | |
| Subject Sample | # 1 | 0.256 | 22:12 | |
| Breath Volume | | 1.261 | 22:12 | |
| Air Blank | | 0.000 | 22:13 | |
| Air Blank | | 0.000 | 22:15 | |
| Subject Sample | # 2 | 0.247 | 22:17 | |
| Breath Volume | | 1.445 | 22:17 | |
| Air Blank | | 0.000 | 22:18 | |
| Control Test | | 0.073 | 22:19 | Control Outside Tolerance |
| Air Blank | | 0.000 | 22:19 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/25/2026 22:22:47

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 22:21 | |
| Air Blank | 0.000 | 22:21 | |
| Control Test | 0.074 | 22:22 | Control Outside Tolerance |
| Air Blank | 0.000 | 22:22 | |

Data Download Date/Time: 06/05/2026 06:02


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: YOUNG,ROBERT
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 05/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403549457
DRY GAS EXPIRY DATE: 01/29/2029
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 05/25/2026 22:34:48

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:25 | |
| Air Blank | | 0.000 | 22:25 | |
| Control Test | | 0.076 | 22:26 | |
| Air Blank | | 0.000 | 22:26 | |
| Subject Sample | # 1 | 0.250 | 22:28 | |
| Breath Volume | | 1.406 | 22:28 | |
| Air Blank | | 0.000 | 22:29 | |
| Air Blank | | 0.000 | 22:31 | |
| Subject Sample | # 2 | 0.256 | 22:32 | |
| Breath Volume | | 1.703 | 22:32 | |
| Air Blank | | 0.000 | 22:33 | |
| Control Test | | 0.072 | 22:34 | Control Outside Tolerance |
| Air Blank | | 0.000 | 22:34 | |

Data Download Date/Time: 06/05/2026 06:02

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARKIE,GEORGE,P
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: TONA,MICHAEL
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/29/2026 04:24:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:18 |
| Air Blank | | 0.000 | 04:18 |
| Control Test | | 0.081 | 04:18 |
| Air Blank | | 0.000 | 04:19 |
| Subject Sample | # 1 | 0.159 | 04:20 |
| Breath Volume | | 2.734 | 04:20 |
| Air Blank | | 0.000 | 04:20 |
| Air Blank | | 0.000 | 04:22 |
| Subject Sample | # 2 | 0.155 | 04:23 |
| Breath Volume | | 2.906 | 04:23 |
| Air Blank | | 0.000 | 04:23 |
| Control Test | | 0.080 | 04:24 |
| Air Blank | | 0.000 | 04:24 |
| Diagnostics Check | | OK | 04:24 |

Data Download Date/Time: 06/29/2026 10:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HASWELL,AUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: DEANGELIS

DATE/TIME OF TEST: 05/29/2026 04:43:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:37 |
| Air Blank | | 0.000 | 04:37 |
| Control Test | | 0.081 | 04:38 |
| Air Blank | | 0.000 | 04:38 |
| Subject Sample | # 1 | 0.142 | 04:39 |
| Breath Volume | | 3.378 | 04:39 |
| Air Blank | | 0.000 | 04:39 |
| Air Blank | | 0.000 | 04:41 |
| Subject Sample | # 2 | 0.142 | 04:42 |
| Breath Volume | | 3.382 | 04:42 |
| Air Blank | | 0.000 | 04:42 |
| Control Test | | 0.081 | 04:42 |
| Air Blank | | 0.000 | 04:43 |
| Diagnostics Check | | OK | 04:43 |

Data Download Date/Time: 06/29/2026 10:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007362

INSTRUMENT REGISTERED TO: PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 05/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: CHANCEY

DATE/TIME OF TEST: 05/31/2026 02:11:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:05 |
| Air Blank | | 0.000 | 02:05 |
| Control Test | | 0.079 | 02:06 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 1 | 0.226 | 02:07 |
| Breath Volume | | 2.363 | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Air Blank | | 0.000 | 02:09 |
| Subject Sample | # 2 | 0.225 | 02:09 |
| Breath Volume | | 2.363 | 02:09 |
| Air Blank | | 0.000 | 02:10 |
| Control Test | | 0.079 | 02:10 |
| Air Blank | | 0.000 | 02:11 |
| Diagnostics Check | | OK | 02:11 |

Data Download Date/Time: 06/29/2026 10:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RIGGINS,BRENDAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: PARISI

DATE/TIME OF TEST: 05/05/2026 23:24:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:15 |
| Air Blank | | 0.000 | 23:15 |
| Control Test | | 0.081 | 23:15 |
| Air Blank | | 0.000 | 23:16 |
| Subject Sample | # 1 | 0.217 | 23:16 |
| Breath Volume | | 1.613 | 23:16 |
| Air Blank | | 0.000 | 23:17 |
| Air Blank | | 0.000 | 23:19 |
| Subject Sample | # 2 | 0.185 | 23:19 |
| Breath Volume | | 2.031 | 23:19 |
| Air Blank | | 0.000 | 23:20 |
| Air Blank | | 0.000 | 23:22 |
| Subject Sample | # 3 | 0.186 | 23:22 |
| Breath Volume | | 1.609 | 23:22 |
| Air Blank | | 0.000 | 23:23 |
| Control Test | | 0.080 | 23:23 |
| Air Blank | | 0.000 | 23:24 |
| Diagnostics Check | | OK | 23:24 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,BRENDAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MARTINEZ,CHEYLA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: WATT

DATE/TIME OF TEST: 05/09/2026 04:07:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:00 |
| Air Blank | | 0.000 | 04:01 |
| Control Test | | 0.082 | 04:01 |
| Air Blank | | 0.000 | 04:02 |
| Subject Sample | # 1 | 0.000 | 04:02 |
| Breath Volume | | 2.171 | 04:02 |
| Air Blank | | 0.000 | 04:03 |
| Air Blank | | 0.000 | 04:05 |
| Subject Sample | # 2 | 0.000 | 04:05 |
| Breath Volume | | 2.023 | 04:05 |
| Air Blank | | 0.000 | 04:06 |
| Control Test | | 0.081 | 04:06 |
| Air Blank | | 0.000 | 04:06 |
| Diagnostics Check | | OK | 04:07 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,MIKO,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BOESDORFER,CURTIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: ADAMS 3RD

DATE/TIME OF TEST: 05/09/2026 06:33:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:27 |
| Air Blank | | 0.000 | 06:28 |
| Control Test | | 0.082 | 06:28 |
| Air Blank | | 0.000 | 06:29 |
| Subject Sample | # 1 | 0.154 | 06:29 |
| Breath Volume | | 1.714 | 06:29 |
| Air Blank | | 0.000 | 06:30 |
| Air Blank | | 0.000 | 06:31 |
| Subject Sample | # 2 | 0.152 | 06:32 |
| Breath Volume | | 1.804 | 06:32 |
| Air Blank | | 0.000 | 06:33 |
| Control Test | | 0.080 | 06:33 |
| Air Blank | | 0.000 | 06:33 |
| Diagnostics Check | | OK | 06:33 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUMAKER,THOMAS,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: PARKER,DARBY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RAMIREZ VELAZQUEZ

DATE/TIME OF TEST: 05/09/2026 19:25:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:18 |
| Air Blank | | 0.000 | 19:18 |
| Control Test | | 0.081 | 19:19 |
| Air Blank | | 0.000 | 19:19 |
| Subject Sample | # 1 | 0.203 | 19:20 |
| Breath Volume | | 1.257 | 19:20 |
| Air Blank | | 0.000 | 19:20 |
| Air Blank | | 0.000 | 19:22 |
| Subject Sample | # 2 | 0.195 | 19:23 |
| Breath Volume | | 1.527 | 19:23 |
| Air Blank | | 0.000 | 19:24 |
| Control Test | | 0.080 | 19:24 |
| Air Blank | | 0.000 | 19:25 |
| Diagnostics Check | | OK | 19:25 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRONSON,TAYLOR,K
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: GALLAGHER,KYLEIGH
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: DIPAOLO

DATE/TIME OF TEST: 05/10/2026 00:35:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:28 |
| Air Blank | | 0.000 | 00:28 |
| Control Test | | 0.082 | 00:29 |
| Air Blank | | 0.000 | 00:29 |
| Subject Sample | # 1 | 0.287 | 00:30 |
| Breath Volume | | 1.671 | 00:30 |
| Air Blank | | 0.000 | 00:31 |
| Air Blank | | 0.000 | 00:32 |
| Subject Sample | # 2 | 0.270 | 00:33 |
| Breath Volume | | 2.035 | 00:33 |
| Air Blank | | 0.000 | 00:34 |
| Control Test | | 0.080 | 00:34 |
| Air Blank | | 0.000 | 00:34 |
| Diagnostics Check | | OK | 00:35 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,LETOSHUA,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREENE,LETOSHUA
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: R426265000000

DATE/TIME OF TEST: 05/11/2026 02:56:10

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:44 | |
| Air Blank | | 0.000 | 02:44 | |
| Control Test | | 0.081 | 02:44 | |
| Air Blank | | 0.000 | 02:45 | |
| Subject Sample | # 1 | 0.117 | 02:48 | Volume Not Met |
| Breath Volume | | 0.238 | 02:48 | |
| Air Blank | | 0.000 | 02:49 | |
| Air Blank | | 0.000 | 02:51 | |
| Subject Sample | # 2 | 0.088 | 02:54 | Volume Not Met |
| Breath Volume | | 0.226 | 02:54 | |
| Air Blank | | 0.000 | 02:55 | |
| Control Test | | 0.080 | 02:55 | |
| Air Blank | | 0.000 | 02:56 | |
| Diagnostics Check | | OK | 02:56 | |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,LETOSHUA,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREENE,LETOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: R426265000000

DATE/TIME OF TEST: 05/11/2026 03:10:29

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:58 | |
| Air Blank | | 0.000 | 02:59 | |
| Control Test | | 0.078 | 02:59 | |
| Air Blank | | 0.000 | 03:00 | |
| Subject Sample | # 1 | 0.118 | 03:03 | Volume Not Met |
| Breath Volume | | 0.050 | 03:03 | |
| Air Blank | | 0.000 | 03:03 | |
| Air Blank | | 0.000 | 03:05 | |
| Subject Sample | # 2 | 0.108 | 03:08 | Volume Not Met |
| Breath Volume | | 0.054 | 03:08 | |
| Air Blank | | 0.000 | 03:09 | |
| Control Test | | 0.079 | 03:09 | |
| Air Blank | | 0.000 | 03:10 | |
| Diagnostics Check | | OK | 03:10 | |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: JAROUSSE,DANIELLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 08:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:07

SUBJECT'S LAST NAME: LOW

DATE/TIME OF TEST: 05/11/2026 09:34:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 09:27 |
| Air Blank | | 0.000 | 09:28 |
| Control Test | | 0.080 | 09:28 |
| Air Blank | | 0.000 | 09:28 |
| Subject Sample | # 1 | 0.206 | 09:29 |
| Breath Volume | | 2.441 | 09:29 |
| Air Blank | | 0.000 | 09:30 |
| Air Blank | | 0.000 | 09:31 |
| Subject Sample | # 2 | 0.203 | 09:32 |
| Breath Volume | | 1.253 | 09:32 |
| Air Blank | | 0.000 | 09:33 |
| Control Test | | 0.080 | 09:33 |
| Air Blank | | 0.000 | 09:34 |
| Diagnostics Check | | OK | 09:34 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FREEMAN,MATTHEW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: HIOTT

DATE/TIME OF TEST: 05/12/2026 19:10:28

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:03 |
| Air Blank | | 0.000 | 19:03 |
| Control Test | | 0.080 | 19:04 |
| Air Blank | | 0.000 | 19:04 |
| Subject Sample | # 1 | 0.143 | 19:05 |
| Breath Volume | | 2.984 | 19:05 |
| Air Blank | | 0.000 | 19:06 |
| Air Blank | | 0.000 | 19:08 |
| Subject Sample | # 2 | 0.139 | 19:08 |
| Breath Volume | | 3.046 | 19:08 |
| Air Blank | | 0.000 | 19:09 |
| Control Test | | 0.080 | 19:09 |
| Air Blank | | 0.000 | 19:10 |
| Diagnostics Check | | OK | 19:10 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,MYREON,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PIERCE,MYREON
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: TATUM

DATE/TIME OF TEST: 05/12/2026 22:43:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:36 |
| Air Blank | | 0.000 | 22:37 |
| Control Test | | 0.081 | 22:37 |
| Air Blank | | 0.000 | 22:38 |
| Subject Sample | # 1 | 0.402 | 22:39 |
| Breath Volume | | 2.625 | 22:39 |
| Air Blank | | 0.000 | 22:39 |
| Air Blank | | 0.000 | 22:41 |
| Subject Sample | # 2 | 0.386 | 22:42 |
| Breath Volume | | 2.273 | 22:42 |
| Air Blank | | 0.000 | 22:43 |
| Control Test | | 0.079 | 22:43 |
| Air Blank | | 0.000 | 22:43 |
| Diagnostics Check | | OK | 22:43 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRONSON,TAYLOR,K
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRONSON,TAYLOR
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: RADBURN

DATE/TIME OF TEST: 05/13/2026 19:28:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:22 |
| Air Blank | | 0.000 | 19:22 |
| Control Test | | 0.081 | 19:22 |
| Air Blank | | 0.000 | 19:23 |
| Subject Sample | # 1 | 0.250 | 19:24 |
| Breath Volume | | 2.496 | 19:24 |
| Air Blank | | 0.000 | 19:25 |
| Air Blank | | 0.000 | 19:26 |
| Subject Sample | # 2 | 0.261 | 19:27 |
| Breath Volume | | 2.574 | 19:27 |
| Air Blank | | 0.000 | 19:27 |
| Control Test | | 0.080 | 19:28 |
| Air Blank | | 0.000 | 19:28 |
| Diagnostics Check | | OK | 19:28 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUCH,DOMANIK,G
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: COLIN,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/15/2026 03:18:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:12 |
| Air Blank | | 0.000 | 03:12 |
| Control Test | | 0.081 | 03:13 |
| Air Blank | | 0.000 | 03:13 |
| Subject Sample | # 1 | 0.105 | 03:14 |
| Breath Volume | | 2.570 | 03:14 |
| Air Blank | | 0.000 | 03:14 |
| Air Blank | | 0.000 | 03:16 |
| Subject Sample | # 2 | 0.099 | 03:17 |
| Breath Volume | | 2.218 | 03:17 |
| Air Blank | | 0.000 | 03:17 |
| Control Test | | 0.081 | 03:18 |
| Air Blank | | 0.000 | 03:18 |
| Diagnostics Check | | OK | 03:18 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,MYREON,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PIERCE,MYREON
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/17/2026 20:40:27

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:36 | |
| Air Blank | | 0.000 | 20:37 | |
| Control Test | | 0.082 | 20:37 | |
| Air Blank | | 0.000 | 20:38 | |
| Subject Sample | # 1 | | 20:39 | Subject Test Refused |
| Breath Volume | | 0.000 | 20:39 | |
| Air Blank | | 0.000 | 20:39 | |
| Control Test | | 0.081 | 20:39 | |
| Air Blank | | 0.000 | 20:40 | |
| Diagnostics Check | | OK | 20:40 | |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BROWNETT

DATE/TIME OF TEST: 05/23/2026 00:04:07

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:59 | |
| Air Blank | | 0.000 | 23:59 | |
| Control Test | | 0.082 | 23:59 | |
| Air Blank | | 0.000 | 00:00 | |
| Subject Sample | # 1 | 0.000 | 00:00 | |
| Breath Volume | | 3.531 | 00:00 | |
| Air Blank | | 0.000 | 00:01 | |
| Air Blank | | 0.000 | 00:03 | |
| Subject Sample | # 2 | | 00:03 | RFI Detect |
| Breath Volume | | 0.000 | 00:03 | |
| Air Blank | | | 00:03 | RFI Detect |
| Air Blank | | 0.000 | 00:04 | |

Data Download Date/Time: 05/29/2026 20:51


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BROWNETT

DATE/TIME OF TEST: 05/23/2026 00:11:47

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:05 |
| Air Blank | | 0.000 | 00:06 |
| Control Test | | 0.081 | 00:06 |
| Air Blank | | 0.000 | 00:07 |
| Subject Sample | # 1 | 0.000 | 00:07 |
| Breath Volume | | 3.574 | 00:07 |
| Air Blank | | 0.000 | 00:08 |
| Air Blank | | 0.000 | 00:10 |
| Subject Sample | # 2 | 0.000 | 00:10 |
| Breath Volume | | 3.625 | 00:10 |
| Air Blank | | 0.000 | 00:10 |
| Control Test | | 0.080 | 00:11 |
| Air Blank | | 0.000 | 00:11 |
| Diagnostics Check | | OK | 00:11 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUMAKER,THOMAS,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SHUMAKER,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 05/23/2026 23:40:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:33 |
| Air Blank | | 0.000 | 23:34 |
| Control Test | | 0.082 | 23:34 |
| Air Blank | | 0.000 | 23:35 |
| Subject Sample | # 1 | 0.171 | 23:35 |
| Breath Volume | | 1.375 | 23:35 |
| Air Blank | | 0.000 | 23:36 |
| Air Blank | | 0.000 | 23:38 |
| Subject Sample | # 2 | 0.162 | 23:38 |
| Breath Volume | | 2.039 | 23:38 |
| Air Blank | | 0.000 | 23:39 |
| Control Test | | 0.080 | 23:40 |
| Air Blank | | 0.000 | 23:40 |
| Diagnostics Check | | OK | 23:40 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: RODISH

DATE/TIME OF TEST: 05/24/2026 00:10:28

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:07 | |
| Air Blank | | 0.000 | 00:07 | |
| Control Test | | 0.081 | 00:08 | |
| Air Blank | | 0.000 | 00:08 | |
| Subject Sample | # 1 | | 00:09 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:09 | |
| Air Blank | | 0.000 | 00:09 | |
| Control Test | | 0.081 | 00:09 | |
| Air Blank | | 0.000 | 00:10 | |
| Diagnostics Check | | OK | 00:10 | |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAROUSSE,DANIELLE,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WELLS,JOSEPH
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: UHLAND

DATE/TIME OF TEST: 05/25/2026 05:21:27

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:14 |
| Air Blank | | 0.000 | 05:14 |
| Control Test | | 0.082 | 05:15 |
| Air Blank | | 0.000 | 05:15 |
| Subject Sample | # 1 | 0.190 | 05:16 |
| Breath Volume | | 2.003 | 05:16 |
| Air Blank | | 0.000 | 05:17 |
| Air Blank | | 0.000 | 05:18 |
| Subject Sample | # 2 | 0.191 | 05:19 |
| Breath Volume | | 1.433 | 05:19 |
| Air Blank | | 0.000 | 05:20 |
| Control Test | | 0.081 | 05:20 |
| Air Blank | | 0.000 | 05:21 |
| Diagnostics Check | | OK | 05:21 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SIENA,MICHAEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CHAMPINE

DATE/TIME OF TEST: 05/26/2026 01:50:23

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:43 |
| Air Blank | | 0.000 | 01:43 |
| Control Test | | 0.081 | 01:44 |
| Air Blank | | 0.000 | 01:44 |
| Subject Sample | # 1 | 0.104 | 01:45 |
| Breath Volume | | 1.824 | 01:45 |
| Air Blank | | 0.000 | 01:46 |
| Air Blank | | 0.000 | 01:48 |
| Subject Sample | # 2 | 0.100 | 01:48 |
| Breath Volume | | 1.449 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Control Test | | 0.080 | 01:49 |
| Air Blank | | 0.000 | 01:50 |
| Diagnostics Check | | OK | 01:50 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARROLL,JERRY,F
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CARROLL,JERRY
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MYSLIWIEC

DATE/TIME OF TEST: 05/28/2026 02:39:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:31 |
| Air Blank | | 0.000 | 02:31 |
| Control Test | | 0.081 | 02:31 |
| Air Blank | | 0.000 | 02:32 |
| Subject Sample | # 1 | 0.182 | 02:34 |
| Breath Volume | | 1.558 | 02:34 |
| Air Blank | | 0.000 | 02:35 |
| Air Blank | | 0.000 | 02:37 |
| Subject Sample | # 2 | 0.180 | 02:38 |
| Breath Volume | | 2.250 | 02:38 |
| Air Blank | | 0.000 | 02:38 |
| Control Test | | 0.079 | 02:39 |
| Air Blank | | 0.000 | 02:39 |
| Diagnostics Check | | OK | 02:39 |

Data Download Date/Time: 05/29/2026 20:51

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,N
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GARCIA HERNANDEZ

DATE/TIME OF TEST: 05/07/2026 02:21:24

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:14 | |
| Air Blank | | 0.000 | 02:15 | |
| Control Test | | 0.081 | 02:15 | |
| Air Blank | | 0.000 | 02:15 | |
| Subject Sample | # 1 | 0.232 | 02:16 | |
| Breath Volume | | 1.218 | 02:16 | |
| Air Blank | | 0.000 | 02:17 | |
| Air Blank | | 0.000 | 02:19 | |
| Subject Sample | # 2 | 0.226 | 02:19 | |
| Breath Volume | | 2.101 | 02:19 | |
| Air Blank | | 0.000 | 02:20 | |
| Control Test | | 0.069 | 02:20 | Control Outside Tolerance |
| Air Blank | | 0.000 | 02:21 | |

Data Download Date/Time: 05/30/2026 18:41

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,OSCAR,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROQUE,OSCAR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:54

SUBJECT'S LAST NAME: MENDOZA ISAGUIRRE

DATE/TIME OF TEST: 05/07/2026 13:22:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 13:15 |
| Air Blank | | 0.000 | 13:15 |
| Control Test | | 0.080 | 13:15 |
| Air Blank | | 0.000 | 13:16 |
| Subject Sample | # 1 | 0.165 | 13:17 |
| Breath Volume | | 1.597 | 13:17 |
| Air Blank | | 0.000 | 13:18 |
| Air Blank | | 0.000 | 13:19 |
| Subject Sample | # 2 | 0.159 | 13:20 |
| Breath Volume | | 2.851 | 13:20 |
| Air Blank | | 0.000 | 13:21 |
| Control Test | | 0.079 | 13:21 |
| Air Blank | | 0.000 | 13:22 |
| Diagnostics Check | | OK | 13:22 |

Data Download Date/Time: 05/30/2026 18:41

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,NIVIA,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SERRANO GONZALEZ

DATE/TIME OF TEST: 05/11/2026 02:05:13

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:02 | |
| Air Blank | | 0.000 | 02:02 | |
| Control Test | | 0.081 | 02:03 | |
| Air Blank | | 0.000 | 02:03 | |
| Subject Sample | # 1 | | 02:03 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:03 | |
| Air Blank | | 0.000 | 02:04 | |
| Control Test | | 0.081 | 02:04 | |
| Air Blank | | 0.000 | 02:05 | |
| Diagnostics Check | | OK | 02:05 | |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: IRBY,D
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: WOODALL

DATE/TIME OF TEST: 05/16/2026 03:17:26

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:08 | |
| Air Blank | | 0.000 | 03:09 | |
| Control Test | | 0.081 | 03:09 | |
| Air Blank | | 0.000 | 03:10 | |
| Subject Sample | # 1 | | 03:13 | No Sample Provided |
| Breath Volume | | 0.000 | 03:13 | |
| Air Blank | | 0.000 | 03:13 | |
| Air Blank | | 0.000 | 03:15 | |
| Subject Sample | # 2 | | 03:16 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:16 | |
| Air Blank | | 0.000 | 03:16 | |
| Control Test | | 0.081 | 03:16 | |
| Air Blank | | 0.000 | 03:17 | |
| Diagnostics Check | | OK | 03:17 | |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ACKERMAN,ELVIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: JUAN JUAN

DATE/TIME OF TEST: 05/17/2026 00:42:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:35 |
| Air Blank | | 0.000 | 00:35 |
| Control Test | | 0.081 | 00:36 |
| Air Blank | | 0.000 | 00:36 |
| Subject Sample | # 1 | 0.065 | 00:37 |
| Breath Volume | | 3.343 | 00:37 |
| Air Blank | | 0.000 | 00:38 |
| Air Blank | | 0.000 | 00:39 |
| Subject Sample | # 2 | 0.064 | 00:40 |
| Breath Volume | | 3.355 | 00:40 |
| Air Blank | | 0.000 | 00:41 |
| Control Test | | 0.080 | 00:41 |
| Air Blank | | 0.000 | 00:42 |
| Diagnostics Check | | OK | 00:42 |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,ELVIS,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GARCIA GOMEZ

DATE/TIME OF TEST: 05/17/2026 01:42:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:35 |
| Air Blank | | 0.000 | 01:35 |
| Control Test | | 0.081 | 01:35 |
| Air Blank | | 0.000 | 01:36 |
| Subject Sample | # 1 | 0.270 | 01:37 |
| Breath Volume | | 1.535 | 01:37 |
| Air Blank | | 0.000 | 01:37 |
| Air Blank | | 0.000 | 01:39 |
| Subject Sample | # 2 | 0.260 | 01:40 |
| Breath Volume | | 1.347 | 01:40 |
| Air Blank | | 0.000 | 01:41 |
| Control Test | | 0.079 | 01:41 |
| Air Blank | | 0.000 | 01:41 |
| Diagnostics Check | | OK | 01:42 |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: INNOCENT,L,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: HILSON,R
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: FONSECA CRUZ

DATE/TIME OF TEST: 05/17/2026 04:23:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:20 | |
| Air Blank | | 0.000 | 04:21 | |
| Control Test | | 0.080 | 04:21 | |
| Air Blank | | 0.000 | 04:22 | |
| Subject Sample | # 1 | | 04:22 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:22 | |
| Air Blank | | 0.000 | 04:22 | |
| Control Test | | 0.080 | 04:23 | |
| Air Blank | | 0.000 | 04:23 | |
| Diagnostics Check | | OK | 04:23 | |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANECA,DANILO,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GONZALEZ,JHAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: MATEOFRANCISCO

DATE/TIME OF TEST: 05/17/2026 22:39:25

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:33 | |
| Air Blank | | 0.000 | 22:33 | |
| Control Test | | 0.081 | 22:34 | |
| Air Blank | | 0.000 | 22:34 | |
| Subject Sample | # 1 | 0.304 | 22:35 | |
| Breath Volume | | 3.003 | 22:35 | |
| Air Blank | | 0.000 | 22:36 | |
| Air Blank | | 0.000 | 22:37 | |
| Subject Sample | # 2 | 0.295 | 22:38 | |
| Breath Volume | | 2.898 | 22:38 | |
| Air Blank | | | 22:38 | Purge Fail |
| Air Blank | | 0.000 | 22:39 | |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANECA,DANILO,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: HASMI,CHRISTIAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: PAVON

DATE/TIME OF TEST: 05/18/2026 00:16:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:09 |
| Air Blank | | 0.000 | 00:10 |
| Control Test | | 0.080 | 00:10 |
| Air Blank | | 0.000 | 00:11 |
| Subject Sample | # 1 | 0.241 | 00:11 |
| Breath Volume | | 3.761 | 00:11 |
| Air Blank | | 0.000 | 00:12 |
| Air Blank | | 0.000 | 00:14 |
| Subject Sample | # 2 | 0.244 | 00:14 |
| Breath Volume | | 3.726 | 00:14 |
| Air Blank | | 0.000 | 00:15 |
| Control Test | | 0.078 | 00:16 |
| Air Blank | | 0.000 | 00:16 |
| Diagnostics Check | | OK | 00:16 |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IRBY,DOMINIC,R
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,SEBASTIAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 05/22/2026 04:40:54

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:29 | |
| Air Blank | | 0.000 | 04:29 | |
| Control Test | | 0.081 | 04:29 | |
| Air Blank | | 0.000 | 04:30 | |
| Subject Sample | # 1 | 0.195 | 04:33 | Volume Not Met |
| Breath Volume | | 0.066 | 04:33 | |
| Air Blank | | 0.000 | 04:34 | |
| Air Blank | | 0.000 | 04:36 | |
| Subject Sample | # 2 | | 04:39 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:39 | |
| Air Blank | | 0.000 | 04:39 | |
| Control Test | | 0.077 | 04:40 | |
| Air Blank | | 0.000 | 04:40 | |
| Diagnostics Check | | OK | 04:40 | |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,OSCAR,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROQUE,OSCAR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: TZUNUN

DATE/TIME OF TEST: 05/23/2026 23:04:13

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:57 |
| Air Blank | | 0.000 | 22:57 |
| Control Test | | 0.080 | 22:58 |
| Air Blank | | 0.000 | 22:58 |
| Subject Sample | # 1 | 0.253 | 22:59 |
| Breath Volume | | 3.332 | 22:59 |
| Air Blank | | 0.000 | 23:00 |
| Air Blank | | 0.000 | 23:01 |
| Subject Sample | # 2 | 0.256 | 23:02 |
| Breath Volume | | 3.000 | 23:02 |
| Air Blank | | 0.000 | 23:03 |
| Control Test | | 0.078 | 23:03 |
| Air Blank | | 0.000 | 23:04 |
| Diagnostics Check | | OK | 23:04 |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,OSCAR,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROQUE,OSCAR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: MENDEZ LUCAS

DATE/TIME OF TEST: 05/24/2026 10:59:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 10:50 | |
| Air Blank | | 0.000 | 10:50 | |
| Control Test | | 0.081 | 10:51 | |
| Air Blank | | 0.000 | 10:51 | |
| Subject Sample | # 1 | | 10:54 | No Sample Provided |
| Breath Volume | | 0.000 | 10:54 | |
| Air Blank | | 0.000 | 10:55 | |
| Air Blank | | 0.000 | 10:57 | |
| Subject Sample | # 2 | | 10:58 | Subject Test Refused |
| Breath Volume | | 0.000 | 10:58 | |
| Air Blank | | 0.000 | 10:58 | |
| Control Test | | 0.080 | 10:59 | |
| Air Blank | | 0.000 | 10:59 | |
| Diagnostics Check | | OK | 10:59 | |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: CHRISTIAN,HASSMI
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MARTIN GARCIA

DATE/TIME OF TEST: 05/28/2026 00:31:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:24 |
| Air Blank | | 0.000 | 00:25 |
| Control Test | | 0.080 | 00:25 |
| Air Blank | | 0.000 | 00:26 |
| Subject Sample | # 1 | 0.170 | 00:26 |
| Breath Volume | | 2.574 | 00:26 |
| Air Blank | | 0.000 | 00:27 |
| Air Blank | | 0.000 | 00:29 |
| Subject Sample | # 2 | 0.178 | 00:29 |
| Breath Volume | | 2.476 | 00:29 |
| Air Blank | | 0.000 | 00:30 |
| Control Test | | 0.080 | 00:31 |
| Air Blank | | 0.000 | 00:31 |
| Diagnostics Check | | OK | 00:31 |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007380

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIROGA,DIEGO,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: QUIROGA,DIEGO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: LOPEZ CASTRO

DATE/TIME OF TEST: 05/30/2026 02:00:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:53 |
| Air Blank | | 0.000 | 01:54 |
| Control Test | | 0.081 | 01:54 |
| Air Blank | | 0.000 | 01:55 |
| Subject Sample | # 1 | 0.179 | 01:55 |
| Breath Volume | | 1.531 | 01:55 |
| Air Blank | | 0.000 | 01:56 |
| Air Blank | | 0.000 | 01:58 |
| Subject Sample | # 2 | 0.183 | 01:58 |
| Breath Volume | | 1.113 | 01:58 |
| Air Blank | | 0.000 | 01:59 |
| Control Test | | 0.079 | 01:59 |
| Air Blank | | 0.000 | 02:00 |
| Diagnostics Check | | OK | 02:00 |

Data Download Date/Time: 05/30/2026 18:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,MARK,A
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: FIGUEROA,ALVARO
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 05/20/2026 00:28:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:21 |
| Air Blank | | 0.000 | 00:21 |
| Control Test | | 0.080 | 00:22 |
| Air Blank | | 0.000 | 00:22 |
| Subject Sample | # 1 | 0.237 | 00:23 |
| Breath Volume | | 3.246 | 00:23 |
| Air Blank | | 0.000 | 00:23 |
| Air Blank | | 0.000 | 00:25 |
| Subject Sample | # 2 | 0.239 | 00:26 |
| Breath Volume | | 4.484 | 00:26 |
| Air Blank | | 0.000 | 00:27 |
| Control Test | | 0.078 | 00:27 |
| Air Blank | | 0.000 | 00:28 |
| Diagnostics Check | | OK | 00:28 |

Data Download Date/Time: 06/30/2026 12:45

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFEK,GENEVA,K
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: FIGUREROA,ALVARO
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: BRONSON

DATE/TIME OF TEST: 05/22/2026 19:18:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:12 |
| Air Blank | | 0.000 | 19:13 |
| Control Test | | 0.080 | 19:13 |
| Air Blank | | 0.000 | 19:13 |
| Subject Sample | # 1 | 0.204 | 19:14 |
| Breath Volume | | 2.089 | 19:14 |
| Air Blank | | 0.000 | 19:15 |
| Air Blank | | 0.000 | 19:16 |
| Subject Sample | # 2 | 0.186 | 19:17 |
| Breath Volume | | 1.371 | 19:17 |
| Air Blank | | 0.000 | 19:18 |
| Control Test | | 0.079 | 19:18 |
| Air Blank | | 0.000 | 19:18 |
| Diagnostics Check | | OK | 19:18 |

Data Download Date/Time: 06/30/2026 12:45

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTRUNK,TRENT,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SISTRUNK,TRENT
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/25/2026 15:36:27

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:30 |
| Air Blank | | 0.000 | 15:30 |
| Control Test | | 0.080 | 15:31 |
| Air Blank | | 0.000 | 15:31 |
| Subject Sample | # 1 | 0.160 | 15:32 |
| Breath Volume | | 1.417 | 15:32 |
| Air Blank | | 0.000 | 15:32 |
| Air Blank | | 0.000 | 15:34 |
| Subject Sample | # 2 | 0.161 | 15:34 |
| Breath Volume | | 2.511 | 15:34 |
| Air Blank | | 0.000 | 15:35 |
| Control Test | | 0.079 | 15:35 |
| Air Blank | | 0.000 | 15:36 |
| Diagnostics Check | | OK | 15:36 |

Data Download Date/Time: 06/30/2026 12:45

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BARNETT

DATE/TIME OF TEST: 05/30/2026 04:01:08

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:58 | |
| Air Blank | | 0.000 | 03:58 | |
| Control Test | | 0.080 | 03:59 | |
| Air Blank | | 0.000 | 03:59 | |
| Subject Sample | # 1 | | 03:59 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:59 | |
| Air Blank | | 0.000 | 04:00 | |
| Control Test | | 0.080 | 04:00 | |
| Air Blank | | 0.000 | 04:01 | |
| Diagnostics Check | | OK | 04:01 | |

Data Download Date/Time: 06/30/2026 12:45

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: BONEY

DATE/TIME OF TEST: 05/30/2026 05:07:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:01 |
| Air Blank | | 0.000 | 05:02 |
| Control Test | | 0.079 | 05:02 |
| Air Blank | | 0.000 | 05:03 |
| Subject Sample | # 1 | 0.173 | 05:03 |
| Breath Volume | | 2.613 | 05:03 |
| Air Blank | | 0.000 | 05:04 |
| Air Blank | | 0.000 | 05:06 |
| Subject Sample | # 2 | 0.174 | 05:06 |
| Breath Volume | | 1.531 | 05:06 |
| Air Blank | | 0.000 | 05:06 |
| Control Test | | 0.079 | 05:07 |
| Air Blank | | 0.000 | 05:07 |
| Diagnostics Check | | OK | 05:07 |

Data Download Date/Time: 06/30/2026 12:45

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: BRANNON

DATE/TIME OF TEST: 05/31/2026 02:57:16

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:54 | |
| Air Blank | | 0.000 | 02:54 | |
| Control Test | | 0.080 | 02:55 | |
| Air Blank | | 0.000 | 02:55 | |
| Subject Sample | # 1 | | 02:55 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:55 | |
| Air Blank | | 0.000 | 02:56 | |
| Control Test | | 0.080 | 02:56 | |
| Air Blank | | 0.000 | 02:57 | |
| Diagnostics Check | | OK | 02:57 | |

Data Download Date/Time: 06/30/2026 12:45

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FISHER,RYAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: HOSSAIN

DATE/TIME OF TEST: 05/31/2026 05:59:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:52 |
| Air Blank | | 0.000 | 05:53 |
| Control Test | | 0.079 | 05:53 |
| Air Blank | | 0.000 | 05:54 |
| Subject Sample | # 1 | 0.055 | 05:54 |
| Breath Volume | | 1.503 | 05:54 |
| Air Blank | | 0.000 | 05:55 |
| Air Blank | | 0.000 | 05:57 |
| Subject Sample | # 2 | 0.055 | 05:58 |
| Breath Volume | | 1.652 | 05:58 |
| Air Blank | | 0.000 | 05:58 |
| Control Test | | 0.078 | 05:58 |
| Air Blank | | 0.000 | 05:59 |
| Diagnostics Check | | OK | 05:59 |

Data Download Date/Time: 06/30/2026 12:45

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403602530
DRY GAS EXPIRY DATE: 04/01/2029
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: GAINEY

DATE/TIME OF TEST: 05/16/2026 23:46:10

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 23:44 | |
| Air Blank | 0.000 | 23:45 | |
| Control Test | 0.000 | 23:45 | Control Outside Tolerance |
| Air Blank | 0.000 | 23:46 | |

Data Download Date/Time: 06/02/2026 02:57

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

INSTRUMENT REGISTERED TO: FHP TRP B LAKE CITY
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,SHANE,G
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TAYLOR,SHANE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403602530
DRY GAS EXPIRY DATE: 04/01/2029
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: GAINEY

DATE/TIME OF TEST: 05/16/2026 23:50:00

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 23:48 | |
| Air Blank | 0.000 | 23:49 | |
| Control Test | 0.000 | 23:49 | Control Outside Tolerance |
| Air Blank | 0.000 | 23:50 | |

Data Download Date/Time: 06/02/2026 02:57

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,COLBY,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SEARS,COLBY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 05/03/2026 02:41:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:35 |
| Air Blank | | 0.000 | 02:35 |
| Control Test | | 0.080 | 02:36 |
| Air Blank | | 0.000 | 02:36 |
| Subject Sample | # 1 | 0.159 | 02:37 |
| Breath Volume | | 1.292 | 02:37 |
| Air Blank | | 0.000 | 02:37 |
| Air Blank | | 0.000 | 02:39 |
| Subject Sample | # 2 | 0.150 | 02:39 |
| Breath Volume | | 1.312 | 02:39 |
| Air Blank | | 0.000 | 02:40 |
| Control Test | | 0.079 | 02:40 |
| Air Blank | | 0.000 | 02:41 |
| Diagnostics Check | | OK | 02:41 |

Data Download Date/Time: 05/06/2026 20:12

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 05/18/2026 23:10:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:04 |
| Air Blank | | 0.000 | 23:05 |
| Control Test | | 0.078 | 23:05 |
| Air Blank | | 0.000 | 23:06 |
| Subject Sample | # 1 | 0.201 | 23:06 |
| Breath Volume | | 2.066 | 23:06 |
| Air Blank | | 0.000 | 23:07 |
| Air Blank | | 0.000 | 23:08 |
| Subject Sample | # 2 | 0.200 | 23:09 |
| Breath Volume | | 1.515 | 23:09 |
| Air Blank | | 0.000 | 23:09 |
| Control Test | | 0.081 | 23:10 |
| Air Blank | | 0.000 | 23:10 |
| Diagnostics Check | | OK | 23:10 |

Data Download Date/Time: 06/16/2026 20:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: KLINE

DATE/TIME OF TEST: 05/29/2026 00:46:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:40 |
| Air Blank | | 0.000 | 00:40 |
| Control Test | | 0.080 | 00:40 |
| Air Blank | | 0.000 | 00:41 |
| Subject Sample | # 1 | 0.000 | 00:42 |
| Breath Volume | | 5.398 | 00:42 |
| Air Blank | | 0.000 | 00:42 |
| Air Blank | | 0.000 | 00:44 |
| Subject Sample | # 2 | 0.000 | 00:45 |
| Breath Volume | | 5.117 | 00:45 |
| Air Blank | | 0.000 | 00:45 |
| Control Test | | 0.079 | 00:45 |
| Air Blank | | 0.000 | 00:46 |
| Diagnostics Check | | OK | 00:46 |

Data Download Date/Time: 06/16/2026 20:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516666
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: STUNKARD

DATE/TIME OF TEST: 05/07/2026 02:58:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:51 |
| Air Blank | | 0.000 | 02:51 |
| Control Test | | 0.077 | 02:52 |
| Air Blank | | 0.000 | 02:52 |
| Subject Sample | # 1 | 0.180 | 02:54 |
| Breath Volume | | 1.605 | 02:54 |
| Air Blank | | 0.000 | 02:54 |
| Air Blank | | 0.000 | 02:56 |
| Subject Sample | # 2 | 0.194 | 02:57 |
| Breath Volume | | 1.441 | 02:57 |
| Air Blank | | 0.000 | 02:57 |
| Control Test | | 0.078 | 02:58 |
| Air Blank | | 0.000 | 02:58 |
| Diagnostics Check | | OK | 02:58 |

Data Download Date/Time: 06/29/2026 05:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DRIGGERS,BRANDON
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516666
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BROKER

DATE/TIME OF TEST: 05/12/2026 02:48:11

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:36 | |
| Air Blank | | 0.000 | 02:37 | |
| Control Test | | 0.078 | 02:37 | |
| Air Blank | | 0.000 | 02:38 | |
| Subject Sample | # 1 | | 02:41 | No Sample Provided |
| Breath Volume | | 0.000 | 02:41 | |
| Air Blank | | 0.000 | 02:41 | |
| Air Blank | | 0.000 | 02:43 | |
| Subject Sample | # 2 | | 02:46 | No Sample Provided |
| Breath Volume | | 0.000 | 02:46 | |
| Air Blank | | 0.000 | 02:47 | |
| Control Test | | 0.079 | 02:47 | |
| Air Blank | | 0.000 | 02:48 | |
| Diagnostics Check | | OK | 02:48 | |

Data Download Date/Time: 06/29/2026 05:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516666
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/29/2026 23:56:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:49 |
| Air Blank | | 0.000 | 23:50 |
| Control Test | | 0.079 | 23:50 |
| Air Blank | | 0.000 | 23:51 |
| Subject Sample | # 1 | 0.181 | 23:51 |
| Breath Volume | | 1.292 | 23:51 |
| Air Blank | | 0.000 | 23:52 |
| Air Blank | | 0.000 | 23:54 |
| Subject Sample | # 2 | 0.167 | 23:55 |
| Breath Volume | | 2.156 | 23:55 |
| Air Blank | | 0.000 | 23:55 |
| Control Test | | 0.079 | 23:56 |
| Air Blank | | 0.000 | 23:56 |
| Diagnostics Check | | OK | 23:56 |

Data Download Date/Time: 06/29/2026 05:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516666
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 05/30/2026 22:09:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:01 |
| Air Blank | | 0.000 | 22:02 |
| Control Test | | 0.078 | 22:02 |
| Air Blank | | 0.000 | 22:03 |
| Subject Sample | # 1 | 0.123 | 22:03 |
| Breath Volume | | 1.910 | 22:03 |
| Air Blank | | 0.000 | 22:04 |
| Air Blank | | 0.000 | 22:06 |
| Subject Sample | # 2 | 0.131 | 22:07 |
| Breath Volume | | 2.230 | 22:07 |
| Air Blank | | 0.000 | 22:08 |
| Control Test | | 0.078 | 22:08 |
| Air Blank | | 0.000 | 22:08 |
| Diagnostics Check | | OK | 22:09 |

Data Download Date/Time: 06/29/2026 05:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516666
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DE SANTO

DATE/TIME OF TEST: 05/31/2026 22:13:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:06 |
| Air Blank | | 0.000 | 22:06 |
| Control Test | | 0.078 | 22:07 |
| Air Blank | | 0.000 | 22:07 |
| Subject Sample | # 1 | 0.188 | 22:08 |
| Breath Volume | | 1.679 | 22:08 |
| Air Blank | | 0.000 | 22:09 |
| Air Blank | | 0.000 | 22:11 |
| Subject Sample | # 2 | 0.188 | 22:11 |
| Breath Volume | | 1.558 | 22:11 |
| Air Blank | | 0.000 | 22:12 |
| Control Test | | 0.079 | 22:12 |
| Air Blank | | 0.000 | 22:13 |
| Diagnostics Check | | OK | 22:13 |

Data Download Date/Time: 06/29/2026 05:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SCOTT II

DATE/TIME OF TEST: 05/01/2026 02:56:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:48 |
| Air Blank | | 0.000 | 02:49 |
| Control Test | | 0.079 | 02:49 |
| Air Blank | | 0.000 | 02:50 |
| Subject Sample | # 1 | 0.177 | 02:51 |
| Breath Volume | | 1.789 | 02:51 |
| Air Blank | | 0.000 | 02:51 |
| Air Blank | | 0.000 | 02:53 |
| Subject Sample | # 2 | 0.167 | 02:54 |
| Breath Volume | | 2.050 | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Control Test | | 0.079 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Diagnostics Check | | OK | 02:56 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: GARCED

DATE/TIME OF TEST: 05/01/2026 04:12:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:05 |
| Air Blank | | 0.000 | 04:05 |
| Control Test | | 0.080 | 04:06 |
| Air Blank | | 0.000 | 04:06 |
| Subject Sample | # 1 | 0.105 | 04:08 |
| Breath Volume | | 1.398 | 04:08 |
| Air Blank | | 0.000 | 04:08 |
| Air Blank | | 0.000 | 04:10 |
| Subject Sample | # 2 | 0.101 | 04:11 |
| Breath Volume | | 1.769 | 04:11 |
| Air Blank | | 0.000 | 04:11 |
| Control Test | | 0.078 | 04:12 |
| Air Blank | | 0.000 | 04:12 |
| Diagnostics Check | | OK | 04:12 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PINCAY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 05/01/2026 23:37:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:31 |
| Air Blank | | 0.000 | 23:31 |
| Control Test | | 0.080 | 23:32 |
| Air Blank | | 0.000 | 23:32 |
| Subject Sample | # 1 | 0.000 | 23:33 |
| Breath Volume | | 2.687 | 23:33 |
| Air Blank | | 0.000 | 23:33 |
| Air Blank | | 0.000 | 23:35 |
| Subject Sample | # 2 | 0.000 | 23:36 |
| Breath Volume | | 2.257 | 23:36 |
| Air Blank | | 0.000 | 23:36 |
| Control Test | | 0.080 | 23:37 |
| Air Blank | | 0.000 | 23:37 |
| Diagnostics Check | | OK | 23:37 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: LOZADA

DATE/TIME OF TEST: 05/02/2026 00:00:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:53 |
| Air Blank | | 0.000 | 23:54 |
| Control Test | | 0.080 | 23:54 |
| Air Blank | | 0.000 | 23:55 |
| Subject Sample | # 1 | 0.224 | 23:55 |
| Breath Volume | | 1.707 | 23:55 |
| Air Blank | | 0.000 | 23:56 |
| Air Blank | | 0.000 | 23:57 |
| Subject Sample | # 2 | 0.226 | 23:58 |
| Breath Volume | | 2.878 | 23:58 |
| Air Blank | | 0.000 | 23:59 |
| Control Test | | 0.079 | 23:59 |
| Air Blank | | 0.000 | 00:00 |
| Diagnostics Check | | OK | 00:00 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: STRIGLE

DATE/TIME OF TEST: 05/02/2026 01:50:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:43 |
| Air Blank | | 0.000 | 01:43 |
| Control Test | | 0.079 | 01:43 |
| Air Blank | | 0.000 | 01:44 |
| Subject Sample | # 1 | 0.128 | 01:45 |
| Breath Volume | | 1.335 | 01:45 |
| Air Blank | | 0.000 | 01:45 |
| Air Blank | | 0.000 | 01:47 |
| Subject Sample | # 2 | 0.136 | 01:48 |
| Breath Volume | | 1.949 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Control Test | | 0.079 | 01:49 |
| Air Blank | | 0.000 | 01:49 |
| Diagnostics Check | | OK | 01:50 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: WELLING

DATE/TIME OF TEST: 05/02/2026 02:50:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:43 |
| Air Blank | | 0.000 | 02:43 |
| Control Test | | 0.080 | 02:44 |
| Air Blank | | 0.000 | 02:44 |
| Subject Sample | # 1 | 0.151 | 02:45 |
| Breath Volume | | 1.441 | 02:45 |
| Air Blank | | 0.000 | 02:46 |
| Air Blank | | 0.000 | 02:47 |
| Subject Sample | # 2 | 0.156 | 02:48 |
| Breath Volume | | 1.878 | 02:48 |
| Air Blank | | 0.000 | 02:49 |
| Control Test | | 0.079 | 02:49 |
| Air Blank | | 0.000 | 02:49 |
| Diagnostics Check | | OK | 02:50 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SARRASSIN,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: COSBY III

DATE/TIME OF TEST: 05/02/2026 04:56:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:49 |
| Air Blank | | 0.000 | 04:49 |
| Control Test | | 0.080 | 04:49 |
| Air Blank | | 0.000 | 04:50 |
| Subject Sample | # 1 | 0.087 | 04:51 |
| Breath Volume | | 1.500 | 04:51 |
| Air Blank | | 0.000 | 04:51 |
| Air Blank | | 0.000 | 04:53 |
| Subject Sample | # 2 | 0.083 | 04:54 |
| Breath Volume | | 2.003 | 04:54 |
| Air Blank | | 0.000 | 04:55 |
| Control Test | | 0.079 | 04:55 |
| Air Blank | | 0.000 | 04:56 |
| Diagnostics Check | | OK | 04:56 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MORALES MUSTELIER

DATE/TIME OF TEST: 05/02/2026 05:12:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:05 |
| Air Blank | | 0.000 | 05:05 |
| Control Test | | 0.079 | 05:06 |
| Air Blank | | 0.000 | 05:06 |
| Subject Sample | # 1 | 0.151 | 05:07 |
| Breath Volume | | 1.167 | 05:07 |
| Air Blank | | 0.000 | 05:08 |
| Air Blank | | 0.000 | 05:09 |
| Subject Sample | # 2 | 0.147 | 05:10 |
| Breath Volume | | 1.121 | 05:10 |
| Air Blank | | 0.000 | 05:11 |
| Control Test | | 0.078 | 05:11 |
| Air Blank | | 0.000 | 05:12 |
| Diagnostics Check | | OK | 05:12 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,ANTHONY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,ANTHONY
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 05/02/2026 06:18:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:11 |
| Air Blank | | 0.000 | 06:11 |
| Control Test | | 0.079 | 06:12 |
| Air Blank | | 0.000 | 06:12 |
| Subject Sample | # 1 | 0.056 | 06:13 |
| Breath Volume | | 1.234 | 06:13 |
| Air Blank | | 0.000 | 06:13 |
| Air Blank | | 0.000 | 06:15 |
| Subject Sample | # 2 | 0.058 | 06:17 |
| Breath Volume | | 1.640 | 06:17 |
| Air Blank | | 0.000 | 06:17 |
| Control Test | | 0.079 | 06:17 |
| Air Blank | | 0.000 | 06:18 |
| Diagnostics Check | | OK | 06:18 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DE JESUS

DATE/TIME OF TEST: 05/02/2026 22:50:54

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:42 | |
| Air Blank | | 0.000 | 22:42 | |
| Control Test | | 0.079 | 22:43 | |
| Air Blank | | 0.000 | 22:43 | |
| Subject Sample | # 1 | | 22:46 | No Sample Provided |
| Breath Volume | | 0.000 | 22:46 | |
| Air Blank | | 0.000 | 22:47 | |
| Air Blank | | 0.000 | 22:49 | |
| Subject Sample | # 2 | | 22:49 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:49 | |
| Air Blank | | 0.000 | 22:49 | |
| Control Test | | 0.079 | 22:50 | |
| Air Blank | | 0.000 | 22:50 | |
| Diagnostics Check | | OK | 22:50 | |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MIKSCH,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: CORONA-MORO

DATE/TIME OF TEST: 05/03/2026 00:38:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:31 |
| Air Blank | | 0.000 | 00:31 |
| Control Test | | 0.079 | 00:32 |
| Air Blank | | 0.000 | 00:32 |
| Subject Sample | # 1 | 0.178 | 00:33 |
| Breath Volume | | 2.511 | 00:33 |
| Air Blank | | 0.000 | 00:34 |
| Air Blank | | 0.000 | 00:36 |
| Subject Sample | # 2 | 0.176 | 00:36 |
| Breath Volume | | 2.535 | 00:36 |
| Air Blank | | 0.000 | 00:37 |
| Control Test | | 0.078 | 00:37 |
| Air Blank | | 0.000 | 00:38 |
| Diagnostics Check | | OK | 00:38 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: PICKARD

DATE/TIME OF TEST: 05/03/2026 02:01:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:54 |
| Air Blank | | 0.000 | 01:54 |
| Control Test | | 0.079 | 01:55 |
| Air Blank | | 0.000 | 01:55 |
| Subject Sample | # 1 | 0.128 | 01:56 |
| Breath Volume | | 1.667 | 01:56 |
| Air Blank | | 0.000 | 01:56 |
| Air Blank | | 0.000 | 01:58 |
| Subject Sample | # 2 | 0.135 | 01:59 |
| Breath Volume | | 1.714 | 01:59 |
| Air Blank | | 0.000 | 02:00 |
| Control Test | | 0.079 | 02:00 |
| Air Blank | | 0.000 | 02:01 |
| Diagnostics Check | | OK | 02:01 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SIFRALIEN

DATE/TIME OF TEST: 05/03/2026 03:01:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:54 |
| Air Blank | | 0.000 | 02:54 |
| Control Test | | 0.079 | 02:55 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 1 | 0.133 | 02:56 |
| Breath Volume | | 1.132 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Air Blank | | 0.000 | 02:59 |
| Subject Sample | # 2 | 0.140 | 02:59 |
| Breath Volume | | 2.027 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Control Test | | 0.078 | 03:00 |
| Air Blank | | 0.000 | 03:01 |
| Diagnostics Check | | OK | 03:01 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: WONG

DATE/TIME OF TEST: 05/03/2026 04:53:28

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:46 |
| Air Blank | | 0.000 | 04:47 |
| Control Test | | 0.079 | 04:47 |
| Air Blank | | 0.000 | 04:47 |
| Subject Sample | # 1 | 0.116 | 04:48 |
| Breath Volume | | 3.566 | 04:48 |
| Air Blank | | 0.000 | 04:49 |
| Air Blank | | 0.000 | 04:51 |
| Subject Sample | # 2 | 0.113 | 04:51 |
| Breath Volume | | 3.640 | 04:51 |
| Air Blank | | 0.000 | 04:52 |
| Control Test | | 0.079 | 04:52 |
| Air Blank | | 0.000 | 04:53 |
| Diagnostics Check | | OK | 04:53 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PINCAY,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 05/03/2026 06:07:22

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:59 |
| Air Blank | | 0.000 | 05:59 |
| Control Test | | 0.080 | 05:59 |
| Air Blank | | 0.000 | 06:00 |
| Subject Sample | # 1 | 0.115 | 06:02 |
| Breath Volume | | 1.671 | 06:02 |
| Air Blank | | 0.000 | 06:03 |
| Air Blank | | 0.000 | 06:05 |
| Subject Sample | # 2 | 0.112 | 06:05 |
| Breath Volume | | 1.535 | 06:05 |
| Air Blank | | 0.000 | 06:06 |
| Control Test | | 0.078 | 06:06 |
| Air Blank | | 0.000 | 06:07 |
| Diagnostics Check | | OK | 06:07 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAZQUEZ,V
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/03/2026 08:17:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:11 |
| Air Blank | | 0.000 | 08:11 |
| Control Test | | 0.080 | 08:12 |
| Air Blank | | 0.000 | 08:12 |
| Subject Sample | # 1 | 0.103 | 08:13 |
| Breath Volume | | 1.445 | 08:13 |
| Air Blank | | 0.000 | 08:13 |
| Air Blank | | 0.000 | 08:15 |
| Subject Sample | # 2 | 0.104 | 08:16 |
| Breath Volume | | 1.562 | 08:16 |
| Air Blank | | 0.000 | 08:16 |
| Control Test | | 0.079 | 08:16 |
| Air Blank | | 0.000 | 08:17 |
| Diagnostics Check | | OK | 08:17 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: REBOLLO

DATE/TIME OF TEST: 05/04/2026 02:24:48

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:19 | |
| Air Blank | | 0.000 | 02:19 | |
| Control Test | | 0.079 | 02:20 | |
| Air Blank | | 0.000 | 02:20 | |
| Subject Sample | # 1 | | 02:23 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:23 | |
| Air Blank | | 0.000 | 02:23 | |
| Control Test | | 0.079 | 02:24 | |
| Air Blank | | 0.000 | 02:24 | |
| Diagnostics Check | | OK | 02:24 | |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: SUAREZ LOZANO

DATE/TIME OF TEST: 05/04/2026 04:23:34

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:17 |
| Air Blank | | 0.000 | 04:17 |
| Control Test | | 0.079 | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Subject Sample | # 1 | 0.089 | 04:19 |
| Breath Volume | | 2.652 | 04:19 |
| Air Blank | | 0.000 | 04:19 |
| Air Blank | | 0.000 | 04:21 |
| Subject Sample | # 2 | 0.088 | 04:22 |
| Breath Volume | | 2.558 | 04:22 |
| Air Blank | | 0.000 | 04:22 |
| Control Test | | 0.079 | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Diagnostics Check | | OK | 04:23 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 05/05/2026 02:13:07

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:05 |
| Air Blank | | 0.000 | 02:06 |
| Control Test | | 0.080 | 02:06 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 1 | 0.185 | 02:08 |
| Breath Volume | | 1.304 | 02:08 |
| Air Blank | | 0.000 | 02:08 |
| Air Blank | | 0.000 | 02:10 |
| Subject Sample | # 2 | 0.173 | 02:11 |
| Breath Volume | | 2.097 | 02:11 |
| Air Blank | | 0.000 | 02:12 |
| Control Test | | 0.079 | 02:12 |
| Air Blank | | 0.000 | 02:13 |
| Diagnostics Check | | OK | 02:13 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: FACEY

DATE/TIME OF TEST: 05/05/2026 04:24:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:17 |
| Air Blank | | 0.000 | 04:18 |
| Control Test | | 0.079 | 04:18 |
| Air Blank | | 0.000 | 04:19 |
| Subject Sample | # 1 | 0.184 | 04:20 |
| Breath Volume | | 2.152 | 04:20 |
| Air Blank | | 0.000 | 04:20 |
| Air Blank | | 0.000 | 04:22 |
| Subject Sample | # 2 | 0.188 | 04:23 |
| Breath Volume | | 2.796 | 04:23 |
| Air Blank | | 0.000 | 04:23 |
| Control Test | | 0.079 | 04:24 |
| Air Blank | | 0.000 | 04:24 |
| Diagnostics Check | | OK | 04:24 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 05/05/2026 05:25:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:19 |
| Air Blank | | 0.000 | 05:19 |
| Control Test | | 0.079 | 05:20 |
| Air Blank | | 0.000 | 05:20 |
| Subject Sample | # 1 | 0.280 | 05:21 |
| Breath Volume | | 2.730 | 05:21 |
| Air Blank | | 0.000 | 05:21 |
| Air Blank | | 0.000 | 05:23 |
| Subject Sample | # 2 | 0.281 | 05:24 |
| Breath Volume | | 3.046 | 05:24 |
| Air Blank | | 0.000 | 05:24 |
| Control Test | | 0.079 | 05:25 |
| Air Blank | | 0.000 | 05:25 |
| Diagnostics Check | | OK | 05:25 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ACEVEDO,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/05/2026 21:48:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:39 |
| Air Blank | | 0.000 | 21:40 |
| Control Test | | 0.079 | 21:40 |
| Air Blank | | 0.000 | 21:41 |
| Subject Sample | # 1 | 0.288 | 21:42 |
| Breath Volume | | 1.558 | 21:42 |
| Air Blank | | 0.000 | 21:43 |
| Air Blank | | 0.000 | 21:45 |
| Subject Sample | # 2 | 0.272 | 21:46 |
| Breath Volume | | 1.167 | 21:46 |
| Air Blank | | 0.000 | 21:47 |
| Control Test | | 0.077 | 21:47 |
| Air Blank | | 0.000 | 21:48 |
| Diagnostics Check | | OK | 21:48 |

Data Download Date/Time: 05/15/2026 09:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: STROUD

DATE/TIME OF TEST: 05/06/2026 02:50:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:40 |
| Air Blank | | 0.000 | 02:41 |
| Control Test | | 0.079 | 02:41 |
| Air Blank | | 0.000 | 02:42 |
| Subject Sample | # 1 | 0.117 | 02:44 |
| Breath Volume | | 2.386 | 02:44 |
| Air Blank | | 0.000 | 02:45 |
| Air Blank | | 0.000 | 02:47 |
| Subject Sample | # 2 | 0.121 | 02:48 |
| Breath Volume | | 2.304 | 02:48 |
| Air Blank | | 0.000 | 02:49 |
| Control Test | | 0.081 | 02:49 |
| Air Blank | | 0.000 | 02:50 |
| Diagnostics Check | | OK | 02:50 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PLATT,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 05/06/2026 04:26:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:18 |
| Air Blank | | 0.000 | 04:18 |
| Control Test | | 0.079 | 04:18 |
| Air Blank | | 0.000 | 04:19 |
| Subject Sample | # 1 | 0.141 | 04:21 |
| Breath Volume | | 1.296 | 04:21 |
| Air Blank | | 0.000 | 04:22 |
| Air Blank | | 0.000 | 04:23 |
| Subject Sample | # 2 | 0.153 | 04:24 |
| Breath Volume | | 1.644 | 04:24 |
| Air Blank | | 0.000 | 04:25 |
| Control Test | | 0.078 | 04:25 |
| Air Blank | | 0.000 | 04:26 |
| Diagnostics Check | | OK | 04:26 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:26

SUBJECT'S LAST NAME: JEFFERS

DATE/TIME OF TEST: 05/06/2026 06:52:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:46 |
| Air Blank | | 0.000 | 06:46 |
| Control Test | | 0.079 | 06:47 |
| Air Blank | | 0.000 | 06:47 |
| Subject Sample | # 1 | 0.119 | 06:48 |
| Breath Volume | | 3.242 | 06:48 |
| Air Blank | | 0.000 | 06:48 |
| Air Blank | | 0.000 | 06:50 |
| Subject Sample | # 2 | 0.114 | 06:50 |
| Breath Volume | | 2.488 | 06:50 |
| Air Blank | | 0.000 | 06:51 |
| Control Test | | 0.079 | 06:51 |
| Air Blank | | 0.000 | 06:52 |
| Diagnostics Check | | OK | 06:52 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: ILDEMIR

DATE/TIME OF TEST: 05/06/2026 21:18:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:12 |
| Air Blank | | 0.000 | 21:12 |
| Control Test | | 0.079 | 21:13 |
| Air Blank | | 0.000 | 21:13 |
| Subject Sample | # 1 | 0.000 | 21:14 |
| Breath Volume | | 2.160 | 21:14 |
| Air Blank | | 0.000 | 21:14 |
| Air Blank | | 0.000 | 21:16 |
| Subject Sample | # 2 | 0.000 | 21:17 |
| Breath Volume | | 2.707 | 21:17 |
| Air Blank | | 0.000 | 21:17 |
| Control Test | | 0.079 | 21:17 |
| Air Blank | | 0.000 | 21:18 |
| Diagnostics Check | | OK | 21:18 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GAROFALO,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: TALAVERA

DATE/TIME OF TEST: 05/06/2026 23:41:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:34 |
| Air Blank | | 0.000 | 23:35 |
| Control Test | | 0.080 | 23:35 |
| Air Blank | | 0.000 | 23:35 |
| Subject Sample | # 1 | 0.149 | 23:36 |
| Breath Volume | | 2.070 | 23:36 |
| Air Blank | | 0.000 | 23:37 |
| Air Blank | | 0.000 | 23:38 |
| Subject Sample | # 2 | 0.145 | 23:40 |
| Breath Volume | | 1.167 | 23:40 |
| Air Blank | | 0.000 | 23:40 |
| Control Test | | 0.078 | 23:41 |
| Air Blank | | 0.000 | 23:41 |
| Diagnostics Check | | OK | 23:41 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: KERWIN

DATE/TIME OF TEST: 05/07/2026 00:22:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:15 |
| Air Blank | | 0.000 | 00:16 |
| Control Test | | 0.079 | 00:16 |
| Air Blank | | 0.000 | 00:17 |
| Subject Sample | # 1 | 0.109 | 00:18 |
| Breath Volume | | 1.804 | 00:18 |
| Air Blank | | 0.000 | 00:18 |
| Air Blank | | 0.000 | 00:20 |
| Subject Sample | # 2 | 0.108 | 00:21 |
| Breath Volume | | 1.144 | 00:21 |
| Air Blank | | 0.000 | 00:21 |
| Control Test | | 0.079 | 00:22 |
| Air Blank | | 0.000 | 00:22 |
| Diagnostics Check | | OK | 00:22 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: CRUZ TAMAYO

DATE/TIME OF TEST: 05/07/2026 03:21:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:15 |
| Air Blank | | 0.000 | 03:15 |
| Control Test | | 0.079 | 03:16 |
| Air Blank | | 0.000 | 03:16 |
| Subject Sample | # 1 | 0.108 | 03:17 |
| Breath Volume | | 2.218 | 03:17 |
| Air Blank | | 0.000 | 03:17 |
| Air Blank | | 0.000 | 03:19 |
| Subject Sample | # 2 | 0.111 | 03:20 |
| Breath Volume | | 2.589 | 03:20 |
| Air Blank | | 0.000 | 03:20 |
| Control Test | | 0.079 | 03:21 |
| Air Blank | | 0.000 | 03:21 |
| Diagnostics Check | | OK | 03:21 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CARDONA CASTILLO

DATE/TIME OF TEST: 05/08/2026 00:09:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:01 |
| Air Blank | | 0.000 | 00:02 |
| Control Test | | 0.080 | 00:02 |
| Air Blank | | 0.000 | 00:03 |
| Subject Sample | # 1 | 0.051 | 00:04 |
| Breath Volume | | 1.425 | 00:04 |
| Air Blank | | 0.000 | 00:04 |
| Air Blank | | 0.000 | 00:06 |
| Subject Sample | # 2 | 0.050 | 00:07 |
| Breath Volume | | 1.421 | 00:07 |
| Air Blank | | 0.000 | 00:08 |
| Control Test | | 0.079 | 00:08 |
| Air Blank | | 0.000 | 00:09 |
| Diagnostics Check | | OK | 00:09 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: MENENDEZ-CUARTA

DATE/TIME OF TEST: 05/09/2026 00:33:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:24 |
| Air Blank | | 0.000 | 00:25 |
| Control Test | | 0.079 | 00:25 |
| Air Blank | | 0.000 | 00:26 |
| Subject Sample | # 1 | 0.227 | 00:28 |
| Breath Volume | | 1.183 | 00:28 |
| Air Blank | | 0.000 | 00:28 |
| Air Blank | | 0.000 | 00:30 |
| Subject Sample | # 2 | 0.237 | 00:31 |
| Breath Volume | | 1.257 | 00:31 |
| Air Blank | | 0.000 | 00:32 |
| Control Test | | 0.079 | 00:32 |
| Air Blank | | 0.000 | 00:33 |
| Diagnostics Check | | OK | 00:33 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIBBINS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/09/2026 04:16:36

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:03 | |
| Air Blank | | 0.000 | 04:04 | |
| Control Test | | 0.079 | 04:04 | |
| Air Blank | | 0.000 | 04:05 | |
| Subject Sample | # 1 | | 04:08 | No Sample Provided |
| Breath Volume | | 0.000 | 04:08 | |
| Air Blank | | 0.000 | 04:08 | |
| Air Blank | | 0.000 | 04:10 | |
| Subject Sample | # 2 | 0.197 | 04:12 | No .020 agreement |
| Breath Volume | | 1.582 | 04:12 | |
| Air Blank | | 0.000 | 04:12 | |
| Air Blank | | 0.000 | 04:14 | |
| Subject Sample | # 3 | 0.173 | 04:15 | No .020 agreement |
| Breath Volume | | 2.507 | 04:15 | |
| Air Blank | | 0.000 | 04:15 | |
| Control Test | | 0.079 | 04:16 | |
| Air Blank | | 0.000 | 04:16 | |
| Diagnostics Check | | OK | 04:16 | |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIBBINS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/09/2026 04:25:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:19 |
| Air Blank | | 0.000 | 04:19 |
| Control Test | | 0.079 | 04:19 |
| Air Blank | | 0.000 | 04:20 |
| Subject Sample | # 1 | 0.183 | 04:21 |
| Breath Volume | | 1.292 | 04:21 |
| Air Blank | | 0.000 | 04:21 |
| Air Blank | | 0.000 | 04:23 |
| Subject Sample | # 2 | 0.180 | 04:24 |
| Breath Volume | | 1.781 | 04:24 |
| Air Blank | | 0.000 | 04:24 |
| Control Test | | 0.079 | 04:25 |
| Air Blank | | 0.000 | 04:25 |
| Diagnostics Check | | OK | 04:25 |

Data Download Date/Time: 05/15/2026 09:18

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EHRHART,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 05/09/2026 04:53:51

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:46 |
| Air Blank | | 0.000 | 04:47 |
| Control Test | | 0.079 | 04:47 |
| Air Blank | | 0.000 | 04:48 |
| Subject Sample | # 1 | 0.089 | 04:49 |
| Breath Volume | | 2.589 | 04:49 |
| Air Blank | | 0.000 | 04:49 |
| Air Blank | | 0.000 | 04:51 |
| Subject Sample | # 2 | 0.088 | 04:52 |
| Breath Volume | | 1.324 | 04:52 |
| Air Blank | | 0.000 | 04:52 |
| Control Test | | 0.079 | 04:53 |
| Air Blank | | 0.000 | 04:53 |
| Diagnostics Check | | OK | 04:53 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARRASCO,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: OWENS II

DATE/TIME OF TEST: 05/09/2026 23:39:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:33 |
| Air Blank | | 0.000 | 23:33 |
| Control Test | | 0.080 | 23:34 |
| Air Blank | | 0.000 | 23:34 |
| Subject Sample | # 1 | 0.000 | 23:35 |
| Breath Volume | | 2.226 | 23:35 |
| Air Blank | | 0.000 | 23:35 |
| Air Blank | | 0.000 | 23:37 |
| Subject Sample | # 2 | 0.000 | 23:38 |
| Breath Volume | | 2.480 | 23:38 |
| Air Blank | | 0.000 | 23:38 |
| Control Test | | 0.079 | 23:38 |
| Air Blank | | 0.000 | 23:39 |
| Diagnostics Check | | OK | 23:39 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: HERNANDEZ ROMERO

DATE/TIME OF TEST: 05/10/2026 03:04:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:58 |
| Air Blank | | 0.000 | 02:59 |
| Control Test | | 0.078 | 02:59 |
| Air Blank | | 0.000 | 02:59 |
| Subject Sample | # 1 | 0.146 | 03:00 |
| Breath Volume | | 1.792 | 03:00 |
| Air Blank | | 0.000 | 03:00 |
| Air Blank | | 0.000 | 03:02 |
| Subject Sample | # 2 | 0.152 | 03:03 |
| Breath Volume | | 2.273 | 03:03 |
| Air Blank | | 0.000 | 03:03 |
| Control Test | | 0.078 | 03:04 |
| Air Blank | | 0.000 | 03:04 |
| Diagnostics Check | | OK | 03:04 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEACH,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: RAVELO CALERO

DATE/TIME OF TEST: 05/10/2026 04:46:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:39 |
| Air Blank | | 0.000 | 04:40 |
| Control Test | | 0.079 | 04:40 |
| Air Blank | | 0.000 | 04:41 |
| Subject Sample | # 1 | 0.113 | 04:42 |
| Breath Volume | | 2.164 | 04:42 |
| Air Blank | | 0.000 | 04:42 |
| Air Blank | | 0.000 | 04:44 |
| Subject Sample | # 2 | 0.106 | 04:45 |
| Breath Volume | | 2.746 | 04:45 |
| Air Blank | | 0.000 | 04:45 |
| Control Test | | 0.078 | 04:46 |
| Air Blank | | 0.000 | 04:46 |
| Diagnostics Check | | OK | 04:46 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/10/2026 22:02:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:55 |
| Air Blank | | 0.000 | 21:56 |
| Control Test | | 0.080 | 21:56 |
| Air Blank | | 0.000 | 21:57 |
| Subject Sample | # 1 | 0.141 | 21:57 |
| Breath Volume | | 2.683 | 21:57 |
| Air Blank | | 0.000 | 21:58 |
| Air Blank | | 0.000 | 21:59 |
| Subject Sample | # 2 | 0.146 | 22:00 |
| Breath Volume | | 2.910 | 22:00 |
| Air Blank | | 0.000 | 22:01 |
| Control Test | | 0.079 | 22:01 |
| Air Blank | | 0.000 | 22:02 |
| Diagnostics Check | | OK | 22:02 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TELLEZ-ALFARO,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 05/11/2026 00:29:13

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:17 | |
| Air Blank | | 0.000 | 00:18 | |
| Control Test | | 0.080 | 00:18 | |
| Air Blank | | 0.000 | 00:18 | |
| Subject Sample | # 1 | 0.093 | 00:22 | Volume Not Met |
| Breath Volume | | 0.128 | 00:22 | |
| Air Blank | | 0.000 | 00:22 | |
| Air Blank | | 0.000 | 00:24 | |
| Subject Sample | # 2 | 0.115 | 00:27 | Volume Not Met |
| Breath Volume | | 0.265 | 00:27 | |
| Air Blank | | 0.000 | 00:28 | |
| Control Test | | 0.079 | 00:28 | |
| Air Blank | | 0.000 | 00:29 | |
| Diagnostics Check | | OK | 00:29 | |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ERHART,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 05/11/2026 03:34:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:26 |
| Air Blank | | 0.000 | 03:27 |
| Control Test | | 0.079 | 03:27 |
| Air Blank | | 0.000 | 03:28 |
| Subject Sample | # 1 | 0.154 | 03:28 |
| Breath Volume | | 1.621 | 03:28 |
| Air Blank | | 0.000 | 03:29 |
| Air Blank | | 0.000 | 03:31 |
| Subject Sample | # 2 | 0.152 | 03:32 |
| Breath Volume | | 2.082 | 03:32 |
| Air Blank | | 0.000 | 03:33 |
| Control Test | | 0.079 | 03:33 |
| Air Blank | | 0.000 | 03:34 |
| Diagnostics Check | | OK | 03:34 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: ARNETT

DATE/TIME OF TEST: 05/11/2026 05:21:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 05:14 |
| Air Blank | | 0.000 | 05:15 |
| Control Test | | 0.079 | 05:15 |
| Air Blank | | 0.000 | 05:16 |
| Subject Sample | # 1 | 0.246 | 05:17 |
| Breath Volume | | 2.023 | 05:17 |
| Air Blank | | 0.000 | 05:17 |
| Air Blank | | 0.000 | 05:19 |
| Subject Sample | # 2 | 0.245 | 05:20 |
| Breath Volume | | 1.738 | 05:20 |
| Air Blank | | 0.000 | 05:20 |
| Control Test | | 0.079 | 05:21 |
| Air Blank | | 0.000 | 05:21 |
| Diagnostics Check | | OK | 05:21 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REICHERT,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 05/11/2026 15:08:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:02 |
| Air Blank | | 0.000 | 15:03 |
| Control Test | | 0.080 | 15:03 |
| Air Blank | | 0.000 | 15:03 |
| Subject Sample | # 1 | 0.000 | 15:04 |
| Breath Volume | | 2.968 | 15:04 |
| Air Blank | | 0.000 | 15:04 |
| Air Blank | | 0.000 | 15:06 |
| Subject Sample | # 2 | 0.000 | 15:07 |
| Breath Volume | | 2.718 | 15:07 |
| Air Blank | | 0.000 | 15:07 |
| Control Test | | 0.080 | 15:08 |
| Air Blank | | 0.000 | 15:08 |
| Diagnostics Check | | OK | 15:08 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LUGO,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: GROGGS

DATE/TIME OF TEST: 05/11/2026 19:14:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:04 |
| Air Blank | | 0.000 | 19:05 |
| Control Test | | 0.079 | 19:05 |
| Air Blank | | 0.000 | 19:05 |
| Subject Sample | # 1 | 0.311 | 19:06 |
| Breath Volume | | 1.558 | 19:06 |
| Air Blank | | 0.000 | 19:07 |
| Air Blank | | 0.000 | 19:09 |
| Subject Sample | # 2 | 0.289 | 19:09 |
| Breath Volume | | 2.585 | 19:09 |
| Air Blank | | 0.000 | 19:10 |
| Air Blank | | 0.000 | 19:12 |
| Subject Sample | # 3 | 0.284 | 19:12 |
| Breath Volume | | 1.710 | 19:12 |
| Air Blank | | 0.000 | 19:13 |
| Control Test | | 0.079 | 19:13 |
| Air Blank | | 0.000 | 19:14 |
| Diagnostics Check | | OK | 19:14 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 05/12/2026 01:24:53

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:20 | |
| Air Blank | | 0.000 | 01:21 | |
| Control Test | | 0.079 | 01:21 | |
| Air Blank | | 0.000 | 01:22 | |
| Subject Sample | # 1 | | 01:23 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:23 | |
| Air Blank | | 0.000 | 01:23 | |
| Control Test | | 0.079 | 01:24 | |
| Air Blank | | 0.000 | 01:24 | |
| Diagnostics Check | | OK | 01:24 | |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: COSTLOW

DATE/TIME OF TEST: 05/12/2026 02:51:16

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:43 |
| Air Blank | | 0.000 | 02:44 |
| Control Test | | 0.079 | 02:44 |
| Air Blank | | 0.000 | 02:45 |
| Subject Sample | # 1 | 0.189 | 02:46 |
| Breath Volume | | 2.765 | 02:46 |
| Air Blank | | 0.000 | 02:47 |
| Air Blank | | 0.000 | 02:49 |
| Subject Sample | # 2 | 0.173 | 02:49 |
| Breath Volume | | 1.175 | 02:49 |
| Air Blank | | 0.000 | 02:50 |
| Control Test | | 0.079 | 02:50 |
| Air Blank | | 0.000 | 02:51 |
| Diagnostics Check | | OK | 02:51 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007480

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:02

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 05/12/2026 13:31:50

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 13:24 |
| Air Blank | | 0.000 | 13:24 |
| Control Test | | 0.079 | 13:25 |
| Air Blank | | 0.000 | 13:25 |
| Subject Sample | # 1 | 0.000 | 13:27 |
| Breath Volume | | 1.417 | 13:27 |
| Air Blank | | 0.000 | 13:27 |
| Air Blank | | 0.000 | 13:29 |
| Subject Sample | # 2 | 0.000 | 13:30 |
| Breath Volume | | 1.808 | 13:30 |
| Air Blank | | 0.000 | 13:30 |
| Control Test | | 0.079 | 13:31 |
| Air Blank | | 0.000 | 13:31 |
| Diagnostics Check | | OK | 13:31 |

Data Download Date/Time: 05/15/2026 09:19

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PATLAN NAVARRETE

DATE/TIME OF TEST: 05/06/2026 02:11:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:05 |
| Air Blank | | 0.000 | 02:05 |
| Control Test | | 0.077 | 02:05 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 1 | 0.146 | 02:06 |
| Breath Volume | | 1.734 | 02:06 |
| Air Blank | | 0.000 | 02:07 |
| Air Blank | | 0.000 | 02:09 |
| Subject Sample | # 2 | 0.141 | 02:09 |
| Breath Volume | | 1.847 | 02:09 |
| Air Blank | | 0.000 | 02:10 |
| Control Test | | 0.077 | 02:10 |
| Air Blank | | 0.000 | 02:11 |
| Diagnostics Check | | OK | 02:11 |

Data Download Date/Time: 05/15/2026 10:59

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CAUSEY

DATE/TIME OF TEST: 05/10/2026 04:56:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:50 |
| Air Blank | | 0.000 | 04:50 |
| Control Test | | 0.077 | 04:50 |
| Air Blank | | 0.000 | 04:51 |
| Subject Sample | # 1 | 0.085 | 04:51 |
| Breath Volume | | 2.355 | 04:51 |
| Air Blank | | 0.000 | 04:52 |
| Air Blank | | 0.000 | 04:54 |
| Subject Sample | # 2 | 0.088 | 04:55 |
| Breath Volume | | 2.074 | 04:55 |
| Air Blank | | 0.000 | 04:56 |
| Control Test | | 0.077 | 04:56 |
| Air Blank | | 0.000 | 04:56 |
| Diagnostics Check | | OK | 04:56 |

Data Download Date/Time: 05/15/2026 10:59

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: FRANCISCO MIGUEL

DATE/TIME OF TEST: 05/11/2026 00:51:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:44 |
| Air Blank | | 0.000 | 00:45 |
| Control Test | | 0.077 | 00:45 |
| Air Blank | | 0.000 | 00:46 |
| Subject Sample | # 1 | 0.262 | 00:46 |
| Breath Volume | | 1.863 | 00:46 |
| Air Blank | | 0.000 | 00:47 |
| Air Blank | | 0.000 | 00:49 |
| Subject Sample | # 2 | 0.254 | 00:50 |
| Breath Volume | | 2.722 | 00:50 |
| Air Blank | | 0.000 | 00:50 |
| Control Test | | 0.077 | 00:51 |
| Air Blank | | 0.000 | 00:51 |
| Diagnostics Check | | OK | 00:51 |

Data Download Date/Time: 05/15/2026 10:59

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/18/2026 22:59:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:53 |
| Air Blank | | 0.000 | 22:54 |
| Control Test | | 0.077 | 22:54 |
| Air Blank | | 0.000 | 22:54 |
| Subject Sample | # 1 | 0.042 | 22:55 |
| Breath Volume | | 3.554 | 22:55 |
| Air Blank | | 0.000 | 22:56 |
| Air Blank | | 0.000 | 22:57 |
| Subject Sample | # 2 | 0.041 | 22:58 |
| Breath Volume | | 3.488 | 22:58 |
| Air Blank | | 0.000 | 22:59 |
| Control Test | | 0.077 | 22:59 |
| Air Blank | | 0.000 | 22:59 |
| Diagnostics Check | | OK | 22:59 |

Data Download Date/Time: 06/24/2026 12:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEACH,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: GALINDO MENDOZA

DATE/TIME OF TEST: 05/25/2026 00:30:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:22 |
| Air Blank | | 0.000 | 00:22 |
| Control Test | | 0.078 | 00:22 |
| Air Blank | | 0.000 | 00:23 |
| Subject Sample | # 1 | 0.293 | 00:25 |
| Breath Volume | | 1.539 | 00:25 |
| Air Blank | | 0.000 | 00:26 |
| Air Blank | | 0.000 | 00:28 |
| Subject Sample | # 2 | 0.302 | 00:29 |
| Breath Volume | | 1.402 | 00:29 |
| Air Blank | | 0.000 | 00:30 |
| Control Test | | 0.076 | 00:30 |
| Air Blank | | 0.000 | 00:30 |
| Diagnostics Check | | OK | 00:30 |

Data Download Date/Time: 06/24/2026 12:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEACH,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: CHACON SANTOS

DATE/TIME OF TEST: 05/05/2026 21:16:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:10 |
| Air Blank | | 0.000 | 21:10 |
| Control Test | | 0.080 | 21:11 |
| Air Blank | | 0.000 | 21:11 |
| Subject Sample | # 1 | 0.158 | 21:12 |
| Breath Volume | | 1.625 | 21:12 |
| Air Blank | | 0.000 | 21:12 |
| Air Blank | | 0.000 | 21:14 |
| Subject Sample | # 2 | 0.141 | 21:15 |
| Breath Volume | | 2.667 | 21:15 |
| Air Blank | | 0.000 | 21:15 |
| Control Test | | 0.078 | 21:16 |
| Air Blank | | 0.000 | 21:16 |
| Diagnostics Check | | OK | 21:16 |

Data Download Date/Time: 05/15/2026 10:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 05/13/2026 02:18:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:11 |
| Air Blank | | 0.000 | 02:11 |
| Control Test | | 0.079 | 02:11 |
| Air Blank | | 0.000 | 02:12 |
| Subject Sample | # 1 | 0.191 | 02:13 |
| Breath Volume | | 1.257 | 02:13 |
| Air Blank | | 0.000 | 02:14 |
| Air Blank | | 0.000 | 02:16 |
| Subject Sample | # 2 | 0.183 | 02:17 |
| Breath Volume | | 1.855 | 02:17 |
| Air Blank | | 0.000 | 02:17 |
| Control Test | | 0.078 | 02:17 |
| Air Blank | | 0.000 | 02:18 |
| Diagnostics Check | | OK | 02:18 |

Data Download Date/Time: 05/15/2026 10:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: PINEDA MELGAR

DATE/TIME OF TEST: 05/14/2026 02:39:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:33 |
| Air Blank | | 0.000 | 02:33 |
| Control Test | | 0.080 | 02:34 |
| Air Blank | | 0.000 | 02:34 |
| Subject Sample | # 1 | 0.131 | 02:35 |
| Breath Volume | | 3.277 | 02:35 |
| Air Blank | | 0.000 | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Subject Sample | # 2 | 0.128 | 02:38 |
| Breath Volume | | 3.597 | 02:38 |
| Air Blank | | 0.000 | 02:38 |
| Control Test | | 0.079 | 02:39 |
| Air Blank | | 0.000 | 02:39 |
| Diagnostics Check | | OK | 02:39 |

Data Download Date/Time: 05/15/2026 10:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MUNCH

DATE/TIME OF TEST: 05/14/2026 03:04:24

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:57 |
| Air Blank | | 0.000 | 02:57 |
| Control Test | | 0.079 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Subject Sample | # 1 | 0.203 | 02:59 |
| Breath Volume | | 2.046 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Air Blank | | 0.000 | 03:02 |
| Subject Sample | # 2 | 0.211 | 03:02 |
| Breath Volume | | 2.136 | 03:02 |
| Air Blank | | 0.000 | 03:03 |
| Control Test | | 0.078 | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Diagnostics Check | | OK | 03:04 |

Data Download Date/Time: 05/15/2026 10:42

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CARDOSO

DATE/TIME OF TEST: 05/16/2026 02:38:12

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:31 |
| Air Blank | | 0.000 | 02:32 |
| Control Test | | 0.079 | 02:32 |
| Air Blank | | 0.000 | 02:32 |
| Subject Sample | # 1 | 0.139 | 02:33 |
| Breath Volume | | 3.183 | 02:33 |
| Air Blank | | 0.000 | 02:34 |
| Air Blank | | 0.000 | 02:36 |
| Subject Sample | # 2 | 0.130 | 02:36 |
| Breath Volume | | 2.015 | 02:36 |
| Air Blank | | 0.000 | 02:37 |
| Control Test | | 0.080 | 02:37 |
| Air Blank | | 0.000 | 02:38 |
| Diagnostics Check | | OK | 02:38 |

Data Download Date/Time: 06/23/2026 11:12

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST. WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DACIUS

DATE/TIME OF TEST: 05/16/2026 04:54:42

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:42 | |
| Air Blank | | 0.000 | 04:42 | |
| Control Test | | 0.079 | 04:42 | |
| Air Blank | | 0.000 | 04:43 | |
| Subject Sample | # 1 | 0.175 | 04:46 | Volume Not Met |
| Breath Volume | | 0.250 | 04:46 | |
| Air Blank | | 0.000 | 04:47 | |
| Air Blank | | 0.000 | 04:48 | |
| Subject Sample | # 2 | 0.197 | 04:49 | |
| Breath Volume | | 1.984 | 04:49 | |
| Air Blank | | 0.000 | 04:50 | |
| Air Blank | | 0.000 | 04:52 | |
| Subject Sample | # 3 | 0.205 | 04:53 | |
| Breath Volume | | 1.785 | 04:53 | |
| Air Blank | | 0.000 | 04:53 | |
| Control Test | | 0.078 | 04:54 | |
| Air Blank | | 0.000 | 04:54 | |
| Diagnostics Check | | OK | 04:54 | |

Data Download Date/Time: 06/23/2026 11:12

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:43

SUBJECT'S LAST NAME: RAINS

DATE/TIME OF TEST: 05/16/2026 08:15:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:09 |
| Air Blank | | 0.000 | 08:09 |
| Control Test | | 0.079 | 08:10 |
| Air Blank | | 0.000 | 08:10 |
| Subject Sample | # 1 | 0.162 | 08:11 |
| Breath Volume | | 1.496 | 08:11 |
| Air Blank | | 0.000 | 08:11 |
| Air Blank | | 0.000 | 08:13 |
| Subject Sample | # 2 | 0.171 | 08:14 |
| Breath Volume | | 1.777 | 08:14 |
| Air Blank | | 0.000 | 08:14 |
| Control Test | | 0.078 | 08:15 |
| Air Blank | | 0.000 | 08:15 |
| Diagnostics Check | | OK | 08:15 |

Data Download Date/Time: 06/23/2026 11:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: PATISHTAN PATISHTAN

DATE/TIME OF TEST: 05/17/2026 02:16:04

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:59 | |
| Air Blank | | 0.000 | 01:59 | |
| Control Test | | 0.080 | 02:00 | |
| Air Blank | | 0.000 | 02:00 | |
| Subject Sample | # 1 | 0.233 | 02:03 | |
| Breath Volume | | 2.792 | 02:03 | |
| Air Blank | | 0.000 | 02:04 | |
| Air Blank | | 0.000 | 02:05 | |
| Subject Sample | # 2 | | 02:09 | No Sample Provided |
| Breath Volume | | 0.000 | 02:09 | |
| Air Blank | | 0.000 | 02:09 | |
| Air Blank | | 0.000 | 02:11 | |
| Subject Sample | # 3 | | 02:14 | No Sample Provided |
| Breath Volume | | 0.000 | 02:14 | |
| Air Blank | | 0.000 | 02:15 | |
| Control Test | | 0.078 | 02:15 | |
| Air Blank | | 0.000 | 02:15 | |
| Diagnostics Check | | OK | 02:16 | |

Data Download Date/Time: 06/23/2026 11:13


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: HARWOOD

DATE/TIME OF TEST: 05/17/2026 03:12:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:03 |
| Air Blank | | 0.000 | 03:03 |
| Control Test | | 0.079 | 03:03 |
| Air Blank | | 0.000 | 03:04 |
| Subject Sample | # 1 | 0.141 | 03:07 |
| Breath Volume | | 1.933 | 03:07 |
| Air Blank | | 0.000 | 03:08 |
| Air Blank | | 0.000 | 03:09 |
| Subject Sample | # 2 | 0.139 | 03:10 |
| Breath Volume | | 1.988 | 03:10 |
| Air Blank | | 0.000 | 03:11 |
| Control Test | | 0.079 | 03:11 |
| Air Blank | | 0.000 | 03:12 |
| Diagnostics Check | | OK | 03:12 |

Data Download Date/Time: 06/23/2026 11:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ACEVEDO,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: CORVINGTON

DATE/TIME OF TEST: 05/17/2026 06:29:20

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:22 |
| Air Blank | | 0.000 | 06:23 |
| Control Test | | 0.079 | 06:23 |
| Air Blank | | 0.000 | 06:23 |
| Subject Sample | # 1 | 0.167 | 06:24 |
| Breath Volume | | 1.437 | 06:24 |
| Air Blank | | 0.000 | 06:25 |
| Air Blank | | 0.000 | 06:27 |
| Subject Sample | # 2 | 0.167 | 06:27 |
| Breath Volume | | 1.726 | 06:27 |
| Air Blank | | 0.000 | 06:28 |
| Control Test | | 0.078 | 06:28 |
| Air Blank | | 0.000 | 06:29 |
| Diagnostics Check | | OK | 06:29 |

Data Download Date/Time: 06/23/2026 11:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: SCHARES

DATE/TIME OF TEST: 05/17/2026 23:27:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:21 |
| Air Blank | | 0.000 | 23:21 |
| Control Test | | 0.079 | 23:22 |
| Air Blank | | 0.000 | 23:22 |
| Subject Sample | # 1 | 0.210 | 23:23 |
| Breath Volume | | 1.539 | 23:23 |
| Air Blank | | 0.000 | 23:24 |
| Air Blank | | 0.000 | 23:25 |
| Subject Sample | # 2 | 0.219 | 23:26 |
| Breath Volume | | 1.648 | 23:26 |
| Air Blank | | 0.000 | 23:27 |
| Control Test | | 0.079 | 23:27 |
| Air Blank | | 0.000 | 23:27 |
| Diagnostics Check | | OK | 23:27 |

Data Download Date/Time: 06/23/2026 11:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CARR,X
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HOANG

DATE/TIME OF TEST: 05/18/2026 03:45:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:39 |
| Air Blank | | 0.000 | 03:39 |
| Control Test | | 0.079 | 03:39 |
| Air Blank | | 0.000 | 03:40 |
| Subject Sample | # 1 | 0.000 | 03:41 |
| Breath Volume | | 2.097 | 03:41 |
| Air Blank | | 0.000 | 03:42 |
| Air Blank | | 0.000 | 03:44 |
| Subject Sample | # 2 | 0.000 | 03:44 |
| Breath Volume | | 2.937 | 03:44 |
| Air Blank | | 0.000 | 03:44 |
| Control Test | | 0.079 | 03:45 |
| Air Blank | | 0.000 | 03:45 |
| Diagnostics Check | | OK | 03:45 |

Data Download Date/Time: 06/23/2026 11:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/19/2026 01:46:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:33 | |
| Air Blank | | 0.000 | 01:34 | |
| Control Test | | 0.080 | 01:34 | |
| Air Blank | | 0.000 | 01:35 | |
| Subject Sample | # 1 | 0.170 | 01:38 | Volume Not Met |
| Breath Volume | | 0.464 | 01:38 | |
| Air Blank | | 0.000 | 01:38 | |
| Air Blank | | 0.000 | 01:40 | |
| Subject Sample | # 2 | 0.162 | 01:42 | No .020 agreement |
| Breath Volume | | 1.160 | 01:42 | |
| Air Blank | | 0.000 | 01:42 | |
| Air Blank | | 0.000 | 01:44 | |
| Subject Sample | # 3 | 0.184 | 01:45 | No .020 agreement |
| Breath Volume | | 1.132 | 01:45 | |
| Air Blank | | 0.000 | 01:45 | |
| Control Test | | 0.078 | 01:46 | |
| Air Blank | | 0.000 | 01:46 | |
| Diagnostics Check | | OK | 01:46 | |

Data Download Date/Time: 06/23/2026 11:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/19/2026 01:55:40

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:47 |
| Air Blank | | 0.000 | 01:48 |
| Control Test | | 0.079 | 01:48 |
| Air Blank | | 0.000 | 01:49 |
| Subject Sample | # 1 | 0.169 | 01:49 |
| Breath Volume | | 1.218 | 01:49 |
| Air Blank | | 0.000 | 01:50 |
| Air Blank | | 0.000 | 01:52 |
| Subject Sample | # 2 | 0.163 | 01:54 |
| Breath Volume | | 1.410 | 01:54 |
| Air Blank | | 0.000 | 01:54 |
| Control Test | | 0.078 | 01:55 |
| Air Blank | | 0.000 | 01:55 |
| Diagnostics Check | | OK | 01:55 |

Data Download Date/Time: 06/23/2026 11:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SHAY

DATE/TIME OF TEST: 05/20/2026 23:27:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:20 |
| Air Blank | | 0.000 | 23:21 |
| Control Test | | 0.080 | 23:21 |
| Air Blank | | 0.000 | 23:21 |
| Subject Sample | # 1 | 0.056 | 23:22 |
| Breath Volume | | 2.535 | 23:22 |
| Air Blank | | 0.000 | 23:23 |
| Air Blank | | 0.000 | 23:25 |
| Subject Sample | # 2 | 0.057 | 23:25 |
| Breath Volume | | 2.238 | 23:25 |
| Air Blank | | 0.000 | 23:26 |
| Control Test | | 0.079 | 23:26 |
| Air Blank | | 0.000 | 23:27 |
| Diagnostics Check | | OK | 23:27 |

Data Download Date/Time: 06/23/2026 11:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: DE JESUS SANTANA

DATE/TIME OF TEST: 05/22/2026 03:51:19

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:44 |
| Air Blank | | 0.000 | 03:45 |
| Control Test | | 0.079 | 03:45 |
| Air Blank | | 0.000 | 03:46 |
| Subject Sample | # 1 | 0.008 | 03:47 |
| Breath Volume | | 2.210 | 03:47 |
| Air Blank | | 0.000 | 03:47 |
| Air Blank | | 0.000 | 03:49 |
| Subject Sample | # 2 | 0.007 | 03:49 |
| Breath Volume | | 1.496 | 03:49 |
| Air Blank | | 0.000 | 03:50 |
| Control Test | | 0.079 | 03:50 |
| Air Blank | | 0.000 | 03:51 |
| Diagnostics Check | | OK | 03:51 |

Data Download Date/Time: 06/23/2026 11:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: OILER,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 05/22/2026 20:12:36

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 19:57 | |
| Air Blank | | 0.000 | 19:58 | |
| Control Test | | 0.079 | 19:58 | |
| Air Blank | | 0.000 | 19:58 | |
| Subject Sample | # 1 | 0.226 | 19:59 | |
| Breath Volume | | 1.226 | 19:59 | |
| Air Blank | | 0.000 | 20:00 | |
| Air Blank | | 0.000 | 20:02 | |
| Subject Sample | # 2 | | 20:05 | No Sample Provided |
| Breath Volume | | 0.000 | 20:05 | |
| Air Blank | | 0.000 | 20:06 | |
| Air Blank | | 0.000 | 20:08 | |
| Subject Sample | # 3 | | 20:11 | No Sample Provided |
| Breath Volume | | 0.000 | 20:11 | |
| Air Blank | | 0.000 | 20:11 | |
| Control Test | | 0.079 | 20:12 | |
| Air Blank | | 0.000 | 20:12 | |
| Diagnostics Check | | OK | 20:12 | |

Data Download Date/Time: 06/23/2026 11:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: OILER,B
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 05/22/2026 20:29:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:22 |
| Air Blank | | 0.000 | 20:22 |
| Control Test | | 0.079 | 20:22 |
| Air Blank | | 0.000 | 20:23 |
| Subject Sample | # 1 | 0.216 | 20:24 |
| Breath Volume | | 2.187 | 20:24 |
| Air Blank | | 0.000 | 20:25 |
| Air Blank | | 0.000 | 20:26 |
| Subject Sample | # 2 | 0.209 | 20:27 |
| Breath Volume | | 1.648 | 20:27 |
| Air Blank | | 0.000 | 20:28 |
| Control Test | | 0.078 | 20:28 |
| Air Blank | | 0.000 | 20:28 |
| Diagnostics Check | | OK | 20:29 |

Data Download Date/Time: 06/23/2026 11:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BUCK

DATE/TIME OF TEST: 05/23/2026 02:15:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:07 |
| Air Blank | | 0.000 | 02:08 |
| Control Test | | 0.079 | 02:08 |
| Air Blank | | 0.000 | 02:09 |
| Subject Sample | # 1 | 0.195 | 02:10 |
| Breath Volume | | 2.441 | 02:10 |
| Air Blank | | 0.000 | 02:10 |
| Air Blank | | 0.000 | 02:12 |
| Subject Sample | # 2 | 0.186 | 02:14 |
| Breath Volume | | 1.683 | 02:14 |
| Air Blank | | 0.000 | 02:14 |
| Control Test | | 0.079 | 02:15 |
| Air Blank | | 0.000 | 02:15 |
| Diagnostics Check | | OK | 02:15 |

Data Download Date/Time: 06/23/2026 11:14

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/23/2026 02:44:46

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:37 |
| Air Blank | | 0.000 | 02:37 |
| Control Test | | 0.078 | 02:38 |
| Air Blank | | 0.000 | 02:38 |
| Subject Sample | # 1 | 0.181 | 02:40 |
| Breath Volume | | 1.742 | 02:40 |
| Air Blank | | 0.000 | 02:40 |
| Air Blank | | 0.000 | 02:42 |
| Subject Sample | # 2 | 0.171 | 02:43 |
| Breath Volume | | 2.523 | 02:43 |
| Air Blank | | 0.000 | 02:43 |
| Control Test | | 0.079 | 02:44 |
| Air Blank | | 0.000 | 02:44 |
| Diagnostics Check | | OK | 02:44 |

Data Download Date/Time: 06/23/2026 11:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/23/2026 05:08:14

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:51 | |
| Air Blank | | 0.000 | 04:52 | |
| Control Test | | 0.079 | 04:52 | |
| Air Blank | | 0.000 | 04:53 | |
| Subject Sample | # 1 | 0.119 | 04:56 | Volume Not Met |
| Breath Volume | | 0.113 | 04:56 | |
| Air Blank | | 0.000 | 04:56 | |
| Air Blank | | 0.000 | 04:58 | |
| Subject Sample | # 2 | 0.141 | 05:00 | |
| Breath Volume | | 2.113 | 05:00 | |
| Air Blank | | 0.000 | 05:01 | |
| Air Blank | | 0.000 | 05:03 | |
| Subject Sample | # 3 | 0.111 | 05:06 | Volume Not Met |
| Breath Volume | | 0.000 | 05:06 | |
| Air Blank | | 0.000 | 05:07 | |
| Control Test | | 0.078 | 05:07 | |
| Air Blank | | 0.000 | 05:08 | |
| Diagnostics Check | | OK | 05:08 | |

Data Download Date/Time: 06/23/2026 11:15


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: BRITTON

DATE/TIME OF TEST: 05/24/2026 00:29:53

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:22 |
| Air Blank | | 0.000 | 00:23 |
| Control Test | | 0.079 | 00:23 |
| Air Blank | | 0.000 | 00:24 |
| Subject Sample | # 1 | 0.175 | 00:25 |
| Breath Volume | | 2.199 | 00:25 |
| Air Blank | | 0.000 | 00:25 |
| Air Blank | | 0.000 | 00:27 |
| Subject Sample | # 2 | 0.175 | 00:28 |
| Breath Volume | | 2.742 | 00:28 |
| Air Blank | | 0.000 | 00:28 |
| Control Test | | 0.079 | 00:29 |
| Air Blank | | 0.000 | 00:29 |
| Diagnostics Check | | OK | 00:29 |

Data Download Date/Time: 06/23/2026 11:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 05/27/2026 01:42:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:35 |
| Air Blank | | 0.000 | 01:36 |
| Control Test | | 0.079 | 01:36 |
| Air Blank | | 0.000 | 01:36 |
| Subject Sample | # 1 | 0.152 | 01:37 |
| Breath Volume | | 1.832 | 01:37 |
| Air Blank | | 0.000 | 01:38 |
| Air Blank | | 0.000 | 01:40 |
| Subject Sample | # 2 | 0.151 | 01:40 |
| Breath Volume | | 1.617 | 01:40 |
| Air Blank | | 0.000 | 01:41 |
| Control Test | | 0.079 | 01:41 |
| Air Blank | | 0.000 | 01:42 |
| Diagnostics Check | | OK | 01:42 |

Data Download Date/Time: 06/23/2026 11:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORRIS,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: HERRERA RIOS

DATE/TIME OF TEST: 05/28/2026 04:53:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:46 |
| Air Blank | | 0.000 | 04:46 |
| Control Test | | 0.079 | 04:47 |
| Air Blank | | 0.000 | 04:47 |
| Subject Sample | # 1 | 0.224 | 04:48 |
| Breath Volume | | 2.269 | 04:48 |
| Air Blank | | 0.000 | 04:49 |
| Air Blank | | 0.000 | 04:50 |
| Subject Sample | # 2 | 0.220 | 04:51 |
| Breath Volume | | 1.355 | 04:51 |
| Air Blank | | 0.000 | 04:52 |
| Control Test | | 0.078 | 04:52 |
| Air Blank | | 0.000 | 04:53 |
| Diagnostics Check | | OK | 04:53 |

Data Download Date/Time: 06/23/2026 11:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/30/2026 05:00:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:53 |
| Air Blank | | 0.000 | 04:53 |
| Control Test | | 0.079 | 04:54 |
| Air Blank | | 0.000 | 04:54 |
| Subject Sample | # 1 | 0.099 | 04:55 |
| Breath Volume | | 1.496 | 04:55 |
| Air Blank | | 0.000 | 04:56 |
| Air Blank | | 0.000 | 04:58 |
| Subject Sample | # 2 | 0.098 | 04:59 |
| Breath Volume | | 1.875 | 04:59 |
| Air Blank | | 0.000 | 04:59 |
| Control Test | | 0.079 | 04:59 |
| Air Blank | | 0.000 | 05:00 |
| Diagnostics Check | | OK | 05:00 |

Data Download Date/Time: 06/23/2026 11:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/30/2026 06:34:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:27 |
| Air Blank | | 0.000 | 06:27 |
| Control Test | | 0.079 | 06:28 |
| Air Blank | | 0.000 | 06:28 |
| Subject Sample | # 1 | 0.072 | 06:29 |
| Breath Volume | | 1.824 | 06:29 |
| Air Blank | | 0.000 | 06:29 |
| Air Blank | | 0.000 | 06:31 |
| Subject Sample | # 2 | 0.072 | 06:32 |
| Breath Volume | | 2.230 | 06:32 |
| Air Blank | | 0.000 | 06:33 |
| Control Test | | 0.079 | 06:33 |
| Air Blank | | 0.000 | 06:34 |
| Diagnostics Check | | OK | 06:34 |

Data Download Date/Time: 06/23/2026 11:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST.WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: LOZADA VARGAS

DATE/TIME OF TEST: 05/31/2026 01:13:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:07 |
| Air Blank | | 0.000 | 01:07 |
| Control Test | | 0.079 | 01:07 |
| Air Blank | | 0.000 | 01:08 |
| Subject Sample | # 1 | 0.192 | 01:09 |
| Breath Volume | | 2.164 | 01:09 |
| Air Blank | | 0.000 | 01:09 |
| Air Blank | | 0.000 | 01:11 |
| Subject Sample | # 2 | 0.195 | 01:12 |
| Breath Volume | | 1.972 | 01:12 |
| Air Blank | | 0.000 | 01:12 |
| Control Test | | 0.079 | 01:13 |
| Air Blank | | 0.000 | 01:13 |
| Diagnostics Check | | OK | 01:13 |

Data Download Date/Time: 06/23/2026 11:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: MURRIAS

DATE/TIME OF TEST: 05/31/2026 02:57:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:50 |
| Air Blank | | 0.000 | 02:50 |
| Control Test | | 0.079 | 02:51 |
| Air Blank | | 0.000 | 02:51 |
| Subject Sample | # 1 | 0.000 | 02:52 |
| Breath Volume | | 2.539 | 02:52 |
| Air Blank | | 0.000 | 02:53 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 2 | 0.000 | 02:55 |
| Breath Volume | | 2.753 | 02:55 |
| Air Blank | | 0.000 | 02:56 |
| Control Test | | 0.078 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Diagnostics Check | | OK | 02:57 |

Data Download Date/Time: 06/23/2026 11:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: CUCURELLA

DATE/TIME OF TEST: 05/31/2026 04:08:44

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:02 |
| Air Blank | | 0.000 | 04:02 |
| Control Test | | 0.079 | 04:02 |
| Air Blank | | 0.000 | 04:03 |
| Subject Sample | # 1 | 0.207 | 04:04 |
| Breath Volume | | 1.746 | 04:04 |
| Air Blank | | 0.000 | 04:04 |
| Air Blank | | 0.000 | 04:06 |
| Subject Sample | # 2 | 0.206 | 04:07 |
| Breath Volume | | 1.773 | 04:07 |
| Air Blank | | 0.000 | 04:07 |
| Control Test | | 0.078 | 04:08 |
| Air Blank | | 0.000 | 04:08 |
| Diagnostics Check | | OK | 04:08 |

Data Download Date/Time: 06/23/2026 11:16

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007484

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,PRISCILLA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: HONG

DATE/TIME OF TEST: 05/31/2026 04:31:35

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:24 |
| Air Blank | | 0.000 | 04:25 |
| Control Test | | 0.079 | 04:25 |
| Air Blank | | 0.000 | 04:25 |
| Subject Sample | # 1 | 0.161 | 04:26 |
| Breath Volume | | 1.406 | 04:26 |
| Air Blank | | 0.000 | 04:27 |
| Air Blank | | 0.000 | 04:29 |
| Subject Sample | # 2 | 0.170 | 04:29 |
| Breath Volume | | 1.417 | 04:29 |
| Air Blank | | 0.000 | 04:30 |
| Control Test | | 0.078 | 04:30 |
| Air Blank | | 0.000 | 04:31 |
| Diagnostics Check | | OK | 04:31 |

Data Download Date/Time: 06/23/2026 11:16

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007491

INSTRUMENT REGISTERED TO: ORANGE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRISERA,M,J
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: TAVERAS,T
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 05/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3652508A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: BARKSDALE

DATE/TIME OF TEST: 05/28/2026 21:29:34

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:17 | |
| Air Blank | | 0.000 | 21:17 | |
| Control Test | | 0.077 | 21:18 | |
| Air Blank | | 0.000 | 21:18 | |
| Subject Sample | # 1 | 0.206 | 21:21 | Volume Not Met |
| Breath Volume | | 0.656 | 21:21 | |
| Air Blank | | 0.000 | 21:22 | |
| Air Blank | | 0.000 | 21:24 | |
| Subject Sample | # 2 | 0.207 | 21:27 | Volume Not Met |
| Breath Volume | | 0.761 | 21:27 | |
| Air Blank | | 0.000 | 21:28 | |
| Control Test | | 0.076 | 21:28 | |
| Air Blank | | 0.000 | 21:29 | |
| Diagnostics Check | | OK | 21:29 | |

Data Download Date/Time: 06/27/2026 04:54

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,ROBERT,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLEN,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: BUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/23/2026 20:06:04

| | | | |
|---|---|---|---|
| Diagnostics Check | Fail | 20:05 | Internal Printer Test |
| Air Blank | 0.000 | 20:06 | |

Data Download Date/Time: 05/29/2026 13:03

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,ROBERT,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLEN,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: BUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/23/2026 20:08:14

| | | | |
|---|---|---|---|
| Diagnostics Check | Fail | 20:07 | Internal Printer Test |
| Air Blank | 0.000 | 20:08 | |

Data Download Date/Time: 05/29/2026 13:03

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007499

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,ROBERT,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLEN,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: BUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 05/23/2026 20:16:54

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:10 |
| Air Blank | | 0.000 | 20:10 |
| Control Test | | 0.078 | 20:10 |
| Air Blank | | 0.000 | 20:11 |
| Subject Sample | # 1 | 0.054 | 20:12 |
| Breath Volume | | 1.453 | 20:12 |
| Air Blank | | 0.000 | 20:13 |
| Air Blank | | 0.000 | 20:14 |
| Subject Sample | # 2 | 0.063 | 20:15 |
| Breath Volume | | 3.652 | 20:15 |
| Air Blank | | 0.000 | 20:15 |
| Control Test | | 0.079 | 20:16 |
| Air Blank | | 0.000 | 20:16 |
| Diagnostics Check | | OK | 20:16 |

Data Download Date/Time: 05/29/2026 13:03

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLIARD,TYLER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WHITE-SCOTT

DATE/TIME OF TEST: 05/06/2026 03:31:09

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:28 | |
| Air Blank | | 0.000 | 03:28 | |
| Control Test | | 0.080 | 03:28 | |
| Air Blank | | 0.000 | 03:29 | |
| Subject Sample | # 1 | | 03:29 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:29 | |
| Air Blank | | 0.000 | 03:30 | |
| Control Test | | 0.080 | 03:30 | |
| Air Blank | | 0.000 | 03:31 | |
| Diagnostics Check | | OK | 03:31 | |

Data Download Date/Time: 05/10/2026 20:32

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ARMSTEAD,CAMERON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: DEFORD

DATE/TIME OF TEST: 05/08/2026 18:16:11

| | | | |
|---|---|---|---|
| Diagnostics Check | OK | 18:14 | |
| Air Blank | 0.000 | 18:15 | |
| Control Test | 0.031 | 18:15 | RFI Detect |
| Air Blank | 0.000 | 18:16 | |

Data Download Date/Time: 05/10/2026 20:32


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ARMSTEAD,CAMERON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: DEFORD

DATE/TIME OF TEST: 05/08/2026 18:23:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:18 |
| Air Blank | | 0.000 | 18:18 |
| Control Test | | 0.080 | 18:18 |
| Air Blank | | 0.000 | 18:19 |
| Subject Sample | # 1 | 0.000 | 18:19 |
| Breath Volume | | 2.945 | 18:19 |
| Air Blank | | 0.000 | 18:20 |
| Air Blank | | 0.000 | 18:22 |
| Subject Sample | # 2 | 0.000 | 18:22 |
| Breath Volume | | 2.906 | 18:22 |
| Air Blank | | 0.000 | 18:23 |
| Control Test | | 0.079 | 18:23 |
| Air Blank | | 0.000 | 18:23 |
| Diagnostics Check | | OK | 18:23 |

Data Download Date/Time: 05/10/2026 20:32


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ARMSTEAD,CAMERON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 05/08/2026 19:02:09

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:56 |
| Air Blank | | 0.000 | 18:56 |
| Control Test | | 0.080 | 18:57 |
| Air Blank | | 0.000 | 18:57 |
| Subject Sample | # 1 | 0.000 | 18:57 |
| Breath Volume | | 2.535 | 18:57 |
| Air Blank | | 0.000 | 18:58 |
| Air Blank | | 0.000 | 19:00 |
| Subject Sample | # 2 | 0.000 | 19:00 |
| Breath Volume | | 2.632 | 19:00 |
| Air Blank | | 0.000 | 19:01 |
| Control Test | | 0.079 | 19:01 |
| Air Blank | | 0.000 | 19:02 |
| Diagnostics Check | | OK | 19:02 |

Data Download Date/Time: 05/10/2026 20:32

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLIARD,TYLER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 05/09/2026 03:47:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:39 |
| Air Blank | | 0.000 | 03:39 |
| Control Test | | 0.079 | 03:39 |
| Air Blank | | 0.000 | 03:40 |
| Subject Sample | # 1 | 0.143 | 03:42 |
| Breath Volume | | 1.320 | 03:42 |
| Air Blank | | 0.000 | 03:43 |
| Air Blank | | 0.000 | 03:44 |
| Subject Sample | # 2 | 0.146 | 03:45 |
| Breath Volume | | 1.542 | 03:45 |
| Air Blank | | 0.000 | 03:46 |
| Control Test | | 0.078 | 03:46 |
| Air Blank | | 0.000 | 03:47 |
| Diagnostics Check | | OK | 03:47 |

Data Download Date/Time: 05/10/2026 20:32

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLIARD,TYLER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 05/11/2026 00:30:28

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:23 | |
| Air Blank | | 0.000 | 00:24 | |
| Control Test | | 0.079 | 00:24 | |
| Air Blank | | 0.000 | 00:25 | |
| Subject Sample | # 1 | 0.039 | 00:25 | |
| Breath Volume | | 1.414 | 00:25 | |
| Air Blank | | 0.000 | 00:26 | |
| Air Blank | | 0.000 | 00:27 | |
| Subject Sample | # 2 | 0.042 | 00:28 | |
| Breath Volume | | 1.511 | 00:28 | |
| Air Blank | | 0.000 | 00:29 | |
| Control Test | | 0.080 | 00:29 | |
| Air Blank | | 0.000 | 00:29 | |
| Diagnostics Check | | Fail | 00:30 | DSP Test |
| Air Blank | | 0.000 | 00:30 | |

Data Download Date/Time: 05/29/2026 06:21

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ARMSTEAD,CAMERON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: VICENTY

DATE/TIME OF TEST: 05/13/2026 20:44:23

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:38 | |
| Air Blank | | 0.000 | 20:38 | |
| Control Test | | 0.080 | 20:38 | |
| Air Blank | | 0.000 | 20:39 | |
| Subject Sample | # 1 | 0.000 | 20:39 | |
| Breath Volume | | 2.773 | 20:39 | |
| Air Blank | | 0.000 | 20:40 | |
| Air Blank | | 0.000 | 20:42 | |
| Subject Sample | # 2 | 0.000 | 20:42 | |
| Breath Volume | | 2.406 | 20:42 | |
| Air Blank | | 0.000 | 20:42 | |
| Control Test | | 0.079 | 20:43 | |
| Air Blank | | 0.000 | 20:43 | |
| Diagnostics Check | | Fail | 20:43 | DSP Test |
| Air Blank | | 0.000 | 20:44 | |

Data Download Date/Time: 05/29/2026 06:21

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 12:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: TATTERSALL

DATE/TIME OF TEST: 05/15/2026 13:40:11

| | | | |
|---|---|---|---|
| Diagnostics Check | Fail | 13:39 | DSP Test |
| Air Blank | 0.000 | 13:40 | |

Data Download Date/Time: 05/29/2026 06:21

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 12:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: TATTERSALL

DATE/TIME OF TEST: 05/15/2026 13:49:02

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 13:42 | |
| Air Blank | | 0.000 | 13:42 | |
| Control Test | | 0.080 | 13:43 | |
| Air Blank | | 0.000 | 13:43 | |
| Subject Sample | # 1 | 0.000 | 13:44 | |
| Breath Volume | | 1.531 | 13:44 | |
| Air Blank | | 0.000 | 13:44 | |
| Air Blank | | 0.000 | 13:46 | |
| Subject Sample | # 2 | 0.000 | 13:47 | |
| Breath Volume | | 1.628 | 13:47 | |
| Air Blank | | 0.000 | 13:47 | |
| Control Test | | 0.080 | 13:47 | |
| Air Blank | | 0.000 | 13:48 | |
| Diagnostics Check | | Fail | 13:48 | DSP Test |
| Air Blank | | 0.000 | 13:49 | |

Data Download Date/Time: 05/29/2026 06:21

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,RODNEY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:51

SUBJECT'S LAST NAME: SUMAR

DATE/TIME OF TEST: 05/01/2026 08:17:49

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:11 |
| Air Blank | | 0.000 | 08:11 |
| Control Test | | 0.081 | 08:12 |
| Air Blank | | 0.000 | 08:12 |
| Subject Sample | # 1 | 0.099 | 08:13 |
| Breath Volume | | 1.707 | 08:13 |
| Air Blank | | 0.000 | 08:13 |
| Air Blank | | 0.000 | 08:15 |
| Subject Sample | # 2 | 0.102 | 08:16 |
| Breath Volume | | 1.429 | 08:16 |
| Air Blank | | 0.000 | 08:16 |
| Control Test | | 0.081 | 08:17 |
| Air Blank | | 0.000 | 08:17 |
| Diagnostics Check | | OK | 08:17 |

Data Download Date/Time: 05/10/2026 20:24

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: NEW,STEVEN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: MAST

DATE/TIME OF TEST: 05/02/2026 03:59:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:51 |
| Air Blank | | 0.000 | 03:52 |
| Control Test | | 0.081 | 03:52 |
| Air Blank | | 0.000 | 03:53 |
| Subject Sample | # 1 | 0.206 | 03:54 |
| Breath Volume | | 1.910 | 03:54 |
| Air Blank | | 0.000 | 03:54 |
| Air Blank | | 0.000 | 03:56 |
| Subject Sample | # 2 | 0.200 | 03:57 |
| Breath Volume | | 2.027 | 03:57 |
| Air Blank | | 0.000 | 03:58 |
| Control Test | | 0.081 | 03:58 |
| Air Blank | | 0.000 | 03:58 |
| Diagnostics Check | | OK | 03:59 |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: ALVELO RODRIGUEZ

DATE/TIME OF TEST: 05/02/2026 10:41:03

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 10:35 |
| Air Blank | | 0.000 | 10:35 |
| Control Test | | 0.082 | 10:35 |
| Air Blank | | 0.000 | 10:36 |
| Subject Sample | # 1 | 0.000 | 10:36 |
| Breath Volume | | 3.042 | 10:36 |
| Air Blank | | 0.000 | 10:37 |
| Air Blank | | 0.000 | 10:39 |
| Subject Sample | # 2 | 0.000 | 10:39 |
| Breath Volume | | 2.367 | 10:39 |
| Air Blank | | 0.000 | 10:40 |
| Control Test | | 0.082 | 10:40 |
| Air Blank | | 0.000 | 10:40 |
| Diagnostics Check | | OK | 10:41 |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALACIOS-FLORES,CRISTIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/03/2026 08:53:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:47 |
| Air Blank | | 0.000 | 08:47 |
| Control Test | | 0.081 | 08:48 |
| Air Blank | | 0.000 | 08:48 |
| Subject Sample | # 1 | 0.000 | 08:49 |
| Breath Volume | | 2.738 | 08:49 |
| Air Blank | | 0.000 | 08:49 |
| Air Blank | | 0.000 | 08:51 |
| Subject Sample | # 2 | 0.000 | 08:51 |
| Breath Volume | | 1.796 | 08:51 |
| Air Blank | | 0.000 | 08:52 |
| Control Test | | 0.080 | 08:52 |
| Air Blank | | 0.000 | 08:53 |
| Diagnostics Check | | OK | 08:53 |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/03/2026 13:22:06

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 13:16 |
| Air Blank | | 0.000 | 13:16 |
| Control Test | | 0.081 | 13:16 |
| Air Blank | | 0.000 | 13:17 |
| Subject Sample | # 1 | 0.000 | 13:17 |
| Breath Volume | | 2.281 | 13:17 |
| Air Blank | | 0.000 | 13:18 |
| Air Blank | | 0.000 | 13:20 |
| Subject Sample | # 2 | 0.000 | 13:20 |
| Breath Volume | | 2.292 | 13:20 |
| Air Blank | | 0.000 | 13:21 |
| Control Test | | 0.081 | 13:21 |
| Air Blank | | 0.000 | 13:22 |
| Diagnostics Check | | OK | 13:22 |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FILLMORE,DONALD
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/04/2026 03:27:55

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:24 | |
| Air Blank | | 0.000 | 03:25 | |
| Control Test | | 0.081 | 03:25 | |
| Air Blank | | 0.000 | 03:26 | |
| Subject Sample | # 1 | | 03:26 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:26 | |
| Air Blank | | 0.000 | 03:26 | |
| Control Test | | 0.081 | 03:27 | |
| Air Blank | | 0.000 | 03:27 | |
| Diagnostics Check | | OK | 03:27 | |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: TATTERSALL

DATE/TIME OF TEST: 05/04/2026 13:52:00

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 13:46 |
| Air Blank | | 0.000 | 13:46 |
| Control Test | | 0.081 | 13:46 |
| Air Blank | | 0.000 | 13:47 |
| Subject Sample | # 1 | 0.000 | 13:47 |
| Breath Volume | | 1.324 | 13:47 |
| Air Blank | | 0.000 | 13:48 |
| Air Blank | | 0.000 | 13:50 |
| Subject Sample | # 2 | 0.000 | 13:50 |
| Breath Volume | | 1.125 | 13:50 |
| Air Blank | | 0.000 | 13:51 |
| Control Test | | 0.081 | 13:51 |
| Air Blank | | 0.000 | 13:51 |
| Diagnostics Check | | OK | 13:52 |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: PUJOL

DATE/TIME OF TEST: 05/05/2026 16:25:31

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 16:19 | |
| Air Blank | | 0.000 | 16:20 | |
| Control Test | | 0.081 | 16:20 | |
| Air Blank | | 0.000 | 16:21 | |
| Subject Sample | # 1 | 0.000 | 16:21 | |
| Breath Volume | | 1.523 | 16:21 | |
| Air Blank | | 0.000 | 16:21 | |
| Air Blank | | 0.000 | 16:23 | |
| Subject Sample | # 2 | 0.000 | 16:24 | |
| Breath Volume | | 1.710 | 16:24 | |
| Air Blank | | | 16:24 | Ambient Fail |
| Air Blank | | | 16:25 | Purge Fail |

Data Download Date/Time: 05/10/2026 20:25


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,NEAL,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: YARBROUGH,G
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 05/07/2026 01:56:36

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:50 |
| Air Blank | | 0.000 | 01:50 |
| Control Test | | 0.081 | 01:50 |
| Air Blank | | 0.000 | 01:51 |
| Subject Sample | # 1 | 0.079 | 01:51 |
| Breath Volume | | 1.203 | 01:51 |
| Air Blank | | 0.000 | 01:52 |
| Air Blank | | 0.000 | 01:54 |
| Subject Sample | # 2 | 0.074 | 01:55 |
| Breath Volume | | 2.121 | 01:55 |
| Air Blank | | 0.000 | 01:55 |
| Control Test | | 0.082 | 01:55 |
| Air Blank | | 0.000 | 01:56 |
| Diagnostics Check | | OK | 01:56 |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,N,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MCDANIEL,N
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: DIMURO

DATE/TIME OF TEST: 05/07/2026 22:48:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 22:42 |
| Air Blank | | 0.000 | 22:42 |
| Control Test | | 0.081 | 22:43 |
| Air Blank | | 0.000 | 22:43 |
| Subject Sample | # 1 | 0.000 | 22:44 |
| Breath Volume | | 2.468 | 22:44 |
| Air Blank | | 0.000 | 22:44 |
| Air Blank | | 0.000 | 22:46 |
| Subject Sample | # 2 | 0.000 | 22:46 |
| Breath Volume | | 2.507 | 22:46 |
| Air Blank | | 0.000 | 22:47 |
| Control Test | | 0.081 | 22:47 |
| Air Blank | | 0.000 | 22:48 |
| Diagnostics Check | | OK | 22:48 |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GAMEZ,ALEJANDRO
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: UEBELACKER

DATE/TIME OF TEST: 05/08/2026 15:20:21

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 15:16 | |
| Air Blank | | 0.000 | 15:16 | |
| Control Test | | 0.082 | 15:16 | |
| Air Blank | | 0.000 | 15:17 | |
| Subject Sample | # 1 | | 15:18 | Subject Test Refused |
| Breath Volume | | 0.000 | 15:18 | |
| Air Blank | | 0.000 | 15:19 | |
| Control Test | | 0.083 | 15:19 | |
| Air Blank | | 0.000 | 15:20 | |
| Diagnostics Check | | OK | 15:20 | |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MILLER,PRESTON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 04/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 05/09/2026 22:28:19

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 22:25 | |
| Air Blank | | 0.000 | 22:25 | |
| Control Test | | 0.081 | 22:26 | |
| Air Blank | | 0.000 | 22:26 | |
| Subject Sample | # 1 | | 22:26 | Subject Test Refused |
| Breath Volume | | 0.000 | 22:26 | |
| Air Blank | | 0.000 | 22:27 | |
| Control Test | | 0.081 | 22:27 | |
| Air Blank | | 0.000 | 22:28 | |
| Diagnostics Check | | OK | 22:28 | |

Data Download Date/Time: 05/10/2026 20:25

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLARD,TYLER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 05/11/2026 00:18:56

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:15 | |
| Air Blank | | 0.000 | 00:16 | |
| Control Test | | 0.081 | 00:16 | |
| Air Blank | | 0.000 | 00:16 | |
| Subject Sample | # 1 | | 00:17 | Interferent Detect |
| Breath Volume | | 1.597 | 00:17 | |
| Air Blank | | | 00:18 | Purge Fail |
| Air Blank | | | 00:18 | Purge Fail |

Data Download Date/Time: 06/07/2026 20:30


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: KNIGHT',J
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 05/11/2026 01:01:28

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:55 |
| Air Blank | | 0.000 | 00:55 |
| Control Test | | 0.082 | 00:56 |
| Air Blank | | 0.000 | 00:56 |
| Subject Sample | # 1 | 0.128 | 00:56 |
| Breath Volume | | 1.269 | 00:56 |
| Air Blank | | 0.000 | 00:57 |
| Air Blank | | 0.000 | 00:59 |
| Subject Sample | # 2 | 0.131 | 00:59 |
| Breath Volume | | 2.546 | 00:59 |
| Air Blank | | 0.000 | 01:00 |
| Control Test | | 0.081 | 01:00 |
| Air Blank | | 0.000 | 01:01 |
| Diagnostics Check | | OK | 01:01 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,ANTHONY,K
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROBERTS,P
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: KIRKEJARD

DATE/TIME OF TEST: 05/11/2026 10:56:30

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 10:50 | |
| Air Blank | | 0.000 | 10:51 | |
| Control Test | | 0.081 | 10:51 | |
| Air Blank | | 0.000 | 10:52 | |
| Subject Sample | # 1 | 0.000 | 10:52 | |
| Breath Volume | | 3.093 | 10:52 | |
| Air Blank | | 0.000 | 10:53 | |
| Air Blank | | 0.000 | 10:55 | |
| Subject Sample | # 2 | 0.000 | 10:55 | |
| Breath Volume | | 3.125 | 10:55 | |
| Air Blank | | | 10:56 | RFI Detect |
| Air Blank | | 0.000 | 10:56 | |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUTLER,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROBERTS,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: KIRKEGARD

DATE/TIME OF TEST: 05/11/2026 11:08:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 11:02 |
| Air Blank | | 0.000 | 11:03 |
| Control Test | | 0.081 | 11:03 |
| Air Blank | | 0.000 | 11:03 |
| Subject Sample | # 1 | 0.000 | 11:04 |
| Breath Volume | | 2.113 | 11:04 |
| Air Blank | | 0.000 | 11:04 |
| Air Blank | | 0.000 | 11:06 |
| Subject Sample | # 2 | 0.000 | 11:07 |
| Breath Volume | | 1.933 | 11:07 |
| Air Blank | | 0.000 | 11:07 |
| Control Test | | 0.081 | 11:08 |
| Air Blank | | 0.000 | 11:08 |
| Diagnostics Check | | OK | 11:08 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: PUJOL

DATE/TIME OF TEST: 05/14/2026 15:45:59

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:40 |
| Air Blank | | 0.000 | 15:40 |
| Control Test | | 0.082 | 15:40 |
| Air Blank | | 0.000 | 15:41 |
| Subject Sample | # 1 | 0.000 | 15:41 |
| Breath Volume | | 1.644 | 15:41 |
| Air Blank | | 0.000 | 15:42 |
| Air Blank | | 0.000 | 15:44 |
| Subject Sample | # 2 | 0.000 | 15:44 |
| Breath Volume | | 1.515 | 15:44 |
| Air Blank | | 0.000 | 15:45 |
| Control Test | | 0.082 | 15:45 |
| Air Blank | | 0.000 | 15:45 |
| Diagnostics Check | | OK | 15:45 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALACIOS-FLORES,CRISTIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/16/2026 08:59:17

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:53 |
| Air Blank | | 0.000 | 08:54 |
| Control Test | | 0.081 | 08:54 |
| Air Blank | | 0.000 | 08:54 |
| Subject Sample | # 1 | 0.000 | 08:55 |
| Breath Volume | | 2.320 | 08:55 |
| Air Blank | | 0.000 | 08:55 |
| Air Blank | | 0.000 | 08:57 |
| Subject Sample | # 2 | 0.000 | 08:57 |
| Breath Volume | | 2.160 | 08:57 |
| Air Blank | | 0.000 | 08:58 |
| Control Test | | 0.079 | 08:58 |
| Air Blank | | 0.000 | 08:59 |
| Diagnostics Check | | OK | 08:59 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALACIOS-FLORES,CRISTIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/16/2026 11:55:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 11:49 |
| Air Blank | | 0.000 | 11:49 |
| Control Test | | 0.080 | 11:50 |
| Air Blank | | 0.000 | 11:50 |
| Subject Sample | # 1 | 0.000 | 11:51 |
| Breath Volume | | 1.707 | 11:51 |
| Air Blank | | 0.000 | 11:51 |
| Air Blank | | 0.000 | 11:53 |
| Subject Sample | # 2 | 0.000 | 11:53 |
| Breath Volume | | 1.812 | 11:53 |
| Air Blank | | 0.000 | 11:54 |
| Control Test | | 0.079 | 11:54 |
| Air Blank | | 0.000 | 11:55 |
| Diagnostics Check | | OK | 11:55 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: FILLMORE,DONALD
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/18/2026 00:53:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:46 |
| Air Blank | | 0.000 | 00:47 |
| Control Test | | 0.081 | 00:47 |
| Air Blank | | 0.000 | 00:48 |
| Subject Sample | # 1 | 0.070 | 00:48 |
| Breath Volume | | 1.488 | 00:48 |
| Air Blank | | 0.000 | 00:49 |
| Air Blank | | 0.000 | 00:51 |
| Subject Sample | # 2 | 0.068 | 00:51 |
| Breath Volume | | 1.625 | 00:51 |
| Air Blank | | 0.000 | 00:52 |
| Control Test | | 0.081 | 00:52 |
| Air Blank | | 0.000 | 00:53 |
| Diagnostics Check | | OK | 00:53 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: ALVELO RODRIGUEZ

DATE/TIME OF TEST: 05/18/2026 10:27:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 10:21 |
| Air Blank | | 0.000 | 10:21 |
| Control Test | | 0.081 | 10:22 |
| Air Blank | | 0.000 | 10:22 |
| Subject Sample | # 1 | 0.000 | 10:23 |
| Breath Volume | | 3.046 | 10:23 |
| Air Blank | | 0.000 | 10:23 |
| Air Blank | | 0.000 | 10:25 |
| Subject Sample | # 2 | 0.000 | 10:25 |
| Breath Volume | | 2.980 | 10:25 |
| Air Blank | | 0.000 | 10:26 |
| Control Test | | 0.081 | 10:26 |
| Air Blank | | 0.000 | 10:27 |
| Diagnostics Check | | OK | 10:27 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: DEFORD

DATE/TIME OF TEST: 05/18/2026 16:26:15

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:20 |
| Air Blank | | 0.000 | 16:20 |
| Control Test | | 0.081 | 16:20 |
| Air Blank | | 0.000 | 16:21 |
| Subject Sample | # 1 | 0.000 | 16:22 |
| Breath Volume | | 3.019 | 16:22 |
| Air Blank | | 0.000 | 16:22 |
| Air Blank | | 0.000 | 16:24 |
| Subject Sample | # 2 | 0.000 | 16:24 |
| Breath Volume | | 2.726 | 16:24 |
| Air Blank | | 0.000 | 16:25 |
| Control Test | | 0.080 | 16:25 |
| Air Blank | | 0.000 | 16:26 |
| Diagnostics Check | | OK | 16:26 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/18/2026 16:40:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:34 |
| Air Blank | | 0.000 | 16:34 |
| Control Test | | 0.081 | 16:34 |
| Air Blank | | 0.000 | 16:35 |
| Subject Sample | # 1 | 0.000 | 16:36 |
| Breath Volume | | 1.554 | 16:36 |
| Air Blank | | 0.000 | 16:36 |
| Air Blank | | 0.000 | 16:38 |
| Subject Sample | # 2 | 0.000 | 16:39 |
| Breath Volume | | 1.703 | 16:39 |
| Air Blank | | 0.000 | 16:39 |
| Control Test | | 0.080 | 16:39 |
| Air Blank | | 0.000 | 16:40 |
| Diagnostics Check | | OK | 16:40 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,N,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,R
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: EARNEST

DATE/TIME OF TEST: 05/19/2026 21:48:43

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 21:42 | |
| Air Blank | | 0.000 | 21:43 | |
| Control Test | | 0.081 | 21:43 | |
| Air Blank | | 0.000 | 21:44 | |
| Subject Sample | # 1 | 0.085 | 21:45 | |
| Breath Volume | | 1.394 | 21:45 | |
| Air Blank | | 0.000 | 21:45 | |
| Air Blank | | | 21:47 | RFI Detect |
| Air Blank | | 0.000 | 21:47 | |
| Control Test | | 0.081 | 21:48 | |
| Air Blank | | 0.000 | 21:48 | |
| Diagnostics Check | | OK | 21:48 | |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,N,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,R
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: EARNEST

DATE/TIME OF TEST: 05/19/2026 21:58:56

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 21:52 |
| Air Blank | | 0.000 | 21:52 |
| Control Test | | 0.080 | 21:52 |
| Air Blank | | 0.000 | 21:53 |
| Subject Sample | # 1 | 0.090 | 21:54 |
| Breath Volume | | 1.750 | 21:54 |
| Air Blank | | 0.000 | 21:54 |
| Air Blank | | 0.000 | 21:56 |
| Subject Sample | # 2 | 0.084 | 21:57 |
| Breath Volume | | 1.726 | 21:57 |
| Air Blank | | 0.000 | 21:58 |
| Control Test | | 0.079 | 21:58 |
| Air Blank | | 0.000 | 21:58 |
| Diagnostics Check | | OK | 21:58 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,N,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HODGE,K
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 05/20/2026 01:08:37

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:01 |
| Air Blank | | 0.000 | 01:02 |
| Control Test | | 0.080 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Subject Sample | # 1 | 0.181 | 01:03 |
| Breath Volume | | 1.394 | 01:03 |
| Air Blank | | 0.000 | 01:04 |
| Air Blank | | 0.000 | 01:06 |
| Subject Sample | # 2 | 0.191 | 01:07 |
| Breath Volume | | 2.066 | 01:07 |
| Air Blank | | 0.000 | 01:07 |
| Control Test | | 0.080 | 01:08 |
| Air Blank | | 0.000 | 01:08 |
| Diagnostics Check | | OK | 01:08 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GLASCO,GREGORY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 05/22/2026 18:53:04

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 18:47 | |
| Air Blank | | 0.000 | 18:47 | |
| Control Test | | 0.081 | 18:48 | |
| Air Blank | | 0.000 | 18:48 | |
| Subject Sample | # 1 | 0.000 | 18:49 | |
| Breath Volume | | 1.765 | 18:49 | |
| Air Blank | | 0.000 | 18:49 | |
| Air Blank | | 0.000 | 18:51 | |
| Subject Sample | # 2 | | 18:51 | RFI Detect |
| Breath Volume | | 0.000 | 18:51 | |
| Air Blank | | 0.000 | 18:52 | |
| Control Test | | 0.080 | 18:52 | |
| Air Blank | | 0.000 | 18:52 | |
| Diagnostics Check | | OK | 18:53 | |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GLASCO,GREGORY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 05/22/2026 19:01:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:55 |
| Air Blank | | 0.000 | 18:55 |
| Control Test | | 0.080 | 18:56 |
| Air Blank | | 0.000 | 18:56 |
| Subject Sample | # 1 | 0.000 | 18:57 |
| Breath Volume | | 1.593 | 18:57 |
| Air Blank | | 0.000 | 18:57 |
| Air Blank | | 0.000 | 18:59 |
| Subject Sample | # 2 | 0.000 | 18:59 |
| Breath Volume | | 1.667 | 18:59 |
| Air Blank | | 0.000 | 19:00 |
| Control Test | | 0.081 | 19:00 |
| Air Blank | | 0.000 | 19:01 |
| Diagnostics Check | | OK | 19:01 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: KNAPP,JAMIE
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CRISTOBAL-ROJA

DATE/TIME OF TEST: 05/23/2026 03:42:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:36 |
| Air Blank | | 0.000 | 03:36 |
| Control Test | | 0.081 | 03:36 |
| Air Blank | | 0.000 | 03:37 |
| Subject Sample | # 1 | 0.112 | 03:37 |
| Breath Volume | | 2.062 | 03:37 |
| Air Blank | | 0.000 | 03:38 |
| Air Blank | | 0.000 | 03:40 |
| Subject Sample | # 2 | 0.116 | 03:40 |
| Breath Volume | | 2.222 | 03:40 |
| Air Blank | | 0.000 | 03:41 |
| Control Test | | 0.080 | 03:41 |
| Air Blank | | 0.000 | 03:41 |
| Diagnostics Check | | OK | 03:42 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,SUSAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 10:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: DIMURO

DATE/TIME OF TEST: 05/23/2026 10:31:41

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 10:25 |
| Air Blank | | 0.000 | 10:26 |
| Control Test | | 0.080 | 10:26 |
| Air Blank | | 0.000 | 10:27 |
| Subject Sample | # 1 | 0.000 | 10:27 |
| Breath Volume | | 2.445 | 10:27 |
| Air Blank | | 0.000 | 10:27 |
| Air Blank | | 0.000 | 10:29 |
| Subject Sample | # 2 | 0.000 | 10:30 |
| Breath Volume | | 2.386 | 10:30 |
| Air Blank | | 0.000 | 10:30 |
| Control Test | | 0.081 | 10:31 |
| Air Blank | | 0.000 | 10:31 |
| Diagnostics Check | | OK | 10:31 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,SUSAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: PUJOL

DATE/TIME OF TEST: 05/23/2026 17:07:31

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 17:01 |
| Air Blank | | 0.000 | 17:02 |
| Control Test | | 0.081 | 17:02 |
| Air Blank | | 0.000 | 17:02 |
| Subject Sample | # 1 | 0.000 | 17:03 |
| Breath Volume | | 1.503 | 17:03 |
| Air Blank | | 0.000 | 17:03 |
| Air Blank | | 0.000 | 17:05 |
| Subject Sample | # 2 | 0.000 | 17:06 |
| Breath Volume | | 1.347 | 17:06 |
| Air Blank | | 0.000 | 17:06 |
| Control Test | | 0.080 | 17:06 |
| Air Blank | | 0.000 | 17:07 |
| Diagnostics Check | | OK | 17:07 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,RODNEY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: DELP

DATE/TIME OF TEST: 05/23/2026 23:12:38

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:06 |
| Air Blank | | 0.000 | 23:06 |
| Control Test | | 0.081 | 23:06 |
| Air Blank | | 0.000 | 23:07 |
| Subject Sample | # 1 | 0.162 | 23:07 |
| Breath Volume | | 2.914 | 23:07 |
| Air Blank | | 0.000 | 23:08 |
| Air Blank | | 0.000 | 23:10 |
| Subject Sample | # 2 | 0.157 | 23:10 |
| Breath Volume | | 2.972 | 23:10 |
| Air Blank | | 0.000 | 23:11 |
| Control Test | | 0.079 | 23:12 |
| Air Blank | | 0.000 | 23:12 |
| Diagnostics Check | | OK | 23:12 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: KNAPP,JAMIE
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: ROONEY

DATE/TIME OF TEST: 05/24/2026 00:15:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:08 |
| Air Blank | | 0.000 | 00:09 |
| Control Test | | 0.080 | 00:09 |
| Air Blank | | 0.000 | 00:10 |
| Subject Sample | # 1 | 0.165 | 00:10 |
| Breath Volume | | 1.933 | 00:10 |
| Air Blank | | 0.000 | 00:11 |
| Air Blank | | 0.000 | 00:12 |
| Subject Sample | # 2 | 0.155 | 00:13 |
| Breath Volume | | 1.773 | 00:13 |
| Air Blank | | 0.000 | 00:14 |
| Control Test | | 0.078 | 00:14 |
| Air Blank | | 0.000 | 00:15 |
| Diagnostics Check | | OK | 00:15 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ZEPADA

DATE/TIME OF TEST: 05/24/2026 01:15:08

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:08 |
| Air Blank | | 0.000 | 01:09 |
| Control Test | | 0.081 | 01:09 |
| Air Blank | | 0.000 | 01:10 |
| Subject Sample | # 1 | 0.083 | 01:10 |
| Breath Volume | | 1.675 | 01:10 |
| Air Blank | | 0.000 | 01:11 |
| Air Blank | | 0.000 | 01:13 |
| Subject Sample | # 2 | 0.080 | 01:13 |
| Breath Volume | | 2.441 | 01:13 |
| Air Blank | | 0.000 | 01:14 |
| Control Test | | 0.079 | 01:14 |
| Air Blank | | 0.000 | 01:15 |
| Diagnostics Check | | OK | 01:15 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,RODNEY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FERNANDO

DATE/TIME OF TEST: 05/24/2026 02:03:59

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 02:00 | |
| Air Blank | | 0.000 | 02:01 | |
| Control Test | | 0.080 | 02:01 | |
| Air Blank | | 0.000 | 02:01 | |
| Subject Sample | # 1 | | 02:02 | Subject Test Refused |
| Breath Volume | | 0.000 | 02:02 | |
| Air Blank | | 0.000 | 02:03 | |
| Control Test | | 0.080 | 02:03 | |
| Air Blank | | 0.000 | 02:03 | |
| Diagnostics Check | | OK | 02:03 | |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 05/25/2026 01:16:11

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:13 | |
| Air Blank | | 0.000 | 01:13 | |
| Control Test | | 0.081 | 01:14 | |
| Air Blank | | 0.000 | 01:14 | |
| Subject Sample | # 1 | | 01:14 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:14 | |
| Air Blank | | 0.000 | 01:15 | |
| Control Test | | 0.081 | 01:15 | |
| Air Blank | | 0.000 | 01:16 | |
| Diagnostics Check | | OK | 01:16 | |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALACIOS-FLORES,CRISTIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: TATTERSALL

DATE/TIME OF TEST: 05/26/2026 14:43:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 14:37 |
| Air Blank | | 0.000 | 14:37 |
| Control Test | | 0.080 | 14:38 |
| Air Blank | | 0.000 | 14:38 |
| Subject Sample | # 1 | 0.000 | 14:39 |
| Breath Volume | | 1.492 | 14:39 |
| Air Blank | | 0.000 | 14:39 |
| Air Blank | | 0.000 | 14:41 |
| Subject Sample | # 2 | 0.000 | 14:41 |
| Breath Volume | | 1.113 | 14:41 |
| Air Blank | | 0.000 | 14:42 |
| Control Test | | 0.079 | 14:42 |
| Air Blank | | 0.000 | 14:43 |
| Diagnostics Check | | OK | 14:43 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALACIOS-FLORES,CRISTIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: DEFORD

DATE/TIME OF TEST: 05/26/2026 15:36:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:30 |
| Air Blank | | 0.000 | 15:30 |
| Control Test | | 0.080 | 15:31 |
| Air Blank | | 0.000 | 15:31 |
| Subject Sample | # 1 | 0.000 | 15:32 |
| Breath Volume | | 2.375 | 15:32 |
| Air Blank | | 0.000 | 15:32 |
| Air Blank | | 0.000 | 15:34 |
| Subject Sample | # 2 | 0.000 | 15:34 |
| Breath Volume | | 2.437 | 15:34 |
| Air Blank | | 0.000 | 15:35 |
| Control Test | | 0.080 | 15:35 |
| Air Blank | | 0.000 | 15:36 |
| Diagnostics Check | | OK | 15:36 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,N,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: TONEY,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: HICKEY

DATE/TIME OF TEST: 05/27/2026 01:41:39

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 01:34 |
| Air Blank | | 0.000 | 01:35 |
| Control Test | | 0.081 | 01:35 |
| Air Blank | | 0.000 | 01:35 |
| Subject Sample | # 1 | 0.300 | 01:36 |
| Breath Volume | | 1.195 | 01:36 |
| Air Blank | | 0.000 | 01:37 |
| Air Blank | | 0.000 | 01:39 |
| Subject Sample | # 2 | 0.284 | 01:39 |
| Breath Volume | | 1.136 | 01:39 |
| Air Blank | | 0.000 | 01:40 |
| Control Test | | 0.078 | 01:41 |
| Air Blank | | 0.000 | 01:41 |
| Diagnostics Check | | OK | 01:41 |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: KNAPP,Z
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:57

SUBJECT'S LAST NAME: FABIAN TOMAS

DATE/TIME OF TEST: 05/27/2026 13:20:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 13:17 | |
| Air Blank | | 0.000 | 13:18 | |
| Control Test | | 0.080 | 13:18 | |
| Air Blank | | 0.000 | 13:19 | |
| Subject Sample | # 1 | | 13:19 | Subject Test Refused |
| Breath Volume | | 0.000 | 13:19 | |
| Air Blank | | 0.000 | 13:19 | |
| Control Test | | 0.080 | 13:20 | |
| Air Blank | | 0.000 | 13:20 | |
| Diagnostics Check | | OK | 13:20 | |

Data Download Date/Time: 06/07/2026 20:30

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,RODNEY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: HERNANDEZ LARA

DATE/TIME OF TEST: 05/28/2026 03:15:21

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 03:12 | |
| Air Blank | | 0.000 | 03:12 | |
| Control Test | | 0.080 | 03:13 | |
| Air Blank | | 0.000 | 03:13 | |
| Subject Sample | # 1 | | 03:13 | Subject Test Refused |
| Breath Volume | | 0.000 | 03:13 | |
| Air Blank | | 0.000 | 03:14 | |
| Control Test | | 0.080 | 03:14 | |
| Air Blank | | 0.000 | 03:15 | |
| Diagnostics Check | | OK | 03:15 | |

Data Download Date/Time: 06/07/2026 20:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTEAD,CAMERON,P
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: EARLE,RODNEY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/29/2026 04:35:52

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:32 | |
| Air Blank | | 0.000 | 04:33 | |
| Control Test | | 0.081 | 04:33 | |
| Air Blank | | 0.000 | 04:34 | |
| Subject Sample | # 1 | | 04:34 | Subject Test Refused |
| Breath Volume | | 0.000 | 04:34 | |
| Air Blank | | 0.000 | 04:34 | |
| Control Test | | 0.080 | 04:35 | |
| Air Blank | | 0.000 | 04:35 | |
| Diagnostics Check | | OK | 04:35 | |

Data Download Date/Time: 06/07/2026 20:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGE,WILLIAM,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HODGE,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/29/2026 12:31:32

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 12:25 |
| Air Blank | | 0.000 | 12:26 |
| Control Test | | 0.080 | 12:26 |
| Air Blank | | 0.000 | 12:26 |
| Subject Sample | # 1 | 0.000 | 12:27 |
| Breath Volume | | 2.531 | 12:27 |
| Air Blank | | 0.000 | 12:27 |
| Air Blank | | 0.000 | 12:29 |
| Subject Sample | # 2 | 0.000 | 12:30 |
| Breath Volume | | 2.085 | 12:30 |
| Air Blank | | 0.000 | 12:30 |
| Control Test | | 0.079 | 12:30 |
| Air Blank | | 0.000 | 12:31 |
| Diagnostics Check | | OK | 12:31 |

Data Download Date/Time: 06/07/2026 20:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 05/29/2026 13:17:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 13:10 |
| Air Blank | | 0.000 | 13:11 |
| Control Test | | 0.080 | 13:11 |
| Air Blank | | 0.000 | 13:12 |
| Subject Sample | # 1 | 0.000 | 13:12 |
| Breath Volume | | 2.132 | 13:12 |
| Air Blank | | 0.000 | 13:13 |
| Air Blank | | 0.000 | 13:15 |
| Subject Sample | # 2 | 0.000 | 13:15 |
| Breath Volume | | 2.195 | 13:15 |
| Air Blank | | 0.000 | 13:16 |
| Control Test | | 0.080 | 13:16 |
| Air Blank | | 0.000 | 13:17 |
| Diagnostics Check | | OK | 13:17 |

Data Download Date/Time: 06/07/2026 20:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: COTTO,GABRIEL
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/30/2026 03:01:10

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:54 |
| Air Blank | | 0.000 | 02:55 |
| Control Test | | 0.081 | 02:55 |
| Air Blank | | 0.000 | 02:55 |
| Subject Sample | # 1 | 0.173 | 02:56 |
| Breath Volume | | 1.222 | 02:56 |
| Air Blank | | 0.000 | 02:57 |
| Air Blank | | 0.000 | 02:58 |
| Subject Sample | # 2 | 0.169 | 02:59 |
| Breath Volume | | 1.398 | 02:59 |
| Air Blank | | 0.000 | 03:00 |
| Control Test | | 0.080 | 03:00 |
| Air Blank | | 0.000 | 03:01 |
| Diagnostics Check | | OK | 03:01 |

Data Download Date/Time: 06/07/2026 20:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGE,WILLIAM,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HODGE,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 05/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: PUJOL

DATE/TIME OF TEST: 05/30/2026 12:15:55

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 12:09 |
| Air Blank | | 0.000 | 12:10 |
| Control Test | | 0.080 | 12:10 |
| Air Blank | | 0.000 | 12:11 |
| Subject Sample | # 1 | 0.000 | 12:11 |
| Breath Volume | | 1.261 | 12:11 |
| Air Blank | | 0.000 | 12:12 |
| Air Blank | | 0.000 | 12:14 |
| Subject Sample | # 2 | 0.000 | 12:14 |
| Breath Volume | | 1.367 | 12:14 |
| Air Blank | | 0.000 | 12:15 |
| Control Test | | 0.080 | 12:15 |
| Air Blank | | 0.000 | 12:15 |
| Diagnostics Check | | OK | 12:15 |

Data Download Date/Time: 06/07/2026 20:31

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JULIAN,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: BIRGE,ZACH
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 05/01/2026 00:53:48

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:50 | |
| Air Blank | | 0.000 | 00:51 | |
| Control Test | | 0.079 | 00:51 | |
| Air Blank | | 0.000 | 00:52 | |
| Subject Sample | # 1 | | 00:52 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:52 | |
| Air Blank | | 0.000 | 00:52 | |
| Control Test | | 0.079 | 00:53 | |
| Air Blank | | 0.000 | 00:53 | |
| Diagnostics Check | | OK | 00:53 | |

Data Download Date/Time: 05/09/2026 12:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: KOWALSKE,MATTHEW
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VILLA

DATE/TIME OF TEST: 05/03/2026 01:14:59

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:12 | |
| Air Blank | | 0.000 | 01:12 | |
| Control Test | | 0.079 | 01:12 | |
| Air Blank | | 0.000 | 01:13 | |
| Subject Sample | # 1 | | 01:13 | Subject Test Refused |
| Breath Volume | | 0.000 | 01:13 | |
| Air Blank | | 0.000 | 01:14 | |
| Control Test | | 0.079 | 01:14 | |
| Air Blank | | 0.000 | 01:14 | |
| Diagnostics Check | | OK | 01:14 | |

Data Download Date/Time: 05/09/2026 12:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SMITH,SAUL
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: PEREZ-CRUZ

DATE/TIME OF TEST: 05/03/2026 02:26:43

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:17 |
| Air Blank | | 0.000 | 02:17 |
| Control Test | | 0.078 | 02:18 |
| Air Blank | | 0.000 | 02:18 |
| Subject Sample | # 1 | 0.230 | 02:19 |
| Breath Volume | | 1.894 | 02:19 |
| Air Blank | | 0.000 | 02:20 |
| Air Blank | | 0.000 | 02:22 |
| Subject Sample | # 2 | 0.240 | 02:25 |
| Breath Volume | | 2.062 | 02:25 |
| Air Blank | | 0.000 | 02:25 |
| Control Test | | 0.077 | 02:26 |
| Air Blank | | 0.000 | 02:26 |
| Diagnostics Check | | OK | 02:26 |

Data Download Date/Time: 05/09/2026 12:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: KHANG,ADAM
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HERNANDEZ-MEXICANO

DATE/TIME OF TEST: 05/03/2026 04:15:01

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:06 |
| Air Blank | | 0.000 | 04:07 |
| Control Test | | 0.078 | 04:07 |
| Air Blank | | 0.000 | 04:08 |
| Subject Sample | # 1 | 0.139 | 04:10 |
| Breath Volume | | 1.617 | 04:10 |
| Air Blank | | 0.000 | 04:11 |
| Air Blank | | 0.000 | 04:12 |
| Subject Sample | # 2 | 0.137 | 04:13 |
| Breath Volume | | 1.710 | 04:13 |
| Air Blank | | 0.000 | 04:14 |
| Control Test | | 0.077 | 04:14 |
| Air Blank | | 0.000 | 04:14 |
| Diagnostics Check | | OK | 04:15 |

Data Download Date/Time: 05/09/2026 12:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: TAYLOR,JORDAN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: SANCHEZ DOMINGO

DATE/TIME OF TEST: 05/03/2026 20:36:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:29 |
| Air Blank | | 0.000 | 20:30 |
| Control Test | | 0.080 | 20:30 |
| Air Blank | | 0.000 | 20:31 |
| Subject Sample | # 1 | 0.150 | 20:31 |
| Breath Volume | | 1.617 | 20:31 |
| Air Blank | | 0.000 | 20:32 |
| Air Blank | | 0.000 | 20:34 |
| Subject Sample | # 2 | 0.152 | 20:34 |
| Breath Volume | | 2.343 | 20:34 |
| Air Blank | | 0.000 | 20:35 |
| Control Test | | 0.079 | 20:35 |
| Air Blank | | 0.000 | 20:35 |
| Diagnostics Check | | OK | 20:36 |

Data Download Date/Time: 05/09/2026 12:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTRERAS,SANTIAGO,
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: JIMENEZ,ANDREW
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: CORRY

DATE/TIME OF TEST: 05/08/2026 15:23:44

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 15:20 | |
| Air Blank | | 0.000 | 15:21 | |
| Control Test | | 0.079 | 15:21 | |
| Air Blank | | 0.000 | 15:22 | |
| Subject Sample | # 1 | | 15:22 | Subject Test Refused |
| Breath Volume | | 0.000 | 15:22 | |
| Air Blank | | 0.000 | 15:22 | |
| Control Test | | 0.078 | 15:23 | |
| Air Blank | | 0.000 | 15:23 | |
| Diagnostics Check | | OK | 15:23 | |

Data Download Date/Time: 05/09/2026 12:13


COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JULIAN,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: RAK,JUSTIN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 04/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/08/2026 20:01:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:55 |
| Air Blank | | 0.000 | 19:56 |
| Control Test | | 0.079 | 19:56 |
| Air Blank | | 0.000 | 19:57 |
| Subject Sample | # 1 | 0.167 | 19:57 |
| Breath Volume | | 3.179 | 19:57 |
| Air Blank | | 0.000 | 19:58 |
| Air Blank | | 0.000 | 19:59 |
| Subject Sample | # 2 | 0.157 | 20:00 |
| Breath Volume | | 2.382 | 20:00 |
| Air Blank | | 0.000 | 20:01 |
| Control Test | | 0.079 | 20:01 |
| Air Blank | | 0.000 | 20:01 |
| Diagnostics Check | | OK | 20:01 |

Data Download Date/Time: 05/09/2026 12:13

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: TAYLOR,JORDAN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RODRIGUEZ-COLORADO

DATE/TIME OF TEST: 05/16/2026 00:02:29

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 23:59 | |
| Air Blank | | 0.000 | 00:00 | |
| Control Test | | 0.078 | 00:00 | |
| Air Blank | | 0.000 | 00:00 | |
| Subject Sample | # 1 | | 00:01 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:01 | |
| Air Blank | | 0.000 | 00:01 | |
| Control Test | | 0.078 | 00:01 | |
| Air Blank | | 0.000 | 00:02 | |
| Diagnostics Check | | OK | 00:02 | |

Data Download Date/Time: 06/06/2026 14:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: TAYLOR,JORDAN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RODRIGUEZ-COLORADO

DATE/TIME OF TEST: 05/16/2026 00:27:42

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:21 |
| Air Blank | | 0.000 | 00:21 |
| Control Test | | 0.078 | 00:21 |
| Air Blank | | 0.000 | 00:22 |
| Subject Sample | # 1 | 0.195 | 00:22 |
| Breath Volume | | 2.089 | 00:22 |
| Air Blank | | 0.000 | 00:23 |
| Air Blank | | 0.000 | 00:25 |
| Subject Sample | # 2 | 0.197 | 00:26 |
| Breath Volume | | 2.285 | 00:26 |
| Air Blank | | 0.000 | 00:26 |
| Control Test | | 0.078 | 00:27 |
| Air Blank | | 0.000 | 00:27 |
| Diagnostics Check | | OK | 00:27 |

Data Download Date/Time: 06/06/2026 14:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: COCANOUGHER,CALEB
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: COCANOUGHER

DATE/TIME OF TEST: 05/16/2026 02:11:39

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 02:05 |
| Air Blank | | 0.000 | 02:05 |
| Control Test | | 0.080 | 02:06 |
| Air Blank | | 0.000 | 02:06 |
| Subject Sample | # 1 | 0.072 | 02:07 |
| Breath Volume | | 1.488 | 02:07 |
| Air Blank | | 0.000 | 02:07 |
| Air Blank | | 0.000 | 02:09 |
| Subject Sample | # 2 | 0.072 | 02:10 |
| Breath Volume | | 1.933 | 02:10 |
| Air Blank | | 0.000 | 02:10 |
| Control Test | | 0.079 | 02:11 |
| Air Blank | | 0.000 | 02:11 |
| Diagnostics Check | | OK | 02:11 |

Data Download Date/Time: 06/06/2026 14:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: LORENZO,GABRIEL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: HEGGAN

DATE/TIME OF TEST: 05/16/2026 04:48:23

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:41 |
| Air Blank | | 0.000 | 04:41 |
| Control Test | | 0.079 | 04:41 |
| Air Blank | | 0.000 | 04:42 |
| Subject Sample | # 1 | 0.088 | 04:43 |
| Breath Volume | | 1.652 | 04:43 |
| Air Blank | | 0.000 | 04:44 |
| Air Blank | | 0.000 | 04:46 |
| Subject Sample | # 2 | 0.086 | 04:46 |
| Breath Volume | | 1.449 | 04:46 |
| Air Blank | | 0.000 | 04:47 |
| Control Test | | 0.079 | 04:47 |
| Air Blank | | 0.000 | 04:48 |
| Diagnostics Check | | OK | 04:48 |

Data Download Date/Time: 06/06/2026 14:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBYNE,CRAIG,A
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SIMS,TERRY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:52

SUBJECT'S LAST NAME: HARKEY

DATE/TIME OF TEST: 05/16/2026 08:26:39

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 08:17 |
| Air Blank | | 0.000 | 08:18 |
| Control Test | | 0.077 | 08:18 |
| Air Blank | | 0.000 | 08:19 |
| Subject Sample | # 1 | 0.104 | 08:20 |
| Breath Volume | | 1.570 | 08:20 |
| Air Blank | | 0.000 | 08:21 |
| Air Blank | | 0.000 | 08:22 |
| Subject Sample | # 2 | 0.115 | 08:24 |
| Breath Volume | | 2.164 | 08:24 |
| Air Blank | | 0.000 | 08:25 |
| Control Test | | 0.075 | 08:26 |
| Air Blank | | 0.000 | 08:26 |
| Diagnostics Check | | OK | 08:26 |

Data Download Date/Time: 06/06/2026 14:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,ROBERTO,R
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: TAYLOR,JORDAN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MERCADO-MEZA

DATE/TIME OF TEST: 05/17/2026 18:54:57

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 18:48 |
| Air Blank | | 0.000 | 18:48 |
| Control Test | | 0.078 | 18:49 |
| Air Blank | | 0.000 | 18:49 |
| Subject Sample | # 1 | 0.152 | 18:50 |
| Breath Volume | | 3.652 | 18:50 |
| Air Blank | | 0.000 | 18:51 |
| Air Blank | | 0.000 | 18:52 |
| Subject Sample | # 2 | 0.161 | 18:53 |
| Breath Volume | | 3.699 | 18:53 |
| Air Blank | | 0.000 | 18:54 |
| Control Test | | 0.078 | 18:54 |
| Air Blank | | 0.000 | 18:54 |
| Diagnostics Check | | OK | 18:54 |

Data Download Date/Time: 06/06/2026 14:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONN,SHARON,K
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: HAGGARD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 05/18/2026 14:56:35

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 14:51 | |
| Air Blank | | 0.000 | 14:51 | |
| Control Test | | 0.078 | 14:51 | |
| Air Blank | | 0.000 | 14:52 | |
| Subject Sample | # 1 | 0.000 | 14:52 | |
| Breath Volume | | 2.609 | 14:52 | |
| Air Blank | | 0.000 | 14:53 | |
| Air Blank | | 0.000 | 14:55 | |
| Subject Sample | # 2 | | 14:55 | RFI Detect |
| Breath Volume | | 1.957 | 14:55 | |
| Air Blank | | 0.000 | 14:56 | |
| | | | 14:56 | RFI Detect |
| Air Blank | | 0.000 | 14:56 | |

Data Download Date/Time: 06/06/2026 14:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONN,SHARON,K
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: HAGGARD,RICKY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22521080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 05/18/2026 15:06:48

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 15:01 |
| Air Blank | | 0.000 | 15:01 |
| Control Test | | 0.078 | 15:01 |
| Air Blank | | 0.000 | 15:02 |
| Subject Sample | # 1 | 0.000 | 15:02 |
| Breath Volume | | 3.023 | 15:02 |
| Air Blank | | 0.000 | 15:03 |
| Air Blank | | 0.000 | 15:05 |
| Subject Sample | # 2 | 0.000 | 15:05 |
| Breath Volume | | 3.351 | 15:05 |
| Air Blank | | 0.000 | 15:05 |
| Control Test | | 0.077 | 15:06 |
| Air Blank | | 0.000 | 15:06 |
| Diagnostics Check | | OK | 15:06 |

Data Download Date/Time: 06/06/2026 14:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: RIOS,DANIELLE
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 05/30/2026 00:46:46

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:43 | |
| Air Blank | | 0.000 | 00:44 | |
| Control Test | | 0.079 | 00:44 | |
| Air Blank | | 0.000 | 00:45 | |
| Subject Sample | # 1 | | 00:45 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:45 | |
| Air Blank | | 0.000 | 00:45 | |
| Control Test | | 0.079 | 00:46 | |
| Air Blank | | 0.000 | 00:46 | |
| Diagnostics Check | | OK | 00:46 | |

Data Download Date/Time: 06/06/2026 14:15

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007509

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: TAYLOR,JORDAN
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 05/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: ZUNIGA-PINTOR

DATE/TIME OF TEST: 05/30/2026 23:20:21

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:13 |
| Air Blank | | 0.000 | 23:14 |
| Control Test | | 0.078 | 23:14 |
| Air Blank | | 0.000 | 23:15 |
| Subject Sample | # 1 | 0.252 | 23:15 |
| Breath Volume | | 2.218 | 23:15 |
| Air Blank | | 0.000 | 23:16 |
| Air Blank | | 0.000 | 23:18 |
| Subject Sample | # 2 | 0.252 | 23:18 |
| Breath Volume | | 2.468 | 23:18 |
| Air Blank | | 0.000 | 23:19 |
| Control Test | | 0.078 | 23:19 |
| Air Blank | | 0.000 | 23:20 |
| Diagnostics Check | | OK | 23:20 |

Data Download Date/Time: 06/06/2026 14:16

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CUMMINGS,W
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/04/2026 20:32:14

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 20:25 |
| Air Blank | | 0.000 | 20:26 |
| Control Test | | 0.080 | 20:26 |
| Air Blank | | 0.000 | 20:27 |
| Subject Sample | # 1 | 0.093 | 20:27 |
| Breath Volume | | 2.441 | 20:27 |
| Air Blank | | 0.000 | 20:28 |
| Air Blank | | 0.000 | 20:30 |
| Subject Sample | # 2 | 0.092 | 20:30 |
| Breath Volume | | 1.738 | 20:30 |
| Air Blank | | 0.000 | 20:31 |
| Control Test | | 0.079 | 20:31 |
| Air Blank | | 0.000 | 20:32 |
| Diagnostics Check | | OK | 20:32 |

Data Download Date/Time: 05/29/2026 20:39

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: OPALINSKI

DATE/TIME OF TEST: 05/09/2026 04:07:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 04:01 |
| Air Blank | | 0.000 | 04:01 |
| Control Test | | 0.081 | 04:01 |
| Air Blank | | 0.000 | 04:02 |
| Subject Sample | # 1 | 0.186 | 04:02 |
| Breath Volume | | 3.843 | 04:02 |
| Air Blank | | 0.000 | 04:03 |
| Air Blank | | 0.000 | 04:05 |
| Subject Sample | # 2 | 0.187 | 04:05 |
| Breath Volume | | 2.726 | 04:05 |
| Air Blank | | 0.000 | 04:06 |
| Control Test | | 0.080 | 04:06 |
| Air Blank | | 0.000 | 04:07 |
| Diagnostics Check | | OK | 04:07 |

Data Download Date/Time: 05/29/2026 20:39

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUI,DANIEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: KEMP,CALEB
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 04/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 05/16/2026 01:25:47

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 01:12 | |
| Air Blank | | 0.000 | 01:13 | |
| Control Test | | 0.080 | 01:13 | |
| Air Blank | | 0.000 | 01:13 | |
| Subject Sample | # 1 | 0.163 | 01:17 | Slope Not Level |
| Breath Volume | | 1.203 | 01:17 | |
| Air Blank | | 0.000 | 01:18 | |
| Air Blank | | 0.000 | 01:19 | |
| Subject Sample | # 2 | 0.189 | 01:20 | |
| Breath Volume | | 2.000 | 01:20 | |
| Air Blank | | 0.000 | 01:21 | |
| Air Blank | | 0.000 | 01:23 | |
| Subject Sample | # 3 | 0.179 | 01:24 | |
| Breath Volume | | 1.687 | 01:24 | |
| Air Blank | | 0.000 | 01:24 | |
| Control Test | | 0.078 | 01:25 | |
| Air Blank | | 0.000 | 01:25 | |
| Diagnostics Check | | OK | 01:25 | |

Data Download Date/Time: 05/29/2026 20:39

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYANT,RUBEN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRYANT,RUBEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BELZ

DATE/TIME OF TEST: 05/30/2026 04:07:41

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 04:03 | |
| Air Blank | | 0.000 | 04:04 | |
| Control Test | | 0.081 | 04:04 | |
| Air Blank | | 0.000 | 04:05 | |
| Subject Sample | # 1 | | 04:06 | RFI Detect |
| Breath Volume | | 0.000 | 04:06 | |
| Air Blank | | 0.000 | 04:06 | |
| Control Test | | 0.080 | 04:07 | |
| Air Blank | | 0.000 | 04:07 | |
| Diagnostics Check | | OK | 04:07 | |

Data Download Date/Time: 06/25/2026 01:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,MYREON,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PIERCE,MYREON
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: NORDLOF

DATE/TIME OF TEST: 05/31/2026 00:13:45

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:07 |
| Air Blank | | 0.000 | 00:07 |
| Control Test | | 0.080 | 00:07 |
| Air Blank | | 0.000 | 00:08 |
| Subject Sample | # 1 | 0.147 | 00:09 |
| Breath Volume | | 4.058 | 00:09 |
| Air Blank | | 0.000 | 00:09 |
| Air Blank | | 0.000 | 00:11 |
| Subject Sample | # 2 | 0.144 | 00:12 |
| Breath Volume | | 4.011 | 00:12 |
| Air Blank | | 0.000 | 00:12 |
| Control Test | | 0.078 | 00:13 |
| Air Blank | | 0.000 | 00:13 |
| Diagnostics Check | | OK | 00:13 |

Data Download Date/Time: 06/25/2026 01:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,MYREON,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RUCCI,MATTHEW
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: BASS III

DATE/TIME OF TEST: 05/31/2026 01:04:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 00:57 |
| Air Blank | | 0.000 | 00:57 |
| Control Test | | 0.080 | 00:58 |
| Air Blank | | 0.000 | 00:58 |
| Subject Sample | # 1 | 0.216 | 00:59 |
| Breath Volume | | 2.777 | 00:59 |
| Air Blank | | 0.000 | 01:00 |
| Air Blank | | 0.000 | 01:02 |
| Subject Sample | # 2 | 0.221 | 01:02 |
| Breath Volume | | 3.402 | 01:02 |
| Air Blank | | 0.000 | 01:03 |
| Control Test | | 0.079 | 01:03 |
| Air Blank | | 0.000 | 01:04 |
| Diagnostics Check | | OK | 01:04 |

Data Download Date/Time: 06/25/2026 01:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUI,DANIEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RUCCI,MATTHEW
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: VEENSTRA

DATE/TIME OF TEST: 05/31/2026 03:32:33

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 03:25 |
| Air Blank | | 0.000 | 03:25 |
| Control Test | | 0.081 | 03:25 |
| Air Blank | | 0.000 | 03:26 |
| Subject Sample | # 1 | 0.174 | 03:27 |
| Breath Volume | | 1.152 | 03:27 |
| Air Blank | | 0.000 | 03:28 |
| Air Blank | | 0.000 | 03:29 |
| Subject Sample | # 2 | 0.166 | 03:30 |
| Breath Volume | | 2.285 | 03:30 |
| Air Blank | | 0.000 | 03:31 |
| Control Test | | 0.080 | 03:31 |
| Air Blank | | 0.000 | 03:32 |
| Diagnostics Check | | OK | 03:32 |

Data Download Date/Time: 06/25/2026 01:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREGER,TENLEY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: GREGER,TENLEY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: PERMENTER

DATE/TIME OF TEST: 05/31/2026 06:38:29

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 06:31 |
| Air Blank | | 0.000 | 06:32 |
| Control Test | | 0.080 | 06:32 |
| Air Blank | | 0.000 | 06:33 |
| Subject Sample | # 1 | 0.119 | 06:33 |
| Breath Volume | | 1.121 | 06:33 |
| Air Blank | | 0.000 | 06:34 |
| Air Blank | | 0.000 | 06:36 |
| Subject Sample | # 2 | 0.122 | 06:36 |
| Breath Volume | | 1.816 | 06:36 |
| Air Blank | | 0.000 | 06:37 |
| Control Test | | 0.080 | 06:37 |
| Air Blank | | 0.000 | 06:38 |
| Diagnostics Check | | OK | 06:38 |

Data Download Date/Time: 06/25/2026 01:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BURKE,LYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: ARVIZU

DATE/TIME OF TEST: 05/31/2026 19:46:02

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 19:39 |
| Air Blank | | 0.000 | 19:39 |
| Control Test | | 0.079 | 19:40 |
| Air Blank | | 0.000 | 19:40 |
| Subject Sample | # 1 | 0.148 | 19:41 |
| Breath Volume | | 1.109 | 19:41 |
| Air Blank | | 0.000 | 19:42 |
| Air Blank | | 0.000 | 19:43 |
| Subject Sample | # 2 | 0.149 | 19:44 |
| Breath Volume | | 1.238 | 19:44 |
| Air Blank | | 0.000 | 19:45 |
| Control Test | | 0.079 | 19:45 |
| Air Blank | | 0.000 | 19:45 |
| Diagnostics Check | | OK | 19:46 |

Data Download Date/Time: 06/25/2026 01:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007515

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MIGLIACCIO,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403525131
DRY GAS EXPIRY DATE: 01/02/2029
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: TELLO WILLARD

DATE/TIME OF TEST: 05/31/2026 23:17:26

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 23:02 |
| Air Blank | | 0.000 | 23:03 |
| Control Test | | 0.080 | 23:03 |
| Air Blank | | 0.000 | 23:04 |
| Subject Sample | # 1 | 0.259 | 23:06 |
| Breath Volume | | 2.546 | 23:06 |
| Air Blank | | 0.000 | 23:07 |
| Air Blank | | 0.000 | 23:08 |
| Subject Sample | # 2 | 0.235 | 23:10 |
| Breath Volume | | 2.664 | 23:10 |
| Air Blank | | 0.000 | 23:11 |
| Air Blank | | 0.000 | 23:12 |
| Subject Sample | # 3 | 0.254 | 23:15 |
| Breath Volume | | 2.968 | 23:15 |
| Air Blank | | 0.000 | 23:16 |
| Control Test | | 0.079 | 23:16 |
| Air Blank | | 0.000 | 23:17 |
| Diagnostics Check | | OK | 23:17 |

Data Download Date/Time: 06/25/2026 01:17

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/16/2026 00:51:10

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 00:47 | |
| Air Blank | | 0.000 | 00:47 | |
| Control Test | | 0.081 | 00:48 | |
| Air Blank | | 0.000 | 00:48 | |
| Subject Sample | # 1 | | 00:49 | Subject Test Refused |
| Breath Volume | | 0.000 | 00:49 | |
| Air Blank | | 0.000 | 00:50 | |
| Control Test | | 0.080 | 00:50 | |
| Air Blank | | 0.000 | 00:51 | |
| Diagnostics Check | | OK | 00:51 | |

Data Download Date/Time: 05/29/2026 19:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,ORESTES,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,ORESTES
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CALDERON BEECH

DATE/TIME OF TEST: 05/17/2026 20:31:59

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:27 | |
| Air Blank | | 0.000 | 20:28 | |
| Control Test | | 0.082 | 20:28 | |
| Air Blank | | 0.000 | 20:29 | |
| Subject Sample | # 1 | | 20:31 | RFI Detect |
| Breath Volume | | 0.000 | 20:31 | |
| Air Blank | | | 20:31 | RFI Detect |
| Air Blank | | 0.000 | 20:31 | |

Data Download Date/Time: 05/29/2026 19:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,ORESTES,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,ORESTES
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CALDERON BEECH

DATE/TIME OF TEST: 05/17/2026 20:42:46

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 20:38 | |
| Air Blank | | 0.000 | 20:38 | |
| Control Test | | 0.081 | 20:38 | |
| Air Blank | | 0.000 | 20:39 | |
| Subject Sample | # 1 | | 20:41 | Subject Test Refused |
| Breath Volume | | 0.000 | 20:41 | |
| Air Blank | | 0.000 | 20:41 | |
| Control Test | | 0.082 | 20:42 | |
| Air Blank | | 0.000 | 20:42 | |
| Diagnostics Check | | OK | 20:42 | |

Data Download Date/Time: 05/29/2026 19:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SABATER,JUSTIN,Y
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SABATER,JUSTIN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/23/2026 05:29:35

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:26 | |
| Air Blank | | 0.000 | 05:26 | |
| Control Test | | 0.081 | 05:27 | |
| Air Blank | | 0.000 | 05:27 | |
| Subject Sample | # 1 | | 05:28 | Sequence Aborted |
| Breath Volume | | 0.000 | 05:28 | |
| Air Blank | | 0.000 | 05:29 | |

Data Download Date/Time: 05/29/2026 19:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SABATER,JUSTIN,Y
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SABATER,JUSTIN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/23/2026 05:37:30

| | | | | |
|---|---|---|---|---|
| Diagnostics Check | | OK | 05:32 | |
| Air Blank | | 0.000 | 05:32 | |
| Control Test | | 0.080 | 05:33 | |
| Air Blank | | 0.000 | 05:33 | |
| Subject Sample | # 1 | | 05:35 | Subject Test Refused |
| Breath Volume | | 0.000 | 05:35 | |
| Air Blank | | 0.000 | 05:36 | |
| Control Test | | 0.080 | 05:36 | |
| Air Blank | | 0.000 | 05:37 | |
| Diagnostics Check | | OK | 05:37 | |

Data Download Date/Time: 05/29/2026 19:26

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.

# *Florida Department of Law Enforcement*

## *Alcohol Testing Program*

## Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ANGEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 04/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:06

SUBJECT'S LAST NAME: REYES-ROMERO

DATE/TIME OF TEST: 05/26/2026 16:34:05

| | | | |
|---|---|---|---|
| Diagnostics Check | | OK | 16:27 |
| Air Blank | | 0.000 | 16:27 |
| Control Test | | 0.080 | 16:27 |
| Air Blank | | 0.000 | 16:28 |
| Subject Sample | # 1 | 0.291 | 16:29 |
| Breath Volume | | 2.527 | 16:29 |
| Air Blank | | 0.000 | 16:29 |
| Air Blank | | 0.000 | 16:31 |
| Subject Sample | # 2 | 0.286 | 16:32 |
| Breath Volume | | 2.890 | 16:32 |
| Air Blank | | 0.000 | 16:33 |
| Control Test | | 0.079 | 16:33 |
| Air Blank | | 0.000 | 16:34 |
| Diagnostics Check | | OK | 16:34 |

Data Download Date/Time: 05/29/2026 18:56

**Printed: 07 August 2026**

COBRA ©1997-2014 CMI, Inc.