

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ANGEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/08/2025 21:31:20

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:27	
Control Test	0.080	21:28	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:29	Subject Test Refused
Breath Volume	0.000	21:29	
Air Blank	0.000	21:30	
Control Test	0.080	21:30	
Air Blank	0.000	21:31	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 08/22/2025 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,ANGEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,ANGEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402732434

DRY GAS EXPIRY DATE: 04/28/2026

VIOLATION CODE: DUI

ARREST TIME: 00:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 08/16/2025 00:41:45

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1 0.143	00:35	Volume Not Met
Breath Volume	0.476	00:35	
Air Blank	0.000	00:35	
Air Blank	0.000	00:37	
Subject Sample	# 2	00:40	Subject Test Refused
Breath Volume	0.855	00:40	
Air Blank	0.000	00:40	
Control Test	0.078	00:41	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 08/22/2025 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,ANGEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,ANGEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402732434

DRY GAS EXPIRY DATE: 04/28/2026

VIOLATION CODE: DUI

ARREST TIME: 01:19

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: NICHOLAS

DATE/TIME OF TEST: 08/17/2025 01:59:25

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.215	01:54
Breath Volume	1.617	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.211	01:57
Breath Volume	2.675	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 08/22/2025 15:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA-RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: DELGADO-SALGADO

DATE/TIME OF TEST: 08/31/2025 04:55:47

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.078	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.256	04:51
Breath Volume	3.632	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.260	04:54
Breath Volume	3.496	04:54
Air Blank	0.000	04:54
Control Test	0.078	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 09/12/2025 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALZATE,MELENNI,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: ALVARADO,ELMER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: ZACARIAS HERNANDEZ

DATE/TIME OF TEST: 08/02/2025 15:43:54

Diagnostics Check	OK	15:37
Air Blank	0.000	15:37
Control Test	0.078	15:38
Air Blank	0.000	15:38
Subject Sample # 1	0.253	15:39
Breath Volume	2.125	15:39
Air Blank	0.000	15:40
Air Blank	0.000	15:41
Subject Sample # 2	0.252	15:42
Breath Volume	2.304	15:42
Air Blank	0.000	15:43
Control Test	0.076	15:43
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

Data Download Date/Time: 08/05/2025 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ANDREW,P
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: QUIROZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/02/2025 23:13:14

Diagnostics Check	OK	23:10	
Air Blank	0.000	23:10	
Control Test	0.079	23:11	
Air Blank	0.000	23:11	
Subject Sample	# 1	23:11	Subject Test Refused
Breath Volume	0.000	23:11	
Air Blank	0.000	23:12	
Control Test	0.079	23:12	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 08/05/2025 01:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CALZADA GUERRERO

DATE/TIME OF TEST: 08/23/2025 01:31:28

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:24	
Control Test	0.079	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1 0.139	01:27	
Breath Volume	1.664	01:27	
Air Blank	0.000	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 2	01:30	RFI Detect
Breath Volume	0.000	01:30	
Air Blank	0.000	01:30	
Control Test	0.000	01:30	Control Outside Tolerance
Air Blank	0.000	01:31	

Data Download Date/Time: 08/27/2025 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEW,REECE,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCGHEE,JOSH
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 07/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 08/09/2025 04:48:31

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.082	04:42
Air Blank	0.000	04:43
Subject Sample # 1	0.114	04:43
Breath Volume	2.660	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.101	04:46
Breath Volume	1.507	04:46
Air Blank	0.000	04:47
Control Test	0.082	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 08/23/2025 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRIDO,JEYNY,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MUNTASER

DATE/TIME OF TEST: 08/02/2025 04:04:25

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.078	04:02	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.077	04:03	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 08/14/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,MANUEL,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: LANDRIAN,FRANCISCO
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: SPIVEY

DATE/TIME OF TEST: 08/04/2025 20:38:15

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:35	
Control Test	0.078	20:36	
Air Blank	0.000	20:36	
Subject Sample	# 1	20:36	Subject Test Refused
Breath Volume	0.000	20:36	
Air Blank	0.000	20:37	
Control Test	0.078	20:37	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 08/14/2025 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CUAREZMA,MICHAEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DAYTON

DATE/TIME OF TEST: 08/06/2025 01:53:11

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.079	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.078	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 08/14/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,MARCOS,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CRUZ,JASON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: BOUCHER

DATE/TIME OF TEST: 08/08/2025 01:10:06

Diagnostics Check	OK	00:59	
Air Blank	0.000	00:59	
Control Test	0.077	01:00	
Air Blank	0.000	01:00	
Subject Sample	# 1 0.158	01:03	Volume Not Met
Breath Volume	0.117	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample	# 2	01:08	Subject Test Refused
Breath Volume	0.757	01:08	
Air Blank	0.000	01:09	
Control Test	0.077	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 08/14/2025 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,MARCOS,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ALVAREZ,D
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: ECKARDT

DATE/TIME OF TEST: 08/09/2025 00:23:36

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.078	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:22	
Control Test	0.078	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 08/14/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CUAREZMA,MICHAEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: JEFFRIES-NIXON

DATE/TIME OF TEST: 08/12/2025 23:22:37

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.115	23:17
Breath Volume	2.468	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.118	23:21
Breath Volume	2.183	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 08/14/2025 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,KELVIN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CUAREZMA,MICHAEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: JEFFRIES-NIXON

DATE/TIME OF TEST: 08/13/2025 00:11:21

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.078	00:09	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:09	Subject Test Refused
Breath Volume	0.000	00:09	
Air Blank	0.000	00:10	
Control Test	0.078	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 08/14/2025 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEZ BARROSO NATARET,JULIO,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: HECHAVARRIA,EDENIS
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: DIEZ BARROSO NATARET

DATE/TIME OF TEST: 08/15/2025 05:06:04

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:01	
Control Test	0.078	05:01	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.136	05:04	
Air Blank	0.000	05:05	
Control Test	0.075	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:06	

Data Download Date/Time: 09/18/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,MARCOS,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ALVAREZ,DANNY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: TUSHINA

DATE/TIME OF TEST: 08/17/2025 02:23:08

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.078	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.000	02:18
Breath Volume	1.937	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.000	02:21
Breath Volume	2.222	02:21
Air Blank	0.000	02:22
Control Test	0.078	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 09/18/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ,JORGE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SONY,ROBENSON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 08/23/2025 03:38:09

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:34	
Control Test	0.078	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Control Test	0.078	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 09/18/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,KENY,
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: GONZALEZ,KENY
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 08/31/2025 05:54:21

Diagnostics Check	OK	05:42	
Air Blank	0.000	05:43	
Control Test	0.078	05:43	
Air Blank	0.000	05:44	
Subject Sample	# 1 0.125	05:47	Volume Not Met
Breath Volume	0.972	05:47	
Air Blank	0.000	05:48	
Air Blank	0.000	05:49	
Subject Sample	# 2 0.153	05:50	
Breath Volume	3.304	05:50	
Air Blank	0.000	05:51	
Air Blank	0.000	05:52	
Subject Sample	# 3 0.148	05:53	
Breath Volume	3.367	05:53	
Air Blank		05:53	RFI Detect
Air Blank	0.000	05:54	

Data Download Date/Time: 09/18/2025 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,KENY,
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: GONZALEZ,KENY
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: JUANAN

DATE/TIME OF TEST: 08/17/2025 05:27:41

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.104	05:22
Breath Volume	1.410	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.107	05:26
Breath Volume	1.882	05:26
Air Blank	0.000	05:26
Control Test	0.080	05:27
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 09/04/2025 03:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MILLAN,ROBERTO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/21/2025 23:01:46

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:59	
Control Test	0.079	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:00	Subject Test Refused
Breath Volume	0.000	23:00	
Air Blank	0.000	23:00	
Control Test	0.080	23:01	
Air Blank	0.000	23:01	
Diagnostics Check	OK	23:01	

Data Download Date/Time: 09/04/2025 03:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FESSOCK,KAILEE
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: EXPOSITO

DATE/TIME OF TEST: 08/22/2025 02:12:19

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.138	02:07
Breath Volume	1.699	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.136	02:10
Breath Volume	4.109	02:10
Air Blank	0.000	02:11
Control Test	0.079	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 09/04/2025 03:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ZARATE,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MARTINEZ MARTINEZ

DATE/TIME OF TEST: 08/24/2025 02:50:02

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:46	
Control Test	0.079	02:47	
Air Blank	0.000	02:47	
Subject Sample	# 1 0.143	02:48	
Breath Volume	1.832	02:48	
Air Blank		02:48	RFI Detect
Air Blank	0.000	02:49	
Control Test	0.081	02:49	Interferent Detect
Air Blank	0.000	02:50	

Data Download Date/Time: 09/04/2025 03:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ZARATE,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MARTINEZ MARTINEZ

DATE/TIME OF TEST: 08/24/2025 02:59:04

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.140	02:54
Breath Volume	2.019	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.140	02:57
Breath Volume	1.929	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:59

Data Download Date/Time: 09/04/2025 03:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/31/2025 15:09:21

Diagnostics Check	OK	15:06	
Air Blank	0.000	15:07	
Control Test	0.079	15:07	
Air Blank	0.000	15:07	
Subject Sample	# 1	15:08	Subject Test Refused
Breath Volume	0.000	15:08	
Air Blank	0.000	15:08	
Control Test	0.080	15:08	
Air Blank	0.000	15:09	
Diagnostics Check	OK	15:09	

Data Download Date/Time: 09/04/2025 03:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PAZ

DATE/TIME OF TEST: 08/09/2025 03:15:05

Diagnostics Check	OK	03:12	
Air Blank	0.000	03:12	
Control Test	0.079	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:14	
Control Test	0.079	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:15	

Data Download Date/Time: 09/04/2025 03:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: TROMBETTA

DATE/TIME OF TEST: 08/15/2025 01:43:30

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.000	01:39
Breath Volume	1.597	01:39
Air Blank	0.000	01:39
Air Blank	0.000	01:41
Subject Sample # 2	0.000	01:42
Breath Volume	1.324	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 09/04/2025 03:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JENKINS,B
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: TANIS

DATE/TIME OF TEST: 08/16/2025 06:55:36

Diagnostics Check	OK	06:52	
Air Blank	0.000	06:53	
Control Test	0.081	06:53	
Air Blank	0.000	06:53	
Subject Sample	# 1	06:54	Subject Test Refused
Breath Volume	0.000	06:54	
Air Blank	0.000	06:54	
Control Test	0.080	06:54	
Air Blank	0.000	06:55	
Diagnostics Check	OK	06:55	

Data Download Date/Time: 09/04/2025 03:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: KILBY

DATE/TIME OF TEST: 08/17/2025 05:10:13

Diagnostics Check	OK	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Subject Sample # 1	0.148	05:05
Breath Volume	1.851	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.155	05:08
Breath Volume	2.175	05:08
Air Blank	0.000	05:09
Control Test	0.079	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 09/04/2025 03:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ROIG,D
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 08/23/2025 04:47:27

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.081	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.249	04:41
Breath Volume	1.226	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.239	04:45
Breath Volume	1.601	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 09/04/2025 03:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,7C02
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VICENTE CHUN

DATE/TIME OF TEST: 08/24/2025 02:00:44

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:57	
Control Test	0.079	01:57	
Air Blank	0.000	01:58	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.081	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 09/04/2025 03:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/29/2025 20:52:37

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.081	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1 0.180	20:42	
Breath Volume	1.242	20:42	
Air Blank	0.000	20:43	
Air Blank	0.000	20:44	
Subject Sample	# 2 0.156	20:48	Volume Not Met
Breath Volume	0.035	20:48	
Air Blank	0.000	20:48	
Air Blank	0.000	20:50	
Subject Sample	# 3 0.175	20:51	
Breath Volume	1.691	20:51	
Air Blank	0.000	20:51	
Control Test	0.077	20:52	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 09/04/2025 03:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RICHARDSON,E
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: CHERLO

DATE/TIME OF TEST: 08/30/2025 02:53:23

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:50	
Control Test	0.080	02:51	
Air Blank	0.000	02:51	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:52	
Control Test	0.078	02:52	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 09/04/2025 03:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RICHARDSON,E
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: CHERLO

DATE/TIME OF TEST: 08/30/2025 03:03:05

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.081	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1 0.117	02:57	
Breath Volume	1.218	02:57	
Air Blank	0.000	02:58	
Air Blank	0.000	03:00	
Subject Sample	# 2 0.118	03:01	
Breath Volume	1.437	03:01	
Air Blank		03:02	RFI Detect
Air Blank	0.000	03:03	

Data Download Date/Time: 09/04/2025 03:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RICHARDSON,E
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 08/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: CHERLO

DATE/TIME OF TEST: 08/30/2025 03:12:19

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.113	03:07
Breath Volume	1.175	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.115	03:10
Breath Volume	1.179	03:10
Air Blank	0.000	03:11
Control Test	0.077	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 09/04/2025 03:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,CRISTIAN,X
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DAVIS,S
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: NEVILLE

DATE/TIME OF TEST: 08/03/2025 17:34:09

Diagnostics Check	OK	17:30	
Air Blank	0.000	17:31	
Control Test	0.080	17:31	
Air Blank	0.000	17:32	
Subject Sample	# 1	17:32	Subject Test Refused
Breath Volume	0.000	17:32	
Air Blank	0.000	17:33	
Control Test	0.080	17:33	
Air Blank	0.000	17:34	
Diagnostics Check	OK	17:34	

Data Download Date/Time: 08/21/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGGINS,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ARAUJO,K
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: MONTES JACINTO

DATE/TIME OF TEST: 08/18/2025 21:41:05

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.081	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.389	21:36
Breath Volume	3.238	21:36
Air Blank	0.000	21:37
Air Blank	0.000	21:38
Subject Sample # 2	0.384	21:39
Breath Volume	3.203	21:39
Air Blank	0.000	21:40
Control Test	0.079	21:40
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 08/21/2025 02:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: OCAMPO

DATE/TIME OF TEST: 08/30/2025 22:27:07

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.081	22:16	
Air Blank	0.000	22:17	
Subject Sample	# 1	22:20	No Sample Provided
Breath Volume	0.000	22:20	
Air Blank	0.000	22:20	
Air Blank	0.000	22:22	
Subject Sample	# 2	22:25	No Sample Provided
Breath Volume	0.000	22:25	
Air Blank	0.000	22:26	
Control Test	0.081	22:26	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 09/04/2025 03:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GARCIA,DAVID
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/31/2025 18:33:45

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.082	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.126	18:28
Breath Volume	2.414	18:28
Air Blank	0.000	18:29
Air Blank	0.000	18:31
Subject Sample # 2	0.118	18:32
Breath Volume	2.207	18:32
Air Blank	0.000	18:32
Control Test	0.079	18:33
Air Blank	0.000	18:33
Diagnostics Check	OK	18:33

Data Download Date/Time: 09/04/2025 03:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: CASTRO,S
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 07/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 08/01/2025 14:22:01

Diagnostics Check	OK	14:15
Air Blank	0.000	14:16
Control Test	0.082	14:16
Air Blank	0.000	14:16
Subject Sample # 1	0.252	14:17
Breath Volume	2.953	14:17
Air Blank	0.000	14:18
Air Blank	0.000	14:20
Subject Sample # 2	0.253	14:20
Breath Volume	1.964	14:20
Air Blank	0.000	14:21
Control Test	0.082	14:21
Air Blank	0.000	14:21
Diagnostics Check	OK	14:22

Data Download Date/Time: 08/07/2025 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGSON,ROLAND,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SINGSON,ROLAND
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: BELTRAN-FORERO

DATE/TIME OF TEST: 08/31/2025 00:21:12

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.078	00:12	
Air Blank	0.000	00:12	
Subject Sample	# 1 0.249	00:15	Volume Not Met
Breath Volume	0.351	00:15	
Air Blank	0.000	00:16	
Air Blank	0.000	00:18	
Subject Sample	# 2	00:19	Subject Test Refused
Breath Volume	0.234	00:19	
Air Blank	0.000	00:20	
Control Test	0.000	00:20	RFI Detect
Air Blank	0.000	00:21	

Data Download Date/Time: 09/17/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGSON,ROLAND,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SINGSON,ROLAND
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: BELTRAN-FOREO

DATE/TIME OF TEST: 08/31/2025 00:27:15

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.076	00:25	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:25	Subject Test Refused
Breath Volume	0.000	00:25	
Air Blank	0.000	00:26	
Control Test	0.077	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 09/17/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: CAMPO

DATE/TIME OF TEST: 08/31/2025 01:05:10

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.077	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Control Test	0.076	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 09/17/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GOODPASTER,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 08/31/2025 02:11:05

Diagnostics Check	OK	02:08	
Air Blank	0.000	02:08	
Control Test	0.077	02:09	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:10	
Control Test	0.077	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:11	

Data Download Date/Time: 09/17/2025 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PETER

DATE/TIME OF TEST: 08/31/2025 03:00:31

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.077	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.061	02:56
Breath Volume	2.945	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.060	02:58
Breath Volume	2.988	02:58
Air Blank	0.000	02:59
Control Test	0.076	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 09/17/2025 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,ANTHONY,R
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GARDNER,ANTHONY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: STANISLAUS

DATE/TIME OF TEST: 08/30/2025 22:56:08

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:53	
Control Test	0.079	22:54	
Air Blank	0.000	22:54	
Subject Sample	# 1	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:55	
Control Test	0.079	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 09/17/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZSIDO,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HICKS,JASON
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 08/31/2025 01:14:04

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:11	
Control Test	0.079	01:11	
Air Blank	0.000	01:12	
Subject Sample	# 1	01:12	Subject Test Refused
Breath Volume	0.000	01:12	
Air Blank	0.000	01:13	
Control Test	0.078	01:13	
Air Blank	0.000	01:13	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 09/17/2025 15:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CRISTIAN,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MILLER,CRISTIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG500903
DRY GAS EXPIRY DATE: 01/09/2027
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: TUNLEY

DATE/TIME OF TEST: 08/31/2025 02:51:07

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.043	02:46
Breath Volume	3.468	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.038	02:49
Breath Volume	2.640	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 09/17/2025 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HANSARD,JOSHUA,D

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: PRITCHARD,THOMAS

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403207508

DRY GAS EXPIRY DATE: 06/27/2026

VIOLATION CODE: BUI

ARREST TIME: 16:52

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 08/03/2025 17:43:20

Diagnostics Check	OK	17:40	
Air Blank	0.000	17:40	
Control Test	0.082	17:41	
Air Blank	0.000	17:41	
Subject Sample	# 1	17:41	Subject Test Refused
Breath Volume	0.000	17:41	
Air Blank	0.000	17:42	
Control Test	0.081	17:42	
Air Blank	0.000	17:43	
Diagnostics Check	OK	17:43	

Data Download Date/Time: 08/19/2025 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 403207508
BREATH TEST OPERATOR: HANSARD,JOSHUA,D	DRY GAS EXPIRY DATE: 06/27/2026
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: PRITCHARD,THOMAS	ARREST TIME: 17:22
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 07/09/2025	OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: MARTIN	DATE/TIME OF TEST: 08/16/2025	18:16:02
-----------------------------	-------------------------------	----------

Diagnostics Check	OK	18:09
Air Blank	0.000	18:09
Control Test	0.081	18:09
Air Blank	0.000	18:10
Subject Sample # 1	0.208	18:11
Breath Volume	3.460	18:11
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample # 2	0.195	18:14
Breath Volume	1.847	18:14
Air Blank	0.000	18:15
Control Test	0.081	18:15
Air Blank	0.000	18:15
Diagnostics Check	OK	18:16

Data Download Date/Time: 08/19/2025 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,DANIEL,S

BREATH TEST OPERATOR AGENCY: Okaloosa County SO

ARRESTING OFFICER: WILLIAMS,DANIEL

ARRESTING OFFICER AGENCY: Okaloosa County SO

LAST AGENCY INSPECTION: 08/19/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403207508

DRY GAS EXPIRY DATE: 12/02/2027

VIOLATION CODE: DUI

ARREST TIME: 19:03

GENDER: Female

OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: HESLER

DATE/TIME OF TEST: 08/23/2025 19:46:06

Diagnostics Check	OK	19:33	
Air Blank	0.000	19:34	
Control Test	0.082	19:34	
Air Blank	0.000	19:34	
Subject Sample	# 1 0.149	19:38	Slope Not Level
Breath Volume	1.324	19:38	
Air Blank	0.000	19:38	
Air Blank	0.000	19:40	
Subject Sample	# 2 0.155	19:41	No .020 agreement
Breath Volume	1.324	19:41	
Air Blank	0.000	19:41	
Air Blank	0.000	19:43	
Subject Sample	# 3 0.177	19:44	No .020 agreement
Breath Volume	1.167	19:44	
Air Blank	0.000	19:45	
Control Test	0.081	19:45	
Air Blank	0.000	19:46	
Diagnostics Check	OK	19:46	

Data Download Date/Time: 09/11/2025 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,E,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HRONCICH,N
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: SANCHEZ PEREZ

DATE/TIME OF TEST: 08/17/2025 00:02:37

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.079	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 1 0.160	23:58	
Breath Volume	3.144	23:58	
Air Blank	0.000	23:59	
Air Blank	0.000	00:00	
		00:00	Improper Sample
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 09/12/2025 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER, ERIC, M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HRONCICH, NIKO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: SANCHEZ PEREZ

DATE/TIME OF TEST: 08/17/2025 00:34:43

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.078	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.138	00:30
Breath Volume	2.585	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.136	00:33
Breath Volume	1.546	00:33
Air Blank	0.000	00:33
Control Test	0.077	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 09/12/2025 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BRIDGE,COLE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 08/17/2025 04:18:23

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:14	No Sample Provided
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 2	04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.079	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 09/12/2025 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FULLER,ZACHARY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: CONSUEGRA ALFONSO

DATE/TIME OF TEST: 08/17/2025 08:04:18

Diagnostics Check	OK	07:58
Air Blank	0.000	07:58
Control Test	0.080	07:58
Air Blank	0.000	07:59
Subject Sample # 1	0.170	07:59
Breath Volume	3.492	07:59
Air Blank	0.000	08:00
Air Blank	0.000	08:02
Subject Sample # 2	0.158	08:02
Breath Volume	2.828	08:02
Air Blank	0.000	08:03
Control Test	0.079	08:03
Air Blank	0.000	08:04
Diagnostics Check	OK	08:04

Data Download Date/Time: 09/12/2025 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,KWESI,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JOHNSON,KWESI
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: MENDOZA ISAGUIRRE

DATE/TIME OF TEST: 08/20/2025 03:02:00

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.139	02:57
Breath Volume	3.074	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.131	03:00
Breath Volume	3.449	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:02

Data Download Date/Time: 09/12/2025 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,ERIC,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BRIDGE,COLE
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: GORMAN

DATE/TIME OF TEST: 08/22/2025 01:17:54

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.080	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.112	01:13
Breath Volume	2.753	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:15
Subject Sample # 2	0.102	01:16
Breath Volume	1.289	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 09/12/2025 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SULZMANN,SAMUEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SACRISTAN

DATE/TIME OF TEST: 08/23/2025 03:22:51

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.182	03:18
Breath Volume	2.636	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.189	03:21
Breath Volume	3.140	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 09/12/2025 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WALLACE, NICHOLAS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: BOLLER

DATE/TIME OF TEST: 08/25/2025 21:13:07

Diagnostics Check	OK	21:06
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.199	21:08
Breath Volume	3.359	21:08
Air Blank	0.000	21:09
Air Blank	0.000	21:10
Subject Sample # 2	0.204	21:11
Breath Volume	3.378	21:11
Air Blank	0.000	21:12
Control Test	0.079	21:12
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 09/12/2025 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDEMANN, ANTHONY, M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: KAVANAGH, SHAWN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BENJAMIN

DATE/TIME OF TEST: 08/25/2025 22:46:31

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.080	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.254	22:41
Breath Volume	2.742	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.241	22:44
Breath Volume	1.183	22:44
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 09/12/2025 20:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGH, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LINDEMANN, ANTHONY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: JARAMILLO

DATE/TIME OF TEST: 08/26/2025 04:53:53

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.275	04:45
Breath Volume	1.207	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.244	04:49
Breath Volume	1.460	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 3	0.231	04:52
Breath Volume	1.742	04:52
Air Blank	0.000	04:52
Control Test	0.076	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 09/12/2025 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: MIDDLETON

DATE/TIME OF TEST: 08/27/2025 20:33:44

Diagnostics Check	OK	20:26
Air Blank	0.000	20:26
Control Test	0.080	20:27
Air Blank	0.000	20:27
Subject Sample # 1	0.230	20:28
Breath Volume	1.750	20:28
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.244	20:32
Breath Volume	2.667	20:32
Air Blank	0.000	20:32
Control Test	0.080	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 09/12/2025 20:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,TASHA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: AZOFEIFA

DATE/TIME OF TEST: 08/28/2025 03:27:06

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.080	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.182	03:22
Breath Volume	3.394	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.184	03:25
Breath Volume	2.039	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 09/12/2025 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VAIL

DATE/TIME OF TEST: 08/30/2025 03:46:56

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.162	03:42
Breath Volume	3.292	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.159	03:45
Breath Volume	3.433	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 09/12/2025 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIETO PEREZ,CARLOS,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARQUEZ
TRIGUERO,EDUANYS
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26824080A1
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:19
GENDER: Female

OBSERVATION PERIOD BEGAN: 14:54

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 08/31/2025 15:20:56

Diagnostics Check	OK	15:14
Air Blank	0.000	15:15
Control Test	0.081	15:15
Air Blank	0.000	15:15
Subject Sample # 1	0.289	15:16
Breath Volume	2.500	15:16
Air Blank	0.000	15:17
Air Blank	0.000	15:18
Subject Sample # 2	0.292	15:19
Breath Volume	2.519	15:19
Air Blank	0.000	15:20
Control Test	0.079	15:20
Air Blank	0.000	15:20
Diagnostics Check	OK	15:20

Data Download Date/Time: 09/12/2025 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,TASHA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DELIGNY,JONATHAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LISI

DATE/TIME OF TEST: 08/09/2025 00:52:36

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.154	00:47
Breath Volume	1.832	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 2	0.159	00:50
Breath Volume	2.113	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Diagnostics Check	OK	00:52

Data Download Date/Time: 08/14/2025 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,TASHA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 08/09/2025 03:44:02

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.104	03:38
Breath Volume	1.867	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.103	03:42
Breath Volume	3.757	03:42
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:44

Data Download Date/Time: 08/14/2025 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOULIHAN,RILEY,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CAMPOS,ABRIL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 08/09/2025 04:54:06

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:40	
Control Test	0.080	04:40	
Air Blank	0.000	04:40	
Subject Sample # 1	0.159	04:42	
Breath Volume	1.316	04:42	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
Subject Sample # 2	0.165	04:48	Volume Not Met
Breath Volume	0.101	04:48	
Air Blank	0.000	04:49	
Air Blank	0.000	04:50	
Subject Sample # 3		04:52	Subject Test Refused
Breath Volume	0.074	04:52	
Air Blank	0.000	04:53	
Control Test	0.077	04:53	
Air Blank	0.000	04:54	
Diagnostics Check	OK	04:54	

Data Download Date/Time: 08/14/2025 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,TASHA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: NATT

DATE/TIME OF TEST: 08/10/2025 02:11:31

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.213	02:04
Breath Volume	3.023	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.195	02:09
Breath Volume	1.644	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 08/14/2025 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, KWESI, L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SULZMAN, SAMUEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 08/10/2025 23:32:07

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:17	
Control Test	0.081	23:17	
Air Blank	0.000	23:17	
Subject Sample	# 1 0.163	23:19	
Breath Volume	1.480	23:19	
Air Blank	0.000	23:20	
Air Blank	0.000	23:21	
Subject Sample	# 2	23:25	No Sample Provided
Breath Volume	0.000	23:25	
Air Blank	0.000	23:25	
Air Blank	0.000	23:27	
Subject Sample	# 3	23:30	No Sample Provided
Breath Volume	0.000	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Diagnostics Check	OK	23:32	

Data Download Date/Time: 08/14/2025 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,KAYLEE,M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FULLER,ZACHARY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 08/11/2025 17:52:48

Diagnostics Check	OK	17:46
Air Blank	0.000	17:46
Control Test	0.080	17:47
Air Blank	0.000	17:47
Subject Sample # 1	0.112	17:48
Breath Volume	3.292	17:48
Air Blank	0.000	17:48
Air Blank	0.000	17:50
Subject Sample # 2	0.113	17:51
Breath Volume	3.222	17:51
Air Blank	0.000	17:51
Control Test	0.079	17:52
Air Blank	0.000	17:52
Diagnostics Check	OK	17:52

Data Download Date/Time: 08/14/2025 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTER,ERIC,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SAINVAL,P
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/09/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: CASTILHO CORREIA

DATE/TIME OF TEST: 08/12/2025 02:44:25

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:40
Breath Volume	3.441	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.000	02:43
Breath Volume	1.449	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 08/14/2025 02:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIAS,HUNTER,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIAS,HUNTER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: CARVAJAL

DATE/TIME OF TEST: 08/01/2025 06:02:40

Diagnostics Check	OK	05:51	
Air Blank	0.000	05:51	
Control Test	0.080	05:51	
Air Blank	0.000	05:52	
Subject Sample	# 1 0.205	05:55	Volume Not Met
Breath Volume	0.332	05:55	
Air Blank	0.000	05:56	
Air Blank	0.000	05:57	
Subject Sample	# 2 0.153	06:01	Volume Not Met
Breath Volume	0.117	06:01	
Air Blank	0.000	06:01	
Control Test	0.078	06:02	
Air Blank	0.000	06:02	
Diagnostics Check	OK	06:02	

Data Download Date/Time: 08/13/2025 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIAS,HUNTER,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIAS,HUNTER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: CARVAJAL

DATE/TIME OF TEST: 08/01/2025 06:16:33

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:05	
Control Test	0.079	06:06	
Air Blank	0.000	06:06	
Subject Sample	# 1 0.237	06:09	Volume Not Met
Breath Volume	0.074	06:09	
Air Blank	0.000	06:10	
Air Blank	0.000	06:12	
Subject Sample	# 2	06:15	Subject Test Refused
Breath Volume	0.000	06:15	
Air Blank	0.000	06:15	
Control Test	0.078	06:15	
Air Blank	0.000	06:16	
Diagnostics Check	OK	06:16	

Data Download Date/Time: 08/13/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MATTHEWS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:24

SUBJECT'S LAST NAME: PAYEUR

DATE/TIME OF TEST: 08/02/2025 10:51:44

Diagnostics Check	OK	10:45
Air Blank	0.000	10:45
Control Test	0.081	10:46
Air Blank	0.000	10:46
Subject Sample # 1	0.000	10:47
Breath Volume	1.566	10:47
Air Blank	0.000	10:47
Air Blank	0.000	10:49
Subject Sample # 2	0.000	10:50
Breath Volume	1.132	10:50
Air Blank	0.000	10:50
Control Test	0.079	10:51
Air Blank	0.000	10:51
Diagnostics Check	OK	10:51

Data Download Date/Time: 08/13/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REINHARDT,KYLE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: REINHARDT,KYLE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: PINO

DATE/TIME OF TEST: 08/03/2025 15:54:40

Diagnostics Check	OK	15:48
Air Blank	0.000	15:48
Control Test	0.081	15:48
Air Blank	0.000	15:49
Subject Sample # 1	0.000	15:50
Breath Volume	1.988	15:50
Air Blank	0.000	15:50
Air Blank	0.000	15:52
Subject Sample # 2	0.000	15:53
Breath Volume	1.378	15:53
Air Blank	0.000	15:53
Control Test	0.079	15:54
Air Blank	0.000	15:54
Diagnostics Check	OK	15:54

Data Download Date/Time: 08/13/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY,RORY,P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DAVIS,CHANDLER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: BUCKNER

DATE/TIME OF TEST: 08/14/2025 14:29:12

Diagnostics Check	OK	14:21
Air Blank	0.000	14:22
Control Test	0.082	14:22
Air Blank	0.000	14:22
Subject Sample # 1	0.239	14:23
Breath Volume	1.609	14:23
Air Blank	0.000	14:24
Air Blank	0.000	14:26
Subject Sample # 2	0.236	14:27
Breath Volume	2.035	14:27
Air Blank	0.000	14:28
Control Test	0.083	14:28
Air Blank	0.000	14:29
Diagnostics Check	OK	14:29

Data Download Date/Time: 09/30/2025 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNEYCUTT,ROBERT,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DUSTMAN,CAYLA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: HUNEYCUTT

DATE/TIME OF TEST: 08/16/2025 23:08:20

Diagnostics Check	OK	23:02
Air Blank	0.000	23:02
Control Test	0.081	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.245	23:03
Breath Volume	4.273	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.226	23:06
Breath Volume	2.792	23:06
Air Blank	0.000	23:07
Control Test	0.078	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 09/30/2025 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEBRETON,QUINN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HINKELL,JACKSON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/17/2025 01:33:29

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.080	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.168	01:29
Breath Volume	2.261	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.160	01:31
Breath Volume	1.949	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 09/30/2025 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: RODRIGUEZ,LORENZO
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 07/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: OTHER
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 08/22/2025 20:53:15

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.079	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.084	20:48
Breath Volume	1.164	20:48
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.089	20:51
Breath Volume	1.132	20:51
Air Blank	0.000	20:52
Control Test	0.077	20:52
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 09/06/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COATNEY,MARANDA,D
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: COATNEY,MARANDA
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: SHIVER

DATE/TIME OF TEST: 08/14/2025 16:54:14

Diagnostics Check	OK	16:45
Air Blank	0.000	16:45
Control Test	0.078	16:45
Air Blank	0.000	16:46
Subject Sample # 1	0.230	16:46
Breath Volume	2.816	16:46
Air Blank	0.000	16:47
Air Blank	0.000	16:49
Subject Sample # 2	0.208	16:49
Breath Volume	2.183	16:49
Air Blank	0.000	16:50
Air Blank	0.000	16:52
Subject Sample # 3	0.213	16:52
Breath Volume	1.351	16:52
Air Blank	0.000	16:53
Control Test	0.076	16:53
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

Data Download Date/Time: 10/29/2025 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIVER,ROBERT,D
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: NASH,CHRISTOPHER
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 07/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HOFFMIRE

DATE/TIME OF TEST: 08/15/2025 23:18:07

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.191	23:13
Breath Volume	1.492	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.190	23:16
Breath Volume	2.355	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 10/29/2025 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SHAWN,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GOODE,BLAZE
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 08/02/2025 07:34:40

Diagnostics Check	OK	07:28
Air Blank	0.000	07:28
Control Test	0.080	07:29
Air Blank	0.000	07:29
Subject Sample # 1	0.066	07:30
Breath Volume	2.500	07:30
Air Blank	0.000	07:30
Air Blank	0.000	07:32
Subject Sample # 2	0.060	07:33
Breath Volume	1.101	07:33
Air Blank	0.000	07:33
Control Test	0.079	07:34
Air Blank	0.000	07:34
Diagnostics Check	OK	07:34

Data Download Date/Time: 09/16/2025 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA,COLBY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LONG,KATHY
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 08/02/2025 17:56:25

Diagnostics Check	OK	17:49
Air Blank	0.000	17:50
Control Test	0.079	17:50
Air Blank	0.000	17:50
Subject Sample # 1	0.334	17:51
Breath Volume	2.089	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:53
Subject Sample # 2	0.315	17:54
Breath Volume	1.453	17:54
Air Blank	0.000	17:55
Control Test	0.078	17:55
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 09/16/2025 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,DAVID,H
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SMITH,DAVID
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/08/2025 04:23:23

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.210	04:18
Breath Volume	4.988	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.204	04:21
Breath Volume	4.816	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 09/16/2025 15:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA,COLBY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BETHEA,COLBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: HIGHTOWER JR

DATE/TIME OF TEST: 08/31/2025 05:11:01

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:08	
Control Test	0.081	05:08	
Air Blank	0.000	05:09	
Subject Sample	# 1	05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:10	
Control Test	0.079	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:11	

Data Download Date/Time: 09/30/2025 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HALL,NOAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 08/04/2025 04:50:14

Diagnostics Check	OK	04:44
Air Blank	0.000	04:44
Control Test	0.082	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.132	04:45
Breath Volume	2.210	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.131	04:48
Breath Volume	1.292	04:48
Air Blank	0.000	04:49
Control Test	0.082	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 08/11/2025 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PHELPS,WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/06/2025 03:05:45

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.244	03:00
Breath Volume	2.656	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.237	03:04
Breath Volume	2.609	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 08/11/2025 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUCHOCKI,RAYMOND,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SUCHOCKI,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/07/2025 04:55:58

Diagnostics Check	OK	04:43	
Air Blank	0.000	04:43	
Control Test	0.082	04:43	
Air Blank	0.000	04:44	
Subject Sample	# 1 0.167	04:44	
Breath Volume	1.976	04:44	
Air Blank	0.000	04:45	
Air Blank	0.000	04:47	
Subject Sample	# 2 0.154	04:50	Volume Not Met
Breath Volume	1.019	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:53	
Subject Sample	# 3 0.170	04:54	
Breath Volume	2.144	04:54	
Air Blank	0.000	04:55	
Control Test	0.080	04:55	
Air Blank	0.000	04:55	
Diagnostics Check	OK	04:55	

Data Download Date/Time: 08/11/2025 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SMITH,ALEXANDER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MITCHNER

DATE/TIME OF TEST: 08/10/2025 03:54:06

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.200	03:49
Breath Volume	1.472	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.193	03:52
Breath Volume	2.191	03:52
Air Blank	0.000	03:53
Control Test	0.081	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 08/11/2025 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROGERSON,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: DAWKINS

DATE/TIME OF TEST: 08/10/2025 05:28:41

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.081	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.132	05:24
Breath Volume	2.914	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.128	05:27
Breath Volume	2.218	05:27
Air Blank	0.000	05:27
Control Test	0.081	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 08/11/2025 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: MONTGOMERY,WILLIAM
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DIEZ ZAPATA

DATE/TIME OF TEST: 08/11/2025 01:58:22

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.081	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.189	01:53
Breath Volume	1.960	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.192	01:56
Breath Volume	2.609	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 08/11/2025 11:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN,DANIEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: CHATHAM

DATE/TIME OF TEST: 08/18/2025 16:26:23

Diagnostics Check	OK	16:19
Air Blank	0.000	16:20
Control Test	0.080	16:20
Air Blank	0.000	16:20
Subject Sample # 1	0.267	16:21
Breath Volume	2.730	16:21
Air Blank	0.000	16:22
Air Blank	0.000	16:23
Subject Sample # 2	0.259	16:24
Breath Volume	2.582	16:24
Air Blank	0.000	16:25
Control Test	0.081	16:25
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 09/15/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ROGERSON,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: CHAIRES

DATE/TIME OF TEST: 08/19/2025 06:42:57

Diagnostics Check	OK	06:35
Air Blank	0.000	06:36
Control Test	0.081	06:36
Air Blank	0.000	06:37
Subject Sample # 1	0.103	06:38
Breath Volume	1.750	06:38
Air Blank	0.000	06:38
Air Blank	0.000	06:40
Subject Sample # 2	0.092	06:41
Breath Volume	1.289	06:41
Air Blank	0.000	06:42
Control Test	0.081	06:42
Air Blank	0.000	06:42
Diagnostics Check	OK	06:42

Data Download Date/Time: 09/15/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KING,CASEY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: DALLAS

DATE/TIME OF TEST: 08/20/2025 04:52:08

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.147	04:47
Breath Volume	2.933	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.138	04:50
Breath Volume	2.406	04:50
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 09/15/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HALL,NOAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 08/22/2025 03:03:39

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.082	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.166	02:59
Breath Volume	1.890	02:59
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.169	03:02
Breath Volume	2.109	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:03
Air Blank	0.000	03:03
Diagnostics Check	OK	03:03

Data Download Date/Time: 09/15/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: PHELPS,WILLIAM
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: HENLEY

DATE/TIME OF TEST: 08/22/2025 05:04:50

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.081	04:59
Air Blank	0.000	04:59
Subject Sample # 1	0.258	05:00
Breath Volume	2.699	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample # 2	0.258	05:03
Breath Volume	2.863	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 09/15/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/23/2025 05:05:33

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.081	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.089	05:01
Breath Volume	1.648	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.087	05:04
Breath Volume	1.421	05:04
Air Blank	0.000	05:04
Control Test	0.082	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 09/15/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABB,CALEB,W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BABB,CALEB
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 08/23/2025 07:09:48

Diagnostics Check	OK	07:03
Air Blank	0.000	07:04
Control Test	0.081	07:04
Air Blank	0.000	07:05
Subject Sample # 1	0.110	07:05
Breath Volume	1.558	07:05
Air Blank	0.000	07:06
Air Blank	0.000	07:07
Subject Sample # 2	0.107	07:08
Breath Volume	2.386	07:08
Air Blank	0.000	07:08
Control Test	0.080	07:09
Air Blank	0.000	07:09
Diagnostics Check	OK	07:09

Data Download Date/Time: 09/15/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD, N
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 08/24/2025 18:37:54

Diagnostics Check	OK	18:24
Air Blank	0.000	18:25
Control Test	0.080	18:25
Air Blank	0.000	18:25
Subject Sample # 1	0.238	18:27
Breath Volume	1.152	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.281	18:32
Breath Volume	2.421	18:32
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 3	0.263	18:36
Breath Volume	1.210	18:36
Air Blank	0.000	18:36
Control Test	0.079	18:37
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

Data Download Date/Time: 09/15/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JONATHAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 08/27/2025 01:57:48

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.081	01:42	
Air Blank	0.000	01:43	
Subject Sample	# 1 0.225	01:46	Volume Not Met
Breath Volume	0.390	01:46	
Air Blank	0.000	01:47	
Air Blank	0.000	01:48	
Subject Sample	# 2 0.274	01:50	
Breath Volume	1.656	01:50	
Air Blank	0.000	01:50	
Air Blank	0.000	01:52	
Subject Sample	# 3 0.248	01:55	
Breath Volume	1.535	01:55	
Air Blank	0.000	01:56	
Control Test	0.073	01:57	Control Outside Tolerance
Air Blank	0.000	01:57	

Data Download Date/Time: 09/15/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: WILL,HEATHER
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 08/09/2025 12:00:01

Diagnostics Check	OK	11:57	
Air Blank	0.000	11:57	
Control Test	0.081	11:57	
Air Blank	0.000	11:58	
Subject Sample	# 1	11:58	Subject Test Refused
Breath Volume	0.000	11:58	
Air Blank	0.000	11:59	
Control Test	0.081	11:59	
Air Blank	0.000	11:59	
Diagnostics Check	OK	12:00	

Data Download Date/Time: 09/11/2025 11:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SUPLICE, LOVINGSON
ARRESTING OFFICER AGENCY: Eatonville PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: LAROACH

DATE/TIME OF TEST: 08/01/2025 22:23:09

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.119	22:18
Breath Volume	3.941	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.120	22:21
Breath Volume	4.046	22:21
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 08/13/2025 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NILSSON,ERIK
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 07/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/08/2025 00:07:34

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.082	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.201	00:03
Breath Volume	2.113	00:03
Air Blank	0.000	00:03
Air Blank	0.000	00:05
Subject Sample # 2	0.204	00:06
Breath Volume	2.394	00:06
Air Blank	0.000	00:06
Control Test	0.081	00:06
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 08/13/2025 10:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHEWS,ZACHARY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ORTIZ RIVAS

DATE/TIME OF TEST: 08/14/2025 22:53:01

Diagnostics Check	OK	22:43
Air Blank	0.000	22:44
Control Test	0.082	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.065	22:46
Breath Volume	3.347	22:46
Air Blank	0.000	22:47
Air Blank	0.000	22:48
Subject Sample # 2	0.060	22:51
Breath Volume	2.027	22:51
Air Blank	0.000	22:52
Control Test	0.082	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:53

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCMORROW,RYAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 08/15/2025 02:14:21

Diagnostics Check	OK	02:07
Air Blank	0.000	02:07
Control Test	0.082	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.207	02:09
Breath Volume	1.226	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.220	02:12
Breath Volume	2.308	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WRAY,JACOB
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: BAQUE INIGUEZ

DATE/TIME OF TEST: 08/15/2025 06:07:01

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.082	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.136	06:02
Breath Volume	2.035	06:02
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.135	06:05
Breath Volume	1.300	06:05
Air Blank	0.000	06:06
Control Test	0.081	06:06
Air Blank	0.000	06:06
Diagnostics Check	OK	06:07

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: EBBIN

DATE/TIME OF TEST: 08/16/2025 00:26:13

Diagnostics Check	OK	00:19
Air Blank	0.000	00:20
Control Test	0.082	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.000	00:21
Breath Volume	2.359	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.000	00:24
Breath Volume	2.414	00:24
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONTE,COREY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: CHINO RODRIGUEZ

DATE/TIME OF TEST: 08/16/2025 02:37:25

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.082	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.212	02:32
Breath Volume	1.859	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.226	02:35
Breath Volume	1.234	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIELS,DAMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: DEPINTO

DATE/TIME OF TEST: 08/16/2025 05:28:55

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.082	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.178	05:24
Breath Volume	5.132	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.180	05:27
Breath Volume	4.546	05:27
Air Blank	0.000	05:28
Control Test	0.081	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONTE,COREY
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CLAYTONBRYSON

DATE/TIME OF TEST: 08/16/2025 20:31:29

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:28	
Control Test	0.082	20:29	
Air Blank	0.000	20:29	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.082	20:30	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARVAJAL,JUAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ORTEGALOPEZ

DATE/TIME OF TEST: 08/17/2025 01:07:05

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.082	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.224	01:02
Breath Volume	2.148	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.227	01:05
Breath Volume	1.929	01:05
Air Blank	0.000	01:06
Control Test	0.081	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONERLY,BRANDON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MAHAMAD

DATE/TIME OF TEST: 08/17/2025 03:13:02

Diagnostics Check	OK	03:01	
Air Blank	0.000	03:02	
Control Test	0.082	03:02	
Air Blank	0.000	03:02	
Subject Sample	# 1	03:06	No Sample Provided
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Air Blank	0.000	03:08	
Subject Sample	# 2	03:11	No Sample Provided
Breath Volume	0.000	03:11	
Air Blank	0.000	03:12	
Control Test	0.082	03:12	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORRERO,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: BARKSDALE

DATE/TIME OF TEST: 08/19/2025 20:06:44

Diagnostics Check	OK	20:00
Air Blank	0.000	20:00
Control Test	0.083	20:00
Air Blank	0.000	20:01
Subject Sample # 1	0.000	20:02
Breath Volume	2.707	20:02
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.000	20:05
Breath Volume	1.117	20:05
Air Blank	0.000	20:05
Control Test	0.081	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 09/01/2025 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: VAZQUEZGODINEZ

DATE/TIME OF TEST: 08/19/2025 23:38:58

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.082	23:32
Air Blank	0.000	23:33
Subject Sample # 1	0.201	23:34
Breath Volume	1.207	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.195	23:37
Breath Volume	1.171	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCHMALL,NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: SLEIGHT

DATE/TIME OF TEST: 08/21/2025 02:14:30

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.081	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.143	02:10
Breath Volume	1.867	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.139	02:13
Breath Volume	1.152	02:13
Air Blank	0.000	02:13
Control Test	0.081	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,ANTONIO
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ZULUAGA

DATE/TIME OF TEST: 08/21/2025 03:43:46

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.182	03:39
Breath Volume	3.441	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.187	03:42
Breath Volume	3.472	03:42
Air Blank	0.000	03:42
Control Test	0.081	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,ANTONIO
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 08/21/2025 23:49:48

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.082	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.179	23:45
Breath Volume	1.468	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.175	23:48
Breath Volume	2.078	23:48
Air Blank	0.000	23:48
Control Test	0.081	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRUITT,JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: RODRIGUEZSALGADO

DATE/TIME OF TEST: 08/22/2025 03:56:33

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.082	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.102	03:52
Breath Volume	2.394	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.100	03:55
Breath Volume	2.488	03:55
Air Blank	0.000	03:55
Control Test	0.082	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARRERO,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: LOZANOMOYA

DATE/TIME OF TEST: 08/22/2025 20:57:54

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.082	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.220	20:53
Breath Volume	1.398	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.223	20:56
Breath Volume	1.531	20:56
Air Blank	0.000	20:56
Control Test	0.079	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEITT,DAVIN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ETHRIDGE

DATE/TIME OF TEST: 08/23/2025 00:38:05

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.082	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1 0.157	00:30	Volume Not Met
Breath Volume	0.035	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample	# 2 0.155	00:36	Volume Not Met
Breath Volume	0.281	00:36	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GUERRIER,JEFFERSON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: JACOBO LLANAS

DATE/TIME OF TEST: 08/23/2025 03:48:43

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.082	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.040	03:44
Breath Volume	1.390	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.042	03:47
Breath Volume	1.742	03:47
Air Blank	0.000	03:47
Control Test	0.081	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSA,BRANDON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LALLHARRY

DATE/TIME OF TEST: 08/23/2025 04:14:36

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.082	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.094	04:10
Breath Volume	2.121	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.096	04:13
Breath Volume	2.578	04:13
Air Blank	0.000	04:13
Control Test	0.081	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KLINE, CHRISTINA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: SEMEXANT

DATE/TIME OF TEST: 08/23/2025 06:37:19

Diagnostics Check	OK	06:31
Air Blank	0.000	06:31
Control Test	0.082	06:32
Air Blank	0.000	06:32
Subject Sample # 1	0.130	06:32
Breath Volume	2.128	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:35
Subject Sample # 2	0.128	06:35
Breath Volume	2.285	06:35
Air Blank	0.000	06:36
Control Test	0.081	06:36
Air Blank	0.000	06:37
Diagnostics Check	OK	06:37

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIDOMENICO, GINO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: JUAREZ MEZA

DATE/TIME OF TEST: 08/23/2025 23:18:14

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.172	23:13
Breath Volume	1.574	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.172	23:16
Breath Volume	1.265	23:16
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAAVEDRA, NIKOLAS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 08/24/2025 02:50:18

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.082	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.091	02:45
Breath Volume	1.824	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.087	02:48
Breath Volume	1.871	02:48
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIOS,JOSHUA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 08/24/2025 04:15:08

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.222	04:10
Breath Volume	3.375	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.228	04:13
Breath Volume	3.671	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WRAY, JACOB
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: DIAZ MARTINEZ

DATE/TIME OF TEST: 08/24/2025 04:44:43

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.082	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.188	04:39
Breath Volume	1.226	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.191	04:43
Breath Volume	2.671	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:44
Air Blank	0.000	04:44
Diagnostics Check	OK	04:44

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSA,BRANDON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 08/24/2025 06:01:16

Diagnostics Check	OK	05:44	
Air Blank	0.000	05:44	
Control Test	0.081	05:45	
Air Blank	0.000	05:45	
Subject Sample	# 1 0.098	05:48	Volume Not Met
Breath Volume	0.476	05:48	
Air Blank	0.000	05:49	
Air Blank	0.000	05:51	
Subject Sample	# 2 0.131	05:54	
Breath Volume	1.316	05:54	
Air Blank	0.000	05:54	
Air Blank	0.000	05:56	
Subject Sample	# 3	05:59	No Sample Provided
Breath Volume	0.000	05:59	
Air Blank	0.000	06:00	
Control Test	0.079	06:00	
Air Blank	0.000	06:01	
Diagnostics Check	OK	06:01	

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PATTERSON,ZAKAI
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 08/24/2025 21:57:09

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:46	
Control Test	0.082	21:46	
Air Blank	0.000	21:47	
Subject Sample	# 1	21:50	No Sample Provided
Breath Volume	0.000	21:50	
Air Blank	0.000	21:50	
Air Blank	0.000	21:52	
Subject Sample	# 2	21:55	No Sample Provided
Breath Volume	0.000	21:55	
Air Blank	0.000	21:56	
Control Test	0.083	21:56	
Air Blank	0.000	21:57	
Diagnostics Check	OK	21:57	

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAHLAU,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: AYALA ELIAS

DATE/TIME OF TEST: 08/24/2025 22:17:29

Diagnostics Check	OK	22:10
Air Blank	0.000	22:11
Control Test	0.083	22:11
Air Blank	0.000	22:12
Subject Sample # 1	0.243	22:12
Breath Volume	2.179	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.229	22:15
Breath Volume	2.097	22:15
Air Blank	0.000	22:16
Control Test	0.082	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIOS,JOSHUA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 08/25/2025 03:35:51

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.083	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.187	03:31
Breath Volume	3.230	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.188	03:34
Breath Volume	3.917	03:34
Air Blank	0.000	03:34
Control Test	0.081	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NUNEZ,JAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 08/25/2025 04:28:46

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.083	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.149	04:24
Breath Volume	1.929	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.146	04:27
Breath Volume	2.710	04:27
Air Blank	0.000	04:27
Control Test	0.082	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/26/2025 00:02:59

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.082	23:57
Air Blank	0.000	23:58
Subject Sample # 2	0.000	23:58
Breath Volume	2.644	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.000	00:01
Breath Volume	1.460	00:01
Air Blank	0.000	00:02
Control Test	0.082	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,AUSTIN,T
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MEZQUITA,FRANCISCO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: CARRO

DATE/TIME OF TEST: 08/26/2025 06:29:07

Diagnostics Check	OK	06:19
Air Blank	0.000	06:20
Control Test	0.082	06:20
Air Blank	0.000	06:21
Subject Sample # 1	0.227	06:21
Breath Volume	2.203	06:21
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.202	06:24
Breath Volume	2.054	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 3	0.226	06:27
Breath Volume	2.132	06:27
Air Blank	0.000	06:28
Control Test	0.081	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSA,BRANDON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: FIGUEROA JR

DATE/TIME OF TEST: 08/27/2025 04:33:33

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.082	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.177	04:29
Breath Volume	1.738	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.180	04:32
Breath Volume	2.660	04:32
Air Blank	0.000	04:32
Control Test	0.082	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 09/01/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WOMACK, ANTIA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: HEALY

DATE/TIME OF TEST: 08/27/2025 20:44:20

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.083	20:39
Air Blank	0.000	20:39
Subject Sample # 1	0.175	20:39
Breath Volume	2.484	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.175	20:42
Breath Volume	2.574	20:42
Air Blank	0.000	20:43
Control Test	0.082	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RILEY,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 08/27/2025 21:51:37

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.082	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.201	21:47
Breath Volume	2.222	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.201	21:50
Breath Volume	3.574	21:50
Air Blank	0.000	21:50
Control Test	0.081	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECKER, THOMAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: POGUE

DATE/TIME OF TEST: 08/28/2025 03:00:50

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:58	
Control Test	0.083	02:58	
Air Blank	0.000	02:59	
Subject Sample	# 1	02:59	Subject Test Refused
Breath Volume	0.000	02:59	
Air Blank	0.000	02:59	
Control Test	0.082	03:00	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUFKIN,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: DESUEZACABREJA

DATE/TIME OF TEST: 08/28/2025 05:03:52

Diagnostics Check	OK	04:57
Air Blank	0.000	04:58
Control Test	0.082	04:58
Air Blank	0.000	04:59
Subject Sample # 1	0.151	04:59
Breath Volume	3.125	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.142	05:02
Breath Volume	1.757	05:02
Air Blank	0.000	05:02
Control Test	0.081	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: MACHIAVELLO

DATE/TIME OF TEST: 08/28/2025 17:06:15

Diagnostics Check	OK	16:59
Air Blank	0.000	16:59
Control Test	0.081	17:00
Air Blank	0.000	17:00
Subject Sample # 1	0.232	17:01
Breath Volume	1.230	17:01
Air Blank	0.000	17:02
Air Blank	0.000	17:04
Subject Sample # 2	0.236	17:04
Breath Volume	1.300	17:04
Air Blank	0.000	17:05
Control Test	0.080	17:05
Air Blank	0.000	17:06
Diagnostics Check	OK	17:06

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA, AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BOLUFER SANABRIA

DATE/TIME OF TEST: 08/28/2025 20:01:56

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.082	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.276	19:57
Breath Volume	1.558	19:57
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.275	20:00
Breath Volume	2.570	20:00
Air Blank	0.000	20:01
Control Test	0.080	20:01
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANTOS, STEVEN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: ROQUE

DATE/TIME OF TEST: 08/29/2025 21:47:31

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.082	21:41
Air Blank	0.000	21:41
Subject Sample # 1	0.134	21:42
Breath Volume	1.218	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.141	21:46
Breath Volume	1.441	21:46
Air Blank	0.000	21:46
Control Test	0.081	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DURAN,ALFREDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BARRETO DIAZ

DATE/TIME OF TEST: 08/30/2025 02:51:38

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.082	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.130	02:47
Breath Volume	2.582	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.139	02:50
Breath Volume	2.656	02:50
Air Blank	0.000	02:50
Control Test	0.081	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIELS,DAMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/30/2025 04:28:22

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.082	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.035	04:23
Breath Volume	2.511	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.035	04:26
Breath Volume	2.765	04:26
Air Blank	0.000	04:27
Control Test	0.082	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRORY, DIARMUID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 08/30/2025 04:59:06

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Subject Sample # 1	0.159	04:54
Breath Volume	2.738	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.156	04:57
Breath Volume	2.566	04:57
Air Blank	0.000	04:58
Control Test	0.081	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALICEA GALLOZA,MARK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: CANO JIMENEZ

DATE/TIME OF TEST: 08/31/2025 03:57:35

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.082	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.231	03:52
Breath Volume	1.585	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.227	03:56
Breath Volume	2.023	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 08/31/2025 06:58:14

Diagnostics Check	OK	06:52
Air Blank	0.000	06:52
Control Test	0.082	06:53
Air Blank	0.000	06:53
Subject Sample # 1	0.106	06:53
Breath Volume	2.738	06:53
Air Blank	0.000	06:54
Air Blank	0.000	06:56
Subject Sample # 2	0.109	06:56
Breath Volume	2.820	06:56
Air Blank	0.000	06:57
Control Test	0.081	06:57
Air Blank	0.000	06:58
Diagnostics Check	OK	06:58

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARION,MICHAEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: GARZON

DATE/TIME OF TEST: 08/31/2025 21:56:00

Diagnostics Check	OK	21:49
Air Blank	0.000	21:50
Control Test	0.082	21:50
Air Blank	0.000	21:51
Subject Sample # 1	0.283	21:51
Breath Volume	2.343	21:51
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.293	21:54
Breath Volume	1.882	21:54
Air Blank	0.000	21:55
Control Test	0.081	21:55
Air Blank	0.000	21:55
Diagnostics Check	OK	21:56

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BANOS,FRANK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 08/31/2025 23:00:22

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.082	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.281	22:55
Breath Volume	1.859	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.285	22:58
Breath Volume	1.636	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 09/01/2025 11:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BENAFIELD,RYAN,E

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: BENAFIELD,RYAN

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 05:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: PEREZ ZELEDON

DATE/TIME OF TEST: 08/06/2025 06:04:54

Diagnostics Check	OK	05:58
Air Blank	0.000	05:59
Control Test	0.079	05:59
Air Blank	0.000	05:59
Subject Sample # 1	0.159	06:00
Breath Volume	2.562	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.144	06:03
Breath Volume	1.789	06:03
Air Blank	0.000	06:03
Control Test	0.078	06:04
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 08/25/2025 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,R,R
BREATH TEST OPERATOR AGENCY: Midway PD
ARRESTING OFFICER: PIERRE,E
ARRESTING OFFICER AGENCY: Midway PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: DENNARD

DATE/TIME OF TEST: 08/15/2025 01:38:25

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.171	01:34
Breath Volume	1.335	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.173	01:36
Breath Volume	1.171	01:36
Air Blank	0.000	01:37
Control Test	0.079	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 08/25/2025 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALDAY,SETH,A
BREATH TEST OPERATOR AGENCY: Chattahoochee PD
ARRESTING OFFICER: ALDAY,SETH
ARRESTING OFFICER AGENCY: Chattahoochee PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 08/23/2025 01:54:14

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.271	01:49
Breath Volume	1.667	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.272	01:52
Breath Volume	1.347	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 08/25/2025 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARABALLO, DOMINICK, A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HOWARD, JESSE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DUCKER

DATE/TIME OF TEST: 08/01/2025 02:10:45

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.199	02:06
Breath Volume	2.585	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.190	02:09
Breath Volume	2.320	02:09
Air Blank	0.000	02:09
Control Test	0.078	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 08/31/2025 14:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON, WILLIAM, M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: WATSON, WAILLIAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 08/02/2025 17:05:58

Diagnostics Check	OK	16:59
Air Blank	0.000	16:59
Control Test	0.079	17:00
Air Blank	0.000	17:00
Subject Sample # 1	0.339	17:01
Breath Volume	2.875	17:01
Air Blank	0.000	17:01
Air Blank	0.000	17:03
Subject Sample # 2	0.336	17:04
Breath Volume	2.187	17:04
Air Blank	0.000	17:05
Control Test	0.078	17:05
Air Blank	0.000	17:05
Diagnostics Check	OK	17:05

Data Download Date/Time: 09/05/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ-GUILLEN,MARCOS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LOPEZ-GUILLEN,MARCOS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 10:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: SCIAL LIS

DATE/TIME OF TEST: 08/07/2025 11:16:59

Diagnostics Check	OK	11:10
Air Blank	0.000	11:10
Control Test	0.079	11:11
Air Blank	0.000	11:11
Subject Sample # 1	0.000	11:12
Breath Volume	4.011	11:12
Air Blank	0.000	11:13
Air Blank	0.000	11:15
Subject Sample # 2	0.000	11:15
Breath Volume	3.996	11:15
Air Blank	0.000	11:16
Control Test	0.080	11:16
Air Blank	0.000	11:16
Diagnostics Check	OK	11:16

Data Download Date/Time: 09/05/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: ROUMELIOTIS

DATE/TIME OF TEST: 08/07/2025 19:25:46

Diagnostics Check	OK	19:17
Air Blank	0.000	19:18
Control Test	0.079	19:18
Air Blank	0.000	19:19
Subject Sample # 1	0.217	19:20
Breath Volume	1.136	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.207	19:24
Breath Volume	1.417	19:24
Air Blank	0.000	19:24
Control Test	0.076	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 09/05/2025 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARABALLO,FTO D.,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: REINERT,DEPUTY P.
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 08/09/2025 04:49:02

Diagnostics Check	OK	04:37	
Air Blank	0.000	04:37	
Control Test	0.079	04:38	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:41	No Sample Provided
Breath Volume	0.000	04:41	
Air Blank	0.000	04:42	
Air Blank	0.000	04:44	
Subject Sample	# 2	04:47	Volume Not Met
Breath Volume	0.621	04:47	
Air Blank	0.000	04:48	
Control Test	0.080	04:48	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 09/05/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRACCICA,CARMEN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERARDI,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WOODEN

DATE/TIME OF TEST: 08/14/2025 03:55:52

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.074	03:55	Control Outside Tolerance
Air Blank	0.000	03:55	

Data Download Date/Time: 09/05/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRACCICA,CARMEN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERARDI,JOHN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WOODEN

DATE/TIME OF TEST: 08/14/2025 04:05:03

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.203	04:00
Breath Volume	2.289	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:02
Subject Sample # 2	0.214	04:03
Breath Volume	2.386	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:05

Data Download Date/Time: 09/05/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRACCICA,CARMEN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GIRARDI,JOHN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: NORTON

DATE/TIME OF TEST: 08/15/2025 01:42:10

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.038	01:37
Breath Volume	3.910	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:39
Subject Sample # 2	0.034	01:40
Breath Volume	2.027	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 09/05/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRACCICA,CARMEN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FREDRICKSON,JARROD
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 08/15/2025 02:03:51

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.078	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.135	01:59
Breath Volume	2.203	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:01
Subject Sample # 2	0.129	02:02
Breath Volume	1.757	02:02
Air Blank	0.000	02:02
Control Test	0.079	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 09/05/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 08/16/2025 04:00:04

Diagnostics Check	OK	03:48	
Air Blank	0.000	03:49	
Control Test	0.079	03:49	
Air Blank	0.000	03:49	
Subject Sample	# 1	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:53	
Air Blank	0.000	03:55	
Subject Sample	# 2	03:58	No Sample Provided
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	03:59	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 09/05/2025 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: BRAVO

DATE/TIME OF TEST: 08/16/2025 05:16:12

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.089	05:11
Breath Volume	3.406	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.088	05:14
Breath Volume	3.800	05:14
Air Blank	0.000	05:15
Control Test	0.078	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 09/05/2025 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARABLO,FTO D.,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HOWARD,DEPUTY J.
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 08/24/2025 02:22:33

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.079	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:14	No Sample Provided
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Air Blank	0.000	02:17	
Subject Sample	# 2	02:20	Volume Not Met
Breath Volume	0.078	02:20	
Air Blank	0.000	02:21	
Control Test	0.079	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 09/05/2025 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARABALLO,FTO D.,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: HOWARD,DEPUTY J.
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 08/24/2025 02:35:35

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.077	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.165	02:29
Breath Volume	2.710	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.177	02:33
Breath Volume	2.617	02:33
Air Blank	0.000	02:34
Control Test	0.077	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 09/05/2025 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: HILTON

DATE/TIME OF TEST: 08/30/2025 23:55:17

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.138	23:50
Breath Volume	3.027	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.138	23:53
Breath Volume	2.894	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 09/05/2025 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOE,MARISA,P
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 07/31/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403096064
DRY GAS EXPIRY DATE: 07/16/2027
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: LAKHANI

DATE/TIME OF TEST: 08/31/2025 00:34:50

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.068	00:30
Breath Volume	1.500	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:32
Subject Sample # 2	0.067	00:33
Breath Volume	1.304	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 09/05/2025 09:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO-PEREZ,MICHAEL,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ALVAREZ,GUMERSINDO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 08/01/2025 22:56:11

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.080	22:50
Air Blank	0.000	22:50
Subject Sample # 1	0.091	22:51
Breath Volume	1.140	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.088	22:54
Breath Volume	1.558	22:54
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ROSA,DANIEL
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: ASIS SILAH

DATE/TIME OF TEST: 08/02/2025 05:34:39

Diagnostics Check	OK	05:27
Air Blank	0.000	05:27
Control Test	0.079	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.158	05:30
Breath Volume	1.527	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.166	05:33
Breath Volume	1.945	05:33
Air Blank	0.000	05:33
Control Test	0.078	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NIEVES, CHARLIE,

BREATH TEST OPERATOR AGENCY: Kissimmee PD

ARRESTING OFFICER: CAMACHO, FRANCES

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403207507

DRY GAS EXPIRY DATE: 11/26/2027

VIOLATION CODE: DUI

ARREST TIME: 18:56

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: CARRION DIAZ

DATE/TIME OF TEST: 08/02/2025 20:12:12

Diagnostics Check	OK	19:59	
Air Blank	0.000	20:00	
Control Test	0.080	20:00	
Air Blank	0.000	20:00	
Subject Sample	# 1 0.111	20:04	Volume Not Met
Breath Volume	0.765	20:04	
Air Blank	0.000	20:04	
Air Blank	0.000	20:06	
Subject Sample	# 2 0.149	20:06	
Breath Volume	2.234	20:06	
Air Blank	0.000	20:07	
Air Blank	0.000	20:09	
Subject Sample	# 3 0.146	20:10	
Breath Volume	2.375	20:10	
Air Blank	0.000	20:11	
Control Test	0.078	20:11	
Air Blank	0.000	20:12	
Diagnostics Check	OK	20:12	

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 08/03/2025 03:32:54

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.081	03:28
Breath Volume	1.144	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.079	03:31
Breath Volume	1.417	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMERO-PEREZ,MICHAEL,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLADES,JASON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: OHARA

DATE/TIME OF TEST: 08/03/2025 05:27:26

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.080	05:21
Air Blank	0.000	05:22
Subject Sample # 1	0.144	05:22
Breath Volume	4.023	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:25
Subject Sample # 2	0.140	05:25
Breath Volume	2.582	05:25
Air Blank	0.000	05:26
Control Test	0.078	05:26
Air Blank	0.000	05:27
Diagnostics Check	OK	05:27

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPINSKI,LUKE,W
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: KRUPINSKI,LUKE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: BRAZ

DATE/TIME OF TEST: 08/04/2025 06:36:03

Diagnostics Check	OK	06:24	
Air Blank	0.000	06:24	
Control Test	0.080	06:25	
Air Blank	0.000	06:25	
Subject Sample	# 1	06:28	No Sample Provided
Breath Volume	0.000	06:28	
Air Blank	0.000	06:29	
Air Blank	0.000	06:31	
Subject Sample	# 2	06:34	Volume Not Met
Breath Volume	0.050	06:34	
Air Blank	0.000	06:35	
Control Test	0.078	06:35	
Air Blank	0.000	06:35	
Diagnostics Check	OK	06:36	

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LOPEZ,DERICK
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LONGTINE

DATE/TIME OF TEST: 08/07/2025 00:56:32

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.079	00:49
Air Blank	0.000	00:50
Subject Sample # 1	0.116	00:52
Breath Volume	1.558	00:52
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.115	00:55
Breath Volume	1.710	00:55
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: COLON CRUZ

DATE/TIME OF TEST: 08/09/2025 03:48:39

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.173	03:43
Breath Volume	1.199	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.166	03:47
Breath Volume	1.242	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARZYGnat,DESTINY,P
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MOYA,NATHANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 06:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: LAFLAM

DATE/TIME OF TEST: 08/09/2025 07:56:51

Diagnostics Check	OK	07:48
Air Blank	0.000	07:49
Control Test	0.079	07:49
Air Blank	0.000	07:50
Subject Sample # 1	0.048	07:51
Breath Volume	1.523	07:51
Air Blank	0.000	07:51
Air Blank	0.000	07:53
Subject Sample # 2	0.044	07:55
Breath Volume	1.101	07:55
Air Blank	0.000	07:55
Control Test	0.078	07:56
Air Blank	0.000	07:56
Diagnostics Check	OK	07:56

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HERRADA, J
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ZORRILLA DIAZ

DATE/TIME OF TEST: 08/10/2025 01:39:03

Diagnostics Check	OK	01:29
Air Blank	0.000	01:29
Control Test	0.079	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.031	01:32
Breath Volume	1.253	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.029	01:37
Breath Volume	1.246	01:37
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:39

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BENAVIDES CUARAN

DATE/TIME OF TEST: 08/10/2025 04:21:39

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:10	
Control Test	0.079	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 1 0.129	04:14	Volume Not Met
Breath Volume	0.316	04:14	
Air Blank	0.000	04:15	
Air Blank	0.000	04:16	
Subject Sample	# 2 0.126	04:19	Volume Not Met
Breath Volume	0.015	04:19	
Air Blank	0.000	04:20	
Control Test	0.076	04:21	
Air Blank	0.000	04:21	
Diagnostics Check	OK	04:21	

Data Download Date/Time: 08/21/2025 13:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BENAVIDES CUARAN

DATE/TIME OF TEST: 08/10/2025 04:34:18

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.076	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.128	04:29
Breath Volume	1.644	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:31
Subject Sample # 2	0.121	04:32
Breath Volume	1.660	04:32
Air Blank	0.000	04:33
Control Test	0.076	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMALDI, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TAYLOR, JAKE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HOIST

DATE/TIME OF TEST: 08/11/2025 01:28:55

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.047	01:24
Breath Volume	2.609	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.043	01:27
Breath Volume	2.351	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VICKERS, S
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 08/14/2025 02:04:34

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:58
Subject Sample # 1	0.245	01:59
Breath Volume	2.234	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.245	02:03
Breath Volume	2.105	02:03
Air Blank	0.000	02:03
Control Test	0.079	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD, A
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: SANTANA WILLIAMS

DATE/TIME OF TEST: 08/14/2025 23:02:57

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.038	22:57
Breath Volume	1.355	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.036	23:01
Breath Volume	1.457	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,AVERY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: CACHAGO-TANDAZO

DATE/TIME OF TEST: 08/16/2025 02:23:44

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.188	02:19
Breath Volume	1.378	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.189	02:22
Breath Volume	1.636	02:22
Air Blank	0.000	02:22
Control Test	0.077	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: QUIJANO,STEVEN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: NIETO LINARES

DATE/TIME OF TEST: 08/16/2025 03:29:44

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.112	03:24
Breath Volume	2.000	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.120	03:28
Breath Volume	1.207	03:28
Air Blank	0.000	03:28
Control Test	0.078	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: RAYO

DATE/TIME OF TEST: 08/17/2025 01:18:31

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.079	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.177	01:13
Breath Volume	1.644	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.181	01:17
Breath Volume	1.851	01:17
Air Blank	0.000	01:17
Control Test	0.077	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MESTRE,EZEQUIEL
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MATA ORTIZ

DATE/TIME OF TEST: 08/17/2025 04:24:01

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.081	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.185	04:19
Breath Volume	1.613	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:21
Subject Sample # 2	0.178	04:22
Breath Volume	1.703	04:22
Air Blank	0.000	04:23
Control Test	0.078	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:24

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VICKERS, S
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: JIRAU KRUMEICH

DATE/TIME OF TEST: 08/19/2025 03:59:39

Diagnostics Check	OK	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.000	03:54
Breath Volume	1.296	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.000	03:58
Breath Volume	1.269	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MATUS

DATE/TIME OF TEST: 08/20/2025 04:04:27

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.265	03:56
Breath Volume	1.246	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.286	03:59
Breath Volume	2.257	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 3	0.278	04:02
Breath Volume	1.867	04:02
Air Blank	0.000	04:03
Control Test	0.077	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/21/2025 13:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,CORINA,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: BARBOZA,GUILLERMO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:54

SUBJECT'S LAST NAME: UTITIAJ

DATE/TIME OF TEST: 08/26/2025 15:26:40

Diagnostics Check	OK	15:18
Air Blank	0.000	15:19
Control Test	0.080	15:19
Air Blank	0.000	15:19
Subject Sample # 1	0.000	15:22
Breath Volume	1.242	15:22
Air Blank	0.000	15:22
Air Blank	0.000	15:24
Subject Sample # 2	0.000	15:25
Breath Volume	2.246	15:25
Air Blank	0.000	15:25
Control Test	0.080	15:26
Air Blank	0.000	15:26
Diagnostics Check	OK	15:26

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ESTRADA,JERIAH
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 08/27/2025 05:49:46

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.079	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.168	05:45
Breath Volume	2.644	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.168	05:48
Breath Volume	2.468	05:48
Air Blank	0.000	05:48
Control Test	0.079	05:49
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: NIEVES, CHARLIE
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: VERA MENDIETA

DATE/TIME OF TEST: 08/30/2025 01:31:15

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.050	01:24
Breath Volume	1.179	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.051	01:29
Breath Volume	1.742	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,GABRIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 08/30/2025 22:25:11

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.180	22:20
Breath Volume	1.519	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.182	22:23
Breath Volume	1.183	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AL-ANBKI,HAEDER,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,AVERY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: POST

DATE/TIME OF TEST: 08/31/2025 00:30:27

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.166	00:25
Breath Volume	3.542	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.180	00:28
Breath Volume	2.308	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAVILANES, MONICA,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GAVILANES, MONICA
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: RAMOS ORTIZ

DATE/TIME OF TEST: 08/31/2025 01:51:33

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.031	01:47
Breath Volume	2.246	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.031	01:50
Breath Volume	2.449	01:50
Air Blank	0.000	01:50
Control Test	0.080	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: NIEVES, CHARLIE
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: METAYER

DATE/TIME OF TEST: 08/31/2025 03:06:14

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.254	03:01
Breath Volume	2.183	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.248	03:04
Breath Volume	1.867	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIEVES, CHARLIE,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: COUCH, BILLON
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 08/31/2025 03:21:14

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.167	03:16
Breath Volume	1.957	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.157	03:19
Breath Volume	1.472	03:19
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: NIEVES, CHARLIE,

BREATH TEST OPERATOR AGENCY: Kissimmee PD

ARRESTING OFFICER: GALEANO, NICHOLAS

ARRESTING OFFICER AGENCY: St. Cloud PD

LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403207507

DRY GAS EXPIRY DATE: 11/26/2027

VIOLATION CODE: DUI

ARREST TIME: 03:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 08/31/2025 04:12:34

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.134	04:08
Breath Volume	2.015	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.135	04:11
Breath Volume	1.710	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,GABRIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ALVAREZ,GUMERSINDO
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: NETTA

DATE/TIME OF TEST: 08/31/2025 20:27:45

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.164	20:23
Breath Volume	2.648	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.160	20:26
Breath Volume	1.769	20:26
Air Blank	0.000	20:26
Control Test	0.080	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARLES,DJENY,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CHARLES,DJENY
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SEPULVEDA

DATE/TIME OF TEST: 08/31/2025 21:48:41

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.143	21:43
Breath Volume	1.937	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.147	21:47
Breath Volume	2.523	21:47
Air Blank	0.000	21:47
Control Test	0.078	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 09/16/2025 12:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BARNES,NOAH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HAMMOCK

DATE/TIME OF TEST: 08/08/2025 21:56:45

Diagnostics Check	OK	21:50
Air Blank	0.000	21:50
Control Test	0.077	21:51
Air Blank	0.000	21:51
Subject Sample # 1	0.159	21:52
Breath Volume	3.406	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 2	0.165	21:55
Breath Volume	2.496	21:55
Air Blank	0.000	21:55
Control Test	0.078	21:56
Air Blank	0.000	21:56
Diagnostics Check	OK	21:56

Data Download Date/Time: 08/14/2025 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIVANS,GARRETH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: GIVANS,GARRETH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: FELSOT

DATE/TIME OF TEST: 08/09/2025 22:03:06

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.078	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.000	21:58
Breath Volume	1.527	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.000	22:01
Breath Volume	1.406	22:01
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 08/14/2025 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 07/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: ST ONGE

DATE/TIME OF TEST: 08/12/2025 22:56:45

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.078	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.182	22:50
Breath Volume	1.300	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.172	22:55
Breath Volume	1.632	22:55
Air Blank	0.000	22:55
Control Test	0.077	22:56
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 08/14/2025 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WERDA,JACOB,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: ROBERTSON,AARAN
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: PEEKS

DATE/TIME OF TEST: 08/22/2025 00:05:02

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.078	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.197	00:00
Breath Volume	1.183	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.202	00:03
Breath Volume	1.785	00:03
Air Blank	0.000	00:04
Control Test	0.078	00:04
Air Blank	0.000	00:04
Diagnostics Check	OK	00:05

Data Download Date/Time: 09/11/2025 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: PRADIPCUMAR

DATE/TIME OF TEST: 08/22/2025 01:03:44

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.077	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.130	00:58
Breath Volume	1.554	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.121	01:02
Breath Volume	1.707	01:02
Air Blank	0.000	01:02
Control Test	0.078	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 09/11/2025 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARIV,ALI,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BARNES,N
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BASALLO

DATE/TIME OF TEST: 08/24/2025 21:13:45

Diagnostics Check	OK	21:07
Air Blank	0.000	21:08
Control Test	0.077	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.239	21:09
Breath Volume	2.253	21:09
Air Blank	0.000	21:09
Air Blank	0.000	21:11
Subject Sample # 2	0.243	21:12
Breath Volume	2.496	21:12
Air Blank	0.000	21:12
Control Test	0.084	21:13
Air Blank	0.000	21:13
Diagnostics Check	OK	21:13

Data Download Date/Time: 09/11/2025 17:09