

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FEITO,EDUARDO,L

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: ADLER,COREY

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402732434

DRY GAS EXPIRY DATE: 04/28/2026

VIOLATION CODE: BUI

ARREST TIME: 21:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 05/07/2025 22:29:00

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.078	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1	22:25	No Sample Provided
Breath Volume	0.000	22:25	
Air Blank	0.000	22:25	
Air Blank	0.000	22:27	
Subject Sample	# 2	22:27	Subject Test Refused
Breath Volume	0.000	22:27	
Air Blank	0.000	22:28	
Control Test	0.079	22:28	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 05/31/2025 01:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FEITO,EDUARDO,L

BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)

ARRESTING OFFICER: ADLER,COREY

ARRESTING OFFICER AGENCY: FFWCC (West Palm)

LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 402732434

DRY GAS EXPIRY DATE: 04/28/2026

VIOLATION CODE: BUI

ARREST TIME: 21:34

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 05/07/2025 22:35:05

Diagnostics Check	OK	22:31	
Air Blank	0.000	22:31	
Control Test	0.080	22:32	
Air Blank	0.000	22:32	
Subject Sample	# 1	22:33	Subject Test Refused
Breath Volume	0.000	22:33	
Air Blank	0.000	22:34	
Control Test	0.078	22:34	
Air Blank	0.000	22:34	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 05/31/2025 01:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,KRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: JORDAN,KRISTOPHER
ARRESTING OFFICER AGENCY: Blountstown PD
LAST AGENCY INSPECTION: 04/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: AGUIRRE ORTEGA

DATE/TIME OF TEST: 05/11/2025 22:07:38

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.082	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.062	22:03
Breath Volume	2.335	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.063	22:06
Breath Volume	1.359	22:06
Air Blank	0.000	22:06
Control Test	0.082	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/12/2025 16:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402956578
DRY GAS EXPIRY DATE: 01/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: TRUAX

DATE/TIME OF TEST: 05/24/2025 23:17:56

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.165	23:13
Breath Volume	1.796	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:15
Subject Sample # 2	0.163	23:16
Breath Volume	2.371	23:16
Air Blank	0.000	23:17
Control Test	0.082	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 06/20/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ,JORGE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: NUNEZ,JORGE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: JAKIE

DATE/TIME OF TEST: 05/02/2025 01:35:19

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.078	01:33	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.078	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 05/22/2025 11:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOSEPH,HERMANN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: SALAS AGUDELO

DATE/TIME OF TEST: 05/03/2025 05:48:55

Diagnostics Check	OK	05:45	
Air Blank	0.000	05:46	
Control Test	0.078	05:46	
Air Blank	0.000	05:46	
Subject Sample	# 1	05:48	Sequence Aborted
Breath Volume	0.000	05:48	
Air Blank	0.000	05:48	

Data Download Date/Time: 05/22/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOSEPH,HERMANN,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: SALAS AGUDELO

DATE/TIME OF TEST: 05/03/2025 05:50:56

Diagnostics Check	OK	05:47	
Air Blank	0.000	05:47	
Control Test	0.078	05:47	
Air Blank	0.000	05:48	
Subject Sample	# 1	05:50	Sequence Aborted
Breath Volume	0.000	05:50	
Air Blank	0.000	05:50	

Data Download Date/Time: 05/22/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 05/05/2025 19:55:44

Diagnostics Check	OK	19:46	
Air Blank	0.000	19:46	
Control Test	0.078	19:46	
Air Blank	0.000	19:47	
Subject Sample	# 1 0.176	19:50	Volume Not Met
Breath Volume	0.000	19:50	
Air Blank	0.000	19:51	
Air Blank	0.000	19:52	
Subject Sample	# 2	19:54	Subject Test Refused
Breath Volume	0.035	19:54	
Air Blank	0.000	19:54	
Control Test	0.075	19:55	
Air Blank	0.000	19:55	
Diagnostics Check	OK	19:55	

Data Download Date/Time: 05/22/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERRERA TERRY, MELANIE, M
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE, RENATO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: PINCHEIRA DONOSO

DATE/TIME OF TEST: 05/11/2025 06:17:14

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.078	06:09
Air Blank	0.000	06:09
Subject Sample # 1	0.030	06:10
Breath Volume	1.441	06:10
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.028	06:15
Breath Volume	1.113	06:15
Air Blank	0.000	06:16
Control Test	0.078	06:16
Air Blank	0.000	06:17
Diagnostics Check	OK	06:17

Data Download Date/Time: 05/22/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ,JORGE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: NUNEZ,JORGE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ANDOR

DATE/TIME OF TEST: 05/16/2025 23:05:13

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:50	
Control Test	0.078	22:51	
Air Blank	0.000	22:51	
Subject Sample	# 1 0.172	22:54	Volume Not Met
Breath Volume	0.085	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 2 0.154	23:00	
Breath Volume	2.152	23:00	
Air Blank	0.000	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 3 0.168	23:03	
Breath Volume	1.875	23:03	
Air Blank	0.000	23:04	
Control Test	0.075	23:04	
Air Blank	0.000	23:05	
Diagnostics Check	OK	23:05	

Data Download Date/Time: 05/22/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STENBACK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: STENBACK,AUSTIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PALUMBO

DATE/TIME OF TEST: 05/18/2025 01:48:16

Diagnostics Check	OK	01:39	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1 0.154	01:43	Volume Not Met
Breath Volume	0.355	01:43	
Air Blank	0.000	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 2	01:46	Subject Test Refused
Breath Volume	0.058	01:46	
Air Blank	0.000	01:47	
Control Test	0.076	01:47	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 05/22/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRIDO,JEYNY,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,R
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: KOKKINOS

DATE/TIME OF TEST: 05/18/2025 03:57:23

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.078	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.113	03:52
Breath Volume	2.445	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.121	03:55
Breath Volume	1.945	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 05/22/2025 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRIDO,JEYNY,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MAURE,R
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: KOKKINOS

DATE/TIME OF TEST: 05/18/2025 04:39:18

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.078	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.104	04:34
Breath Volume	1.839	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.106	04:37
Breath Volume	2.351	04:37
Air Blank	0.000	04:38
Control Test	0.077	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 05/22/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, HECTOR, L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: BLANCO, YASNIER
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/25/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SEVILLA

DATE/TIME OF TEST: 05/18/2025 21:23:45

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.077	21:21	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:22	Subject Test Refused
Breath Volume	0.000	21:22	
Air Blank	0.000	21:22	
Control Test	0.077	21:23	
Air Blank	0.000	21:23	
Diagnostics Check	OK	21:23	

Data Download Date/Time: 05/22/2025 11:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISNEY,WHITNEY,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SISNEY,WHITNEY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 10:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:39

SUBJECT'S LAST NAME: SOCKEL

DATE/TIME OF TEST: 05/28/2025 12:12:59

Diagnostics Check	OK	11:59	
Air Blank	0.000	12:00	
Control Test	0.079	12:00	
Air Blank	0.000	12:01	
Subject Sample	# 1	12:04	No Sample Provided
Breath Volume	0.000	12:04	
Air Blank	0.000	12:04	
Air Blank	0.000	12:06	
Subject Sample	# 2	0.000	12:07
Breath Volume	3.363	12:07	
Air Blank	0.000	12:07	
Air Blank	0.000	12:09	
Subject Sample	# 3	0.000	12:11
Breath Volume	2.648	12:11	
Air Blank	0.000	12:12	
Control Test	0.079	12:12	
Air Blank	0.000	12:12	
Diagnostics Check	OK	12:12	

Data Download Date/Time: 06/12/2025 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRANDON,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MILLER,BRANDON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 05/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402648278
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SVENDBLAD

DATE/TIME OF TEST: 05/31/2025 23:47:58

Diagnostics Check	OK	23:38	
Air Blank	0.000	23:39	
Control Test	0.078	23:39	
Air Blank	0.000	23:40	
Subject Sample	# 1 0.282	23:40	
Breath Volume	1.335	23:40	
Air Blank	0.000	23:41	
Air Blank	0.000	23:43	
Subject Sample	# 2 0.256	23:43	
Breath Volume	1.488	23:43	
Air Blank	0.000	23:44	
Air Blank	0.000	23:46	
Subject Sample	# 3 0.253	23:46	
Breath Volume	1.496	23:46	
Air Blank	0.000	23:47	
Control Test	0.067	23:47	Control Outside Tolerance
Air Blank	0.000	23:47	

Data Download Date/Time: 06/12/2025 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,J,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ZARATE,J
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BELHOMME

DATE/TIME OF TEST: 05/02/2025 01:44:38

Diagnostics Check	OK	01:37
Air Blank	0.000	01:37
Control Test	0.080	01:38
Air Blank	0.000	01:38
Subject Sample # 1	0.159	01:39
Breath Volume	1.628	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.158	01:43
Breath Volume	1.179	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 05/03/2025 22:25:26

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.081	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1 0.153	22:15	Volume Not Met
Breath Volume	0.488	22:15	
Air Blank	0.000	22:16	
Air Blank	0.000	22:17	
Subject Sample	# 2 0.155	22:19	
Breath Volume	1.398	22:19	
Air Blank	0.000	22:20	
Air Blank	0.000	22:22	
Subject Sample	# 3 0.159	22:23	
Breath Volume	1.312	22:23	
Air Blank	0.000	22:24	
Control Test	0.078	22:24	
Air Blank	0.000	22:25	
Diagnostics Check	OK	22:25	

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 05/11/2025 04:10:09

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.081	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 1 0.152	04:08	Volume Not Met
Breath Volume	0.484	04:08	
Air Blank		04:09	Purge Fail
Air Blank	0.000	04:10	

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 05/11/2025 04:23:33

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.169	04:16
Breath Volume	2.082	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.168	04:21
Breath Volume	1.433	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: TAMAYO,KEVIN
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: RIVERA FLORES

DATE/TIME OF TEST: 05/11/2025 15:29:31

Diagnostics Check	OK	15:26	
Air Blank	0.000	15:27	
Control Test	0.079	15:27	
Air Blank	0.000	15:27	
Subject Sample	# 1	15:28	Subject Test Refused
Breath Volume	0.000	15:28	
Air Blank	0.000	15:28	
Control Test	0.081	15:28	
Air Blank	0.000	15:29	
Diagnostics Check	OK	15:29	

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABIB,EMAD,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LABIB,EMAD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ALBELO

DATE/TIME OF TEST: 05/14/2025 18:00:07

Diagnostics Check	OK	17:54
Air Blank	0.000	17:54
Control Test	0.080	17:54
Air Blank	0.000	17:55
Subject Sample # 1	0.000	17:55
Breath Volume	2.503	17:55
Air Blank	0.000	17:56
Air Blank	0.000	17:58
Subject Sample # 2	0.000	17:58
Breath Volume	2.980	17:58
Air Blank	0.000	17:59
Control Test	0.082	17:59
Air Blank	0.000	18:00
Diagnostics Check	OK	18:00

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BAKEWELL,JACHARY
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: ORTIZ ABRAHAM

DATE/TIME OF TEST: 05/14/2025 22:22:02

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.167	22:17
Breath Volume	2.171	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.158	22:20
Breath Volume	1.718	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:22

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,MARIO,
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: RODRIGUEZ,I
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 07:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:27

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 05/18/2025 08:53:54

Diagnostics Check	OK	08:51	
Air Blank	0.000	08:51	
Control Test	0.080	08:51	
Air Blank	0.000	08:52	
Subject Sample	# 1	08:52	Subject Test Refused
Breath Volume	0.000	08:52	
Air Blank	0.000	08:53	
Control Test	0.079	08:53	
Air Blank	0.000	08:53	
Diagnostics Check	OK	08:53	

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PIERRE LOUIS

DATE/TIME OF TEST: 05/24/2025 01:37:27

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.116	01:33
Breath Volume	1.468	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.121	01:35
Breath Volume	2.468	01:35
Air Blank	0.000	01:36
Control Test	0.080	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JONATHAN,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: CASTILLO OLIVAS

DATE/TIME OF TEST: 05/31/2025 04:13:37

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.082	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.165	04:08
Breath Volume	1.542	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.156	04:12
Breath Volume	2.128	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CELINA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,CELINA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/31/2025 05:22:41

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.082	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.118	05:18
Breath Volume	3.250	05:18
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.117	05:21
Breath Volume	3.273	05:21
Air Blank	0.000	05:21
Control Test	0.080	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 06/12/2025 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,CRISTIAN,X
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SANTOS,CRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:01

SUBJECT'S LAST NAME: SERRANO GOMEZ

DATE/TIME OF TEST: 05/03/2025 14:28:50

Diagnostics Check	OK	14:21
Air Blank	0.000	14:22
Control Test	0.081	14:22
Air Blank	0.000	14:23
Subject Sample # 1	0.202	14:24
Breath Volume	1.789	14:24
Air Blank	0.000	14:24
Air Blank	0.000	14:26
Subject Sample # 2	0.199	14:27
Breath Volume	1.398	14:27
Air Blank	0.000	14:27
Control Test	0.080	14:28
Air Blank	0.000	14:28
Diagnostics Check	OK	14:28

Data Download Date/Time: 06/05/2025 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORETON,MYKELL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FUENTES,A
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SENA

DATE/TIME OF TEST: 05/04/2025 02:05:48

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:04	
Control Test	0.081	02:04	
Air Blank		02:05	RFI Detect
Air Blank	0.000	02:05	

Data Download Date/Time: 06/05/2025 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORETON,MYKELL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FUENTES,ANGELA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SENA

DATE/TIME OF TEST: 05/04/2025 02:10:13

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:07	
Control Test	0.081	02:08	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:09	
Control Test	0.079	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 06/05/2025 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCAULEY, DEVERO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ARAUJO, KAYLA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:23

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 05/10/2025 09:53:12

Diagnostics Check	OK	09:50	
Air Blank	0.000	09:50	
Control Test	0.081	09:51	
Air Blank	0.000	09:51	
Subject Sample	# 1	09:51	Subject Test Refused
Breath Volume	0.000	09:51	
Air Blank	0.000	09:52	
Control Test	0.081	09:52	
Air Blank	0.000	09:53	
Diagnostics Check	OK	09:53	

Data Download Date/Time: 06/05/2025 02:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORETON,MYKELL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FUENTES,ANGELA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/18/2025 03:45:52

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.082	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.081	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 06/05/2025 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GARCIA,JONATHAN]
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: RAMOS CRUZ

DATE/TIME OF TEST: 05/25/2025 20:32:42

Diagnostics Check	OK	20:26
Air Blank	0.000	20:26
Control Test	0.082	20:26
Air Blank	0.000	20:27
Subject Sample # 1	0.196	20:28
Breath Volume	3.640	20:28
Air Blank	0.000	20:28
Air Blank	0.000	20:30
Subject Sample # 2	0.189	20:31
Breath Volume	2.007	20:31
Air Blank	0.000	20:31
Control Test	0.080	20:32
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 06/05/2025 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RUIZ,SAMUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 05/04/2025 02:09:34

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.081	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1 0.134	02:08	
Breath Volume	1.539	02:08	
Air Blank		02:08	Purge Fail
Air Blank	0.000	02:09	

Data Download Date/Time: 06/05/2025 02:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEANDRE,GERALD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RUIZ,SAMUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 05/04/2025 02:26:23

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:12	
Subject Sample	# 1	02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:16	
Air Blank	0.000	02:18	
Subject Sample	# 2	0.130	02:20
Breath Volume	1.566	02:20	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 3	0.130	02:24
Breath Volume	1.773	02:24	
Air Blank	0.000	02:25	
Control Test	0.078	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 06/05/2025 02:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GIBSON,NICHOLAS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: TZUL-CHAMORRO

DATE/TIME OF TEST: 05/04/2025 02:45:04

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.082	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.190	02:40
Breath Volume	2.031	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.200	02:43
Breath Volume	1.949	02:43
Air Blank	0.000	02:44
Control Test	0.081	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:45

Data Download Date/Time: 06/05/2025 02:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON III, EMANUEL,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WALTON III, EMANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MANZANARES QUINTANIL

DATE/TIME OF TEST: 05/10/2025 23:07:15

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.082	23:05	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:06	
Control Test	0.082	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 06/05/2025 02:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GAMEZ

DATE/TIME OF TEST: 05/11/2025 01:41:20

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.083	01:39	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.082	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 06/05/2025 02:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORTEGA,WILFREDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:12

SUBJECT'S LAST NAME: ESQUIVEL GARCIA

DATE/TIME OF TEST: 05/11/2025 08:40:35

Diagnostics Check	OK	08:37	
Air Blank	0.000	08:38	
Control Test	0.082	08:38	
Air Blank	0.000	08:38	
Subject Sample	# 1	08:39	Subject Test Refused
Breath Volume	0.000	08:39	
Air Blank	0.000	08:39	
Control Test	0.082	08:39	
Air Blank	0.000	08:40	
Diagnostics Check	OK	08:40	

Data Download Date/Time: 06/05/2025 02:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ORTEGA,WILFREDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 08:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:43

SUBJECT'S LAST NAME: CORDOVA DE LA CRUZ

DATE/TIME OF TEST: 05/17/2025 09:07:36

Diagnostics Check	OK	09:04	
Air Blank	0.000	09:05	
Control Test	0.082	09:05	
Air Blank	0.000	09:05	
Subject Sample	# 1	09:06	Subject Test Refused
Breath Volume	0.000	09:06	
Air Blank	0.000	09:06	
Control Test	0.082	09:07	
Air Blank	0.000	09:07	
Diagnostics Check	OK	09:07	

Data Download Date/Time: 06/05/2025 02:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GATTORNO,ALEXANDER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: CEPERO ABREU

DATE/TIME OF TEST: 05/25/2025 12:54:16

Diagnostics Check	OK	12:48
Air Blank	0.000	12:48
Control Test	0.082	12:49
Air Blank	0.000	12:49
Subject Sample # 1	0.000	12:49
Breath Volume	3.640	12:49
Air Blank	0.000	12:50
Air Blank	0.000	12:52
Subject Sample # 2	0.000	12:52
Breath Volume	3.175	12:52
Air Blank	0.000	12:53
Control Test	0.083	12:53
Air Blank	0.000	12:54
Diagnostics Check	OK	12:54

Data Download Date/Time: 06/05/2025 02:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GODBEE,CYTERIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:19

SUBJECT'S LAST NAME: CAYADO

DATE/TIME OF TEST: 05/26/2025 13:46:34

Diagnostics Check	OK	13:40
Air Blank	0.000	13:40
Control Test	0.082	13:40
Air Blank	0.000	13:41
Subject Sample # 1	0.115	13:42
Breath Volume	3.421	13:42
Air Blank	0.000	13:42
Air Blank	0.000	13:44
Subject Sample # 2	0.120	13:45
Breath Volume	2.847	13:45
Air Blank	0.000	13:45
Control Test	0.083	13:45
Air Blank	0.000	13:46
Diagnostics Check	OK	13:46

Data Download Date/Time: 06/05/2025 02:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAIKEL,H
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RODRIGUEZ,MAIKEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 06/05/2025
VIOLATION CODE: DUI
ARREST TIME: 15:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:07

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/29/2025 17:33:16

Diagnostics Check	OK	17:28	
Air Blank	0.000	17:29	
Control Test	0.081	17:29	
Air Blank	0.000	17:29	
Subject Sample	# 1	17:31	Subject Test Refused
Breath Volume	0.000	17:31	
Air Blank	0.000	17:32	
Control Test	0.081	17:32	
Air Blank	0.000	17:33	
Diagnostics Check	OK	17:33	

Data Download Date/Time: 06/05/2025 02:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: WILLEY

DATE/TIME OF TEST: 05/05/2025 21:45:45

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:43	
Control Test	0.077	21:43	
Air Blank	0.000	21:44	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:44	
Control Test	0.077	21:45	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 05/21/2025 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KILIAN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 05/06/2025 02:09:31

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.077	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.186	02:04
Breath Volume	1.617	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.189	02:07
Breath Volume	1.335	02:07
Air Blank	0.000	02:08
Control Test	0.077	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 05/21/2025 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: WILLOUGHBY

DATE/TIME OF TEST: 05/10/2025 02:02:52

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.076	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:01	
Control Test	0.077	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 05/21/2025 13:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FAILS,CHASE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: CONORD

DATE/TIME OF TEST: 05/10/2025 04:11:51

Diagnostics Check	OK	04:09	
Air Blank	0.000	04:09	
Control Test	0.077	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 1	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Control Test	0.077	04:11	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 05/21/2025 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCHILL,CAMERON,P
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CHURCHILL,CAMERON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG329101
DRY GAS EXPIRY DATE: 10/18/2025
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 05/11/2025 01:58:35

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.076	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 1	01:57	Subject Test Refused
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Control Test	0.077	01:57	
Air Blank	0.000	01:58	
Diagnostics Check	OK	01:58	

Data Download Date/Time: 05/21/2025 13:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: PHILLIPS,CHARLES,B

BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)

ARRESTING OFFICER: PHILLIPS,CHARLES

ARRESTING OFFICER AGENCY: FFWCC (Panama City)

LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403207508

DRY GAS EXPIRY DATE: 06/27/2026

VIOLATION CODE: BUI

ARREST TIME: 19:47

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: TOUCHTON

DATE/TIME OF TEST: 05/08/2025 20:27:56

Diagnostics Check	OK	20:25	
Air Blank	0.000	20:25	
Control Test	0.082	20:25	
Air Blank	0.000	20:26	
Subject Sample	# 1	20:26	Subject Test Refused
Breath Volume	0.000	20:26	
Air Blank	0.000	20:27	
Control Test	0.083	20:27	
Air Blank	0.000	20:27	
Diagnostics Check	OK	20:27	

Data Download Date/Time: 05/29/2025 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PHILLIPS,CHARLES
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: TOUCHTON

DATE/TIME OF TEST: 05/08/2025 20:39:51

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:37	
Control Test	0.082	20:37	
Air Blank	0.000	20:38	
Subject Sample	# 1	20:38	Subject Test Refused
Breath Volume	0.000	20:38	
Air Blank	0.000	20:38	
Control Test	0.082	20:39	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 05/29/2025 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: POWELL,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: TIANO JR

DATE/TIME OF TEST: 05/16/2025 17:00:22

Diagnostics Check	OK	16:54	
Air Blank	0.000	16:55	
Control Test	0.081	16:55	
Air Blank	0.000	16:56	
Subject Sample	# 1	16:59	Subject Test Refused
Breath Volume	0.000	16:59	
Air Blank	0.000	16:59	
Control Test	0.082	16:59	
Air Blank	0.000	17:00	
Diagnostics Check	OK	17:00	

Data Download Date/Time: 05/29/2025 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBLIN,ZACHARY,W
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GAMBLIN,ZACHARY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: CHESTANG

DATE/TIME OF TEST: 05/17/2025 19:18:24

Diagnostics Check	OK	19:15	
Air Blank	0.000	19:15	
Control Test	0.082	19:16	
Air Blank	0.000	19:16	
Subject Sample	# 1	19:17	Subject Test Refused
Breath Volume	0.000	19:17	
Air Blank	0.000	19:17	
Control Test	0.082	19:17	
Air Blank	0.000	19:18	
Diagnostics Check	OK	19:18	

Data Download Date/Time: 05/29/2025 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBIN,ZACHARY,W
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: GAMBLIN,ZACHARY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: CHESTANG

DATE/TIME OF TEST: 05/17/2025 19:28:23

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:25	
Control Test	0.082	19:26	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:27	Subject Test Refused
Breath Volume	0.000	19:27	
Air Blank	0.000	19:27	
Control Test	0.081	19:27	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 05/29/2025 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: STINSON,ERIC
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: ZITO

DATE/TIME OF TEST: 05/24/2025 17:14:08

Diagnostics Check	OK	17:10	
Air Blank	0.000	17:10	
Control Test	0.082	17:11	
Air Blank	0.000	17:11	
Subject Sample	# 1	17:12	Subject Test Refused
Breath Volume	0.000	17:12	
Air Blank	0.000	17:13	
Control Test	0.083	17:13	
Air Blank	0.000	17:14	
Diagnostics Check	OK	17:14	

Data Download Date/Time: 05/29/2025 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 403207508
BREATH TEST OPERATOR: REEVES,JAMES,F	DRY GAS EXPIRY DATE: 06/27/2026
BREATH TEST OPERATOR AGENCY: Okaloosa County SO	VIOLATION CODE: BUI
ARRESTING OFFICER: REEVES,JAMES	ARREST TIME: 22:44
ARRESTING OFFICER AGENCY: Okaloosa County SO	GENDER: Male
LAST AGENCY INSPECTION: 05/29/2025	OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: BROOKS DATE/TIME OF TEST: 05/29/2025 23:25:22

Diagnostics Check	OK	23:22	
Air Blank	0.000	23:22	
Control Test	0.081	23:23	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:23	Subject Test Refused
Breath Volume	0.000	23:23	
Air Blank	0.000	23:24	
Control Test	0.084	23:24	
Air Blank	0.000	23:25	
Diagnostics Check	OK	23:25	

Data Download Date/Time: 06/25/2025 17:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHARD, WILLIAM, J
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BOLEY, CHRISTOPHER
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:14

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 05/24/2025 17:48:25

Diagnostics Check	OK	17:45	
Air Blank	0.000	17:46	
Control Test	0.081	17:46	
Air Blank	0.000	17:46	
Subject Sample	# 1	17:47	Subject Test Refused
Breath Volume	0.000	17:47	
Air Blank	0.000	17:47	
Control Test	0.081	17:47	
Air Blank	0.000	17:48	
Diagnostics Check	OK	17:48	

Data Download Date/Time: 07/01/2025 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 402783528
BREATH TEST OPERATOR: BASFORD,NICOLE,M	DRY GAS EXPIRY DATE: 06/27/2026
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: JARRELL,CHRISTIAN	ARREST TIME: 17:30
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 04/21/2025	OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: BALDWIN	DATE/TIME OF TEST: 05/24/2025 18:02:21
------------------------------	--

Diagnostics Check	OK	17:55
Air Blank	0.000	17:55
Control Test	0.081	17:55
Air Blank	0.000	17:56
Subject Sample # 1	0.155	17:57
Breath Volume	1.808	17:57
Air Blank	0.000	17:58
Air Blank	0.000	17:59
Subject Sample # 2	0.157	18:00
Breath Volume	1.261	18:00
Air Blank	0.000	18:01
Control Test	0.080	18:01
Air Blank	0.000	18:02
Diagnostics Check	OK	18:02

Data Download Date/Time: 07/01/2025 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,NICOLE,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAFRENZ,HANNAH
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: VAN LIEROP

DATE/TIME OF TEST: 05/24/2025 18:25:55

Diagnostics Check	OK	18:22	
Air Blank	0.000	18:23	
Control Test	0.081	18:23	
Air Blank	0.000	18:24	
Subject Sample	# 1	18:24	Subject Test Refused
Breath Volume	0.000	18:24	
Air Blank	0.000	18:25	
Control Test	0.080	18:25	
Air Blank	0.000	18:25	
Diagnostics Check	OK	18:25	

Data Download Date/Time: 07/01/2025 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAGG,DEVIN,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: RAMOSACEVEDO

DATE/TIME OF TEST: 05/25/2025 18:43:49

Diagnostics Check	OK	18:40	
Air Blank	0.000	18:40	
Control Test	0.080	18:40	
Air Blank	0.000	18:41	
Subject Sample	# 1	18:42	Subject Test Refused
Breath Volume	0.000	18:42	
Air Blank	0.000	18:42	
Control Test	0.080	18:43	
Air Blank	0.000	18:43	
Diagnostics Check	OK	18:43	

Data Download Date/Time: 07/01/2025 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAGG,DEVIN,M
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,NICOLE
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: RAMOS ACEVEDO

DATE/TIME OF TEST: 05/25/2025 18:51:27

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:49	
Control Test	0.081	18:49	
Air Blank	0.000	18:49	
Subject Sample	# 1	18:50	Subject Test Refused
Breath Volume	0.000	18:50	
Air Blank	0.000	18:50	
Control Test	0.080	18:50	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 07/01/2025 12:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: AIGNER,BRIAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 05/01/2025 22:03:45

Diagnostics Check	OK	21:56
Air Blank	0.000	21:57
Control Test	0.084	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.213	21:58
Breath Volume	1.132	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.220	22:02
Breath Volume	1.632	22:02
Air Blank	0.000	22:02
Control Test	0.082	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADFORD, HARLEE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: IMBODY, M
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:06

SUBJECT'S LAST NAME: VECCHIO

DATE/TIME OF TEST: 05/02/2025 14:41:03

Diagnostics Check	OK	14:33
Air Blank	0.000	14:33
Control Test	0.083	14:34
Air Blank	0.000	14:34
Subject Sample # 1	0.000	14:35
Breath Volume	2.398	14:35
Air Blank	0.000	14:35
Air Blank	0.000	14:38
Subject Sample # 2	0.000	14:39
Breath Volume	2.234	14:39
Air Blank	0.000	14:40
Control Test	0.082	14:40
Air Blank	0.000	14:40
Diagnostics Check	OK	14:41

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HERNANDEZ RODRIGUEZ

DATE/TIME OF TEST: 05/03/2025 04:08:09

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.082	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.186	04:03
Breath Volume	1.816	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.195	04:06
Breath Volume	1.523	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: JANVIER

DATE/TIME OF TEST: 05/04/2025 00:35:02

Diagnostics Check	OK	00:32	
Air Blank	0.000	00:32	
Control Test	0.083	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.082	00:34	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:35	

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRARO,PAUL,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MALLONE,N
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 05/04/2025 03:38:06

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.084	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.292	03:33
Breath Volume	2.832	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:35
Subject Sample # 2	0.281	03:36
Breath Volume	2.707	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRARO,PAUL,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CAMILLERI,Z
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 05/04/2025 03:53:51

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.061	03:49
Breath Volume	2.386	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.058	03:52
Breath Volume	2.066	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GORNEY,LINDSAY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: SAINT JEAN

DATE/TIME OF TEST: 05/04/2025 04:58:51

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.083	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.204	04:54
Breath Volume	2.042	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.219	04:57
Breath Volume	2.246	04:57
Air Blank	0.000	04:57
Control Test	0.082	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:58

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: RAMJATTAN, KARON
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 05/05/2025 01:52:02

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.083	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.186	01:46
Breath Volume	1.203	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.174	01:50
Breath Volume	1.246	01:50
Air Blank	0.000	01:51
Control Test	0.083	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:52

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRARO,PAUL,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CAMILLERI,Z
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: GARCIA BACALLAO

DATE/TIME OF TEST: 05/05/2025 02:33:38

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.082	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.237	02:28
Breath Volume	1.789	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.244	02:32
Breath Volume	1.609	02:32
Air Blank	0.000	02:32
Control Test	0.082	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BARCO CASTRO

DATE/TIME OF TEST: 05/05/2025 03:29:17

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.082	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.180	03:24
Breath Volume	3.285	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.180	03:27
Breath Volume	3.777	03:27
Air Blank	0.000	03:28
Control Test	0.083	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FANEGO, DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARNIMONT, MASON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: PUGA

DATE/TIME OF TEST: 05/10/2025 05:49:54

Diagnostics Check	OK	05:43
Air Blank	0.000	05:44
Control Test	0.081	05:44
Air Blank	0.000	05:44
Subject Sample # 1	0.178	05:45
Breath Volume	2.406	05:45
Air Blank	0.000	05:46
Air Blank	0.000	05:47
Subject Sample # 2	0.176	05:48
Breath Volume	2.742	05:48
Air Blank	0.000	05:48
Control Test	0.082	05:49
Air Blank	0.000	05:49
Diagnostics Check	OK	05:49

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEACHMAN, MICHAEL, W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PEACHMAN, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: GIL

DATE/TIME OF TEST: 05/11/2025 17:16:01

Diagnostics Check	OK	17:09
Air Blank	0.000	17:10
Control Test	0.082	17:10
Air Blank	0.000	17:11
Subject Sample # 1	0.194	17:11
Breath Volume	1.808	17:11
Air Blank	0.000	17:12
Air Blank	0.000	17:13
Subject Sample # 2	0.175	17:14
Breath Volume	1.960	17:14
Air Blank	0.000	17:15
Control Test	0.082	17:15
Air Blank	0.000	17:15
Diagnostics Check	OK	17:16

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,JACOB,C
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MYNAUGH,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: OWEN

DATE/TIME OF TEST: 05/11/2025 23:04:36

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.081	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.193	23:00
Breath Volume	3.027	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.206	23:03
Breath Volume	2.832	23:03
Air Blank	0.000	23:03
Control Test	0.083	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 05/13/2025 09:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: KING,JONATHAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GOODCHILD

DATE/TIME OF TEST: 05/14/2025 00:58:41

Diagnostics Check	OK	00:52
Air Blank	0.000	00:52
Control Test	0.084	00:53
Air Blank	0.000	00:53
Subject Sample # 1	0.177	00:54
Breath Volume	2.546	00:54
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.177	00:57
Breath Volume	2.246	00:57
Air Blank	0.000	00:57
Control Test	0.083	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 05/15/2025 01:25:12

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.084	01:23	
Air Blank	0.000	01:23	
Subject Sample	# 1	01:24	RFI Detect
Breath Volume	1.003	01:24	
Air Blank		01:24	RFI Detect
Air Blank	0.000	01:25	

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH, TRAEVONNE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 05/15/2025 01:32:58

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.083	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.177	01:28
Breath Volume	1.300	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:30
Subject Sample # 2	0.168	01:31
Breath Volume	1.425	01:31
Air Blank	0.000	01:32
Control Test	0.083	01:32
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA,S,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: GODINEZ,O
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RAMIREZ CHIVALAN

DATE/TIME OF TEST: 05/18/2025 01:04:13

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.083	00:58
Air Blank	0.000	00:59
Subject Sample # 1	0.232	00:59
Breath Volume	2.949	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.227	01:02
Breath Volume	2.675	01:02
Air Blank	0.000	01:03
Control Test	0.083	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 05/18/2025 03:43:50

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.083	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.088	03:38
Breath Volume	1.351	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.083	03:42
Breath Volume	1.289	03:42
Air Blank	0.000	03:42
Control Test	0.084	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: PERERA, BRIAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:44

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 05/21/2025 20:48:07

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.083	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.000	20:43
Breath Volume	2.917	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.000	20:46
Breath Volume	2.832	20:46
Air Blank	0.000	20:47
Control Test	0.084	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRARO,PAUL,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TIRADO,B
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: SALES

DATE/TIME OF TEST: 05/21/2025 23:18:22

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.085	23:11
Air Blank	0.000	23:11
Subject Sample # 1	0.255	23:13
Breath Volume	1.675	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.253	23:16
Breath Volume	1.562	23:16
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MENDEZ,VIOLET
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PEREZ OJEDA

DATE/TIME OF TEST: 05/22/2025 04:00:38

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.084	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.069	03:56
Breath Volume	1.816	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.066	03:59
Breath Volume	1.593	03:59
Air Blank	0.000	03:59
Control Test	0.085	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: ACOSTA,CARLOS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: SCHNELL

DATE/TIME OF TEST: 05/23/2025 00:55:03

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.084	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.000	00:50
Breath Volume	2.003	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:53
Breath Volume	1.382	00:53
Air Blank	0.000	00:54
Control Test	0.084	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:55

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZUNIGA,S,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: AIGNER,B
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CASSIDY

DATE/TIME OF TEST: 05/23/2025 23:21:32

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.084	23:15
Air Blank	0.000	23:16
Subject Sample # 1	0.155	23:17
Breath Volume	2.113	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.159	23:20
Breath Volume	2.437	23:20
Air Blank	0.000	23:20
Control Test	0.084	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FANEGO, DAVID,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MATERA, DEREK
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: KEOUGH

DATE/TIME OF TEST: 05/24/2025 00:56:11

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.084	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.167	00:51
Breath Volume	1.410	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:54
Subject Sample # 2	0.166	00:54
Breath Volume	2.546	00:54
Air Blank	0.000	00:55
Control Test	0.084	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,LINDSAY,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOFF,DAVID
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 05/31/2025 08:43:46

Diagnostics Check	OK	08:34	
Air Blank	0.000	08:34	
Control Test	0.085	08:35	
Air Blank	0.000	08:35	
Subject Sample	# 1 0.177	08:37	
Breath Volume	1.222	08:37	
Air Blank	0.000	08:38	
Air Blank	0.000	08:40	
Subject Sample	# 2 0.166	08:42	
Breath Volume	1.351	08:42	
Air Blank	0.000	08:42	
Control Test	0.000	08:43	RFI Detect
Air Blank	0.000	08:43	

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,LINDSAY,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOFF,DAVID
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 05/31/2025 08:59:39

Diagnostics Check	OK	08:46	
Air Blank	0.000	08:47	
Control Test	0.083	08:47	
Air Blank	0.000	08:47	
Subject Sample	# 1 0.157	08:51	Volume Not Met
Breath Volume	0.730	08:51	
Air Blank	0.000	08:51	
Air Blank	0.000	08:53	
Subject Sample	# 2 0.165	08:54	
Breath Volume	1.390	08:54	
Air Blank	0.000	08:54	
Air Blank	0.000	08:56	
Subject Sample	# 3 0.174	08:58	
Breath Volume	2.355	08:58	
Air Blank	0.000	08:58	
Control Test	0.082	08:59	
Air Blank	0.000	08:59	
Diagnostics Check	OK	08:59	

Data Download Date/Time: 06/17/2025 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: DELVA

DATE/TIME OF TEST: 05/03/2025 01:35:51

Diagnostics Check	OK	01:29
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:30
Subject Sample # 1	0.159	01:31
Breath Volume	2.992	01:31
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.157	01:34
Breath Volume	3.066	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERO,L
ARRESTING OFFICER AGENCY: Miccosukee PD
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: RICARDO

DATE/TIME OF TEST: 05/09/2025 23:39:14

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.080	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.165	23:34
Breath Volume	1.289	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.166	23:37
Breath Volume	1.273	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HERNANDEZ PEREZ

DATE/TIME OF TEST: 05/10/2025 03:07:08

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.247	03:01
Breath Volume	1.269	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.253	03:05
Breath Volume	2.300	03:05
Air Blank	0.000	03:06
Control Test	0.078	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 04/05/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ZERON

DATE/TIME OF TEST: 05/10/2025 03:32:44

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.185	03:27
Breath Volume	1.250	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.181	03:31
Breath Volume	1.234	03:31
Air Blank	0.000	03:31
Control Test	0.078	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 05/10/2025 20:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,HENRY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANDOVAL,LUIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:23

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 05/11/2025 06:55:17

Diagnostics Check	OK	06:48
Air Blank	0.000	06:49
Control Test	0.080	06:49
Air Blank	0.000	06:50
Subject Sample # 1	0.169	06:50
Breath Volume	3.210	06:50
Air Blank	0.000	06:51
Air Blank	0.000	06:52
Subject Sample # 2	0.174	06:53
Breath Volume	3.023	06:53
Air Blank	0.000	06:54
Control Test	0.078	06:54
Air Blank	0.000	06:55
Diagnostics Check	OK	06:55

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/11/2025 22:24:06

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.172	22:18
Breath Volume	1.593	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.173	22:22
Breath Volume	2.140	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: REBOLLEDO

DATE/TIME OF TEST: 05/17/2025 02:27:02

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.124	02:22
Breath Volume	3.070	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.120	02:25
Breath Volume	2.320	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCESCHI,VANESSA,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: BARRERAS,RAFAEL
ARRESTING OFFICER AGENCY: Miami Springs PD
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/17/2025 08:28:17

Diagnostics Check	OK	08:25	
Air Blank	0.000	08:25	
Control Test	0.081	08:26	
Air Blank	0.000	08:26	
Subject Sample	# 1	08:26	Subject Test Refused
Breath Volume	0.000	08:26	
Air Blank	0.000	08:27	
Control Test	0.080	08:27	
Air Blank	0.000	08:28	
Diagnostics Check	OK	08:28	

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 05/18/2025 03:24:27

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.173	03:19
Breath Volume	2.906	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.171	03:22
Breath Volume	1.128	03:22
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LOPEZ DELGADILLO

DATE/TIME OF TEST: 05/18/2025 04:48:56

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.243	04:44
Breath Volume	3.339	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample # 2	0.239	04:47
Breath Volume	2.335	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDOVAL,LUIS,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANDOVAL,LUIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: SANCHEZ TOLOZA

DATE/TIME OF TEST: 05/19/2025 05:10:15

Diagnostics Check	OK	05:04
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Subject Sample # 1	0.216	05:05
Breath Volume	1.324	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:08
Subject Sample # 2	0.219	05:08
Breath Volume	1.558	05:08
Air Blank	0.000	05:09
Control Test	0.078	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VARGAS,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: CADAVIECOSALGADO

DATE/TIME OF TEST: 05/21/2025 05:53:05

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.080	05:43
Air Blank	0.000	05:44
Subject Sample # 1	0.250	05:44
Breath Volume	2.277	05:44
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.225	05:48
Breath Volume	2.238	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:50
Subject Sample # 3	0.246	05:51
Breath Volume	2.523	05:51
Air Blank	0.000	05:52
Control Test	0.079	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:53

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: FERNANDEZ ALVARADO

DATE/TIME OF TEST: 05/23/2025 05:14:57

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.081	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.181	05:10
Breath Volume	3.343	05:10
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.182	05:13
Breath Volume	3.894	05:13
Air Blank	0.000	05:14
Control Test	0.079	05:14
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: NARVAEZ

DATE/TIME OF TEST: 05/24/2025 03:44:42

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.109	03:40
Breath Volume	1.230	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.111	03:43
Breath Volume	2.113	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 05/24/2025 22:46:52

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.081	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.210	22:42
Breath Volume	1.722	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:44
Subject Sample # 2	0.209	22:45
Breath Volume	2.183	22:45
Air Blank	0.000	22:45
Control Test	0.078	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: DIAZASCANIO

DATE/TIME OF TEST: 05/26/2025 01:34:54

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.192	01:27
Breath Volume	1.250	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.226	01:30
Breath Volume	3.039	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 3	0.217	01:33
Breath Volume	1.648	01:33
Air Blank	0.000	01:34
Control Test	0.079	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 40305679
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 12:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DIAZ FONSECA

DATE/TIME OF TEST: 05/29/2025 02:09:04

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.081	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.187	02:04
Breath Volume	1.925	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:06
Subject Sample # 2	0.196	02:07
Breath Volume	1.640	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:09

Data Download Date/Time: 06/03/2025 19:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: SELBY

DATE/TIME OF TEST: 05/01/2025 22:29:20

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.081	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.212	22:24
Breath Volume	1.457	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.206	22:27
Breath Volume	1.617	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: AMELIO

DATE/TIME OF TEST: 05/03/2025 23:49:35

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.082	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.236	23:45
Breath Volume	1.441	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.226	23:48
Breath Volume	1.562	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANKOSKY,D,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LEIVA,J
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 05/06/2025 00:33:30

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1 0.198	00:25	Volume Not Met
Breath Volume	0.406	00:25	
Air Blank	0.000	00:26	
Air Blank	0.000	00:27	
Subject Sample	# 2 0.217	00:28	
Breath Volume	2.207	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 3 0.216	00:31	
Breath Volume	2.023	00:31	
Air Blank	0.000	00:32	
Control Test	0.078	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINI,RYAN,H
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MARTINI,RYAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 13:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: LOOMIS

DATE/TIME OF TEST: 05/08/2025 17:25:10

Diagnostics Check	OK	17:19
Air Blank	0.000	17:19
Control Test	0.082	17:19
Air Blank	0.000	17:20
Subject Sample # 1	0.000	17:20
Breath Volume	1.488	17:20
Air Blank	0.000	17:21
Air Blank	0.000	17:23
Subject Sample # 2	0.000	17:23
Breath Volume	1.453	17:23
Air Blank	0.000	17:24
Control Test	0.081	17:24
Air Blank	0.000	17:25
Diagnostics Check	OK	17:25

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,TASHA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CASSINGHAM

DATE/TIME OF TEST: 05/09/2025 03:00:24

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.082	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.163	02:52
Breath Volume	2.621	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.138	02:55
Breath Volume	1.199	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 3	0.151	02:58
Breath Volume	3.609	02:58
Air Blank	0.000	02:59
Control Test	0.081	02:59
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: DERoy

DATE/TIME OF TEST: 05/09/2025 18:47:04

Diagnostics Check	OK	18:43	
Air Blank	0.000	18:44	
Control Test	0.082	18:44	
Air Blank	0.000	18:44	
Subject Sample	# 1	18:45	Interferent Detect
Breath Volume	1.863	18:45	
Air Blank	0.000	18:46	
Control Test	0.081	18:46	
Air Blank	0.000	18:46	
Diagnostics Check	OK	18:47	

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:22

SUBJECT'S LAST NAME: DERoy

DATE/TIME OF TEST: 05/09/2025 18:51:55

Diagnostics Check	OK	18:48	
Air Blank	0.000	18:49	
Control Test	0.081	18:49	
Air Blank	0.000	18:49	
Subject Sample	# 1	18:50	Interferent Detect
Breath Volume	1.781	18:50	
Air Blank	0.000	18:51	
Control Test	0.081	18:51	
Air Blank	0.000	18:51	
Diagnostics Check	OK	18:51	

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PETERSEN,BRENDEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: ROSSO

DATE/TIME OF TEST: 05/09/2025 23:06:30

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Subject Sample # 1	0.177	23:01
Breath Volume	1.457	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.185	23:04
Breath Volume	1.269	23:04
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: LOESCHNER

DATE/TIME OF TEST: 05/10/2025 01:28:35

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.081	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.212	01:23
Breath Volume	1.406	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.219	01:26
Breath Volume	3.273	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDEMANN, ANTHONY, M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HILL, TIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: POULTON

DATE/TIME OF TEST: 05/10/2025 03:46:25

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.082	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.116	03:41
Breath Volume	2.007	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.117	03:44
Breath Volume	2.152	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 05/13/2025 02:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SERRES,DEVAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: SANCHEZ MORALES

DATE/TIME OF TEST: 05/13/2025 06:13:50

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.082	06:08
Air Blank	0.000	06:08
Subject Sample # 1	0.180	06:09
Breath Volume	1.691	06:09
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample # 2	0.188	06:12
Breath Volume	2.292	06:12
Air Blank	0.000	06:12
Control Test	0.080	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: CHMIELEWSKI

DATE/TIME OF TEST: 05/17/2025 00:53:02

Diagnostics Check	OK	00:46
Air Blank	0.000	00:47
Control Test	0.082	00:47
Air Blank	0.000	00:47
Subject Sample # 1	0.107	00:48
Breath Volume	1.281	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:50
Subject Sample # 2	0.100	00:51
Breath Volume	2.226	00:51
Air Blank	0.000	00:52
Control Test	0.081	00:52
Air Blank	0.000	00:52
Diagnostics Check	OK	00:53

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,TASHA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DEL TORO ABAD

DATE/TIME OF TEST: 05/17/2025 03:36:43

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.216	03:32
Breath Volume	2.867	03:32
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.207	03:35
Breath Volume	2.070	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: VAZQUEZ,HUMBERTO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: LOPEZ GARCIA

DATE/TIME OF TEST: 05/18/2025 01:08:20

Diagnostics Check	OK	00:58
Air Blank	0.000	00:59
Control Test	0.082	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.188	01:01
Breath Volume	1.613	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:03
Subject Sample # 2	0.180	01:06
Breath Volume	1.265	01:06
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FRAZIN,SHAWN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BURNETTE

DATE/TIME OF TEST: 05/18/2025 02:00:21

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.000	01:55
Breath Volume	2.371	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.000	01:58
Breath Volume	3.160	01:58
Air Blank	0.000	01:59
Control Test	0.082	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CLAUSNITZER,CHAD
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: SICKLER

DATE/TIME OF TEST: 05/19/2025 00:14:33

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.082	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.085	00:10
Breath Volume	2.113	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.087	00:12
Breath Volume	3.292	00:12
Air Blank	0.000	00:13
Control Test	0.082	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ,FRANCISCO,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: PETERSEN,BRENDEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 05/21/2025 03:58:32

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.115	03:54
Breath Volume	4.429	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.113	03:57
Breath Volume	3.613	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENUP, RONNY,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GREENUP, RONNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: FAUST

DATE/TIME OF TEST: 05/23/2025 21:47:18

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.082	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.225	21:42
Breath Volume	2.566	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.236	21:45
Breath Volume	2.886	21:45
Air Blank	0.000	21:46
Control Test	0.081	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDEMANN, ANTHONY, M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CRUZ-MERCHAN, WINSTON
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: URDANETA GONZALEZ

DATE/TIME OF TEST: 05/25/2025 00:17:48

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.082	00:10
Air Blank	0.000	00:11
Subject Sample # 1	0.146	00:13
Breath Volume	2.859	00:13
Air Blank	0.000	00:13
Air Blank	0.000	00:15
Subject Sample # 2	0.155	00:16
Breath Volume	2.859	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HERMOSO BRITO

DATE/TIME OF TEST: 05/25/2025 02:55:44

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.081	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.222	02:50
Breath Volume	3.328	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.219	02:54
Breath Volume	2.777	02:54
Air Blank	0.000	02:54
Control Test	0.081	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDEMANN, ANTHONY, M
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LINDEMANN, ANTHONY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: JACOB

DATE/TIME OF TEST: 05/26/2025 07:00:07

Diagnostics Check	OK	06:52
Air Blank	0.000	06:53
Control Test	0.082	06:53
Air Blank	0.000	06:53
Subject Sample # 1	0.039	06:55
Breath Volume	1.511	06:55
Air Blank	0.000	06:56
Air Blank	0.000	06:57
Subject Sample # 2	0.035	06:58
Breath Volume	1.613	06:58
Air Blank	0.000	06:59
Control Test	0.081	06:59
Air Blank	0.000	07:00
Diagnostics Check	OK	07:00

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,T,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,T
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: SECCHIAROLI

DATE/TIME OF TEST: 05/26/2025 21:12:22

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.082	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.258	21:07
Breath Volume	2.390	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.267	21:10
Breath Volume	2.359	21:10
Air Blank	0.000	21:11
Control Test	0.081	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,TASHA,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ALLEN,TASHA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: PARENTI

DATE/TIME OF TEST: 05/31/2025 18:57:35

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.082	18:51
Air Blank	0.000	18:52
Subject Sample # 1	0.233	18:52
Breath Volume	2.503	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.242	18:55
Breath Volume	3.035	18:55
Air Blank	0.000	18:56
Control Test	0.082	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 06/06/2025 02:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MYERS,JUSTIN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 05/31/2025 19:28:03

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.080	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.172	19:23
Breath Volume	2.332	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:25
Subject Sample # 2	0.175	19:26
Breath Volume	1.203	19:26
Air Blank	0.000	19:27
Control Test	0.080	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:28

Data Download Date/Time: 06/06/2025 02:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: STRYBEL,BRIAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 05/31/2025 22:04:47

Diagnostics Check	OK	21:58
Air Blank	0.000	21:58
Control Test	0.082	21:58
Air Blank	0.000	21:59
Subject Sample # 1	0.138	22:00
Breath Volume	3.519	22:00
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.148	22:03
Breath Volume	3.464	22:03
Air Blank	0.000	22:03
Control Test	0.082	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 06/06/2025 02:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,C,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PHILLIPS,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: DAVILALEON

DATE/TIME OF TEST: 05/18/2025 03:46:13

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.084	03:41
Breath Volume	2.996	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.073	03:44
Breath Volume	1.160	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 06/28/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,JUAN,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: LOPEZ LUIZ

DATE/TIME OF TEST: 05/22/2025 22:29:03

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.162	22:24
Breath Volume	2.824	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.161	22:27
Breath Volume	2.765	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:29

Data Download Date/Time: 06/28/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIS,MEGAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,TINA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 05/26/2025 02:22:08

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.199	02:16
Breath Volume	1.605	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.207	02:20
Breath Volume	2.230	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 06/28/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATES,SHEBRETA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GULOTTA,JUSTIN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:36

SUBJECT'S LAST NAME: OZUNA

DATE/TIME OF TEST: 05/26/2025 17:11:54

Diagnostics Check	OK	17:05
Air Blank	0.000	17:06
Control Test	0.079	17:06
Air Blank	0.000	17:06
Subject Sample # 1	0.200	17:07
Breath Volume	2.839	17:07
Air Blank	0.000	17:08
Air Blank	0.000	17:09
Subject Sample # 2	0.201	17:10
Breath Volume	2.636	17:10
Air Blank	0.000	17:10
Control Test	0.078	17:11
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

Data Download Date/Time: 06/28/2025 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,CHRIS,W
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: BARTON,LORI
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: KARSHE

DATE/TIME OF TEST: 05/04/2025 09:34:13

Diagnostics Check	OK	09:31	
Air Blank	0.000	09:31	
Control Test	0.077	09:32	
Air Blank	0.000	09:32	
Subject Sample	# 1	09:32	Subject Test Refused
Breath Volume	0.000	09:32	
Air Blank	0.000	09:33	
Control Test	0.078	09:33	
Air Blank	0.000	09:34	
Diagnostics Check	OK	09:34	

Data Download Date/Time: 06/30/2025 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CISNEROS,JACOB,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVAREZ,A
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: ADMIN
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 05/05/2025 04:32:09

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.076	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.000	04:28
Breath Volume	2.476	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.000	04:30
Breath Volume	2.218	04:30
Air Blank	0.000	04:31
Control Test	0.077	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 06/30/2025 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVAREZ,ALEXANDER
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1524080A1
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: RUIZ AMBROSIO

DATE/TIME OF TEST: 05/16/2025 01:59:35

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.077	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.153	01:54
Breath Volume	2.203	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.146	01:57
Breath Volume	2.343	01:57
Air Blank	0.000	01:58
Control Test	0.077	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 06/30/2025 14:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,ROLF,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: GORDON,ROLF
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/05/2025 20:33:04

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:30	
Control Test	0.081	20:31	
Air Blank	0.000	20:31	
Subject Sample	# 1	20:31	Subject Test Refused
Breath Volume	0.000	20:31	
Air Blank	0.000	20:32	
Control Test	0.081	20:32	
Air Blank	0.000	20:32	
Diagnostics Check	OK	20:33	

Data Download Date/Time: 05/28/2025 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,JERALD,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SCARABIN,JIM
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 05/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A3
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: ROBISON

DATE/TIME OF TEST: 05/29/2025 02:06:40

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.000	02:02
Breath Volume	2.457	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.000	02:05
Breath Volume	1.500	02:05
Air Blank	0.000	02:05
Control Test	0.079	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 06/24/2025 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FENTON, ANTHONY, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LANCASTER, CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14323080A2
DRY GAS EXPIRY DATE: 09/05/2025
VIOLATION CODE: DUI
ARREST TIME: 11:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ALMONTE

DATE/TIME OF TEST: 05/26/2025 01:29:09

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.080	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.198	01:24
Breath Volume	4.234	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.181	01:27
Breath Volume	2.277	01:27
Air Blank	0.000	01:28
Control Test	0.078	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 06/26/2025 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERSON,CAMERON,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HOPKINS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 05/18/2025 22:48:10

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.081	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.000	22:43
Breath Volume	2.167	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.000	22:46
Breath Volume	1.382	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 06/10/2025 15:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: KALLENBORN

DATE/TIME OF TEST: 05/04/2025 03:14:41

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.081	03:09	
Air Blank	0.000	03:10	
Subject Sample # 1	0.211	03:10	
Breath Volume	2.500	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:13	
Subject Sample # 2	0.210	03:13	
Breath Volume	2.421	03:13	
Air Blank		03:14	RFI Detect
Air Blank	0.000	03:14	

Data Download Date/Time: 05/13/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: KELLENBORN

DATE/TIME OF TEST: 05/04/2025 03:22:56

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.080	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.208	03:18
Breath Volume	2.253	03:18
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.207	03:21
Breath Volume	2.402	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 05/13/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: GOUIN, M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ELLIOT

DATE/TIME OF TEST: 05/04/2025 04:03:20

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.177	03:58
Breath Volume	1.425	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.171	04:01
Breath Volume	1.402	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 05/13/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON, GERALD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRISON, GERALD
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: JHA

DATE/TIME OF TEST: 05/04/2025 22:49:50

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1 0.210	22:43	Volume Not Met
Breath Volume	0.753	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 2	22:48	Subject Test Refused
Breath Volume	0.546	22:48	
Air Blank	0.000	22:48	
Control Test	0.078	22:49	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 05/13/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JONATHAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 05/05/2025 19:56:52

Diagnostics Check	OK	19:54	
Air Blank	0.000	19:55	
Control Test	0.080	19:55	
Air Blank		19:56	Ambient Fail
Air Blank	0.000	19:56	

Data Download Date/Time: 05/13/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OWENS,JONATHAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08223080A2
DRY GAS EXPIRY DATE: 05/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 05/05/2025 20:05:49

Diagnostics Check	OK	19:59
Air Blank	0.000	19:59
Control Test	0.081	19:59
Air Blank	0.000	20:00
Subject Sample # 1	0.213	20:01
Breath Volume	2.484	20:01
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.197	20:04
Breath Volume	2.019	20:04
Air Blank	0.000	20:04
Control Test	0.081	20:05
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 05/13/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: RILEY, T
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 05/10/2025 04:15:31

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.152	04:11
Breath Volume	1.765	04:11
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.154	04:14
Breath Volume	1.148	04:14
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 05/13/2025 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WILLIAMS,DEVONTE
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 05/14/2025 23:48:54

Diagnostics Check	OK	23:47	
Air Blank	0.000	23:48	
Control Test	0.000	23:48	Control Outside Tolerance
Air Blank	0.000	23:48	

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WILLIAMS,DEVONTE
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 05/14/2025 23:57:28

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.078	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.401	23:52
Breath Volume	1.332	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:54
Subject Sample # 2	0.399	23:55
Breath Volume	1.351	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,THOMAS,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/17/2025 04:26:47

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.082	04:13	
Air Blank	0.000	04:13	
Subject Sample	# 1 0.108	04:14	
Breath Volume	3.140	04:14	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 2	04:19	No Sample Provided
Breath Volume	0.000	04:19	
Air Blank	0.000	04:20	
Air Blank	0.000	04:22	
Subject Sample	# 3	04:25	No Sample Provided
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Control Test	0.082	04:26	
Air Blank	0.000	04:26	
Diagnostics Check	OK	04:26	

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,THOMAS,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/17/2025 04:48:52

Diagnostics Check	OK	04:42
Air Blank	0.000	04:43
Control Test	0.082	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.098	04:44
Breath Volume	3.847	04:44
Air Blank	0.000	04:45
Air Blank	0.000	04:46
Subject Sample # 2	0.093	04:47
Breath Volume	3.242	04:47
Air Blank	0.000	04:47
Control Test	0.081	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HALL,NOAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HAND

DATE/TIME OF TEST: 05/19/2025 00:07:10

Diagnostics Check	OK	00:06	
Air Blank		00:06	RFI Detect
Air Blank	0.000	00:07	

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HALL,NOAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HAND

DATE/TIME OF TEST: 05/19/2025 00:18:20

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.081	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.300	00:10
Breath Volume	2.722	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.278	00:13
Breath Volume	2.410	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 3	0.298	00:16
Breath Volume	2.281	00:16
Air Blank	0.000	00:17
Control Test	0.081	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,THOMAS,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: YOUNG,THOMAS
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/20/2025 05:08:30

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.200	05:03
Breath Volume	4.039	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.202	05:06
Breath Volume	3.613	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIDDLE,MICAH,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LANGLEY,CODY
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/26/2025 02:39:10

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.155	02:34
Breath Volume	1.136	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.156	02:37
Breath Volume	1.468	02:37
Air Blank	0.000	02:38
Control Test	0.081	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 05/30/2025 01:08:08

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.081	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1 0.145	00:58	Volume Not Met
Breath Volume	0.304	00:58	
Air Blank	0.000	00:58	
Air Blank	0.000	01:00	
Subject Sample	# 2 0.141	01:02	
Breath Volume	1.136	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 3	01:06	Subject Test Refused
Breath Volume	0.000	01:06	
Air Blank	0.000	01:07	
Control Test	0.079	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 05/30/2025 01:19:46

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:18	RFI Detect
Breath Volume	1.000	01:18	
Air Blank	0.000	01:18	
Control Test	0.081	01:19	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIGANCE, MICHAEL, G
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: NELSON, COLIN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 05/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 05/30/2025 01:32:15

Diagnostics Check	OK	01:21
Air Blank	0.000	01:22
Control Test	0.080	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.137	01:24
Breath Volume	1.386	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.136	01:30
Breath Volume	1.433	01:30
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 06/10/2025 11:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,NATHAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SABA,DUSTIN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 05/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: IRVIN

DATE/TIME OF TEST: 05/31/2025 23:48:02

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.030	23:43
Breath Volume	1.308	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.029	23:46
Breath Volume	1.519	23:46
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:48

Data Download Date/Time: 06/24/2025 12:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS, KANARVIS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: TAYLOR IV

DATE/TIME OF TEST: 05/06/2025 02:17:58

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.082	02:07	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:11	No Sample Provided
Breath Volume	0.000	02:11	
Air Blank	0.000	02:11	
Air Blank	0.000	02:13	
Subject Sample	# 2	02:16	No Sample Provided
Breath Volume	0.000	02:16	
Air Blank	0.000	02:17	
Control Test	0.082	02:17	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 05/14/2025 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORDOVA,ODELIA
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: FAJARDO

DATE/TIME OF TEST: 05/06/2025 03:46:07

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.082	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.195	03:41
Breath Volume	1.175	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.185	03:44
Breath Volume	1.835	03:44
Air Blank	0.000	03:45
Control Test	0.082	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 05/14/2025 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MEROLA,GEORGE
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28423080A4
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 05/06/2025 05:32:14

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.082	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.141	05:27
Breath Volume	1.988	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.137	05:30
Breath Volume	2.933	05:30
Air Blank	0.000	05:31
Control Test	0.081	05:31
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 05/14/2025 10:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALLE,OMAR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: KHAN

DATE/TIME OF TEST: 05/01/2025 01:57:12

Diagnostics Check	OK	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.156	01:51
Breath Volume	1.941	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.154	01:55
Breath Volume	2.371	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILSON,IAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: BIGELOW

DATE/TIME OF TEST: 05/01/2025 02:33:27

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.000	02:29
Breath Volume	1.101	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.000	02:32
Breath Volume	1.773	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONTE,COREY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: SPIVY

DATE/TIME OF TEST: 05/01/2025 21:51:57

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:38	
Control Test	0.079	21:39	
Air Blank	0.000	21:39	
Subject Sample	# 1 0.245	21:42	Volume Not Met
Breath Volume	0.929	21:42	
Air Blank	0.000	21:43	
Air Blank	0.000	21:45	
Subject Sample	# 2 0.248	21:47	
Breath Volume	1.312	21:47	
Air Blank	0.000	21:48	
Air Blank	0.000	21:50	
Subject Sample	# 3 0.268	21:50	
Breath Volume	1.832	21:50	
Air Blank	0.000	21:51	
Control Test	0.078	21:51	
Air Blank	0.000	21:51	
Diagnostics Check	OK	21:51	

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/02/2025 01:59:07

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.239	01:51
Breath Volume	1.320	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.241	01:57
Breath Volume	1.691	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: AL-NASER

DATE/TIME OF TEST: 05/02/2025 02:28:20

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:24	
Control Test	0.079	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1	02:27	Subject Test Refused
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.080	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALDES,BRYAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: VELASCOVAZQUEZ

DATE/TIME OF TEST: 05/02/2025 03:35:36

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.224	03:31
Breath Volume	2.230	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.216	03:34
Breath Volume	2.351	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 05/14/2025 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO,SERGIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: RAMADHAR

DATE/TIME OF TEST: 05/03/2025 02:51:17

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.124	02:46
Breath Volume	1.414	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.124	02:49
Breath Volume	1.500	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRUITT,JAMES
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: MOYER

DATE/TIME OF TEST: 05/03/2025 15:45:56

Diagnostics Check	OK	15:39
Air Blank	0.000	15:40
Control Test	0.080	15:40
Air Blank	0.000	15:41
Subject Sample # 1	0.166	15:41
Breath Volume	3.828	15:41
Air Blank	0.000	15:42
Air Blank	0.000	15:44
Subject Sample # 2	0.168	15:44
Breath Volume	3.921	15:44
Air Blank	0.000	15:45
Control Test	0.079	15:45
Air Blank	0.000	15:45
Diagnostics Check	OK	15:45

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALFRED,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: OHLHAUSEN

DATE/TIME OF TEST: 05/04/2025 05:56:58

Diagnostics Check	OK	05:50
Air Blank	0.000	05:51
Control Test	0.080	05:51
Air Blank	0.000	05:52
Subject Sample # 1	0.160	05:52
Breath Volume	1.359	05:52
Air Blank	0.000	05:53
Air Blank	0.000	05:55
Subject Sample # 2	0.161	05:55
Breath Volume	2.042	05:55
Air Blank	0.000	05:56
Control Test	0.080	05:56
Air Blank	0.000	05:56
Diagnostics Check	OK	05:56

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JABLONSKI,ADAM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: ROBLERODIAZ

DATE/TIME OF TEST: 05/05/2025 00:08:10

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.078	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.267	00:03
Breath Volume	3.121	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.272	00:06
Breath Volume	2.335	00:06
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/05/2025 04:19:42

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.197	04:15
Breath Volume	1.289	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.196	04:18
Breath Volume	1.625	04:18
Air Blank	0.000	04:18
Control Test	0.079	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO, WILFRED
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 05/05/2025 10:53:34

Diagnostics Check	OK	10:47
Air Blank	0.000	10:48
Control Test	0.080	10:48
Air Blank	0.000	10:48
Subject Sample # 1	0.204	10:49
Breath Volume	2.812	10:49
Air Blank	0.000	10:49
Air Blank	0.000	10:51
Subject Sample # 2	0.205	10:52
Breath Volume	1.300	10:52
Air Blank	0.000	10:52
Control Test	0.079	10:52
Air Blank	0.000	10:53
Diagnostics Check	OK	10:53

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PUTMAN,FORREST
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: MATTAR

DATE/TIME OF TEST: 05/05/2025 21:14:16

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.079	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.000	21:10
Breath Volume	3.472	21:10
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.000	21:12
Breath Volume	3.410	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: GOMEZHERNANDEZ

DATE/TIME OF TEST: 05/05/2025 21:59:17

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.205	21:55
Breath Volume	2.359	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.198	21:57
Breath Volume	2.105	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO,SERGIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: LOPEZMORALES

DATE/TIME OF TEST: 05/06/2025 00:29:44

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.080	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.185	00:25
Breath Volume	2.675	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.189	00:28
Breath Volume	2.488	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 05/06/2025 00:54:27

Diagnostics Check	OK	00:48
Air Blank	0.000	00:49
Control Test	0.080	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.077	00:50
Breath Volume	1.765	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.076	00:53
Breath Volume	1.523	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COOKE,SPENCER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: KUDRIN

DATE/TIME OF TEST: 05/06/2025 05:21:47

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.169	05:17
Breath Volume	3.343	05:17
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.168	05:20
Breath Volume	3.507	05:20
Air Blank	0.000	05:20
Control Test	0.079	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEBSTER,BRAD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/06/2025 23:13:56

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.079	23:09	
Air Blank	0.000	23:09	
Subject Sample	# 1	23:12	Subject Test Refused
Breath Volume	0.000	23:12	
Air Blank	0.000	23:13	
Control Test	0.080	23:13	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LESSMAN,DAVID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: RANGEL

DATE/TIME OF TEST: 05/07/2025 03:52:58

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.018	03:48
Breath Volume	2.187	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.017	03:51
Breath Volume	1.757	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: PEDRAZA

DATE/TIME OF TEST: 05/07/2025 23:52:24

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.080	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.017	23:48
Breath Volume	2.855	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.016	23:50
Breath Volume	3.109	23:50
Air Blank	0.000	23:51
Control Test	0.081	23:51
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWLING,PETER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: CLAUDE

DATE/TIME OF TEST: 05/09/2025 00:45:50

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.080	00:34	
Air Blank	0.000	00:34	
Subject Sample	# 1 0.167	00:38	Volume Not Met
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Air Blank	0.000	00:40	
Subject Sample	# 2 0.224	00:41	
Breath Volume	1.582	00:41	
Air Blank	0.000	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 3 0.229	00:44	
Breath Volume	1.484	00:44	
Air Blank	0.000	00:44	
Control Test	0.078	00:45	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BHATIA,AJAY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: GOLDEN

DATE/TIME OF TEST: 05/09/2025 05:14:34

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.115	05:10
Breath Volume	1.761	05:10
Air Blank	0.000	05:10
Air Blank	0.000	05:12
Subject Sample # 2	0.116	05:13
Breath Volume	1.316	05:13
Air Blank	0.000	05:13
Control Test	0.079	05:13
Air Blank	0.000	05:14
Diagnostics Check	OK	05:14

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,ERNESTO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DIAZORELLANA

DATE/TIME OF TEST: 05/09/2025 22:52:59

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.169	22:48
Breath Volume	1.382	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.182	22:51
Breath Volume	2.011	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEBSTER,BRAD
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: SHEATZLEY

DATE/TIME OF TEST: 05/10/2025 00:46:07

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:39	No Sample Provided
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 2	00:44	No Sample Provided
Breath Volume	0.000	00:44	
Air Blank	0.000	00:45	
Control Test	0.080	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: FLORENCE III

DATE/TIME OF TEST: 05/10/2025 01:22:16

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.056	01:17
Breath Volume	1.328	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.054	01:20
Breath Volume	1.726	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: AZ BAEZA

DATE/TIME OF TEST: 05/10/2025 02:30:54

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.144	02:26
Breath Volume	1.386	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.143	02:29
Breath Volume	2.691	02:29
Air Blank	0.000	02:30
Control Test	0.080	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SELLS

DATE/TIME OF TEST: 05/10/2025 03:19:56

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.224	03:12
Breath Volume	2.027	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.203	03:15
Breath Volume	1.632	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 3	0.212	03:18
Breath Volume	2.492	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARCIA,ROGERS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MENDEZ ZAMORA

DATE/TIME OF TEST: 05/10/2025 04:02:36

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.156	03:58
Breath Volume	2.351	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 2	0.159	04:01
Breath Volume	2.507	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TRENDAFILOV,NICHOLAS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: CHALUISAN

DATE/TIME OF TEST: 05/10/2025 14:13:45

Diagnostics Check	OK	14:07
Air Blank	0.000	14:08
Control Test	0.080	14:08
Air Blank	0.000	14:09
Subject Sample # 1	0.141	14:09
Breath Volume	2.863	14:09
Air Blank	0.000	14:10
Air Blank	0.000	14:11
Subject Sample # 2	0.141	14:12
Breath Volume	1.652	14:12
Air Blank	0.000	14:12
Control Test	0.080	14:13
Air Blank	0.000	14:13
Diagnostics Check	OK	14:13

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BISPO,STEPHANIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: KOTULA

DATE/TIME OF TEST: 05/10/2025 22:02:08

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.080	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.242	21:57
Breath Volume	3.464	21:57
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.252	22:00
Breath Volume	3.343	22:00
Air Blank	0.000	22:01
Control Test	0.079	22:01
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,EDWIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: THIGPEN

DATE/TIME OF TEST: 05/11/2025 01:09:16

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.158	01:04
Breath Volume	2.253	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.149	01:07
Breath Volume	1.164	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/11/2025 02:18:30

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.045	02:14
Breath Volume	1.972	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.043	02:17
Breath Volume	3.550	02:17
Air Blank	0.000	02:17
Control Test	0.080	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELTORO,MICHAEL
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BONILLACARDONA

DATE/TIME OF TEST: 05/11/2025 03:35:01

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.100	03:30
Breath Volume	1.152	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.092	03:33
Breath Volume	1.109	03:33
Air Blank	0.000	03:34
Control Test	0.081	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:35

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: OLIVERIA ALFONZO

DATE/TIME OF TEST: 05/11/2025 03:56:27

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.088	03:50
Breath Volume	1.277	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.081	03:55
Breath Volume	1.464	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARCIA,ROGERS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: CORREA JR

DATE/TIME OF TEST: 05/11/2025 04:11:37

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.120	04:07
Breath Volume	1.476	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.121	04:10
Breath Volume	1.437	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PATTERSON,JACAARA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: DELMEDICO

DATE/TIME OF TEST: 05/11/2025 04:25:50

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.101	04:21
Breath Volume	1.210	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.101	04:24
Breath Volume	1.886	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 05/14/2025 13:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARDY,BRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BANNIS

DATE/TIME OF TEST: 05/12/2025 05:27:04

Diagnostics Check	OK	05:19
Air Blank	0.000	05:20
Control Test	0.080	05:20
Air Blank	0.000	05:21
Subject Sample # 1	0.099	05:22
Breath Volume	2.031	05:22
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.106	05:25
Breath Volume	3.070	05:25
Air Blank	0.000	05:26
Control Test	0.081	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:27

Data Download Date/Time: 05/14/2025 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EDOUARD,WENDY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: COUTU

DATE/TIME OF TEST: 05/13/2025 00:50:14

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.080	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.000	00:46
Breath Volume	2.121	00:46
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.000	00:48
Breath Volume	1.746	00:48
Air Blank	0.000	00:49
Control Test	0.081	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 05/14/2025 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MANUCHA,SAMIRA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 05/13/2025 01:46:47

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.313	01:42
Breath Volume	1.535	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.333	01:45
Breath Volume	2.218	01:45
Air Blank	0.000	01:45
Control Test	0.078	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 05/14/2025 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUFKIN,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 05/14/2025 22:52:29

Diagnostics Check	OK	22:44
Air Blank	0.000	22:45
Control Test	0.080	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.214	22:47
Breath Volume	1.195	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.202	22:51
Breath Volume	1.390	22:51
Air Blank	0.000	22:51
Control Test	0.079	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVA,LUIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: AGUILAROROZCO

DATE/TIME OF TEST: 05/18/2025 02:55:36

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.083	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.263	02:51
Breath Volume	2.117	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.248	02:54
Breath Volume	1.851	02:54
Air Blank	0.000	02:54
Control Test	0.077	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORENO,PAYTON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GARCIAPOLANCO

DATE/TIME OF TEST: 05/18/2025 04:04:03

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.150	03:58
Breath Volume	3.976	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.139	04:02
Breath Volume	2.000	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:04

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CRESPO,BRYAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: GARCIA JR

DATE/TIME OF TEST: 05/19/2025 01:58:04

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:52
Air Blank	0.000	01:53
Subject Sample # 1	0.231	01:53
Breath Volume	3.117	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.232	01:56
Breath Volume	2.468	01:56
Air Blank	0.000	01:57
Control Test	0.080	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:58

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATT,ANTONIO
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MINCY

DATE/TIME OF TEST: 05/21/2025 03:29:26

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.080	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.134	03:24
Breath Volume	1.847	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.131	03:27
Breath Volume	1.781	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: DE LA CRUZ LIND

DATE/TIME OF TEST: 05/23/2025 03:02:17

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:55	No Sample Provided
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample	# 2	03:00	No Sample Provided
Breath Volume	0.000	03:00	
Air Blank	0.000	03:01	
Control Test	0.080	03:01	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPENCER,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/24/2025 02:42:00

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.120	02:36
Breath Volume	2.035	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.119	02:40
Breath Volume	3.109	02:40
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:42

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VALLE,OMAR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: SHABBIR

DATE/TIME OF TEST: 05/24/2025 03:34:16

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.172	03:29
Breath Volume	1.675	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.185	03:32
Breath Volume	1.339	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orlando PD
ARRESTING OFFICER: LUFKIN,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 05/24/2025 04:23:48

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:12	
Control Test	0.079	04:12	
Air Blank	0.000	04:12	
Subject Sample	# 1 0.142	04:16	Volume Not Met
Breath Volume	0.343	04:16	
Air Blank	0.000	04:16	
Air Blank	0.000	04:18	
Subject Sample	# 2 0.150	04:19	
Breath Volume	1.128	04:19	
Air Blank	0.000	04:19	
Air Blank	0.000	04:21	
Subject Sample	# 3 0.164	04:22	
Breath Volume	1.285	04:22	
Air Blank	0.000	04:22	
Control Test	0.080	04:23	
Air Blank	0.000	04:23	
Diagnostics Check	OK	04:23	

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIERRA,EDGAR
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GRIFFIS

DATE/TIME OF TEST: 05/24/2025 04:40:53

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.080	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.120	04:36
Breath Volume	1.292	04:36
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.115	04:39
Breath Volume	1.457	04:39
Air Blank	0.000	04:39
Control Test	0.080	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEN,KEVIN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MORAESQUIRINODESILVA

DATE/TIME OF TEST: 05/26/2025 02:38:13

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.078	02:22	
Air Blank	0.000	02:23	
Subject Sample	# 1 0.150	02:26	Volume Not Met
Breath Volume	0.082	02:26	
Air Blank	0.000	02:27	
Air Blank	0.000	02:28	
Subject Sample	# 2 0.156	02:31	No .020 agreement
Breath Volume	1.285	02:31	
Air Blank	0.000	02:32	
Air Blank	0.000	02:34	
Subject Sample	# 3 0.135	02:36	No .020 agreement
Breath Volume	1.183	02:36	
Air Blank	0.000	02:37	
Control Test	0.079	02:37	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUIS,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEN,KEVIN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MORAESQUIRINODASILVA

DATE/TIME OF TEST: 05/26/2025 02:55:42

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.077	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.158	02:45
Breath Volume	1.355	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.136	02:51
Breath Volume	1.253	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 3	0.157	02:54
Breath Volume	1.406	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/26/2025 03:56:02

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.205	03:51
Breath Volume	1.859	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.200	03:54
Breath Volume	2.031	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:56

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CREWS,CURT
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 05/26/2025 05:26:16

Diagnostics Check	OK	05:19
Air Blank	0.000	05:20
Control Test	0.081	05:20
Air Blank	0.000	05:20
Subject Sample # 1	0.211	05:21
Breath Volume	1.132	05:21
Air Blank	0.000	05:22
Air Blank	0.000	05:24
Subject Sample # 2	0.214	05:24
Breath Volume	1.328	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWLING,PETER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: STELLA DIAZ

DATE/TIME OF TEST: 05/27/2025 22:51:44

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.000	22:47
Breath Volume	3.253	22:47
Air Blank	0.000	22:47
Air Blank	0.000	22:49
Subject Sample # 2	0.000	22:50
Breath Volume	3.757	22:50
Air Blank	0.000	22:50
Control Test	0.080	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORENO,PAYTON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: PERDUE

DATE/TIME OF TEST: 05/28/2025 03:51:19

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.014	03:47
Breath Volume	2.070	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.014	03:49
Breath Volume	1.855	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIELS,DAMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 05/30/2025 02:57:22

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.204	02:53
Breath Volume	2.234	02:53
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.195	02:55
Breath Volume	1.863	02:55
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 06/25/2025 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BENAFIELD,RYAN,E

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: BENAFIELD,RYAN

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 10124080A3

DRY GAS EXPIRY DATE: 06/05/2026

VIOLATION CODE: DUI

ARREST TIME: 02:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/04/2025 03:03:05

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.065	02:58
Breath Volume	2.644	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.065	03:01
Breath Volume	2.765	03:01
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:03

Data Download Date/Time: 06/05/2025 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMILLIAN,BRANDON,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MCMILLIAN,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/12/2025 01:00:43

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.080	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.198	00:56
Breath Volume	3.828	00:56
Air Blank	0.000	00:57
Air Blank	0.000	00:58
Subject Sample # 2	0.189	00:59
Breath Volume	2.878	00:59
Air Blank	0.000	00:59
Control Test	0.079	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 06/05/2025 10:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARZYGnat,DESTINY,P
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCCARLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: MARMANN

DATE/TIME OF TEST: 05/26/2025 08:37:30

Diagnostics Check	OK	08:30
Air Blank	0.000	08:31
Control Test	0.082	08:31
Air Blank	0.000	08:31
Subject Sample # 1	0.092	08:32
Breath Volume	2.378	08:32
Air Blank	0.000	08:33
Air Blank	0.000	08:35
Subject Sample # 2	0.087	08:35
Breath Volume	2.726	08:35
Air Blank	0.000	08:36
Control Test	0.081	08:36
Air Blank	0.000	08:37
Diagnostics Check	OK	08:37

Data Download Date/Time: 06/23/2025 13:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZIZ,D,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: AZIZ,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/03/2025 18:46:49

Diagnostics Check	OK	18:40
Air Blank	0.000	18:41
Control Test	0.078	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.138	18:42
Breath Volume	2.628	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:44
Subject Sample # 2	0.128	18:45
Breath Volume	2.972	18:45
Air Blank	0.000	18:45
Control Test	0.076	18:46
Air Blank	0.000	18:46
Diagnostics Check	OK	18:46

Data Download Date/Time: 05/16/2025 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY, GRANT, D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY, GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DIAZ DOMINGUEZ

DATE/TIME OF TEST: 05/04/2025 03:01:07

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.078	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.150	02:56
Breath Volume	1.382	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.148	02:59
Breath Volume	3.324	02:59
Air Blank	0.000	03:00
Control Test	0.077	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 05/16/2025 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,THOMAS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FRAKING,JAMES
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: PENNY

DATE/TIME OF TEST: 05/11/2025 15:52:46

Diagnostics Check	OK	15:46	
Air Blank	0.000	15:46	
Control Test	0.076	15:46	
Air Blank	0.000	15:47	
Subject Sample	# 1 0.244	15:48	
Breath Volume	1.371	15:48	
Air Blank	0.000	15:48	
Air Blank	0.000	15:50	
Subject Sample	# 2 0.236	15:51	
Breath Volume	1.285	15:51	
Air Blank	0.000	15:51	
Control Test	0.074	15:52	Control Outside Tolerance
Air Blank	0.000	15:52	

Data Download Date/Time: 05/16/2025 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAKE,P,A
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: FREDRICKSON,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: SALCEDO

DATE/TIME OF TEST: 05/12/2025 22:13:43

Diagnostics Check	OK	22:07
Air Blank	0.000	22:07
Control Test	0.076	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.124	22:09
Breath Volume	1.347	22:09
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.112	22:12
Breath Volume	3.425	22:12
Air Blank	0.000	22:12
Control Test	0.077	22:13
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 05/16/2025 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SZWAREK

DATE/TIME OF TEST: 05/13/2025 23:16:14

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.077	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.184	23:10
Breath Volume	1.195	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.179	23:14
Breath Volume	1.812	23:14
Air Blank	0.000	23:15
Control Test	0.075	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 05/16/2025 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: COTTER

DATE/TIME OF TEST: 05/15/2025 22:34:53

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.077	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.265	22:29
Breath Volume	1.878	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.260	22:33
Breath Volume	1.679	22:33
Air Blank	0.000	22:33
Control Test	0.077	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 05/16/2025 13:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: NEWKIRK

DATE/TIME OF TEST: 05/22/2025 00:18:34

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.077	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.000	00:13
Breath Volume	1.437	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.000	00:17
Breath Volume	2.539	00:17
Air Blank	0.000	00:17
Control Test	0.078	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SYMONS,DEPUTY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 05/23/2025 01:17:38

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.077	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.302	01:12
Breath Volume	3.242	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.300	01:16
Breath Volume	3.515	01:16
Air Blank	0.000	01:16
Control Test	0.076	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IVEY,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 05/23/2025 03:12:58

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.077	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.202	03:08
Breath Volume	3.683	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.190	03:11
Breath Volume	2.406	03:11
Air Blank	0.000	03:12
Control Test	0.075	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ-GUILLEN,MARCOS,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LOPEZ-GUILLEN,MARCOS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: MANZOLATI

DATE/TIME OF TEST: 05/27/2025 19:19:47

Diagnostics Check	OK	19:12
Air Blank	0.000	19:12
Control Test	0.077	19:12
Air Blank	0.000	19:13
Subject Sample # 1	0.126	19:14
Breath Volume	1.218	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample # 2	0.111	19:18
Breath Volume	1.500	19:18
Air Blank	0.000	19:18
Control Test	0.076	19:19
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,B,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GERARDI,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: ZARRO

DATE/TIME OF TEST: 05/30/2025 23:43:38

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.076	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.201	23:38
Breath Volume	1.972	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.192	23:41
Breath Volume	1.136	23:41
Air Blank	0.000	23:42
Control Test	0.077	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 06/29/2025 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTEGA,MANUEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 05/15/2025 22:10:44

Diagnostics Check	OK	22:01	
Air Blank	0.000	22:02	
Control Test	0.079	22:02	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:06	No Sample Provided
Breath Volume	0.000	22:06	
Air Blank	0.000	22:06	
Air Blank	0.000	22:08	
Subject Sample	# 2	22:10	Sequence Aborted
Breath Volume	0.000	22:10	
Air Blank	0.000	22:10	

Data Download Date/Time: 06/23/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,DERICK,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MCDONALD,AVERY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ALERS

DATE/TIME OF TEST: 05/16/2025 02:59:59

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.081	02:55
Breath Volume	2.277	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:57
Subject Sample # 2	0.076	02:58
Breath Volume	1.945	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 06/23/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: WOLFE,K
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: CRUZ FONSECA

DATE/TIME OF TEST: 05/18/2025 03:58:13

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.323	03:52
Breath Volume	1.242	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.335	03:56
Breath Volume	1.203	03:56
Air Blank	0.000	03:57
Control Test	0.076	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 06/23/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ALVES-FERRER, C
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: TAYLOR JR

DATE/TIME OF TEST: 05/20/2025 19:32:54

Diagnostics Check	OK	19:29	
Air Blank	0.000	19:29	
Control Test	0.080	19:30	
Air Blank	0.000	19:30	
Subject Sample	# 1	19:31	Subject Test Refused
Breath Volume	0.000	19:31	
Air Blank	0.000	19:31	
Control Test	0.079	19:32	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:32	

Data Download Date/Time: 06/23/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLEGAS,BRIAN,A
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: GALARZ, TIFFANY
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: COLON ALVAREZ

DATE/TIME OF TEST: 05/21/2025 21:10:14

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.080	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.267	21:04
Breath Volume	1.277	21:04
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.256	21:08
Breath Volume	1.812	21:08
Air Blank	0.000	21:09
Control Test	0.080	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 06/23/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLER,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMACHO,FRANCES
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 05/24/2025 02:27:50

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.192	02:23
Breath Volume	2.726	02:23
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.196	02:26
Breath Volume	3.210	02:26
Air Blank	0.000	02:26
Control Test	0.079	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 06/23/2025 13:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEARSON, K
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: GUZMAN RODRIGUEZ

DATE/TIME OF TEST: 05/25/2025 00:58:40

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.080	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.182	00:53
Breath Volume	2.207	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:56
Subject Sample # 2	0.178	00:57
Breath Volume	1.781	00:57
Air Blank	0.000	00:57
Control Test	0.079	00:58
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MONTALVO, J
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 05/25/2025 02:50:15

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.186	02:45
Breath Volume	1.738	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.184	02:48
Breath Volume	1.144	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AL-ANBKI,HAEDER,N
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMACHO,FRANCES
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BRIZUELA MORENO

DATE/TIME OF TEST: 05/30/2025 03:53:38

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.061	03:49
Breath Volume	1.875	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.065	03:52
Breath Volume	2.644	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BILLITTIER,JOSEPH
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: VAZQUEZ PEREZ

DATE/TIME OF TEST: 05/31/2025 04:13:55

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1 0.149	04:05	Volume Not Met
Breath Volume	0.046	04:05	
Air Blank	0.000	04:06	
Air Blank	0.000	04:08	
Subject Sample	# 2 0.126	04:08	
Breath Volume	1.726	04:08	
Air Blank	0.000	04:09	
Air Blank	0.000	04:10	
Subject Sample	# 3 0.139	04:12	
Breath Volume	1.738	04:12	
Air Blank	0.000	04:13	
Control Test	0.077	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:13	

Data Download Date/Time: 06/23/2025 13:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MOTILALSINGH

DATE/TIME OF TEST: 05/01/2025 23:56:35

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:43	
Air Blank	0.000	23:44	
Subject Sample	# 1 0.109	23:47	Volume Not Met
Breath Volume	0.179	23:47	
Air Blank	0.000	23:48	
Air Blank	0.000	23:49	
Subject Sample	# 2 0.106	23:51	
Breath Volume	1.332	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:54	
Subject Sample	# 3 0.104	23:55	
Breath Volume	1.292	23:55	
Air Blank	0.000	23:55	
Control Test	0.077	23:56	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: FAHSI

DATE/TIME OF TEST: 05/02/2025 02:39:01

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.035	02:34
Breath Volume	2.867	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.034	02:37
Breath Volume	2.988	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:39

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 04/08/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: FERGUSON

DATE/TIME OF TEST: 05/05/2025 23:39:14

Diagnostics Check	OK	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.129	23:34
Breath Volume	1.203	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 2	0.137	23:37
Breath Volume	1.460	23:37
Air Blank	0.000	23:38
Control Test	0.079	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 05/12/2025 17:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIRANDA,JAVIER,A
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: MIRANDA,JAVIER
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: BROWNING

DATE/TIME OF TEST: 05/24/2025 00:23:09

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.079	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1 0.158	00:10	
Breath Volume	1.191	00:10	
Air Blank	0.000	00:11	
Air Blank	0.000	00:13	
Subject Sample	# 2 0.150	00:16	Volume Not Met
Breath Volume	0.816	00:16	
Air Blank	0.000	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 3 0.166	00:21	
Breath Volume	2.035	00:21	
Air Blank	0.000	00:22	
Control Test	0.077	00:22	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 06/13/2025 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WERDA,JACOB,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: WERDA,JACOB
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 05/24/2025 22:09:40

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.077	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.231	22:04
Breath Volume	1.542	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:06
Subject Sample # 2	0.238	22:08
Breath Volume	1.488	22:08
Air Blank	0.000	22:08
Control Test	0.078	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 06/13/2025 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONNES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: VALDES GARCIA

DATE/TIME OF TEST: 05/26/2025 01:32:28

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.244	01:27
Breath Volume	1.882	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.232	01:31
Breath Volume	3.746	01:31
Air Blank	0.000	01:31
Control Test	0.078	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 06/13/2025 17:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARIV,ALI,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: BARNES,NOAH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/27/2025 03:14:59

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.077	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.200	03:10
Breath Volume	1.828	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.207	03:13
Breath Volume	1.394	03:13
Air Blank	0.000	03:14
Control Test	0.076	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 06/13/2025 17:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WERDA,JACOB,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: EASTMAN,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 05/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/29/2027
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 05/31/2025 03:32:49

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.133	03:28
Breath Volume	2.757	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.130	03:31
Breath Volume	3.269	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 06/13/2025 17:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: NIGH,ALEXANDER
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 05/03/2025 23:36:51

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.082	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.014	23:32
Breath Volume	1.367	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.015	23:35
Breath Volume	1.246	23:35
Air Blank	0.000	23:35
Control Test	0.082	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 05/28/2025 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/10/2025 20:46:32

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.082	20:40	
Air Blank	0.000	20:40	
Subject Sample # 1	0.000	20:41	
Breath Volume	2.988	20:41	
Air Blank	0.000	20:42	
Air Blank	0.000	20:44	
Subject Sample # 2	0.000	20:44	
Breath Volume	3.148	20:44	
Air Blank	0.000	20:45	
Control Test	0.081	20:45	
Air Blank	0.000	20:45	
Diagnostics Check	Fail	20:46	Temperature Regulation Test
Air Blank	0.000	20:46	

Data Download Date/Time: 05/28/2025 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DASZUTA,DYLAN
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: TRIBUE

DATE/TIME OF TEST: 05/12/2025 21:05:56

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.081	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.144	21:01
Breath Volume	2.406	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.139	21:04
Breath Volume	2.324	21:04
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEDERMAN,CASEY,L
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: BAKER,DALTON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: MASTROIANNI

DATE/TIME OF TEST: 05/17/2025 02:44:00

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.233	02:39
Breath Volume	2.359	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.229	02:42
Breath Volume	1.945	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KELLY, KAITLYN, V

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: BELFORD, DALEN

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 231592

DRY GAS EXPIRY DATE: 11/05/2025

VIOLATION CODE: DUI

ARREST TIME: 12:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: MAZUR

DATE/TIME OF TEST: 05/18/2025 13:26:58

Diagnostics Check	OK	13:20
Air Blank	0.000	13:20
Control Test	0.081	13:21
Air Blank	0.000	13:21
Subject Sample # 1	0.123	13:22
Breath Volume	2.371	13:22
Air Blank	0.000	13:22
Air Blank	0.000	13:24
Subject Sample # 2	0.122	13:25
Breath Volume	2.691	13:25
Air Blank	0.000	13:26
Control Test	0.080	13:26
Air Blank	0.000	13:26
Diagnostics Check	OK	13:26

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DOMINGUEZ,ANDREW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: DEFREITAS

DATE/TIME OF TEST: 05/20/2025 23:41:34

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.082	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.192	23:29
Breath Volume	1.847	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.169	23:34
Breath Volume	1.617	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:36
Subject Sample # 3	0.169	23:39
Breath Volume	1.300	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: JABLONSKY,NICOLAS
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: BRUCE

DATE/TIME OF TEST: 05/23/2025 20:40:25

Diagnostics Check	OK	20:34
Air Blank	0.000	20:34
Control Test	0.081	20:34
Air Blank	0.000	20:35
Subject Sample # 1	0.000	20:35
Breath Volume	2.175	20:35
Air Blank	0.000	20:36
Air Blank	0.000	20:38
Subject Sample # 2	0.000	20:39
Breath Volume	2.195	20:39
Air Blank	0.000	20:39
Control Test	0.081	20:39
Air Blank	0.000	20:40
Diagnostics Check	OK	20:40

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: SWINTON

DATE/TIME OF TEST: 05/23/2025 23:11:41

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.081	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.178	23:06
Breath Volume	3.390	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:08
Subject Sample # 2	0.185	23:10
Breath Volume	3.039	23:10
Air Blank	0.000	23:10
Control Test	0.081	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: CARLTON,DONOVON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: PETTIT

DATE/TIME OF TEST: 05/24/2025 02:59:42

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.081	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1 0.104	02:54	
Breath Volume	1.816	02:54	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample	# 2 0.104	02:57	
Breath Volume	1.687	02:57	
Air Blank	0.000	02:58	
Control Test	0.080	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	Fail	02:59	Temperature Regulation Test
Air Blank	0.000	02:59	

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 05/24/2025 03:43:57

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.080	03:29	
Air Blank	0.000	03:30	
Subject Sample # 1	0.043	03:33	
Breath Volume	1.886	03:33	
Air Blank	0.000	03:33	
Air Blank	0.000	03:35	
Subject Sample # 2	0.039	03:38	Volume Not Met
Breath Volume	0.015	03:38	
Air Blank	0.000	03:39	
Air Blank	0.000	03:41	
Subject Sample # 3	0.039	03:42	
Breath Volume	1.765	03:42	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BAKER,DALTON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 05/25/2025 00:59:21

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.081	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.157	00:54
Breath Volume	3.800	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.155	00:57
Breath Volume	3.472	00:57
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SUPPA

DATE/TIME OF TEST: 05/25/2025 05:05:36

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.081	04:57	
Air Blank	0.000	04:57	
Subject Sample	# 1 0.185	04:59	
Breath Volume	1.972	04:59	
Air Blank	0.000	05:00	
Air Blank	0.000	05:02	
Subject Sample	# 2 0.195	05:03	
Breath Volume	2.222	05:03	
Air Blank	0.000	05:04	
Control Test	0.080	05:04	
Air Blank	0.000	05:05	
Diagnostics Check	Fail	05:05	Temperature Regulation Test
Air Blank	0.000	05:05	

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLARD,FILIFE,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: POLLARD,FILIFE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: LOUDERMILK

DATE/TIME OF TEST: 05/25/2025 19:08:12

Diagnostics Check	OK	19:01	
Air Blank	0.000	19:01	
Control Test	0.081	19:02	
Air Blank	0.000	19:02	
Subject Sample	# 1 0.241	19:03	
Breath Volume	2.355	19:03	
Air Blank	0.000	19:03	
Air Blank	0.000	19:05	
Subject Sample	# 2 0.246	19:06	
Breath Volume	1.421	19:06	
Air Blank	0.000	19:06	
Control Test	0.080	19:07	
Air Blank	0.000	19:07	
Diagnostics Check	Fail	19:07	Temperature Regulation Test
Air Blank	0.000	19:08	

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: RASMUSSEN,MATTHEW
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: SHORTT

DATE/TIME OF TEST: 05/25/2025 20:57:54

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:45	
Control Test	0.081	20:45	
Air Blank	0.000	20:46	
Subject Sample	# 1 0.115	20:49	Slope Not Level
Breath Volume	1.335	20:49	
Air Blank	0.000	20:50	
Air Blank	0.000	20:51	
Subject Sample	# 2 0.119	20:52	
Breath Volume	1.601	20:52	
Air Blank	0.000	20:53	
Air Blank	0.000	20:55	
Subject Sample	# 3 0.125	20:56	
Breath Volume	1.285	20:56	
Air Blank	0.000	20:56	
Control Test	0.081	20:57	
Air Blank	0.000	20:57	
Diagnostics Check	OK	20:57	

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: CARLTON,DONOVON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: JARED

DATE/TIME OF TEST: 05/26/2025 00:38:08

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.100	00:32
Breath Volume	1.378	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.104	00:36
Breath Volume	1.195	00:36
Air Blank	0.000	00:37
Control Test	0.080	00:37
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 05/28/2025 20:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RASMUSSEN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: DUMMER,LUKE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 04/30/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 231592
DRY GAS EXPIRY DATE: 11/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/26/2025 00:53:54

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.080	00:46	
Air Blank	0.000	00:46	
Subject Sample	# 1 0.070	00:48	
Breath Volume	2.371	00:48	
Air Blank	0.000	00:49	
Air Blank	0.000	00:51	
Subject Sample	# 2 0.060	00:51	
Breath Volume	1.519	00:51	
Air Blank	0.000	00:52	
Control Test	0.080	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	Fail	00:53	Temperature Regulation Test
Air Blank	0.000	00:53	

Data Download Date/Time: 05/28/2025 20:08