# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO **Date** 01/12/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 16:51:30 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/12/2026 16:51

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO **Date** 02/04/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 22:05:21 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/04/2026 22:08

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO **Date** 03/05/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 19:08:29 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/05/2026 19:52

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO **Date** 05/04/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 22:40:33 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail LET ROOM AIR OUT AND RETESTED.

**Data Download Date/Time** 05/04/2026 22:41

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO **Date** 06/12/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 01:09:13 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/12/2026 01:06

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO **Date** 07/14/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 02:40:30 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/14/2026 02:39

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO **Date** 08/25/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 03:20:55 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/25/2026 03:18

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO **Date** 09/14/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 02:20:46 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31425080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 01/05/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/14/2026 02:16

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MARTIN COUNTY SO | **Date** | 02/11/2026 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 11:07:04 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 03/06/2026 13:41

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

**Agency** MARTIN COUNTY SO **Date** 04/07/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 00:37:09 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/07/2026 00:37

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

**Agency** MARTIN COUNTY SO **Date** 05/04/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 22:53:36 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/04/2026 22:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

**Agency** MARTIN COUNTY SO **Date** 06/12/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 01:19:38 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/12/2026 01:16

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

**Agency** MARTIN COUNTY SO **Date** 07/14/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 02:18:14 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/14/2026 02:15

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

**Agency** MARTIN COUNTY SO **Date** 08/24/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 23:11:37 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/24/2026 23:12

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

**Agency** MARTIN COUNTY SO **Date** 09/13/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 23:02:05 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31425080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 01/05/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/13/2026 23:02

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

**Agency** MARTIN COUNTY SO **Date** 07/13/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 23:46:16 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/13/2026 23:45

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

**Agency** MARTIN COUNTY SO **Date** 08/24/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 23:51:55 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31124080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/24/2026 23:48

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

**Agency** MARTIN COUNTY SO **Date** 09/13/2026 **Type** Agency Inspection
**Inspector** THOMAS,CHRISTOPHER,L **Time** 23:38:54 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 31425080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 01/05/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/13/2026 23:39

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD **Date** 01/22/2026 **Type** Agency Inspection
**Inspector** ACKERMAN,ELVIS,A **Time** 19:11:39 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402806340
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 01/22/2026 19:11

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD **Date** 02/16/2026 **Type** Agency Inspection
**Inspector** ACKERMAN,ELVIS,A **Time** 18:11:26 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402806340
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2026 18:10

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD **Date** 03/23/2026 **Type** Agency Inspection
**Inspector** ACKERMAN,ELVIS,A **Time** 19:56:11 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402806340
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided.

**Data Download Date/Time** 03/23/2026 19:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD **Date** 05/01/2026 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 00:36:42 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402806340
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail.MOUTH ALCOHOL IN THE AIR

**Data Download Date/Time** 05/01/2026 00:34

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD **Date** 06/30/2026 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 17:50:25 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403008479
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 03/28/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 06/30/2026 17:45

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD **Date** 07/31/2026 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 22:34:37 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403008479
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 03/28/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 07/31/2026 22:29

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD **Date** 08/29/2026 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 18:44:39 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403008479
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 03/28/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 08/29/2026 18:38

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

**Agency** HOMESTEAD PD **Date** 09/26/2026 **Type** Agency Inspection
**Inspector** PEREZ,RAUL, **Time** 15:23:51 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403008479
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 03/28/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 09/26/2026 15:20

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

**Agency** OCALA PD **Date** 03/04/2026 **Type** Agency Inspection
**Inspector** JOEDICKE,WILLIAM,R **Time** 00:09:47 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | INT |
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.075 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Interferent Detect HOOKED UP WRONG SIM.

**Data Download Date/Time** 03/04/2026 19:17

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

**Agency** OCALA PD **Date** 04/23/2026 **Type** Agency Inspection
**Inspector** JOEDICKE,WILLIAM,R **Time** 00:27:02 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.000 |
| 0.05 g/210L Test # (Repeat) 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.000 |
| 0.05 g/210L Test # (Repeat) 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.000 |
| 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance - HOOKED UP WROMG SIM.

**Data Download Date/Time** 04/23/2026 17:11

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

**Agency** OCALA PD **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** JOEDICKE,WILLIAM,R **Time** 21:59:04 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: Ambient Fail CLEARED AREA OF MOUTH ALCOHOL BOTTLE.

**Data Download Date/Time** 06/10/2026 17:25

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

**Agency** OCALA PD **Date** 06/05/2026 **Type** Agency Inspection
**Inspector** JOEDICKE,WILLIAM,R **Time** 15:55:10 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 31124080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NONE

**Data Download Date/Time** 06/10/2026 17:25

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006221

**Agency** FHP TROOP B **Date** 03/25/2026 **Type** Agency Inspection
**Inspector** MURPHY,JAVON,D **Time** 01:05:44 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402909409
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/29/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: NOT ENOUGH MA

**Data Download Date/Time** 03/25/2026 01:17

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006221

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP B | **Date** | 07/31/2026 | **Type** | Agency Inspection |
| **Inspector** | DAVIS,KEVIN,M | **Time** | 11:33:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 403602530 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 04/01/2029 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/31/2026 12:48

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006221

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP TROOP B | **Date** | 09/14/2026 | **Type** | Agency Inspection |
| **Inspector** | DAVIS,KEVIN,M | **Time** | 02:02:38 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 403602530 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 04/01/2029 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/14/2026 14:23

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

**Agency** LAKE COUNTY SO **Date** 06/29/2026 **Type** Agency Inspection
**Inspector** BOWDEN,MATTHEW,C **Time** 10:32:58 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 26-4437
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 01/29/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/29/2026 10:33

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 08/25/2026 | **Type** | Agency Inspection |
| **Inspector** | BOWDEN,MATTHEW,C | **Time** | 10:45:04 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 26-4944 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 05/12/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.079 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.199 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.191 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 08/25/2026 10:41

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | LAKE COUNTY SO | **Date** | 09/21/2026 | **Type** | Agency Inspection |
| **Inspector** | BOWDEN,MATTHEW,C | **Time** | 09:46:00 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 26-4944 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 05/12/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/21/2026 09:43

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BREVARD COUNTY SO | **Date** | 02/11/2026 | **Type** | Agency Inspection |
| **Inspector** | HIGGINBOTHAM,LEANDRA, | **Time** | 10:56:22 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

**Data Download Date/Time** 02/12/2026 17:04

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO **Date** 03/25/2026 **Type** Agency Inspection
**Inspector** WAGNER,ROBERT,F **Time** 11:37:52 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 25925080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2025 **Expiration** 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/25/2026 11:37

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO **Date** 04/27/2026 **Type** Agency Inspection
**Inspector** LEME,SANTIAGO, **Time** 09:26:24 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 25925080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/27/2026 09:25

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO **Date** 05/17/2026 **Type** Agency Inspection
**Inspector** LEME,SANTIAGO, **Time** 11:02:52 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 25925080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/17/2026 10:59

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO **Date** 06/28/2026 **Type** Agency Inspection
**Inspector** LEME,SANTIAGO, **Time** 15:19:37 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 25925080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.044 |
| 0.05 g/210L Test # (Repeat) 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.05 g/210L Test # (Repeat) 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 06/28/2026 15:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO
**Inspector** WAGNER,ROBERT,F
**Date** 07/27/2026
**Time** 12:40:17
**Type** Agency Inspection
**Software** 8100.27

**0.05 Lot #** 202506E
**Expiration** 06/24/2027
**0.08 Lot #**
**Expiration**
**0.20 Lot #**
**Expiration**
**0.08 Gas Lot #**
**Expiration**

**Number of Simulators** 5
**Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | No |
| 0.08 g/210L Test: OK | No |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.043 |
| 0.05 g/210L Test # (Repeat) 1 | 0.029 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.044 |
| 0.05 g/210L Test # (Repeat) 2 | 0.034 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.05 g/210L Test # (Repeat) 3 | 0.037 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

COT - COLD SIM, Control Outside Tolerance. COT .05 SIM.

**Data Download Date/Time** 07/27/2026 12:38

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO **Date** 07/27/2026 **Type** Agency Inspection
**Inspector** LEME,SANTIAGO, **Time** 15:03:31 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #**
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.040 |
| 0.05 g/210L Test # (Repeat) 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.135 |
| 0.20 g/210L Test # (Repeat) 1 | 0.181 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.042 |
| 0.05 g/210L Test # (Repeat) 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.151 |
| 0.20 g/210L Test # (Repeat) 2 | 0.188 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.043 |
| 0.05 g/210L Test # (Repeat) 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.159 |
| 0.20 g/210L Test # (Repeat) 3 | 0.192 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05:COT COLD SIM. 20: COT-OPEN SIM, ALCOHOL ESCAPED 2ND;COLD SIM . COT.

**Data Download Date/Time** 07/27/2026 16:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO
**Inspector** WAGNER,ROBERT,F
**Date** 07/27/2026
**Time** 15:52:25
**Type** Agency Inspection
**Software** 8100.27

**0.05 Lot #** 202506E
**Expiration** 06/24/2027
**0.08 Lot #**
**Expiration**
**0.20 Lot #**
**Expiration**
**0.08 Gas Lot #**
**Expiration**

**Number of Simulators** 5
**Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | No |
| 0.08 g/210L Test: OK | No |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.063 |
| 0.05 g/210L Test # (Repeat) 1 | 0.042 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.059 |
| 0.05 g/210L Test # (Repeat) 2 | 0.046 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.056 |
| 0.05 g/210L Test # (Repeat) 3 | 0.046 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

COT-HOT SIM, 2ND COLD SIM. COT.

**Data Download Date/Time** 07/27/2026 16:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO **Date** 07/27/2026 **Type** Agency Inspection
**Inspector** WAGNER,ROBERT,F **Time** 16:27:18 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 29525080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/27/2026 16:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO **Date** 08/24/2026 **Type** Agency Inspection
**Inspector** LEME,SANTIAGO, **Time** 10:40:17 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 25925080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/24/2026 10:36

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO **Date** 09/30/2026 **Type** Agency Inspection
**Inspector** WAGNER,ROBERT,F **Time** 09:13:48 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 25925080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/30/2026 10:48

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO **Date** 01/16/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:11:24 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.085 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.085 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.085 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/16/2026 10:01

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO **Date** 02/13/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:54:42 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2026 10:16

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO **Date** 03/06/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:13:27 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.085 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.084 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.084 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/06/2026 10:15

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO **Date** 05/20/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:53:19 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/20/2026 10:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:12:55 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/19/2026 09:47

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO **Date** 07/17/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:16:31 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/17/2026 10:03

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO **Date** 08/03/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:15:26 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/03/2026 10:34

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO **Date** 09/18/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 06:46:04 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/18/2026 10:07

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

**Agency** MANATEE COUNTY SO **Date** 01/12/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 10:05:59 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/10/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/12/2026 10:01

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 04/22/2026 | **Type** | Agency Inspection |
| **Inspector** | NOVAK,CHRISTOPHER, | **Time** | 10:25:43 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403411042 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 08/02/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/22/2026 10:22

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 05/22/2026 | **Type** | Agency Inspection |
| **Inspector** | COLEMAN,WILLIAM, | **Time** | 09:18:26 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403419188 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 08/13/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 09:15

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

**Agency** MANATEE COUNTY SO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:12:58 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.000 |
| 0.20 g/210L Test # (Repeat) 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.000 |
| 0.20 g/210L Test # (Repeat) 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.000 |
| 0.20 g/210L Test # (Repeat) 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 06/19/2026 09:10

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

**Agency** MANATEE COUNTY SO **Date** 07/15/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 10:03:43 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403411042
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WIPED CLEAN

**Data Download Date/Time** 07/15/2026 10:02

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 08/20/2026 | **Type** | Agency Inspection |
| **Inspector** | COLEMAN,WILLIAM, | **Time** | 09:22:01 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 260363 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 05/12/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/20/2026 09:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

**Agency** MANATEE COUNTY SO **Date** 09/16/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 14:21:52 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 260363
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/12/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/16/2026 14:39

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

**Agency** MANATEE COUNTY SO **Date** 10/12/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:28:34 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 260363
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/12/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2026 09:25

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

**Agency** CLAY COUNTY SO **Date** 01/29/2026 **Type** Agency Inspection
**Inspector** JOHNSON,MATTHEW,P **Time** 12:54:35 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 01524080A1
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 02/03/2026 18:23

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

**Agency** CLAY COUNTY SO **Date** 08/19/2026 **Type** Agency Inspection
**Inspector** TURNER,JOSEPH,B **Time** 15:30:10 **Software** 8100.27

**0.05 Lot #** 20250E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 36525080A4
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/05/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

RETURN TO EVI USE FROM DEPT INSPECTION

**Data Download Date/Time** 08/19/2026 15:27

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

**Agency** CLAY COUNTY SO **Date** 09/29/2026 **Type** Agency Inspection
**Inspector** JOHNSON,MATTHEW,P **Time** 09:47:55 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 36525080A4
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/05/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 09/29/2026 09:45

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH COUNTY SO **Date** 01/16/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:56:42 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/16/2026 12:55

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH COUNTY SO **Date** 02/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:41:15 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2026 12:40

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH CSO **Date** 03/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:21:02 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/13/2026 13:19

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH CSO **Date** 04/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:26:31 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/17/2026 15:25

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH CSO **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:50:52 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 15:48

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH CSO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:47:52 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/19/2026 12:45

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH CSO **Date** 07/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:59:01 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/17/2026 12:57

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH CSO **Date** 08/14/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:54:52 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 25-3342
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2026 12:51

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PALM BEACH CSO | **Date** | 09/25/2026 | **Type** | Agency Inspection |
| **Inspector** | KARLECKE,JASON,C | **Time** | 12:26:58 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 25-3342 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/29/2027 | **Expiration** | 02/09/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/25/2026 12:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO **Date** 01/16/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 11:58:02 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/16/2026 11:57

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO **Date** 02/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:20:48 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2026 13:18

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO **Date** 03/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 14:02:38 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/13/2026 14:02

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO **Date** 04/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:10:40 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance DRY GAS TUBE NOT CONNECTED.

**Data Download Date/Time** 04/17/2026 15:14

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 16:00:36 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 15:59

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:35:42 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/19/2026 15:32

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO **Date** 07/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:41:10 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/17/2026 13:38

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO **Date** 08/14/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:07:35 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2026 13:04

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO **Date** 09/25/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:13:52 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202510F **0.08 Gas Lot #** 25-3342
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/29/2027 **Expiration** 02/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/25/2026 12:10

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH COUNTY SO **Date** 01/16/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 16:39:01 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/16/2026 16:37

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH COUNTY SO **Date** 02/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:42:28 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2026 15:40

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH CSO **Date** 03/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:59:13 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/13/2026 13:57

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH CSO **Date** 04/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:47:59 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/17/2026 12:51

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH CSO **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:59:06 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 14:00

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH CSO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:22:44 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 25-3342
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/19/2026 13:20

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH CSO **Date** 07/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 14:09:55 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 25-3342
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/17/2026 14:08

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH CSO **Date** 08/14/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:34:12 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 25-3342
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2026 15:32

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 01/22/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 00:45:23 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 3024028434
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 01/22/2026 01:56

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 02/18/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 23:00:16 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 3024028434
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 02/19/2026 01:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 03/11/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 20:56:57 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 3024028434
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.000 |
| 0.08 g/210L Test # (Repeat) 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.000 |
| 0.08 g/210L Test # (Repeat) 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.000 |
| 0.08 g/210L Test # (Repeat) 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: DID NOT CONNECT HOSE.

**Data Download Date/Time** 03/16/2026 12:46

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 04/02/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 01:37:15 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 3024028434
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 04/03/2026 18:31

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 05/23/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 21:17:39 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 3024028434
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 05/23/2026 22:07

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 06/17/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 22:55:29 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 3024028434
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 06/17/2026 23:39

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 07/24/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 22:56:20 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 3024028434
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 07/25/2026 00:21

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 08/27/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 21:21:42 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 3024028434
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/27/2026 21:49

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 09/30/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 21:15:21 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 26-4820
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 04/17/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 10/02/2026 14:39

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL **Date** 10/01/2026 **Type** Agency Inspection
**Inspector** BLAKE,DAVID,A **Time** 00:36:27 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 26-4820
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 04/17/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

INTOX FREE OF CRACKS AND DAMAGE

**Data Download Date/Time** 10/02/2026 14:39

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

**Agency** FL HIGHWAY PATROL **Date** 01/09/2026 **Type** Agency Inspection
**Inspector** SEARS,COLBY,D **Time** 18:28:51 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.204 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/09/2026 18:41

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

**Agency** FL HIGHWAY PATROL **Date** 02/06/2026 **Type** Agency Inspection
**Inspector** SEARS,COLBY,D **Time** 17:42:06 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/06/2026 17:44

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

**Agency** FL HIGHWAY PATROL **Date** 03/06/2026 **Type** Agency Inspection
**Inspector** SEARS,COLBY,D **Time** 17:56:28 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | |
| 0.08 g/210L Test # (Repeat) 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # (Repeat) 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # (Repeat) 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: RFI Detect.

**Data Download Date/Time** 03/06/2026 18:02

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

**Agency** FL HIGHWAY PATROL **Date** 04/04/2026 **Type** Agency Inspection
**Inspector** SEARS,COLBY,D **Time** 19:50:20 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/04/2026 19:56

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

**Agency** FL HIGHWAY PATROL **Date** 06/21/2026 **Type** Agency Inspection
**Inspector** MONTGOMERY,KENNETH,R **Time** 20:55:59 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403135736
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/29/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.207 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.205 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

00: RFI Detect.

**Data Download Date/Time** 06/24/2026 12:13

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

**Agency** VOLUSIA COUNTY SO **Date** 01/30/2026 **Type** Agency Inspection
**Inspector** MALETTO,NICHOLAS,R **Time** 16:06:05 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403135737
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.000 |
| 0.08 g/210L Test # (Repeat) 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.000 |
| 0.08 g/210L Test # (Repeat) 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.000 |
| 0.08 g/210L Test # (Repeat) 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside TOL. SOL CH LOT202510E 10/28/2027

**Data Download Date/Time** 02/02/2026 19:11

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | VOLUSIA COUNTY SO | **Date** | 02/27/2026 | **Type** | Agency Inspection |
| **Inspector** | MALETTO,NICHOLAS,R | **Time** | 16:14:39 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 403135737 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 08/30/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/27/2026 16:13

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

**Agency** VOLUSIA COUNTY SO **Date** 03/29/2026 **Type** Agency Inspection
**Inspector** MALETTO,NICHOLAS,R **Time** 16:34:34 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403135737
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 03/29/2026 16:36

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

**Agency** VOLUSIA COUNTY SO **Date** 04/29/2026 **Type** Agency Inspection
**Inspector** MALETTO,NICHOLAS,R **Time** 18:21:45 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403135737
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/29/2026 18:19

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

**Agency** VOLUSIA COUNTY SO **Date** 05/24/2026 **Type** Agency Inspection
**Inspector** MALETTO,NICHOLAS,R **Time** 15:50:20 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403135737
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 05/24/2026 15:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

**Agency** VOLUSIA COUNTY SO **Date** 06/22/2026 **Type** Agency Inspection
**Inspector** MALETTO,NICHOLAS,R **Time** 10:19:31 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403135737
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/22/2026 10:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

**Agency** VOLUSIA COUNTY SO **Date** 07/27/2026 **Type** Agency Inspection
**Inspector** MALETTO,NICHOLAS,R **Time** 13:57:56 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403135737
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/27/2026 14:03

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

**Agency** VOLUSIA COUNTY SO **Date** 08/24/2026 **Type** Agency Inspection
**Inspector** MALETTO,NICHOLAS,R **Time** 11:07:25 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403135737
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/24/2026 11:17

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

**Agency** VOLUSIA COUNTY SO **Date** 09/30/2026 **Type** Agency Inspection
**Inspector** MALETTO,NICHOLAS,R **Time** 18:50:46 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403135737
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/04/2026 18:00

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD **Date** 01/28/2026 **Type** Agency Inspection
**Inspector** STEMEN,AARON,M **Time** 20:02:20 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/25/2026 19:03

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD **Date** 02/25/2026 **Type** Agency Inspection
**Inspector** STEMEN,AARON,M **Time** 20:07:07 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/25/2026 20:10

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD **Date** 03/30/2026 **Type** Agency Inspection
**Inspector** STEMEN,AARON,M **Time** 21:42:56 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 03/30/2026 21:43

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD **Date** 04/24/2026 **Type** Agency Inspection
**Inspector** STEMEN,AARON,M **Time** 01:49:22 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 04/24/2026 01:58

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD **Date** 05/27/2026 **Type** Agency Inspection
**Inspector** STEMEN,AARON,M **Time** 05:08:54 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 05/27/2026 05:10

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD **Date** 06/24/2026 **Type** Agency Inspection
**Inspector** STEMEN,AARON,M **Time** 02:43:48 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 06/24/2026 02:42

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD **Date** 07/26/2026 **Type** Agency Inspection
**Inspector** STEMEN,AARON,M **Time** 16:54:16 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 07/26/2026 17:13

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD **Date** 08/22/2026 **Type** Agency Inspection
**Inspector** STEMEN,AARON,M **Time** 19:48:31 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 08/22/2026 19:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

**Agency** GAINESVILLE PD **Date** 09/23/2026 **Type** Agency Inspection
**Inspector** STEMEN,AARON,M **Time** 17:16:44 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 19825080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 09/25/2026 05:49

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO **Date** 01/16/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:43:02 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/16/2026 09:56

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO **Date** 02/13/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 08:25:45 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2026 10:19

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO **Date** 03/06/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:43:30 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/06/2026 10:11

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO **Date** 05/20/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 08:25:22 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 05/20/2026 10:52

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 06:36:51 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.074 |
| 0.08 g/210L Test # (Repeat) 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.08 g/210L Test # (Repeat) 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.08 g/210L Test # (Repeat) 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 06/19/2026 09:49

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO **Date** 07/17/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 06:41:17 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/17/2026 10:05

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO **Date** 08/03/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 06:40:57 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/03/2026 10:35

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO **Date** 09/18/2026 **Type** Agency Inspection
**Inspector** LINK,CHRIS,D **Time** 07:20:02 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/18/2026 10:04

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

**Agency** ESCAMBIA COUNTY SO **Date** 01/14/2026 **Type** Agency Inspection
**Inspector** SHELLEY,SAMUEL,L **Time** 09:25:42 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402987895
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/06/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.195 |
| 0.05 g/210L Test # (Repeat) 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.196 |
| 0.05 g/210L Test # (Repeat) 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.197 |
| 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

TESTED INCORRECT SIM.

**Data Download Date/Time** 01/14/2026 10:37

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

**Agency** ESCAMBIA COUNTY SO **Date** 02/09/2026 **Type** Agency Inspection
**Inspector** SHELLEY,SAMUEL,L **Time** 12:38:45 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402987895
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/06/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/09/2026 13:46

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 03/17/2026 | **Type** | Agency Inspection |
| **Inspector** | SHELLEY,SAMUEL,L | **Time** | 11:01:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 402987895 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 03/06/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 03/17/2026 12:05

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 06/26/2026 | **Type** | Agency Inspection |
| **Inspector** | PIERSON,WEN-CHI,K | **Time** | 08:49:38 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 07/15/2026 08:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 01/05/2026 | **Type** | Agency Inspection |
| **Inspector** | PIERSON,WEN-CHI,K | **Time** | 09:00:48 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 01/05/2026 14:43

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

**Agency** ESCAMBIA COUNTY SO **Date** 02/09/2026 **Type** Agency Inspection
**Inspector** SHELLEY,SAMUEL,L **Time** 11:45:26 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402987895
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/06/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.053 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.053 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.053 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/09/2026 12:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 03/17/2026 | **Type** | Agency Inspection |
| **Inspector** | SHELLEY,SAMUEL,L | **Time** | 11:50:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 402987895 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 03/06/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.084 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 03/17/2026 14:28

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 04/21/2026 | **Type** | Agency Inspection |
| **Inspector** | GRUS,TRENTON,L | **Time** | 10:33:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 402987895 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 03/06/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 04/21/2026 12:16

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 05/15/2026 | **Type** | Agency Inspection |
| **Inspector** | GRUS,TRENTON,L | **Time** | 15:38:13 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 402987895 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 03/06/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 05/15/2026 17:33

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 06/17/2026 | **Type** | Agency Inspection |
| **Inspector** | GRUS,TRENTON,L | **Time** | 08:17:14 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 402987895 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 03/06/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/17/2026 10:13

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

**Agency** ESCAMBIA COUNTY SO **Date** 07/13/2026 **Type** Agency Inspection
**Inspector** GRUS,TRENTON,L **Time** 08:56:51 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 402987895
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 03/06/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/13/2026 10:16

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 08/24/2026 | **Type** | Agency Inspection |
| **Inspector** | GRUS,TRENTON,L | **Time** | 09:13:13 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 402987895 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 03/06/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.190 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/24/2026 10:27

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 09/21/2026 | **Type** | Agency Inspection |
| **Inspector** | GRUS,TRENTON,L | **Time** | 09:25:59 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 403525131 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 01/02/2029 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/25/2026 12:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | ESCAMBIA COUNTY SO | **Date** | 10/08/2026 | **Type** | Agency Inspection |
| **Inspector** | COLLINS,CHASE,M | **Time** | 09:59:41 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | Yes | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: Ambient FailAmbient Fail. Non-compliance:AMBIENT FAIL.

**Data Download Date/Time** 10/08/2026 10:57

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

**Agency** MIAMI PD **Date** 02/16/2026 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 11:45:28 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402732434
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/28/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2026 11:45

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

**Agency** MIAMI PD **Date** 03/23/2026 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 18:06:34 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403318190
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/11/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/23/2026 18:06

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

**Agency** MIAMI PD **Date** 04/24/2026 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 10:25:18 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403318190
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/11/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.074 |
| 0.08 g/210L Test # (Repeat) 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.182 |
| 0.20 g/210L Test # (Repeat) 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.045 |
| 0.08 g/210L Test # 2 | 0.074 |
| 0.08 g/210L Test # (Repeat) 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.182 |
| 0.20 g/210L Test # (Repeat) 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test # 3 | 0.074 |
| 0.08 g/210L Test # (Repeat) 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.183 |
| 0.20 g/210L Test # (Repeat) 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

| |
|---|
| 08: Control Outside Tolerance. 20: CHANGE SOLUTIONS e. |

**Data Download Date/Time** 04/24/2026 10:27

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MIAMI PD | **Date** | 05/29/2026 | **Type** | Agency Inspection |
| **Inspector** | FERNANDEZ,ANGEL, | **Time** | 20:53:48 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403318190 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 04/11/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 05/29/2026 20:55

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

**Agency** MIAMI PD **Date** 06/01/2026 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 11:37:05 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403318190
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/11/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.192 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/01/2026 11:35

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

**Agency** MIAMI PD **Date** 07/24/2026 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 12:06:41 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403318190
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/11/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.074 |
| 0.08 g/210L Test # (Repeat) 1 | 0.075 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.074 |
| 0.08 g/210L Test # (Repeat) 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.075 |
| 0.08 g/210L Test # (Repeat) 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.CHANGE SOLUTIONS

**Data Download Date/Time** 07/24/2026 14:40

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

**Agency** MIAMI PD **Date** 08/12/2026 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 10:39:35 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403319190
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/11/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.182 |
| 0.20 g/210L Test # (Repeat) 1 | 0.208 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.20 g/210L Test # (Repeat) 2 | 0.208 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.193 |
| 0.20 g/210L Test # (Repeat) 3 | 0.209 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance CHANGE SOLUTIONS.

**Data Download Date/Time** 08/12/2026 10:35

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

**Agency** MIAMI PD **Date** 09/29/2026 **Type** Agency Inspection
**Inspector** FERNANDEZ,ANGEL, **Time** 18:30:47 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403318190
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 04/11/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.185 |
| 0.20 g/210L Test # (Repeat) 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.20 g/210L Test # (Repeat) 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.20 g/210L Test # (Repeat) 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

20: Control Outside Tolerance CHANGE SOLUTIONS.

**Data Download Date/Time** 10/02/2026 12:57

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 01/22/2026 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 11:05:06 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 01/22/2026 11:03

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 02/16/2026 | **Type** | Agency Inspection |
| **Inspector** | ZITO,ANTHONY,P | **Time** | 12:41:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403207508 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 12/02/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PK

**Data Download Date/Time** 02/16/2026 12:39

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 03/12/2026 | **Type** | Agency Inspection |
| **Inspector** | ZITO,ANTHONY,P | **Time** | 14:41:19 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403207508 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 12/02/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

PK MAIN

**Data Download Date/Time** 03/12/2026 14:38

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

**Agency** MONROE COUNTY SO **Date** 03/25/2026 **Type** Agency Inspection
**Inspector** ZITO,ANTHONY,P **Time** 14:30:07 **Software** 8100.27

**0.05 Lot #** 202410F **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
**Expiration** 10/30/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.191 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

POST EVIDENTIARY AI. GOING TO FDLE

**Data Download Date/Time** 03/25/2026 14:30

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 05/21/2026 | **Type** | Agency Inspection |
| **Inspector** | ZITO,ANTHONY,P | **Time** | 13:06:30 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 403207508 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 12/02/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PK MAIN

**Data Download Date/Time** 05/21/2026 13:08

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 06/26/2026 | **Type** | Agency Inspection |
| **Inspector** | ZITO,ANTHONY,P | **Time** | 13:34:37 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | No |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | No |
| 0.08 g/210L Test: OK | No |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| Interferent Detect Test # 1 | 0.000 |
| Interferent Detect Test # (Repeat) 1 | 0.000 |
| Alcohol Free Test # 2 | 0.000 |
| Interferent Detect Test # 2 | 0.000 |
| Interferent Detect Test # (Repeat) 2 | 0.000 |
| Alcohol Free Test # 3 | 0.000 |
| Interferent Detect Test # 3 | 0.000 |
| Interferent Detect Test # (Repeat) 3 | 0.000 |

**Remarks:**

Int Det:SEAL ON SIM WAS NOT SEALED CORRECTLY . Non-compliance: .

**Data Download Date/Time** 06/26/2026 16:18

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 06/26/2026 | **Type** | Agency Inspection |
| **Inspector** | ZITO,ANTHONY,P | **Time** | 14:18:41 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 403411042 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 08/02/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PK PRIMARY INTOX

**Data Download Date/Time** 06/26/2026 16:18

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 07/20/2026 | **Type** | Agency Inspection |
| **Inspector** | HERNANDEZ,JOSE,C | **Time** | 17:56:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 403411042 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 08/02/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PK MAIN

**Data Download Date/Time** 07/20/2026 17:57

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MONROE COUNTY SO | **Date** | 08/27/2026 | **Type** | Agency Inspection |
| **Inspector** | HERNANDEZ,JOSE,C | **Time** | 11:02:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 403476659 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 10/28/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PK MAIN

**Data Download Date/Time** 08/27/2026 11:11

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

**Agency** MONROE COUNTY SO **Date** 09/26/2026 **Type** Agency Inspection
**Inspector** HERNANDEZ,JOSE,C **Time** 18:23:15 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403476659
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 10/28/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.08 g/210L Test # (Repeat) 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test # (Repeat) 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # (Repeat) 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: RFI DetecT/ PHONE WAS NEAR INTOX

**Data Download Date/Time** 09/26/2026 18:22

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 01/09/2026 | **Type** | Agency Inspection |
| **Inspector** | RUBEL,NEIL,F | **Time** | 11:55:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403374313 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 06/25/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.084 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.084 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.084 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 01/09/2026 12:58

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

**Agency** FL HIGHWAY PATROL **Date** 02/06/2026 **Type** Agency Inspection
**Inspector** RUBEL,NEIL,F **Time** 13:02:43 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402946336
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 01/17/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 02/06/2026 14:06

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 03/12/2026 | **Type** | Agency Inspection |
| **Inspector** | RUBEL,NEIL,F | **Time** | 12:08:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 402946336 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 01/17/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 03/12/2026 14:03

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

**Agency** FL HIGHWAY PATROL **Date** 04/04/2026 **Type** Agency Inspection
**Inspector** MCDOWELL,MICHAEL,J **Time** 09:01:58 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402946336
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 01/17/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 04/04/2026 10:02

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 05/04/2026 | **Type** | Agency Inspection |
| **Inspector** | MCDOWELL,MICHAEL,J | **Time** | 22:56:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 402946336 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 01/17/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.206 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.203 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 05/05/2026 00:09

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

**Agency** FL HIGHWAY PATROL **Date** 06/02/2026 **Type** Agency Inspection
**Inspector** MCDOWELL,MICHAEL,J **Time** 06:00:01 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 402946336
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 01/17/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.045 |
| 0.08 g/210L Test # 1 | 0.074 |
| 0.08 g/210L Test # (Repeat) 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.190 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.08 g/210L Test # (Repeat) 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.08 g/210L Test # (Repeat) 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance - TIGHTENED LID ON SIM.

**Data Download Date/Time** 06/02/2026 07:05

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

**Agency** FL HIGHWAY PATROL **Date** 07/03/2026 **Type** Agency Inspection
**Inspector** BRAMMER,DALTON,J **Time** 18:44:18 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402946336
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/17/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/08/2026 18:17

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 08/13/2026 | **Type** | Agency Inspection |
| **Inspector** | MCDOWELL,MICHAEL,J | **Time** | 20:31:13 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 402946336 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 01/17/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 08/13/2026 21:28

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 09/04/2026 | **Type** | Agency Inspection |
| **Inspector** | MCDOWELL,MICHAEL,J | **Time** | 22:12:08 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 402946336 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 01/17/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 09/04/2026 23:15

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 10/04/2026 | **Type** | Agency Inspection |
| **Inspector** | MCDOWELL,MICHAEL,J | **Time** | 08:45:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 402946336 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 01/17/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 10/08/2026 09:52

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

**Agency** PINECREST PD **Date** 01/15/2026 **Type** Agency Inspection
**Inspector** WADDELL,NATHANIEL,J **Time** 09:10:19 **Software** 8100.27

**0.05 Lot #** 202410E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403268835
**Expiration** 10/22/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/13/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/15/2026 09:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

**Agency** PINECREST PD **Date** 02/13/2026 **Type** Agency Inspection
**Inspector** VEGA,OBEL, **Time** 15:40:39 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403268835
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/13/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2026 15:43

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PINECREST PD | **Date** | 09/25/2026 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 11:27:35 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 09/25/2026 16:21

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

**Agency** BROWARD COUNTY SO **Date** 01/14/2026 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:54:03 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402783528
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 01/14/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

S TILMAN IN TRAINING

**Data Download Date/Time** 01/14/2026 11:14

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BROWARD COUNTY SO | **Date** | 02/04/2026 | **Type** | Agency Inspection |
| **Inspector** | FRAZIER,ANAYA,S | **Time** | 09:38:15 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 402783528 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 06/27/2026 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

S. TILLMAN IN TRAINING.

**Data Download Date/Time** 02/04/2026 09:38

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

**Agency** BROWARD COUNTY SO **Date** 03/04/2026 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 08:09:31 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402783528
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 06/27/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

S. TILLAN IN TRAINING.

**Data Download Date/Time** 03/04/2026 08:25

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BROWARD COUNTY SO | **Date** | 04/01/2026 | **Type** | Agency Inspection |
| **Inspector** | FRAZIER,ANAYA,S | **Time** | 08:24:41 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 402783528 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 06/27/2026 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

S. TILLMAN IN TRAINING

**Data Download Date/Time** 04/01/2026 08:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

**Agency** BROWARD COUNTY SO **Date** 05/30/2026 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 22:55:38 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/30/2026 22:55

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BROWARD COUNTY SO | **Date** | 06/10/2026 | **Type** | Agency Inspection |
| **Inspector** | FRAZIER,ANAYA,S | **Time** | 08:33:03 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 403411042 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 08/02/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

S. TILLMAN IN TRAINING.

**Data Download Date/Time** 06/10/2026 08:34

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

**Agency** BROWARD COUNTY SO **Date** 07/01/2026 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 10:51:28 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403411042
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

S. TILLMAN IN TRAINING.

**Data Download Date/Time** 07/01/2026 11:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

**Agency** BROWARD COUNTY SO **Date** 08/05/2026 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 08:20:59 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403411042
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

S. TILLMAN IN TRAINING.

**Data Download Date/Time** 08/05/2026 08:32

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

**Agency** BROWARD COUNTY SO **Date** 09/02/2026 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 08:22:25 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403411042
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/03/2026 03:41

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

**Agency** BROWARD COUNTY SO **Date** 10/07/2026 **Type** Agency Inspection
**Inspector** FRAZIER,ANAYA,S **Time** 07:47:27 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403411042
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

S. TILLMAN IN TRAINING

**Data Download Date/Time** 10/07/2026 07:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO **Date** 01/16/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:57:11 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/16/2026 15:57

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO **Date** 02/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:07:22 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2026 15:09

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO **Date** 03/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 16:22:12 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/13/2026 16:20

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO **Date** 04/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:04:13 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/17/2026 13:02

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:55:35 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 13:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:10:59 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/19/2026 15:10

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO **Date** 07/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 16:45:43 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/17/2026 16:53

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO **Date** 08/14/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:16:08 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 25-3342
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2026 15:13

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO **Date** 09/25/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 14:14:57 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202510F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/29/2027 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/25/2026 14:11

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO **Date** 01/16/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:28:24 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/16/2026 13:28

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO **Date** 02/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:05:55 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2026 13:15

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO **Date** 03/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:43:29 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/13/2026 13:42

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO **Date** 04/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:42:34 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/17/2026 12:41

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:19:02 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 13:16

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 14:54:02 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/19/2026 14:51

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO **Date** 07/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 16:18:22 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/17/2026 16:15

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO **Date** 08/14/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 16:22:38 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2026 16:20

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO **Date** 09/25/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 14:44:13 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202510F **0.08 Gas Lot #** 25-3342
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/29/2027 **Expiration** 02/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/25/2026 14:52

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

**Agency** PALM BEACH CSO **Date** 01/16/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 13:23:49 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/16/2026 13:30

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

**Agency** PALM BEACH CSO **Date** 02/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:24:32 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/13/2026 15:23

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

**Agency** PALM BEACH CSO **Date** 03/13/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 15:54:22 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/13/2026 15:52

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

**Agency** PALM BEACH CSO **Date** 04/17/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 14:37:07 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 32324080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/17/2026 14:35

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

**Agency** PALM BEACH CSO **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 16:19:48 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 16:19

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

**Agency** PALM BEACH CSO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:27:28 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 19825080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/19/2026 12:25

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | PALM BEACH CSO | **Date** | 07/17/2026 | **Type** | Agency Inspection |
| **Inspector** | KARLECKE,JASON,C | **Time** | 13:57:25 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 25-3342 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 02/09/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/17/2026 13:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

**Agency** PALM BEACH CSO **Date** 08/14/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 12:57:18 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 25-3342
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 02/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/14/2026 13:16

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

**Agency** PALM BEACH CSO **Date** 09/25/2026 **Type** Agency Inspection
**Inspector** KARLECKE,JASON,C **Time** 14:41:26 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202510F **0.08 Gas Lot #** 25-3342
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/29/2027 **Expiration** 02/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/25/2026 15:05

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

**Agency** MANATEE COUNTY SO **Date** 01/12/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:32:27 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 09/10/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 01/12/2026 09:29

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

**Agency** MANATEE COUNTY SO **Date** 02/16/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 13:53:00 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/10/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2026 13:51

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

**Agency** MANATEE COUNTY SO **Date** 03/16/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 13:54:04 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/10/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/16/2026 13:53

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

**Agency** MANATEE COUNTY SO **Date** 04/22/2026 **Type** Agency Inspection
**Inspector** NOVAK,CHRISTOPHER, **Time** 09:28:14 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
**Expiration** 06/24/2026 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/10/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.193 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: .

**Data Download Date/Time** 04/22/2026 09:26

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

**Agency** MANATEE COUNTY SO **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 14:24:11 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/10/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 14:22

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

**Agency** MANATEE COUNTY SO **Date** 07/15/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 14:38:59 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 264944
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/12/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WIPED CLEAN

**Data Download Date/Time** 07/15/2026 14:35

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

**Agency** MANATEE COUNTY SO **Date** 09/16/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:27:53 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403374313
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 06/25/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/16/2026 09:26

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

**Agency** MANATEE COUNTY SO **Date** 10/12/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 13:30:04 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403374313
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 06/25/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2026 13:29

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

**Agency** MANATEE COUNTY SO **Date** 01/12/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:38:39 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/12/2026 09:34

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

**Agency** MANATEE COUNTY SO **Date** 02/16/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 14:00:32 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2026 13:59

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

**Agency** MANATEE COUNTY SO **Date** 03/16/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:10:04 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/16/2026 09:06

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

**Agency** MANATEE COUNTY SO **Date** 04/22/2026 **Type** Agency Inspection
**Inspector** NOVAK,CHRISTOPHER, **Time** 09:13:51 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/10/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/22/2026 09:12

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

**Agency** MANATEE COUNTY SO **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 14:22:09 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/10/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 14:19

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 07/21/2026 | **Type** | Agency Inspection |
| **Inspector** | SPEARIN,KATIE,T | **Time** | 10:31:12 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED TO OPERATE. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 07/21/2026 13:57

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

**Agency** MANATEE COUNTY SO **Date** 08/20/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:10:54 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 260493
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 04/17/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.052 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.052 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/20/2026 10:31

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

**Agency** MANATEE COUNTY SO **Date** 09/16/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:22:50 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 260493
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 04/17/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/16/2026 09:22

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 10/12/2026 | **Type** | Agency Inspection |
| **Inspector** | COLEMAN,WILLIAM, | **Time** | 13:18:52 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 260493 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 04/17/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2026 13:17

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

**Agency** HILLSBOROUGH CSO **Date** 01/15/2026 **Type** Agency Inspection
**Inspector** BROWN,KEVIN,C **Time** 08:23:18 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 29425080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 12/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B KROME #263389 PRESENT FOR INSPECTION

**Data Download Date/Time** 01/15/2026 09:22

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

**Agency** HILLSBOROUGH CSO **Date** 03/12/2026 **Type** Agency Inspection
**Inspector** BROWN,KEVIN,C **Time** 10:19:37 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 04125080A6
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B KROME #263389 PRESENT FOR INSPECTION

**Data Download Date/Time** 03/12/2026 11:39

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

**Agency** HILLSBOROUGH CSO **Date** 04/09/2026 **Type** Agency Inspection
**Inspector** BROWN,KEVIN,C **Time** 08:40:26 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 04125080A6
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 03/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI B KROME #263389 PRESENT FOR INSPECTION

**Data Download Date/Time** 04/09/2026 09:55

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

**Agency** HILLSBOROUGH CSO **Date** 05/14/2026 **Type** Agency Inspection
**Inspector** BROWN,KEVIN,C **Time** 11:03:45 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** **0.08 Gas Lot #**
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | No |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

AI J HAMMOND #264409 PRESENT FOR INSPECTIONNon-compliance:AMB FAIL.

**Data Download Date/Time** 05/15/2026 12:46

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

**Agency** FLAGLER COUNTY SO **Date** 01/16/2026 **Type** Agency Inspection
**Inspector** WEAVER,DANIEL, **Time** 13:48:42 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 23725080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 01/16/2026 13:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

**Agency** FLAGLER COUNTY SO **Date** 02/28/2026 **Type** Agency Inspection
**Inspector** WEAVER,DANIEL, **Time** 14:24:12 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 23725080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Interferent Detect Test # (Repeat) 1 | INT |
| Alcohol Free Test # (Repeat) 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Interferent Detect Test # (Repeat) 2 | INT |
| Alcohol Free Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | |
| Interferent Detect Test # (Repeat) 3 | INT |

**Remarks:**

00: Ambient Fail. Int Det: RFI Detect - RADIO. MONTHLY INSP

**Data Download Date/Time** 02/28/2026 14:23

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

**Agency** FLAGLER COUNTY SO **Date** 04/17/2026 **Type** Agency Inspection
**Inspector** FAIRCLOTH,EDWARD,W **Time** 10:15:24 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202510F **0.08 Gas Lot #** 23725080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/29/2027 **Expiration** 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 04/17/2026 10:12

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

**Agency** FLAGLER COUNTY SO **Date** 05/30/2026 **Type** Agency Inspection
**Inspector** FAIRCLOTH,EDWARD,W **Time** 10:54:38 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202510F **0.08 Gas Lot #** 23725080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/29/2027 **Expiration** 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 06/02/2026 15:09

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

**Agency** FLAGLER COUNTY SO **Date** 06/11/2026 **Type** Agency Inspection
**Inspector** FAIRCLOTH,EDWARD,W **Time** 14:24:27 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 23725080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 06/12/2026 12:56

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

**Agency** FLAGLER COUNTY SO **Date** 07/16/2026 **Type** Agency Inspection
**Inspector** FAIRCLOTH,EDWARD,W **Time** 09:30:32 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 23725080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 07/16/2026 09:43

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

**Agency** FLAGLER COUNTY SO **Date** 08/17/2026 **Type** Agency Inspection
**Inspector** WEAVER,DANIEL, **Time** 10:26:45 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 23725080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 08/17/2026 10:23

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

**Agency** FLAGLER COUNTY SO **Date** 09/18/2026 **Type** Agency Inspection
**Inspector** FAIRCLOTH,EDWARD,W **Time** 11:14:31 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 23725080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 09/18/2026 11:38

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | KENNEDY SPACE CENTER | **Date** | 01/28/2026 | **Type** | Agency Inspection |
| **Inspector** | ETTMAN,KRISTOFER,C | **Time** | 03:05:14 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

A F / M A: . Non-compliance: .

**Data Download Date/Time** 01/28/2026 11:41

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | KENNEDY SPACE CENTER | **Date** | 01/28/2026 | **Type** | Agency Inspection |
| **Inspector** | ETTMAN,KRISTOFER,C | **Time** | 04:17:40 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 08425080A1 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 05/05/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided.

**Data Download Date/Time** 01/28/2026 11:41

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

**Agency** KENNEDY SPACE CENTER **Date** 02/24/2026 **Type** Agency Inspection
**Inspector** SAVARD,CHARLES,M **Time** 11:55:12 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.194 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/24/2026 12:46

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | KENNEDY SPACE CENTER | **Date** | 03/25/2026 | **Type** | Agency Inspection |
| **Inspector** | ETTMAN,KRISTOFER,C | **Time** | 02:45:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 08425080A1 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 05/05/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/05/2026 23:51

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

**Agency** KENNEDY SPACE CENTER **Date** 04/25/2026 **Type** Agency Inspection
**Inspector** SAVARD,CHARLES,M **Time** 07:54:17 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/27/2026 23:01

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

**Agency** KENNEDY SPACE CENTER **Date** 05/24/2026 **Type** Agency Inspection
**Inspector** ETTMAN,KRISTOFER,C **Time** 00:08:43 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A1
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.046 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.08 g/210L Test # 2 | 0.076 |
| 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.076 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/24/2026 00:05

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

**Agency** KENNEDY SPACE CENTER **Date** 06/25/2026 **Type** Agency Inspection
**Inspector** SAVARD,CHARLES,M **Time** 11:55:39 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 08425080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/25/2026 12:15

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

**Agency** KENNEDY SPACE CENTER **Date** 07/25/2026 **Type** Agency Inspection
**Inspector** SAVARD,CHARLES,M **Time** 08:20:24 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 08425080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/25/2026 08:36

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | KENNEDY SPACE CENTER | **Date** | 09/01/2026 | **Type** | Agency Inspection |
| **Inspector** | HIGGINBOTHAM,LEANDRA, | **Time** | 13:05:09 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

**Data Download Date/Time** 09/01/2026 16:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

**Agency** KENNEDY SPACE CENTER **Date** 09/17/2026 **Type** Agency Inspection
**Inspector** SAVARD,CHARLES,M **Time** 11:36:34 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 08425080A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/17/2026 11:35

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

**Agency** BOCA RATON PD **Date** 01/13/2026 **Type** Agency Inspection
**Inspector** CRAWFORD,ANDREW, **Time** 16:11:22 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.076 |
| 0.08 g/210L Test # (Repeat) 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test # (Repeat) 2 | 0.077 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.077 |
| 0.08 g/210L Test # (Repeat) 3 | 0.077 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: RFI Detect IN CAR RADIO SALLYPORT. PASS ALL OKAY

**Data Download Date/Time** 01/13/2026 16:15

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

**Agency** BOCA RATON PD **Date** 02/17/2026 **Type** Agency Inspection
**Inspector** CRAWFORD,ANDREW, **Time** 10:10:02 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASS

**Data Download Date/Time** 02/17/2026 10:09

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

**Agency** BOCA RATON PD **Date** 03/11/2026 **Type** Agency Inspection
**Inspector** CRAWFORD,ANDREW, **Time** 15:18:58 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASS

**Data Download Date/Time** 03/11/2026 15:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

**Agency** BOCA RATON PD **Date** 04/16/2026 **Type** Agency Inspection
**Inspector** CRAWFORD,ANDREW, **Time** 10:48:05 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A2
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASS

**Data Download Date/Time** 04/16/2026 10:53

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

**Agency** BOCA RATON PD **Date** 05/21/2026 **Type** Agency Inspection
**Inspector** CRAWFORD,ANDREW, **Time** 16:27:52 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASS

**Data Download Date/Time** 05/21/2026 16:31

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

**Agency** BOCA RATON PD **Date** 06/25/2026 **Type** Agency Inspection
**Inspector** CRAWFORD,ANDREW, **Time** 14:58:06 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASS

**Data Download Date/Time** 06/25/2026 14:56

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

**Agency** BOCA RATON PD **Date** 07/22/2026 **Type** Agency Inspection
**Inspector** CRAWFORD,ANDREW, **Time** 09:57:00 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASS

**Data Download Date/Time** 07/22/2026 09:58

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | BOCA RATON PD | **Date** | 07/23/2026 | **Type** | Agency Inspection |
| **Inspector** | CRAWFORD,ANDREW, | **Time** | 18:17:11 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 08425080A2 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 05/05/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASS

**Data Download Date/Time** 07/23/2026 18:13

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

**Agency** BOCA RATON PD **Date** 08/17/2026 **Type** Agency Inspection
**Inspector** CRAWFORD,ANDREW, **Time** 10:47:55 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 08425080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASS

**Data Download Date/Time** 08/17/2026 10:44

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

**Agency** BOCA RATON PD **Date** 09/24/2026 **Type** Agency Inspection
**Inspector** CRAWFORD,ANDREW, **Time** 10:12:06 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 08425080A2
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

PASS

**Data Download Date/Time** 09/24/2026 10:09

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 01/30/2026 | **Type** | Agency Inspection |
| **Inspector** | KEMP,TREVOR,D | **Time** | 17:00:18 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403135736 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 08/29/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 01/30/2026 18:13

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 02/22/2026 | **Type** | Agency Inspection |
| **Inspector** | KEMP,TREVOR,D | **Time** | 16:39:23 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403135736 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 08/29/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 02/22/2026 17:18

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

**Agency** FL HIGHWAY PATROL **Date** 03/20/2026 **Type** Agency Inspection
**Inspector** KEMP,TREVOR,D **Time** 13:18:48 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403135736
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/29/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 03/23/2026 16:12

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

**Agency** FL HIGHWAY PATROL **Date** 04/26/2026 **Type** Agency Inspection
**Inspector** KEMP,TREVOR,D **Time** 19:27:06 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403135736
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/29/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 04/26/2026 19:52

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

**Agency** FL HIGHWAY PATROL **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** KEMP,TREVOR,D **Time** 16:45:25 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403135736
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/29/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 16:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

**Agency** FL HIGHWAY PATROL **Date** 06/28/2026 **Type** Agency Inspection
**Inspector** KEMP,TREVOR,D **Time** 16:41:17 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403135736
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/29/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 06/28/2026 17:27

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

**Agency** FL HIGHWAY PATROL **Date** 07/13/2026 **Type** Agency Inspection
**Inspector** KEMP,TREVOR,D **Time** 16:18:09 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403135736
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/29/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 07/13/2026 17:00

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 09/09/2026 | **Type** | Agency Inspection |
| **Inspector** | GUTSCHOW,TAYLOR,D | **Time** | 13:16:54 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED

**Data Download Date/Time** 09/10/2026 14:42

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

**Agency** FHP **Date** 01/22/2026 **Type** Agency Inspection
**Inspector** SOUZA,MAYKOLL,G **Time** 10:28:23 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403342588
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 05/09/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/22/2026 10:37

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FHP | **Date** | 09/09/2026 | **Type** | Agency Inspection |
| **Inspector** | NOVAK,TYLER,B | **Time** | 12:54:02 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403342588 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 05/09/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 09/09/2026 13:33

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 01/22/2026 | **Type** | Agency Inspection |
| **Inspector** | SOUZA,MAYKOLL,G | **Time** | 10:22:54 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403342588 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 05/09/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.077 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/22/2026 10:39

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 09/09/2026 | **Type** | Agency Inspection |
| **Inspector** | NOVAK,TYLER,B | **Time** | 12:00:32 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403342588 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 05/09/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 09/09/2026 13:20

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 01/30/2026 | **Type** | Agency Inspection |
| **Inspector** | KEMP,TREVOR,D | **Time** | 17:36:09 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403268835 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 02/13/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 01/30/2026 18:04

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 02/22/2026 | **Type** | Agency Inspection |
| **Inspector** | KEMP,TREVOR,D | **Time** | 16:51:27 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403268835 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 02/13/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 02/22/2026 17:23

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 03/20/2026 | **Type** | Agency Inspection |
| **Inspector** | KEMP,TREVOR,D | **Time** | 13:33:06 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403268835 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 02/13/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 03/23/2026 16:23

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 04/26/2026 | **Type** | Agency Inspection |
| **Inspector** | KEMP,TREVOR,D | **Time** | 19:15:52 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403268835 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 02/13/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

NA

**Data Download Date/Time** 04/26/2026 19:48

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

**Agency** FL HIGHWAY PATROL **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** KEMP,TREVOR,D **Time** 16:52:26 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403268835
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 02/13/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 05/22/2026 16:55

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

**Agency** FL HIGHWAY PATROL **Date** 06/28/2026 **Type** Agency Inspection
**Inspector** KEMP,TREVOR,D **Time** 16:52:12 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403268835
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 02/13/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.052 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 06/28/2026 17:24

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 07/13/2026 | **Type** | Agency Inspection |
| **Inspector** | PEREZ MORALES,JAVIER,D | **Time** | 15:36:30 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** No

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

Non-compliance: .

**Data Download Date/Time** 07/13/2026 16:58

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

**Agency** FL HIGHWAY PATROL **Date** 07/13/2026 **Type** Agency Inspection
**Inspector** PEREZ MORALES,JAVIER,D **Time** 16:13:51 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202610E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403268835
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 02/13/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

N/A

**Data Download Date/Time** 07/13/2026 16:58

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

**Agency** MANATEE COUNTY SO **Date** 01/12/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 15:53:01 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403419188
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/13/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/12/2026 15:52

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 02/16/2026 | **Type** | Agency Inspection |
| **Inspector** | COLEMAN,WILLIAM, | **Time** | 09:41:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403419188 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 08/13/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2026 09:37

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

**Agency** MANATEE COUNTY SO **Date** 03/16/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 10:20:05 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403419188
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/13/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/16/2026 10:17

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

**Agency** MANATEE COUNTY SO **Date** 04/22/2026 **Type** Agency Inspection
**Inspector** NOVAK,CHRISTOPHER, **Time** 13:47:02 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403419188
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/13/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 04/22/2026 13:45

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

**Agency** MANATEE COUNTY SO **Date** 05/22/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 15:08:44 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403141809
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 09/10/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/22/2026 15:06

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

**Agency** MANATEE COUNTY SO **Date** 06/19/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:29:11 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 1502508A1
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 07/05/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 06/19/2026 09:26

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

**Agency** MANATEE COUNTY SO **Date** 07/15/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 14:33:13 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403374313
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 06/25/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

WIPED CLEAN

**Data Download Date/Time** 07/15/2026 14:33

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 08/20/2026 | **Type** | Agency Inspection |
| **Inspector** | COLEMAN,WILLILAM, | **Time** | 09:15:26 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 264944 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 05/12/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/20/2026 10:25

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 09/16/2026 | **Type** | Agency Inspection |
| **Inspector** | COLEMAN,WILLIAM, | **Time** | 14:15:05 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 264944 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 05/12/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 09/16/2026 14:37

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 10/12/2026 | **Type** | Agency Inspection |
| **Inspector** | COLEMAN,WILLIAM, | **Time** | 09:49:57 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 264944 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 05/12/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2026 09:46

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

**Agency** FL HIGHWAY PATROL **Date** 02/11/2026 **Type** Agency Inspection
**Inspector** PALESE,ANTHONY,W **Time** 20:21:54 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

M/A.

**Data Download Date/Time** 02/14/2026 17:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

**Agency** FL HIGHWAY PATROL **Date** 03/26/2026 **Type** Agency Inspection
**Inspector** PALESE,ANTHONY,W **Time** 17:35:26 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/26/2026 16:36

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 04/15/2026 | **Type** | Agency Inspection |
| **Inspector** | PALESE,ANTHONY,W | **Time** | 15:17:08 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 402872586 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 10/12/2026 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | Yes | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 04/15/2026 15:22

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 05/07/2026 | **Type** | Agency Inspection |
| **Inspector** | SPEARIN,KATIE,T | **Time** | 11:12:18 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED TO CONFIRM FORMS LOADED PROPERLY.

**Data Download Date/Time** 05/07/2026 11:14

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

**Agency** FL HIGHWAY PATROL **Date** 07/21/2026 **Type** Agency Inspection
**Inspector** PALESE,ANTHONY,W **Time** 17:36:14 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/21/2026 17:34

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

**Agency** FL HIGHWAY PATROL **Date** 08/09/2026 **Type** Agency Inspection
**Inspector** PALESE,ANTHONY,W **Time** 22:01:47 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 402872586
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/09/2026 21:58

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 09/12/2026 | **Type** | Agency Inspection |
| **Inspector** | PALESE,ANTHONY,W | **Time** | 17:55:23 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** 5 **Compliance** No

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | No |
| Diagnostic Check (Post-Inspection): OK | No |
| Alcohol Free Test: OK | No |
| 0.05 g/210L Test: OK | No |
| 0.08 g/210L Test: OK | No |
| 0.20 g/210L Test: OK | No |
| 0.08 g/210L Dry Gas Std Test: OK | No |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | INT |
| Alcohol Free Test # (Repeat) 1 | 0.000 |
| Alcohol Free Test # (Repeat) 2 | INT |

**Remarks:**

Time-Date changed. USED DIFF H20 SIM, Interferent Detect. Non-compliance:DET INT ON H20.

**Data Download Date/Time** 09/12/2026 21:00

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 09/12/2026 | **Type** | Agency Inspection |
| **Inspector** | PALESE,ANTHONY,W | **Time** | 21:02:50 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403135736 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 08/29/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/12/2026 21:00

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

**Agency** FL HIGHWAY PATROL **Date** 09/12/2026 **Type** Agency Inspection
**Inspector** PALESE,ANTHONY,W **Time** 21:09:34 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403135736
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/29/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/12/2026 21:06

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

**Agency** MANATEE COUNTY SO **Date** 01/12/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 15:59:33 **Software** 8100.27

**0.05 Lot #** 202402E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
**Expiration** 02/27/2026 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 01/12/2026 15:55

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 02/16/2026 | **Type** | Agency Inspection |
| **Inspector** | COLEMAN,WILLIAM, | **Time** | 09:47:21 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403411042 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/30/2026 | **Expiration** | 08/02/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/16/2026 09:43

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

**Agency** MANATEE COUNTY SO **Date** 03/16/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 08:50:05 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403411042
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/30/2026 **Expiration** 08/02/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/16/2026 08:47

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 05/19/2026 | **Type** | Agency Inspection |
| **Inspector** | SPEARIN,KATIE,T | **Time** | 10:10:47 | **Software** | 8100.27 |

| | | | |
|---|---|---|---|
| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASSED TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 05/28/2026 08:04

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | MANATEE COUNTY SO | **Date** | 07/15/2026 | **Type** | Agency Inspection |
| **Inspector** | COLEMAN,WILLIAM, | **Time** | 09:39:28 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202510E | **0.20 Lot #** | 202510F | **0.08 Gas Lot #** | 264944 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/28/2027 | **Expiration** | 10/29/2027 | **Expiration** | 05/12/2028 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | Yes |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 07/15/2026 09:36

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

**Agency** MANATEE COUNTY SO **Date** 08/20/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 14:52:59 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 264944
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/12/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

A F / M A: No Sample Provided.

**Data Download Date/Time** 08/20/2026 14:50

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

**Agency** MANATEE COUNTY SO **Date** 09/17/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 14:38:40 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 264944
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/12/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/17/2026 14:35

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

**Agency** MANATEE COUNTY SO **Date** 10/12/2026 **Type** Agency Inspection
**Inspector** COLEMAN,WILLIAM, **Time** 09:39:56 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 264944
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 05/12/2028

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/12/2026 09:36

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 02/11/2026 | **Type** | Agency Inspection |
| **Inspector** | HIGGINBOTHAM,LEANDRA, | **Time** | 11:01:22 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

**Data Download Date/Time** 02/12/2026 17:21

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

**Agency** FL HIGHWAY PATROL **Date** 03/31/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 23:16:30 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.048 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/31/2026 23:13

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

**Agency** FL HIGHWAY PATROL **Date** 04/29/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 21:24:07 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.044 |
| 0.05 g/210L Test # (Repeat) 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.046 |
| 0.05 g/210L Test # (Repeat) 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.046 |
| 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

05: Control Outside Tolerance.REPEATED OK

**Data Download Date/Time** 04/29/2026 21:21

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

**Agency** FL HIGHWAY PATROL **Date** 05/26/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 11:00:53 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.082 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/27/2026 07:54

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

**Agency** FL HIGHWAY PATROL **Date** 06/28/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 09:52:31 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 07/30/2026 12:33

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

**Agency** FL HIGHWAY PATROL **Date** 07/30/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 14:01:23 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.051 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/28/2026 13:51

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

**Agency** FL HIGHWAY PATROL **Date** 08/29/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 08:24:35 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/29/2026 08:53

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

**Agency** FL HIGHWAY PATROL **Date** 09/29/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 14:43:47 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510E **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/28/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.195 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.048 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.048 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/29/2026 16:31

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

**Agency** FL HIGHWAY PATROL **Date** 01/30/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 00:04:20 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.083 |
| 0.20 g/210L Test # 1 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.082 |
| 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.081 |
| 0.20 g/210L Test # 3 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/02/2026 10:58

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

**Agency** FL HIGHWAY PATROL **Date** 02/28/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 13:05:12 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202410E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.078 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.049 |
| 0.08 g/210L Test # 2 | 0.078 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Test # 3 | 0.078 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 02/28/2026 20:15

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 03/31/2026 | **Type** | Agency Inspection |
| **Inspector** | HITCHCOCK,WILLIAM,D | **Time** | 15:59:49 | **Software** | 8100.27 |

| | | | | | | | |
|---|---|---|---|---|---|---|---|
| **0.05 Lot #** | 202506E | **0.08 Lot #** | 202410E | **0.20 Lot #** | 202410F | **0.08 Gas Lot #** | 403207507 |
| **Expiration** | 06/24/2027 | **Expiration** | 10/22/2026 | **Expiration** | 10/30/2026 | **Expiration** | 11/26/2027 |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 03/31/2026 23:08

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

**Agency** FL HIGHWAY PATROL **Date** 05/26/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 10:55:43 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 201510E **0.20 Lot #** 201510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.082 |
| 0.20 g/210L Test # 3 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 05/29/2026 09:44

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

**Agency** FL HIGHWAY PATROL **Date** 06/28/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 09:59:42 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.050 |
| 0.08 g/210L Test # 1 | 0.080 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.050 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/02/2026 14:17

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

**Agency** FL HIGHWAY PATROL **Date** 07/30/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 14:06:47 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.051 |
| 0.08 g/210L Test # 1 | 0.081 |
| 0.20 g/210L Test # 1 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.051 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.052 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 08/02/2026 14:17

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

**Agency** FL HIGHWAY PATROL **Date** 08/29/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 08:19:23 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.049 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.050 |
| 0.08 g/210L Test # 2 | 0.080 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.052 |
| 0.08 g/210L Test # 3 | 0.080 |
| 0.20 g/210L Test # 3 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 09/05/2026 11:26

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

**Agency** FL HIGHWAY PATROL **Date** 09/29/2026 **Type** Agency Inspection
**Inspector** HITCHCOCK,WILLIAM,D **Time** 06:54:57 **Software** 8100.27

**0.05 Lot #** 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202510F **0.08 Gas Lot #** 403207507
**Expiration** 06/24/2027 **Expiration** 10/28/2027 **Expiration** 10/29/2027 **Expiration** 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

| Check or Test | Result |
|---|---|
| Diagnostic Check (Pre-Inspection): OK | Yes |
| Date and/or Time Adjusted | No |
| Minimum Sample Volume Check: OK | |
| Barometric Pressure Sensor Check: OK | |
| Alcohol Free Subject Test: 0.000 | Yes |
| Mouth Alcohol Test: Slope Not Met | Yes |
| Interferent Detect Test: Interferent Detect | Yes |
| Diagnostic Check (Post-Inspection): OK | Yes |
| Alcohol Free Test: OK | Yes |
| 0.05 g/210L Test: OK | Yes |
| 0.08 g/210L Test: OK | Yes |
| 0.20 g/210L Test: OK | Yes |
| 0.08 g/210L Dry Gas Std Test: OK | Yes |

| Test | Result |
|---|---|
| Alcohol Free Test # 1 | 0.000 |
| 0.05 g/210L Test # 1 | 0.047 |
| 0.08 g/210L Test # 1 | 0.079 |
| 0.20 g/210L Test # 1 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Interferent Detect Test # 1 | INT |
| Alcohol Free Test # 2 | 0.000 |
| 0.05 g/210L Test # 2 | 0.047 |
| 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| Interferent Detect Test # 2 | INT |
| Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test # 3 | 0.047 |
| 0.08 g/210L Test # 3 | 0.079 |
| 0.20 g/210L Test # 3 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| Interferent Detect Test # 3 | INT |

**Remarks:**

**Data Download Date/Time** 10/06/2026 10:51

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

| | | | | | |
|---|---|---|---|---|---|
| **Agency** | FL HIGHWAY PATROL | **Date** | 10/01/2026 | **Type** | Agency Inspection |
| **Inspector** | SMYKOWSKI,ISABELLA,A | **Time** | 13:18:39 | **Software** | 8100.27 |

| **0.05 Lot #** | **0.08 Lot #** | **0.20 Lot #** | **0.08 Gas Lot #** |
|---|---|---|---|
| **Expiration** | **Expiration** | **Expiration** | **Expiration** |

**Number of Simulators** **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|---|---|---|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

**Remarks:**

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 10/12/2026 06:45

**Printed: 13 October 2026**

COBRA ©1997-2014 CMI, Inc.