Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD
Time of Inspection: 18:24 Date of Inspection: 06/20/2022
Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:33321080A2 Exp: 01/05/2024
0.000	0.048	0.076	0.193	0.078
0.000	0.048	0.076	0.194	0.078
0.000	0.048	0.076	0.194	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Carly Rosas #219

CARLY A ROSAS
Signature and Printed Name

06/20/2022
Date

E [signature] 197
PALM BAY PD

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD
Time of Inspection: 17:26 Date of Inspection: 08/09/2022
Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:33321080A2 Exp: 01/05/2024
0.000	0.048	0.075	0.197	0.078
0.000	0.049	0.076	0.199	0.078
0.000	0.048	0.076	0.198	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GEOFFERY E MCDOLE
Signature and Printed Name

08/09/2022
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD
Time of Inspection: 19:24 Date of Inspection: 08/31/2022
Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:
A F / M A:NOT ENOUGH MA USED. GM.164 08/31/2022
Agency Inspection redone

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GEOFFERY E MCDOLE
Signature and Printed Name

08/31/2022
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD
Time of Inspection: 20:06 Date of Inspection: 08/31/2022
Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:33321080A2 Exp: 01/05/2024
0.015 / 0.000	0.048	0.077	0.190	0.078
0.000 / 0.000	0.049	0.076	0.193	0.078
0.000 / 0.000	0.049	0.077	0.195	0.078

Number of Simulators Used: 5

Remarks:
00: CONTROL OUTSIDE TOLERANCE DUE TO RESIDUAL ALCOHOL VAPOR.

Test Redone GM1644 08/31/2022

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GEOFFERY E MCDOLE
Signature and Printed Name

08/31/2022
Date

FDLE/ATP Form 40 – March 2004

Initial Diagnostic

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
08/31/2022
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Alcohol Free Stability Test
0.00 g/210L

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
08/31/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:56
Control Test	0.000	18:57
Air Blank	0.000	18:57
Control Test	0.000	18:58
Air Blank	0.000	18:58
Control Test	0.000	18:59
Air Blank	0.000	18:59
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

EMC 164
Operator's Signature

0.05 g/210L stability Test

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
08/31/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	19:01
Control Test	0.048	19:02
Air Blank	0.000	19:02
Control Test	0.048	19:03
Air Blank	0.000	19:03
Control Test	0.047	19:04
Air Blank	0.000	19:05
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

Operator's Signature

0.08 g/210L Stability Test

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
08/31/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	19:06
Control Test	0.077	19:06
Air Blank	0.000	19:07
Control Test	0.078	19:08
Air Blank	0.000	19:08
Control Test	0.077	19:09
Air Blank	0.000	19:09
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

EMT 1164
Operator's Signature

0.20 g/210L Stability Test

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
08/31/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	19:11
Control Test	0.192	19:11
Air Blank	0.000	19:12
Control Test	0.195	19:13
Air Blank	0.000	19:13
Control Test	0.196	19:14
Air Blank	0.000	19:14
Control Test Stats		
Average	0.1943	
Std Dev	0.0021	
Rel Std Dev(%)	1.0712	

164

Operator's Signature

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD
Time of Inspection: 14:50

Date of Inspection: 09/18/2022

Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 08/17/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:202106B Exp: 06/22/2023
0.000	0.049	0.076	0.195	0.078
0.000	0.050	0.076	0.196	0.078
0.000	0.050	0.076	0.197	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Carly Rosas #219

CARLY A ROSAS

Signature and Printed Name

09/18/2022

Date

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD
Time of Inspection: 16:29 Date of Inspection: 10/23/2022
Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:33321080A2 Exp: 01/05/2024
0.000	0.049	0.077	0.195	0.078
0.000	0.048	0.077	0.195	0.078
0.000	0.048	0.077	0.196	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CarlyRosas #219 CARLY A ROSAS

Signature and Printed Name

10/23/2022
Date

FDLE/ATP Form 40 – March 2004

November Inspection
Attempt 1

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD
Time of Inspection: 18:24 Date of Inspection: 12/07/2022
Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: ____

Remarks:
DIAGNOSTIC FAILURENon-compliance:DIAGNOSTIC FAILURE AT INSPECTION GM164

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GM164 GEOFFERY E MCDOLE
Signature and Printed Name

12/07/2022
Date

November Inspection
Attempt 2

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD
Time of Inspection: 19:14 Date of Inspection: 12/07/2022
Serial Number: 80-000996
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:33321080A2 Exp: 01/05/2024
0.000	0.048	0.078	0.194	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GEOFFERY E MCDOLE
Signature and Printed Name

12/07/2022
Date

TITUSVILLE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000996
12/07/2022
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK