Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARES,RAMEZ,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 09/06/2025 01:10:39

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.081	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.212	01:05
Breath Volume		1.910	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.213	01:08
Breath Volume		2.089	01:08
Air Blank		0.000	01:09
Control Test		0.079	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 09/15/2025 13:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: RIVERA ALMENDARES

DATE/TIME OF TEST: 09/12/2025 23:02:27

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.081	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.110	22:57
Breath Volume		1.113	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.106	23:00
Breath Volume		1.402	23:00
Air Blank		0.000	23:01
Control Test		0.078	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 09/15/2025 13:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,COLIN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CALDWELL,COLIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: LEIVA MAIRENA

DATE/TIME OF TEST: 09/13/2025 02:35:51

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.080	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.121	02:31
Breath Volume		4.558	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.120	02:34
Breath Volume		4.863	02:34
Air Blank		0.000	02:34
Control Test		0.080	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 09/15/2025 13:42

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: BAUTISTA FUENTES

DATE/TIME OF TEST: 09/01/2025 00:40:02

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.139	00:34
Breath Volume		1.613	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:36
Subject Sample	# 2	0.144	00:38
Breath Volume		1.886	00:38
Air Blank		0.000	00:39
Control Test		0.078	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:40

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: VALLE CRUZ

DATE/TIME OF TEST: 09/01/2025 01:01:44

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.079	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.203	00:56
Breath Volume		1.886	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.202	01:00
Breath Volume		1.417	01:00
Air Blank		0.000	01:00
Control Test		0.076	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAULUS,CASEY,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PAULUS,CASEY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: MEJIA ROMERO

DATE/TIME OF TEST: 09/01/2025 04:45:49

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.079	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.141	04:41
Breath Volume		2.621	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:43
Subject Sample	# 2	0.132	04:44
Breath Volume		3.035	04:44
Air Blank		0.000	04:44
Control Test		0.076	04:45
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: VASQUEZ-BAUTISTA

DATE/TIME OF TEST: 09/01/2025 23:53:53

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.079	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.043	23:49
Breath Volume		3.546	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:51
Subject Sample	# 2	0.039	23:52
Breath Volume		2.175	23:52
Air Blank		0.000	23:52
Control Test		0.079	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELACRUZ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FARES,RAMEZ
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: SNODGRASS

DATE/TIME OF TEST: 09/02/2025 01:26:19

Diagnostics Check		OK	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.086	01:20
Breath Volume		1.390	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.080	01:24
Breath Volume		1.488	01:24
Air Blank		0.000	01:25
Control Test		0.077	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNUTSON,TRISTAN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KNUTSON,TRISTAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 09/05/2025 03:32:15

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.139	03:27
Breath Volume		1.773	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.131	03:30
Breath Volume		1.523	03:30
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: STEFENACK

DATE/TIME OF TEST: 09/05/2025 21:05:41

Diagnostics Check		OK	20:59
Air Blank		0.000	20:59
Control Test		0.080	21:00
Air Blank		0.000	21:00
Subject Sample	# 1	0.000	21:01
Breath Volume		2.812	21:01
Air Blank		0.000	21:01
Air Blank		0.000	21:03
Subject Sample	# 2	0.000	21:04
Breath Volume		3.980	21:04
Air Blank		0.000	21:04
Control Test		0.081	21:05
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELACRUZ,JONATHAN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DELACRUZ,JONATHAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 09/06/2025 02:25:35

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.081	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.131	02:20
Breath Volume		2.289	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.129	02:23
Breath Volume		1.925	02:23
Air Blank		0.000	02:24
Control Test		0.078	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: TELLEZ

DATE/TIME OF TEST: 09/06/2025 05:12:28

Diagnostics Check		OK	05:04
Air Blank		0.000	05:05
Control Test		0.080	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.160	05:07
Breath Volume		2.445	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:09
Subject Sample	# 2	0.163	05:10
Breath Volume		3.664	05:10
Air Blank		0.000	05:11
Control Test		0.079	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN],M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 09/07/2025 00:45:08

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.081	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.170	00:40
Breath Volume		3.160	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.168	00:43
Breath Volume		3.285	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,LOGAN,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: WHEELER,LOGAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ACKERSON

DATE/TIME OF TEST: 09/07/2025 21:32:30

Diagnostics Check		OK	21:25
Air Blank		0.000	21:26
Control Test		0.081	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.148	21:27
Breath Volume		3.746	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.144	21:30
Breath Volume		2.738	21:30
Air Blank		0.000	21:31
Control Test		0.078	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNUTSON,TRISTAN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KNUTSON,TRISTAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: BOGGS

DATE/TIME OF TEST: 09/09/2025 05:16:03

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.081	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.000	05:11
Breath Volume		3.199	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.000	05:14
Breath Volume		3.039	05:14
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:16

Data Download Date/Time: 09/15/2025 13:55

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HAUSER

DATE/TIME OF TEST: 09/09/2025 23:08:25

Diagnostics Check		OK	23:01
Air Blank		0.000	23:02
Control Test		0.081	23:02
Air Blank		0.000	23:03
Subject Sample	# 1	0.080	23:03
Breath Volume		2.765	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.077	23:06
Breath Volume		2.515	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 09/15/2025 13:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,ROBERT,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: STAPLETON,JACK
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 11:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: MCKANNA

DATE/TIME OF TEST: 09/10/2025 12:57:14

Diagnostics Check		OK	12:50
Air Blank		0.000	12:51
Control Test		0.080	12:51
Air Blank		0.000	12:52
Subject Sample	# 1	0.276	12:52
Breath Volume		2.316	12:52
Air Blank		0.000	12:53
Air Blank		0.000	12:54
Subject Sample	# 2	0.281	12:55
Breath Volume		1.863	12:55
Air Blank		0.000	12:56
Control Test		0.078	12:56
Air Blank		0.000	12:57
Diagnostics Check		OK	12:57

Data Download Date/Time: 09/15/2025 13:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNUTSON,TRISTAN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KNUTSON,TRISTAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BLOSKI

DATE/TIME OF TEST: 09/13/2025 00:18:23

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.200	00:13
Breath Volume		2.136	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.200	00:16
Breath Volume		1.703	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 09/15/2025 13:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LUMBRERAS

DATE/TIME OF TEST: 09/13/2025 02:36:40

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.197	02:31
Breath Volume		2.753	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.190	02:35
Breath Volume		2.968	02:35
Air Blank		0.000	02:35
Control Test		0.079	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 09/15/2025 13:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: DE LA RIVA GUZMAN

DATE/TIME OF TEST: 09/13/2025 21:47:35

Diagnostics Check		OK	21:39
Air Blank		0.000	21:40
Control Test		0.080	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.169	21:42
Breath Volume		2.375	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:44
Subject Sample	# 2	0.178	21:45
Breath Volume		1.175	21:45
Air Blank		0.000	21:46
Control Test		0.077	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 09/15/2025 13:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 09/14/2025 01:33:54

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:19	
Control Test		0.080	01:20	
Air Blank		0.000	01:20	
Subject Sample	# 1	0.071	01:23	Slope Not Level
Breath Volume		1.855	01:23	
Air Blank		0.000	01:24	
Air Blank		0.000	01:26	
Subject Sample	# 2	0.066	01:26	
Breath Volume		1.101	01:26	
Air Blank		0.000	01:27	
Air Blank		0.000	01:29	
Subject Sample	# 3	0.061	01:32	Volume Not Met
Breath Volume		0.937	01:32	
Air Blank		0.000	01:32	
Control Test		0.079	01:33	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 09/15/2025 13:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 09/14/2025 01:43:14

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.065	01:37
Breath Volume		1.148	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.064	01:41
Breath Volume		1.214	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 09/15/2025 13:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,JONATHAN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: TAYLOR,JONATHAN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: LOPEZ VAZQUEZ

DATE/TIME OF TEST: 09/14/2025 04:18:44

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.234	04:13
Breath Volume		1.125	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.236	04:16
Breath Volume		3.398	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 09/15/2025 13:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNUTSON,TRISTAN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BONNET,GREGORY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 08/14/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SALES DE SOUSA NETO

DATE/TIME OF TEST: 09/14/2025 22:19:22

Diagnostics Check		OK	22:12
Air Blank		0.000	22:13
Control Test		0.080	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.129	22:14
Breath Volume		1.367	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.136	22:17
Breath Volume		1.445	22:17
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 09/15/2025 13:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOGLUND,WALKER,D
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: NGUYEN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/03/2025 02:53:48

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.081	02:39	
Air Blank		0.000	02:39	
Subject Sample	# 1	0.122	02:43	Volume Not Met
Breath Volume		0.113	02:43	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 2	0.119	02:47	
Breath Volume		1.621	02:47	
Air Blank		0.000	02:47	
Air Blank		0.000	02:49	
Subject Sample	# 3	0.111	02:52	
Breath Volume		1.433	02:52	
Air Blank		0.000	02:52	
Control Test		0.078	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GREEN,SETH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:56

SUBJECT'S LAST NAME: WINCHESTER

DATE/TIME OF TEST: 09/03/2025 19:19:40

Diagnostics Check		OK	19:16	
Air Blank		0.000	19:17	
Control Test		0.080	19:17	
Air Blank		0.000	19:18	
Subject Sample	# 1		19:18	Subject Test Refused
Breath Volume		0.000	19:18	
Air Blank		0.000	19:18	
Control Test		0.080	19:19	
Air Blank		0.000	19:19	
Diagnostics Check		OK	19:19	

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REXFORD,MARK,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GIBBS,MICHAEL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 12:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 09/05/2025 13:47:14

Diagnostics Check		OK	13:40
Air Blank		0.000	13:41
Control Test		0.081	13:41
Air Blank		0.000	13:42
Subject Sample	# 1	0.000	13:42
Breath Volume		1.875	13:42
Air Blank		0.000	13:43
Air Blank		0.000	13:45
Subject Sample	# 2	0.000	13:45
Breath Volume		1.585	13:45
Air Blank		0.000	13:46
Control Test		0.080	13:46
Air Blank		0.000	13:47
Diagnostics Check		OK	13:47

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLS,WILLIAM,W
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: SNYDER,MICHAEL
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 09/07/2025 02:52:06

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.214	02:47
Breath Volume		1.640	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.214	02:50
Breath Volume		1.128	02:50
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFMAN,DYLAN,R
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: COFFMAN,DYLAN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: HOPPER

DATE/TIME OF TEST: 09/07/2025 03:56:50

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.114	03:51
Breath Volume		1.125	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:53
Subject Sample	# 2	0.104	03:55
Breath Volume		1.265	03:55
Air Blank		0.000	03:55
Control Test		0.080	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONG,J,A
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GUINUP,Z
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 16:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: WAHLER

DATE/TIME OF TEST: 09/09/2025 17:03:24

Diagnostics Check		OK	16:57
Air Blank		0.000	16:57
Control Test		0.080	16:57
Air Blank		0.000	16:58
Subject Sample	# 1	0.000	16:59
Breath Volume		1.273	16:59
Air Blank		0.000	16:59
Air Blank		0.000	17:01
Subject Sample	# 2	0.000	17:02
Breath Volume		1.878	17:02
Air Blank		0.000	17:02
Control Test		0.079	17:02
Air Blank		0.000	17:03
Diagnostics Check		OK	17:03

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SMITH,L
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: LEDUC

DATE/TIME OF TEST: 09/09/2025 20:01:35

Diagnostics Check		OK	19:58	
Air Blank		0.000	19:59	
Control Test		0.080	19:59	
Air Blank		0.000	20:00	
Subject Sample	# 1		20:00	Subject Test Refused
Breath Volume		0.000	20:00	
Air Blank		0.000	20:00	
Control Test		0.081	20:00	
Air Blank		0.000	20:01	
Diagnostics Check		OK	20:01	

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DOGGETT,SEAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: WALDHAUER

DATE/TIME OF TEST: 09/10/2025 00:46:07

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.163	00:41
Breath Volume		1.769	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.153	00:44
Breath Volume		1.777	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROME,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: FLORES,EMMANUEL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 09/10/2025 19:13:52

Diagnostics Check		OK	19:05	
Air Blank		0.000	19:05	
Control Test		0.081	19:06	
Air Blank		0.000	19:06	
Subject Sample	# 1		19:09	No Sample Provided
Breath Volume		0.000	19:09	
Air Blank		0.000	19:10	
Air Blank		0.000	19:12	
Subject Sample	# 2		19:12	Subject Test Refused
Breath Volume		0.000	19:12	
Air Blank		0.000	19:12	
Control Test		0.080	19:13	
Air Blank		0.000	19:13	
Diagnostics Check		OK	19:13	

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWERS,ABIGAYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: POWERS,ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A2
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: TIMPER

DATE/TIME OF TEST: 09/13/2025 20:46:02

Diagnostics Check		OK	20:39
Air Blank		0.000	20:40
Control Test		0.080	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.307	20:41
Breath Volume		1.996	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:43
Subject Sample	# 2	0.299	20:44
Breath Volume		1.644	20:44
Air Blank		0.000	20:45
Control Test		0.080	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:46

Data Download Date/Time: 09/16/2025 12:53

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIMA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: LIMA,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: URIAS

DATE/TIME OF TEST: 09/26/2025 05:11:09

Diagnostics Check		OK	05:04
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:05
Subject Sample	# 1	0.205	05:05
Breath Volume		2.515	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 2	0.199	05:09
Breath Volume		2.781	05:09
Air Blank		0.000	05:10
Control Test		0.080	05:10
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 10/28/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: DOMINICK,JOSEPH
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RIDDICK

DATE/TIME OF TEST: 09/28/2025 02:48:02

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.149	02:42
Breath Volume		1.855	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.150	02:46
Breath Volume		2.164	02:46
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:48

Data Download Date/Time: 10/28/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR,EVAN,M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: DAROVITZ,SHELBY
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: PALIZANO

DATE/TIME OF TEST: 09/28/2025 22:26:15

Diagnostics Check		OK	22:19
Air Blank		0.000	22:20
Control Test		0.080	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.170	22:21
Breath Volume		2.285	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.168	22:24
Breath Volume		2.066	22:24
Air Blank		0.000	22:25
Control Test		0.080	22:25
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 10/28/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROME,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LIPTAK,LAUREN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 09/29/2025 18:59:08

Diagnostics Check		OK	18:51
Air Blank		0.000	18:51
Control Test		0.081	18:51
Air Blank		0.000	18:52
Subject Sample	# 1	0.227	18:53
Breath Volume		1.992	18:53
Air Blank		0.000	18:54
Air Blank		0.000	18:56
Subject Sample	# 2	0.226	18:57
Breath Volume		2.246	18:57
Air Blank		0.000	18:58
Control Test		0.077	18:58
Air Blank		0.000	18:59
Diagnostics Check		OK	18:59

Data Download Date/Time: 10/28/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROME,NICHOLAS,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BLAKE,PAIGE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/30/2025 03:36:19

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.081	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.000	03:31
Breath Volume		2.406	03:31
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.000	03:34
Breath Volume		1.503	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 10/28/2025 16:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMUNNI,FRANCESCO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RAMUNNI,FRANCESCO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 09/02/2025 03:47:56

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.078	03:45	
Air Blank		0.000	03:46	
Subject Sample	# 1		03:46	Subject Test Refused
Breath Volume		0.000	03:46	
Air Blank		0.000	03:47	
Control Test		0.079	03:47	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 09/23/2025 11:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMUNNI,FRANCESCO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ROSARIO,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 09/10/2025 03:31:50

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.078	03:29	
Air Blank		0.000	03:30	
Subject Sample	# 1		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Control Test		0.079	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 09/23/2025 11:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVA,TIMO,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALMEIDA,TIMOTHY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: HYNDMAN

DATE/TIME OF TEST: 09/12/2025 19:05:40

Diagnostics Check		OK	18:57
Air Blank		0.000	18:57
Control Test		0.079	18:58
Air Blank		0.000	18:58
Subject Sample	# 1	0.162	19:00
Breath Volume		2.339	19:00
Air Blank		0.000	19:01
Air Blank		0.000	19:03
Subject Sample	# 2	0.164	19:03
Breath Volume		2.835	19:03
Air Blank		0.000	19:04
Control Test		0.079	19:05
Air Blank		0.000	19:05
Diagnostics Check		OK	19:05

Data Download Date/Time: 09/23/2025 11:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMUNNI,FRANCESCO,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: HARRISON,DANIELLE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RAK

DATE/TIME OF TEST: 09/15/2025 22:36:02

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.078	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.181	22:31
Breath Volume		2.164	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.176	22:34
Breath Volume		1.667	22:34
Air Blank		0.000	22:35
Control Test		0.078	22:35
Air Blank		0.000	22:35
Diagnostics Check		OK	22:36

Data Download Date/Time: 09/23/2025 11:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA RUA,BRYAN,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: LEONARD,SETH
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: GAYOSSO-MARQUEZ

DATE/TIME OF TEST: 09/21/2025 07:50:55

Diagnostics Check		OK	07:47	
Air Blank		0.000	07:47	
Control Test		0.078	07:48	
Air Blank		0.000	07:48	
Subject Sample	# 1		07:49	Subject Test Refused
Breath Volume		0.000	07:49	
Air Blank		0.000	07:50	
Control Test		0.078	07:50	
Air Blank		0.000	07:50	
Diagnostics Check		OK	07:50	

Data Download Date/Time: 09/23/2025 11:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,DANIELLE,M
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: TAYLOR,ZACHARY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BISSON

DATE/TIME OF TEST: 09/24/2025 03:22:43

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:20	
Control Test		0.078	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:21	Subject Test Refused
Breath Volume		0.000	03:21	
Air Blank		0.000	03:21	
Control Test		0.078	03:22	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANGER,MOOKIE,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: TAYLOR,ZACHARY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: LEVINE

DATE/TIME OF TEST: 09/25/2025 22:11:16

Diagnostics Check		OK	22:02	
Air Blank		0.000	22:03	
Control Test		0.077	22:03	
Air Blank		0.000	22:03	
Subject Sample	# 1		22:07	No Sample Provided
Breath Volume		0.000	22:07	
Air Blank		0.000	22:07	
Air Blank		0.000	22:09	
Subject Sample	# 2		22:09	Subject Test Refused
Breath Volume		0.000	22:09	
Air Blank		0.000	22:10	
Control Test		0.078	22:10	
Air Blank		0.000	22:11	
Diagnostics Check		OK	22:11	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALMEIDA,TIMOTHY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 09/28/2025 00:31:51

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:29	
Control Test		0.077	00:29	
Air Blank		0.000	00:30	
Subject Sample	# 1		00:30	Subject Test Refused
Breath Volume		0.000	00:30	
Air Blank		0.000	00:30	
Control Test		0.078	00:31	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DANIEL,X
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: RAMUNNI,FRANCESCO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: GRYGA

DATE/TIME OF TEST: 09/28/2025 00:46:53

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.078	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.169	00:42
Breath Volume		2.691	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:44
Subject Sample	# 2	0.171	00:45
Breath Volume		1.812	00:45
Air Blank		0.000	00:45
Control Test		0.076	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 10/28/2025 15:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUNEZ RIVA,PIERO,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: NUNEZ RIVA

DATE/TIME OF TEST: 09/01/2025 02:02:31

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:51	
Control Test		0.080	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.186	01:55	Volume Not Met
Breath Volume		0.031	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.175	02:00	Volume Not Met
Breath Volume		0.164	02:00	
Air Blank		0.000	02:01	
Control Test		0.078	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 09/30/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: CISNEROS ARGUETA

DATE/TIME OF TEST: 09/14/2025 03:33:59

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.144	03:29
Breath Volume		3.062	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:31
Subject Sample	# 2	0.146	03:32
Breath Volume		3.859	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 09/30/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/21/2025 03:51:37

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.081	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.253	03:44
Breath Volume		1.332	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:46
Subject Sample	# 2	0.251	03:49
Breath Volume		1.332	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 09/30/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICARDO,EDWARD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RICARDO,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: PAURA LOPEZ

DATE/TIME OF TEST: 09/21/2025 20:01:47

Diagnostics Check		OK	19:54
Air Blank		0.000	19:54
Control Test		0.081	19:54
Air Blank		0.000	19:55
Subject Sample	# 1	0.170	19:56
Breath Volume		1.527	19:56
Air Blank		0.000	19:57
Air Blank		0.000	19:59
Subject Sample	# 2	0.166	20:00
Breath Volume		2.472	20:00
Air Blank		0.000	20:00
Control Test		0.077	20:01
Air Blank		0.000	20:01
Diagnostics Check		OK	20:01

Data Download Date/Time: 09/30/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ROSAS

DATE/TIME OF TEST: 09/22/2025 02:48:51

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.070	02:44
Breath Volume		2.523	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:46
Subject Sample	# 2	0.065	02:47
Breath Volume		1.636	02:47
Air Blank		0.000	02:47
Control Test		0.081	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 09/30/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: FERTILIEN

DATE/TIME OF TEST: 09/22/2025 05:09:46

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.081	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.034	05:04
Breath Volume		2.058	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.033	05:08
Breath Volume		1.679	05:08
Air Blank		0.000	05:08
Control Test		0.081	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 09/30/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MORA

DATE/TIME OF TEST: 09/28/2025 22:38:25

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.080	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.166	22:34
Breath Volume		2.042	22:34
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.174	22:36
Breath Volume		1.703	22:36
Air Blank		0.000	22:37
Control Test		0.080	22:37
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 09/30/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICARDO,EDWARD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RICARDO,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: VARFOLOMEEV

DATE/TIME OF TEST: 09/29/2025 03:11:25

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.027	03:04
Breath Volume		1.382	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.027	03:09
Breath Volume		3.238	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 09/30/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ESCALANTE

DATE/TIME OF TEST: 09/30/2025 00:41:56

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.080	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.119	00:37
Breath Volume		2.589	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.122	00:40
Breath Volume		2.417	00:40
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 09/30/2025 17:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: RAGHUNANAN

DATE/TIME OF TEST: 09/07/2025 02:37:22

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.079	02:25	
Air Blank		0.000	02:25	
Subject Sample	# 1	0.000	02:27	
Breath Volume		1.222	02:27	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 2	0.000	02:33	Volume Not Met
Breath Volume		0.781	02:33	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 3		02:35	Subject Test Refused
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Control Test		0.079	02:36	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 09/15/2025 15:15

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: PALLON

DATE/TIME OF TEST: 09/19/2025 02:35:04

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:33	
Control Test		0.080	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 1		02:34	Sequence Aborted
Breath Volume		0.000	02:34	
Air Blank		0.000	02:35	

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PALLON

DATE/TIME OF TEST: 09/19/2025 02:51:52

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.108	02:47
Breath Volume		1.726	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.101	02:50
Breath Volume		1.550	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: BENCE

DATE/TIME OF TEST: 09/20/2025 01:30:56

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.079	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.209	01:25
Breath Volume		1.789	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.213	01:29
Breath Volume		2.722	01:29
Air Blank		0.000	01:29
Control Test		0.080	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: STARKEY

DATE/TIME OF TEST: 09/27/2025 03:42:33

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.079	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.126	03:37
Breath Volume		1.492	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.133	03:40
Breath Volume		2.187	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 09/28/2025 01:11:54

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.188	01:07
Breath Volume		3.839	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.201	01:10
Breath Volume		3.437	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: ESTRELLA

DATE/TIME OF TEST: 09/28/2025 05:12:08

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.079	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.179	05:07
Breath Volume		3.957	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:09
Subject Sample	# 2	0.175	05:10
Breath Volume		3.656	05:10
Air Blank		0.000	05:11
Control Test		0.077	05:11
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 11/13/2025 17:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 09/16/2025 01:48:50

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.079	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:47	Subject Test Refused
Breath Volume		0.000	01:47	
Air Blank		0.000	01:47	
Control Test		0.080	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:48	

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: UPCHURCH

DATE/TIME OF TEST: 09/16/2025 16:25:43

Diagnostics Check		OK	16:14	
Air Blank		0.000	16:14	
Control Test		0.079	16:15	
Air Blank		0.000	16:15	
Subject Sample	# 1		16:18	No Sample Provided
Breath Volume		0.000	16:18	
Air Blank		0.000	16:19	
Air Blank		0.000	16:21	
Subject Sample	# 2		16:24	No Sample Provided
Breath Volume		0.000	16:24	
Air Blank		0.000	16:24	
Control Test		0.080	16:25	
Air Blank		0.000	16:25	
Diagnostics Check		OK	16:25	

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: UPCHURCH

DATE/TIME OF TEST: 09/16/2025 16:57:43

Diagnostics Check		OK	16:51
Air Blank		0.000	16:51
Control Test		0.080	16:52
Air Blank		0.000	16:52
Subject Sample	# 1	0.000	16:53
Breath Volume		3.304	16:53
Air Blank		0.000	16:53
Air Blank		0.000	16:55
Subject Sample	# 2	0.000	16:56
Breath Volume		2.757	16:56
Air Blank		0.000	16:56
Control Test		0.079	16:57
Air Blank		0.000	16:57
Diagnostics Check		OK	16:57

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/17/2025 16:53:45

Diagnostics Check		OK	16:46
Air Blank		0.000	16:46
Control Test		0.079	16:46
Air Blank		0.000	16:47
Subject Sample	# 1	0.249	16:48
Breath Volume		1.785	16:48
Air Blank		0.000	16:49
Air Blank		0.000	16:51
Subject Sample	# 2	0.245	16:51
Breath Volume		1.945	16:51
Air Blank		0.000	16:52
Control Test		0.075	16:53
Air Blank		0.000	16:53
Diagnostics Check		OK	16:53

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 09/17/2025 22:54:10

Diagnostics Check		OK	22:47
Air Blank		0.000	22:48
Control Test		0.079	22:48
Air Blank		0.000	22:49
Subject Sample	# 1	0.221	22:49
Breath Volume		1.933	22:49
Air Blank		0.000	22:50
Air Blank		0.000	22:51
Subject Sample	# 2	0.217	22:52
Breath Volume		2.292	22:52
Air Blank		0.000	22:53
Control Test		0.078	22:53
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNUTSON,TRISTAN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KNUTSON,TRISTAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SIZEMORE

DATE/TIME OF TEST: 09/18/2025 02:31:52

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.205	02:26
Breath Volume		2.167	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.209	02:30
Breath Volume		2.292	02:30
Air Blank		0.000	02:30
Control Test		0.079	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNUTSON,TRISTAN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KNUTSON,TRISTAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 09/18/2025 21:07:15

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.079	21:00
Air Blank		0.000	21:01
Subject Sample	# 1	0.166	21:02
Breath Volume		1.457	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:04
Subject Sample	# 2	0.156	21:05
Breath Volume		1.136	21:05
Air Blank		0.000	21:06
Control Test		0.078	21:06
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 09/19/2025 02:37:36

Diagnostics Check		OK	02:30
Air Blank		0.000	02:30
Control Test		0.079	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.179	02:32
Breath Volume		1.804	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.170	02:35
Breath Volume		1.761	02:35
Air Blank		0.000	02:36
Control Test		0.077	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: CHANFRANTE

DATE/TIME OF TEST: 09/19/2025 03:08:47

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.079	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.110	03:04
Breath Volume		2.734	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.107	03:07
Breath Volume		2.222	03:07
Air Blank		0.000	03:07
Control Test		0.078	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUJO,MARCOS,V
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ARAUJO,MARCOS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 09/19/2025 04:30:54

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.078	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.113	04:26
Breath Volume		1.281	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.105	04:29
Breath Volume		1.664	04:29
Air Blank		0.000	04:29
Control Test		0.076	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/19/2025 23:06:41

Diagnostics Check		OK	22:56
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.124	22:59
Breath Volume		1.261	22:59
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.114	23:04
Breath Volume		1.445	23:04
Air Blank		0.000	23:05
Control Test		0.078	23:06
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATT,DEREK,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PLATT,DEREK
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 09/20/2025 06:53:16

Diagnostics Check		OK	06:46
Air Blank		0.000	06:46
Control Test		0.079	06:47
Air Blank		0.000	06:47
Subject Sample	# 1	0.124	06:48
Breath Volume		3.324	06:48
Air Blank		0.000	06:48
Air Blank		0.000	06:50
Subject Sample	# 2	0.127	06:51
Breath Volume		4.820	06:51
Air Blank		0.000	06:52
Control Test		0.078	06:52
Air Blank		0.000	06:53
Diagnostics Check		OK	06:53

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,AUSTIN,W
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COOK,AUSTIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: DE LEON VASQUEZ

DATE/TIME OF TEST: 09/21/2025 02:20:42

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.184	02:15
Breath Volume		1.152	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:17
Subject Sample	# 2	0.178	02:18
Breath Volume		1.132	02:18
Air Blank		0.000	02:19
Control Test		0.077	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 09/23/2025 16:42:46

Diagnostics Check		OK	16:36
Air Blank		0.000	16:37
Control Test		0.080	16:37
Air Blank		0.000	16:37
Subject Sample	# 1	0.000	16:38
Breath Volume		2.589	16:38
Air Blank		0.000	16:38
Air Blank		0.000	16:40
Subject Sample	# 2	0.000	16:41
Breath Volume		2.785	16:41
Air Blank		0.000	16:41
Control Test		0.078	16:42
Air Blank		0.000	16:42
Diagnostics Check		OK	16:42

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDWELL,COLIN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 09/24/2025 04:05:32

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.102	04:00
Breath Volume		1.828	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.107	04:03
Breath Volume		2.351	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY,QUINN,D
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: COTTOM,CALEB
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MILEK

DATE/TIME OF TEST: 09/24/2025 23:40:53

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.180	23:36
Breath Volume		2.628	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:38
Subject Sample	# 2	0.166	23:39
Breath Volume		2.722	23:39
Air Blank		0.000	23:39
Control Test		0.079	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMANDL,REAGHN,S
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COBB,JAMEICA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 09/25/2025 00:56:33

Diagnostics Check		OK	00:44	
Air Blank		0.000	00:44	
Control Test		0.080	00:45	
Air Blank		0.000	00:45	
Subject Sample	# 1	0.117	00:48	Volume Not Met
Breath Volume		0.308	00:48	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 2	0.098	00:54	Volume Not Met
Breath Volume		0.250	00:54	
Air Blank		0.000	00:55	
Control Test		0.075	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ZELAYA VINDEL

DATE/TIME OF TEST: 09/26/2025 04:01:09

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.171	03:56
Breath Volume		1.179	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:58
Subject Sample	# 2	0.154	03:59
Breath Volume		1.601	03:59
Air Blank		0.000	04:00
Control Test		0.077	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPRINGMAN,JASON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: ESTEBAN MAS

DATE/TIME OF TEST: 09/26/2025 23:53:37

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.080	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.215	23:49
Breath Volume		2.230	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:51
Subject Sample	# 2	0.208	23:51
Breath Volume		2.046	23:51
Air Blank		0.000	23:52
Control Test		0.077	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: ALANIA

DATE/TIME OF TEST: 09/27/2025 02:43:15

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.080	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:35	No Sample Provided
Breath Volume		0.000	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 2	0.087	02:41	Volume Not Met
Breath Volume		0.503	02:41	
Air Blank		0.000	02:42	
Control Test		0.081	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 09/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: ALANIA

DATE/TIME OF TEST: 09/27/2025 02:48:18

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:44	
Control Test		0.077	02:45	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:46	Subject Test Refused
Breath Volume		1.078	02:46	
Air Blank		0.000	02:47	
Control Test		0.079	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 10/17/2025 14:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBOSIEWICZ,ANTHONY,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2402649340
DRY GAS EXPIRY DATE: 01/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: HADZALIC

DATE/TIME OF TEST: 09/20/2025 02:31:54

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.078	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.023	02:27
Breath Volume		2.578	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:29
Subject Sample	# 2	0.023	02:30
Breath Volume		2.562	02:30
Air Blank		0.000	02:30
Control Test		0.079	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 10/29/2025 21:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HRADECKY,ALEXANDRIA,N
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARIN,MARIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 09/02/2025 02:09:55

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.000	02:05
Breath Volume		1.234	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.000	02:08
Breath Volume		2.156	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 09/16/2025 14:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LEON GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 09/03/2025 22:19:54

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.079	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.192	22:14
Breath Volume		1.480	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.204	22:18
Breath Volume		2.394	22:18
Air Blank		0.000	22:18
Control Test		0.078	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 09/16/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,SHANE,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RANCEL,RANDY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: INCLAINOCEGUERA

DATE/TIME OF TEST: 09/05/2025 03:04:57

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:02	
Control Test		0.078	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:03	Subject Test Refused
Breath Volume		0.000	03:03	
Air Blank		0.000	03:04	
Control Test		0.079	03:04	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 09/16/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: CATO

DATE/TIME OF TEST: 09/07/2025 01:41:35

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.172	01:36
Breath Volume		2.433	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.169	01:40
Breath Volume		1.933	01:40
Air Blank		0.000	01:40
Control Test		0.078	01:40
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 09/16/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUFIN,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: RUFIN,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: BORRERO GRANADO

DATE/TIME OF TEST: 09/08/2025 15:36:33

Diagnostics Check		OK	15:33	
Air Blank		0.000	15:34	
Control Test		0.079	15:34	
Air Blank		0.000	15:35	
Subject Sample	# 1		15:35	Subject Test Refused
Breath Volume		0.000	15:35	
Air Blank		0.000	15:35	
Control Test		0.079	15:35	
Air Blank		0.000	15:36	
Diagnostics Check		OK	15:36	

Data Download Date/Time: 09/16/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCINTYRE,DELLVON,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SIMPSON,JAMES
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: ZARESK

DATE/TIME OF TEST: 09/08/2025 22:06:54

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.079	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.053	22:02
Breath Volume		1.289	22:02
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.050	22:05
Breath Volume		1.464	22:05
Air Blank		0.000	22:05
Control Test		0.079	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 09/16/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MCINTYRE,DELLVON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: STARCHER

DATE/TIME OF TEST: 09/10/2025 03:49:03

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.079	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:48	
Control Test		0.079	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 09/16/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LEON GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SCHOFIELD

DATE/TIME OF TEST: 09/14/2025 03:10:13

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.079	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Control Test		0.080	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 09/16/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,JORGE,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MORENO,JORGE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/15/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 09/15/2025 21:55:51

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:53	
Control Test		0.079	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:54	Subject Test Refused
Breath Volume		0.000	21:54	
Air Blank		0.000	21:54	
Control Test		0.078	21:55	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:55	

Data Download Date/Time: 09/16/2025 14:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,OSCAR,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARIN,MARIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: LOVING

DATE/TIME OF TEST: 09/16/2025 21:46:34

Diagnostics Check		OK	21:43	
Air Blank		0.000	21:44	
Control Test		0.079	21:44	
Air Blank		0.000	21:45	
Subject Sample	# 1		21:45	Subject Test Refused
Breath Volume		0.000	21:45	
Air Blank		0.000	21:45	
Control Test		0.078	21:45	
Air Blank		0.000	21:46	
Diagnostics Check		OK	21:46	

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/19/2025 02:24:46

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:22	
Control Test		0.079	02:22	
Air Blank		0.000	02:22	
Subject Sample	# 1		02:23	Subject Test Refused
Breath Volume		0.000	02:23	
Air Blank		0.000	02:23	
Control Test		0.079	02:24	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDON DEL RIO,RODOLFO,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RONDON DEL RIO,RODOLFO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: ESQUIVEL

DATE/TIME OF TEST: 09/19/2025 21:44:53

Diagnostics Check		OK	21:41	
Air Blank		0.000	21:42	
Control Test		0.079	21:42	
Air Blank		0.000	21:43	
Subject Sample	# 1		21:43	Subject Test Refused
Breath Volume		0.000	21:43	
Air Blank		0.000	21:43	
Control Test		0.079	21:44	
Air Blank		0.000	21:44	
Diagnostics Check		OK	21:44	

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDON DEL RIO,RODOLFO,E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RONDON DEL RIO,RODOLFO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: ESQUIVEL

DATE/TIME OF TEST: 09/19/2025 22:11:38

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:09	
Control Test		0.079	22:09	
Air Blank		0.000	22:10	
Subject Sample	# 1		22:10	Subject Test Refused
Breath Volume		0.000	22:10	
Air Blank		0.000	22:10	
Control Test		0.078	22:11	
Air Blank		0.000	22:11	
Diagnostics Check		OK	22:11	

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 09/19/2025 23:42:33

Diagnostics Check		OK	23:36
Air Blank		0.000	23:36
Control Test		0.078	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.089	23:37
Breath Volume		2.003	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.086	23:41
Breath Volume		2.527	23:41
Air Blank		0.000	23:41
Control Test		0.078	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,OSCAR,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MARIN,MARIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: ZAVALETA DAVILA

DATE/TIME OF TEST: 09/20/2025 05:08:37

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:06	
Control Test		0.079	05:06	
Air Blank		0.000	05:07	
Subject Sample	# 1		05:07	Subject Test Refused
Breath Volume		0.000	05:07	
Air Blank		0.000	05:07	
Control Test		0.079	05:08	
Air Blank			05:08	RFI Detect
Air Blank		0.000	05:08	

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LEON GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/23/2025 21:31:01

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:28	
Control Test		0.079	21:28	
Air Blank		0.000	21:29	
Subject Sample	# 1		21:29	Subject Test Refused
Breath Volume		0.000	21:29	
Air Blank		0.000	21:30	
Control Test		0.079	21:30	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:31	

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,SHANE,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CASTRO,SHANE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402558895
DRY GAS EXPIRY DATE: 10/03/2025
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: THOMAN

DATE/TIME OF TEST: 09/26/2025 00:25:25

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:23	
Control Test		0.079	00:23	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:24	Subject Test Refused
Breath Volume		0.000	00:24	
Air Blank		0.000	00:24	
Control Test		0.079	00:24	
Air Blank		0.000	00:25	
Diagnostics Check		OK	00:25	

Data Download Date/Time: 10/17/2025 14:34

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATHIEU,REGINALD,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MATHIEU,REGINALD
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DEL SOLAR REYES

DATE/TIME OF TEST: 09/03/2025 03:25:01

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.083	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.143	03:20
Breath Volume		2.707	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 2	0.144	03:23
Breath Volume		2.613	03:23
Air Blank		0.000	03:24
Control Test		0.078	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:25

Data Download Date/Time: 09/30/2025 17:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,KAREN,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FERNANDEZ,KAREN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: SUAZO

DATE/TIME OF TEST: 09/03/2025 18:25:23

Diagnostics Check		OK	18:22	
Air Blank		0.000	18:22	
Control Test		0.082	18:22	
Air Blank		0.000	18:23	
Subject Sample	# 1		18:24	Subject Test Refused
Breath Volume		0.000	18:24	
Air Blank		0.000	18:24	
Control Test		0.082	18:24	
Air Blank		0.000	18:25	
Diagnostics Check		OK	18:25	

Data Download Date/Time: 09/30/2025 17:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN,EMMANUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ,FERNANDO
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 09/06/2025 11:10:40

Diagnostics Check		OK	11:01
Air Blank		0.000	11:02
Control Test		0.082	11:02
Air Blank		0.000	11:02
Subject Sample	# 1	0.079	11:05
Breath Volume		2.109	11:05
Air Blank		0.000	11:05
Air Blank		0.000	11:07
Subject Sample	# 2	0.080	11:09
Breath Volume		2.484	11:09
Air Blank		0.000	11:09
Control Test		0.078	11:10
Air Blank		0.000	11:10
Diagnostics Check		OK	11:10

Data Download Date/Time: 09/30/2025 17:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTHUR,MELVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 09/13/2025 01:08:59

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:02	
Control Test		0.082	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1	0.166	01:03	
Breath Volume		2.062	01:03	
Air Blank		0.000	01:04	
Air Blank		0.000	01:06	
Subject Sample	# 2	0.169	01:07	
Breath Volume		2.316	01:07	
Air Blank		0.000	01:08	
Control Test		0.086	01:08	Control Outside Tolerance
Air Blank		0.000	01:08	

Data Download Date/Time: 09/30/2025 17:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVIL,LUXERNE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 09/13/2025 01:40:19

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.082	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.162	01:35
Breath Volume		2.023	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.158	01:38
Breath Volume		1.929	01:38
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVIL,LUXERNE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PREVIL,LUXERNE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: TABARES

DATE/TIME OF TEST: 09/16/2025 03:09:59

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.081	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.305	03:05
Breath Volume		3.355	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:07
Subject Sample	# 2	0.296	03:08
Breath Volume		2.910	03:08
Air Blank		0.000	03:09
Control Test		0.082	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVIL,LUXERNE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/19/2025 04:03:36

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.082	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.339	03:58
Breath Volume		1.292	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.354	04:01
Breath Volume		2.085	04:01
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: YUVASHEVA

DATE/TIME OF TEST: 09/19/2025 23:20:21

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.081	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.281	23:15
Breath Volume		1.898	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.270	23:18
Breath Volume		1.488	23:18
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ MORALES,JAVIER,D
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PEREZ MORALES,JAVIER
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/21/2025 02:29:12

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.083	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.146	02:24
Breath Volume		1.355	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.144	02:27
Breath Volume		1.406	02:27
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/21/2025 05:20:57

Diagnostics Check		OK	05:14
Air Blank		0.000	05:14
Control Test		0.081	05:15
Air Blank		0.000	05:15
Subject Sample	# 1	0.182	05:16
Breath Volume		1.554	05:16
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.184	05:19
Breath Volume		2.035	05:19
Air Blank		0.000	05:20
Control Test		0.079	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANTHONY,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: NAVARRO MELENDEZ

DATE/TIME OF TEST: 09/27/2025 03:22:25

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.082	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.113	03:17
Breath Volume		1.378	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.108	03:20
Breath Volume		1.386	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,JOSE,F
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DOMINGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: HERNANDEZ DIAZ

DATE/TIME OF TEST: 09/27/2025 14:44:13

Diagnostics Check		OK	14:36
Air Blank		0.000	14:36
Control Test		0.083	14:37
Air Blank		0.000	14:37
Subject Sample	# 1	0.196	14:39
Breath Volume		2.261	14:39
Air Blank		0.000	14:40
Air Blank		0.000	14:42
Subject Sample	# 2	0.197	14:42
Breath Volume		1.316	14:42
Air Blank		0.000	14:43
Control Test		0.079	14:43
Air Blank		0.000	14:44
Diagnostics Check		OK	14:44

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICARDO,EDWARD,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERA,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ANDRE

DATE/TIME OF TEST: 09/28/2025 04:29:33

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.082	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.180	04:24
Breath Volume		1.320	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.174	04:27
Breath Volume		1.160	04:27
Air Blank		0.000	04:28
Control Test		0.078	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006736

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,KAREN,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ARTHUR,MELVIN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: GRANJA GAITAN

DATE/TIME OF TEST: 09/28/2025 05:00:04

Diagnostics Check		OK	04:53
Air Blank		0.000	04:54
Control Test		0.081	04:54
Air Blank		0.000	04:55
Subject Sample	# 1	0.156	04:55
Breath Volume		1.394	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:57
Subject Sample	# 2	0.151	04:58
Breath Volume		1.359	04:58
Air Blank		0.000	04:59
Control Test		0.081	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	05:00

Data Download Date/Time: 09/30/2025 17:28

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SHUMATE,TODD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 09/17/2025 16:52:53

Diagnostics Check		OK	16:49	
Air Blank		0.000	16:49	
Control Test		0.080	16:50	
Air Blank		0.000	16:50	
Subject Sample	# 1		16:51	Interferent Detect
Breath Volume		3.277	16:51	
Air Blank		0.000	16:51	
Control Test		0.082	16:52	
Air Blank		0.000	16:52	
Diagnostics Check		OK	16:52	

Data Download Date/Time: 09/22/2025 19:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SHUMATE,TODD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 09/17/2025 17:01:05

Diagnostics Check		OK	16:54
Air Blank		0.000	16:55
Control Test		0.081	16:55
Air Blank		0.000	16:56
Subject Sample	# 1	0.000	16:56
Breath Volume		3.246	16:56
Air Blank		0.000	16:57
Air Blank		0.000	16:59
Subject Sample	# 2	0.000	16:59
Breath Volume		3.125	16:59
Air Blank		0.000	17:00
Control Test		0.084	17:00
Air Blank		0.000	17:00
Diagnostics Check		OK	17:01

Data Download Date/Time: 09/22/2025 19:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ETTMAN,KRISTOFER,C
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: MCGEHEE,NICHOLAS
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08425080A1
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: BON CZEK

DATE/TIME OF TEST: 09/01/2025 04:48:05

Diagnostics Check	OK	04:46	
Air Blank	0.000	04:47	
Control Test	0.026	04:47	RFI Detect
Air Blank	0.000	04:48	

Data Download Date/Time: 09/28/2025 03:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,PATRICK,K
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: STEPHENS,JONATHAN
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A1
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: LUGO YNOA

DATE/TIME OF TEST: 09/06/2025 21:13:47

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.079	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.095	21:08
Breath Volume		1.808	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:10
Subject Sample	# 2	0.087	21:11
Breath Volume		1.152	21:11
Air Blank		0.000	21:12
Control Test		0.076	21:13
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 09/28/2025 03:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ERIC,C
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: STEPHENS,JONATHON
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A1
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: CUZOVOV

DATE/TIME OF TEST: 09/16/2025 21:11:17

Diagnostics Check		OK	21:03	
Air Blank		0.000	21:03	
Control Test		0.079	21:04	
Air Blank		0.000	21:04	
Subject Sample	# 1	0.196	21:05	
Breath Volume		1.230	21:05	
Air Blank		0.000	21:05	
Air Blank		0.000	21:07	
Subject Sample	# 2	0.201	21:09	
Breath Volume		2.609	21:09	
Air Blank		0.000	21:10	
Control Test		0.070	21:10	Control Outside Tolerance
Air Blank			21:11	Purge Fail

Data Download Date/Time: 09/28/2025 03:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANGAS,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: MCGEHEE,NICHOLAS
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A1
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: KANGAS

DATE/TIME OF TEST: 09/21/2025 06:34:47

Diagnostics Check	OK	06:34	
Air Blank		06:34	RFI Detect
Air Blank	0.000	06:34	

Data Download Date/Time: 09/28/2025 03:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHAS,SAMSON,P
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: MCGEHEE,NICHOLAS
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 08425080A1
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: KANGAS

DATE/TIME OF TEST: 09/21/2025 06:46:14

Diagnostics Check	OK	06:44	
Air Blank	0.000	06:44	
Control Test	0.079	06:44	
Air Blank		06:45	RFI Detect
Air Blank		06:46	Purge Fail

Data Download Date/Time: 09/28/2025 03:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

INSTRUMENT REGISTERED TO: KENNEDY SPACE CENTER
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHAS,SAMSON,P
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: MCGEHEE,NICHOLAS
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 08425080A1
DRY GAS EXPIRY DATE: 05/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: KANGAS

DATE/TIME OF TEST: 09/21/2025 06:55:31

Diagnostics Check	OK	06:54	
Air Blank	0.000	06:54	
Control Test	0.007	06:54	Control Outside Tolerance
Air Blank	0.000	06:55	

Data Download Date/Time: 09/28/2025 03:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENSON,JAY,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENSON,JAY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/28/2025 03:08:32

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.263	02:57
Breath Volume		1.457	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.235	03:01
Breath Volume		2.343	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:03
Subject Sample	# 3	0.240	03:06
Breath Volume		2.425	03:06
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 09/29/2025 19:59

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,RYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,RYAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 09/18/2025 02:39:05

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:26	
Control Test		0.080	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:30	No Sample Provided
Breath Volume		0.000	02:30	
Air Blank		0.000	02:30	
Air Blank		0.000	02:32	
Subject Sample	# 2	0.360	02:34	No .020 agreement
Breath Volume		1.699	02:34	
Air Blank		0.000	02:34	
Air Blank		0.000	02:36	
Subject Sample	# 3	0.314	02:37	No .020 agreement
Breath Volume		1.792	02:37	
Air Blank		0.000	02:38	
Control Test		0.078	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 09/25/2025 11:08

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/21/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 09/19/2025 01:48:08

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:39	
Control Test		0.079	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1	0.171	01:42	Volume Not Met
Breath Volume		0.585	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2		01:47	Subject Test Refused
Breath Volume		0.128	01:47	
Air Blank			01:47	Sequence Aborted
Air Blank		0.000	01:48	

Data Download Date/Time: 09/25/2025 11:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006761

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOPE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KOPE,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402590345
DRY GAS EXPIRY DATE: 11/04/2025
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: RIDDLE

DATE/TIME OF TEST: 09/28/2025 02:41:56

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.079	02:35
Air Blank		0.000	02:36
Subject Sample	# 1	0.035	02:37
Breath Volume		2.363	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:39
Subject Sample	# 2	0.037	02:40
Breath Volume		1.109	02:40
Air Blank		0.000	02:40
Control Test		0.078	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 10/17/2025 09:41

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: BALLADARES CALERO

DATE/TIME OF TEST: 09/20/2025 01:14:31

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.077	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.173	01:09
Breath Volume		1.289	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.175	01:12
Breath Volume		1.332	01:12
Air Blank		0.000	01:13
Control Test		0.076	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 10/22/2025 17:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAPELA,JORDAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CAPELA,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135736
DRY GAS EXPIRY DATE: 08/29/2027
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: COLCORD

DATE/TIME OF TEST: 09/21/2025 03:11:43

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.077	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.080	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 10/22/2025 17:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POLO,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: OSPINA

DATE/TIME OF TEST: 09/01/2025 03:47:23

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.095	03:42
Breath Volume		1.156	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.104	03:45
Breath Volume		2.382	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 09/03/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006924

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 08/06/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/01/2025 04:17:18

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.078	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.282	04:12
Breath Volume		2.789	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.270	04:15
Breath Volume		2.691	04:15
Air Blank		0.000	04:16
Control Test		0.077	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 09/03/2025 09:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006938

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,JAMES,W
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 09/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1474747
DRY GAS EXPIRY DATE: 04/12/2024
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/23/2025 23:35:01

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.078	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.172	23:30
Breath Volume		2.058	23:30
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.172	23:33
Breath Volume		2.378	23:33
Air Blank		0.000	23:34
Control Test		0.078	23:34
Air Blank		0.000	23:34
Diagnostics Check		OK	23:35

Data Download Date/Time: 12/02/2025 12:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006939

INSTRUMENT REGISTERED TO: FORT LAUDERDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEATING,KAYLA,C
BREATH TEST OPERATOR AGENCY: Fort Lauderdale PD
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: SIGLER

DATE/TIME OF TEST: 09/13/2025 00:06:25

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.078	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.224	00:01
Breath Volume		1.785	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.244	00:04
Breath Volume		1.574	00:04
Air Blank		0.000	00:05
Control Test		0.077	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 09/27/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/04/2025 23:47:08

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.020	23:42
Breath Volume		1.285	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.021	23:45
Breath Volume		1.441	23:45
Air Blank		0.000	23:46
Control Test		0.081	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GIL,RAMON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: NUNEZ SERRATO

DATE/TIME OF TEST: 09/05/2025 01:06:10

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.209	01:01
Breath Volume		1.753	01:01
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.205	01:04
Breath Volume		3.343	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MINIHAN,O
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: JAIJAIRAM

DATE/TIME OF TEST: 09/05/2025 19:30:28

Diagnostics Check		OK	19:22
Air Blank		0.000	19:23
Control Test		0.080	19:23
Air Blank		0.000	19:24
Subject Sample	# 1	0.203	19:25
Breath Volume		2.542	19:25
Air Blank		0.000	19:26
Air Blank		0.000	19:28
Subject Sample	# 2	0.193	19:28
Breath Volume		1.164	19:28
Air Blank		0.000	19:29
Control Test		0.080	19:29
Air Blank		0.000	19:30
Diagnostics Check		OK	19:30

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 09/06/2025 07:52:37

Diagnostics Check		OK	07:45
Air Blank		0.000	07:45
Control Test		0.080	07:46
Air Blank		0.000	07:46
Subject Sample	# 1	0.150	07:47
Breath Volume		1.324	07:47
Air Blank		0.000	07:48
Air Blank		0.000	07:50
Subject Sample	# 2	0.149	07:51
Breath Volume		1.441	07:51
Air Blank		0.000	07:51
Control Test		0.081	07:52
Air Blank		0.000	07:52
Diagnostics Check		OK	07:52

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: DETATA

DATE/TIME OF TEST: 09/06/2025 22:06:44

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.081	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.000	22:02
Breath Volume		1.718	22:02
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.000	22:05
Breath Volume		1.179	22:05
Air Blank		0.000	22:05
Control Test		0.081	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PAEZ,ANTHONY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RICKETTS

DATE/TIME OF TEST: 09/07/2025 03:40:30

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.184	03:35
Breath Volume		1.351	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.180	03:38
Breath Volume		1.824	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LITTLE,HEATHER
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/08/2025 03:18:04

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.225	03:13
Breath Volume		1.578	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.230	03:16
Breath Volume		1.867	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:18

Data Download Date/Time: 10/01/2025 10:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 09/08/2025 04:03:58

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.125	03:58
Breath Volume		1.714	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.140	04:02
Breath Volume		1.515	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HIPPOLYTE,VENEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: BRISSON

DATE/TIME OF TEST: 09/08/2025 06:12:49

Diagnostics Check		OK	06:06
Air Blank		0.000	06:07
Control Test		0.081	06:07
Air Blank		0.000	06:07
Subject Sample	# 1	0.175	06:08
Breath Volume		1.578	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:10
Subject Sample	# 2	0.173	06:11
Breath Volume		1.597	06:11
Air Blank		0.000	06:11
Control Test		0.080	06:12
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DUQUE,BRYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 11:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:47

SUBJECT'S LAST NAME: LEONOV

DATE/TIME OF TEST: 09/09/2025 13:20:51

Diagnostics Check		OK	13:13
Air Blank		0.000	13:13
Control Test		0.080	13:14
Air Blank		0.000	13:14
Subject Sample	# 1	0.015	13:15
Breath Volume		2.347	13:15
Air Blank		0.000	13:15
Air Blank		0.000	13:17
Subject Sample	# 2	0.014	13:19
Breath Volume		1.136	13:19
Air Blank		0.000	13:19
Control Test		0.082	13:20
Air Blank		0.000	13:20
Diagnostics Check		OK	13:20

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: KOPYLOV

DATE/TIME OF TEST: 09/10/2025 00:41:20

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.080	00:30	
Air Blank		0.000	00:30	
Subject Sample	# 1	0.151	00:34	Volume Not Met
Breath Volume		0.464	00:34	
Air Blank		0.000	00:34	
Air Blank		0.000	00:36	
Subject Sample	# 2	0.191	00:36	
Breath Volume		1.363	00:36	
Air Blank		0.000	00:37	
Air Blank		0.000	00:39	
Subject Sample	# 3	0.199	00:39	
Breath Volume		1.957	00:39	
Air Blank		0.000	00:40	
Control Test		0.079	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOPEZ,BRYAN
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: TOLENTINO HILARES

DATE/TIME OF TEST: 09/10/2025 02:20:46

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.160	02:16
Breath Volume		1.156	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:18
Subject Sample	# 2	0.164	02:19
Breath Volume		1.890	02:19
Air Blank		0.000	02:19
Control Test		0.078	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 10/01/2025 10:32

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: PINEDA LEMUS

DATE/TIME OF TEST: 09/12/2025 06:10:09

Diagnostics Check		OK	06:02
Air Blank		0.000	06:02
Control Test		0.080	06:03
Air Blank		0.000	06:03
Subject Sample	# 1	0.206	06:05
Breath Volume		2.101	06:05
Air Blank		0.000	06:06
Air Blank		0.000	06:07
Subject Sample	# 2	0.217	06:08
Breath Volume		2.386	06:08
Air Blank		0.000	06:09
Control Test		0.078	06:09
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

Data Download Date/Time: 10/01/2025 10:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POLO,ANDREW
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 09/13/2025 23:44:45

Diagnostics Check		OK	23:36
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.152	23:39
Breath Volume		1.718	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.155	23:43
Breath Volume		1.449	23:43
Air Blank		0.000	23:43
Control Test		0.080	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 10/01/2025 10:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BARBERIO,MADISON
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PELAEZ

DATE/TIME OF TEST: 09/14/2025 03:46:13

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.134	03:41
Breath Volume		2.628	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.123	03:44
Breath Volume		1.214	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 10/01/2025 10:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOPEZ,SAMANTHA
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: COBO JARAMILLO

DATE/TIME OF TEST: 09/16/2025 02:06:12

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.081	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.121	02:00
Breath Volume		1.402	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.113	02:04
Breath Volume		1.246	02:04
Air Blank		0.000	02:05
Control Test		0.080	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 10/01/2025 10:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: OCAL,BERK
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: HARTMANN

DATE/TIME OF TEST: 09/20/2025 08:29:01

Diagnostics Check		OK	08:20
Air Blank		0.000	08:21
Control Test		0.081	08:21
Air Blank		0.000	08:21
Subject Sample	# 1	0.232	08:24
Breath Volume		2.144	08:24
Air Blank		0.000	08:25
Air Blank		0.000	08:26
Subject Sample	# 2	0.241	08:27
Breath Volume		2.675	08:27
Air Blank		0.000	08:28
Control Test		0.079	08:28
Air Blank		0.000	08:28
Diagnostics Check		OK	08:29

Data Download Date/Time: 10/01/2025 10:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BRADOR,GIANLUIS
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:39

SUBJECT'S LAST NAME: CHALAS SANTOS

DATE/TIME OF TEST: 09/20/2025 09:11:08

Diagnostics Check		OK	09:04
Air Blank		0.000	09:05
Control Test		0.080	09:05
Air Blank		0.000	09:05
Subject Sample	# 1	0.187	09:06
Breath Volume		1.621	09:06
Air Blank		0.000	09:07
Air Blank		0.000	09:09
Subject Sample	# 2	0.186	09:09
Breath Volume		1.410	09:09
Air Blank		0.000	09:10
Control Test		0.080	09:10
Air Blank		0.000	09:11
Diagnostics Check		OK	09:11

Data Download Date/Time: 10/01/2025 10:33

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 07:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 09/20/2025 09:54:07

Diagnostics Check		OK	09:46
Air Blank		0.000	09:47
Control Test		0.080	09:47
Air Blank		0.000	09:48
Subject Sample	# 1	0.076	09:49
Breath Volume		1.250	09:49
Air Blank		0.000	09:49
Air Blank		0.000	09:51
Subject Sample	# 2	0.076	09:52
Breath Volume		1.164	09:52
Air Blank		0.000	09:53
Control Test		0.080	09:53
Air Blank		0.000	09:54
Diagnostics Check		OK	09:54

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MONESTIME

DATE/TIME OF TEST: 09/22/2025 04:14:41

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.234	04:09
Breath Volume		1.832	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.236	04:13
Breath Volume		1.230	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BELONY,ANATUS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: KUPINSKI

DATE/TIME OF TEST: 09/22/2025 17:10:50

Diagnostics Check		OK	17:03
Air Blank		0.000	17:04
Control Test		0.081	17:04
Air Blank		0.000	17:04
Subject Sample	# 1	0.025	17:05
Breath Volume		2.199	17:05
Air Blank		0.000	17:06
Air Blank		0.000	17:08
Subject Sample	# 2	0.023	17:09
Breath Volume		1.878	17:09
Air Blank		0.000	17:09
Control Test		0.082	17:10
Air Blank		0.000	17:10
Diagnostics Check		OK	17:10

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: AMPARO,AURIS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 09/22/2025 21:11:00

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.115	21:05
Breath Volume		1.132	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.122	21:09
Breath Volume		1.835	21:09
Air Blank		0.000	21:10
Control Test		0.080	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:11

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SANABRIA

DATE/TIME OF TEST: 09/25/2025 22:36:06

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:33	
Control Test		0.081	22:33	
Air Blank		0.000	22:34	
Subject Sample	# 1		22:34	RFI Detect
Breath Volume		0.000	22:34	
Air Blank		0.000	22:35	
Control Test		0.080	22:35	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:36	

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MATTERN

DATE/TIME OF TEST: 09/26/2025 03:32:09

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.081	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.126	03:27
Breath Volume		2.042	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.125	03:30
Breath Volume		2.277	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TOWNSEND,KIRKPATRICK
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SANON

DATE/TIME OF TEST: 09/26/2025 04:27:32

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.081	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.168	04:22
Breath Volume		2.164	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.170	04:25
Breath Volume		1.824	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPICK,SCOTT,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPIC,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MAS GENTA

DATE/TIME OF TEST: 09/27/2025 01:05:07

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.105	01:00
Breath Volume		1.636	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.099	01:03
Breath Volume		1.289	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: ARAUJO

DATE/TIME OF TEST: 09/28/2025 23:59:57

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.081	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.168	23:55
Breath Volume		1.722	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.166	23:58
Breath Volume		2.535	23:58
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNAPP,ANDREW,J
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402783528
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: PEREZ CORNIEL

DATE/TIME OF TEST: 09/29/2025 03:49:35

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.081	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.124	03:45
Breath Volume		1.738	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.126	03:48
Breath Volume		1.937	03:48
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 10/01/2025 10:39

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,NICKOLAUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SUEROLANTIGUA,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: DUNMARS-DANTLEY

DATE/TIME OF TEST: 09/02/2025 08:36:33

Diagnostics Check		OK	08:25	
Air Blank		0.000	08:25	
Control Test		0.079	08:25	
Air Blank		0.000	08:26	
Subject Sample	# 1	0.000	08:29	Volume Not Met
Breath Volume		0.097	08:29	
Air Blank		0.000	08:30	
Air Blank		0.000	08:32	
Subject Sample	# 2	0.000	08:35	Volume Not Met
Breath Volume		0.277	08:35	
Air Blank		0.000	08:35	
Control Test		0.079	08:35	
Air Blank		0.000	08:36	
Diagnostics Check		OK	08:36	

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSSO,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HAMPTON,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: RAYMUNDOMENDOZA

DATE/TIME OF TEST: 09/03/2025 23:24:11

Diagnostics Check		OK	23:17
Air Blank		0.000	23:18
Control Test		0.079	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.131	23:19
Breath Volume		2.636	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:21
Subject Sample	# 2	0.128	23:22
Breath Volume		3.089	23:22
Air Blank		0.000	23:23
Control Test		0.079	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,M,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HILL,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: RODRIGUEZ VILLATORO

DATE/TIME OF TEST: 09/04/2025 02:15:10

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.172	02:10
Breath Volume		2.468	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.174	02:13
Breath Volume		2.414	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EWERT,ZACHARY,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GRISALES,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 09/05/2025 22:40:59

Diagnostics Check		OK	22:25	
Air Blank		0.000	22:26	
Control Test		0.079	22:26	
Air Blank		0.000	22:27	
Subject Sample	# 1	0.205	22:27	
Breath Volume		1.152	22:27	
Air Blank		0.000	22:28	
Air Blank		0.000	22:30	
Subject Sample	# 2	0.178	22:33	Volume Not Met
Breath Volume		0.074	22:33	
Air Blank		0.000	22:34	
Air Blank		0.000	22:36	
Subject Sample	# 3	0.161	22:39	Volume Not Met
Breath Volume		0.308	22:39	
Air Blank		0.000	22:40	
Control Test		0.077	22:40	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EWERT,ZACHARY,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GRISALES,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 09/05/2025 22:51:09

Diagnostics Check		OK	22:46	
Air Blank		0.000	22:47	
Control Test		0.077	22:47	
Air Blank		0.000	22:48	
Subject Sample	# 1		22:49	Subject Test Refused
Breath Volume		0.058	22:49	
Air Blank		0.000	22:50	
Control Test		0.077	22:50	
Air Blank		0.000	22:51	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,AVERY,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,AVERY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RAMEY

DATE/TIME OF TEST: 09/06/2025 02:02:05

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.079	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.084	01:57
Breath Volume		3.078	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.083	02:00
Breath Volume		2.898	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUGG,MEGAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: REIFF,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: LIVIDINI

DATE/TIME OF TEST: 09/06/2025 21:37:13

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.079	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.145	21:32
Breath Volume		2.523	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.138	21:35
Breath Volume		1.765	21:35
Air Blank		0.000	21:36
Control Test		0.078	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ TRIANA,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TROY,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SALAZAR FUNES

DATE/TIME OF TEST: 09/08/2025 03:46:09

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.216	03:41
Breath Volume		2.960	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.210	03:44
Breath Volume		2.003	03:44
Air Blank		0.000	03:45
Control Test		0.078	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDERE,ANTONIO,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SHADLEY,O
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 09/12/2025 20:48:56

Diagnostics Check		OK	20:41
Air Blank		0.000	20:42
Control Test		0.079	20:42
Air Blank		0.000	20:43
Subject Sample	# 1	0.353	20:44
Breath Volume		2.031	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.357	20:47
Breath Volume		1.699	20:47
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGBY,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAREY,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ALEXANDRA

DATE/TIME OF TEST: 09/14/2025 04:46:22

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.032	04:41
Breath Volume		1.167	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.030	04:44
Breath Volume		1.617	04:44
Air Blank		0.000	04:45
Control Test		0.080	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: DELUNA

DATE/TIME OF TEST: 09/14/2025 08:08:01

Diagnostics Check		OK	08:00
Air Blank		0.000	08:01
Control Test		0.079	08:01
Air Blank		0.000	08:02
Subject Sample	# 1	0.256	08:03
Breath Volume		3.281	08:03
Air Blank		0.000	08:04
Air Blank		0.000	08:05
Subject Sample	# 2	0.256	08:06
Breath Volume		1.894	08:06
Air Blank		0.000	08:07
Control Test		0.078	08:07
Air Blank		0.000	08:07
Diagnostics Check		OK	08:08

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSTABLE,AVRY,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FAREZ,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: BOLTZ

DATE/TIME OF TEST: 09/15/2025 03:38:28

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.000	03:33
Breath Volume		1.656	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.000	03:37
Breath Volume		2.062	03:37
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREGOSO,DANIEL,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VANN,JOSHUA
ARRESTING OFFICER AGENCY: FDLE
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: JAVOR

DATE/TIME OF TEST: 09/15/2025 14:57:57

Diagnostics Check		OK	14:51
Air Blank		0.000	14:51
Control Test		0.079	14:52
Air Blank		0.000	14:52
Subject Sample	# 1	0.093	14:53
Breath Volume		1.382	14:53
Air Blank		0.000	14:53
Air Blank		0.000	14:55
Subject Sample	# 2	0.093	14:56
Breath Volume		2.093	14:56
Air Blank		0.000	14:57
Control Test		0.079	14:57
Air Blank		0.000	14:57
Diagnostics Check		OK	14:57

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPISE,STEPHEN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LING,JONATHAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 09/15/2025 20:39:00

Diagnostics Check		OK	20:30	
Air Blank		0.000	20:31	
Control Test		0.079	20:31	
Air Blank		0.000	20:32	
Subject Sample	# 1	0.119	20:33	
Breath Volume		1.718	20:33	
Air Blank		0.000	20:34	
Air Blank		0.000	20:35	
Subject Sample	# 2		20:38	Sequence Aborted
Breath Volume		0.000	20:38	
Air Blank		0.000	20:39	

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTT,RYAN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LING,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 09/16/2025 04:14:23

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.192	04:09
Breath Volume		1.644	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.187	04:12
Breath Volume		1.914	04:12
Air Blank		0.000	04:13
Control Test		0.075	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAND-JEAN,DUDELEY-JACQUES,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: MCDERMOTT

DATE/TIME OF TEST: 09/18/2025 18:12:56

Diagnostics Check		OK	18:06
Air Blank		0.000	18:06
Control Test		0.080	18:06
Air Blank		0.000	18:07
Subject Sample	# 1	0.269	18:07
Breath Volume		1.562	18:07
Air Blank		0.000	18:08
Air Blank		0.000	18:10
Subject Sample	# 2	0.262	18:11
Breath Volume		1.875	18:11
Air Blank		0.000	18:11
Control Test		0.079	18:12
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,T,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: AMMIN,TATE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: PROB/PAROL
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: KANIA

DATE/TIME OF TEST: 09/19/2025 19:12:31

Diagnostics Check		OK	19:06
Air Blank		0.000	19:06
Control Test		0.080	19:06
Air Blank		0.000	19:07
Subject Sample	# 1	0.039	19:07
Breath Volume		2.050	19:07
Air Blank		0.000	19:08
Air Blank		0.000	19:10
Subject Sample	# 2	0.037	19:10
Breath Volume		2.289	19:10
Air Blank		0.000	19:11
Control Test		0.079	19:11
Air Blank		0.000	19:12
Diagnostics Check		OK	19:12

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,AVERY,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,AVERY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LACCETTI

DATE/TIME OF TEST: 09/20/2025 04:18:07

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.080	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.156	04:13
Breath Volume		1.687	04:13
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.155	04:16
Breath Volume		1.699	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,ADAM,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PELLEGRINO,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 09/20/2025 05:03:55

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.079	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.183	04:59
Breath Volume		1.207	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.183	05:02
Breath Volume		1.601	05:02
Air Blank		0.000	05:02
Control Test		0.078	05:03
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,ADAM,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/20/2025 05:40:22

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.079	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.126	05:35
Breath Volume		2.113	05:35
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.122	05:38
Breath Volume		2.003	05:38
Air Blank		0.000	05:39
Control Test		0.077	05:39
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEVER,ZACKERY,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SANDERS,S
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: SCHREKENSTEIN

DATE/TIME OF TEST: 09/20/2025 17:50:44

Diagnostics Check		OK	17:44
Air Blank		0.000	17:44
Control Test		0.079	17:45
Air Blank		0.000	17:45
Subject Sample	# 1	0.220	17:46
Breath Volume		1.875	17:46
Air Blank		0.000	17:46
Air Blank		0.000	17:48
Subject Sample	# 2	0.221	17:49
Breath Volume		1.718	17:49
Air Blank		0.000	17:49
Control Test		0.079	17:50
Air Blank		0.000	17:50
Diagnostics Check		OK	17:50

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACUNA,V,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOTTLE,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: MENDOZA ESCALERA

DATE/TIME OF TEST: 09/20/2025 22:29:15

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:24
Subject Sample	# 1	0.204	22:24
Breath Volume		2.871	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.208	22:27
Breath Volume		3.113	22:27
Air Blank		0.000	22:28
Control Test		0.078	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MENNINI,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: EMHOFF

DATE/TIME OF TEST: 09/22/2025 02:11:30

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.205	02:06
Breath Volume		1.402	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.205	02:09
Breath Volume		3.628	02:09
Air Blank		0.000	02:10
Control Test		0.078	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SAINT HILAIRE

DATE/TIME OF TEST: 09/22/2025 23:49:47

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.156	23:43
Breath Volume		1.667	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.155	23:48
Breath Volume		1.914	23:48
Air Blank		0.000	23:48
Control Test		0.079	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007077

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EWERT,ZACHARY,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BYERS,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: CARDIN

DATE/TIME OF TEST: 09/25/2025 03:56:23

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:42	
Control Test		0.079	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1	0.230	03:46	Volume Not Met
Breath Volume		1.003	03:46	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 2	0.236	03:50	
Breath Volume		2.164	03:50	
Air Blank		0.000	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 3	0.230	03:54	
Breath Volume		2.378	03:54	
Air Blank		0.000	03:55	
Control Test		0.078	03:55	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 09/27/2025 13:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 09/04/2025 22:18:21

Diagnostics Check		OK	22:11
Air Blank		0.000	22:12
Control Test		0.082	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.000	22:13
Breath Volume		1.785	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.000	22:16
Breath Volume		1.652	22:16
Air Blank		0.000	22:17
Control Test		0.082	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/05/2025 03:02:49

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.081	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.143	02:58
Breath Volume		1.664	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.133	03:01
Breath Volume		1.796	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIINANEN,JOSIAH,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: VARONOS,EVANGELOS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MCCRAY

DATE/TIME OF TEST: 09/06/2025 03:43:44

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.083	03:38
Breath Volume		1.542	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.079	03:42
Breath Volume		1.695	03:42
Air Blank		0.000	03:42
Control Test		0.079	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIINANEN,JOSIAH,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: THOMPSON,GREGORY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/06/2025 04:22:42

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.081	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.008	04:18
Breath Volume		2.902	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.008	04:21
Breath Volume		2.359	04:21
Air Blank		0.000	04:21
Control Test		0.081	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALENTINE,ROBIN,F
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: VALENTINE,ROBIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 09/07/2025 04:32:17

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.082	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.148	04:26
Breath Volume		1.554	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:28
Subject Sample	# 2	0.147	04:30
Breath Volume		1.761	04:30
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIINANEN,JOSIAH,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WIINANEN,JOSIAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 09/10/2025 20:19:02

Diagnostics Check		OK	20:11
Air Blank		0.000	20:12
Control Test		0.083	20:12
Air Blank		0.000	20:13
Subject Sample	# 1	0.086	20:14
Breath Volume		1.515	20:14
Air Blank		0.000	20:15
Air Blank		0.000	20:16
Subject Sample	# 2	0.090	20:17
Breath Volume		1.453	20:17
Air Blank		0.000	20:18
Control Test		0.080	20:18
Air Blank		0.000	20:18
Diagnostics Check		OK	20:19

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLEY,RASHELL,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WALSTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 09/11/2025 17:32:54

Diagnostics Check		OK	17:24
Air Blank		0.000	17:24
Control Test		0.082	17:25
Air Blank		0.000	17:25
Subject Sample	# 1	0.285	17:27
Breath Volume		2.417	17:27
Air Blank		0.000	17:28
Air Blank		0.000	17:30
Subject Sample	# 2	0.284	17:31
Breath Volume		2.484	17:31
Air Blank		0.000	17:31
Control Test		0.080	17:32
Air Blank		0.000	17:32
Diagnostics Check		OK	17:32

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASWELL,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HASWELL,AUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 09/13/2025 02:55:21

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.083	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.000	02:51
Breath Volume		2.671	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.000	02:53
Breath Volume		2.894	02:53
Air Blank		0.000	02:54
Control Test		0.082	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILSON,WILLIAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: KEOUGH

DATE/TIME OF TEST: 09/13/2025 03:14:48

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.082	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.081	03:10
Breath Volume		3.371	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.081	03:13
Breath Volume		4.570	03:13
Air Blank		0.000	03:13
Control Test		0.082	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTTON,PIERCE,A
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MAZELLA

DATE/TIME OF TEST: 09/14/2025 02:49:00

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.082	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.191	02:40
Breath Volume		1.714	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:42
Subject Sample	# 2	0.169	02:43
Breath Volume		1.226	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:45
Subject Sample	# 3	0.196	02:47
Breath Volume		2.472	02:47
Air Blank		0.000	02:48
Control Test		0.080	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:49

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: REEL,R
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: KALICHMAN

DATE/TIME OF TEST: 09/15/2025 00:39:07

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.081	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.106	00:34
Breath Volume		1.609	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.102	00:37
Breath Volume		1.167	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,TORRIE,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: COTTON,PIERCE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/15/2025 04:09:33

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:01	
Control Test		0.082	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:05	No Sample Provided
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 2		04:08	RFI Detect
Breath Volume		0.000	04:08	
Air Blank		0.000	04:08	
Control Test		0.082	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIINANEN,JOSIAH,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WIINANEN,JOSIAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 12:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: RABBYSMITH

DATE/TIME OF TEST: 09/17/2025 01:37:00

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.082	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.045	01:31
Breath Volume		3.054	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.047	01:35
Breath Volume		3.398	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:37

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,NOAH,I
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DUFOUR,NOAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: VENTURA-AUSENCIO

DATE/TIME OF TEST: 09/19/2025 23:53:40

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.082	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.179	23:48
Breath Volume		3.250	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.171	23:51
Breath Volume		3.246	23:51
Air Blank		0.000	23:52
Control Test		0.080	23:53
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLEY,RASHELL,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WRIGHT,DUSTIN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: PRADO JR

DATE/TIME OF TEST: 09/20/2025 17:54:35

Diagnostics Check		OK	17:48
Air Blank		0.000	17:48
Control Test		0.082	17:49
Air Blank		0.000	17:49
Subject Sample	# 1	0.285	17:50
Breath Volume		1.867	17:50
Air Blank		0.000	17:50
Air Blank		0.000	17:52
Subject Sample	# 2	0.288	17:52
Breath Volume		1.710	17:52
Air Blank		0.000	17:53
Control Test		0.080	17:53
Air Blank		0.000	17:54
Diagnostics Check		OK	17:54

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLEY,RASHELL,E
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: KELLEY,R
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 11:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/21/2025 00:24:35

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.083	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.097	00:19
Breath Volume		1.621	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.097	00:22
Breath Volume		2.234	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,NOAH,I
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DUFOUR,NOAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/26/2025 02:40:09

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.083	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1	0.194	02:32	Volume Not Met
Breath Volume		0.101	02:32	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 2	0.184	02:38	Volume Not Met
Breath Volume		0.085	02:38	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFOUR,NOAH,I
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DUFOUR,NOAH
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/26/2025 02:54:39

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:43	
Control Test		0.079	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 1	0.191	02:47	Volume Not Met
Breath Volume		0.015	02:47	
Air Blank		0.000	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 2	0.161	02:52	Volume Not Met
Breath Volume		0.132	02:52	
Air Blank		0.000	02:53	
Control Test		0.078	02:53	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,BENJAMIN,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: WILSON,WILLIAM
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: SPRING

DATE/TIME OF TEST: 09/27/2025 03:23:50

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.082	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.038	03:19
Breath Volume		1.437	03:19
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.039	03:22
Breath Volume		1.386	03:22
Air Blank		0.000	03:22
Control Test		0.083	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,TORRIE,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: THOMPSON,TORRIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CONTO

DATE/TIME OF TEST: 09/27/2025 04:36:56

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:31	
Control Test		0.082	04:32	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.221	04:34	
Breath Volume		2.527	04:34	
Air Blank		0.000	04:35	
Air Blank			04:36	Ambient Fail
Air Blank		0.000	04:36	

Data Download Date/Time: 10/16/2025 13:45

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

INSTRUMENT REGISTERED TO: PENSACOLA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,TORRIE,R
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: THOMPSON,TORRIE
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 09/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 232353
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CONTO

DATE/TIME OF TEST: 09/27/2025 04:47:44

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.081	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.226	04:41
Breath Volume		1.824	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.208	04:45
Breath Volume		3.164	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 10/16/2025 13:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: LOSEK,CALEB
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 09/04/2025 21:13:02

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:10	
Control Test		0.079	21:10	
Air Blank		0.000	21:11	
Subject Sample	# 1		21:11	Subject Test Refused
Breath Volume		0.000	21:11	
Air Blank		0.000	21:12	
Control Test		0.079	21:12	
Air Blank		0.000	21:12	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 10/06/2025 03:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONEY,MARY,L
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: SHEAHAN,BRENDON
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: ESPINOZA TORRES

DATE/TIME OF TEST: 09/14/2025 03:49:52

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:47	
Control Test		0.078	03:47	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:48	Subject Test Refused
Breath Volume		0.000	03:48	
Air Blank		0.000	03:48	
Control Test		0.078	03:49	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 10/06/2025 03:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALBRITTON,GARRETT,M
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: PATTERSON,CHANDLER
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GIL VASQUEZ

DATE/TIME OF TEST: 09/20/2025 00:43:26

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.193	00:38
Breath Volume		3.375	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.193	00:41
Breath Volume		2.339	00:41
Air Blank		0.000	00:42
Control Test		0.077	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 10/06/2025 03:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,SOLEY,G
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: KALLMEYER,DANIEL
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SANTIS MENDEZ

DATE/TIME OF TEST: 09/22/2025 04:05:58

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.080	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.137	04:01
Breath Volume		2.160	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:03
Subject Sample	# 2	0.135	04:04
Breath Volume		1.914	04:04
Air Blank		0.000	04:05
Control Test		0.076	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 10/06/2025 03:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007160

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,SOLEY,G
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WATERS,BRODY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: CAMBRAY

DATE/TIME OF TEST: 09/30/2025 23:43:27

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.078	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.173	23:38
Breath Volume		5.117	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.174	23:41
Breath Volume		5.328	23:41
Air Blank		0.000	23:42
Control Test		0.076	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 10/06/2025 03:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GOMEZ SANCHEZ

DATE/TIME OF TEST: 09/18/2025 23:59:26

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.078	23:52
Air Blank		0.000	23:53
Subject Sample	# 1	0.087	23:54
Breath Volume		3.558	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.085	23:57
Breath Volume		2.828	23:57
Air Blank		0.000	23:58
Control Test		0.079	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 10/16/2025 11:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 09/19/2025 03:54:01

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.079	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.187	03:49
Breath Volume		2.605	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:51
Subject Sample	# 2	0.192	03:52
Breath Volume		2.574	03:52
Air Blank		0.000	03:53
Control Test		0.078	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:54

Data Download Date/Time: 10/16/2025 11:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PETERS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 09/20/2025 03:58:58

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.079	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.124	03:54
Breath Volume		1.875	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.123	03:57
Breath Volume		1.640	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 10/16/2025 11:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIBBINS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MARRA II

DATE/TIME OF TEST: 09/20/2025 04:23:39

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:07	
Control Test		0.078	04:08	
Air Blank		0.000	04:08	
Subject Sample	# 1	0.148	04:11	Volume Not Met
Breath Volume		0.328	04:11	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2	0.176	04:16	
Breath Volume		2.796	04:16	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 3	0.163	04:21	
Breath Volume		1.652	04:21	
Air Blank		0.000	04:22	
Control Test		0.082	04:22	
Air Blank		0.000	04:23	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 10/16/2025 11:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: ROSARIO RIVERA

DATE/TIME OF TEST: 09/20/2025 05:21:08

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:15
Subject Sample	# 1	0.027	05:16
Breath Volume		1.773	05:16
Air Blank		0.000	05:17
Air Blank		0.000	05:18
Subject Sample	# 2	0.028	05:19
Breath Volume		2.230	05:19
Air Blank		0.000	05:20
Control Test		0.078	05:20
Air Blank		0.000	05:21
Diagnostics Check		OK	05:21

Data Download Date/Time: 10/16/2025 11:58

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVERES,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/20/2025 22:25:18

Diagnostics Check		OK	22:18
Air Blank		0.000	22:19
Control Test		0.078	22:19
Air Blank		0.000	22:19
Subject Sample	# 1	0.148	22:20
Breath Volume		3.320	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.145	22:23
Breath Volume		1.976	22:23
Air Blank		0.000	22:24
Control Test		0.078	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: HERNANDEZ GOMEZ

DATE/TIME OF TEST: 09/21/2025 00:42:50

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.078	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.270	00:37
Breath Volume		1.636	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.266	00:41
Breath Volume		2.046	00:41
Air Blank		0.000	00:41
Control Test		0.077	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 10/16/2025 12:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST.WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: MODESTI

DATE/TIME OF TEST: 09/21/2025 01:39:30

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.077	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.095	01:34
Breath Volume		1.152	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.097	01:37
Breath Volume		1.566	01:37
Air Blank		0.000	01:38
Control Test		0.078	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 09/21/2025 03:56:55

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.078	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.048	03:52
Breath Volume		3.019	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.048	03:55
Breath Volume		3.839	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PORTER,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: ESPINOSAMACDONALD

DATE/TIME OF TEST: 09/21/2025 17:06:06

Diagnostics Check		OK	16:59
Air Blank		0.000	16:59
Control Test		0.078	17:00
Air Blank		0.000	17:00
Subject Sample	# 1	0.194	17:01
Breath Volume		1.597	17:01
Air Blank		0.000	17:01
Air Blank		0.000	17:03
Subject Sample	# 2	0.197	17:04
Breath Volume		1.902	17:04
Air Blank		0.000	17:05
Control Test		0.079	17:05
Air Blank		0.000	17:06
Diagnostics Check		OK	17:06

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALVA-ARIAS,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: MCLEAN

DATE/TIME OF TEST: 09/21/2025 19:39:14

Diagnostics Check		OK	19:32
Air Blank		0.000	19:33
Control Test		0.079	19:33
Air Blank		0.000	19:34
Subject Sample	# 1	0.243	19:34
Breath Volume		3.289	19:34
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.262	19:37
Breath Volume		3.273	19:37
Air Blank		0.000	19:38
Control Test		0.078	19:38
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: VERA

DATE/TIME OF TEST: 09/21/2025 22:00:12

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:48	
Control Test		0.079	21:48	
Air Blank		0.000	21:49	
Subject Sample	# 1	0.097	21:52	Volume Not Met
Breath Volume		0.433	21:52	
Air Blank		0.000	21:53	
Air Blank		0.000	21:55	
Subject Sample	# 2	0.129	21:58	Volume Not Met
Breath Volume		0.679	21:58	
Air Blank		0.000	21:59	
Control Test		0.082	21:59	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:00	

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: POWARD

DATE/TIME OF TEST: 09/22/2025 04:14:27

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.079	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.097	04:09
Breath Volume		1.769	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.097	04:12
Breath Volume		2.093	04:12
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: COMM VEHIC
ARREST TIME: 04:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 09/22/2025 05:48:42

Diagnostics Check		OK	05:42
Air Blank		0.000	05:42
Control Test		0.078	05:42
Air Blank		0.000	05:43
Subject Sample	# 1	0.091	05:43
Breath Volume		1.664	05:43
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.090	05:47
Breath Volume		1.800	05:47
Air Blank		0.000	05:47
Control Test		0.078	05:48
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: COMM VEHIC
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MESA PEREZ

DATE/TIME OF TEST: 09/25/2025 04:27:35

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.077	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.187	04:22
Breath Volume		1.871	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.175	04:25
Breath Volume		1.562	04:25
Air Blank		0.000	04:26
Control Test		0.077	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: DIEZ CLEMENTE

DATE/TIME OF TEST: 09/26/2025 04:23:48

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.072	04:18
Breath Volume		1.406	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.069	04:22
Breath Volume		1.257	04:22
Air Blank		0.000	04:22
Control Test		0.078	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: RUTLEDGE

DATE/TIME OF TEST: 09/27/2025 01:57:27

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.079	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.141	01:52
Breath Volume		1.445	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.147	01:55
Breath Volume		1.125	01:55
Air Blank		0.000	01:56
Control Test		0.081	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BENAVENTE ORDONEZ

DATE/TIME OF TEST: 09/28/2025 00:24:11

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.200	00:19
Breath Volume		3.011	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:21
Subject Sample	# 2	0.195	00:22
Breath Volume		1.925	00:22
Air Blank		0.000	00:23
Control Test		0.077	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 10/16/2025 12:29

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MAROCOLA,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: PREDMORE

DATE/TIME OF TEST: 09/28/2025 00:48:56

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.078	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.217	00:44
Breath Volume		3.039	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.213	00:47
Breath Volume		2.566	00:47
Air Blank		0.000	00:47
Control Test		0.077	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 10/16/2025 12:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ DIAZ

DATE/TIME OF TEST: 09/28/2025 03:13:35

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.188	03:08
Breath Volume		1.320	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.170	03:11
Breath Volume		1.609	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GIBBINS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: SWEDENBURG

DATE/TIME OF TEST: 09/28/2025 04:21:45

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.078	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.195	04:17
Breath Volume		1.121	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.193	04:20
Breath Volume		1.300	04:20
Air Blank		0.000	04:20
Control Test		0.078	04:21
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DURAN,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 09/29/2025 01:15:54

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.109	01:10
Breath Volume		1.519	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.110	01:14
Breath Volume		1.199	01:14
Air Blank		0.000	01:14
Control Test		0.081	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: COMM VEHIC
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/29/2025 01:43:24

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.149	01:38
Breath Volume		2.394	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:40
Subject Sample	# 2	0.132	01:41
Breath Volume		1.906	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FRAHER

DATE/TIME OF TEST: 09/30/2025 03:49:03

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:43
Subject Sample	# 1	0.075	03:44
Breath Volume		1.921	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.068	03:47
Breath Volume		1.382	03:47
Air Blank		0.000	03:48
Control Test		0.079	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:49

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007163

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 09/30/2025 23:04:44

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.079	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.009	23:00
Breath Volume		1.656	23:00
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.008	23:03
Breath Volume		1.804	23:03
Air Blank		0.000	23:03
Control Test		0.079	23:04
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 10/16/2025 12:30

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: TANKSLEY

DATE/TIME OF TEST: 09/01/2025 20:46:01

Diagnostics Check		OK	20:39
Air Blank		0.000	20:39
Control Test		0.080	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.036	20:41
Breath Volume		2.324	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:43
Subject Sample	# 2	0.035	20:44
Breath Volume		2.417	20:44
Air Blank		0.000	20:45
Control Test		0.078	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:46

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 09/02/2025 04:30:19

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.249	04:25
Breath Volume		3.097	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.243	04:28
Breath Volume		2.953	04:28
Air Blank		0.000	04:29
Control Test		0.077	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: TEET

DATE/TIME OF TEST: 09/03/2025 00:51:39

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:49	
Control Test		0.080	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:50	
Control Test		0.080	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNBRUGH,E,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ROMAN,L
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:04

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 09/03/2025 16:31:51

Diagnostics Check		OK	16:26	
Air Blank		0.000	16:26	
Control Test		0.080	16:27	
Air Blank		0.000	16:27	
Subject Sample	# 1	0.233	16:27	
Breath Volume		1.886	16:27	
Air Blank		0.000	16:28	
Air Blank		0.000	16:30	
Subject Sample	# 2	0.248	16:31	
Breath Volume		1.910	16:31	
Air Blank			16:31	RFI Detect
Air Blank		0.000	16:31	

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: SHARP,ELIOT
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: RASK

DATE/TIME OF TEST: 09/04/2025 21:58:36

Diagnostics Check		OK	21:52
Air Blank		0.000	21:52
Control Test		0.080	21:53
Air Blank		0.000	21:53
Subject Sample	# 1	0.125	21:54
Breath Volume		1.187	21:54
Air Blank		0.000	21:54
Air Blank		0.000	21:56
Subject Sample	# 2	0.110	21:56
Breath Volume		1.140	21:56
Air Blank		0.000	21:57
Control Test		0.079	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,AYDEN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FLETCHER,AYDEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 09/05/2025 01:50:00

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:39	
Control Test		0.080	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:43	No Sample Provided
Breath Volume		0.000	01:43	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2		01:48	No Sample Provided
Breath Volume		0.000	01:48	
Air Blank		0.000	01:49	
Control Test		0.081	01:49	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,TY,D
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,TY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DOLQUEIST

DATE/TIME OF TEST: 09/06/2025 02:54:38

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.269	02:49
Breath Volume		1.617	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.277	02:52
Breath Volume		1.988	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 09/06/2025 04:00:37

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.080	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1		03:59	Subject Test Refused
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.081	03:59	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 09/07/2025 19:51:46

Diagnostics Check		OK	19:48	
Air Blank		0.000	19:49	
Control Test		0.079	19:49	
Air Blank		0.000	19:50	
Subject Sample	# 1		19:50	Subject Test Refused
Breath Volume		0.000	19:50	
Air Blank		0.000	19:50	
Control Test		0.079	19:51	
Air Blank		0.000	19:51	
Diagnostics Check		OK	19:51	

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: HOLMES,RUSSELL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 09/07/2025 23:01:55

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:59	
Control Test		0.079	22:59	
Air Blank		0.000	23:00	
Subject Sample	# 1		23:00	Subject Test Refused
Breath Volume		0.000	23:00	
Air Blank		0.000	23:00	
Control Test		0.079	23:01	
Air Blank		0.000	23:01	
Diagnostics Check		OK	23:01	

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ROSS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HUGHES,ROSS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: POTTER

DATE/TIME OF TEST: 09/08/2025 21:09:33

Diagnostics Check		OK	21:02
Air Blank		0.000	21:02
Control Test		0.079	21:02
Air Blank		0.000	21:03
Subject Sample	# 1	0.159	21:04
Breath Volume		2.550	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.158	21:07
Breath Volume		3.277	21:07
Air Blank		0.000	21:08
Control Test		0.077	21:08
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,AYDEN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FLETCHER,AYDEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: AVERY

DATE/TIME OF TEST: 09/09/2025 00:07:04

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.078	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.267	00:02
Breath Volume		1.660	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:04
Subject Sample	# 2	0.270	00:05
Breath Volume		4.183	00:05
Air Blank		0.000	00:06
Control Test		0.078	00:06
Air Blank		0.000	00:06
Diagnostics Check		OK	00:07

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,RUSSELL,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WRIGHT,DANIEL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 09/10/2025 23:14:04

Diagnostics Check		OK	23:07
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Subject Sample	# 1	0.039	23:09
Breath Volume		1.382	23:09
Air Blank		0.000	23:10
Air Blank		0.000	23:11
Subject Sample	# 2	0.039	23:12
Breath Volume		1.554	23:12
Air Blank		0.000	23:13
Control Test		0.077	23:13
Air Blank		0.000	23:13
Diagnostics Check		OK	23:14

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,AYDEN,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: JESUS,ROJAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/13/2025 23:41:04

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.080	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Control Test		0.080	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTINO-LYDA,RIVER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CUTINO-LYDA,RIVER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:29

SUBJECT'S LAST NAME: BARROS

DATE/TIME OF TEST: 09/14/2025 07:06:11

Diagnostics Check		OK	06:58
Air Blank		0.000	06:58
Control Test		0.080	06:59
Air Blank		0.000	06:59
Subject Sample	# 1	0.126	07:01
Breath Volume		2.226	07:01
Air Blank		0.000	07:02
Air Blank		0.000	07:03
Subject Sample	# 2	0.121	07:04
Breath Volume		2.464	07:04
Air Blank		0.000	07:05
Control Test		0.078	07:05
Air Blank		0.000	07:06
Diagnostics Check		OK	07:06

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: OSBORNE,JEROME
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: HALL-FREEMAN

DATE/TIME OF TEST: 09/15/2025 00:51:53

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:40	
Control Test		0.079	00:40	
Air Blank		0.000	00:41	
Subject Sample	# 1	0.214	00:41	
Breath Volume		2.167	00:41	
Air Blank		0.000	00:42	
Air Blank		0.000	00:44	
Subject Sample	# 2	0.161	00:47	Volume Not Met
Breath Volume		0.015	00:47	
Air Blank		0.000	00:48	
Air Blank		0.000	00:49	
Subject Sample	# 3	0.203	00:50	
Breath Volume		2.351	00:50	
Air Blank		0.000	00:50	
Control Test		0.073	00:51	Control Outside Tolerance
Air Blank		0.000	00:51	

Data Download Date/Time: 09/17/2025 11:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ROSS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HUGHES,ROSS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: FLINT

DATE/TIME OF TEST: 09/19/2025 00:33:25

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.079	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:32	Subject Test Refused
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Control Test		0.078	00:32	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 15:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:34

SUBJECT'S LAST NAME: MICKEL

DATE/TIME OF TEST: 09/19/2025 15:58:51

Diagnostics Check		OK	15:55	
Air Blank		0.000	15:56	
Control Test		0.079	15:56	
Air Blank		0.000	15:57	
Subject Sample	# 1		15:57	Subject Test Refused
Breath Volume		0.000	15:57	
Air Blank		0.000	15:57	
Control Test		0.080	15:58	
Air Blank		0.000	15:58	
Diagnostics Check		OK	15:58	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,BRANDON,H
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KNOWLES,BRANDON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/21/2025 20:47:59

Diagnostics Check		OK	20:45	
Air Blank		0.000	20:45	
Control Test		0.079	20:45	
Air Blank		0.000	20:46	
Subject Sample	# 1		20:46	Subject Test Refused
Breath Volume		0.000	20:46	
Air Blank		0.000	20:47	
Control Test		0.079	20:47	
Air Blank		0.000	20:47	
Diagnostics Check		OK	20:47	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO PEREZ,RODOLFO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CASTILLO PEREZ,RODOLFO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 09/24/2025 21:33:33

Diagnostics Check		OK	21:24	
Air Blank		0.000	21:25	
Control Test		0.078	21:25	
Air Blank		0.000	21:26	
Subject Sample	# 1	0.188	21:28	
Breath Volume		1.554	21:28	
Air Blank		0.000	21:29	
Air Blank		0.000	21:30	
Subject Sample	# 2		21:31	Subject Test Refused
Breath Volume		0.949	21:31	
Air Blank		0.000	21:32	
Control Test		0.075	21:32	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO PEREZ,RODOLFO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: JEAN BAPTISTE,STANLEY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 09/25/2025 01:02:52

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:59	
Control Test		0.078	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:01	
Control Test		0.078	01:02	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JAMES,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: WRIGHT,DANIEL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/26/2025 03:25:50

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:13	
Control Test		0.079	03:14	
Air Blank		0.000	03:14	
Subject Sample	# 1	0.155	03:17	Volume Not Met
Breath Volume		0.710	03:17	
Air Blank		0.000	03:18	
Air Blank		0.000	03:20	
Subject Sample	# 2	0.160	03:21	
Breath Volume		1.941	03:21	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 3	0.147	03:24	
Breath Volume		1.679	03:24	
Air Blank		0.000	03:24	
Control Test		0.077	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVAL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PREVAL,MICHAEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/26/2025 21:28:08

Diagnostics Check		OK	21:21
Air Blank		0.000	21:22
Control Test		0.079	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.085	21:23
Breath Volume		1.484	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.083	21:26
Breath Volume		1.867	21:26
Air Blank		0.000	21:27
Control Test		0.077	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 10/18/2025 10:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ROSS,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HUGHES,ROSS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: ZAVALA

DATE/TIME OF TEST: 09/27/2025 05:12:39

Diagnostics Check		OK	05:09	
Air Blank		0.000	05:10	
Control Test		0.079	05:10	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:11	Subject Test Refused
Breath Volume		0.000	05:11	
Air Blank		0.000	05:11	
Control Test		0.078	05:12	
Air Blank		0.000	05:12	
Diagnostics Check		OK	05:12	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ANDRADE,ESTEBAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: OTT

DATE/TIME OF TEST: 09/27/2025 22:46:08

Diagnostics Check		OK	22:43	
Air Blank		0.000	22:43	
Control Test		0.079	22:43	
Air Blank		0.000	22:44	
Subject Sample	# 1		22:44	Subject Test Refused
Breath Volume		0.000	22:44	
Air Blank		0.000	22:45	
Control Test		0.079	22:45	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 10/18/2025 10:13

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007167

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDRADE,ESTEBAN,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ANDRADE,ESTEBAN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: MYER

DATE/TIME OF TEST: 09/28/2025 04:32:23

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.215	04:27
Breath Volume		1.488	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.214	04:30
Breath Volume		1.433	04:30
Air Blank		0.000	04:31
Control Test		0.077	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 10/18/2025 10:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOCKE,BRIANNA,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LOCKE,BRIANNA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 09/21/2025 05:23:47

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.079	05:17
Air Blank		0.000	05:18
Subject Sample	# 1	0.083	05:18
Breath Volume		2.101	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.077	05:21
Breath Volume		1.785	05:21
Air Blank		0.000	05:22
Control Test		0.079	05:23
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINNERT,AJAY,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MEINNERT,AJAY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: CRUMEDY

DATE/TIME OF TEST: 09/21/2025 06:01:30

Diagnostics Check		OK	05:52
Air Blank		0.000	05:52
Control Test		0.079	05:53
Air Blank		0.000	05:53
Subject Sample	# 1	0.190	05:54
Breath Volume		1.277	05:54
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.180	05:59
Breath Volume		1.203	05:59
Air Blank		0.000	06:00
Control Test		0.077	06:00
Air Blank		0.000	06:01
Diagnostics Check		OK	06:01

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 09/22/2025 21:21:04

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:16	
Control Test		0.079	21:16	
Air Blank		0.000	21:17	
Subject Sample	# 1	0.225	21:19	
Breath Volume		1.906	21:19	
Air Blank			21:20	Purge Fail
Air Blank		0.000	21:21	

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRINGTON,ALEX,
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HARRINGTON,ALEX
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403318190
DRY GAS EXPIRY DATE: 04/10/2028
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 09/22/2025 21:34:29

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.076	21:23
Air Blank		0.000	21:23
Subject Sample	# 1	0.222	21:25
Breath Volume		1.292	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.245	21:28
Breath Volume		1.222	21:28
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 3	0.231	21:32
Breath Volume		1.410	21:32
Air Blank		0.000	21:33
Control Test		0.077	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,COY,
BREATH TEST OPERATOR AGENCY: Belleview PD
ARRESTING OFFICER: HUNT,COY
ARRESTING OFFICER AGENCY: Belleview PD
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 09/24/2025 00:10:32

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.163	00:05
Breath Volume		2.632	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.164	00:08
Breath Volume		2.328	00:08
Air Blank		0.000	00:09
Control Test		0.078	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,ROSA,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: BROWN,JESSE
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 09/25/2025 20:50:04

Diagnostics Check		OK	20:43
Air Blank		0.000	20:43
Control Test		0.080	20:43
Air Blank		0.000	20:44
Subject Sample	# 1	0.144	20:45
Breath Volume		2.222	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:47
Subject Sample	# 2	0.147	20:48
Breath Volume		2.027	20:48
Air Blank		0.000	20:49
Control Test		0.077	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:50

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HATCH,VIRGINIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403315737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: CALLAWAY

DATE/TIME OF TEST: 09/26/2025 14:16:14

Diagnostics Check	OK	14:15	
Air Blank		14:15	RFI Detect
Air Blank	0.000	14:16	

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,CHRISTIAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HATCH,VIRGINIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403315737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: CALLAWAY

DATE/TIME OF TEST: 09/26/2025 14:25:11

Diagnostics Check		OK	14:17
Air Blank		0.000	14:18
Control Test		0.080	14:18
Air Blank		0.000	14:19
Subject Sample	# 1	0.065	14:19
Breath Volume		1.324	14:19
Air Blank		0.000	14:20
Air Blank		0.000	14:22
Subject Sample	# 2	0.065	14:23
Breath Volume		1.394	14:23
Air Blank		0.000	14:24
Control Test		0.079	14:24
Air Blank		0.000	14:25
Diagnostics Check		OK	14:25

Data Download Date/Time: 10/21/2025 12:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINNERT,AJAY,M
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MEINNERT,AJAY
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 09/03/2025 04:14:40

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.202	04:09
Breath Volume		1.144	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.198	04:12
Breath Volume		1.195	04:12
Air Blank		0.000	04:13
Control Test		0.075	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTS,C,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ARIAS,C
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:59

SUBJECT'S LAST NAME: JETER

DATE/TIME OF TEST: 09/03/2025 11:42:24

Diagnostics Check		OK	11:36
Air Blank		0.000	11:36
Control Test		0.079	11:36
Air Blank		0.000	11:37
Subject Sample	# 1	0.000	11:38
Breath Volume		2.203	11:38
Air Blank		0.000	11:38
Air Blank		0.000	11:40
Subject Sample	# 2	0.000	11:40
Breath Volume		1.859	11:40
Air Blank		0.000	11:41
Control Test		0.077	11:41
Air Blank		0.000	11:42
Diagnostics Check		OK	11:42

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,SEBASTIAN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MENDOZA,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/04/2025 23:34:03

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:28
Air Blank		0.000	23:28
Subject Sample	# 1	0.126	23:29
Breath Volume		1.585	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:31
Subject Sample	# 2	0.126	23:32
Breath Volume		1.570	23:32
Air Blank		0.000	23:33
Control Test		0.076	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:34

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,JUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LONEY,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: SILVIA

DATE/TIME OF TEST: 09/06/2025 19:08:51

Diagnostics Check		OK	19:01
Air Blank		0.000	19:02
Control Test		0.079	19:02
Air Blank		0.000	19:03
Subject Sample	# 1	0.000	19:04
Breath Volume		1.500	19:04
Air Blank		0.000	19:04
Air Blank		0.000	19:06
Subject Sample	# 2	0.000	19:07
Breath Volume		3.378	19:07
Air Blank		0.000	19:07
Control Test		0.078	19:08
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTIS,LAYNE,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MCGEE,RYAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: BANAS

DATE/TIME OF TEST: 09/07/2025 19:03:51

Diagnostics Check		OK	18:57
Air Blank		0.000	18:57
Control Test		0.079	18:57
Air Blank		0.000	18:58
Subject Sample	# 1	0.164	18:59
Breath Volume		1.640	18:59
Air Blank		0.000	19:00
Air Blank		0.000	19:01
Subject Sample	# 2	0.157	19:02
Breath Volume		1.511	19:02
Air Blank		0.000	19:02
Control Test		0.077	19:03
Air Blank		0.000	19:03
Diagnostics Check		OK	19:03

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADARA,MATTHEW,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MADARA,MATTHEW
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MASSENGALE

DATE/TIME OF TEST: 09/08/2025 02:54:02

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:43	
Control Test		0.079	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 1	0.174	02:47	Volume Not Met
Breath Volume		0.273	02:47	
Air Blank		0.000	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 2	0.170	02:52	Volume Not Met
Breath Volume		0.097	02:52	
Air Blank			02:53	Purge Fail
Air Blank		0.000	02:54	

Data Download Date/Time: 09/19/2025 16:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESANDRE,DAVID,H
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: DESANDRE,DAVID
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: SZCZEBAK

DATE/TIME OF TEST: 09/08/2025 23:32:47

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.079	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.195	23:27
Breath Volume		1.703	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.194	23:31
Breath Volume		2.335	23:31
Air Blank		0.000	23:31
Control Test		0.078	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,ROBERT,S
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: DAVILA,DULCE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/11/2025 01:16:13

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.079	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.213	01:11
Breath Volume		2.734	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:13
Subject Sample	# 2	0.197	01:14
Breath Volume		1.789	01:14
Air Blank		0.000	01:15
Control Test		0.075	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,DAVID,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: SANTANA,DAVID
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 09/11/2025 20:49:09

Diagnostics Check		OK	20:41
Air Blank		0.000	20:42
Control Test		0.079	20:42
Air Blank		0.000	20:42
Subject Sample	# 1	0.000	20:44
Breath Volume		3.312	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.000	20:47
Breath Volume		2.714	20:47
Air Blank		0.000	20:48
Control Test		0.076	20:48
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEZUSKI,JIMMY,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEZUSKI,JIMMY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/13/2025 03:58:18

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.079	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.122	03:53
Breath Volume		3.179	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.132	03:56
Breath Volume		4.023	03:56
Air Blank		0.000	03:57
Control Test		0.078	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEGUERO,FEDERICO,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: PEGUERO,FEDERICO
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: DARIENZO

DATE/TIME OF TEST: 09/13/2025 14:58:07

Diagnostics Check		OK	14:50
Air Blank		0.000	14:51
Control Test		0.079	14:51
Air Blank		0.000	14:52
Subject Sample	# 1	0.064	14:52
Breath Volume		2.402	14:52
Air Blank		0.000	14:53
Air Blank		0.000	14:54
Subject Sample	# 2	0.066	14:56
Breath Volume		1.753	14:56
Air Blank		0.000	14:57
Control Test		0.077	14:57
Air Blank		0.000	14:58
Diagnostics Check		OK	14:58

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,A
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MURRY

DATE/TIME OF TEST: 09/14/2025 04:15:46

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.079	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.134	04:11
Breath Volume		1.519	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.132	04:14
Breath Volume		1.144	04:14
Air Blank		0.000	04:14
Control Test		0.078	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONEY,JUSTIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: LONEY,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:43

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/16/2025 16:10:41

Diagnostics Check		OK	16:03
Air Blank		0.000	16:04
Control Test		0.079	16:04
Air Blank		0.000	16:04
Subject Sample	# 1	0.000	16:06
Breath Volume		2.367	16:06
Air Blank		0.000	16:06
Air Blank		0.000	16:08
Subject Sample	# 2	0.000	16:09
Breath Volume		1.859	16:09
Air Blank		0.000	16:09
Control Test		0.077	16:10
Air Blank		0.000	16:10
Diagnostics Check		OK	16:10

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007172

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COON,AUSTIN,D
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: COON,AUSTIN
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 08/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: DOLLMAN

DATE/TIME OF TEST: 09/16/2025 20:42:28

Diagnostics Check		OK	20:39	
Air Blank		0.000	20:40	
Control Test		0.079	20:40	
Air Blank		0.000	20:40	
Subject Sample	# 1		20:41	Subject Test Refused
Breath Volume		0.000	20:41	
Air Blank		0.000	20:41	
Control Test		0.078	20:41	
Air Blank		0.000	20:42	
Diagnostics Check		OK	20:42	

Data Download Date/Time: 09/19/2025 16:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: HAMMOND

DATE/TIME OF TEST: 09/09/2025 17:34:01

Diagnostics Check		OK	17:25
Air Blank		0.000	17:26
Control Test		0.078	17:26
Air Blank		0.000	17:27
Subject Sample	# 1	0.097	17:28
Breath Volume		2.101	17:28
Air Blank		0.000	17:29
Air Blank		0.000	17:30
Subject Sample	# 2	0.097	17:32
Breath Volume		1.812	17:32
Air Blank		0.000	17:33
Control Test		0.077	17:33
Air Blank		0.000	17:33
Diagnostics Check		OK	17:34

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAVES,J,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: WILLIS,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/12/2025 15:22:57

Diagnostics Check		OK	15:16
Air Blank		0.000	15:17
Control Test		0.078	15:17
Air Blank		0.000	15:18
Subject Sample	# 1	0.000	15:18
Breath Volume		2.179	15:18
Air Blank		0.000	15:18
Air Blank		0.000	15:20
Subject Sample	# 2	0.000	15:21
Breath Volume		2.621	15:21
Air Blank		0.000	15:21
Control Test		0.079	15:22
Air Blank		0.000	15:22
Diagnostics Check		OK	15:22

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,BRIANNA,M
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: URQUHART,CLAY
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 09/13/2025 23:47:40

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.145	23:42
Breath Volume		2.125	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.141	23:45
Breath Volume		2.015	23:45
Air Blank		0.000	23:46
Control Test		0.077	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: DOUGLAS,JORDAN
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: ROBINS

DATE/TIME OF TEST: 09/14/2025 05:22:14

Diagnostics Check		OK	05:15
Air Blank		0.000	05:15
Control Test		0.077	05:16
Air Blank		0.000	05:16
Subject Sample	# 1	0.139	05:17
Breath Volume		2.468	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:19
Subject Sample	# 2	0.139	05:20
Breath Volume		1.812	05:20
Air Blank		0.000	05:21
Control Test		0.078	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PENA,RUSSEL,S
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: MARRERO,E
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: URUENA

DATE/TIME OF TEST: 09/20/2025 02:24:39

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.034	02:19
Breath Volume		3.019	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.035	02:22
Breath Volume		3.386	02:22
Air Blank		0.000	02:23
Control Test		0.078	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,J
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: HOLMES,A
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HOWER

DATE/TIME OF TEST: 09/20/2025 22:58:24

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.079	22:55	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:56	Subject Test Refused
Breath Volume		0.000	22:56	
Air Blank		0.000	22:57	
Control Test		0.078	22:57	
Air Blank		0.000	22:58	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,J
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ALVARADO,M
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: GOMEZ HERNANDEZ

DATE/TIME OF TEST: 09/21/2025 00:53:35

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:52	
Control Test	0.079	00:52	RFI Detect
Air Blank	0.000	00:53	

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOL,J,J
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: ALVARADO,M
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GOMEZ HERNANDEZ

DATE/TIME OF TEST: 09/21/2025 01:04:43

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.164	00:59
Breath Volume		2.042	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.164	01:02
Breath Volume		2.195	01:02
Air Blank		0.000	01:03
Control Test		0.076	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACAVOY,J,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BEENER,K
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 09/21/2025 04:03:40

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:01	
Control Test		0.076	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:02	
Control Test		0.077	04:02	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOCKETT-WRIGHT,D,
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/21/2025 06:13:32

Diagnostics Check		OK	06:05
Air Blank		0.000	06:05
Control Test		0.078	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.000	06:08
Breath Volume		1.496	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.000	06:12
Breath Volume		1.332	06:12
Air Blank		0.000	06:12
Control Test		0.079	06:12
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,BRIANNA,M
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LAWLES,BILLY
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 09/23/2025 20:40:56

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.079	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.093	20:35
Breath Volume		1.832	20:35
Air Blank		0.000	20:36
Air Blank		0.000	20:37
Subject Sample	# 2	0.094	20:39
Breath Volume		1.804	20:39
Air Blank		0.000	20:39
Control Test		0.079	20:40
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELS,JESSICA,N
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: WORTH,M
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: RONDON

DATE/TIME OF TEST: 09/23/2025 23:33:04

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.077	23:32	
Air Blank		23:32	RFI Detect
Air Blank	0.000	23:33	

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELS,JESSICA,N
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: WORTH,M
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RONDON

DATE/TIME OF TEST: 09/23/2025 23:36:51

Diagnostics Check	OK	23:36	
Air Blank		23:36	RFI Detect
Air Blank	0.000	23:36	

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELS,JESSICA,N
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: WORTH,M
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RONDON

DATE/TIME OF TEST: 09/23/2025 23:41:49

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.079	23:39	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:40	Subject Test Refused
Breath Volume		0.000	23:40	
Air Blank		0.000	23:40	
Control Test		0.078	23:41	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACAVOY,J,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: URQUHART,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 09/25/2025 03:01:37

Diagnostics Check	OK	03:00	
Air Blank		03:01	Ambient Fail
Air Blank	0.000	03:01	

Data Download Date/Time: 10/28/2025 12:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACAVOY,J,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: URQUHART,C
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: AEVEDO

DATE/TIME OF TEST: 09/25/2025 03:10:54

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.078	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.204	03:05
Breath Volume		2.269	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.204	03:08
Breath Volume		2.687	03:08
Air Blank		0.000	03:09
Control Test		0.077	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,BRIANNA,M
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: LOWE,MAYCEE
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 09/28/2025 01:14:39

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:11	
Control Test		0.079	01:12	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:13	
Control Test		0.078	01:13	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 10/28/2025 12:51

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: WORTH,M
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 09/28/2025 02:59:30

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:51	
Control Test		0.079	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.183	02:54	
Breath Volume		1.734	02:54	
Air Blank		0.000	02:55	
Air Blank		0.000	02:56	
Subject Sample	# 2	0.178	02:57	
Breath Volume		1.351	02:57	
Air Blank		0.000	02:58	
Control Test		0.073	02:58	Control Outside Tolerance
Air Blank		0.000	02:59	

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,ZAHIRYD,M
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: WOOD,A
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: D HAITI

DATE/TIME OF TEST: 09/28/2025 03:19:03

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.075	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:18	
Control Test		0.078	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

INSTRUMENT REGISTERED TO: GROVELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELS,JESSICA,N
BREATH TEST OPERATOR AGENCY: Groveland PD
ARRESTING OFFICER: LOWE,M
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: SVENDSEN

DATE/TIME OF TEST: 09/29/2025 00:50:17

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.078	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.033	00:45
Breath Volume		2.246	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:47
Subject Sample	# 2	0.033	00:48
Breath Volume		2.566	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 10/28/2025 12:52

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRABTREE,JEREMIAH,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CRABTREE,JEREMIAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: BRAWNER

DATE/TIME OF TEST: 09/01/2025 00:03:15

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:00	
Control Test		0.081	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:01	Subject Test Refused
Breath Volume		0.000	00:01	
Air Blank		0.000	00:02	
Control Test		0.081	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 09/22/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,RYAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DAVIS,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 09/01/2025 13:22:03

Diagnostics Check		OK	13:15
Air Blank		0.000	13:16
Control Test		0.082	13:16
Air Blank		0.000	13:16
Subject Sample	# 1	0.195	13:17
Breath Volume		5.343	13:17
Air Blank		0.000	13:18
Air Blank		0.000	13:19
Subject Sample	# 2	0.197	13:20
Breath Volume		5.246	13:20
Air Blank		0.000	13:21
Control Test		0.080	13:21
Air Blank		0.000	13:21
Diagnostics Check		OK	13:22

Data Download Date/Time: 09/22/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDENAS,ADAM,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CARDENAS,A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:33

SUBJECT'S LAST NAME: BOONE

DATE/TIME OF TEST: 09/07/2025 17:01:09

Diagnostics Check		OK	16:53
Air Blank		0.000	16:53
Control Test		0.081	16:53
Air Blank		0.000	16:54
Subject Sample	# 1	0.254	16:56
Breath Volume		3.289	16:56
Air Blank		0.000	16:57
Air Blank		0.000	16:58
Subject Sample	# 2	0.243	16:59
Breath Volume		1.828	16:59
Air Blank		0.000	17:00
Control Test		0.080	17:00
Air Blank		0.000	17:01
Diagnostics Check		OK	17:01

Data Download Date/Time: 09/22/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOMAN,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HOMAN,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: BUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 09/07/2025 20:56:07

Diagnostics Check		OK	20:49
Air Blank		0.000	20:50
Control Test		0.081	20:50
Air Blank		0.000	20:51
Subject Sample	# 1	0.087	20:51
Breath Volume		4.472	20:51
Air Blank		0.000	20:52
Air Blank		0.000	20:54
Subject Sample	# 2	0.087	20:54
Breath Volume		4.320	20:54
Air Blank		0.000	20:55
Control Test		0.078	20:55
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 09/22/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,TRISTAN,T
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: ROGERS,TRISTAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: IVERSON

DATE/TIME OF TEST: 09/18/2025 01:26:09

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:23	
Control Test		0.081	01:24	
Air Blank		0.000	01:24	
Subject Sample	# 1		01:24	Subject Test Refused
Breath Volume		0.000	01:24	
Air Blank		0.000	01:25	
Control Test		0.082	01:25	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 09/22/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: LARGEN,JACKSON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: ETHRIDGE

DATE/TIME OF TEST: 09/19/2025 20:36:57

Diagnostics Check	Fail	20:36	Internal Printer Test
Air Blank	0.000	20:36	

Data Download Date/Time: 09/22/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNAHAN,TREVOR,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: LARGEN,JACKSON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: ETHRIDGE

DATE/TIME OF TEST: 09/19/2025 20:53:30

Diagnostics Check		OK	20:47
Air Blank		0.000	20:47
Control Test		0.081	20:48
Air Blank		0.000	20:48
Subject Sample	# 1	0.229	20:49
Breath Volume		2.160	20:49
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.221	20:51
Breath Volume		1.539	20:51
Air Blank		0.000	20:52
Control Test		0.080	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 09/22/2025 11:50

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 09/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: PREJS

DATE/TIME OF TEST: 09/29/2025 18:42:56

Diagnostics Check		OK	18:36
Air Blank		0.000	18:36
Control Test		0.081	18:37
Air Blank		0.000	18:37
Subject Sample	# 1	0.345	18:38
Breath Volume		2.617	18:38
Air Blank		0.000	18:39
Air Blank		0.000	18:40
Subject Sample	# 2	0.328	18:41
Breath Volume		1.773	18:41
Air Blank		0.000	18:42
Control Test		0.080	18:42
Air Blank		0.000	18:42
Diagnostics Check		OK	18:42

Data Download Date/Time: 10/16/2025 10:27

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: PHILLIPS,DONNIE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/01/2025 02:11:10

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.082	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.078	02:06
Breath Volume		1.414	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.077	02:09
Breath Volume		1.316	02:09
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 09/30/2025 22:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RUCCI,MATTHEW
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: CULPEPPER

DATE/TIME OF TEST: 09/05/2025 21:16:52

Diagnostics Check		OK	21:10
Air Blank		0.000	21:10
Control Test		0.081	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.217	21:12
Breath Volume		3.378	21:12
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.200	21:15
Breath Volume		2.734	21:15
Air Blank		0.000	21:15
Control Test		0.080	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:16

Data Download Date/Time: 09/30/2025 22:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DA SILVA,ANDRE,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: DA SILVA,ANDRE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 09/06/2025 18:33:59

Diagnostics Check		OK	18:30	
Air Blank		0.000	18:31	
Control Test		0.082	18:31	
Air Blank		0.000	18:32	
Subject Sample	# 1		18:32	Subject Test Refused
Breath Volume		0.000	18:32	
Air Blank		0.000	18:33	
Control Test		0.081	18:33	
Air Blank		0.000	18:33	
Diagnostics Check		OK	18:33	

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: VELEZ,SETH
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: GISPANSKI-WILFONG

DATE/TIME OF TEST: 09/07/2025 01:18:51

Diagnostics Check		OK	01:06
Air Blank		0.000	01:06
Control Test		0.081	01:07
Air Blank		0.000	01:07
Subject Sample	# 1	0.250	01:08
Breath Volume		1.480	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:10
Subject Sample	# 2	0.217	01:13
Breath Volume		1.125	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:15
Subject Sample	# 3	0.222	01:17
Breath Volume		1.343	01:17
Air Blank		0.000	01:17
Control Test		0.078	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYANT,RUBEN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 09/08/2025 03:39:40

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.082	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.151	03:35
Breath Volume		1.488	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.167	03:37
Breath Volume		1.425	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: DELOACH

DATE/TIME OF TEST: 09/10/2025 21:49:53

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:47	
Control Test		0.081	21:47	
Air Blank		0.000	21:47	
Subject Sample	# 1		21:49	Sequence Aborted
Breath Volume		0.000	21:49	
Air Blank		0.000	21:49	

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: ROCKEFELLER

DATE/TIME OF TEST: 09/11/2025 21:42:03

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.081	21:41	
Air Blank		21:41	RFI Detect
Air Blank	0.000	21:42	

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: ROCKEFELLER

DATE/TIME OF TEST: 09/11/2025 21:51:42

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.081	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.173	21:46
Breath Volume		1.414	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:48
Subject Sample	# 2	0.179	21:50
Breath Volume		2.531	21:50
Air Blank		0.000	21:50
Control Test		0.082	21:51
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,M,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: PHILLIPS,DONNIE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 09/14/2025 02:01:01

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.203	01:56
Breath Volume		1.351	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:58
Subject Sample	# 2	0.202	01:59
Breath Volume		1.285	01:59
Air Blank		0.000	02:00
Control Test		0.081	02:00
Air Blank		0.000	02:00
Diagnostics Check		OK	02:01

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,JONATHAN,E
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CORREA,JONATHAN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: EATON

DATE/TIME OF TEST: 09/15/2025 04:07:03

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.082	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.111	04:02
Breath Volume		2.132	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.115	04:05
Breath Volume		2.449	04:05
Air Blank		0.000	04:06
Control Test		0.082	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:07

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MILLER,CHRIS
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 09/16/2025 14:16:59

Diagnostics Check		OK	14:10
Air Blank		0.000	14:11
Control Test		0.081	14:11
Air Blank		0.000	14:12
Subject Sample	# 1	0.077	14:12
Breath Volume		2.000	14:12
Air Blank		0.000	14:13
Air Blank		0.000	14:14
Subject Sample	# 2	0.081	14:15
Breath Volume		2.203	14:15
Air Blank		0.000	14:16
Control Test		0.081	14:16
Air Blank		0.000	14:16
Diagnostics Check		OK	14:16

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: BROWNING

DATE/TIME OF TEST: 09/17/2025 17:29:36

Diagnostics Check		OK	17:21
Air Blank		0.000	17:21
Control Test		0.081	17:22
Air Blank		0.000	17:22
Subject Sample	# 1	0.227	17:24
Breath Volume		3.109	17:24
Air Blank		0.000	17:25
Air Blank		0.000	17:26
Subject Sample	# 2	0.224	17:27
Breath Volume		3.207	17:27
Air Blank		0.000	17:28
Control Test		0.080	17:28
Air Blank		0.000	17:29
Diagnostics Check		OK	17:29

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEROS,KOBE,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MARTINEZ,CHEYLA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 09:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: SEALE

DATE/TIME OF TEST: 09/18/2025 11:35:33

Diagnostics Check		OK	11:29
Air Blank		0.000	11:29
Control Test		0.081	11:30
Air Blank		0.000	11:30
Subject Sample	# 1	0.000	11:31
Breath Volume		3.015	11:31
Air Blank		0.000	11:31
Air Blank		0.000	11:33
Subject Sample	# 2	0.000	11:34
Breath Volume		2.437	11:34
Air Blank		0.000	11:34
Control Test		0.080	11:34
Air Blank		0.000	11:35
Diagnostics Check		OK	11:35

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,CHRIS,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: RUCCI,MATHEW
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: MAERZ

DATE/TIME OF TEST: 09/18/2025 16:26:29

Diagnostics Check		OK	16:20
Air Blank		0.000	16:20
Control Test		0.081	16:20
Air Blank		0.000	16:21
Subject Sample	# 1	0.175	16:21
Breath Volume		2.250	16:21
Air Blank		0.000	16:22
Air Blank		0.000	16:24
Subject Sample	# 2	0.171	16:24
Breath Volume		1.554	16:24
Air Blank		0.000	16:25
Control Test		0.080	16:25
Air Blank		0.000	16:26
Diagnostics Check		OK	16:26

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWERS,ABIGAYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: POWERS,ABIGAYLE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/18/2025 22:45:04

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.081	22:42	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.081	22:44	
Air Blank		0.000	22:44	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: LASTINGER

DATE/TIME OF TEST: 09/21/2025 00:38:25

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.268	00:33
Breath Volume		3.023	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.248	00:36
Breath Volume		1.765	00:36
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: CORSI

DATE/TIME OF TEST: 09/21/2025 21:17:01

Diagnostics Check		OK	21:11
Air Blank		0.000	21:11
Control Test		0.081	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.000	21:12
Breath Volume		1.460	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.000	21:15
Breath Volume		2.082	21:15
Air Blank		0.000	21:16
Control Test		0.081	21:16
Air Blank		0.000	21:16
Diagnostics Check		OK	21:17

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: RAINES

DATE/TIME OF TEST: 09/22/2025 00:28:56

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.081	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.174	00:24
Breath Volume		2.417	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.174	00:27
Breath Volume		2.160	00:27
Air Blank		0.000	00:28
Control Test		0.081	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIGLIACCIO,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: LARKIN

DATE/TIME OF TEST: 09/22/2025 03:45:19

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.000	03:41
Breath Volume		1.972	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.000	03:43
Breath Volume		1.996	03:43
Air Blank		0.000	03:44
Control Test		0.082	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOMBARDI,MADISON,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MARTINEZ,CHEYLA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 13:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:57

SUBJECT'S LAST NAME: SINCLAIR

DATE/TIME OF TEST: 09/22/2025 15:37:44

Diagnostics Check		OK	15:31
Air Blank		0.000	15:31
Control Test		0.081	15:32
Air Blank		0.000	15:32
Subject Sample	# 1	0.057	15:33
Breath Volume		3.515	15:33
Air Blank		0.000	15:33
Air Blank		0.000	15:35
Subject Sample	# 2	0.050	15:36
Breath Volume		3.203	15:36
Air Blank		0.000	15:36
Control Test		0.078	15:37
Air Blank		0.000	15:37
Diagnostics Check		OK	15:37

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHWINE,CHARLES,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MARQUEZ,JARED
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: DAVIES

DATE/TIME OF TEST: 09/22/2025 23:08:17

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.081	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.151	23:03
Breath Volume		1.417	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:05
Subject Sample	# 2	0.145	23:06
Breath Volume		2.035	23:06
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: CAMDEN,JAMES
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: BAUMGARTEN

DATE/TIME OF TEST: 09/25/2025 21:53:39

Diagnostics Check		OK	21:46
Air Blank		0.000	21:46
Control Test		0.082	21:47
Air Blank		0.000	21:47
Subject Sample	# 1	0.051	21:48
Breath Volume		1.382	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.055	21:52
Breath Volume		2.222	21:52
Air Blank		0.000	21:52
Control Test		0.080	21:53
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZ,BRANDON,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HARTZ,BRANDON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 12:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/26/2025 01:55:33

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.082	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.000	01:50
Breath Volume		1.242	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.000	01:54
Breath Volume		1.453	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHUMAKER,THOMAS,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SHUMAKER,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: EMIG

DATE/TIME OF TEST: 09/26/2025 17:28:51

Diagnostics Check		OK	17:22
Air Blank		0.000	17:22
Control Test		0.081	17:23
Air Blank		0.000	17:23
Subject Sample	# 1	0.246	17:24
Breath Volume		3.273	17:24
Air Blank		0.000	17:25
Air Blank		0.000	17:26
Subject Sample	# 2	0.247	17:27
Breath Volume		2.816	17:27
Air Blank		0.000	17:27
Control Test		0.080	17:28
Air Blank		0.000	17:28
Diagnostics Check		OK	17:28

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,M,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HINES,MIKO
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: VEIT

DATE/TIME OF TEST: 09/26/2025 19:29:32

Diagnostics Check		OK	19:23
Air Blank		0.000	19:23
Control Test		0.081	19:23
Air Blank		0.000	19:24
Subject Sample	# 1	0.231	19:24
Breath Volume		1.484	19:24
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.230	19:27
Breath Volume		1.734	19:27
Air Blank		0.000	19:28
Control Test		0.081	19:28
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,MIKO,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FRYER,JEROD
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: BAKER JR

DATE/TIME OF TEST: 09/26/2025 22:17:53

Diagnostics Check		OK	22:08
Air Blank		0.000	22:09
Control Test		0.081	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.252	22:10
Breath Volume		1.597	22:10
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.225	22:13
Breath Volume		2.410	22:13
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 3	0.207	22:16
Breath Volume		1.777	22:16
Air Blank		0.000	22:16
Control Test		0.080	22:17
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHWINE,CHARLES,B
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RICHWINE,CHARLES
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: OSVALDO

DATE/TIME OF TEST: 09/27/2025 05:06:58

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.081	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.115	05:02
Breath Volume		1.308	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.115	05:05
Breath Volume		1.457	05:05
Air Blank		0.000	05:06
Control Test		0.081	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ROMANO

DATE/TIME OF TEST: 09/28/2025 01:18:37

Diagnostics Check	OK	01:17	
Air Blank		01:18	RFI Detect
Air Blank	0.000	01:18	

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUSTER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SCHUSTER,ZACHARY
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ROMANO

DATE/TIME OF TEST: 09/28/2025 01:27:05

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.082	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.153	01:22
Breath Volume		2.175	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:24
Subject Sample	# 2	0.163	01:25
Breath Volume		2.601	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGOLD,K,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: DANIELS,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TINDLE

DATE/TIME OF TEST: 09/28/2025 03:37:35

Diagnostics Check		OK	03:26	
Air Blank		0.000	03:26	
Control Test		0.082	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 2		03:36	No Sample Provided
Breath Volume		0.000	03:36	
Air Blank		0.000	03:36	
Control Test		0.082	03:36	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORAN,SEAN,W
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: COLIN,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/28/2025 17:26:17

Diagnostics Check		OK	17:13	
Air Blank		0.000	17:14	
Control Test		0.082	17:14	
Air Blank		0.000	17:14	
Subject Sample	# 1	0.112	17:18	Volume Not Met
Breath Volume		0.914	17:18	
Air Blank		0.000	17:18	
Air Blank		0.000	17:20	
Subject Sample	# 2	0.125	17:21	
Breath Volume		1.359	17:21	
Air Blank		0.000	17:22	
Air Blank		0.000	17:24	
Subject Sample	# 3	0.116	17:24	
Breath Volume		1.785	17:24	
Air Blank		0.000	17:25	
Control Test		0.079	17:25	
Air Blank		0.000	17:26	
Diagnostics Check		OK	17:26	

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOFFORD,JOSEPH,T
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HARRIS,RYAN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 09/29/2025 04:53:32

Diagnostics Check		OK	04:46
Air Blank		0.000	04:47
Control Test		0.082	04:47
Air Blank		0.000	04:48
Subject Sample	# 1	0.238	04:48
Breath Volume		1.992	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:51
Subject Sample	# 2	0.251	04:51
Breath Volume		2.664	04:51
Air Blank		0.000	04:52
Control Test		0.079	04:52
Air Blank		0.000	04:53
Diagnostics Check		OK	04:53

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,LETOSHUA,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GREENE,LETOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:14

SUBJECT'S LAST NAME: STEPP

DATE/TIME OF TEST: 09/29/2025 15:37:35

Diagnostics Check		OK	15:30
Air Blank		0.000	15:31
Control Test		0.082	15:31
Air Blank		0.000	15:32
Subject Sample	# 1	0.303	15:32
Breath Volume		2.964	15:32
Air Blank		0.000	15:33
Air Blank		0.000	15:34
Subject Sample	# 2	0.302	15:35
Breath Volume		3.054	15:35
Air Blank		0.000	15:36
Control Test		0.079	15:36
Air Blank		0.000	15:37
Diagnostics Check		OK	15:37

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

INSTRUMENT REGISTERED TO: ST JOHNS SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALIFU,EZULDEEN,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SALIFU,EZULDEEN
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 08/28/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 09/29/2025 23:05:23

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:49	
Control Test		0.081	22:49	
Air Blank		0.000	22:50	
Subject Sample	# 1	0.147	22:53	Volume Not Met
Breath Volume		0.832	22:53	
Air Blank		0.000	22:54	
Air Blank		0.000	22:56	
Subject Sample	# 2	0.157	22:58	
Breath Volume		1.144	22:58	
Air Blank		0.000	22:59	
Air Blank		0.000	23:00	
Subject Sample	# 3	0.165	23:03	
Breath Volume		1.531	23:03	
Air Blank		0.000	23:04	
Control Test		0.076	23:04	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 09/30/2025 22:38

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

INSTRUMENT REGISTERED TO: SUWANNEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WAGNER,DUSTIN
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 09/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04120080A1
DRY GAS EXPIRY DATE: 03/05/2022
VIOLATION CODE: DUI
ARREST TIME: 10:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:29

SUBJECT'S LAST NAME: GOLENIEWSKI

DATE/TIME OF TEST: 09/30/2025 12:32:52

Diagnostics Check		OK	12:30	
Air Blank		0.000	12:30	
Control Test		0.079	12:30	
Air Blank		0.000	12:31	
Subject Sample	# 1		12:31	Subject Test Refused
Breath Volume		0.000	12:31	
Air Blank		0.000	12:31	
Control Test		0.078	12:32	
Air Blank		0.000	12:32	
Diagnostics Check		OK	12:32	

Data Download Date/Time: 10/31/2025 14:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: HADZALIC

DATE/TIME OF TEST: 09/20/2025 02:17:25

Diagnostics Check	Fail	02:16	Internal Printer Test
Air Blank	0.000	02:17	

Data Download Date/Time: 10/29/2025 21:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LOGAN,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: HADZALIC

DATE/TIME OF TEST: 09/20/2025 02:21:58

Diagnostics Check	Fail	02:21	Internal Printer Test
Air Blank	0.000	02:21	

Data Download Date/Time: 10/29/2025 21:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOGAN,ZACHARY,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/01/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: BOSCACCI

DATE/TIME OF TEST: 09/28/2025 22:08:06

Diagnostics Check	Fail	22:07	Internal Printer Test
Air Blank	0.000	22:08	

Data Download Date/Time: 10/29/2025 21:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 09/08/2025 01:00:26

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.078	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.000	00:55
Breath Volume		3.074	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.000	00:59
Breath Volume		2.472	00:59
Air Blank		0.000	00:59
Control Test		0.077	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 11/06/2025 04:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: DIAZ MONCHO

DATE/TIME OF TEST: 09/11/2025 02:01:23

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:51	
Control Test		0.078	01:51	
Air Blank		0.000	01:51	
Subject Sample	# 1	0.164	01:55	Volume Not Met
Breath Volume		0.015	01:55	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 2		01:59	Subject Test Refused
Breath Volume		0.843	01:59	
Air Blank		0.000	02:00	
Control Test		0.075	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 11/06/2025 04:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: BLANKENSHIP

DATE/TIME OF TEST: 09/16/2025 20:34:21

Diagnostics Check		OK	20:27
Air Blank		0.000	20:28
Control Test		0.077	20:28
Air Blank		0.000	20:29
Subject Sample	# 1	0.000	20:29
Breath Volume		1.562	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.000	20:32
Breath Volume		1.683	20:32
Air Blank		0.000	20:33
Control Test		0.077	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 11/06/2025 04:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007457

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDEMAN-STUART,JORDAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WILDEMAN-STUART,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 09/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: WALTON JR

DATE/TIME OF TEST: 09/21/2025 19:27:41

Diagnostics Check		OK	19:21
Air Blank		0.000	19:22
Control Test		0.079	19:22
Air Blank		0.000	19:23
Subject Sample	# 1	0.000	19:23
Breath Volume		2.351	19:23
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.000	19:26
Breath Volume		1.945	19:26
Air Blank		0.000	19:26
Control Test		0.078	19:27
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 11/06/2025 04:40

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007479

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: GALLON

DATE/TIME OF TEST: 09/02/2025 05:13:36

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:08	
Control Test		0.080	05:09	
Air Blank		0.000	05:09	
Subject Sample	# 1	0.212	05:11	
Breath Volume		1.460	05:11	
Air Blank			05:12	Purge Fail
Air Blank		0.000	05:13	

Data Download Date/Time: 09/18/2025 10:24

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KUSTKA

DATE/TIME OF TEST: 09/01/2025 01:12:02

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.258	01:07
Breath Volume		1.750	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:09
Subject Sample	# 2	0.250	01:10
Breath Volume		1.417	01:10
Air Blank		0.000	01:11
Control Test		0.078	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:12

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MORRIS,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: SACUL CABNAL

DATE/TIME OF TEST: 09/01/2025 02:40:51

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.079	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.180	02:36
Breath Volume		2.984	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:38
Subject Sample	# 2	0.183	02:39
Breath Volume		1.109	02:39
Air Blank		0.000	02:39
Control Test		0.078	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/01/2025 02:52:25

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.079	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.041	02:47
Breath Volume		1.597	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.039	02:50
Breath Volume		1.738	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: GAMEZ INFANTE

DATE/TIME OF TEST: 09/01/2025 03:11:36

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.079	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.173	03:06
Breath Volume		1.472	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.170	03:09
Breath Volume		2.019	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 09/01/2025 04:35:17

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:19	
Control Test		0.078	04:20	
Air Blank		0.000	04:20	
Subject Sample	# 1	0.183	04:23	
Breath Volume		1.160	04:23	
Air Blank		0.000	04:24	
Air Blank		0.000	04:26	
Subject Sample	# 2	0.161	04:29	Slope Not Level
Breath Volume		1.203	04:29	
Air Blank		0.000	04:30	
Air Blank		0.000	04:32	
Subject Sample	# 3	0.175	04:33	
Breath Volume		1.105	04:33	
Air Blank		0.000	04:34	
Control Test		0.075	04:34	
Air Blank		0.000	04:35	
Diagnostics Check		OK	04:35	

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: RUETZ

DATE/TIME OF TEST: 09/02/2025 00:36:08

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.078	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1	0.152	00:28	Volume Not Met
Breath Volume		0.050	00:28	
Air Blank		0.000	00:29	
Air Blank		0.000	00:31	
Subject Sample	# 2	0.166	00:34	Volume Not Met
Breath Volume		0.160	00:34	
Air Blank		0.000	00:35	
Control Test		0.077	00:35	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: HERNANDEZ-VENTURA

DATE/TIME OF TEST: 09/02/2025 02:11:51

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.078	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.072	02:06
Breath Volume		1.167	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.070	02:10
Breath Volume		1.902	02:10
Air Blank		0.000	02:10
Control Test		0.078	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: VO

DATE/TIME OF TEST: 09/02/2025 03:11:02

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:02	
Control Test		0.079	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 1	0.092	03:06	Volume Not Met
Breath Volume		0.121	03:06	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 2		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:10	
Control Test		0.076	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST.WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: BUTTON JR

DATE/TIME OF TEST: 09/02/2025 19:11:14

Diagnostics Check		OK	19:04
Air Blank		0.000	19:04
Control Test		0.079	19:05
Air Blank		0.000	19:05
Subject Sample	# 1	0.177	19:06
Breath Volume		1.531	19:06
Air Blank		0.000	19:07
Air Blank		0.000	19:09
Subject Sample	# 2	0.177	19:09
Breath Volume		1.121	19:09
Air Blank		0.000	19:10
Control Test		0.078	19:10
Air Blank		0.000	19:11
Diagnostics Check		OK	19:11

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 09/02/2025 20:16:37

Diagnostics Check		OK	20:09
Air Blank		0.000	20:10
Control Test		0.079	20:10
Air Blank		0.000	20:11
Subject Sample	# 1	0.000	20:12
Breath Volume		1.207	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.000	20:15
Breath Volume		1.828	20:15
Air Blank		0.000	20:15
Control Test		0.079	20:15
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: GAUTHIER

DATE/TIME OF TEST: 09/02/2025 21:23:22

Diagnostics Check		OK	21:12
Air Blank		0.000	21:12
Control Test		0.079	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.137	21:16
Breath Volume		1.359	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.118	21:21
Breath Volume		1.203	21:21
Air Blank		0.000	21:22
Control Test		0.077	21:22
Air Blank		0.000	21:23
Diagnostics Check		OK	21:23

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: JURICH

DATE/TIME OF TEST: 09/02/2025 23:05:26

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:51	
Control Test		0.079	22:51	
Air Blank		0.000	22:52	
Subject Sample	# 1	0.141	22:52	
Breath Volume		4.425	22:52	
Air Blank		0.000	22:53	
Air Blank		0.000	22:55	
Subject Sample	# 2		22:58	No Sample Provided
Breath Volume		0.000	22:58	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 3		23:04	No Sample Provided
Breath Volume		0.000	23:04	
Air Blank		0.000	23:04	
Control Test		0.079	23:04	
Air Blank		0.000	23:05	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: JURICH

DATE/TIME OF TEST: 09/02/2025 23:21:57

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.135	23:17
Breath Volume		4.453	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:19
Subject Sample	# 2	0.130	23:20
Breath Volume		4.765	23:20
Air Blank		0.000	23:21
Control Test		0.078	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: CLEGG

DATE/TIME OF TEST: 09/03/2025 03:57:54

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.078	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.146	03:53
Breath Volume		3.976	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.134	03:56
Breath Volume		1.914	03:56
Air Blank		0.000	03:56
Control Test		0.079	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: RICKETTS KID

DATE/TIME OF TEST: 09/03/2025 23:13:06

Diagnostics Check		OK	23:09	
Air Blank		0.000	23:10	
Control Test		0.079	23:10	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:11	Subject Test Refused
Breath Volume		0.000	23:11	
Air Blank		0.000	23:12	
Control Test		0.079	23:12	
Air Blank		0.000	23:13	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: NEWCOMB

DATE/TIME OF TEST: 09/04/2025 02:02:55

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.079	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1	0.192	01:51	Volume Not Met
Breath Volume		0.406	01:51	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 2	0.208	01:55	
Breath Volume		2.335	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 3	0.197	02:01	Volume Not Met
Breath Volume		0.929	02:01	
Air Blank		0.000	02:01	
Control Test		0.077	02:02	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: NEWCOMB

DATE/TIME OF TEST: 09/04/2025 02:20:29

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.075	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.195	02:07	
Breath Volume		1.492	02:07	
Air Blank		0.000	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 2	0.190	02:12	Volume Not Met
Breath Volume		0.089	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 3	0.204	02:18	Volume Not Met
Breath Volume		0.113	02:18	
Air Blank		0.000	02:19	
Control Test		0.075	02:19	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 09/04/2025 03:58:20

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.168	03:53
Breath Volume		3.625	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.180	03:56
Breath Volume		2.574	03:56
Air Blank		0.000	03:57
Control Test		0.078	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SCHRAMM

DATE/TIME OF TEST: 09/04/2025 04:31:35

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.078	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.122	04:26
Breath Volume		2.585	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.121	04:29
Breath Volume		1.734	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/05/2025 00:52:59

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.079	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:51	Subject Test Refused
Breath Volume		0.000	00:51	
Air Blank		0.000	00:52	
Control Test		0.078	00:52	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JACOBS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: APPLE

DATE/TIME OF TEST: 09/05/2025 03:44:48

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.078	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.173	03:39
Breath Volume		2.207	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.177	03:43
Breath Volume		3.039	03:43
Air Blank		0.000	03:43
Control Test		0.078	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 09/05/2025 21:55:16

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:52	
Control Test		0.079	21:52	
Air Blank		0.000	21:53	
Subject Sample	# 1		21:53	Subject Test Refused
Breath Volume		0.000	21:53	
Air Blank		0.000	21:54	
Control Test		0.079	21:54	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:55	

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PARADO

DATE/TIME OF TEST: 09/06/2025 02:00:20

Diagnostics Check		OK	01:53
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.163	01:55
Breath Volume		3.058	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.158	01:58
Breath Volume		2.480	01:58
Air Blank		0.000	01:59
Control Test		0.078	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 09/06/2025 05:20:00

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.079	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.171	05:15
Breath Volume		1.843	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:17
Subject Sample	# 2	0.170	05:18
Breath Volume		4.132	05:18
Air Blank		0.000	05:19
Control Test		0.079	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:20

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: O'LEARY

DATE/TIME OF TEST: 09/06/2025 08:12:43

Diagnostics Check		OK	08:05
Air Blank		0.000	08:06
Control Test		0.078	08:06
Air Blank		0.000	08:07
Subject Sample	# 1	0.213	08:07
Breath Volume		1.574	08:07
Air Blank		0.000	08:08
Air Blank		0.000	08:10
Subject Sample	# 2	0.215	08:11
Breath Volume		1.753	08:11
Air Blank		0.000	08:11
Control Test		0.077	08:12
Air Blank		0.000	08:12
Diagnostics Check		OK	08:12

Data Download Date/Time: 09/18/2025 10:35

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: PEREZ NOPAL

DATE/TIME OF TEST: 09/06/2025 20:50:40

Diagnostics Check		OK	20:44
Air Blank		0.000	20:44
Control Test		0.079	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.201	20:45
Breath Volume		2.902	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:48
Subject Sample	# 2	0.198	20:48
Breath Volume		3.007	20:48
Air Blank		0.000	20:49
Control Test		0.078	20:50
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: MUSCATO

DATE/TIME OF TEST: 09/06/2025 21:06:49

Diagnostics Check		OK	20:53	
Air Blank		0.000	20:53	
Control Test		0.079	20:53	
Air Blank		0.000	20:54	
Subject Sample	# 1	0.144	20:57	Volume Not Met
Breath Volume		0.031	20:57	
Air Blank		0.000	20:58	
Air Blank		0.000	21:00	
Subject Sample	# 2	0.175	21:00	
Breath Volume		1.402	21:00	
Air Blank		0.000	21:01	
Air Blank		0.000	21:03	
Subject Sample	# 3	0.158	21:05	
Breath Volume		1.449	21:05	
Air Blank		0.000	21:05	
Control Test		0.078	21:06	
Air Blank		0.000	21:06	
Diagnostics Check		OK	21:06	

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 09/07/2025 00:37:56

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.000	00:33
Breath Volume		2.523	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.000	00:36
Breath Volume		1.398	00:36
Air Blank		0.000	00:36
Control Test		0.079	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SUAREZ PEREZ

DATE/TIME OF TEST: 09/07/2025 01:39:01

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.078	01:36	
Air Blank		0.000	01:36	
Subject Sample	# 1		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:38	
Control Test		0.079	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: GARCIA GONZALEZ

DATE/TIME OF TEST: 09/07/2025 02:59:21

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.078	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1	0.084	02:53	Volume Not Met
Breath Volume		0.242	02:53	
Air Blank		0.000	02:53	
Air Blank		0.000	02:55	
Subject Sample	# 2		02:57	Subject Test Refused
Breath Volume		0.464	02:57	
Air Blank		0.000	02:58	
Control Test		0.077	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: GARCIA GONZALEZ

DATE/TIME OF TEST: 09/07/2025 03:08:08

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.078	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:06	Subject Test Refused
Breath Volume		0.355	03:06	
Air Blank		0.000	03:07	
Control Test		0.079	03:07	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:08	

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: WILBURN

DATE/TIME OF TEST: 09/07/2025 04:16:11

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.079	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.124	04:11
Breath Volume		4.113	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.119	04:14
Breath Volume		2.519	04:14
Air Blank		0.000	04:15
Control Test		0.078	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUG,Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: GOMEZ DIAZ

DATE/TIME OF TEST: 09/07/2025 04:27:19

Diagnostics Check	OK	04:25	
Air Blank	0.000	04:26	
Control Test	0.078	04:26	
Air Blank		04:26	RFI Detect
Air Blank	0.000	04:27	

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUG,Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: GOMEZ DIAZ

DATE/TIME OF TEST: 09/07/2025 04:36:10

Diagnostics Check		OK	04:29
Air Blank		0.000	04:29
Control Test		0.078	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.117	04:31
Breath Volume		2.480	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.111	04:34
Breath Volume		2.070	04:34
Air Blank		0.000	04:35
Control Test		0.078	04:35
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: GALAN PEREZ

DATE/TIME OF TEST: 09/07/2025 04:49:03

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.078	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.182	04:44
Breath Volume		1.878	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:46
Subject Sample	# 2	0.186	04:47
Breath Volume		1.386	04:47
Air Blank		0.000	04:48
Control Test		0.077	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:49

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/07/2025 05:00:10

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.078	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.133	04:54
Breath Volume		1.777	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.150	04:58
Breath Volume		2.226	04:58
Air Blank		0.000	04:59
Control Test		0.079	04:59
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/07/2025 05:16:40

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.078	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.188	05:11
Breath Volume		1.335	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.196	05:14
Breath Volume		2.917	05:14
Air Blank		0.000	05:15
Control Test		0.078	05:16
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: FAJIN PEREZ

DATE/TIME OF TEST: 09/07/2025 08:04:54

Diagnostics Check		OK	07:58
Air Blank		0.000	07:58
Control Test		0.078	07:59
Air Blank		0.000	07:59
Subject Sample	# 1	0.072	08:00
Breath Volume		1.179	08:00
Air Blank		0.000	08:00
Air Blank		0.000	08:02
Subject Sample	# 2	0.073	08:03
Breath Volume		1.332	08:03
Air Blank		0.000	08:03
Control Test		0.079	08:04
Air Blank		0.000	08:04
Diagnostics Check		OK	08:04

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: HELLER

DATE/TIME OF TEST: 09/07/2025 20:39:36

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.078	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.042	20:34
Breath Volume		1.812	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.040	20:37
Breath Volume		1.906	20:37
Air Blank		0.000	20:38
Control Test		0.078	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/07/2025 23:43:56

Diagnostics Check		OK	23:32
Air Blank		0.000	23:33
Control Test		0.079	23:33
Air Blank		0.000	23:34
Subject Sample	# 1	0.263	23:36
Breath Volume		1.140	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:38
Subject Sample	# 2	0.294	23:39
Breath Volume		2.082	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:41
Subject Sample	# 3	0.291	23:42
Breath Volume		1.843	23:42
Air Blank		0.000	23:42
Control Test		0.076	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PETERS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 09/08/2025 01:14:38

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.078	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.123	01:10
Breath Volume		2.734	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.120	01:12
Breath Volume		2.027	01:12
Air Blank		0.000	01:13
Control Test		0.078	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ST WILLIAMS,P
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: VALLEJO ROSA

DATE/TIME OF TEST: 09/09/2025 04:16:00

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.124	04:11
Breath Volume		1.992	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:13
Subject Sample	# 2	0.124	04:14
Breath Volume		1.312	04:14
Air Blank		0.000	04:15
Control Test		0.077	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:16

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KISS,L
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 09/09/2025 16:44:12

Diagnostics Check		OK	16:35
Air Blank		0.000	16:36
Control Test		0.078	16:36
Air Blank		0.000	16:37
Subject Sample	# 1	0.215	16:38
Breath Volume		1.597	16:38
Air Blank		0.000	16:39
Air Blank		0.000	16:40
Subject Sample	# 2	0.204	16:42
Breath Volume		1.789	16:42
Air Blank		0.000	16:43
Control Test		0.077	16:43
Air Blank		0.000	16:44
Diagnostics Check		OK	16:44

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TRUJILLO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BABIS

DATE/TIME OF TEST: 09/09/2025 20:13:07

Diagnostics Check		OK	20:00	
Air Blank		0.000	20:01	
Control Test		0.078	20:01	
Air Blank		0.000	20:02	
Subject Sample	# 1	0.222	20:05	Volume Not Met
Breath Volume		0.101	20:05	
Air Blank		0.000	20:05	
Air Blank		0.000	20:07	
Subject Sample	# 2	0.338	20:08	
Breath Volume		1.320	20:08	
Air Blank		0.000	20:09	
Air Blank		0.000	20:10	
Subject Sample	# 3	0.340	20:11	
Breath Volume		1.554	20:11	
Air Blank		0.000	20:12	
Control Test		0.076	20:12	
Air Blank		0.000	20:13	
Diagnostics Check		OK	20:13	

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: VAVRINA

DATE/TIME OF TEST: 09/09/2025 23:28:55

Diagnostics Check		OK	23:16	
Air Blank		0.000	23:16	
Control Test		0.078	23:16	
Air Blank		0.000	23:17	
Subject Sample	# 1	0.054	23:20	Volume Not Met
Breath Volume		1.082	23:20	
Air Blank		0.000	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 2	0.056	23:23	
Breath Volume		1.535	23:23	
Air Blank		0.000	23:24	
Air Blank		0.000	23:26	
Subject Sample	# 3	0.054	23:27	
Breath Volume		1.550	23:27	
Air Blank		0.000	23:27	
Control Test		0.078	23:28	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 09/18/2025 10:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: PEREIRA LETTIERE

DATE/TIME OF TEST: 09/10/2025 23:30:06

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.079	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.052	23:25
Breath Volume		1.460	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:27
Subject Sample	# 2	0.048	23:28
Breath Volume		1.691	23:28
Air Blank		0.000	23:29
Control Test		0.078	23:29
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 09/11/2025 04:10:41

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.119	04:05
Breath Volume		1.644	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.116	04:08
Breath Volume		1.597	04:08
Air Blank		0.000	04:09
Control Test		0.078	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: BEYER

DATE/TIME OF TEST: 09/12/2025 22:23:46

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.079	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.220	22:18
Breath Volume		1.742	22:18
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.210	22:21
Breath Volume		1.765	22:21
Air Blank		0.000	22:22
Control Test		0.078	22:23
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LAITINEN,H
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: COMM VEHIC
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: VOKATY

DATE/TIME OF TEST: 09/13/2025 01:23:57

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.055	01:19
Breath Volume		3.117	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:21
Subject Sample	# 2	0.058	01:22
Breath Volume		3.734	01:22
Air Blank		0.000	01:22
Control Test		0.079	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: COYOTE SIR

DATE/TIME OF TEST: 09/13/2025 02:34:14

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.079	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.187	02:29
Breath Volume		1.484	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.187	02:32
Breath Volume		1.496	02:32
Air Blank		0.000	02:33
Control Test		0.078	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: BEALL

DATE/TIME OF TEST: 09/13/2025 03:25:45

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.079	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.199	03:21
Breath Volume		2.007	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.202	03:24
Breath Volume		1.113	03:24
Air Blank		0.000	03:24
Control Test		0.077	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SERRANO HERRERA

DATE/TIME OF TEST: 09/13/2025 03:41:33

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.213	03:36
Breath Volume		1.207	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.218	03:39
Breath Volume		1.414	03:39
Air Blank		0.000	03:40
Control Test		0.077	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: SANN

DATE/TIME OF TEST: 09/13/2025 04:25:42

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.078	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.252	04:21
Breath Volume		2.027	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.249	04:24
Breath Volume		1.160	04:24
Air Blank		0.000	04:24
Control Test		0.081	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 09/13/2025 21:42:30

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.079	21:35
Air Blank		0.000	21:36
Subject Sample	# 1	0.081	21:37
Breath Volume		1.382	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.081	21:40
Breath Volume		1.347	21:40
Air Blank		0.000	21:41
Control Test		0.078	21:41
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: LOPEZ VASQUEZ

DATE/TIME OF TEST: 09/13/2025 23:50:46

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:39	
Control Test		0.078	23:39	
Air Blank		0.000	23:39	
Subject Sample	# 1	0.156	23:43	Volume Not Met
Breath Volume		0.542	23:43	
Air Blank		0.000	23:43	
Air Blank		0.000	23:45	
Subject Sample	# 2	0.179	23:46	
Breath Volume		1.101	23:46	
Air Blank		0.000	23:46	
Air Blank		0.000	23:48	
Subject Sample	# 3	0.180	23:49	
Breath Volume		1.925	23:49	
Air Blank		0.000	23:49	
Control Test		0.076	23:50	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MORALES MOCTEZUMA

DATE/TIME OF TEST: 09/14/2025 01:32:50

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.225	01:28
Breath Volume		2.515	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.223	01:31
Breath Volume		3.312	01:31
Air Blank		0.000	01:31
Control Test		0.077	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 09/18/2025 10:36

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MORALES AGUIRRE

DATE/TIME OF TEST: 09/14/2025 02:35:21

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.118	02:30
Breath Volume		2.757	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.117	02:33
Breath Volume		2.058	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOUIRBI,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/14/2025 04:03:51

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.145	03:59
Breath Volume		1.335	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:01
Subject Sample	# 2	0.148	04:02
Breath Volume		1.187	04:02
Air Blank		0.000	04:02
Control Test		0.078	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CRUZ BAEZ

DATE/TIME OF TEST: 09/14/2025 04:23:50

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.164	04:18
Breath Volume		1.597	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.167	04:22
Breath Volume		2.035	04:22
Air Blank		0.000	04:22
Control Test		0.078	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MENTION

DATE/TIME OF TEST: 09/14/2025 04:47:54

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.078	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.150	04:42
Breath Volume		1.816	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.149	04:46
Breath Volume		1.984	04:46
Air Blank		0.000	04:46
Control Test		0.078	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: KUHLMAN

DATE/TIME OF TEST: 09/14/2025 05:13:58

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.078	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.115	05:09
Breath Volume		1.828	05:09
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.111	05:12
Breath Volume		1.914	05:12
Air Blank		0.000	05:13
Control Test		0.078	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/14/2025 06:37:35

Diagnostics Check		OK	06:29
Air Blank		0.000	06:30
Control Test		0.078	06:30
Air Blank		0.000	06:31
Subject Sample	# 1	0.127	06:32
Breath Volume		1.894	06:32
Air Blank		0.000	06:33
Air Blank		0.000	06:34
Subject Sample	# 2	0.125	06:35
Breath Volume		1.378	06:35
Air Blank		0.000	06:36
Control Test		0.078	06:36
Air Blank		0.000	06:37
Diagnostics Check		OK	06:37

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MITCHELL,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:31

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 09/14/2025 08:58:19

Diagnostics Check		OK	08:51
Air Blank		0.000	08:52
Control Test		0.078	08:52
Air Blank		0.000	08:53
Subject Sample	# 1	0.120	08:53
Breath Volume		1.726	08:53
Air Blank		0.000	08:54
Air Blank		0.000	08:56
Subject Sample	# 2	0.127	08:56
Breath Volume		1.757	08:56
Air Blank		0.000	08:57
Control Test		0.079	08:57
Air Blank		0.000	08:58
Diagnostics Check		OK	08:58

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: WALDREP

DATE/TIME OF TEST: 09/15/2025 01:08:11

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.157	01:03
Breath Volume		2.656	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.159	01:06
Breath Volume		2.953	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CASTILLO DIAZ

DATE/TIME OF TEST: 09/15/2025 02:28:42

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.190	02:23
Breath Volume		1.695	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.182	02:26
Breath Volume		1.890	02:26
Air Blank		0.000	02:27
Control Test		0.076	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 09/16/2025 05:42:12

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.079	05:35
Air Blank		0.000	05:36
Subject Sample	# 1	0.113	05:37
Breath Volume		1.679	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.112	05:40
Breath Volume		1.714	05:40
Air Blank		0.000	05:41
Control Test		0.079	05:41
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: ZHENG

DATE/TIME OF TEST: 09/16/2025 23:51:38

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.133	23:47
Breath Volume		2.492	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.130	23:49
Breath Volume		2.464	23:49
Air Blank		0.000	23:50
Control Test		0.078	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ALMAGUER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: ORDUNA MARTINEZ

DATE/TIME OF TEST: 09/17/2025 01:20:35

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.183	01:15
Breath Volume		1.429	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.187	01:18
Breath Volume		1.835	01:18
Air Blank		0.000	01:19
Control Test		0.079	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PIERRE,C
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 09/17/2025 03:56:54

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.000	03:52
Breath Volume		2.105	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.000	03:55
Breath Volume		1.800	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JIMENEZ ALVARENGA

DATE/TIME OF TEST: 09/17/2025 22:27:19

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.079	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.246	22:22
Breath Volume		2.000	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.245	22:25
Breath Volume		1.964	22:25
Air Blank		0.000	22:26
Control Test		0.078	22:26
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MAROCOLA,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: DEFILIPPIS

DATE/TIME OF TEST: 09/18/2025 01:02:55

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.000	00:57
Breath Volume		1.406	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.000	01:01
Breath Volume		1.472	01:01
Air Blank		0.000	01:01
Control Test		0.078	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 09/18/2025 10:37

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: HERNANDEZ VELAZQUEZ

DATE/TIME OF TEST: 09/18/2025 19:55:44

Diagnostics Check		OK	19:49
Air Blank		0.000	19:49
Control Test		0.079	19:49
Air Blank		0.000	19:50
Subject Sample	# 1	0.199	19:51
Breath Volume		2.550	19:51
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.200	19:53
Breath Volume		2.121	19:53
Air Blank		0.000	19:54
Control Test		0.078	19:55
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 10/16/2025 12:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JACOBS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: MOHR

DATE/TIME OF TEST: 09/18/2025 21:55:44

Diagnostics Check		OK	21:48
Air Blank		0.000	21:49
Control Test		0.079	21:49
Air Blank		0.000	21:50
Subject Sample	# 1	0.310	21:51
Breath Volume		3.250	21:51
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.303	21:53
Breath Volume		2.621	21:53
Air Blank		0.000	21:54
Control Test		0.078	21:55
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 10/16/2025 12:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 09/18/2025 23:27:28

Diagnostics Check		OK	23:23	
Air Blank		0.000	23:23	
Control Test		0.079	23:24	
Air Blank		0.000	23:24	
Subject Sample	# 1		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:26	
Control Test		0.078	23:26	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 10/16/2025 12:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 09/19/2025 00:00:53

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.079	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.057	23:56
Breath Volume		1.878	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.055	23:59
Breath Volume		1.882	23:59
Air Blank		0.000	23:59
Control Test		0.079	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 10/16/2025 12:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ROMO-SAIFFE

DATE/TIME OF TEST: 09/19/2025 01:51:12

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.181	01:46
Breath Volume		2.109	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:48
Subject Sample	# 2	0.171	01:49
Breath Volume		1.500	01:49
Air Blank		0.000	01:50
Control Test		0.078	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 10/16/2025 12:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMOND,JENNY,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CUTINO,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: MUSTAIN

DATE/TIME OF TEST: 09/19/2025 06:38:38

Diagnostics Check		OK	06:30
Air Blank		0.000	06:31
Control Test		0.079	06:31
Air Blank		0.000	06:32
Subject Sample	# 1	0.112	06:33
Breath Volume		1.667	06:33
Air Blank		0.000	06:33
Air Blank		0.000	06:35
Subject Sample	# 2	0.123	06:36
Breath Volume		2.195	06:36
Air Blank		0.000	06:37
Control Test		0.079	06:37
Air Blank		0.000	06:38
Diagnostics Check		OK	06:38

Data Download Date/Time: 10/16/2025 12:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PETERS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: BOLEN

DATE/TIME OF TEST: 09/19/2025 20:46:02

Diagnostics Check		OK	20:36
Air Blank		0.000	20:37
Control Test		0.079	20:37
Air Blank		0.000	20:38
Subject Sample	# 1	0.198	20:40
Breath Volume		1.984	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.192	20:44
Breath Volume		1.425	20:44
Air Blank		0.000	20:45
Control Test		0.077	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:46

Data Download Date/Time: 10/16/2025 12:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: ROBIRA

DATE/TIME OF TEST: 09/19/2025 23:07:03

Diagnostics Check		OK	22:52	
Air Blank		0.000	22:52	
Control Test		0.079	22:53	
Air Blank		0.000	22:53	
Subject Sample	# 1	0.084	22:55	
Breath Volume		1.117	22:55	
Air Blank		0.000	22:56	
Air Blank		0.000	22:57	
Subject Sample	# 2	0.077	23:00	Volume Not Met
Breath Volume		0.269	23:00	
Air Blank		0.000	23:01	
Air Blank		0.000	23:03	
Subject Sample	# 3		23:05	Subject Test Refused
Breath Volume		1.011	23:05	
Air Blank		0.000	23:06	
Control Test		0.078	23:06	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 10/16/2025 12:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: GONZALEZ GERTZ

DATE/TIME OF TEST: 09/20/2025 00:52:34

Diagnostics Check		OK	00:45
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.077	00:47
Breath Volume		1.554	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.075	00:50
Breath Volume		1.765	00:50
Air Blank		0.000	00:51
Control Test		0.078	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 10/16/2025 12:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/20/2025 04:05:08

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.126	04:00
Breath Volume		1.449	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.138	04:03
Breath Volume		1.753	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 10/16/2025 12:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: PEGUERO

DATE/TIME OF TEST: 09/20/2025 04:20:26

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.078	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.155	04:15
Breath Volume		2.406	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.143	04:18
Breath Volume		2.449	04:18
Air Blank		0.000	04:19
Control Test		0.078	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 10/16/2025 12:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: MEDINA CERON

DATE/TIME OF TEST: 09/20/2025 05:28:02

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:14	
Control Test		0.079	05:14	
Air Blank		0.000	05:15	
Subject Sample	# 1	0.106	05:18	Volume Not Met
Breath Volume		0.734	05:18	
Air Blank		0.000	05:19	
Air Blank		0.000	05:21	
Subject Sample	# 2	0.122	05:21	
Breath Volume		1.132	05:21	
Air Blank		0.000	05:22	
Air Blank		0.000	05:24	
Subject Sample	# 3	0.107	05:26	
Breath Volume		1.289	05:26	
Air Blank		0.000	05:27	
Control Test		0.078	05:27	
Air Blank		0.000	05:27	
Diagnostics Check		OK	05:28	

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUG,Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/20/2025 20:20:15

Diagnostics Check		OK	20:13
Air Blank		0.000	20:13
Control Test		0.079	20:13
Air Blank		0.000	20:14
Subject Sample	# 1	0.108	20:15
Breath Volume		1.195	20:15
Air Blank		0.000	20:15
Air Blank		0.000	20:17
Subject Sample	# 2	0.105	20:18
Breath Volume		1.425	20:18
Air Blank		0.000	20:19
Control Test		0.079	20:19
Air Blank		0.000	20:20
Diagnostics Check		OK	20:20

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 09/21/2025 00:24:40

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.167	00:20
Breath Volume		1.214	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.156	00:22
Breath Volume		2.851	00:22
Air Blank		0.000	00:23
Control Test		0.078	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: APONTE RODRIGUEZ

DATE/TIME OF TEST: 09/21/2025 00:49:33

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.079	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.087	00:44
Breath Volume		2.472	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.086	00:47
Breath Volume		1.593	00:47
Air Blank		0.000	00:48
Control Test		0.078	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CARDONA-AGUILERA

DATE/TIME OF TEST: 09/21/2025 01:49:42

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.078	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.152	01:45
Breath Volume		2.175	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.158	01:48
Breath Volume		2.003	01:48
Air Blank		0.000	01:48
Control Test		0.078	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VELEZ,A
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: YOUNG JR

DATE/TIME OF TEST: 09/21/2025 02:52:36

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.128	02:48
Breath Volume		2.148	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.130	02:50
Breath Volume		1.945	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TOUIRBI,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/21/2025 03:51:18

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.167	03:46
Breath Volume		2.617	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:48
Subject Sample	# 2	0.157	03:49
Breath Volume		1.640	03:49
Air Blank		0.000	03:50
Control Test		0.078	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PETERS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: DORSETT

DATE/TIME OF TEST: 09/21/2025 05:46:23

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.079	05:40
Air Blank		0.000	05:41
Subject Sample	# 1	0.159	05:41
Breath Volume		1.355	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.154	05:44
Breath Volume		1.582	05:44
Air Blank		0.000	05:45
Control Test		0.079	05:45
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAZQUEZ,V
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:13

SUBJECT'S LAST NAME: DIAZ DIAZ

DATE/TIME OF TEST: 09/21/2025 09:41:22

Diagnostics Check		OK	09:34
Air Blank		0.000	09:34
Control Test		0.079	09:35
Air Blank		0.000	09:35
Subject Sample	# 1	0.141	09:36
Breath Volume		1.632	09:36
Air Blank		0.000	09:37
Air Blank		0.000	09:39
Subject Sample	# 2	0.143	09:39
Breath Volume		1.722	09:39
Air Blank		0.000	09:40
Control Test		0.078	09:40
Air Blank		0.000	09:41
Diagnostics Check		OK	09:41

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KELLY,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:51

SUBJECT'S LAST NAME: BERTONE

DATE/TIME OF TEST: 09/21/2025 12:25:29

Diagnostics Check		OK	12:13	
Air Blank		0.000	12:13	
Control Test		0.079	12:13	
Air Blank		0.000	12:14	
Subject Sample	# 1		12:17	No Sample Provided
Breath Volume		0.000	12:17	
Air Blank		0.000	12:17	
Air Blank		0.000	12:19	
Subject Sample	# 2	0.150	12:20	
Breath Volume		1.601	12:20	
Air Blank		0.000	12:21	
Air Blank		0.000	12:23	
Subject Sample	# 3	0.147	12:23	
Breath Volume		1.988	12:23	
Air Blank		0.000	12:24	
Control Test		0.078	12:24	
Air Blank		0.000	12:25	
Diagnostics Check		OK	12:25	

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RYAN,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: SALAS FUENTES

DATE/TIME OF TEST: 09/21/2025 16:49:28

Diagnostics Check		OK	16:45	
Air Blank		0.000	16:46	
Control Test		0.079	16:46	
Air Blank		0.000	16:47	
Subject Sample	# 1		16:47	Interferent Detect
Breath Volume		1.691	16:47	
Air Blank		0.000	16:48	
Control Test		0.079	16:48	
Air Blank		0.000	16:49	
Diagnostics Check		OK	16:49	

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RYAN,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: SALAS FUENTES

DATE/TIME OF TEST: 09/21/2025 17:15:00

Diagnostics Check		OK	17:10	
Air Blank		0.000	17:10	
Control Test		0.079	17:11	
Air Blank		0.000	17:11	
Subject Sample	# 1		17:13	Interferent Detect
Breath Volume		1.851	17:13	
Air Blank		0.000	17:14	
Control Test		0.078	17:14	
Air Blank		0.000	17:14	
Diagnostics Check		OK	17:15	

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKIMS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: REDUCINDO

DATE/TIME OF TEST: 09/22/2025 02:10:03

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.106	02:05
Breath Volume		1.953	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:07
Subject Sample	# 2	0.107	02:08
Breath Volume		1.421	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:10

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MAROCOLA,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: BETANCURT SALCEDO

DATE/TIME OF TEST: 09/22/2025 21:04:13

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.140	20:59
Breath Volume		1.800	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.142	21:02
Breath Volume		2.125	21:02
Air Blank		0.000	21:03
Control Test		0.078	21:03
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: GRANDE BASILIO

DATE/TIME OF TEST: 09/22/2025 23:02:35

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.258	22:58
Breath Volume		2.199	22:58
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.251	23:00
Breath Volume		1.867	23:00
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 10/16/2025 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/23/2025 03:39:29

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.132	03:34
Breath Volume		2.324	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.128	03:37
Breath Volume		2.585	03:37
Air Blank		0.000	03:38
Control Test		0.078	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAVER,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/23/2025 16:57:14

Diagnostics Check		OK	16:50
Air Blank		0.000	16:51
Control Test		0.079	16:51
Air Blank		0.000	16:52
Subject Sample	# 1	0.350	16:52
Breath Volume		2.605	16:52
Air Blank		0.000	16:53
Air Blank		0.000	16:55
Subject Sample	# 2	0.350	16:55
Breath Volume		2.714	16:55
Air Blank		0.000	16:56
Control Test		0.078	16:56
Air Blank		0.000	16:57
Diagnostics Check		OK	16:57

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RENDON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 09/24/2025 05:08:49

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:57	
Control Test		0.079	04:57	
Air Blank		0.000	04:58	
Subject Sample	# 1	0.172	05:01	Volume Not Met
Breath Volume		0.164	05:01	
Air Blank		0.000	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 2		05:07	Subject Test Refused
Breath Volume		0.082	05:07	
Air Blank		0.000	05:07	
Control Test		0.077	05:08	
Air Blank		0.000	05:08	
Diagnostics Check		OK	05:08	

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: NAVA CASARRUBIAS

DATE/TIME OF TEST: 09/25/2025 00:12:24

Diagnostics Check		OK	00:05
Air Blank		0.000	00:06
Control Test		0.079	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.279	00:07
Breath Volume		2.152	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.265	00:10
Breath Volume		1.472	00:10
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 09/25/2025 02:48:11

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.219	02:43
Breath Volume		2.410	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:45
Subject Sample	# 2	0.235	02:46
Breath Volume		1.890	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANT,TIFFANY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/26/2025 00:36:06

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.000	00:31
Breath Volume		2.750	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.000	00:34
Breath Volume		2.109	00:34
Air Blank		0.000	00:35
Control Test		0.079	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 10/16/2025 12:18

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BURKHOLDER

DATE/TIME OF TEST: 09/26/2025 00:50:56

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.041	00:46
Breath Volume		2.757	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:48
Subject Sample	# 2	0.041	00:49
Breath Volume		3.167	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALL,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: BANG

DATE/TIME OF TEST: 09/26/2025 02:48:40

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.079	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.169	02:44
Breath Volume		2.179	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.156	02:46
Breath Volume		1.722	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/26/2025 04:16:10

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:08	
Control Test		0.079	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:12	No Sample Provided
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:15	
Control Test		0.079	04:15	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RAY,K
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: WARE

DATE/TIME OF TEST: 09/26/2025 19:48:16

Diagnostics Check		OK	19:41
Air Blank		0.000	19:42
Control Test		0.080	19:42
Air Blank		0.000	19:43
Subject Sample	# 1	0.139	19:43
Breath Volume		2.304	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.128	19:46
Breath Volume		1.136	19:46
Air Blank		0.000	19:47
Control Test		0.079	19:47
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LIGGINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: COMM VEHIC
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 09/27/2025 01:19:56

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.026	01:15
Breath Volume		1.484	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.025	01:18
Breath Volume		1.101	01:18
Air Blank		0.000	01:18
Control Test		0.079	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: DIAZ VAZQUEZ

DATE/TIME OF TEST: 09/27/2025 02:43:49

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.079	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.240	02:38
Breath Volume		1.945	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.239	02:42
Breath Volume		1.617	02:42
Air Blank		0.000	02:42
Control Test		0.078	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: LOPEZ FRANCO

DATE/TIME OF TEST: 09/27/2025 03:26:05

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.079	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.115	03:21
Breath Volume		2.355	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:23
Subject Sample	# 2	0.112	03:24
Breath Volume		1.828	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KRUG,Z
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: LOPEZ GONZALEZ

DATE/TIME OF TEST: 09/27/2025 04:13:01

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.079	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.122	04:08
Breath Volume		1.468	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.117	04:11
Breath Volume		1.656	04:11
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:13

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROME,BRITTNEY,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: SIMMONS JR

DATE/TIME OF TEST: 09/27/2025 04:45:29

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.079	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.104	04:40
Breath Volume		1.820	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:42
Subject Sample	# 2	0.103	04:43
Breath Volume		1.640	04:43
Air Blank		0.000	04:44
Control Test		0.080	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 09/27/2025 05:09:22

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.079	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.237	05:04
Breath Volume		1.972	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.248	05:07
Breath Volume		1.363	05:07
Air Blank		0.000	05:08
Control Test		0.078	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: CALDERON LOPEZ

DATE/TIME OF TEST: 09/27/2025 21:03:55

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.236	20:59
Breath Volume		1.503	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:01
Subject Sample	# 2	0.219	21:02
Breath Volume		1.457	21:02
Air Blank		0.000	21:02
Control Test		0.078	21:03
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 09/27/2025 23:28:51

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:18	
Control Test		0.079	23:18	
Air Blank		0.000	23:19	
Subject Sample	# 1		23:22	No Sample Provided
Breath Volume		0.000	23:22	
Air Blank		0.000	23:22	
Air Blank		0.000	23:24	
Subject Sample	# 2		23:27	Subject Test Refused
Breath Volume		0.000	23:27	
Air Blank		0.000	23:27	
Control Test		0.079	23:28	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 09/28/2025 02:32:52

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.158	02:27
Breath Volume		1.816	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.159	02:31
Breath Volume		1.179	02:31
Air Blank		0.000	02:31
Control Test		0.078	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JACOBS,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: GAYTAN

DATE/TIME OF TEST: 09/28/2025 03:02:02

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.078	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.172	02:57
Breath Volume		1.597	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.179	03:00
Breath Volume		1.480	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:02

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,DESMAND,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ABNEY

DATE/TIME OF TEST: 09/28/2025 04:25:37

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.096	04:21
Breath Volume		1.753	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.098	04:23
Breath Volume		1.367	04:23
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 10/16/2025 12:19

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/28/2025 06:17:53

Diagnostics Check		OK	06:08
Air Blank		0.000	06:09
Control Test		0.079	06:09
Air Blank		0.000	06:10
Subject Sample	# 1	0.218	06:12
Breath Volume		1.171	06:12
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.220	06:16
Breath Volume		1.789	06:16
Air Blank		0.000	06:16
Control Test		0.075	06:17
Air Blank		0.000	06:17
Diagnostics Check		OK	06:17

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,JESSICA,M
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MAXIME,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:33

SUBJECT'S LAST NAME: VARGAS RAMIREZ

DATE/TIME OF TEST: 09/28/2025 12:00:31

Diagnostics Check		OK	11:54
Air Blank		0.000	11:54
Control Test		0.079	11:55
Air Blank		0.000	11:55
Subject Sample	# 1	0.010	11:56
Breath Volume		1.136	11:56
Air Blank		0.000	11:56
Air Blank		0.000	11:58
Subject Sample	# 2	0.010	11:59
Breath Volume		1.238	11:59
Air Blank		0.000	11:59
Control Test		0.079	11:59
Air Blank		0.000	12:00
Diagnostics Check		OK	12:00

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JENKINS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/28/2025 23:19:24

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.079	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1	0.179	23:09	
Breath Volume		1.867	23:09	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 2	0.142	23:14	Volume Not Met
Breath Volume		0.523	23:14	
Air Blank		0.000	23:15	
Air Blank		0.000	23:17	
Subject Sample	# 3	0.179	23:17	
Breath Volume		2.367	23:17	
Air Blank		0.000	23:18	
Control Test		0.075	23:18	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 10/16/2025 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUMAN ROMERO,ANGIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 09/29/2025 01:20:46

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.138	01:16
Breath Volume		3.015	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.138	01:19
Breath Volume		1.710	01:19
Air Blank		0.000	01:19
Control Test		0.078	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 09/29/2025 02:59:48

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.079	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1	0.186	02:52	Volume Not Met
Breath Volume		0.203	02:52	
Air Blank		0.000	02:53	
Air Blank		0.000	02:54	
Subject Sample	# 2	0.197	02:57	Volume Not Met
Breath Volume		0.226	02:57	
Air Blank		0.000	02:58	
Control Test		0.077	02:59	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 10/16/2025 12:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: URIZAR PEREZ

DATE/TIME OF TEST: 09/29/2025 03:13:26

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.076	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.209	03:08
Breath Volume		1.515	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.197	03:11
Breath Volume		1.707	03:11
Air Blank		0.000	03:12
Control Test		0.077	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PEPEN,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 09/30/2025 01:49:51

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.000	01:45
Breath Volume		1.320	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.000	01:48
Breath Volume		1.492	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,SERENA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BUSWELL,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:24

SUBJECT'S LAST NAME: NETO

DATE/TIME OF TEST: 09/30/2025 11:52:51

Diagnostics Check		OK	11:45
Air Blank		0.000	11:45
Control Test		0.079	11:45
Air Blank		0.000	11:46
Subject Sample	# 1	0.011	11:48
Breath Volume		1.710	11:48
Air Blank		0.000	11:48
Air Blank		0.000	11:50
Subject Sample	# 2	0.010	11:51
Breath Volume		1.742	11:51
Air Blank		0.000	11:51
Control Test		0.078	11:52
Air Blank		0.000	11:52
Diagnostics Check		OK	11:52

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007482

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 09/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: PROST

DATE/TIME OF TEST: 09/30/2025 21:15:39

Diagnostics Check		OK	21:08
Air Blank		0.000	21:09
Control Test		0.080	21:09
Air Blank		0.000	21:09
Subject Sample	# 1	0.150	21:11
Breath Volume		2.207	21:11
Air Blank		0.000	21:11
Air Blank		0.000	21:13
Subject Sample	# 2	0.147	21:13
Breath Volume		2.000	21:13
Air Blank		0.000	21:14
Control Test		0.079	21:15
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 10/16/2025 12:20

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ABROMAVICIUS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ROSEME

DATE/TIME OF TEST: 09/05/2025 01:04:58

Diagnostics Check		OK	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.159	01:00
Breath Volume		1.421	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.160	01:03
Breath Volume		1.535	01:03
Air Blank		0.000	01:04
Control Test		0.077	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/05/2025 01:59:54

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.174	01:55
Breath Volume		2.085	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:57
Subject Sample	# 2	0.168	01:58
Breath Volume		2.152	01:58
Air Blank		0.000	01:58
Control Test		0.078	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/05/2025 05:02:22

Diagnostics Check		OK	04:59	
Air Blank		0.000	04:59	
Control Test		0.079	04:59	
Air Blank		0.000	05:00	
Subject Sample	# 1		05:00	Subject Test Refused
Breath Volume		0.000	05:00	
Air Blank		0.000	05:01	
Control Test		0.078	05:01	
Air Blank		0.000	05:02	
Diagnostics Check		OK	05:02	

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SUAREZ,O
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ESTRELLA

DATE/TIME OF TEST: 09/05/2025 22:12:48

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:06	
Control Test		0.078	22:06	
Air Blank		0.000	22:06	
Subject Sample	# 1	0.000	22:08	
Breath Volume		1.773	22:08	
Air Blank		0.000	22:09	
Air Blank		0.000	22:11	
Subject Sample	# 2		22:11	Subject Test Refused
Breath Volume		0.000	22:11	
Air Blank		0.000	22:11	
Control Test		0.078	22:12	
Air Blank		0.000	22:12	
Diagnostics Check		OK	22:12	

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,ALECIA,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: SANCHEZ SANCHEZ

DATE/TIME OF TEST: 09/05/2025 23:30:50

Diagnostics Check		OK	23:21
Air Blank		0.000	23:22
Control Test		0.078	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.198	23:25
Breath Volume		1.195	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.193	23:29
Breath Volume		1.574	23:29
Air Blank		0.000	23:29
Control Test		0.076	23:30
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARNER,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MARTINEZ RAMIREZ

DATE/TIME OF TEST: 09/05/2025 23:58:32

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.077	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.180	23:53
Breath Volume		2.628	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.182	23:56
Breath Volume		2.652	23:56
Air Blank		0.000	23:57
Control Test		0.078	23:57
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRUEH,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 09/06/2025 00:34:12

Diagnostics Check		OK	00:22	
Air Blank		0.000	00:22	
Control Test		0.078	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1	0.113	00:26	Volume Not Met
Breath Volume		0.140	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:29	
Subject Sample	# 2	0.134	00:29	
Breath Volume		1.414	00:29	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 3	0.135	00:32	
Breath Volume		1.460	00:32	
Air Blank		0.000	00:33	
Control Test		0.078	00:33	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVERAS,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: RODRIGUEZ VILLA

DATE/TIME OF TEST: 09/06/2025 02:10:09

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.078	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.180	02:04
Breath Volume		1.457	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.182	02:08
Breath Volume		1.414	02:08
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELLES,LEXIE,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LANDRY,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: BREINER

DATE/TIME OF TEST: 09/06/2025 03:24:52

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.078	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.202	03:20
Breath Volume		2.074	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 2	0.199	03:23
Breath Volume		1.859	03:23
Air Blank		0.000	03:23
Control Test		0.077	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKINS,KEYANNA,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAKER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DIESEL

DATE/TIME OF TEST: 09/06/2025 04:29:45

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.078	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1	0.125	04:22	Volume Not Met
Breath Volume		0.054	04:22	
Air Blank		0.000	04:23	
Air Blank		0.000	04:25	
Subject Sample	# 2	0.110	04:28	Volume Not Met
Breath Volume		0.804	04:28	
Air Blank		0.000	04:28	
Control Test		0.076	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PETERS,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: ISOM

DATE/TIME OF TEST: 09/06/2025 22:23:11

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.078	22:17
Air Blank		0.000	22:18
Subject Sample	# 1	0.099	22:18
Breath Volume		1.265	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.095	22:21
Breath Volume		1.589	22:21
Air Blank		0.000	22:22
Control Test		0.078	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLEIN,ETHAN,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLINT,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 09/06/2025 23:44:01

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.079	23:33	
Air Blank		0.000	23:33	
Subject Sample	# 1	0.173	23:36	Volume Not Met
Breath Volume		0.062	23:36	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2		23:42	No Sample Provided
Breath Volume		0.000	23:42	
Air Blank		0.000	23:43	
Control Test		0.076	23:43	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:44	

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: STEPHENS,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SERN DIAZ

DATE/TIME OF TEST: 09/07/2025 02:45:20

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.129	02:40
Breath Volume		1.308	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.130	02:43
Breath Volume		1.367	02:43
Air Blank		0.000	02:44
Control Test		0.077	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/07/2025 04:16:33

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:10	
Control Test		0.079	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 1	0.131	04:11	
Breath Volume		2.058	04:11	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2	0.130	04:14	
Breath Volume		2.488	04:14	
Air Blank		0.000	04:15	
Control Test		0.078	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		Fail	04:16	Temperature Regulation Test
Air Blank		0.000	04:16	

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007483

INSTRUMENT REGISTERED TO: HILLSBOROUGH CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,MINDY,L
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 08/18/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A6
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/07/2025 04:18:30

Diagnostics Check	Fail	04:18	Temperature Regulation Test
Air Blank	0.000	04:18	

Data Download Date/Time: 09/18/2025 10:46

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWRENCE,JOSHUA,S
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: LAWRENCE,JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 09/15/2025 00:04:26

Diagnostics Check	Fail	00:03	Internal Printer Test
Air Blank	0.000	00:04	

Data Download Date/Time: 10/31/2025 11:10

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWRENCE,JOSHUA,S
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: LAWRENCE,JOSHUA
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 08/22/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 09/15/2025 00:14:06

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.082	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1	0.112	00:08	
Breath Volume		1.761	00:08	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 2	0.102	00:11	
Breath Volume		1.750	00:11	
Air Blank		0.000	00:12	
Control Test		0.079	00:12	
Air Blank		0.000	00:13	
Diagnostics Check		Fail	00:13	Internal Printer Test
Air Blank		0.000	00:14	

Data Download Date/Time: 10/31/2025 11:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,MATTHEW,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LEE,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 09/01/2025 15:19:33

Diagnostics Check		OK	15:10	
Air Blank		0.000	15:11	
Control Test		0.079	15:11	
Air Blank		0.000	15:12	
Subject Sample	# 1		15:15	No Sample Provided
Breath Volume		0.000	15:15	
Air Blank		0.000	15:15	
Air Blank		0.000	15:17	
Subject Sample	# 2		15:18	Subject Test Refused
Breath Volume		0.000	15:18	
Air Blank		0.000	15:18	
Control Test		0.080	15:18	
Air Blank		0.000	15:19	
Diagnostics Check		OK	15:19	

Data Download Date/Time: 09/27/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: TYCHEWICZ

DATE/TIME OF TEST: 09/05/2025 20:36:51

Diagnostics Check		OK	20:25	
Air Blank		0.000	20:25	
Control Test		0.080	20:25	
Air Blank		0.000	20:26	
Subject Sample	# 1	0.184	20:29	Volume Not Met
Breath Volume		0.222	20:29	
Air Blank		0.000	20:30	
Air Blank		0.000	20:31	
Subject Sample	# 2	0.195	20:35	Volume Not Met
Breath Volume		0.261	20:35	
Air Blank		0.000	20:35	
Control Test		0.078	20:36	
Air Blank		0.000	20:36	
Diagnostics Check		OK	20:36	

Data Download Date/Time: 09/27/2025 09:17

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: TYCHEICZ

DATE/TIME OF TEST: 09/05/2025 20:51:26

Diagnostics Check		OK	20:45	
Air Blank		0.000	20:46	
Control Test		0.080	20:46	
Air Blank		0.000	20:47	
Subject Sample	# 1	0.186	20:49	
Breath Volume		2.136	20:49	
Air Blank			20:50	Purge Fail
Air Blank			20:51	Purge Fail

Data Download Date/Time: 09/27/2025 09:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JANIAK,MALIKHI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 09/06/2025 00:40:40

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:39	
Control Test	0.051	00:40	Control Outside Tolerance
Air Blank	0.000	00:40	

Data Download Date/Time: 09/27/2025 09:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANIAK,MALIKHI,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 08/26/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A2
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SALAZAR CARRASCO

DATE/TIME OF TEST: 09/06/2025 21:43:14

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:42	
Control Test	0.068	21:42	Control Outside Tolerance
Air Blank		21:43	RFI Detect

Data Download Date/Time: 09/27/2025 09:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/02/2025 12:05:49

Diagnostics Check		OK	11:59
Air Blank		0.000	12:00
Control Test		0.081	12:00
Air Blank		0.000	12:01
Subject Sample	# 1	0.000	12:01
Breath Volume		2.710	12:01
Air Blank		0.000	12:01
Air Blank		0.000	12:03
Subject Sample	# 2	0.000	12:04
Breath Volume		1.867	12:04
Air Blank		0.000	12:04
Control Test		0.080	12:05
Air Blank		0.000	12:05
Diagnostics Check		OK	12:05

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 08:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:31

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 09/03/2025 09:02:05

Diagnostics Check		OK	08:56
Air Blank		0.000	08:56
Control Test		0.080	08:56
Air Blank		0.000	08:57
Subject Sample	# 1	0.000	08:57
Breath Volume		1.195	08:57
Air Blank		0.000	08:58
Air Blank		0.000	09:00
Subject Sample	# 2	0.000	09:00
Breath Volume		1.679	09:00
Air Blank		0.000	09:01
Control Test		0.081	09:01
Air Blank		0.000	09:02
Diagnostics Check		OK	09:02

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 14:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:16

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 09/04/2025 14:47:10

Diagnostics Check		OK	14:41
Air Blank		0.000	14:41
Control Test		0.081	14:41
Air Blank		0.000	14:42
Subject Sample	# 1	0.000	14:42
Breath Volume		2.324	14:42
Air Blank		0.000	14:43
Air Blank		0.000	14:45
Subject Sample	# 2	0.000	14:45
Breath Volume		2.144	14:45
Air Blank		0.000	14:46
Control Test		0.081	14:46
Air Blank		0.000	14:47
Diagnostics Check		OK	14:47

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 14:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:28

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 09/04/2025 14:58:58

Diagnostics Check		OK	14:53
Air Blank		0.000	14:53
Control Test		0.080	14:54
Air Blank		0.000	14:54
Subject Sample	# 1	0.000	14:54
Breath Volume		1.234	14:54
Air Blank		0.000	14:55
Air Blank		0.000	14:57
Subject Sample	# 2	0.000	14:57
Breath Volume		1.476	14:57
Air Blank		0.000	14:58
Control Test		0.080	14:58
Air Blank		0.000	14:58
Diagnostics Check		OK	14:58

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: STOCKS

DATE/TIME OF TEST: 09/04/2025 17:11:14

Diagnostics Check		OK	17:05
Air Blank		0.000	17:05
Control Test		0.080	17:05
Air Blank		0.000	17:06
Subject Sample	# 1	0.000	17:06
Breath Volume		1.199	17:06
Air Blank		0.000	17:07
Air Blank		0.000	17:09
Subject Sample	# 2	0.000	17:09
Breath Volume		1.394	17:09
Air Blank		0.000	17:10
Control Test		0.081	17:10
Air Blank		0.000	17:11
Diagnostics Check		OK	17:11

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MILLER,PRESNTON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 09/04/2025 21:20:44

Diagnostics Check		OK	21:14
Air Blank		0.000	21:15
Control Test		0.080	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.097	21:16
Breath Volume		1.992	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:18
Subject Sample	# 2	0.094	21:19
Breath Volume		1.949	21:19
Air Blank		0.000	21:19
Control Test		0.080	21:20
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GLASCO,GREGORY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: RECTOR-REDEAGLE

DATE/TIME OF TEST: 09/04/2025 22:20:22

Diagnostics Check		OK	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.000	22:16
Breath Volume		1.425	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.000	22:18
Breath Volume		1.554	22:18
Air Blank		0.000	22:19
Control Test		0.081	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/05/2025 10:06:38

Diagnostics Check		OK	10:00
Air Blank		0.000	10:00
Control Test		0.080	10:00
Air Blank		0.000	10:01
Subject Sample	# 1	0.000	10:02
Breath Volume		2.277	10:02
Air Blank		0.000	10:02
Air Blank		0.000	10:04
Subject Sample	# 2	0.000	10:05
Breath Volume		1.980	10:05
Air Blank		0.000	10:05
Control Test		0.080	10:06
Air Blank		0.000	10:06
Diagnostics Check		OK	10:06

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOEB,BRANDON,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOEB,BRANDON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 13:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: PIERRE-PAUL

DATE/TIME OF TEST: 09/06/2025 14:09:09

Diagnostics Check		OK	14:03
Air Blank		0.000	14:03
Control Test		0.081	14:04
Air Blank		0.000	14:04
Subject Sample	# 1	0.000	14:04
Breath Volume		1.527	14:04
Air Blank		0.000	14:05
Air Blank		0.000	14:07
Subject Sample	# 2	0.000	14:07
Breath Volume		1.226	14:07
Air Blank		0.000	14:08
Control Test		0.081	14:08
Air Blank		0.000	14:09
Diagnostics Check		OK	14:09

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,NEAL,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GILLARD,TYLER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: KORO

DATE/TIME OF TEST: 09/07/2025 01:49:03

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:47	Subject Test Refused
Breath Volume		0.000	01:47	
Air Blank		0.000	01:48	
Control Test		0.080	01:48	
Air Blank		0.000	01:48	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,NEAL,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALMERTON,KENNETH
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: CEDENO

DATE/TIME OF TEST: 09/07/2025 04:02:33

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:49	
Control Test		0.080	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1	0.000	03:53	Volume Not Met
Breath Volume		0.652	03:53	
Air Blank		0.000	03:53	
Air Blank		0.000	03:55	
Subject Sample	# 2	0.000	03:58	
Breath Volume		1.714	03:58	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 3	0.000	04:01	
Breath Volume		1.730	04:01	
Air Blank		0.000	04:01	
Control Test		0.079	04:01	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,NEAL,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: KNAPP,ZACHARY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: MCABEE

DATE/TIME OF TEST: 09/07/2025 20:40:17

Diagnostics Check		OK	20:36	
Air Blank		0.000	20:36	
Control Test		0.080	20:37	
Air Blank		0.000	20:37	
Subject Sample	# 1		20:38	Subject Test Refused
Breath Volume		0.000	20:38	
Air Blank		0.000	20:39	
Control Test		0.080	20:39	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,SUSAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:08

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/08/2025 09:40:13

Diagnostics Check		OK	09:34
Air Blank		0.000	09:34
Control Test		0.081	09:35
Air Blank		0.000	09:35
Subject Sample	# 1	0.000	09:36
Breath Volume		1.781	09:36
Air Blank		0.000	09:36
Air Blank		0.000	09:38
Subject Sample	# 2	0.000	09:38
Breath Volume		1.359	09:38
Air Blank		0.000	09:39
Control Test		0.081	09:39
Air Blank		0.000	09:40
Diagnostics Check		OK	09:40

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 09/09/2025 11:09:57

Diagnostics Check		OK	11:03
Air Blank		0.000	11:04
Control Test		0.081	11:04
Air Blank		0.000	11:04
Subject Sample	# 1	0.000	11:05
Breath Volume		1.792	11:05
Air Blank		0.000	11:06
Air Blank		0.000	11:08
Subject Sample	# 2	0.000	11:08
Breath Volume		1.707	11:08
Air Blank		0.000	11:09
Control Test		0.081	11:09
Air Blank		0.000	11:09
Diagnostics Check		OK	11:09

Data Download Date/Time: 09/13/2025 21:56

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,SUSAN,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,SUSAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 15:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: STOCKS GOODRICH

DATE/TIME OF TEST: 09/09/2025 15:56:12

Diagnostics Check		OK	15:50
Air Blank		0.000	15:50
Control Test		0.081	15:50
Air Blank		0.000	15:51
Subject Sample	# 1	0.000	15:51
Breath Volume		1.187	15:51
Air Blank		0.000	15:52
Air Blank		0.000	15:54
Subject Sample	# 2	0.000	15:54
Breath Volume		1.144	15:54
Air Blank		0.000	15:55
Control Test		0.081	15:55
Air Blank		0.000	15:56
Diagnostics Check		OK	15:56

Data Download Date/Time: 09/13/2025 21:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYALA,JOSE,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LISTER,DALTON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/10/2025 04:26:12

Diagnostics Check		OK	04:23	
Air Blank		0.000	04:23	
Control Test		0.081	04:24	
Air Blank		0.000	04:24	
Subject Sample	# 1		04:24	Subject Test Refused
Breath Volume		0.000	04:24	
Air Blank		0.000	04:25	
Control Test		0.080	04:25	
Air Blank		0.000	04:26	
Diagnostics Check		OK	04:26	

Data Download Date/Time: 09/13/2025 21:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,DALE,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: VAUGHN,DALE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: NEIDERMAN

DATE/TIME OF TEST: 09/10/2025 11:11:58

Diagnostics Check		OK	11:05
Air Blank		0.000	11:06
Control Test		0.081	11:06
Air Blank		0.000	11:07
Subject Sample	# 1	0.000	11:07
Breath Volume		1.597	11:07
Air Blank		0.000	11:08
Air Blank		0.000	11:10
Subject Sample	# 2	0.000	11:10
Breath Volume		1.546	11:10
Air Blank		0.000	11:11
Control Test		0.080	11:11
Air Blank		0.000	11:11
Diagnostics Check		OK	11:11

Data Download Date/Time: 09/13/2025 21:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/11/2025 10:22:50

Diagnostics Check		OK	10:16
Air Blank		0.000	10:17
Control Test		0.080	10:17
Air Blank		0.000	10:18
Subject Sample	# 1	0.000	10:18
Breath Volume		1.769	10:18
Air Blank		0.000	10:19
Air Blank		0.000	10:21
Subject Sample	# 2	0.000	10:21
Breath Volume		2.089	10:21
Air Blank		0.000	10:21
Control Test		0.081	10:22
Air Blank		0.000	10:22
Diagnostics Check		OK	10:22

Data Download Date/Time: 09/13/2025 21:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALACIOS-FLORES,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PALACIOS-FLORES,CRISTIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: TREYTON XANDER

DATE/TIME OF TEST: 09/11/2025 18:47:37

Diagnostics Check		OK	18:41
Air Blank		0.000	18:42
Control Test		0.080	18:42
Air Blank		0.000	18:43
Subject Sample	# 1	0.000	18:43
Breath Volume		2.128	18:43
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.000	18:46
Breath Volume		1.625	18:46
Air Blank		0.000	18:46
Control Test		0.080	18:47
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 09/13/2025 21:57

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOEB,BRANDON,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOEB,BRANDON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 08:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:57

SUBJECT'S LAST NAME: HELMS

DATE/TIME OF TEST: 09/14/2025 09:26:35

Diagnostics Check		OK	09:20
Air Blank		0.000	09:21
Control Test		0.079	09:21
Air Blank		0.000	09:22
Subject Sample	# 1	0.000	09:22
Breath Volume		2.343	09:22
Air Blank		0.000	09:22
Air Blank		0.000	09:24
Subject Sample	# 2	0.000	09:25
Breath Volume		2.222	09:25
Air Blank		0.000	09:25
Control Test		0.079	09:25
Air Blank		0.000	09:26
Diagnostics Check		OK	09:26

Data Download Date/Time: 10/26/2025 21:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 11:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: STOCKS-GOODRICH

DATE/TIME OF TEST: 09/14/2025 12:21:36

Diagnostics Check		OK	12:15
Air Blank		0.000	12:16
Control Test		0.081	12:16
Air Blank		0.000	12:16
Subject Sample	# 1	0.000	12:17
Breath Volume		1.210	12:17
Air Blank		0.000	12:17
Air Blank		0.000	12:19
Subject Sample	# 2	0.000	12:20
Breath Volume		1.308	12:20
Air Blank		0.000	12:20
Control Test		0.081	12:20
Air Blank		0.000	12:21
Diagnostics Check		OK	12:21

Data Download Date/Time: 10/26/2025 21:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/14/2025 12:50:27

Diagnostics Check		OK	12:44
Air Blank		0.000	12:45
Control Test		0.081	12:45
Air Blank		0.000	12:45
Subject Sample	# 1	0.000	12:46
Breath Volume		1.757	12:46
Air Blank		0.000	12:46
Air Blank		0.000	12:48
Subject Sample	# 2	0.000	12:49
Breath Volume		1.722	12:49
Air Blank		0.000	12:49
Control Test		0.080	12:49
Air Blank		0.000	12:50
Diagnostics Check		OK	12:50

Data Download Date/Time: 10/26/2025 21:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOEB,BRANDON,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOEB,BRANDON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 14:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:13

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 09/14/2025 14:41:25

Diagnostics Check		OK	14:35
Air Blank		0.000	14:36
Control Test		0.081	14:36
Air Blank		0.000	14:36
Subject Sample	# 1	0.000	14:37
Breath Volume		1.960	14:37
Air Blank		0.000	14:37
Air Blank		0.000	14:39
Subject Sample	# 2	0.000	14:40
Breath Volume		2.125	14:40
Air Blank		0.000	14:40
Control Test		0.081	14:40
Air Blank		0.000	14:41
Diagnostics Check		OK	14:41

Data Download Date/Time: 10/26/2025 21:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOEB,BRANDON,C
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOEB,BRANDON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 09/15/2025 12:33:49

Diagnostics Check		OK	12:28
Air Blank		0.000	12:28
Control Test		0.082	12:28
Air Blank		0.000	12:29
Subject Sample	# 1	0.000	12:29
Breath Volume		1.136	12:29
Air Blank		0.000	12:30
Air Blank		0.000	12:32
Subject Sample	# 2	0.000	12:32
Breath Volume		1.191	12:32
Air Blank		0.000	12:32
Control Test		0.083	12:33
Air Blank		0.000	12:33
Diagnostics Check		OK	12:33

Data Download Date/Time: 10/26/2025 21:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERT
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/17/2025 16:57:15

Diagnostics Check		OK	16:51
Air Blank		0.000	16:51
Control Test		0.080	16:52
Air Blank		0.000	16:52
Subject Sample	# 1	0.000	16:52
Breath Volume		2.082	16:52
Air Blank		0.000	16:53
Air Blank		0.000	16:55
Subject Sample	# 2	0.000	16:55
Breath Volume		1.617	16:55
Air Blank		0.000	16:56
Control Test		0.083	16:56
Air Blank		0.000	16:57
Diagnostics Check		OK	16:57

Data Download Date/Time: 10/26/2025 21:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLASCO,GREGORY,S
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WALKER,DUSTIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WYENANDT

DATE/TIME OF TEST: 09/18/2025 01:16:14

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.081	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.171	01:11
Breath Volume		1.226	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.160	01:14
Breath Volume		1.253	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 10/26/2025 21:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AYALA,JOSE,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROBINSON,REESE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: PEREZ DIAZ

DATE/TIME OF TEST: 09/18/2025 01:55:19

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.080	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:53	Subject Test Refused
Breath Volume		0.000	01:53	
Air Blank		0.000	01:54	
Control Test		0.080	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 10/26/2025 21:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: STOCKS-GOODRICH

DATE/TIME OF TEST: 09/19/2025 17:29:36

Diagnostics Check		OK	17:22	
Air Blank		0.000	17:23	
Control Test		0.080	17:23	
Air Blank		0.000	17:24	
Subject Sample	# 1	0.000	17:24	
Breath Volume		1.820	17:24	
Air Blank		0.000	17:25	
Air Blank		0.000	17:27	
Subject Sample	# 2	0.000	17:28	
Breath Volume		1.867	17:28	
Air Blank			17:28	Purge Fail
Air Blank			17:29	Purge Fail

Data Download Date/Time: 10/26/2025 21:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,NEAL,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RODRIGUEZ,MANUEL
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: SERRECCHIA

DATE/TIME OF TEST: 09/21/2025 20:23:41

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.081	20:18
Air Blank		0.000	20:18
Subject Sample	# 1	0.000	20:19
Breath Volume		1.316	20:19
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.000	20:22
Breath Volume		1.402	20:22
Air Blank		0.000	20:22
Control Test		0.080	20:23
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 10/26/2025 21:47

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,NEAL,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANPOJA,Y
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/21/2025 22:41:43

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.080	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.000	22:37
Breath Volume		1.644	22:37
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.000	22:40
Breath Volume		1.882	22:40
Air Blank		0.000	22:40
Control Test		0.081	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: WOLVIN

DATE/TIME OF TEST: 09/24/2025 17:14:41

Diagnostics Check		OK	17:08
Air Blank		0.000	17:08
Control Test		0.081	17:09
Air Blank		0.000	17:09
Subject Sample	# 1	0.000	17:10
Breath Volume		1.671	17:10
Air Blank		0.000	17:10
Air Blank		0.000	17:12
Subject Sample	# 2	0.000	17:13
Breath Volume		1.757	17:13
Air Blank		0.000	17:13
Control Test		0.081	17:14
Air Blank		0.000	17:14
Diagnostics Check		OK	17:14

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: HELMS

DATE/TIME OF TEST: 09/29/2025 12:42:21

Diagnostics Check		OK	12:35
Air Blank		0.000	12:36
Control Test		0.081	12:36
Air Blank		0.000	12:37
Subject Sample	# 1	0.000	12:37
Breath Volume		2.136	12:37
Air Blank		0.000	12:38
Air Blank		0.000	12:40
Subject Sample	# 2	0.000	12:40
Breath Volume		2.207	12:40
Air Blank		0.000	12:41
Control Test		0.080	12:41
Air Blank		0.000	12:42
Diagnostics Check		OK	12:42

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLVIN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WOLVIN,WILLIAM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 09/13/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A4
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: STOCKS-GOODRICH

DATE/TIME OF TEST: 09/29/2025 17:05:03

Diagnostics Check		OK	16:58
Air Blank		0.000	16:59
Control Test		0.081	16:59
Air Blank		0.000	17:00
Subject Sample	# 1	0.000	17:00
Breath Volume		1.808	17:00
Air Blank		0.000	17:01
Air Blank		0.000	17:03
Subject Sample	# 2	0.000	17:03
Breath Volume		1.425	17:03
Air Blank		0.000	17:04
Control Test		0.081	17:04
Air Blank		0.000	17:04
Diagnostics Check		OK	17:05

Data Download Date/Time: 10/26/2025 21:48

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOVAR,LUIS,D
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: SANTANA,ELIO
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 08/02/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: WISEMAN

DATE/TIME OF TEST: 09/06/2025 04:22:05

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:08	
Control Test		0.080	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1	0.065	04:12	Volume Not Met
Breath Volume		0.312	04:12	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2	0.066	04:16	
Breath Volume		1.621	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:18	
Subject Sample	# 3	0.065	04:20	
Breath Volume		1.617	04:20	
Air Blank		0.000	04:21	
Control Test		0.078	04:21	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 09/07/2025 13:00

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOVAR,LUIS,D
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PANDOLFI,CLEBER
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/12/2025 01:34:52

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.079	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.193	01:30
Breath Volume		3.226	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:32
Subject Sample	# 2	0.193	01:33
Breath Volume		3.820	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 10/27/2025 18:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFRESNE,BRANDY,M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: FORNAL,D
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BLANDIN

DATE/TIME OF TEST: 09/13/2025 21:48:46

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.079	21:42
Air Blank		0.000	21:43
Subject Sample	# 1	0.077	21:43
Breath Volume		2.171	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.075	21:47
Breath Volume		2.066	21:47
Air Blank		0.000	21:47
Control Test		0.079	21:48
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 10/27/2025 18:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOVAR,LUIS,D
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PANDOLFI,CLEBER
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: HERNANDEZ-GONZALEZ

DATE/TIME OF TEST: 09/21/2025 22:28:49

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.080	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.147	22:24
Breath Volume		2.671	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:26
Subject Sample	# 2	0.154	22:27
Breath Volume		3.242	22:27
Air Blank		0.000	22:27
Control Test		0.078	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 10/27/2025 18:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007508

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILIPOWICZ,KEVIN,M
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: POLK,ADRIAN
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 09/07/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: CARRILO-JUAN

DATE/TIME OF TEST: 09/26/2025 23:19:19

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.078	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.228	23:14
Breath Volume		1.105	23:14
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.222	23:17
Breath Volume		1.195	23:17
Air Blank		0.000	23:18
Control Test		0.079	23:18
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 10/27/2025 18:11

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMAZANOVA,SHAFIGA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: NUNES,MICHAEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: RAMAZANOVA

DATE/TIME OF TEST: 09/23/2025 02:32:02

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:31	
Control Test		0.080	02:31	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 10/30/2025 16:25

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAPSHUCK,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: NUNES,MICHAEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RAMAZANOVA

DATE/TIME OF TEST: 09/23/2025 02:55:45

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:53	
Control Test		0.081	02:53	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:54	Subject Test Refused
Breath Volume		0.000	02:54	
Air Blank		0.000	02:54	
Control Test		0.081	02:55	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 10/30/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: JARQUIN JARQUIN

DATE/TIME OF TEST: 09/26/2025 23:12:43

Diagnostics Check		OK	22:57	
Air Blank		0.000	22:57	
Control Test		0.081	22:58	
Air Blank		0.000	22:58	
Subject Sample	# 1	0.173	23:01	Volume Not Met
Breath Volume		0.164	23:01	
Air Blank		0.000	23:02	
Air Blank		0.000	23:04	
Subject Sample	# 2	0.184	23:05	
Breath Volume		1.390	23:05	
Air Blank		0.000	23:06	
Air Blank		0.000	23:07	
Subject Sample	# 3	0.164	23:10	Volume Not Met
Breath Volume		0.074	23:10	
Air Blank		0.000	23:11	
Control Test		0.079	23:12	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 10/30/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: JARQUIN JARQUIN

DATE/TIME OF TEST: 09/26/2025 23:21:51

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.078	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.187	23:16
Breath Volume		1.652	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.187	23:20
Breath Volume		2.093	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 10/30/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDRE,BOBBY,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CASTANEDA,EDUARDO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402732434
DRY GAS EXPIRY DATE: 04/28/2026
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SCHARFENBERGER

DATE/TIME OF TEST: 09/29/2025 04:56:04

Diagnostics Check		OK	04:53	
Air Blank		0.000	04:53	
Control Test		0.081	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 1		04:54	Subject Test Refused
Breath Volume		0.000	04:54	
Air Blank		0.000	04:55	
Control Test		0.080	04:55	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:56	

Data Download Date/Time: 10/30/2025 16:26

Printed: 08 December 2025

COBRA ©1997-2014 CMI, Inc.