Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUPA,KYLE,T
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: KRUPA,KYLE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: RANGEL

DATE/TIME OF TEST: 06/01/2026 22:20:58

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:08	
Control Test		0.080	22:08	
Air Blank		0.000	22:09	
Subject Sample	# 1		22:12	No Sample Provided
Breath Volume		0.000	22:12	
Air Blank		0.000	22:13	
Air Blank		0.000	22:15	
Subject Sample	# 2	0.177	22:16	
Breath Volume		2.402	22:16	
Air Blank		0.000	22:17	
Air Blank		0.000	22:18	
Subject Sample	# 3	0.168	22:19	
Breath Volume		1.757	22:19	
Air Blank		0.000	22:20	
Control Test		0.079	22:20	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 07/24/2026 02:34

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,ANTHONY,L
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: LAYNE,NICK
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BOLANOS

DATE/TIME OF TEST: 06/03/2026 02:04:19

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.081	01:57
Air Blank		0.000	01:58
Subject Sample	# 1	0.126	01:59
Breath Volume		2.832	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.127	02:02
Breath Volume		1.609	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 07/24/2026 02:34

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIDEROA,KEVIN,
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: FAULKENBURG,JASON
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 06/03/2026 08:47:20

Diagnostics Check	OK	08:46	
Air Blank		08:46	RFI Detect
Air Blank	0.000	08:47	

Data Download Date/Time: 07/24/2026 02:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIDEROA,KEVIN',
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: FAULKENBURG,JASON
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 06/03/2026 08:49:17

Diagnostics Check	Fail	08:48	Temperature Regulation Test
Air Blank	0.000	08:49	

Data Download Date/Time: 07/24/2026 02:34

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAS,CARLY,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: GAULDING,ELLISON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ADLER

DATE/TIME OF TEST: 06/03/2026 04:18:04

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.080	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.177	04:12
Breath Volume		1.582	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.172	04:16
Breath Volume		1.265	04:16
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:18

Data Download Date/Time: 06/26/2026 20:05

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SELWACH

DATE/TIME OF TEST: 06/12/2026 03:13:51

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.093	03:09
Breath Volume		1.750	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:11
Subject Sample	# 2	0.091	03:12
Breath Volume		1.597	03:12
Air Blank		0.000	03:12
Control Test		0.081	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 06/26/2026 20:05

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNAMARA,KYLE,J
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MCNAMARA,KYLE
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: BIERMAN

DATE/TIME OF TEST: 06/12/2026 16:49:29

Diagnostics Check		OK	16:44	
Air Blank		0.000	16:45	
Control Test		0.078	16:45	
Air Blank		0.000	16:45	
Subject Sample	# 1		16:48	Subject Test Refused
Breath Volume		0.000	16:48	
Air Blank		0.000	16:48	
Control Test		0.078	16:48	
Air Blank		0.000	16:49	
Diagnostics Check		OK	16:49	

Data Download Date/Time: 06/26/2026 20:05

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: LIESKE,H
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: ROSIER

DATE/TIME OF TEST: 06/13/2026 22:45:20

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.080	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.080	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 06/26/2026 20:05

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAS,CARLY,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: LIESKE,HOLDEN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: WILLIAM

DATE/TIME OF TEST: 06/14/2026 21:43:20

Diagnostics Check		OK	21:34
Air Blank		0.000	21:35
Control Test		0.079	21:35
Air Blank		0.000	21:36
Subject Sample	# 1	0.198	21:38
Breath Volume		1.675	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.184	21:41
Breath Volume		1.269	21:41
Air Blank		0.000	21:42
Control Test		0.079	21:42
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 06/26/2026 20:05

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: DELEON BARRIOS

DATE/TIME OF TEST: 06/16/2026 20:40:40

Diagnostics Check		OK	20:37	
Air Blank		0.000	20:38	
Control Test		0.080	20:38	
Air Blank		0.000	20:39	
Subject Sample	# 1		20:39	Subject Test Refused
Breath Volume		0.000	20:39	
Air Blank		0.000	20:39	
Control Test		0.080	20:40	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 06/26/2026 20:05

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 06/20/2026 00:38:25

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.143	00:34
Breath Volume		1.562	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.144	00:36
Breath Volume		1.679	00:36
Air Blank		0.000	00:37
Control Test		0.079	00:37
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 06/26/2026 20:05

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSAS,CARLY,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: LEDOUX,JACOB
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: SMARTT

DATE/TIME OF TEST: 06/20/2026 03:23:09

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.139	03:17
Breath Volume		1.949	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.140	03:21
Breath Volume		1.625	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 06/26/2026 20:05

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAQUAVIOUS,D
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: SMITH,JAQUAVIOUS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MINGO

DATE/TIME OF TEST: 06/21/2026 00:54:05

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.083	00:49
Breath Volume		2.800	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.081	00:52
Breath Volume		2.460	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 06/26/2026 20:05

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: GRAHAM,EMMA
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 05/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: KODLUK

DATE/TIME OF TEST: 06/13/2026 03:36:22

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.078	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.122	03:31
Breath Volume		1.671	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.129	03:34
Breath Volume		1.812	03:34
Air Blank		0.000	03:35
Control Test		0.078	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 06/30/2026 16:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,EMMA,
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: GRAHAM,EMMA
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 05/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GONZALEZ ALVARADO

DATE/TIME OF TEST: 06/13/2026 03:46:00

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.091	03:41
Breath Volume		1.578	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.087	03:44
Breath Volume		1.449	03:44
Air Blank		0.000	03:45
Control Test		0.078	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:46

Data Download Date/Time: 06/30/2026 16:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: MAHLER,CARLITA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: OURADA

DATE/TIME OF TEST: 06/21/2026 03:07:00

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.173	03:02
Breath Volume		1.714	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.157	03:05
Breath Volume		1.382	03:05
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:07

Data Download Date/Time: 06/30/2026 16:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: KEMMERLING,CONNOR
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 05/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 06/30/2026 03:10:18

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.078	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.126	03:01	Volume Not Met
Breath Volume		0.300	03:01	
Air Blank		0.000	03:02	
Air Blank		0.000	03:04	
Subject Sample	# 2	0.153	03:05	
Breath Volume		1.417	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 3	0.146	03:08	
Breath Volume		1.675	03:08	
Air Blank		0.000	03:09	
Control Test		0.075	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 06/30/2026 16:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECORA,DEPUTY,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GROSS,DEPUTY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 06/24/2026 17:39:54

Diagnostics Check		OK	17:31	
Air Blank		0.000	17:31	
Control Test		0.081	17:32	
Air Blank		0.000	17:32	
Subject Sample	# 1		17:35	No Sample Provided
Breath Volume		0.000	17:35	
Air Blank		0.000	17:36	
Air Blank		0.000	17:38	
Subject Sample	# 2		17:38	Subject Test Refused
Breath Volume		0.000	17:38	
Air Blank		0.000	17:38	
Control Test		0.079	17:39	
Air Blank		0.000	17:39	
Diagnostics Check		OK	17:39	

Data Download Date/Time: 07/31/2026 20:14

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: DE FILIPPO

DATE/TIME OF TEST: 06/15/2026 20:00:00

Diagnostics Check		OK	19:52	
Air Blank		0.000	19:52	
Control Test		0.081	19:52	
Air Blank		0.000	19:53	
Subject Sample	# 1	0.077	19:56	
Breath Volume		1.976	19:56	
Air Blank		0.000	19:56	
Air Blank		0.000	19:58	
Subject Sample	# 2	0.076	19:59	
Breath Volume		2.421	19:59	
Air Blank			19:59	RFI Detect
Air Blank		0.000	20:00	

Data Download Date/Time: 07/18/2026 02:38

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BERRIEN

DATE/TIME OF TEST: 06/16/2026 04:14:48

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.176	04:10
Breath Volume		1.273	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:12
Subject Sample	# 2	0.181	04:13
Breath Volume		2.000	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 07/18/2026 02:38

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,STEVEN,J
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: EVANS,STEVEN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: PROB/PAROL
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 06/16/2026 10:18:18

Diagnostics Check		OK	10:12
Air Blank		0.000	10:12
Control Test		0.081	10:13
Air Blank		0.000	10:13
Subject Sample	# 1	0.000	10:14
Breath Volume		1.589	10:14
Air Blank		0.000	10:14
Air Blank		0.000	10:16
Subject Sample	# 2	0.000	10:16
Breath Volume		1.585	10:16
Air Blank		0.000	10:17
Control Test		0.080	10:17
Air Blank		0.000	10:18
Diagnostics Check		OK	10:18

Data Download Date/Time: 07/18/2026 02:38

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 06/17/2026 13:05:43

Diagnostics Check		OK	12:59
Air Blank		0.000	13:00
Control Test		0.080	13:00
Air Blank		0.000	13:01
Subject Sample	# 1	0.000	13:01
Breath Volume		1.558	13:01
Air Blank		0.000	13:01
Air Blank		0.000	13:03
Subject Sample	# 2	0.000	13:04
Breath Volume		1.414	13:04
Air Blank		0.000	13:04
Control Test		0.081	13:05
Air Blank		0.000	13:05
Diagnostics Check		OK	13:05

Data Download Date/Time: 07/18/2026 02:38

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: COURT ORD
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: STEELMAN

DATE/TIME OF TEST: 06/18/2026 11:11:55

Diagnostics Check		OK	11:06
Air Blank		0.000	11:06
Control Test		0.080	11:06
Air Blank		0.000	11:07
Subject Sample	# 1	0.000	11:07
Breath Volume		1.593	11:07
Air Blank		0.000	11:08
Air Blank		0.000	11:10
Subject Sample	# 2	0.000	11:10
Breath Volume		1.652	11:10
Air Blank		0.000	11:11
Control Test		0.079	11:11
Air Blank		0.000	11:11
Diagnostics Check		OK	11:11

Data Download Date/Time: 07/18/2026 02:38

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: COURT ORD
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 06/18/2026 11:29:06

Diagnostics Check	OK	11:27	
Air Blank	0.000	11:28	
Control Test	0.080	11:28	
Air Blank		11:28	RFI Detect
Air Blank		11:29	RFI Detect

Data Download Date/Time: 07/18/2026 02:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: COURT ORD
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 06/18/2026 11:37:20

Diagnostics Check		OK	11:31
Air Blank		0.000	11:31
Control Test		0.080	11:32
Air Blank		0.000	11:32
Subject Sample	# 1	0.000	11:33
Breath Volume		1.398	11:33
Air Blank		0.000	11:33
Air Blank		0.000	11:35
Subject Sample	# 2	0.000	11:35
Breath Volume		1.714	11:35
Air Blank		0.000	11:36
Control Test		0.080	11:36
Air Blank		0.000	11:37
Diagnostics Check		OK	11:37

Data Download Date/Time: 07/18/2026 02:39

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: COURT ORD
ARREST TIME: 13:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: AYEUL

DATE/TIME OF TEST: 06/18/2026 14:06:22

Diagnostics Check		OK	14:00
Air Blank		0.000	14:01
Control Test		0.080	14:01
Air Blank		0.000	14:01
Subject Sample	# 1	0.000	14:02
Breath Volume		1.488	14:02
Air Blank		0.000	14:02
Air Blank		0.000	14:04
Subject Sample	# 2	0.000	14:05
Breath Volume		1.746	14:05
Air Blank		0.000	14:05
Control Test		0.079	14:05
Air Blank		0.000	14:06
Diagnostics Check		OK	14:06

Data Download Date/Time: 07/18/2026 02:40

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: COURT ORD
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 06/18/2026 14:17:41

Diagnostics Check		OK	14:11
Air Blank		0.000	14:12
Control Test		0.080	14:12
Air Blank		0.000	14:13
Subject Sample	# 1	0.000	14:13
Breath Volume		1.261	14:13
Air Blank		0.000	14:13
Air Blank		0.000	14:15
Subject Sample	# 2	0.000	14:16
Breath Volume		1.203	14:16
Air Blank		0.000	14:16
Control Test		0.079	14:17
Air Blank		0.000	14:17
Diagnostics Check		OK	14:17

Data Download Date/Time: 07/18/2026 02:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: COURT ORD
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: RIVERA PENA

DATE/TIME OF TEST: 06/18/2026 14:27:13

Diagnostics Check		OK	14:21
Air Blank		0.000	14:21
Control Test		0.080	14:22
Air Blank		0.000	14:22
Subject Sample	# 1	0.000	14:23
Breath Volume		1.636	14:23
Air Blank		0.000	14:23
Air Blank		0.000	14:25
Subject Sample	# 2	0.000	14:25
Breath Volume		1.851	14:25
Air Blank		0.000	14:26
Control Test		0.080	14:26
Air Blank		0.000	14:27
Diagnostics Check		OK	14:27

Data Download Date/Time: 07/18/2026 02:40

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:43

SUBJECT'S LAST NAME: DEPAZ

DATE/TIME OF TEST: 06/18/2026 16:20:48

Diagnostics Check		OK	16:14
Air Blank		0.000	16:15
Control Test		0.081	16:15
Air Blank		0.000	16:15
Subject Sample	# 1	0.000	16:16
Breath Volume		2.949	16:16
Air Blank		0.000	16:17
Air Blank		0.000	16:19
Subject Sample	# 2	0.000	16:19
Breath Volume		1.800	16:19
Air Blank		0.000	16:19
Control Test		0.081	16:20
Air Blank		0.000	16:20
Diagnostics Check		OK	16:20

Data Download Date/Time: 07/18/2026 02:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: PROB/PAROL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: IGNACIO MATA

DATE/TIME OF TEST: 06/18/2026 16:33:30

Diagnostics Check		OK	16:27
Air Blank		0.000	16:28
Control Test		0.081	16:28
Air Blank		0.000	16:28
Subject Sample	# 1	0.000	16:29
Breath Volume		1.898	16:29
Air Blank		0.000	16:29
Air Blank		0.000	16:31
Subject Sample	# 2	0.000	16:32
Breath Volume		2.046	16:32
Air Blank		0.000	16:32
Control Test		0.080	16:32
Air Blank		0.000	16:33
Diagnostics Check		OK	16:33

Data Download Date/Time: 07/18/2026 02:41

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILMAN,DUSTIN,D
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: THEODORE

DATE/TIME OF TEST: 06/19/2026 20:27:19

Diagnostics Check		OK	20:24	
Air Blank		0.000	20:24	
Control Test		0.080	20:25	
Air Blank		0.000	20:25	
Subject Sample	# 1		20:25	Subject Test Refused
Breath Volume		0.000	20:25	
Air Blank		0.000	20:26	
Control Test		0.080	20:26	
Air Blank		0.000	20:27	
Diagnostics Check		OK	20:27	

Data Download Date/Time: 07/18/2026 02:41

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: MIRO GARCIA

DATE/TIME OF TEST: 06/19/2026 20:58:05

Diagnostics Check		OK	20:50	
Air Blank		0.000	20:50	
Control Test		0.079	20:50	
Air Blank		0.000	20:51	
Subject Sample	# 1	0.039	20:53	
Breath Volume		1.289	20:53	
Air Blank		0.000	20:54	
Air Blank		0.000	20:56	
Subject Sample	# 2		20:56	Subject Test Refused
Breath Volume		0.000	20:56	
Air Blank		0.000	20:57	
Control Test		0.080	20:57	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 07/18/2026 02:41

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: VOLTAIRE

DATE/TIME OF TEST: 06/20/2026 21:47:54

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.080	21:41
Air Blank		0.000	21:41
Subject Sample	# 1	0.188	21:43
Breath Volume		1.492	21:43
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.190	21:46
Breath Volume		1.964	21:46
Air Blank		0.000	21:46
Control Test		0.079	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 07/18/2026 02:42

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ALVARADO PABON

DATE/TIME OF TEST: 06/21/2026 04:13:55

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.104	04:09
Breath Volume		1.324	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.102	04:12
Breath Volume		1.253	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 07/18/2026 02:42

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: NAVARRO FLORES

DATE/TIME OF TEST: 06/22/2026 03:08:18

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.146	03:02
Breath Volume		1.265	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.141	03:06
Breath Volume		1.812	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 08/26/2026 13:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: SAMAYOA PEREZ

DATE/TIME OF TEST: 06/22/2026 03:18:31

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.058	03:14
Breath Volume		1.234	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.067	03:17
Breath Volume		1.148	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 08/26/2026 13:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: RODRIGUEZ,RICHARD
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: COURT ORD
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 06/24/2026 09:53:17

Diagnostics Check		OK	09:47
Air Blank		0.000	09:47
Control Test		0.080	09:48
Air Blank		0.000	09:48
Subject Sample	# 1	0.000	09:49
Breath Volume		1.660	09:49
Air Blank		0.000	09:49
Air Blank		0.000	09:51
Subject Sample	# 2	0.000	09:51
Breath Volume		1.511	09:51
Air Blank		0.000	09:52
Control Test		0.079	09:52
Air Blank		0.000	09:53
Diagnostics Check		OK	09:53

Data Download Date/Time: 08/26/2026 13:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ARMSTRONG,NOAH
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 06/25/2026 04:17:00

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.104	04:12
Breath Volume		1.851	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.099	04:15
Breath Volume		2.023	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:17

Data Download Date/Time: 08/26/2026 13:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: JOHNSON,GREGORY
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 06/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403627559
DRY GAS EXPIRY DATE: 05/07/2029
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: REYES REYES

DATE/TIME OF TEST: 06/26/2026 03:29:26

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.108	03:25
Breath Volume		1.707	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.107	03:27
Breath Volume		1.433	03:27
Air Blank		0.000	03:28
Control Test		0.080	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 08/26/2026 13:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALABISE,THOMAS,A
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MARKLE,WILLIAM
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: DAUGHERTY

DATE/TIME OF TEST: 06/02/2026 03:17:00

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:14	
Control Test		0.080	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	RFI Detect
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 06/14/2026 13:01

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALABISE,THOMAS,A
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MARKLE,WILLIAM
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: DAUGHERTY

DATE/TIME OF TEST: 06/02/2026 03:25:17

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.026	03:20
Breath Volume		1.320	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:22
Subject Sample	# 2	0.025	03:23
Breath Volume		1.605	03:23
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 06/14/2026 13:01

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: FRANCA SANTOS,RAFAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 06/06/2026 04:21:29

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.143	04:17
Breath Volume		2.250	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.141	04:19
Breath Volume		1.546	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 06/14/2026 13:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAL,JACOB,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DIAL,JACOB
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: PALKO

DATE/TIME OF TEST: 06/10/2026 00:46:02

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.104	00:41
Breath Volume		1.285	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.102	00:44
Breath Volume		1.507	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:46

Data Download Date/Time: 06/14/2026 13:01

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAL,J,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DEPRADO,D
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 06/11/2026 16:52:53

Diagnostics Check		OK	16:46
Air Blank		0.000	16:47
Control Test		0.081	16:47
Air Blank		0.000	16:48
Subject Sample	# 1	0.000	16:48
Breath Volume		1.796	16:48
Air Blank		0.000	16:49
Air Blank		0.000	16:51
Subject Sample	# 2	0.000	16:51
Breath Volume		1.710	16:51
Air Blank		0.000	16:51
Control Test		0.079	16:52
Air Blank		0.000	16:52
Diagnostics Check		OK	16:52

Data Download Date/Time: 06/14/2026 13:01

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHLER,CARLITA,Y
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MAHLER,CARLITA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MULLENSKI

DATE/TIME OF TEST: 06/12/2026 01:46:37

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.123	01:41
Breath Volume		3.656	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.121	01:45
Breath Volume		2.632	01:45
Air Blank		0.000	01:45
Control Test		0.079	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 06/14/2026 13:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: WALKER,J
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/12/2026 22:10:43

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.185	22:05
Breath Volume		2.691	22:05
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.176	22:09
Breath Volume		3.121	22:09
Air Blank		0.000	22:09
Control Test		0.078	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 06/14/2026 13:01

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RILEY,ALLISON,F
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HALL,ALANA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 12:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: KEENAN

DATE/TIME OF TEST: 06/13/2026 13:21:59

Diagnostics Check		OK	13:16
Air Blank		0.000	13:16
Control Test		0.080	13:16
Air Blank		0.000	13:17
Subject Sample	# 1	0.000	13:17
Breath Volume		3.343	13:17
Air Blank		0.000	13:18
Air Blank		0.000	13:20
Subject Sample	# 2	0.000	13:20
Breath Volume		3.394	13:20
Air Blank		0.000	13:21
Control Test		0.079	13:21
Air Blank		0.000	13:21
Diagnostics Check		OK	13:21

Data Download Date/Time: 06/14/2026 13:01

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,WILLIAM,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: WENZLOFF,B
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/14/2026 02:42:40

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.080	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.072	02:38
Breath Volume		2.578	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.073	02:41
Breath Volume		3.390	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 06/14/2026 13:01

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: FRANCA SANTOS,RAFAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: KUR

DATE/TIME OF TEST: 06/19/2026 23:00:43

Diagnostics Check		OK	22:54
Air Blank		0.000	22:54
Control Test		0.080	22:55
Air Blank		0.000	22:55
Subject Sample	# 1	0.260	22:55
Breath Volume		2.550	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.257	22:59
Breath Volume		2.320	22:59
Air Blank		0.000	22:59
Control Test		0.079	23:00
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 07/26/2026 07:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKWITH,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: GUZMAN,K
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 10:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: LANDWERMEYER

DATE/TIME OF TEST: 06/20/2026 11:25:01

Diagnostics Check		OK	11:18
Air Blank		0.000	11:19
Control Test		0.080	11:19
Air Blank		0.000	11:20
Subject Sample	# 1	0.000	11:20
Breath Volume		1.632	11:20
Air Blank		0.000	11:21
Air Blank		0.000	11:23
Subject Sample	# 2	0.000	11:23
Breath Volume		1.808	11:23
Air Blank		0.000	11:24
Control Test		0.079	11:24
Air Blank		0.000	11:24
Diagnostics Check		OK	11:25

Data Download Date/Time: 07/26/2026 07:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,JUSTIN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RILEY,ALLISON
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: MUNSON

DATE/TIME OF TEST: 06/22/2026 17:46:02

Diagnostics Check		OK	17:40
Air Blank		0.000	17:40
Control Test		0.079	17:40
Air Blank		0.000	17:41
Subject Sample	# 1	0.000	17:41
Breath Volume		1.191	17:41
Air Blank		0.000	17:42
Air Blank		0.000	17:44
Subject Sample	# 2	0.000	17:44
Breath Volume		1.136	17:44
Air Blank		0.000	17:45
Control Test		0.079	17:45
Air Blank		0.000	17:45
Diagnostics Check		OK	17:46

Data Download Date/Time: 07/26/2026 07:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,DERICK,F
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: PERRY,KAYLA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/23/2026 02:26:41

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.231	02:19
Breath Volume		2.589	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:21
Subject Sample	# 2	0.206	02:22
Breath Volume		1.523	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 3	0.217	02:25
Breath Volume		2.546	02:25
Air Blank		0.000	02:25
Control Test		0.077	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 07/26/2026 07:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAL,J,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: ANDERSON,B
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 09:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:05

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 06/23/2026 09:49:10

Diagnostics Check		OK	09:43
Air Blank		0.000	09:43
Control Test		0.079	09:43
Air Blank		0.000	09:44
Subject Sample	# 1	0.000	09:44
Breath Volume		1.796	09:44
Air Blank		0.000	09:45
Air Blank		0.000	09:47
Subject Sample	# 2	0.000	09:47
Breath Volume		1.410	09:47
Air Blank		0.000	09:48
Control Test		0.078	09:48
Air Blank		0.000	09:49
Diagnostics Check		OK	09:49

Data Download Date/Time: 07/26/2026 07:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDSTEIN,SHELLI,R
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: BURDIGE,L
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 06/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 06/29/2026 19:16:45

Diagnostics Check		OK	19:06
Air Blank		0.000	19:06
Control Test		0.078	19:06
Air Blank		0.000	19:07
Subject Sample	# 1	0.299	19:07
Breath Volume		1.707	19:07
Air Blank		0.000	19:08
Air Blank		0.000	19:10
Subject Sample	# 2	0.263	19:12
Breath Volume		2.488	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:14
Subject Sample	# 3	0.270	19:15
Breath Volume		2.363	19:15
Air Blank		0.000	19:15
Control Test		0.075	19:16
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 07/26/2026 07:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGDON,TODD,S
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: BISEL,BRANDON
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A3
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MCWHORTER

DATE/TIME OF TEST: 06/20/2026 23:05:17

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.080	22:58
Air Blank		0.000	22:58
Subject Sample	# 1	0.100	23:00
Breath Volume		2.281	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:02
Subject Sample	# 2	0.096	23:03
Breath Volume		2.500	23:03
Air Blank		0.000	23:04
Control Test		0.080	23:04
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 06/25/2026 22:27

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A3
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: VARGAS MAGANA

DATE/TIME OF TEST: 06/27/2026 00:32:24

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.163	00:27
Breath Volume		1.402	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.175	00:30
Breath Volume		2.683	00:30
Air Blank		0.000	00:31
Control Test		0.079	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 07/24/2026 20:22

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A3
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: BOOS

DATE/TIME OF TEST: 06/29/2026 23:49:47

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.079	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.137	23:45
Breath Volume		1.203	23:45
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.139	23:48
Breath Volume		1.445	23:48
Air Blank		0.000	23:48
Control Test		0.080	23:49
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 07/24/2026 20:22

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINICIS,LOURDES,P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LOPEZ,TIMOTHY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: DUNCAN CAMPO

DATE/TIME OF TEST: 06/05/2026 21:38:45

Diagnostics Check		OK	21:31
Air Blank		0.000	21:31
Control Test		0.076	21:32
Air Blank		0.000	21:32
Subject Sample	# 1	0.175	21:33
Breath Volume		1.859	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:35
Subject Sample	# 2	0.176	21:37
Breath Volume		2.906	21:37
Air Blank		0.000	21:37
Control Test		0.076	21:38
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 06/28/2026 15:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GRAHAM JR

DATE/TIME OF TEST: 06/05/2026 21:58:56

Diagnostics Check		OK	21:49	
Air Blank		0.000	21:49	
Control Test		0.076	21:49	
Air Blank		0.000	21:50	
Subject Sample	# 1	0.219	21:52	
Breath Volume		1.218	21:52	
Air Blank		0.000	21:52	
Air Blank		0.000	21:54	
Subject Sample	# 2	0.218	21:57	
Breath Volume		1.269	21:57	
Air Blank		0.000	21:57	
Control Test		0.074	21:58	Control Outside Tolerance
Air Blank		0.000	21:58	

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GRAHAM JR

DATE/TIME OF TEST: 06/05/2026 22:20:12

Diagnostics Check		OK	22:04	
Air Blank		0.000	22:04	
Control Test		0.076	22:04	
Air Blank		0.000	22:05	
Subject Sample	# 1	0.206	22:07	
Breath Volume		1.382	22:07	
Air Blank		0.000	22:07	
Air Blank		0.000	22:09	
Subject Sample	# 2	0.187	22:12	Volume Not Met
Breath Volume		0.085	22:12	
Air Blank		0.000	22:13	
Air Blank		0.000	22:15	
Subject Sample	# 3	0.197	22:18	Volume Not Met
Breath Volume		1.097	22:18	
Air Blank		0.000	22:19	
Control Test		0.075	22:19	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 06/28/2026 15:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STONE,KENNETH,T
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GRAHAM JR

DATE/TIME OF TEST: 06/05/2026 22:30:50

Diagnostics Check		OK	22:22	
Air Blank		0.000	22:22	
Control Test		0.075	22:22	
Air Blank		0.000	22:23	
Subject Sample	# 1	0.225	22:26	
Breath Volume		1.722	22:26	
Air Blank		0.000	22:26	
Air Blank		0.000	22:28	
Subject Sample	# 2	0.220	22:29	
Breath Volume		1.632	22:29	
Air Blank		0.000	22:29	
Control Test		0.074	22:30	Control Outside Tolerance
Air Blank		0.000	22:30	

Data Download Date/Time: 06/28/2026 15:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,AXEL,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: ORTIZ,AXEL
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: FAZZOLARE

DATE/TIME OF TEST: 06/06/2026 02:40:07

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.105	02:35
Breath Volume		3.562	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:37
Subject Sample	# 2	0.107	02:38
Breath Volume		3.125	02:38
Air Blank		0.000	02:39
Control Test		0.077	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 06/28/2026 15:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOWICKI,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: NOWICKI,JUSTIN
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BURD

DATE/TIME OF TEST: 06/06/2026 22:20:47

Diagnostics Check		OK	22:14
Air Blank		0.000	22:15
Control Test		0.078	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.123	22:16
Breath Volume		1.656	22:16
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.122	22:19
Breath Volume		1.914	22:19
Air Blank		0.000	22:19
Control Test		0.077	22:20
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINICIS,LOURDES,P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LABOY,LONITA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/07/2026 18:45:37

Diagnostics Check		OK	18:39
Air Blank		0.000	18:39
Control Test		0.077	18:40
Air Blank		0.000	18:40
Subject Sample	# 1	0.268	18:41
Breath Volume		1.222	18:41
Air Blank		0.000	18:41
Air Blank		0.000	18:43
Subject Sample	# 2	0.282	18:43
Breath Volume		1.976	18:43
Air Blank		0.000	18:44
Control Test		0.076	18:44
Air Blank		0.000	18:45
Diagnostics Check		OK	18:45

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLARD,F,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: POLLARD,FILIPE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: AUTRY

DATE/TIME OF TEST: 06/12/2026 03:30:23

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.076	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.163	03:26
Breath Volume		2.257	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.144	03:28
Breath Volume		1.328	03:28
Air Blank		0.000	03:29
Control Test		0.077	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY,RORY,P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: RODENBERGER,MADELEINE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DUNIKOSKI

DATE/TIME OF TEST: 06/13/2026 02:00:39

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.078	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1	0.030	01:51	Volume Not Met
Breath Volume		0.015	01:51	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 2	0.040	01:55	
Breath Volume		1.246	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 3	0.038	01:59	
Breath Volume		1.957	01:59	
Air Blank		0.000	01:59	
Control Test		0.075	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY,RORY,P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCROSKY,SCOTT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/13/2026 05:00:59

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.077	04:55
Air Blank		0.000	04:55
Subject Sample	# 1	0.159	04:56
Breath Volume		2.390	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:58
Subject Sample	# 2	0.159	04:59
Breath Volume		2.054	04:59
Air Blank		0.000	05:00
Control Test		0.077	05:00
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRAGHER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: FARRAGHER,THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:23

SUBJECT'S LAST NAME: BRENGMAN

DATE/TIME OF TEST: 06/18/2026 10:26:01

Diagnostics Check		OK	10:19
Air Blank		0.000	10:20
Control Test		0.078	10:20
Air Blank		0.000	10:21
Subject Sample	# 1	0.118	10:21
Breath Volume		1.417	10:21
Air Blank		0.000	10:22
Air Blank		0.000	10:23
Subject Sample	# 2	0.108	10:24
Breath Volume		1.441	10:24
Air Blank		0.000	10:25
Control Test		0.077	10:25
Air Blank		0.000	10:25
Diagnostics Check		OK	10:26

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLARD,FILIPE,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: POLLARD,FILIPE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 06/20/2026 01:10:53

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.077	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.116	01:06
Breath Volume		4.066	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:08
Subject Sample	# 2	0.113	01:09
Breath Volume		3.699	01:09
Air Blank		0.000	01:09
Control Test		0.077	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,AXEL,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: ORTIZ,AXEL
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/21/2026 00:51:23

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.077	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.172	00:46
Breath Volume		3.175	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.176	00:49
Breath Volume		2.949	00:49
Air Blank		0.000	00:50
Control Test		0.076	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 06/28/2026 15:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRASMANN,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MALDONADO,JOSHUA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: TARALLO

DATE/TIME OF TEST: 06/23/2026 00:17:38

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:09	
Control Test		0.078	00:09	
Air Blank		0.000	00:09	
Subject Sample	# 1	0.340	00:10	
Breath Volume		3.722	00:10	
Air Blank		0.000	00:11	
Air Blank		0.000	00:12	
Subject Sample	# 2	0.319	00:13	
Breath Volume		5.250	00:13	
Air Blank		0.000	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 3	0.321	00:16	
Breath Volume		5.050	00:16	
Air Blank		0.000	00:16	
Control Test		0.072	00:17	Control Outside Tolerance
Air Blank		0.000	00:17	

Data Download Date/Time: 06/28/2026 15:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRASMANN,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BURNS,BLAKE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: LARNER

DATE/TIME OF TEST: 06/23/2026 00:28:16

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.078	00:22
Air Blank		0.000	00:23
Subject Sample	# 1	0.208	00:23
Breath Volume		2.511	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.207	00:26
Breath Volume		2.050	00:26
Air Blank		0.000	00:27
Control Test		0.075	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEAT,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MCCROSKEY,SCOTT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 10/05/2025
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: KOPP

DATE/TIME OF TEST: 06/28/2026 04:10:29

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.078	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.272	04:05
Breath Volume		4.015	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.264	04:08
Breath Volume		3.707	04:08
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 06/28/2026 15:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001010

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIALOBRZESKI,M,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WENZLOFF,B
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24923080A3
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: BARGE

DATE/TIME OF TEST: 06/28/2026 06:57:13

Diagnostics Check		OK	06:51
Air Blank		0.000	06:51
Control Test		0.077	06:51
Air Blank		0.000	06:52
Subject Sample	# 1	0.222	06:52
Breath Volume		1.937	06:52
Air Blank		0.000	06:53
Air Blank		0.000	06:55
Subject Sample	# 2	0.216	06:55
Breath Volume		2.128	06:55
Air Blank		0.000	06:56
Control Test		0.076	06:56
Air Blank		0.000	06:57
Diagnostics Check		OK	06:57

Data Download Date/Time: 06/28/2026 15:55

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAUFIELD,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SALGADO AVILA

DATE/TIME OF TEST: 06/04/2026 01:11:30

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.000	01:07
Breath Volume		2.050	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.000	01:10
Breath Volume		1.734	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 07/01/2026 09:48

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,H,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORETA,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: TRIPLETT

DATE/TIME OF TEST: 06/07/2026 01:30:35

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.079	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:29	Subject Test Refused
Breath Volume		0.000	01:29	
Air Blank		0.000	01:29	
Control Test		0.080	01:29	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 07/01/2026 09:48

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,ZACHARY,G
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/11/2026 23:41:49

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:39	
Control Test		0.080	23:39	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:40	Subject Test Refused
Breath Volume		0.000	23:40	
Air Blank		0.000	23:40	
Control Test		0.080	23:41	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/11/2026 23:47:40

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:44	
Control Test		0.080	23:45	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:46	Subject Test Refused
Breath Volume		0.000	23:46	
Air Blank		0.000	23:46	
Control Test		0.079	23:47	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,J,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GEIS,J
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: MALLORY

DATE/TIME OF TEST: 06/12/2026 04:55:40

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:52	
Control Test		0.080	04:52	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:54	Subject Test Refused
Breath Volume		0.000	04:54	
Air Blank		0.000	04:54	
Control Test		0.080	04:55	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LOKMIC,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: ROSARIO DIAZ

DATE/TIME OF TEST: 06/13/2026 00:09:27

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.079	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.213	00:04
Breath Volume		2.656	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.208	00:07
Breath Volume		2.386	00:07
Air Blank		0.000	00:08
Control Test		0.078	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MANSEN

DATE/TIME OF TEST: 06/13/2026 01:08:11

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.079	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.000	01:03
Breath Volume		2.753	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.000	01:06
Breath Volume		3.156	01:06
Air Blank		0.000	01:07
Control Test		0.079	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VYKHRIST,S,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 06/13/2026 02:28:40

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.237	02:23
Breath Volume		1.433	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.237	02:26
Breath Volume		1.234	02:26
Air Blank		0.000	02:27
Control Test		0.077	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: GUERRA RODRIGUEZ

DATE/TIME OF TEST: 06/18/2026 21:21:45

Diagnostics Check		OK	21:13
Air Blank		0.000	21:14
Control Test		0.080	21:14
Air Blank		0.000	21:15
Subject Sample	# 1	0.129	21:16
Breath Volume		2.132	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.122	21:20
Breath Volume		1.964	21:20
Air Blank		0.000	21:20
Control Test		0.078	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,J,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: VELASQUES-ROBLERO

DATE/TIME OF TEST: 06/20/2026 00:16:09

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.236	00:10
Breath Volume		2.890	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.228	00:14
Breath Volume		2.722	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 07/01/2026 09:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: D OLIVEIRA CARVALHO

DATE/TIME OF TEST: 06/25/2026 01:19:22

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:16	
Control Test		0.079	01:17	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.079	01:18	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/27/2026 01:03:17

Diagnostics Check		OK	00:56
Air Blank		0.000	00:57
Control Test		0.080	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.052	00:58
Breath Volume		1.800	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.051	01:01
Breath Volume		1.789	01:01
Air Blank		0.000	01:02
Control Test		0.079	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRSTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,I
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/27/2026 03:24:47

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.080	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.164	03:20
Breath Volume		1.960	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.165	03:23
Breath Volume		1.957	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 07/01/2026 09:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/05/2026 02:04:37

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:02	
Control Test		0.078	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:03	
Control Test		0.077	02:04	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 06/25/2026 07:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: KOLLORE

DATE/TIME OF TEST: 06/05/2026 23:21:30

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.079	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.150	23:17
Breath Volume		2.285	23:17
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.149	23:20
Breath Volume		2.382	23:20
Air Blank		0.000	23:20
Control Test		0.078	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 06/25/2026 07:09

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 06/06/2026 02:31:29

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.078	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.176	02:26
Breath Volume		1.894	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.177	02:30
Breath Volume		1.714	02:30
Air Blank		0.000	02:30
Control Test		0.078	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 06/25/2026 07:09

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/06/2026 05:50:42

Diagnostics Check		OK	05:44
Air Blank		0.000	05:44
Control Test		0.078	05:45
Air Blank		0.000	05:45
Subject Sample	# 1	0.076	05:46
Breath Volume		1.765	05:46
Air Blank		0.000	05:46
Air Blank		0.000	05:48
Subject Sample	# 2	0.078	05:49
Breath Volume		1.917	05:49
Air Blank		0.000	05:49
Control Test		0.077	05:50
Air Blank		0.000	05:50
Diagnostics Check		OK	05:50

Data Download Date/Time: 06/25/2026 07:09

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: PAPICH

DATE/TIME OF TEST: 06/06/2026 20:42:45

Diagnostics Check		OK	20:36
Air Blank		0.000	20:37
Control Test		0.079	20:37
Air Blank		0.000	20:37
Subject Sample	# 1	0.000	20:38
Breath Volume		1.664	20:38
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.000	20:41
Breath Volume		1.199	20:41
Air Blank		0.000	20:41
Control Test		0.079	20:42
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 06/25/2026 07:09

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: NAJAR

DATE/TIME OF TEST: 06/07/2026 02:56:17

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.078	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.124	02:51
Breath Volume		1.777	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.122	02:54
Breath Volume		2.039	02:54
Air Blank		0.000	02:55
Control Test		0.078	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 06/25/2026 07:09

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: DE OLIVEIRA

DATE/TIME OF TEST: 06/14/2026 02:40:51

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.036	02:36
Breath Volume		1.804	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.033	02:39
Breath Volume		1.148	02:39
Air Blank		0.000	02:39
Control Test		0.078	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 06/25/2026 07:09

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: VENDRELL-SANCHEZ

DATE/TIME OF TEST: 06/18/2026 20:49:26

Diagnostics Check		OK	20:43
Air Blank		0.000	20:43
Control Test		0.079	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.213	20:45
Breath Volume		2.996	20:45
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.219	20:47
Breath Volume		2.765	20:47
Air Blank		0.000	20:48
Control Test		0.078	20:48
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 06/25/2026 07:09

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLIN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/19/2026 01:22:33

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.078	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.051	01:17
Breath Volume		1.644	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.050	01:21
Breath Volume		1.644	01:21
Air Blank		0.000	01:21
Control Test		0.078	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 06/25/2026 07:09

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/19/2026 18:19:25

Diagnostics Check		OK	18:12
Air Blank		0.000	18:13
Control Test		0.078	18:13
Air Blank		0.000	18:14
Subject Sample	# 1	0.218	18:14
Breath Volume		1.597	18:14
Air Blank		0.000	18:15
Air Blank		0.000	18:17
Subject Sample	# 2	0.215	18:17
Breath Volume		2.324	18:17
Air Blank		0.000	18:18
Control Test		0.077	18:18
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 06/25/2026 07:10

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: NOSOWICZ

DATE/TIME OF TEST: 06/19/2026 21:38:46

Diagnostics Check		OK	21:32
Air Blank		0.000	21:33
Control Test		0.079	21:33
Air Blank		0.000	21:34
Subject Sample	# 1	0.049	21:34
Breath Volume		1.675	21:34
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.052	21:37
Breath Volume		1.460	21:37
Air Blank		0.000	21:37
Control Test		0.078	21:38
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 06/25/2026 07:10

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: FEOLA

DATE/TIME OF TEST: 06/20/2026 00:50:15

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.077	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.122	00:45
Breath Volume		1.550	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.123	00:48
Breath Volume		1.566	00:48
Air Blank		0.000	00:49
Control Test		0.078	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 06/25/2026 07:10

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: CALDWELL JR.,GARY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 06/20/2026 02:43:51

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:41	
Control Test		0.078	02:41	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Control Test		0.078	02:43	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 06/25/2026 07:10

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: FLEURIMOND

DATE/TIME OF TEST: 06/20/2026 02:55:29

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.077	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.000	02:48
Breath Volume		1.128	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.000	02:54
Breath Volume		2.062	02:54
Air Blank		0.000	02:54
Control Test		0.077	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 06/25/2026 07:10

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BOHAN

DATE/TIME OF TEST: 06/20/2026 05:25:20

Diagnostics Check		OK	05:19
Air Blank		0.000	05:19
Control Test		0.078	05:20
Air Blank		0.000	05:20
Subject Sample	# 1	0.000	05:21
Breath Volume		2.628	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.000	05:23
Breath Volume		2.234	05:23
Air Blank		0.000	05:24
Control Test		0.078	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 06/25/2026 07:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIEGLAFF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: TRIEGLAFF,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 06/22/2026 06:27:13

Diagnostics Check		OK	06:19
Air Blank		0.000	06:19
Control Test		0.078	06:19
Air Blank		0.000	06:20
Subject Sample	# 1	0.000	06:21
Breath Volume		1.558	06:21
Air Blank		0.000	06:21
Air Blank		0.000	06:23
Subject Sample	# 2	0.000	06:25
Breath Volume		1.406	06:25
Air Blank		0.000	06:26
Control Test		0.078	06:26
Air Blank		0.000	06:27
Diagnostics Check		OK	06:27

Data Download Date/Time: 06/25/2026 07:10

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:47

SUBJECT'S LAST NAME: PAPAGIANI

DATE/TIME OF TEST: 06/24/2026 15:13:16

Diagnostics Check		OK	15:07
Air Blank		0.000	15:07
Control Test		0.078	15:08
Air Blank		0.000	15:08
Subject Sample	# 1	0.000	15:08
Breath Volume		2.414	15:08
Air Blank		0.000	15:09
Air Blank		0.000	15:11
Subject Sample	# 2	0.000	15:11
Breath Volume		2.363	15:11
Air Blank		0.000	15:12
Control Test		0.079	15:12
Air Blank		0.000	15:13
Diagnostics Check		OK	15:13

Data Download Date/Time: 06/25/2026 07:10

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STERLING,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ADAMS,STERLING
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ERAZO

DATE/TIME OF TEST: 06/28/2026 04:00:03

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.078	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.022	03:55
Breath Volume		1.808	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.019	03:58
Breath Volume		1.132	03:58
Air Blank		0.000	03:59
Control Test		0.078	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	04:00

Data Download Date/Time: 07/30/2026 21:29

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPAU,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WALDROP,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HAVENS

DATE/TIME OF TEST: 06/04/2026 02:25:40

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:17	
Control Test		0.080	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:21	No Sample Provided
Breath Volume		0.000	02:21	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:24	
Control Test		0.080	02:25	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 06/28/2026 18:25

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVESAY,M,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/05/2026 19:17:42

Diagnostics Check		OK	19:10	
Air Blank		0.000	19:10	
Control Test		0.078	19:11	
Air Blank		0.000	19:11	
Subject Sample	# 1	0.223	19:13	
Breath Volume		1.507	19:13	
Air Blank		0.000	19:13	
Air Blank		0.000	19:15	
Subject Sample	# 2	0.223	19:16	
Breath Volume		2.023	19:16	
Air Blank		0.000	19:16	
Control Test		0.070	19:17	Control Outside Tolerance
Air Blank		0.000	19:17	

Data Download Date/Time: 06/28/2026 18:25

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ERNEST,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITNEY,K
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 06/09/2026 19:17:02

Diagnostics Check		OK	19:07	
Air Blank		0.000	19:08	
Control Test		0.080	19:08	
Air Blank		0.000	19:09	
Subject Sample	# 1		19:12	No Sample Provided
Breath Volume		0.000	19:12	
Air Blank		0.000	19:12	
Air Blank		0.000	19:14	
Subject Sample	# 2		19:15	Subject Test Refused
Breath Volume		0.000	19:15	
Air Blank		0.000	19:16	
Control Test		0.080	19:16	
Air Blank		0.000	19:16	
Diagnostics Check		OK	19:17	

Data Download Date/Time: 06/28/2026 18:25

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPAU,A,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WALDROP,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 06/13/2026 22:29:42

Diagnostics Check		OK	22:22
Air Blank		0.000	22:22
Control Test		0.079	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.203	22:25
Breath Volume		1.281	22:25
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.205	22:28
Breath Volume		1.398	22:28
Air Blank		0.000	22:28
Control Test		0.078	22:29
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 06/28/2026 18:25

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,J,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: STUMP

DATE/TIME OF TEST: 06/20/2026 03:43:49

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.080	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.170	03:39
Breath Volume		1.519	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.159	03:42
Breath Volume		1.750	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 06/28/2026 18:25

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVESAY,M,E
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WILSON,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: ENTERILI

DATE/TIME OF TEST: 06/26/2026 04:26:01

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.084	04:21
Breath Volume		3.496	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 2	0.083	04:24
Breath Volume		4.003	04:24
Air Blank		0.000	04:25
Control Test		0.079	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:26

Data Download Date/Time: 06/28/2026 18:25

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSE,ERNEST,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: COLEMAN,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 05/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 06/27/2026 01:52:09

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.236	01:47
Breath Volume		3.722	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.231	01:50
Breath Volume		2.667	01:50
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 06/28/2026 18:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,RONALD,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BOLENA,B
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WILBUR

DATE/TIME OF TEST: 06/29/2026 01:25:36

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:21	No Sample Provided
Breath Volume		0.000	01:21	
Air Blank		0.000	01:21	
Air Blank		0.000	01:23	
Subject Sample	# 2		01:24	Subject Test Refused
Breath Volume		0.000	01:24	
Air Blank		0.000	01:24	
Control Test		0.079	01:25	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 07/29/2026 14:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,K,R
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WALDROP,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: GIZAW

DATE/TIME OF TEST: 06/29/2026 06:48:12

Diagnostics Check		OK	06:41
Air Blank		0.000	06:42
Control Test		0.079	06:42
Air Blank		0.000	06:42
Subject Sample	# 1	0.156	06:43
Breath Volume		2.972	06:43
Air Blank		0.000	06:44
Air Blank		0.000	06:45
Subject Sample	# 2	0.145	06:46
Breath Volume		2.316	06:46
Air Blank		0.000	06:47
Control Test		0.079	06:47
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 07/29/2026 14:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADOCCHIO,CASEY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CHUKES,DARREL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 08:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: OCTOBER

DATE/TIME OF TEST: 06/01/2026 09:48:51

Diagnostics Check		OK	09:41
Air Blank		0.000	09:41
Control Test		0.078	09:42
Air Blank		0.000	09:42
Subject Sample	# 1	0.107	09:43
Breath Volume		1.394	09:43
Air Blank		0.000	09:44
Air Blank		0.000	09:46
Subject Sample	# 2	0.105	09:47
Breath Volume		2.351	09:47
Air Blank		0.000	09:47
Control Test		0.078	09:48
Air Blank		0.000	09:48
Diagnostics Check		OK	09:48

Data Download Date/Time: 06/24/2026 09:36

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STARLING,KELLEY,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STARLING,KELLEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: DOMAGALA

DATE/TIME OF TEST: 06/05/2026 14:53:33

Diagnostics Check		OK	14:47
Air Blank		0.000	14:47
Control Test		0.079	14:48
Air Blank		0.000	14:48
Subject Sample	# 1	0.163	14:49
Breath Volume		2.179	14:49
Air Blank		0.000	14:49
Air Blank		0.000	14:51
Subject Sample	# 2	0.168	14:52
Breath Volume		2.082	14:52
Air Blank		0.000	14:52
Control Test		0.077	14:52
Air Blank		0.000	14:53
Diagnostics Check		OK	14:53

Data Download Date/Time: 06/24/2026 09:36

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 06/12/2026 09:23:15

Diagnostics Check		OK	09:15
Air Blank		0.000	09:15
Control Test		0.078	09:16
Air Blank		0.000	09:16
Subject Sample	# 1	0.070	09:17
Breath Volume		1.242	09:17
Air Blank		0.000	09:18
Air Blank		0.000	09:20
Subject Sample	# 2	0.070	09:21
Breath Volume		1.109	09:21
Air Blank		0.000	09:22
Control Test		0.078	09:22
Air Blank		0.000	09:23
Diagnostics Check		OK	09:23

Data Download Date/Time: 06/24/2026 09:36

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: PARADY

DATE/TIME OF TEST: 06/13/2026 01:17:12

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.078	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.177	01:12
Breath Volume		1.503	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.166	01:15
Breath Volume		1.640	01:15
Air Blank		0.000	01:16
Control Test		0.077	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 06/24/2026 09:36

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTO,JOHN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTO,JOHN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 06/20/2026 03:49:02

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.078	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.284	03:44
Breath Volume		2.074	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.274	03:47
Breath Volume		1.492	03:47
Air Blank		0.000	03:48
Control Test		0.077	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:49

Data Download Date/Time: 06/24/2026 09:36

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,ADRIANNE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WATSON,SAMUEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: KATONA

DATE/TIME OF TEST: 06/20/2026 22:17:00

Diagnostics Check		OK	22:10
Air Blank		0.000	22:11
Control Test		0.079	22:11
Air Blank		0.000	22:12
Subject Sample	# 1	0.000	22:12
Breath Volume		4.125	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.000	22:15
Breath Volume		4.046	22:15
Air Blank		0.000	22:16
Control Test		0.079	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:17

Data Download Date/Time: 06/24/2026 09:36

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: CHEEK

DATE/TIME OF TEST: 06/22/2026 21:35:10

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.273	21:30
Breath Volume		1.136	21:30
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.275	21:33
Breath Volume		1.316	21:33
Air Blank		0.000	21:34
Control Test		0.078	21:34
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 06/24/2026 09:37

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIBERLY,RYAN,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WIBERLY,RYAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/23/2026 19:43:53

Diagnostics Check		OK	19:37
Air Blank		0.000	19:38
Control Test		0.079	19:38
Air Blank		0.000	19:39
Subject Sample	# 1	0.000	19:39
Breath Volume		1.820	19:39
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.000	19:42
Breath Volume		1.515	19:42
Air Blank		0.000	19:42
Control Test		0.079	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 06/24/2026 09:37

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY,GRANT,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY,GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: ABDELNOUR

DATE/TIME OF TEST: 06/27/2026 01:42:43

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.160	01:38
Breath Volume		1.632	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.157	01:41
Breath Volume		1.367	01:41
Air Blank		0.000	01:41
Control Test		0.077	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 07/17/2026 08:18

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY,GRANT,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY,GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: IRIZARRY PONCE

DATE/TIME OF TEST: 06/28/2026 01:23:14

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.084	01:17
Breath Volume		1.664	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.088	01:21
Breath Volume		1.269	01:21
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 07/17/2026 08:18

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: MOAK

DATE/TIME OF TEST: 06/28/2026 08:07:46

Diagnostics Check		OK	08:01
Air Blank		0.000	08:02
Control Test		0.078	08:02
Air Blank		0.000	08:02
Subject Sample	# 1	0.177	08:03
Breath Volume		1.523	08:03
Air Blank		0.000	08:03
Air Blank		0.000	08:05
Subject Sample	# 2	0.189	08:06
Breath Volume		1.558	08:06
Air Blank		0.000	08:06
Control Test		0.078	08:07
Air Blank		0.000	08:07
Diagnostics Check		OK	08:07

Data Download Date/Time: 07/17/2026 08:18

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOWAY,GRANT,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GALLOWAY,GRANT
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: GARCIA GONZALEZ

DATE/TIME OF TEST: 06/29/2026 02:12:06

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.201	02:04
Breath Volume		2.078	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.180	02:07
Breath Volume		2.289	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 3	0.164	02:10
Breath Volume		2.609	02:10
Air Blank		0.000	02:11
Control Test		0.077	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 07/17/2026 08:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,KEVIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 13:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: MACKENZIE

DATE/TIME OF TEST: 06/30/2026 14:37:10

Diagnostics Check		OK	14:31
Air Blank		0.000	14:31
Control Test		0.079	14:31
Air Blank		0.000	14:32
Subject Sample	# 1	0.000	14:32
Breath Volume		1.156	14:32
Air Blank		0.000	14:33
Air Blank		0.000	14:35
Subject Sample	# 2	0.000	14:35
Breath Volume		1.105	14:35
Air Blank		0.000	14:36
Control Test		0.079	14:36
Air Blank		0.000	14:37
Diagnostics Check		OK	14:37

Data Download Date/Time: 07/17/2026 08:19

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SAFAR,B
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 06/02/2026 00:38:16

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:25	
Control Test		0.080	00:26	
Air Blank		0.000	00:26	
Subject Sample	# 1	0.163	00:29	Volume Not Met
Breath Volume		0.632	00:29	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 2	0.158	00:33	
Breath Volume		1.140	00:33	
Air Blank		0.000	00:34	
Air Blank		0.000	00:35	
Subject Sample	# 3	0.161	00:36	
Breath Volume		1.261	00:36	
Air Blank		0.000	00:37	
Control Test		0.078	00:37	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 06/25/2026 05:58

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: FONTANA,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 06/02/2026 04:07:57

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.110	04:03
Breath Volume		1.238	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.119	04:06
Breath Volume		1.382	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 06/25/2026 05:58

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOIVIN,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ANSTETT,ANDREW
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: ROOT

DATE/TIME OF TEST: 06/03/2026 08:06:59

Diagnostics Check		OK	08:04	
Air Blank		0.000	08:04	
Control Test		0.079	08:04	
Air Blank		0.000	08:05	
Subject Sample	# 1		08:05	Subject Test Refused
Breath Volume		0.000	08:05	
Air Blank		0.000	08:06	
Control Test		0.079	08:06	
Air Blank		0.000	08:06	
Diagnostics Check		OK	08:06	

Data Download Date/Time: 06/25/2026 05:58

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEARS,COLBY,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SEARS,COLBY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: LEVINE

DATE/TIME OF TEST: 06/05/2026 01:48:01

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.153	01:43
Breath Volume		1.460	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:45
Subject Sample	# 2	0.156	01:46
Breath Volume		1.578	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:48

Data Download Date/Time: 06/25/2026 05:58

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NOAH,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SAFAR,B
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WHITMORE

DATE/TIME OF TEST: 06/05/2026 02:08:43

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:07	Subject Test Refused
Breath Volume		0.000	02:07	
Air Blank		0.000	02:07	
Control Test		0.079	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 06/25/2026 05:58

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,BENJAMIN,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MAGALSKI,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/07/2026 05:38:38

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.080	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.118	05:34
Breath Volume		1.460	05:34
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.109	05:37
Breath Volume		1.335	05:37
Air Blank		0.000	05:37
Control Test		0.080	05:38
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 06/25/2026 05:58

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NOAH,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HILLIARD,N
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: FARRIS

DATE/TIME OF TEST: 06/09/2026 19:17:10

Diagnostics Check		OK	19:10
Air Blank		0.000	19:11
Control Test		0.081	19:11
Air Blank		0.000	19:11
Subject Sample	# 1	0.000	19:12
Breath Volume		1.687	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:15
Subject Sample	# 2	0.000	19:15
Breath Volume		1.625	19:15
Air Blank		0.000	19:16
Control Test		0.081	19:16
Air Blank		0.000	19:17
Diagnostics Check		OK	19:17

Data Download Date/Time: 06/25/2026 05:59

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SAFAR,B
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LUKE

DATE/TIME OF TEST: 06/11/2026 01:49:10

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.091	01:44
Breath Volume		2.195	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:46
Subject Sample	# 2	0.095	01:47
Breath Volume		1.730	01:47
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 06/25/2026 05:59

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOIVIN,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BUTLER,MASON
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: NEWTON

DATE/TIME OF TEST: 06/11/2026 17:54:13

Diagnostics Check		OK	17:51	
Air Blank		0.000	17:51	
Control Test		0.080	17:52	
Air Blank		0.000	17:52	
Subject Sample	# 1		17:52	Subject Test Refused
Breath Volume		0.000	17:52	
Air Blank		0.000	17:53	
Control Test		0.080	17:53	
Air Blank		0.000	17:54	
Diagnostics Check		OK	17:54	

Data Download Date/Time: 06/25/2026 05:59

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HERRIGNTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: BOSTON

DATE/TIME OF TEST: 06/12/2026 15:00:06

Diagnostics Check		OK	14:53
Air Blank		0.000	14:54
Control Test		0.080	14:54
Air Blank		0.000	14:54
Subject Sample	# 1	0.000	14:55
Breath Volume		3.437	14:55
Air Blank		0.000	14:56
Air Blank		0.000	14:58
Subject Sample	# 2	0.000	14:58
Breath Volume		2.394	14:58
Air Blank		0.000	14:59
Control Test		0.080	14:59
Air Blank		0.000	15:00
Diagnostics Check		OK	15:00

Data Download Date/Time: 06/25/2026 05:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMACK,ZACHARY,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/13/2026 01:02:35

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.123	00:58
Breath Volume		2.316	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.121	01:01
Breath Volume		2.320	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 06/25/2026 05:59

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NOAH,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 06/18/2026 04:26:43

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.188	04:21
Breath Volume		1.187	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.181	04:25
Breath Volume		1.257	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 06/25/2026 05:59

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KEENEY,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: MENZ

DATE/TIME OF TEST: 06/19/2026 21:20:13

Diagnostics Check		OK	21:17	
Air Blank		0.000	21:17	
Control Test		0.080	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 1		21:18	Subject Test Refused
Breath Volume		0.000	21:18	
Air Blank		0.000	21:19	
Control Test		0.080	21:19	
Air Blank		0.000	21:20	
Diagnostics Check		OK	21:20	

Data Download Date/Time: 06/25/2026 05:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KEENEY,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SALATA

DATE/TIME OF TEST: 06/20/2026 02:27:46

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.110	02:22
Breath Volume		1.960	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.103	02:26
Breath Volume		1.632	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 06/25/2026 05:59

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SAFAR,B
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 06/20/2026 05:54:18

Diagnostics Check		OK	05:51	
Air Blank		0.000	05:51	
Control Test		0.080	05:52	
Air Blank		0.000	05:52	
Subject Sample	# 1		05:52	Subject Test Refused
Breath Volume		0.000	05:52	
Air Blank		0.000	05:53	
Control Test		0.079	05:53	
Air Blank		0.000	05:54	
Diagnostics Check		OK	05:54	

Data Download Date/Time: 06/25/2026 05:59

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THOMPSON,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: MARCHENA-MELENDEZ

DATE/TIME OF TEST: 06/20/2026 23:10:56

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.079	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.168	23:06
Breath Volume		2.363	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:08
Subject Sample	# 2	0.171	23:09
Breath Volume		1.277	23:09
Air Blank		0.000	23:10
Control Test		0.079	23:10
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 06/25/2026 06:00

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MAGALSKI,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: LUDY

DATE/TIME OF TEST: 06/22/2026 02:06:05

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.000	02:01
Breath Volume		1.976	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.000	02:04
Breath Volume		1.582	02:04
Air Blank		0.000	02:05
Control Test		0.081	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:06

Data Download Date/Time: 06/25/2026 06:00

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAL,JACOB,N
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HERRINGTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: DECKER

DATE/TIME OF TEST: 06/22/2026 17:14:34

Diagnostics Check		OK	17:01	
Air Blank		0.000	17:01	
Control Test		0.080	17:01	
Air Blank		0.000	17:02	
Subject Sample	# 1		17:05	No Sample Provided
Breath Volume		0.000	17:05	
Air Blank		0.000	17:05	
Air Blank		0.000	17:07	
Subject Sample	# 2	0.178	17:08	
Breath Volume		1.476	17:08	
Air Blank		0.000	17:09	
Air Blank		0.000	17:11	
Subject Sample	# 3	0.160	17:12	
Breath Volume		1.339	17:12	
Air Blank		0.000	17:13	
Control Test		0.079	17:13	
Air Blank		0.000	17:14	
Diagnostics Check		OK	17:14	

Data Download Date/Time: 06/25/2026 06:00

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NOAH,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SAFAR,B
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 06/23/2026 03:00:17

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.165	02:55
Breath Volume		1.550	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.167	02:58
Breath Volume		1.589	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 06/25/2026 06:00

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOIVIN,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: SPENCER,MITCHELL
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 06/24/2026 13:30:58

Diagnostics Check		OK	13:28	
Air Blank		0.000	13:28	
Control Test		0.079	13:28	
Air Blank		0.000	13:29	
Subject Sample	# 1		13:29	Subject Test Refused
Breath Volume		0.000	13:29	
Air Blank		0.000	13:30	
Control Test		0.080	13:30	
Air Blank		0.000	13:30	
Diagnostics Check		OK	13:30	

Data Download Date/Time: 06/25/2026 06:00

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,BENJAMIN,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: FONTANA,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: BOLLINGER

DATE/TIME OF TEST: 06/24/2026 23:31:45

Diagnostics Check		OK	23:24
Air Blank		0.000	23:25
Control Test		0.080	23:25
Air Blank		0.000	23:26
Subject Sample	# 1	0.082	23:26
Breath Volume		1.625	23:26
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.084	23:30
Breath Volume		1.101	23:30
Air Blank		0.000	23:30
Control Test		0.080	23:31
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 06/25/2026 06:00

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KEENEY,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 05/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MANDRICK

DATE/TIME OF TEST: 06/25/2026 03:52:03

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.202	03:47
Breath Volume		1.269	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.203	03:50
Breath Volume		1.671	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:52

Data Download Date/Time: 06/25/2026 06:00

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,B,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BRENNAN,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: OBORN

DATE/TIME OF TEST: 06/26/2026 01:30:45

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:28	
Control Test		0.078	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:29	Subject Test Refused
Breath Volume		0.000	01:29	
Air Blank		0.000	01:29	
Control Test		0.079	01:30	
Air Blank		0.000	01:30	
Diagnostics Check		OK	01:30	

Data Download Date/Time: 07/23/2026 23:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MAGALSKI,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 06/26/2026 02:06:00

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.080	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank		0.000	02:05	
Control Test		0.079	02:05	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 07/23/2026 23:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUNNINGHAM,B,T
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BUTLER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: BUGBEE

DATE/TIME OF TEST: 06/26/2026 02:47:50

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:37	
Control Test		0.080	02:37	
Air Blank		0.000	02:38	
Subject Sample	# 1	0.194	02:41	Volume Not Met
Breath Volume		0.097	02:41	
Air Blank		0.000	02:42	
Air Blank		0.000	02:43	
Subject Sample	# 2		02:46	Subject Test Refused
Breath Volume		0.398	02:46	
Air Blank		0.000	02:46	
Control Test		0.081	02:47	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 07/23/2026 23:48

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NOAH,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: WADE

DATE/TIME OF TEST: 06/28/2026 00:36:15

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:23	
Control Test		0.081	00:24	
Air Blank		0.000	00:24	
Subject Sample	# 1	0.172	00:27	Volume Not Met
Breath Volume		1.031	00:27	
Air Blank		0.000	00:28	
Air Blank		0.000	00:30	
Subject Sample	# 2	0.210	00:31	
Breath Volume		2.519	00:31	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 3	0.199	00:34	
Breath Volume		2.191	00:34	
Air Blank		0.000	00:35	
Control Test		0.080	00:35	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 07/23/2026 23:48

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,NOAH,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: LYDON,A
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 06/28/2026 18:44:10

Diagnostics Check		OK	18:37
Air Blank		0.000	18:37
Control Test		0.080	18:38
Air Blank		0.000	18:38
Subject Sample	# 1	0.143	18:39
Breath Volume		2.121	18:39
Air Blank		0.000	18:40
Air Blank		0.000	18:42
Subject Sample	# 2	0.143	18:42
Breath Volume		1.746	18:42
Air Blank		0.000	18:43
Control Test		0.080	18:43
Air Blank		0.000	18:44
Diagnostics Check		OK	18:44

Data Download Date/Time: 07/23/2026 23:48

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PECK,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/29/2026 21:36:27

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.081	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.145	21:32
Breath Volume		1.136	21:32
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.141	21:34
Breath Volume		1.160	21:34
Air Blank		0.000	21:35
Control Test		0.081	21:35
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 07/23/2026 23:48

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTRANA,TREVOR,J
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: KEENEY,T
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26-4437
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 06/30/2026 02:56:00

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.123	02:51
Breath Volume		2.070	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.122	02:54
Breath Volume		1.531	02:54
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:56

Data Download Date/Time: 07/23/2026 23:48

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIER,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BARBEE

DATE/TIME OF TEST: 06/01/2026 03:11:26

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.267	03:07
Breath Volume		1.398	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.264	03:09
Breath Volume		1.230	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SOLLILARO,MARIA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:41

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/02/2026 10:18:30

Diagnostics Check		OK	10:12
Air Blank		0.000	10:12
Control Test		0.080	10:13
Air Blank		0.000	10:13
Subject Sample	# 1	0.010	10:14
Breath Volume		2.144	10:14
Air Blank		0.000	10:14
Air Blank		0.000	10:16
Subject Sample	# 2	0.009	10:17
Breath Volume		2.230	10:17
Air Blank		0.000	10:17
Control Test		0.080	10:17
Air Blank		0.000	10:18
Diagnostics Check		OK	10:18

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/02/2026 21:44:38

Diagnostics Check		OK	21:33	
Air Blank		0.000	21:33	
Control Test		0.080	21:34	
Air Blank		0.000	21:34	
Subject Sample	# 1		21:37	No Sample Provided
Breath Volume		0.000	21:37	
Air Blank		0.000	21:38	
Air Blank		0.000	21:40	
Subject Sample	# 2		21:43	No Sample Provided
Breath Volume		0.000	21:43	
Air Blank		0.000	21:43	
Control Test		0.080	21:44	
Air Blank		0.000	21:44	
Diagnostics Check		OK	21:44	

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ROMAN JR

DATE/TIME OF TEST: 06/03/2026 03:48:25

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.080	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.199	03:44
Breath Volume		3.363	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.200	03:46
Breath Volume		3.222	03:46
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NARVAEZ,VIRNA
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 06/04/2026 03:46:32

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:35	
Control Test		0.080	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1		03:39	No Sample Provided
Breath Volume		0.000	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2		03:45	No Sample Provided
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Control Test		0.080	03:45	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RAO,RAVI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: COLEY

DATE/TIME OF TEST: 06/05/2026 03:39:26

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.269	03:34
Breath Volume		1.789	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:36
Subject Sample	# 2	0.261	03:37
Breath Volume		1.855	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OUTLAW,ADAM
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BASSHAM

DATE/TIME OF TEST: 06/05/2026 22:06:42

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.080	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1	0.087	21:57	Volume Not Met
Breath Volume		0.160	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	21:59	
Subject Sample	# 2	0.096	22:02	
Breath Volume		1.375	22:02	
Air Blank		0.000	22:02	
Air Blank		0.000	22:04	
Subject Sample	# 3	0.094	22:05	
Breath Volume		1.496	22:05	
Air Blank		0.000	22:05	
Control Test		0.079	22:06	
Air Blank		0.000	22:06	
Diagnostics Check		OK	22:06	

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONCION,CHRISTIAN,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MONCION,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PENAHERRERA

DATE/TIME OF TEST: 06/06/2026 00:46:24

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.149	00:41
Breath Volume		1.214	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.140	00:44
Breath Volume		1.113	00:44
Air Blank		0.000	00:45
Control Test		0.077	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 06/06/2026 03:02:46

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.000	02:58
Breath Volume		2.726	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.000	03:01
Breath Volume		2.132	03:01
Air Blank		0.000	03:01
Control Test		0.080	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: GUTIERREZ GOMEZ

DATE/TIME OF TEST: 06/06/2026 03:45:00

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.080	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.198	03:40
Breath Volume		1.871	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.201	03:43
Breath Volume		1.425	03:43
Air Blank		0.000	03:44
Control Test		0.077	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:45

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOSE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 06/06/2026 03:52:29

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:48	
Control Test		0.079	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:51	RFI Detect
Breath Volume		0.000	03:51	
Air Blank		0.000	03:51	
Control Test		0.080	03:51	
Air Blank		0.000	03:52	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOSE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 06/06/2026 04:15:30

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:04	
Control Test		0.081	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:08	No Sample Provided
Breath Volume		0.000	04:08	
Air Blank		0.000	04:09	
Air Blank		0.000	04:11	
Subject Sample	# 2		04:14	No Sample Provided
Breath Volume		0.000	04:14	
Air Blank		0.000	04:14	
Control Test		0.081	04:14	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 06/17/2026 15:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TEFOE,JUSTIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ESCOBAR BRAVO

DATE/TIME OF TEST: 06/06/2026 04:51:13

Diagnostics Check		OK	04:43
Air Blank		0.000	04:44
Control Test		0.078	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.184	04:46
Breath Volume		2.050	04:46
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.186	04:49
Breath Volume		2.093	04:49
Air Blank		0.000	04:50
Control Test		0.079	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: CASANOVA

DATE/TIME OF TEST: 06/06/2026 07:55:59

Diagnostics Check		OK	07:48
Air Blank		0.000	07:49
Control Test		0.080	07:49
Air Blank		0.000	07:50
Subject Sample	# 1	0.152	07:50
Breath Volume		2.066	07:50
Air Blank		0.000	07:51
Air Blank		0.000	07:53
Subject Sample	# 2	0.150	07:54
Breath Volume		1.785	07:54
Air Blank		0.000	07:55
Control Test		0.080	07:55
Air Blank		0.000	07:55
Diagnostics Check		OK	07:55

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: BRAUNSTEIN

DATE/TIME OF TEST: 06/06/2026 16:36:26

Diagnostics Check		OK	16:29
Air Blank		0.000	16:30
Control Test		0.080	16:30
Air Blank		0.000	16:30
Subject Sample	# 1	0.179	16:31
Breath Volume		1.199	16:31
Air Blank		0.000	16:31
Air Blank		0.000	16:33
Subject Sample	# 2	0.192	16:34
Breath Volume		1.609	16:34
Air Blank		0.000	16:35
Control Test		0.080	16:35
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: AILOR

DATE/TIME OF TEST: 06/06/2026 22:22:15

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.081	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.200	22:17
Breath Volume		1.535	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.199	22:20
Breath Volume		1.945	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: COLON,JOSE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: MALDONADO PEREZ

DATE/TIME OF TEST: 06/07/2026 00:24:45

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.077	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.204	00:20
Breath Volume		2.703	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.206	00:23
Breath Volume		2.265	00:23
Air Blank		0.000	00:23
Control Test		0.079	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 06/17/2026 15:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GARCIA,LUIS
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 06/07/2026 00:55:08

Diagnostics Check		OK	00:40	
Air Blank		0.000	00:40	
Control Test		0.080	00:40	
Air Blank		0.000	00:41	
Subject Sample	# 1	0.273	00:44	Volume Not Met
Breath Volume		0.328	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 2	0.260	00:48	No .020 agreement
Breath Volume		2.519	00:48	
Air Blank		0.000	00:48	
Air Blank		0.000	00:50	
Subject Sample	# 3	0.285	00:53	No .020 agreement
Breath Volume		1.910	00:53	
Air Blank		0.000	00:54	
Control Test		0.077	00:54	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,BENALY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: JASMIN

DATE/TIME OF TEST: 06/07/2026 05:26:11

Diagnostics Check		OK	05:23	
Air Blank		0.000	05:23	
Control Test		0.080	05:24	
Air Blank		0.000	05:24	
Subject Sample	# 1		05:24	RFI Detect
Breath Volume		0.000	05:24	
Air Blank		0.000	05:25	
Control Test		0.081	05:25	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,BENALY,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: JASMIN

DATE/TIME OF TEST: 06/07/2026 05:44:10

Diagnostics Check		OK	05:29	
Air Blank		0.000	05:29	
Control Test		0.080	05:29	
Air Blank		0.000	05:30	
Subject Sample	# 1	0.098	05:33	Volume Not Met
Breath Volume		0.074	05:33	
Air Blank		0.000	05:34	
Air Blank		0.000	05:35	
Subject Sample	# 2	0.119	05:38	
Breath Volume		2.179	05:38	
Air Blank		0.000	05:39	
Air Blank		0.000	05:40	
Subject Sample	# 3	0.111	05:42	
Breath Volume		1.183	05:42	
Air Blank		0.000	05:43	
Control Test		0.078	05:43	
Air Blank		0.000	05:44	
Diagnostics Check		OK	05:44	

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 06/07/2026 10:42:41

Diagnostics Check		OK	10:36
Air Blank		0.000	10:36
Control Test		0.080	10:36
Air Blank		0.000	10:37
Subject Sample	# 1	0.146	10:38
Breath Volume		2.582	10:38
Air Blank		0.000	10:38
Air Blank		0.000	10:40
Subject Sample	# 2	0.143	10:41
Breath Volume		1.488	10:41
Air Blank		0.000	10:41
Control Test		0.079	10:42
Air Blank		0.000	10:42
Diagnostics Check		OK	10:42

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLASER,STEVEN
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 06/07/2026 22:31:32

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.080	22:25
Air Blank		0.000	22:26
Subject Sample	# 1	0.095	22:27
Breath Volume		2.039	22:27
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.096	22:30
Breath Volume		1.976	22:30
Air Blank		0.000	22:30
Control Test		0.080	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,DANIEL,N
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: HICKS,DANIEL
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:01

SUBJECT'S LAST NAME: SALAS

DATE/TIME OF TEST: 06/08/2026 11:48:42

Diagnostics Check		OK	11:45	
Air Blank		0.000	11:46	
Control Test		0.079	11:46	
Air Blank		0.000	11:47	
Subject Sample	# 1		11:47	Subject Test Refused
Breath Volume		0.000	11:47	
Air Blank		0.000	11:47	
Control Test		0.079	11:48	
Air Blank		0.000	11:48	
Diagnostics Check		OK	11:48	

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,JACOB,I
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 06/09/2026 19:43:10

Diagnostics Check		OK	19:34
Air Blank		0.000	19:35
Control Test		0.080	19:35
Air Blank		0.000	19:36
Subject Sample	# 1	0.185	19:37
Breath Volume		1.566	19:37
Air Blank		0.000	19:37
Air Blank		0.000	19:39
Subject Sample	# 2	0.193	19:41
Breath Volume		2.519	19:41
Air Blank		0.000	19:42
Control Test		0.079	19:42
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,EDGAR,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: APONTE,EDGAR
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/10/2026 15:22:12

Diagnostics Check		OK	15:13
Air Blank		0.000	15:14
Control Test		0.080	15:14
Air Blank		0.000	15:14
Subject Sample	# 1	0.130	15:16
Breath Volume		1.414	15:16
Air Blank		0.000	15:16
Air Blank		0.000	15:18
Subject Sample	# 2	0.127	15:20
Breath Volume		1.917	15:20
Air Blank		0.000	15:21
Control Test		0.077	15:21
Air Blank		0.000	15:22
Diagnostics Check		OK	15:22

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PEACOCK

DATE/TIME OF TEST: 06/11/2026 01:08:21

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.079	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.088	01:03
Breath Volume		1.277	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:05
Subject Sample	# 2	0.091	01:06
Breath Volume		2.007	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: OLIVERA

DATE/TIME OF TEST: 06/11/2026 20:57:25

Diagnostics Check		OK	20:46	
Air Blank		0.000	20:46	
Control Test		0.080	20:46	
Air Blank		0.000	20:47	
Subject Sample	# 1		20:50	No Sample Provided
Breath Volume		0.000	20:50	
Air Blank		0.000	20:50	
Air Blank		0.000	20:52	
Subject Sample	# 2		20:56	No Sample Provided
Breath Volume		0.000	20:56	
Air Blank		0.000	20:56	
Control Test		0.080	20:56	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:57	

Data Download Date/Time: 06/17/2026 15:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIREBAUGH,JESSICA,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FIREBAUGH,JESSICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: SPROAT

DATE/TIME OF TEST: 06/12/2026 16:38:58

Diagnostics Check		OK	16:26	
Air Blank		0.000	16:27	
Control Test		0.080	16:27	
Air Blank		0.000	16:27	
Subject Sample	# 1		16:30	No Sample Provided
Breath Volume		0.000	16:30	
Air Blank		0.000	16:31	
Air Blank		0.000	16:33	
Subject Sample	# 2	0.077	16:34	
Breath Volume		3.722	16:34	
Air Blank		0.000	16:35	
Air Blank		0.000	16:37	
Subject Sample	# 3	0.076	16:37	
Breath Volume		1.992	16:37	
Air Blank		0.000	16:38	
Control Test		0.080	16:38	
Air Blank		0.000	16:38	
Diagnostics Check		OK	16:38	

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,JACOB,I
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MAISONET,JIMMY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SWANGER

DATE/TIME OF TEST: 06/13/2026 00:04:20

Diagnostics Check		OK	23:56
Air Blank		0.000	23:57
Control Test		0.081	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:59
Breath Volume		2.007	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.000	00:02
Breath Volume		1.664	00:02
Air Blank		0.000	00:03
Control Test		0.080	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SALINAS,JOSE,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: COLON,GUEMIL
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: WENDKOS JACKSON

DATE/TIME OF TEST: 06/13/2026 01:08:53

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.000	01:04
Breath Volume		2.097	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.000	01:07
Breath Volume		1.843	01:07
Air Blank		0.000	01:07
Control Test		0.080	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,JACOB,I
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 06/13/2026 04:09:07

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.188	04:04
Breath Volume		1.757	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.187	04:07
Breath Volume		1.316	04:07
Air Blank		0.000	04:08
Control Test		0.077	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONEY,KELSEY,N
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MERCER,DANIEL
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: BUMPUS

DATE/TIME OF TEST: 06/13/2026 23:51:50

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.079	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.242	23:47
Breath Volume		2.027	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:49
Subject Sample	# 2	0.235	23:50
Breath Volume		2.285	23:50
Air Blank		0.000	23:50
Control Test		0.078	23:51
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,JACOB,I
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ELLIOTT,JACOB
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: LEIGHLY

DATE/TIME OF TEST: 06/14/2026 03:36:58

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.092	03:32
Breath Volume		2.671	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.091	03:35
Breath Volume		2.539	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MARTIN,SAMUEL
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 06/14/2026 18:17:34

Diagnostics Check		OK	18:06	
Air Blank		0.000	18:06	
Control Test		0.081	18:06	
Air Blank		0.000	18:07	
Subject Sample	# 1		18:10	No Sample Provided
Breath Volume		0.000	18:10	
Air Blank		0.000	18:11	
Air Blank		0.000	18:13	
Subject Sample	# 2		18:16	No Sample Provided
Breath Volume		0.000	18:16	
Air Blank		0.000	18:16	
Control Test		0.080	18:16	
Air Blank		0.000	18:17	
Diagnostics Check		OK	18:17	

Data Download Date/Time: 06/17/2026 15:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEROS,JUAN,P
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: PINEROS,JUAN
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PHILIPSON

DATE/TIME OF TEST: 06/02/2026 01:52:44

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.080	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.169	01:47
Breath Volume		2.062	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.160	01:51
Breath Volume		2.531	01:51
Air Blank		0.000	01:51
Control Test		0.080	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 06/29/2026 16:28

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORREA,MATTHEW,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: SIBOLE,BREANNA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 14:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: FRIEDOPFER

DATE/TIME OF TEST: 06/27/2026 15:05:54

Diagnostics Check		OK	15:01	
Air Blank		0.000	15:01	
Control Test		0.080	15:02	
Air Blank		0.000	15:02	
Subject Sample	# 1		15:05	Sequence Aborted
Breath Volume		0.000	15:05	
Air Blank		0.000	15:05	

Data Download Date/Time: 06/29/2026 16:29

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/10/2026 03:06:07

Diagnostics Check	OK	03:04	
Air Blank	0.000	03:05	
Control Test	0.077	03:05	
Air Blank		03:05	RFI Detect
Air Blank	0.000	03:06	

Data Download Date/Time: 06/30/2026 09:15

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,ALISHA,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WEST,ALISHA
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 06/10/2026 03:09:52

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.077	03:09	
Air Blank		03:09	RFI Detect
Air Blank	0.000	03:09	

Data Download Date/Time: 06/30/2026 09:15

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: SALGUEIRO,GUSTAVO
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: FUENTES RAMIREZ

DATE/TIME OF TEST: 06/01/2026 01:23:37

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.078	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.266	01:18
Breath Volume		1.941	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.269	01:22
Breath Volume		2.328	01:22
Air Blank		0.000	01:22
Control Test		0.079	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 06/10/2026 10:17

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: RADUTIU,TOMMY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/10/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: SOSA POLANCO

DATE/TIME OF TEST: 06/13/2026 03:01:43

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.123	02:56
Breath Volume		1.281	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.112	03:00
Breath Volume		2.253	03:00
Air Blank		0.000	03:00
Control Test		0.078	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 07/01/2026 10:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPANA,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: WILSON,ALLEN
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 06/01/2026 19:52:55

Diagnostics Check		OK	19:45
Air Blank		0.000	19:46
Control Test		0.078	19:46
Air Blank		0.000	19:47
Subject Sample	# 1	0.157	19:47
Breath Volume		1.750	19:47
Air Blank		0.000	19:48
Air Blank		0.000	19:50
Subject Sample	# 2	0.151	19:51
Breath Volume		1.554	19:51
Air Blank		0.000	19:52
Control Test		0.079	19:52
Air Blank		0.000	19:52
Diagnostics Check		OK	19:52

Data Download Date/Time: 06/12/2026 08:28

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TYLER,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 05/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 06/02/2026 02:13:09

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.074	02:08
Breath Volume		2.261	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.071	02:11
Breath Volume		2.000	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 06/12/2026 08:28

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GASSEN,STEVE,
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: BRISENO,N
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 05/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MOUGROS

DATE/TIME OF TEST: 06/07/2026 02:17:36

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.077	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.140	02:12
Breath Volume		4.433	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.125	02:16
Breath Volume		2.050	02:16
Air Blank		0.000	02:16
Control Test		0.076	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 06/10/2026 08:28

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENNAN,CONNOR,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 06/01/2026 21:33:20

Diagnostics Check		OK	21:29	
Air Blank		0.000	21:30	
Control Test		0.080	21:30	
Air Blank		0.000	21:31	
Subject Sample	# 1		21:31	Subject Test Refused
Breath Volume		0.000	21:31	
Air Blank		0.000	21:32	
Control Test		0.079	21:32	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUIRGUIS,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 06:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:29

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 06/02/2026 08:10:46

Diagnostics Check		OK	07:57	
Air Blank		0.000	07:57	
Control Test		0.080	07:57	
Air Blank		0.000	07:58	
Subject Sample	# 1	0.049	08:01	Volume Not Met
Breath Volume		0.421	08:01	
Air Blank		0.000	08:02	
Air Blank		0.000	08:03	
Subject Sample	# 2	0.055	08:06	
Breath Volume		1.640	08:06	
Air Blank		0.000	08:06	
Air Blank		0.000	08:08	
Subject Sample	# 3	0.057	08:09	
Breath Volume		1.230	08:09	
Air Blank		0.000	08:09	
Control Test		0.079	08:10	
Air Blank		0.000	08:10	
Diagnostics Check		OK	08:10	

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUNDH,DEP,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SWIER,DEP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:51

SUBJECT'S LAST NAME: PASTIAN

DATE/TIME OF TEST: 06/02/2026 17:17:59

Diagnostics Check		OK	17:11
Air Blank		0.000	17:12
Control Test		0.080	17:12
Air Blank		0.000	17:13
Subject Sample	# 1	0.072	17:13
Breath Volume		1.539	17:13
Air Blank		0.000	17:14
Air Blank		0.000	17:16
Subject Sample	# 2	0.070	17:16
Breath Volume		1.574	17:16
Air Blank		0.000	17:17
Control Test		0.079	17:17
Air Blank		0.000	17:17
Diagnostics Check		OK	17:17

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEDLIE,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: ENRIQUE DOLORES

DATE/TIME OF TEST: 06/03/2026 03:21:28

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.080	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.204	03:16
Breath Volume		4.160	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.208	03:19
Breath Volume		2.644	03:19
Air Blank		0.000	03:20
Control Test		0.079	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ CRUZ,JORDAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: VAZQUEZ CRUZ,JORDAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: MEJIA MELENDEZ

DATE/TIME OF TEST: 06/03/2026 06:02:09

Diagnostics Check		OK	05:55
Air Blank		0.000	05:55
Control Test		0.081	05:56
Air Blank		0.000	05:56
Subject Sample	# 1	0.171	05:57
Breath Volume		1.121	05:57
Air Blank		0.000	05:58
Air Blank		0.000	06:00
Subject Sample	# 2	0.172	06:00
Breath Volume		1.957	06:00
Air Blank		0.000	06:01
Control Test		0.079	06:01
Air Blank		0.000	06:02
Diagnostics Check		OK	06:02

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEWS,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: NORTHUP

DATE/TIME OF TEST: 06/05/2026 00:59:13

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.080	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 1		00:57	Subject Test Refused
Breath Volume		0.000	00:57	
Air Blank		0.000	00:58	
Control Test		0.080	00:58	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ORTIZ,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WOJTKIEWICZ

DATE/TIME OF TEST: 06/06/2026 21:47:38

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.080	21:41
Air Blank		0.000	21:41
Subject Sample	# 1	0.207	21:42
Breath Volume		1.855	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.199	21:46
Breath Volume		1.726	21:46
Air Blank		0.000	21:46
Control Test		0.079	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUHLMAIER,KORY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCOTT,DANDRE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: WOOL

DATE/TIME OF TEST: 06/08/2026 23:04:43

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.080	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.158	23:00
Breath Volume		2.175	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:02
Subject Sample	# 2	0.158	23:03
Breath Volume		2.500	23:03
Air Blank		0.000	23:03
Control Test		0.078	23:04
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: PARTHENAIS

DATE/TIME OF TEST: 06/09/2026 01:03:59

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.119	00:59
Breath Volume		1.843	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.115	01:02
Breath Volume		1.593	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO',JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MESSENGER,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: XAYPANYA

DATE/TIME OF TEST: 06/10/2026 00:31:49

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.114	00:26
Breath Volume		1.863	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.105	00:30
Breath Volume		1.449	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 06/17/2026 08:53

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,S,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RUSHAKOFF,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: LAMASKO

DATE/TIME OF TEST: 06/11/2026 17:33:39

Diagnostics Check		OK	17:30	
Air Blank		0.000	17:31	
Control Test		0.080	17:31	
Air Blank		0.000	17:31	
Subject Sample	# 1		17:32	Subject Test Refused
Breath Volume		0.000	17:32	
Air Blank		0.000	17:32	
Control Test		0.080	17:33	
Air Blank		0.000	17:33	
Diagnostics Check		OK	17:33	

Data Download Date/Time: 06/17/2026 08:54

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLLOWELL,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: FORMOSO

DATE/TIME OF TEST: 06/12/2026 00:06:30

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.180	00:02
Breath Volume		1.160	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.183	00:05
Breath Volume		1.343	00:05
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 06/17/2026 08:54

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOATENG,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: DEMPKEY

DATE/TIME OF TEST: 06/13/2026 04:14:03

Diagnostics Check		OK	04:07
Air Blank		0.000	04:07
Control Test		0.080	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.155	04:09
Breath Volume		1.375	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.166	04:12
Breath Volume		1.402	04:12
Air Blank		0.000	04:13
Control Test		0.079	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:14

Data Download Date/Time: 06/17/2026 08:54

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: DESALVIA

DATE/TIME OF TEST: 06/13/2026 18:02:17

Diagnostics Check		OK	17:59	
Air Blank		0.000	17:59	
Control Test		0.081	18:00	
Air Blank		0.000	18:00	
Subject Sample	# 1		18:00	Subject Test Refused
Breath Volume		0.000	18:00	
Air Blank		0.000	18:01	
Control Test		0.080	18:01	
Air Blank		0.000	18:02	
Diagnostics Check		OK	18:02	

Data Download Date/Time: 06/17/2026 08:54

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NEWTON,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CURVAN

DATE/TIME OF TEST: 06/14/2026 01:05:26

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:03	
Control Test		0.080	01:03	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.000	01:04	
Air Blank		0.000	01:04	
Control Test		0.080	01:04	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 06/17/2026 08:54

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LONGANO,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LACEY

DATE/TIME OF TEST: 06/16/2026 01:35:44

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:25
Subject Sample	# 1	0.186	01:27
Breath Volume		1.609	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.210	01:30
Breath Volume		1.785	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 3	0.193	01:34
Breath Volume		2.453	01:34
Air Blank		0.000	01:34
Control Test		0.079	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 06/17/2026 08:54

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARRIZA,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PUFF

DATE/TIME OF TEST: 06/16/2026 04:01:31

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.079	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 06/17/2026 08:54

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TAVERNARI,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 06/16/2026 23:42:09

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.080	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.220	23:37
Breath Volume		1.757	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.240	23:40
Breath Volume		2.476	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 06/17/2026 08:54

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEDLIE,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 06/17/2026 00:17:07

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.079	00:10
Air Blank		0.000	00:10
Subject Sample	# 1	0.095	00:12
Breath Volume		1.246	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.106	00:15
Breath Volume		1.531	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 06/17/2026 08:54

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ARRIZA,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: EAGAN

DATE/TIME OF TEST: 06/17/2026 00:55:47

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.079	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:54	
Control Test		0.080	00:55	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 06/17/2026 08:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CONNELL,CHASON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 08:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/17/2026 09:27:17

Diagnostics Check		OK	09:20
Air Blank		0.000	09:21
Control Test		0.080	09:21
Air Blank		0.000	09:22
Subject Sample	# 1	0.000	09:22
Breath Volume		1.781	09:22
Air Blank		0.000	09:23
Air Blank		0.000	09:25
Subject Sample	# 2	0.000	09:25
Breath Volume		1.625	09:25
Air Blank		0.000	09:26
Control Test		0.080	09:26
Air Blank		0.000	09:27
Diagnostics Check		OK	09:27

Data Download Date/Time: 07/29/2026 08:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIS,JUSTIN,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BUIS,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: LOTOV

DATE/TIME OF TEST: 06/19/2026 18:06:36

Diagnostics Check		OK	18:03	
Air Blank		0.000	18:04	
Control Test		0.080	18:04	
Air Blank		0.000	18:05	
Subject Sample	# 1		18:05	Subject Test Refused
Breath Volume		0.000	18:05	
Air Blank		0.000	18:05	
Control Test		0.080	18:06	
Air Blank		0.000	18:06	
Diagnostics Check		OK	18:06	

Data Download Date/Time: 07/29/2026 08:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SEVOSTJANOV,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: COLLINS III

DATE/TIME OF TEST: 06/20/2026 00:07:03

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:04	
Control Test		0.081	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:06	
Control Test		0.080	00:06	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 07/29/2026 08:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TAVERNARI,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MCDEVITT

DATE/TIME OF TEST: 06/20/2026 21:39:36

Diagnostics Check		OK	21:36	
Air Blank		0.000	21:37	
Control Test		0.080	21:37	
Air Blank		0.000	21:37	
Subject Sample	# 1		21:38	Subject Test Refused
Breath Volume		0.000	21:38	
Air Blank		0.000	21:38	
Control Test		0.080	21:39	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 07/29/2026 08:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JUNE,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: OLIVETO

DATE/TIME OF TEST: 06/21/2026 02:42:41

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:40	
Control Test		0.080	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.080	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 07/29/2026 08:49

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOLLOWELL,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 06/21/2026 05:11:28

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.081	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.055	05:06
Breath Volume		1.375	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.057	05:09
Breath Volume		2.292	05:09
Air Blank		0.000	05:10
Control Test		0.081	05:10
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 07/29/2026 08:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,DEPUTY,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHANFISH,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PURCELL

DATE/TIME OF TEST: 06/22/2026 04:05:10

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:04	
Control Test		0.080	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 07/29/2026 08:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOUGLAS,MARK
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/22/2026 15:32:11

Diagnostics Check	OK	15:30	
Air Blank	0.000	15:31	
Control Test	0.080	15:31	
Air Blank		15:31	RFI Detect
Air Blank		15:32	RFI Detect

Data Download Date/Time: 07/29/2026 08:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOUGLAS,MARK
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:10

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/22/2026 15:41:51

Diagnostics Check		OK	15:35
Air Blank		0.000	15:35
Control Test		0.080	15:35
Air Blank		0.000	15:36
Subject Sample	# 1	0.182	15:37
Breath Volume		1.554	15:37
Air Blank		0.000	15:37
Air Blank		0.000	15:39
Subject Sample	# 2	0.186	15:40
Breath Volume		1.394	15:40
Air Blank		0.000	15:40
Control Test		0.080	15:41
Air Blank		0.000	15:41
Diagnostics Check		OK	15:41

Data Download Date/Time: 07/29/2026 08:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAZIME,S,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAZIME,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 14:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: WICKMAN

DATE/TIME OF TEST: 06/24/2026 15:48:05

Diagnostics Check		OK	15:41
Air Blank		0.000	15:42
Control Test		0.081	15:42
Air Blank		0.000	15:43
Subject Sample	# 1	0.275	15:43
Breath Volume		3.070	15:43
Air Blank		0.000	15:44
Air Blank		0.000	15:46
Subject Sample	# 2	0.267	15:46
Breath Volume		2.027	15:46
Air Blank		0.000	15:47
Control Test		0.080	15:47
Air Blank		0.000	15:48
Diagnostics Check		OK	15:48

Data Download Date/Time: 07/29/2026 08:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HAEDKE,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 06/25/2026 23:05:01

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:00	
Control Test		0.080	23:01	
Air Blank		0.000	23:01	
Subject Sample	# 1		23:03	Interferent Detect
Breath Volume		2.347	23:03	
Air Blank		0.000	23:04	
Control Test		0.080	23:04	
Air Blank		0.000	23:04	
Diagnostics Check		OK	23:05	

Data Download Date/Time: 07/29/2026 08:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/27/2026 07:39:50

Diagnostics Check		OK	07:30
Air Blank		0.000	07:31
Control Test		0.080	07:31
Air Blank		0.000	07:32
Subject Sample	# 1	0.164	07:35
Breath Volume		1.863	07:35
Air Blank		0.000	07:35
Air Blank		0.000	07:37
Subject Sample	# 2	0.166	07:38
Breath Volume		1.242	07:38
Air Blank		0.000	07:38
Control Test		0.080	07:39
Air Blank		0.000	07:39
Diagnostics Check		OK	07:39

Data Download Date/Time: 07/29/2026 08:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUHLMAIER,KORY,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MASON,MATTHEW
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 06/27/2026 18:25:29

Diagnostics Check		OK	18:22	
Air Blank		0.000	18:23	
Control Test		0.079	18:23	
Air Blank		0.000	18:23	
Subject Sample	# 1		18:24	Subject Test Refused
Breath Volume		0.000	18:24	
Air Blank		0.000	18:24	
Control Test		0.079	18:24	
Air Blank		0.000	18:25	
Diagnostics Check		OK	18:25	

Data Download Date/Time: 07/29/2026 08:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: ASHWORTH

DATE/TIME OF TEST: 06/27/2026 20:11:00

Diagnostics Check		OK	20:07	
Air Blank		0.000	20:08	
Control Test		0.079	20:08	
Air Blank		0.000	20:09	
Subject Sample	# 1		20:09	Subject Test Refused
Breath Volume		0.000	20:09	
Air Blank		0.000	20:10	
Control Test		0.081	20:10	
Air Blank		0.000	20:10	
Diagnostics Check		OK	20:11	

Data Download Date/Time: 07/29/2026 08:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEWS,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: JETT

DATE/TIME OF TEST: 06/28/2026 00:05:19

Diagnostics Check		OK	23:49	
Air Blank		0.000	23:50	
Control Test		0.080	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 2		23:54	No Sample Provided
Breath Volume		0.000	23:54	
Air Blank		0.000	23:54	
Air Blank		0.000	23:56	
Subject Sample	# 3	0.164	23:59	
Breath Volume		1.464	23:59	
Air Blank		0.000	00:00	
Air Blank		0.000	00:02	
Subject Sample	# 1	0.152	00:03	
Breath Volume		1.839	00:03	
Air Blank		0.000	00:04	
Control Test		0.078	00:04	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 07/29/2026 08:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 07:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:59

SUBJECT'S LAST NAME: IACUONE

DATE/TIME OF TEST: 06/28/2026 08:25:17

Diagnostics Check		OK	08:21	
Air Blank		0.000	08:22	
Control Test		0.080	08:22	
Air Blank		0.000	08:23	
Subject Sample	# 1		08:23	Subject Test Refused
Breath Volume		0.000	08:23	
Air Blank		0.000	08:24	
Control Test		0.080	08:24	
Air Blank		0.000	08:25	
Diagnostics Check		OK	08:25	

Data Download Date/Time: 07/29/2026 08:50

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/29/2026 00:23:16

Diagnostics Check	OK	00:22	
Air Blank		00:22	RFI Detect
Air Blank	0.000	00:23	

Data Download Date/Time: 07/29/2026 08:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORO,JARED,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TORO,JARED
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG525301
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/29/2026 00:27:18

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.080	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:26	
Control Test		0.079	00:26	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 07/29/2026 08:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HYLTON,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: GAY,YASKARA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/1018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/05/2026 19:37:41

Diagnostics Check	OK	19:36	
Air Blank		19:37	RFI Detect
Air Blank	0.000	19:37	

Data Download Date/Time: 06/10/2026 10:11

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HYLTON,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: GAY,YASKARA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/1018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/05/2026 19:45:09

Diagnostics Check		OK	19:41	
Air Blank		0.000	19:42	
Control Test		0.079	19:42	
Air Blank		0.000	19:42	
Subject Sample	# 1		19:44	Sequence Aborted
Breath Volume		0.019	19:44	
Air Blank		0.000	19:45	

Data Download Date/Time: 06/10/2026 10:11

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001067

INSTRUMENT REGISTERED TO: CITY OF MIRAMAR POL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HYLTON,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Miramar PD
ARRESTING OFFICER: GAY,YASKARA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/1018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/05/2026 19:59:04

Diagnostics Check		OK	19:47	
Air Blank		0.000	19:48	
Control Test		0.078	19:48	
Air Blank		0.000	19:48	
Subject Sample	# 1	0.134	19:51	Volume Not Met
Breath Volume		0.363	19:51	
Air Blank		0.000	19:52	
Air Blank		0.000	19:54	
Subject Sample	# 2	0.111	19:57	Volume Not Met
Breath Volume		0.015	19:57	
Air Blank		0.000	19:58	
Control Test		0.078	19:58	
Air Blank		0.000	19:58	
Diagnostics Check		OK	19:59	

Data Download Date/Time: 06/10/2026 10:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,IVAN,
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HERNANDEZ,IVAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/02/2026 00:04:04

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.081	23:58
Air Blank		0.000	23:58
Subject Sample	# 2	0.079	23:59
Breath Volume		4.300	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.074	00:02
Breath Volume		3.558	00:02
Air Blank		0.000	00:03
Control Test		0.080	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:04

Data Download Date/Time: 06/18/2026 14:07

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DEUSEN,COLE,C
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VAN DEUSEN,COLE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: BALDREE

DATE/TIME OF TEST: 06/08/2026 16:25:18

Diagnostics Check		OK	16:13	
Air Blank		0.000	16:14	
Control Test		0.081	16:14	
Air Blank		0.000	16:15	
Subject Sample	# 1		16:18	No Sample Provided
Breath Volume		0.000	16:18	
Air Blank		0.000	16:18	
Air Blank		0.000	16:20	
Subject Sample	# 2		16:23	No Sample Provided
Breath Volume		0.000	16:23	
Air Blank		0.000	16:24	
Control Test		0.082	16:24	
Air Blank		0.000	16:25	
Diagnostics Check		OK	16:25	

Data Download Date/Time: 06/18/2026 14:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KROLICK,ALEX,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: KROLICK,ALEX
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:42

SUBJECT'S LAST NAME: ZAPOTE-HAMMOD

DATE/TIME OF TEST: 06/08/2026 19:13:48

Diagnostics Check		OK	19:03
Air Blank		0.000	19:03
Control Test		0.081	19:04
Air Blank		0.000	19:04
Subject Sample	# 1	0.243	19:05
Breath Volume		1.179	19:05
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.220	19:08
Breath Volume		1.191	19:08
Air Blank		0.000	19:09
Air Blank		0.000	19:10
Subject Sample	# 3	0.224	19:12
Breath Volume		1.523	19:12
Air Blank		0.000	19:12
Control Test		0.081	19:13
Air Blank		0.000	19:13
Diagnostics Check		OK	19:13

Data Download Date/Time: 06/18/2026 14:07

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURCIO,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: CURCIO,ANTHONY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: GONZALEZ SERRANO

DATE/TIME OF TEST: 06/12/2026 00:21:44

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.081	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.199	00:16
Breath Volume		1.136	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.216	00:20
Breath Volume		1.746	00:20
Air Blank		0.000	00:20
Control Test		0.081	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 06/18/2026 14:07

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BACON

DATE/TIME OF TEST: 06/13/2026 02:24:39

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.081	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.086	02:19
Breath Volume		1.457	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.082	02:23
Breath Volume		1.250	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 06/18/2026 14:07

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN DEUSEN,COLE,C
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VAN DEUSEN,COLE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: LIGON

DATE/TIME OF TEST: 06/18/2026 00:38:33

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:27	
Control Test		0.081	00:27	
Air Blank		0.000	00:28	
Subject Sample	# 1		00:31	No Sample Provided
Breath Volume		0.000	00:31	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2		00:37	No Sample Provided
Breath Volume		0.000	00:37	
Air Blank		0.000	00:37	
Control Test		0.081	00:37	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 06/18/2026 14:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STYLES,LAUREN,N
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GIARDINO,CARL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 05/13/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 06/18/2026 07:13:55

Diagnostics Check		OK	07:05
Air Blank		0.000	07:05
Control Test		0.081	07:06
Air Blank		0.000	07:06
Subject Sample	# 1	0.062	07:08
Breath Volume		2.855	07:08
Air Blank		0.000	07:09
Air Blank		0.000	07:10
Subject Sample	# 2	0.052	07:12
Breath Volume		1.222	07:12
Air Blank		0.000	07:12
Control Test		0.081	07:13
Air Blank		0.000	07:13
Diagnostics Check		OK	07:13

Data Download Date/Time: 06/18/2026 14:07

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YEATES,SCOTT,W
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: YEATES,SCOTT
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: RIPA

DATE/TIME OF TEST: 06/19/2026 13:25:37

Diagnostics Check		OK	13:18
Air Blank		0.000	13:19
Control Test		0.081	13:19
Air Blank		0.000	13:20
Subject Sample	# 1	0.145	13:20
Breath Volume		2.070	13:20
Air Blank		0.000	13:21
Air Blank		0.000	13:23
Subject Sample	# 2	0.144	13:24
Breath Volume		1.796	13:24
Air Blank		0.000	13:24
Control Test		0.081	13:24
Air Blank		0.000	13:25
Diagnostics Check		OK	13:25

Data Download Date/Time: 07/17/2026 18:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 06/25/2026 17:37:30

Diagnostics Check		OK	17:30
Air Blank		0.000	17:31
Control Test		0.081	17:31
Air Blank		0.000	17:31
Subject Sample	# 1	0.000	17:32
Breath Volume		1.382	17:32
Air Blank		0.000	17:33
Air Blank		0.000	17:35
Subject Sample	# 2	0.000	17:36
Breath Volume		1.832	17:36
Air Blank		0.000	17:36
Control Test		0.081	17:36
Air Blank		0.000	17:37
Diagnostics Check		OK	17:37

Data Download Date/Time: 07/17/2026 18:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILES,OFFICER,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GILES,OFFICER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BELTRAN MARTINEZ

DATE/TIME OF TEST: 06/27/2026 21:50:47

Diagnostics Check		OK	21:37	
Air Blank		0.000	21:38	
Control Test		0.081	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1	0.223	21:39	
Breath Volume		1.695	21:39	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 2	0.203	21:45	Slope Not Level
Breath Volume		1.128	21:45	
Air Blank		0.000	21:46	
Air Blank		0.000	21:48	
Subject Sample	# 3	0.229	21:49	
Breath Volume		1.468	21:49	
Air Blank		0.000	21:49	
Control Test		0.080	21:50	
Air Blank		0.000	21:50	
Diagnostics Check		OK	21:50	

Data Download Date/Time: 07/17/2026 18:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,CONNER,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVIS,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 05/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 06/01/2026 19:20:25

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.081	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.245	19:15
Breath Volume		1.390	19:15
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.247	19:18
Breath Volume		1.628	19:18
Air Blank		0.000	19:19
Control Test		0.080	19:19
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 06/03/2026 13:21

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: ALVAREZ VAZQUEZ

DATE/TIME OF TEST: 06/05/2026 20:48:21

Diagnostics Check		OK	20:42
Air Blank		0.000	20:42
Control Test		0.081	20:42
Air Blank		0.000	20:43
Subject Sample	# 1	0.000	20:44
Breath Volume		2.347	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.000	20:46
Breath Volume		2.003	20:46
Air Blank		0.000	20:47
Control Test		0.080	20:47
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 06/11/2026 13:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,H,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 06/06/2026 02:39:01

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:38	
Control Test		0.081	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 06/11/2026 13:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 06/06/2026 05:44:22

Diagnostics Check		OK	05:41	
Air Blank		0.000	05:41	
Control Test		0.081	05:42	
Air Blank		0.000	05:42	
Subject Sample	# 1		05:42	Subject Test Refused
Breath Volume		0.000	05:42	
Air Blank		0.000	05:43	
Control Test		0.081	05:43	
Air Blank		0.000	05:44	
Diagnostics Check		OK	05:44	

Data Download Date/Time: 06/11/2026 13:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LAQUIDARA,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: EALEY

DATE/TIME OF TEST: 06/07/2026 01:14:53

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:12	
Control Test		0.081	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:13	
Control Test		0.081	01:14	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 06/11/2026 13:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RICH,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 06/07/2026 02:07:23

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.082	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 1	0.157	02:03	
Breath Volume		2.679	02:03	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2	0.155	02:06	
Breath Volume		2.371	02:06	
Air Blank			02:06	RFI Detect
Air Blank		0.000	02:07	

Data Download Date/Time: 06/11/2026 13:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RICH,T
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 06/07/2026 02:17:44

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.154	02:12
Breath Volume		2.003	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.166	02:16
Breath Volume		2.175	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:17
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 06/11/2026 13:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LUGO,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 06/07/2026 04:23:40

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.081	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.170	04:18
Breath Volume		2.175	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.175	04:21
Breath Volume		2.484	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 06/11/2026 13:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARTIN,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 06/07/2026 07:21:18

Diagnostics Check		OK	07:15
Air Blank		0.000	07:15
Control Test		0.081	07:15
Air Blank		0.000	07:16
Subject Sample	# 1	0.071	07:16
Breath Volume		3.152	07:16
Air Blank		0.000	07:17
Air Blank		0.000	07:19
Subject Sample	# 2	0.071	07:19
Breath Volume		3.527	07:19
Air Blank		0.000	07:20
Control Test		0.081	07:20
Air Blank		0.000	07:21
Diagnostics Check		OK	07:21

Data Download Date/Time: 06/11/2026 13:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORRIS,J,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRUBOR,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: BRAWNER

DATE/TIME OF TEST: 06/09/2026 01:00:14

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.082	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.124	00:55
Breath Volume		1.460	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.122	00:58
Breath Volume		1.574	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 06/11/2026 13:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,CORBYN,K
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36525080A4
DRY GAS EXPIRY DATE: 02/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: KATES

DATE/TIME OF TEST: 06/10/2026 03:16:15

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.082	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.238	03:08
Breath Volume		2.648	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.217	03:11
Breath Volume		2.835	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 3	0.230	03:14
Breath Volume		2.941	03:14
Air Blank		0.000	03:15
Control Test		0.078	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 06/11/2026 13:47

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,AUSTIN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:59

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 06/01/2026 15:35:22

Diagnostics Check		OK	15:29
Air Blank		0.000	15:29
Control Test		0.080	15:30
Air Blank		0.000	15:30
Subject Sample	# 1	0.000	15:31
Breath Volume		2.281	15:31
Air Blank		0.000	15:31
Air Blank		0.000	15:33
Subject Sample	# 2	0.000	15:33
Breath Volume		1.980	15:33
Air Blank		0.000	15:34
Control Test		0.080	15:34
Air Blank		0.000	15:35
Diagnostics Check		OK	15:35

Data Download Date/Time: 06/25/2026 09:29

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,OFFICER,Z
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BRAZELL

DATE/TIME OF TEST: 06/03/2026 01:28:35

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:25	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.079	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 06/25/2026 09:29

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHETEE,NICHOLAS,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KUSHNIR,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: ISRAELSON

DATE/TIME OF TEST: 06/08/2026 04:16:02

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.080	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.128	04:11
Breath Volume		2.343	04:11
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.128	04:14
Breath Volume		2.089	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:16

Data Download Date/Time: 06/25/2026 09:29

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,C,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KUSHNIR,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 06/09/2026 22:41:28

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.080	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.274	22:36
Breath Volume		4.816	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.268	22:39
Breath Volume		3.703	22:39
Air Blank		0.000	22:40
Control Test		0.078	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 06/25/2026 09:29

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: DANIELSON

DATE/TIME OF TEST: 06/13/2026 07:45:43

Diagnostics Check		OK	07:42	
Air Blank		0.000	07:43	
Control Test		0.079	07:43	
Air Blank		0.000	07:44	
Subject Sample	# 1		07:44	Subject Test Refused
Breath Volume		0.000	07:44	
Air Blank		0.000	07:44	
Control Test		0.079	07:45	
Air Blank		0.000	07:45	
Diagnostics Check		OK	07:45	

Data Download Date/Time: 06/25/2026 09:29

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 06/15/2026 19:41:28

Diagnostics Check		OK	19:38	
Air Blank		0.000	19:39	
Control Test		0.080	19:39	
Air Blank		0.000	19:39	
Subject Sample	# 1		19:40	Subject Test Refused
Breath Volume		0.000	19:40	
Air Blank		0.000	19:40	
Control Test		0.081	19:40	
Air Blank		0.000	19:41	
Diagnostics Check		OK	19:41	

Data Download Date/Time: 06/25/2026 09:29

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GODERICH,KRISTINA,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: GOLDMAN

DATE/TIME OF TEST: 06/16/2026 21:08:55

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.080	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.000	21:04
Breath Volume		1.910	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.000	21:07
Breath Volume		2.394	21:07
Air Blank		0.000	21:08
Control Test		0.081	21:08
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 06/25/2026 09:29

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORRIGAN,MATTHEW,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HORRIGAN,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 06/16/2026 21:38:32

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.080	21:29
Air Blank		0.000	21:30
Subject Sample	# 1	0.213	21:30
Breath Volume		5.625	21:30
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.173	21:33
Breath Volume		1.394	21:33
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 3	0.195	21:36
Breath Volume		4.320	21:36
Air Blank		0.000	21:37
Control Test		0.078	21:37
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 06/25/2026 09:30

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORRIGAN,MATTHEW,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HORRIGAN,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 06/18/2026 00:05:02

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:03	Subject Test Refused
Breath Volume		0.000	00:03	
Air Blank		0.000	00:04	
Control Test		0.080	00:04	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 06/25/2026 09:30

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORRIGAN,MATTHEW,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HORRIGAN,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: OROZCO RAMIREZ

DATE/TIME OF TEST: 06/18/2026 23:04:20

Diagnostics Check		OK	22:57
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:59
Subject Sample	# 1	0.160	22:59
Breath Volume		2.792	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.153	23:02
Breath Volume		2.851	23:02
Air Blank		0.000	23:03
Control Test		0.079	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 06/25/2026 09:30

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOBART,C,T
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FOXWORTH

DATE/TIME OF TEST: 06/19/2026 02:17:09

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.175	02:12
Breath Volume		2.390	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.173	02:15
Breath Volume		1.308	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 06/25/2026 09:30

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOWALSKI,A,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 06/19/2026 02:54:37

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:43	
Control Test		0.079	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 1	0.121	02:47	Volume Not Met
Breath Volume		0.046	02:47	
Air Blank		0.000	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 2	0.147	02:52	Volume Not Met
Breath Volume		0.015	02:52	
Air Blank		0.000	02:53	
Control Test		0.079	02:54	
Air Blank		0.000	02:54	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 06/25/2026 09:30

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,AUSTIN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: BARGER

DATE/TIME OF TEST: 06/20/2026 01:00:42

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.132	00:56
Breath Volume		4.093	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.117	00:59
Breath Volume		1.441	00:59
Air Blank		0.000	00:59
Control Test		0.079	01:00
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 06/25/2026 09:30

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: MINEAR

DATE/TIME OF TEST: 06/24/2026 01:28:59

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.000	01:24
Breath Volume		2.105	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.000	01:27
Breath Volume		1.910	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 06/25/2026 09:30

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,OFFICER,Z
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KUSHNIR,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SINGLETARY

DATE/TIME OF TEST: 06/26/2026 04:54:35

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:52	
Control Test		0.080	04:52	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:53	Subject Test Refused
Breath Volume		0.000	04:53	
Air Blank		0.000	04:53	
Control Test		0.079	04:53	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 07/23/2026 08:43

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LENAHAN,RYAN,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KUSHNIR,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: RIGGINS

DATE/TIME OF TEST: 06/27/2026 00:04:31

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.080	23:54
Air Blank		0.000	23:55
Subject Sample	# 2	0.000	23:57
Breath Volume		1.160	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.000	00:03
Breath Volume		1.234	00:03
Air Blank		0.000	00:03
Control Test		0.080	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 07/23/2026 08:43

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,W,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BISHOP,C
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:08

SUBJECT'S LAST NAME: JANNINO

DATE/TIME OF TEST: 06/27/2026 13:40:42

Diagnostics Check		OK	13:31
Air Blank		0.000	13:32
Control Test		0.080	13:32
Air Blank		0.000	13:32
Subject Sample	# 1	0.238	13:35
Breath Volume		1.363	13:35
Air Blank		0.000	13:36
Air Blank		0.000	13:38
Subject Sample	# 2	0.244	13:39
Breath Volume		1.167	13:39
Air Blank		0.000	13:39
Control Test		0.077	13:40
Air Blank		0.000	13:40
Diagnostics Check		OK	13:40

Data Download Date/Time: 07/23/2026 08:43

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31124080A1
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: PINI

DATE/TIME OF TEST: 06/27/2026 21:31:55

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:19	
Control Test		0.080	21:19	
Air Blank		0.000	21:19	
Subject Sample	# 1		21:23	No Sample Provided
Breath Volume		0.000	21:23	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 2	0.162	21:27	
Breath Volume		1.867	21:27	
Air Blank		0.000	21:27	
Air Blank		0.000	21:29	
Subject Sample	# 3	0.164	21:30	
Breath Volume		2.425	21:30	
Air Blank		0.000	21:30	
Control Test		0.078	21:31	
Air Blank		0.000	21:31	
Diagnostics Check		OK	21:31	

Data Download Date/Time: 07/23/2026 08:43

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HADDOCK,CLAY,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HICKS,J
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 06/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23725080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: LEFLET

DATE/TIME OF TEST: 06/28/2026 07:32:14

Diagnostics Check		OK	07:26
Air Blank		0.000	07:26
Control Test		0.079	07:26
Air Blank		0.000	07:27
Subject Sample	# 1	0.171	07:27
Breath Volume		2.164	07:27
Air Blank		0.000	07:28
Air Blank		0.000	07:30
Subject Sample	# 2	0.167	07:30
Breath Volume		2.308	07:30
Air Blank		0.000	07:31
Control Test		0.079	07:31
Air Blank		0.000	07:32
Diagnostics Check		OK	07:32

Data Download Date/Time: 07/22/2026 23:59

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 06/18/2026 00:26:14

Diagnostics Check		OK	00:14	
Air Blank		0.000	00:15	
Control Test		0.079	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:19	No Sample Provided
Breath Volume		0.000	00:19	
Air Blank		0.000	00:19	
Air Blank		0.000	00:21	
Subject Sample	# 2		00:24	No Sample Provided
Breath Volume		0.000	00:24	
Air Blank		0.000	00:25	
Control Test		0.079	00:25	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 07/15/2026 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORSBERG,BRIAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FORSBERG,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ESTRADA MILANES

DATE/TIME OF TEST: 06/19/2026 01:39:13

Diagnostics Check	OK	01:38	
Air Blank		01:39	RFI Detect
Air Blank		01:39	RFI Detect

Data Download Date/Time: 07/15/2026 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORSBERG,BRIAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FORSBERG,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ESTRADA MILANES

DATE/TIME OF TEST: 06/19/2026 01:42:43

Diagnostics Check	OK	01:42	
Air Blank		01:42	RFI Detect
Air Blank		01:42	RFI Detect

Data Download Date/Time: 07/15/2026 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORSBERG,BRIAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FORSBERG,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ESTRADA MILANES

DATE/TIME OF TEST: 06/19/2026 01:46:47

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:44	
Control Test		0.079	01:45	
Air Blank		0.000	01:45	
Subject Sample	# 1		01:45	RFI Detect
Breath Volume		0.000	01:45	
Air Blank			01:46	RFI Detect
Air Blank		0.000	01:46	

Data Download Date/Time: 07/15/2026 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORSBERG,BRIAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FORSBERG,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ESTRADA MILANES

DATE/TIME OF TEST: 06/19/2026 01:54:10

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:52	
Control Test		0.080	01:52	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:53	RFI Detect
Breath Volume		0.000	01:53	
Air Blank			01:53	RFI Detect
Air Blank		0.000	01:54	

Data Download Date/Time: 07/15/2026 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEST,TEST,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 06/19/2026 02:12:02

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:01	
Control Test		0.079	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:05	No Sample Provided
Breath Volume		0.000	02:05	
Air Blank		0.000	02:05	
Air Blank		0.000	02:07	
Subject Sample	# 2		02:10	No Sample Provided
Breath Volume		0.000	02:10	
Air Blank		0.000	02:11	
Control Test		0.080	02:11	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 07/15/2026 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ-ALGARIN,NICOLE,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NARVAEZ,VIRNA
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: ADORNO PABON

DATE/TIME OF TEST: 06/19/2026 03:35:25

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.077	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.055	03:30
Breath Volume		1.304	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.056	03:33
Breath Volume		2.292	03:33
Air Blank		0.000	03:34
Control Test		0.076	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 07/15/2026 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: DIAZ-PORTOCARRERO

DATE/TIME OF TEST: 06/19/2026 03:49:01

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:41	
Control Test		0.079	03:42	
Air Blank		0.000	03:42	
Subject Sample	# 1	0.213	03:44	
Breath Volume		1.578	03:44	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.209	03:47	
Breath Volume		1.117	03:47	
Air Blank		0.000	03:48	
Control Test		0.074	03:48	Control Outside Tolerance
Air Blank		0.000	03:49	

Data Download Date/Time: 07/15/2026 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,JORDIN,K
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FRYSIEK,J
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: SKAGGS

DATE/TIME OF TEST: 06/03/2026 22:31:54

Diagnostics Check		OK	22:23
Air Blank		0.000	22:24
Control Test		0.079	22:24
Air Blank		0.000	22:24
Subject Sample	# 1	0.158	22:26
Breath Volume		1.683	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.156	22:30
Breath Volume		2.109	22:30
Air Blank		0.000	22:31
Control Test		0.078	22:31
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 06/24/2026 09:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARABALLO,DOMINICK,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 06/05/2026 19:06:23

Diagnostics Check		OK	18:59
Air Blank		0.000	18:59
Control Test		0.080	19:00
Air Blank		0.000	19:00
Subject Sample	# 1	0.000	19:02
Breath Volume		1.589	19:02
Air Blank		0.000	19:02
Air Blank		0.000	19:04
Subject Sample	# 2	0.000	19:05
Breath Volume		1.222	19:05
Air Blank		0.000	19:05
Control Test		0.080	19:05
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 06/06/2026 04:44:43

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.079	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.111	04:40
Breath Volume		3.105	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.111	04:43
Breath Volume		2.617	04:43
Air Blank		0.000	04:43
Control Test		0.079	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOGELE,KEVIN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: VOGELE,KEVIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 06/07/2026 03:03:57

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.089	02:58
Breath Volume		1.269	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.095	03:02
Breath Volume		2.246	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 06/07/2026 06:03:05

Diagnostics Check		OK	05:48	
Air Blank		0.000	05:48	
Control Test		0.080	05:49	
Air Blank		0.000	05:49	
Subject Sample	# 1	0.222	05:52	
Breath Volume		1.312	05:52	
Air Blank		0.000	05:53	
Air Blank		0.000	05:55	
Subject Sample	# 2	0.187	05:58	Volume Not Met
Breath Volume		0.925	05:58	
Air Blank		0.000	05:59	
Air Blank		0.000	06:00	
Subject Sample	# 3	0.218	06:01	
Breath Volume		1.332	06:01	
Air Blank		0.000	06:02	
Control Test		0.076	06:02	
Air Blank		0.000	06:02	
Diagnostics Check		OK	06:03	

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNY,E,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NGUYEN,M
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: PORRELLO

DATE/TIME OF TEST: 06/07/2026 07:46:57

Diagnostics Check		OK	07:39
Air Blank		0.000	07:40
Control Test		0.078	07:40
Air Blank		0.000	07:40
Subject Sample	# 1	0.222	07:41
Breath Volume		2.414	07:41
Air Blank		0.000	07:42
Air Blank		0.000	07:44
Subject Sample	# 2	0.223	07:45
Breath Volume		1.687	07:45
Air Blank		0.000	07:46
Control Test		0.078	07:46
Air Blank		0.000	07:46
Diagnostics Check		OK	07:46

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNY,E,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CASTILLO,B
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 09:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 06/07/2026 10:09:17

Diagnostics Check		OK	10:01
Air Blank		0.000	10:01
Control Test		0.079	10:01
Air Blank		0.000	10:02
Subject Sample	# 1	0.078	10:04
Breath Volume		1.812	10:04
Air Blank		0.000	10:04
Air Blank		0.000	10:06
Subject Sample	# 2	0.087	10:07
Breath Volume		3.234	10:07
Air Blank		0.000	10:08
Control Test		0.078	10:08
Air Blank		0.000	10:09
Diagnostics Check		OK	10:09

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BARTLETT,GERALD
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/07/2026 23:56:13

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:51
Breath Volume		2.996	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:54
Breath Volume		2.667	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DION,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:17

SUBJECT'S LAST NAME: VARELA

DATE/TIME OF TEST: 06/08/2026 22:26:54

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.170	22:21
Breath Volume		1.566	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.180	22:25
Breath Volume		1.164	22:25
Air Blank		0.000	22:25
Control Test		0.079	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ-GUILLEN,MARCOS,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LOPEZ-GUILLEN,MARCOS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: CURRADO

DATE/TIME OF TEST: 06/12/2026 20:46:24

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.138	20:41
Breath Volume		1.406	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.134	20:44
Breath Volume		2.800	20:44
Air Blank		0.000	20:45
Control Test		0.079	20:45
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: SCHINSKI

DATE/TIME OF TEST: 06/17/2026 06:22:33

Diagnostics Check		OK	06:16
Air Blank		0.000	06:17
Control Test		0.080	06:17
Air Blank		0.000	06:17
Subject Sample	# 1	0.000	06:18
Breath Volume		2.070	06:18
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.000	06:21
Breath Volume		2.292	06:21
Air Blank		0.000	06:21
Control Test		0.080	06:21
Air Blank		0.000	06:22
Diagnostics Check		OK	06:22

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARQUIS,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: BIAGI,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MENDELSOHN

DATE/TIME OF TEST: 06/19/2026 03:45:18

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.081	03:40
Breath Volume		2.578	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.082	03:43
Breath Volume		2.402	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 06/24/2026 09:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRISCO,RICKY,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: ROBERTS,JOSEPH
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:13

SUBJECT'S LAST NAME: GARRISON

DATE/TIME OF TEST: 06/19/2026 10:42:49

Diagnostics Check		OK	10:35
Air Blank		0.000	10:36
Control Test		0.079	10:36
Air Blank		0.000	10:37
Subject Sample	# 1	0.230	10:38
Breath Volume		1.824	10:38
Air Blank		0.000	10:38
Air Blank		0.000	10:40
Subject Sample	# 2	0.222	10:41
Breath Volume		1.839	10:41
Air Blank		0.000	10:41
Control Test		0.077	10:42
Air Blank		0.000	10:42
Diagnostics Check		OK	10:42

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERARDI,JOHN,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: MACHNICKI,AILECE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/20/2026 23:58:37

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.079	23:52
Air Blank		0.000	23:53
Subject Sample	# 1	0.000	23:54
Breath Volume		3.421	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.000	23:57
Breath Volume		3.558	23:57
Air Blank		0.000	23:57
Control Test		0.080	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 06/24/2026 09:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KISTLER,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: KISTLER,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: BEVERLY

DATE/TIME OF TEST: 06/21/2026 22:43:49

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.080	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.000	22:39
Breath Volume		2.210	22:39
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.000	22:42
Breath Volume		1.750	22:42
Air Blank		0.000	22:42
Control Test		0.079	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 06/24/2026 09:46

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DION,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CHRISTENSEN,CALEB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SARDIDO

DATE/TIME OF TEST: 06/27/2026 04:52:50

Diagnostics Check		OK	04:46
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.067	04:48
Breath Volume		2.242	04:48
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.064	04:51
Breath Volume		1.921	04:51
Air Blank		0.000	04:51
Control Test		0.078	04:52
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 07/17/2026 08:34

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOGELE,KEVIN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: VOGELE,KEVIN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DENNY

DATE/TIME OF TEST: 06/28/2026 03:19:24

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.089	03:14
Breath Volume		1.343	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.083	03:17
Breath Volume		1.980	03:17
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 07/17/2026 08:35

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOGELE,KEVIN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GRISMER,GARRETT
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 06/28/2026 03:49:05

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.079	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.293	03:44
Breath Volume		1.351	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.303	03:47
Breath Volume		1.113	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:49

Data Download Date/Time: 07/17/2026 08:35

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYMONS,DEPUTY,E
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NAPOLEON,DEPUTY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 06/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31425080A1
DRY GAS EXPIRY DATE: 01/05/2028
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 06/28/2026 18:17:49

Diagnostics Check		OK	18:06
Air Blank		0.000	18:07
Control Test		0.079	18:07
Air Blank		0.000	18:08
Subject Sample	# 1	0.140	18:11
Breath Volume		1.855	18:11
Air Blank		0.000	18:11
Air Blank		0.000	18:13
Subject Sample	# 2	0.138	18:16
Breath Volume		1.460	18:16
Air Blank		0.000	18:16
Control Test		0.078	18:17
Air Blank		0.000	18:17
Diagnostics Check		OK	18:17

Data Download Date/Time: 07/17/2026 08:35

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: SINGER-TURNER

DATE/TIME OF TEST: 06/06/2026 01:04:24

Diagnostics Check		OK	00:56
Air Blank		0.000	00:57
Control Test		0.081	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.163	00:59
Breath Volume		1.328	00:59
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.164	01:02
Breath Volume		1.468	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 07/09/2026 16:03

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: GOMEZ SANTIZ

DATE/TIME OF TEST: 06/06/2026 04:03:08

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	03:59	
Subject Sample	# 1	0.173	04:01	
Breath Volume		1.484	04:01	
Air Blank			04:02	Purge Fail
Air Blank		0.000	04:03	

Data Download Date/Time: 07/09/2026 16:03

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: GOMEZ SANTIZ

DATE/TIME OF TEST: 06/06/2026 04:59:04

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:54
Subject Sample	# 1	0.168	04:54
Breath Volume		1.984	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.177	04:57
Breath Volume		1.851	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:58
Diagnostics Check		OK	04:59

Data Download Date/Time: 07/09/2026 16:03

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROOME,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 06/16/2026 03:55:29

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.079	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1		03:54	Subject Test Refused
Breath Volume		0.000	03:54	
Air Blank		0.000	03:54	
Control Test		0.079	03:54	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 07/09/2026 16:03

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SNEERINGER

DATE/TIME OF TEST: 06/21/2026 03:42:37

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.076	03:35	
Air Blank		0.000	03:35	
Subject Sample	# 1	0.178	03:38	
Breath Volume		1.121	03:38	
Air Blank		0.000	03:39	
Air Blank		0.000	03:40	
Subject Sample	# 2	0.184	03:41	
Breath Volume		1.136	03:41	
Air Blank			03:42	Purge Fail
Air Blank		0.000	03:42	

Data Download Date/Time: 07/09/2026 16:03

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROOME,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: TYGART

DATE/TIME OF TEST: 06/22/2026 03:38:05

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.078	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1	0.182	03:27	Volume Not Met
Breath Volume		0.546	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2	0.185	03:31	
Breath Volume		1.148	03:31	
Air Blank		0.000	03:32	
Air Blank		0.000	03:34	
Subject Sample	# 3		03:36	Subject Test Refused
Breath Volume		0.480	03:36	
Air Blank		0.000	03:37	
Control Test		0.079	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:38	

Data Download Date/Time: 07/09/2026 16:04

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROOME,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: TYGART

DATE/TIME OF TEST: 06/22/2026 03:48:50

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.188	03:43
Breath Volume		1.140	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:45
Subject Sample	# 2	0.185	03:47
Breath Volume		1.292	03:47
Air Blank		0.000	03:47
Control Test		0.075	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 07/09/2026 16:04

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FLOOD,KRISTIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402843437
DRY GAS EXPIRY DATE: 09/08/2026
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: STANFORD

DATE/TIME OF TEST: 06/24/2026 19:23:23

Diagnostics Check		OK	19:15
Air Blank		0.000	19:15
Control Test		0.079	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.147	19:17
Breath Volume		1.289	19:17
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.139	19:21
Breath Volume		1.109	19:21
Air Blank		0.000	19:22
Control Test		0.080	19:22
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 07/09/2026 16:04

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS,TYLER,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 06/05/2026 23:27:15

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.082	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.189	23:22
Breath Volume		1.468	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.198	23:25
Breath Volume		1.292	23:25
Air Blank		0.000	23:26
Control Test		0.081	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 07/04/2026 07:37

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES-MASSON,B,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WHITFIELD,C
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 06/12/2026 20:13:40

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.081	20:07
Air Blank		0.000	20:08
Subject Sample	# 1	0.044	20:09
Breath Volume		3.058	20:09
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.043	20:12
Breath Volume		1.820	20:12
Air Blank		0.000	20:12
Control Test		0.082	20:13
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 07/04/2026 07:37

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,TYLER,B
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/12/2026 23:09:36

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.081	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.124	23:05
Breath Volume		1.640	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.120	23:08
Breath Volume		1.554	23:08
Air Blank		0.000	23:08
Control Test		0.080	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 07/04/2026 07:37

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,GEORGE,O
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STONE,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: BRAVO SOTO

DATE/TIME OF TEST: 06/20/2026 19:27:17

Diagnostics Check		OK	19:17
Air Blank		0.000	19:18
Control Test		0.081	19:18
Air Blank		0.000	19:18
Subject Sample	# 1	0.213	19:19
Breath Volume		2.292	19:19
Air Blank		0.000	19:20
Air Blank		0.000	19:21
Subject Sample	# 2	0.192	19:22
Breath Volume		1.878	19:22
Air Blank		0.000	19:22
Air Blank		0.000	19:24
Subject Sample	# 3	0.201	19:25
Breath Volume		1.421	19:25
Air Blank		0.000	19:26
Control Test		0.080	19:26
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 07/04/2026 07:37

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,LEANDRA,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: DAVIS,LEANDRA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 06/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 06/15/2026 02:02:51

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.080	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.229	01:57
Breath Volume		1.910	01:57
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.224	02:01
Breath Volume		1.519	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 07/10/2026 07:52

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOWELL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MCDOWELL,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403308953
DRY GAS EXPIRY DATE: 04/01/2028
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 06/23/2026 01:48:29

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.077	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.328	01:43
Breath Volume		2.132	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.321	01:46
Breath Volume		2.425	01:46
Air Blank		0.000	01:47
Control Test		0.078	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 07/10/2026 07:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: VALLE

DATE/TIME OF TEST: 06/16/2026 22:34:57

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.078	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.211	22:30
Breath Volume		2.835	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.219	22:33
Breath Volume		2.992	22:33
Air Blank		0.000	22:34
Control Test		0.078	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 06/17/2026 15:33

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: DE LEON ORTIZ

DATE/TIME OF TEST: 06/17/2026 00:19:58

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.077	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.072	00:15
Breath Volume		2.789	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.068	00:18
Breath Volume		2.105	00:18
Air Blank		0.000	00:19
Control Test		0.078	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 06/17/2026 15:33

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORSBERG,BRIAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FORSBERG,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ESTRADA MILANES

DATE/TIME OF TEST: 06/19/2026 02:01:32

Diagnostics Check	OK	02:00	
Air Blank		02:01	RFI Detect
Air Blank	0.000	02:01	

Data Download Date/Time: 07/15/2026 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORSBERG,BRIAN,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FORSBERG,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: ESTRADA MILANES

DATE/TIME OF TEST: 06/19/2026 02:16:12

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:04	
Control Test		0.078	02:04	
Air Blank		0.000	02:05	
Subject Sample	# 1	0.156	02:08	Volume Not Met
Breath Volume		0.390	02:08	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.151	02:11	
Breath Volume		2.179	02:11	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 3	0.151	02:14	
Breath Volume		1.812	02:14	
Air Blank		0.000	02:15	
Control Test		0.078	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 07/15/2026 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCALDE,ALEXANDER,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HALL,DEVIN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: JONES III

DATE/TIME OF TEST: 06/19/2026 11:12:51

Diagnostics Check		OK	11:01	
Air Blank		0.000	11:01	
Control Test		0.078	11:02	
Air Blank		0.000	11:02	
Subject Sample	# 1		11:05	No Sample Provided
Breath Volume		0.000	11:05	
Air Blank		0.000	11:06	
Air Blank		0.000	11:08	
Subject Sample	# 2		11:11	No Sample Provided
Breath Volume		0.000	11:11	
Air Blank		0.000	11:11	
Control Test		0.078	11:12	
Air Blank		0.000	11:12	
Diagnostics Check		OK	11:12	

Data Download Date/Time: 07/15/2026 12:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: HERNANDEZ OQUENDRO

DATE/TIME OF TEST: 06/20/2026 00:53:54

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.178	00:49
Breath Volume		1.917	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.172	00:52
Breath Volume		1.289	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 07/15/2026 12:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OWENS,WILLIAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: SEPTER

DATE/TIME OF TEST: 06/20/2026 01:03:19

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.078	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.000	00:58
Breath Volume		2.484	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.000	01:01
Breath Volume		2.039	01:01
Air Blank		0.000	01:02
Control Test		0.078	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 07/15/2026 12:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIREBAUGH,JESSICA,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FIREBAUGH,JESSICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: SANTIAGO-BURRELL

DATE/TIME OF TEST: 06/20/2026 02:17:01

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.078	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:10	No Sample Provided
Breath Volume		0.000	02:10	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2		02:15	No Sample Provided
Breath Volume		0.000	02:15	
Air Blank		0.000	02:16	
Control Test		0.078	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 07/15/2026 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: VELAZQUEZ NEGRETE

DATE/TIME OF TEST: 06/20/2026 04:06:20

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.078	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:57	No Sample Provided
Breath Volume		0.000	03:57	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 2	0.124	04:01	
Breath Volume		1.660	04:01	
Air Blank		0.000	04:02	
Air Blank		0.000	04:04	
Subject Sample	# 3	0.115	04:04	
Breath Volume		1.605	04:04	
Air Blank		0.000	04:05	
Control Test		0.077	04:05	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 07/15/2026 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: OUTLAW,ADAM
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 06/20/2026 20:32:29

Diagnostics Check		OK	20:25
Air Blank		0.000	20:26
Control Test		0.079	20:26
Air Blank		0.000	20:26
Subject Sample	# 1	0.095	20:27
Breath Volume		1.238	20:27
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.094	20:31
Breath Volume		1.742	20:31
Air Blank		0.000	20:31
Control Test		0.079	20:31
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 07/15/2026 12:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAMARA,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: NESTER,LOGAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: HYDE

DATE/TIME OF TEST: 06/20/2026 22:29:47

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.078	22:19	
Air Blank		0.000	22:19	
Subject Sample	# 1		22:22	No Sample Provided
Breath Volume		0.000	22:22	
Air Blank		0.000	22:23	
Air Blank		0.000	22:25	
Subject Sample	# 2		22:28	No Sample Provided
Breath Volume		0.000	22:28	
Air Blank		0.000	22:28	
Control Test		0.079	22:29	
Air Blank		0.000	22:29	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 07/15/2026 12:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SANCHEZ BELLAS

DATE/TIME OF TEST: 06/21/2026 02:48:04

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.078	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.219	02:43
Breath Volume		2.078	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.223	02:46
Breath Volume		2.035	02:46
Air Blank		0.000	02:47
Control Test		0.078	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:48

Data Download Date/Time: 07/15/2026 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIREBAUGH,JESSICA,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FIREBAUGH,JESSICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ROSAS-ROSAS

DATE/TIME OF TEST: 06/21/2026 05:14:09

Diagnostics Check		OK	05:02	
Air Blank		0.000	05:03	
Control Test		0.078	05:03	
Air Blank		0.000	05:04	
Subject Sample	# 1		05:07	No Sample Provided
Breath Volume		0.000	05:07	
Air Blank		0.000	05:07	
Air Blank		0.000	05:09	
Subject Sample	# 2		05:12	No Sample Provided
Breath Volume		0.000	05:12	
Air Blank		0.000	05:13	
Control Test		0.079	05:13	
Air Blank		0.000	05:14	
Diagnostics Check		OK	05:14	

Data Download Date/Time: 07/15/2026 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIREBAUGH,JESSICA,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: BLASER,S
ARRESTING OFFICER AGENCY: Lake Wales PD
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: RAMNARINE

DATE/TIME OF TEST: 06/21/2026 06:51:11

Diagnostics Check		OK	06:38	
Air Blank		0.000	06:39	
Control Test		0.078	06:39	
Air Blank		0.000	06:40	
Subject Sample	# 1		06:43	No Sample Provided
Breath Volume		0.000	06:43	
Air Blank		0.000	06:43	
Air Blank		0.000	06:45	
Subject Sample	# 2	0.190	06:46	
Breath Volume		2.355	06:46	
Air Blank		0.000	06:47	
Air Blank		0.000	06:49	
Subject Sample	# 3	0.183	06:49	
Breath Volume		1.265	06:49	
Air Blank		0.000	06:50	
Control Test		0.078	06:50	
Air Blank		0.000	06:51	
Diagnostics Check		OK	06:51	

Data Download Date/Time: 07/15/2026 12:44

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL MORAL,NATANAEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: DEL MORAL,NATANAEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:58

SUBJECT'S LAST NAME: ARIUS

DATE/TIME OF TEST: 06/21/2026 16:44:13

Diagnostics Check		OK	16:32	
Air Blank		0.000	16:33	
Control Test		0.078	16:33	
Air Blank		0.000	16:34	
Subject Sample	# 1		16:37	No Sample Provided
Breath Volume		0.000	16:37	
Air Blank		0.000	16:37	
Air Blank		0.000	16:39	
Subject Sample	# 2		16:42	No Sample Provided
Breath Volume		0.000	16:42	
Air Blank		0.000	16:43	
Control Test		0.078	16:43	
Air Blank		0.000	16:44	
Diagnostics Check		OK	16:44	

Data Download Date/Time: 07/15/2026 12:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIREBAUGH,JESSICA,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FIREBAUGH,JESSICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 06/22/2026 00:31:38

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:20	
Control Test		0.079	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:24	No Sample Provided
Breath Volume		0.000	00:24	
Air Blank		0.000	00:24	
Air Blank		0.000	00:26	
Subject Sample	# 2	0.101	00:30	Volume Not Met
Breath Volume		0.238	00:30	
Air Blank		0.000	00:30	
Control Test		0.080	00:31	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 07/15/2026 12:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ALEJANDRO,ADRIAN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 06/22/2026 18:24:38

Diagnostics Check		OK	18:18
Air Blank		0.000	18:18
Control Test		0.078	18:19
Air Blank		0.000	18:19
Subject Sample	# 1	0.153	18:20
Breath Volume		2.937	18:20
Air Blank		0.000	18:20
Air Blank		0.000	18:22
Subject Sample	# 2	0.149	18:23
Breath Volume		2.492	18:23
Air Blank		0.000	18:23
Control Test		0.077	18:24
Air Blank		0.000	18:24
Diagnostics Check		OK	18:24

Data Download Date/Time: 07/15/2026 12:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VASQUEZ MORALES

DATE/TIME OF TEST: 06/25/2026 01:17:38

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:12	
Control Test		0.078	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1	0.122	01:14	
Breath Volume		3.191	01:14	
Air Blank		0.000	01:14	
Air Blank		0.000	01:16	
Subject Sample	# 2	0.114	01:16	
Breath Volume		1.757	01:16	
Air Blank			01:17	RFI Detect
Air Blank			01:17	Ambient Fail

Data Download Date/Time: 07/15/2026 12:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VASQUEZ MORALES

DATE/TIME OF TEST: 06/25/2026 01:26:09

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.078	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.118	01:21
Breath Volume		2.250	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 2	0.116	01:24
Breath Volume		2.324	01:24
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 07/15/2026 12:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIREBAUGH,JESSICA,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FIREBAUGH,JESSICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: MCMURRAY

DATE/TIME OF TEST: 06/25/2026 03:36:24

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.078	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.252	03:31
Breath Volume		1.925	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:33
Subject Sample	# 2	0.244	03:34
Breath Volume		1.160	03:34
Air Blank		0.000	03:35
Control Test		0.077	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 07/15/2026 12:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIROUX,ANDREW,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: GIROUX,ANDREW
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: GRAZIER

DATE/TIME OF TEST: 06/25/2026 05:19:25

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.079	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.231	05:15
Breath Volume		2.632	05:15
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.220	05:17
Breath Volume		2.027	05:17
Air Blank		0.000	05:18
Control Test		0.078	05:18
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 07/15/2026 12:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIREBAUGH,JESSICA,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FIREBAUGH,JESSICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SCOTNEY

DATE/TIME OF TEST: 06/25/2026 23:28:28

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.000	23:24
Breath Volume		2.054	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.000	23:27
Breath Volume		2.765	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 07/15/2026 12:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORESTEIRE,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FORESTEIRE,JOSEPH
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 06/28/2026 02:34:18

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:23	
Control Test		0.078	02:23	
Air Blank		0.000	02:24	
Subject Sample	# 1		02:27	No Sample Provided
Breath Volume		0.000	02:27	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 2		02:32	No Sample Provided
Breath Volume		0.000	02:32	
Air Blank		0.000	02:33	
Control Test		0.079	02:33	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 07/15/2026 12:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALSTON,SONDRA,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RAO,RAVI
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: REYNOSO

DATE/TIME OF TEST: 06/28/2026 07:01:11

Diagnostics Check		OK	06:58	
Air Blank		0.000	06:58	
Control Test		0.078	06:59	
Air Blank		0.000	06:59	
Subject Sample	# 1		06:59	Subject Test Refused
Breath Volume		0.000	06:59	
Air Blank		0.000	07:00	
Control Test		0.079	07:00	
Air Blank		0.000	07:01	
Diagnostics Check		OK	07:01	

Data Download Date/Time: 07/15/2026 12:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: ROWLAND,BOBBY
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/28/2026 20:53:58

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.079	20:49
Air Blank		0.000	20:49
Subject Sample	# 1	0.000	20:49
Breath Volume		3.531	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.000	20:52
Breath Volume		3.308	20:52
Air Blank		0.000	20:53
Control Test		0.079	20:53
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 07/15/2026 12:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIREBAUGH,JESSICA,E
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: FIREBAUGH,JESSICA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 06/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: BROOME

DATE/TIME OF TEST: 06/30/2026 05:00:30

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.078	04:55
Air Blank		0.000	04:55
Subject Sample	# 1	0.170	04:56
Breath Volume		4.117	04:56
Air Blank		0.000	04:56
Air Blank		0.000	04:58
Subject Sample	# 2	0.171	04:59
Breath Volume		3.847	04:59
Air Blank		0.000	04:59
Control Test		0.078	04:59
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 07/15/2026 12:45

Printed: 29 September 2026

COBRA ©1997-2014 CMI, Inc.